

Results of Proficiency Test
Jet Fuel A1
March 2010

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: Ing. R.J.Starink
Correctors: Dr. R.G. Visser & Ing. L. Sweere
Report: iis10J01X

May 2010

-- empty page --

CONTENTS

1	INTRODUCTION	4
2	SET UP.....	4
2.1	ACCREDITATION.....	4
2.2	PROTOCOL	4
2.3	CONFIDENTIALITY STATEMENT	4
2.4	SAMPLES.....	5
2.5	STABILITY OF THE SAMPLES	6
2.6	ANALYSES	6
3	RESULTS.....	7
3.1	STATISTICS.....	7
3.2	GRAPHICS.....	8
3.3	Z-SCORES.....	8
4	EVALUATION.....	8
4.1	EVALUATION PER TEST	9
4.2	INTEGRAL PERFORMANCE EVALUATION; GENERAL	12
4.3	INTEGRAL PERFORMANCE EVALUATION PER LABORATORY.....	13
Appendices:		
1.	Data, statistical results and graphic results	14
2.	Z-scores for distillation	50
3.	List of participants	51
4.	Abbreviations and literature	52

1 INTRODUCTION

Since 1995, the Institute for Interlaboratory Studies organises every year proficiency tests for Jet Fuel A1. The international Interlaboratory study on Jet Fuel of March 2010 was extended with a PT for the determination for Particle Size Distribution. In the regular PT, 65 laboratories in 31 different countries have participated; in the PT for the Particle Size Distribution 25 laboratories in 18 different countries. See appendix 4 for a list of participants in alphabetical country order. In this report, the results of the proficiency test are presented and discussed.

2 SET UP

The Institute for Interlaboratory Studies (i.i.s.) in Spijkenisse, The Netherlands, was the organiser of this proficiency test. In the regular Jet Fuel round robin, it was decided, depending on the registration, to send the two identical samples (2*1 liter of sample #1017) for the analyses according to the "Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS)", sometimes referred to as the "Joint Fuelling System Check List For Jet A-1" and/or for the Particle Size round robin to send one sample of 0.5 L (#1018).

The participants were requested to report rounded and unrounded results. The unrounded results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO guide 43 and ILAC-G13:2007, (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This ensures 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organization, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2), which can be downloaded from www.iisnl.com. The participants were asked to report the analytical results using the indicated units on the report form.

2.3 CONFIDENTIALITY STATEMENT

All data present in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

2.4.1 REGULAR JET FUEL

The necessary bulk material was obtained from a local refinery. The approx. 200 litre bulk sample was homogenised and divided over 176 amber glass bottles of one litre with inner and outer caps and labelled #1017. The homogeneity of the subsamples #1017 was checked by the determination of Density in accordance with ASTM D 4052:09 on 8 stratified random selected samples.

Sample	Density @ 15°C in kg/m ³
#1017-1	795.01
#1017-2	795.01
#1017-3	795.00
#1017-4	795.00
#1017-5	795.00
#1017-6	795.00
#1017-7	795.00
#1017-8	795.00

table 1: homogeneity test of sub samples #1017

From the above test results, the repeatability was calculated and compared with 0.3 times the reproducibility of the reference method in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	Density @ 15°C in kg/m ³
r (observed)	0.01
reference method	D 4052:09
0.3 x R (ref. method)	0.15

Table 2: evaluation of repeatability of subsamples #0955

The calculated repeatability is less than 0.3 times the reproducibility of the reference method. Therefore, homogeneity of all subsamples was assumed.

2.4.2 JET FUEL PARTICLE SIZE DETERMINATION

The second bulk material was obtained from a participating laboratory. The approx. 200 litre bulk sample was homogenised and divided over 48 amber glass bottles of 500 mL with inner and outer caps and labelled #1018. The homogeneity of the subsamples #1018 was checked by the determination of Particle Size Distribution in accordance with IP564:10 on 7 stratified random selected samples.

Sample	> 4 µm	> 6 µm	> 14 µm	> 21 µm	> 25 µm	> 30 µm
#0957-1	17064	2873	93	18	9	3
#0957-2	17123	2846	83	15	6	2
#0957-3	16813	2681	64	11	5	2
#0957-4	16804	2747	84	19	9	4
#0957-5	16862	2745	79	17	8	3
#0957-6	17354	2959	96	19	10	4
#0957-7	17245	2945	100	21	10	4

Table 3: homogeneity test of sub samples #0957

From the above test results, the repeatability was calculated and compared with 0.3 times the reproducibility of the reference method in agreement with the procedure of ISO 13528, Annex B2 in the next table:

Sample	> 4 µm	> 6 µm	> 14 µm	> 21 µm	> 25 µm	> 30 µm
r (observed)	612	299	34	9.2	5.5	2.5
reference method	IP564:10	IP564:10	IP564:10	IP564:10	IP564:10	IP564:10
0.3 x R (ref. method)	968	270	15	6.5	3.3	1.6

Table 4: evaluation of repeatabilities of subsamples #0957

Regretfully not all repeatabilities of the results for Particle Size are in agreement with 0.3 times the corresponding reproducibility of the target method. However, the repeatabilities of the results do meet the repeatability limits of IP564:10 ($r_{IP564} = 2305, 557, 38, 17.4, 8.5$ and 4.0). Therefore, homogeneity of the subsamples #1018 was assumed.

Depending on the registration of each participant, the following samples were dispatched on March 3, 2010: 2 bottles of 1 litre, labelled #1017 and/or 1*0.5 L, labelled #1018.

2.5 STABILITY OF THE SAMPLES

The stability of Jet Fuel A1, packed in the brown glass bottles was checked. The type of bottle was chosen in accordance with ASTM D4306:07. The material has been found sufficiently stable for the period of the proficiency test.

2.6 ANALYSIS

The participants were requested to determine on sample #1017: Aromatics by FIA, Aromatics by HPLC (in %M/M and %V/V), Colour Saybolt, Density @15°C, Distillation (IBP, 10%, 20%, 50%, 90% recovered and FBP), Existent Gum, Flash Point, Freezing Point, JFTOT, Mercaptans, MSEP, Naphthalenes, Smoke Point, Specific Energy (on Sulphur free basis), Total Acidity, Total Sulphur and Viscosity @ -20°C. The participants were requested to determine Particle Size only on sample #1018.

The analyses should be performed according to the "Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS)", also referred to as the "Joint Fuelling System Check List" or simply "Check List".

To get comparable results a detailed report form, on which the units were prescribed as well as some of the required standards, was sent together with each set of samples. Also a letter of instructions and a SDS were added to the package.

3 RESULTS

During four weeks after sample despatch, the results of the individual laboratories were gathered. The original data are tabulated per determination in the appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after deadline, a reminder fax was sent to those laboratories that had not yet reported. Shortly after the deadline, the available results were screened for suspect data. A result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the (raw data of the) reported results. Additional or corrected results have been used for data analysis and the original results are placed under 'Remarks' in the result tables in Appendix 1.

3.1 STATISTICS

Statistical calculations were performed as described in the report 'iis Interlaboratory Studies: Protocol for the Organization, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded results. Results reported as '<...>' or '>...>' were not used in the statistical evaluation. First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test. After removal of outliers, this check was repeated. Not all data sets proved to have a normal distribution, in which cases the conclusions of statistical evaluation should be used with due care.

In accordance with ISO 5725 (1986 and 1994) the original results per determination were submitted subsequently to Dixon and Grubbs outlier tests.

Outliers are marked by D(0.01) for the Dixon test, by G(0.01) or DG(0.01) for the Grubbs test. Stragglers are marked by D(0.05) for the Dixon test, by G(0.05) or DG(0.05) for the Grubbs test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualise the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported analysis results are plotted. The corresponding laboratory numbers are under the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected standard. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle. Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms (see appendix 4, nr.14-15).

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the spread of this interlaboratory study. The target standard deviation was calculated from the literature reproducibility by division with 2.8. The z-scores were calculated according to:

$$Z_{(\text{target})} = (\text{result} - \text{average of PT}) / \text{target standard deviation}$$

The $Z_{(\text{target})}$ scores are listed in the result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare.

Therefore, the usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$ z > 3$	unsatisfactory

4 EVALUATION

In these interlaboratory studies some major problems with couriers and/or customs clearance were encountered during dispatch of the samples to laboratories in Bulgaria, Côte D'Ivoire, Croatia, Greece, P.R. of China, Sudan and United Kingdom.

For the "regular Jet Fuel A1" PT, 13 participants reported the results after the final reporting date and 5 participants did not report any results at all.

For the PT "Particle Size", 5 participants reported the results after the final reporting date and 5 participants did not report any results at all.

In total, the participants reported 1257 numerical results. Observed were 64 outlying results, which is 5.1%. In proficiency studies, outlier percentages of 3 % - 7.5 % are quite normal.

4.1 EVALUATION PER TEST

Not all original data sets proved to have a normal distribution. Anormal distributions were found for the following determinations on sample #1017: Colour Saybolt, Density, Distillation 10% recovered, Existent Gum, Flash Point, JFTOT, Mercaptan Sulphur, Naphthalenes and, Specific Energy. Therefore, the statistical evaluation for these determinations should be used with care.

In this section, the results are discussed per test.

Since the checklist is continuously updated, the users are advised to monitor the updates. The latest version at this moment is "Issue 24_October_2008". One must keep in mind that ISO-methods are not mentioned in the "Checklist".

Aromatics by: This determination was not problematic. No statistical outliers were
FIA (D1319): observed and the calculated reproducibility is in good agreement with ASTM D1319:08.

Aromatics by: The %M/M determination was problematic for two laboratories. Two
HPLC (D6379) statistical outliers were observed. The calculated reproducibility is, after rejection of the statistical outliers, in good agreement with ASTM D6379:04. The %V/V determination may also be problematic for two laboratories. Regretfully, no precision data for the determination in %V/V is mentioned in ASTM D6379:04, therefore no further significant conclusions were drawn.

Colour Saybolt: This determination was very problematic in this round robin. Obviously, this was not an "easy" sample as reported results vary between 13 and 28. Four statistical outliers were observed and the calculated reproducibility is, after rejection of the statistical outliers, not in agreement with the requirements of ASTM D156:07a.

Density: This determination is rather problematic for several laboratories. Five statistical outliers were observed. However, the calculated reproducibility, after rejection of the statistical outliers, is in good agreement with the requirements of ASTM D4052:09.

Distillation: As only four participants reported their results determined by the manual method, it was decided to evaluate these results together with the automated group. In total four statistical outliers were observed. All calculated reproducibilities (except for 90% recovered) are, after rejection of the statistical outliers in good agreement with the requirements of ASTM D86:09e1 (automated).

Existent Gum: This determination was problematic for some laboratories. Three statistical outliers were observed and the calculated reproducibility, after rejection of the statistical outliers, is in good agreement with ASTM D381:04e1.

- Flash Point: This determination was problematic for some laboratories. Three statistical outliers were observed. The calculated reproducibility, after rejection of the statistical outliers, is in agreement with the new requirements of IP170:09. Three laboratories reported a method, that is not mentioned in the Joint Fuelling System Checklist. These three results were excluded from calculations.
When the results for IP170 and ASTM D56 were evaluated separately, both calculated reproducibilities are in agreement with the requirements of the respective method.
- Freezing Point: This determination was not problematic. Only two statistical outliers were observed and the calculated reproducibility, after rejection of the statistical outliers, is in good agreement with the requirements of ASTM D2386:06.
- JFTOT: Some reporting problems have been observed. Five laboratories reported a higher volume than the maximum allowed (450 ± 45 mL may be pumped in a valid test, see ASTM D 3241:06-B table 2). It should be noted that a pumped volume higher than 495 mL or below 405 mL means that the test is not performed correctly and results obtained are suspect.
- Mercaptan Sulphur: This determination was problematic for several laboratories at this positive level of 0.0015%M/M. Five statistical outliers and one false negative result were observed. The calculated reproducibility, after rejection of the statistical outliers, is in full agreement with the requirements of ASTM D3227:04a.
- MSEP: This determination was problematic. Two statistical outliers were observed. However, the calculated reproducibility, after rejection of the statistical outliers, is not in agreement with the requirements of ASTM D3948:08.
- Naphthalenes: This determination was problematic for several laboratories at this level of 1.0 %V/V. Five, statistical outliers were observed. However, the calculated reproducibility, after rejection of the statistical outliers, is in full agreement with the requirements of ASTM D1840:07-B.
- Smoke Point: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not at all in agreement with the requirements of ASTM D1322:08.
- Specific Energy: This determination was problematic for several laboratories. Only one statistical outlier was observed. Also, two other reported results were excluded from statistical evaluation as the test method used is not compatible with ASTM D3338. The calculated reproducibility is not at all in agreement with the requirements of ASTM D3338:09.

- Total Acidity: This determination was problematic. Four statistical outliers were observed. The calculated reproducibility, after rejection of the statistical outliers, is not at all in agreement with the requirements of ASTM D3242:08.
- Total Sulphur: This determination was problematic for several laboratories at the high level of Sulphur present in this sample (1714 mg/kg). The results seem to be bimodal divided. The results of group 2 seems to be more inline what should be expected, therefore the results of group 1 (nine results) were excluded for statistical evaluation. One statistical outlier was observed. After excluding the nine suspected results and the statistical outlier, the calculated reproducibility is in full agreement with the requirements of ASTM D5453:09.
- Viscosity: This determination was not problematic. Only two statistical outliers were observed and the calculated reproducibility, after rejection of the statistical outliers, is almost in agreement with the requirements of ASTM D445:09.
- Particle Size: This determination is problematic. The deviating results for particle size distribution of two laboratories were excluded from statistical evaluation. One participant analyzed the wrong sample. After rejection of eight statistical outliers, only the calculated reproducibilities for >21 and >30 µm are in full agreement with the requirements of IP564:10. When compared with IP565:10, none of the calculated reproducibilities agrees.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant standard and the reproducibility as found for the group of laboratories that participated. The reproducibilities derived from literature standards (in casu ASTM standards) and the calculated reproducibilities of samples #1017 are compared in the next table.

Parameter	unit	n	Average	2.8 * sd	R (lit)
Aromatics by FIA	%V/V	38	18.07	1.79	3.01
Aromatics by HPLC	%M/M	13	20.83	1.28	1.98
Aromatics by HPLC	%V/V	10	18.71	1.29	unknown
Colour Saybolt		43	18.1	2.9	2.0
Density at 15°C	Kg/m ³	54	795.00	0.26	0.50
Initial Boiling Point (all)	°C	57	151.80	6.07	8.35
10% recovered (all)	°C	57	171.38	2.69	3.77
50% recovered (all)	°C	58	194.41	2.55	2.97
90% recovered (all)	°C	58	229.22	3.61	3.44
Final Boiling Point (all)	°C	56	258.13	4.23	7.10
Existent Gum	mg/100mL	45	1.13	1.62	3.26
Flash Point	°C	51	43.32	2.87	3.20
Freezing Point	°C	52	-55.53	1.55	2.50
Mercaptan Sulphur	%M/M	32	0.00154	0.00040	0.00038
MSEP	rating	43	87.86	17.53	13.03
Naphthalenes	%V/V	32	0.996	0.082	0.094
Smoke Point	mm	49	23.98	4.59	3.00
Specific Energy	MJ/kg	34	43.2764	0.0774	0.0460
Total Acidity	mg KOH/g	37	0.00296	0.00436	0.00221
Total Sulphur	mg/kg	36	1714.9	151.1	154.5
Viscosity @ -20°C	cSt	41	3.6332	0.0739	0.0690

table 5: comparison of the observed and target reproducibilities of sample #1017

Parameter	unit	n	Average	2.8 * sd	R (lit)
Particle Size >4 µm	mL ⁻¹	16	18992	7601	3576
Particle Size >6 µm	mL ⁻¹	15	3034	2121	959
Particle Size >14 µm	mL ⁻¹	16	115	80	63
Particle Size >21 µm	mL ⁻¹	17	29.2	36.6	37.1
Particle Size >25 µm	mL ⁻¹	17	13.9	22.6	18.6
Particle Size >30 µm	mL ⁻¹	17	6.5	12.5	10.7

table 6: comparison of the observed and target reproducibilities of sample #1018

Without further statistical calculations, it can be concluded that for many tests there is a good compliance of the group of participants with the relevant standards. The tests that are problematic have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF MARCH 2010 WITH PREVIOUS PT'S

Parameter	March 2010	September 2009	March 2009	September 2008
Number of reporting labs	65	186	68	122
Number of results reported	1257	2377	1238	2137
Statistical outliers	64	73	43	65
Percentage outliers	5.1%	3.1%	3.5%	3.0%

table 7: Comparison with previous proficiency tests

In proficiency tests outlier percentages of 3 % - 7.5 % are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective standards. The conclusions are given the following table:

Parameter	March 2010	September 2009	March 2009	September 2008
Aromatics by FIA	++	++	++	++
Aromatics by HPLC	++	++	++	--
Colour Saybolt	--	--	--	+
Density at 15°C	++	++	++	++
Distillation automated	++	++	++	++
Distillation manual	n.e.	++	--	--
Existent Gum	++	++	++	++
Flash Point	++	--	--	--
Freezing Point	++	+/-	++	+
Mercaptan Sulphur	+/-	++	++	++
MSEP	--	+	--	+/-
Naphthalenes	++	--	+/-	+/-
Smoke Point	--	--	+/-	+/-
Specific Energy	--	++	-	--
Total Acidity	--	--	+	+/-
Total Sulphur	+	--	+	--
Viscosity @ -20°C	-	--	--	--
Particle Size Distribution	--	+/-	n.e.	n.e.



table 8: comparison determinations against the standard requirements

The performance of the determinations against the requirements of the respective standards is listed in the above table. The following performance categories were used:

- ++: group performed much better than the standard
- + : group performed better than the standard
- +/-: group performance equals the standard
- : group performed worse than the standard
- : group performed much worse than the standard
- n.d.: not determined
- n.e.: not evaluated





APPENDIX 1**Determination of Aromatics by FIA on sample #1017; results in %V/V**

lab	method	value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463	D1319	18.1		0.03	
495		----		----	
496	D1319	18.3		0.21	
851		----		----	
1017	D1319	18.10	C	0.03	First reported 0.0
1021	D1319	18.8		0.68	
1026	IP436	18.8		0.68	
1032	D1319	17.537		-0.50	
1034	D1319	18.64		0.53	
1038	D1319	18.94		0.81	
1039		----		----	
1047	D1319	17.7		-0.34	
1049		----		----	
1059	D1319	17.8		-0.25	
1062	D1319	18.5		0.40	
1064	D1319	18.49		0.39	
1079	EN15553	17.45		-0.58	
1081	D1319	18.5		0.40	
1097	D1319	17.8		-0.25	
1105	D1319	18.29		0.20	
1108		----		----	
1109	D1319	18.26		0.18	
1126		----		----	
1131	D1319	17.9		-0.16	
1140	D1319	17.6		-0.44	
1150		----		----	
1155		----		----	
1161		----		----	
1167		----		----	
1194	D1319	17.3		-0.72	
1201		----		----	
1203	D1319	18.3		0.21	
1264		----		----	
1276	D1319	17.06		-0.94	
1277		----		----	
1279		----		----	
1284		----		----	
1288	D1319	18.8		0.68	
1318		----		----	
1402	D1319	17.7		-0.34	
1428	ISO3837	19.15	C	1.00	First reported 15.91
1528	D1319	18.8		0.68	
1531		----		----	
1538	D1319	18.6		0.49	
1543		----		----	
1610	IP156	17.3		-0.72	
1634		----		----	
1635		----		----	
1636	D1319	17.85		-0.20	
1651		----		----	
1715		----		----	
1720		----		----	
1724	D1319	18.10		0.03	
1730		----		----	
1811	D1319	17.79		-0.26	
1833	D1319	17.09		-0.91	
1842		----		----	
1854	D1319	16.3		-1.65	
1913	D1319	17.26		-0.75	
2129	D1319	18.7		0.59	
2130	D1319	18.3		0.21	
2133	D1319	18.92		0.79	
2136	D1319	17.83		-0.22	
	normality	OK			
	n	38			
	outliers	0			
	mean (n)	18.07			
	st.dev. (n)	0.638			
	R(calc.)	1.79			
	R(D1319:08)	3.01			





Determination of Aromatics by HPLC on sample #1017; results in %M/M & %V/V

lab	method	%M/M	mark	z(targ)	%V/V	mark	z(targ)	Remarks
225		----		----	----		----	
353		----		----	----		----	
463	EN12916	20.815		-0.02	----		----	
495	D6379	20.26		-0.81	18.00		----	
496		----		----	----		----	
851	D6379	20.89		0.09	18.51		----	
1017		----		----	----		----	
1021		----		----	----		----	
1026		----		----	----		----	
1032		----		----	----		----	
1034		----		----	----		----	
1038		----		----	----		----	
1039	D6379	----		----	19.1		----	
1047	EN12916	21.05		0.31	----		----	
1049	D6379	20.279		-0.78	----		----	
1059		----		----	----		----	
1062		----		----	----		----	
1064	D6379	20.78		-0.07	18.75		----	
1079		----		----	----		----	
1081		----		----	----		----	
1097		----		----	----		----	
1105	D6379	21.421		0.84	19.137		----	
1108	D6379	21.6		1.09	19.4		----	
1109	D6591	21.14		0.44	----		----	
1126		----		----	----		----	
1131		----		----	----		----	
1140		----		----	----		----	
1150		----		----	----		----	
1155		----		----	----		----	
1161	EN12916	20.61		-0.31	18.45		----	
1167	D6379	9.35	G(0.01)	-16.25	6.33	G(0.01)	----	
1194		----		----	----		----	
1201	D6379	20.9		0.10	18.6		----	
1203	D6379	20.0		-1.17	18.1		----	
1264		----		----	----		----	
1276		----		----	----		----	
1277		----		----	----		----	
1279		----		----	----		----	
1284	D6379	18.2	G(0.01)	-3.72	----		----	
1288		----		----	----		----	
1318		----		----	----		----	
1402		----		----	----		----	
1428		----		----	----		----	
1528	D6379	----		----	23.53	G(0.01)	----	
1531		----		----	----		----	
1538		----		----	----		----	
1543		----		----	----		----	
1610		----		----	----		----	
1634		----		----	----		----	
1635		----		----	----		----	
1636		----		----	----		----	
1651		----		----	----		----	
1715		----		----	----		----	
1720		----		----	----		----	
1724		----		----	----		----	
1730		----		----	----		----	
1811		----		----	----		----	
1833		----		----	----		----	
1842		----		----	----		----	
1854		----		----	----		----	
1913		----		----	----		----	
2129	D6379	21.03		0.28	19.01		----	
2130		----		----	----		----	
2133		----		----	----		----	
2136		----		----	----		----	
	normality	OK			OK			
	n	13			10			
	outliers	2			2			
	mean (n)	20.83			18.71			
	st.dev. (n)	0.457			0.459			
	R(calc.)	1.28			1.29			
	R(D6379:04)	1.98			unknown			Compare R(iis09J02) = 0.91





Determination of Colour Saybolt on sample #1017;

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	D6045	18		-0.11	
463	D156	18	C	-0.11	First reported 13
495		----		----	
496		----		----	
851	D156	17		-1.51	
1017		----		----	
1021	D6045	18		-0.11	
1026	D156	18		-0.11	
1032	D156	22	DG(0.01)	5.49	
1034	D156	19		1.29	
1038	D6045	18		-0.11	
1039	D156	18		-0.11	
1047		----		----	
1049		----		----	
1059	D156	13	G(0.05)	-7.11	
1062	D156	19		1.29	
1064	D156	18		-0.11	
1079	D156	18		-0.11	
1081	D6045	18		-0.11	
1097	INH003	18		-0.11	
1105	D156	18		-0.11	
1108	D156	18		-0.11	
1109	D156	21		4.09	
1126		----		----	
1131	D156	18		-0.11	
1140	D6045	18		-0.11	
1150		----		----	
1155		----		----	
1161	D6045	19		1.29	
1167	D156	17		-1.51	
1194		----		----	
1201	D156	28	G(0.01)	13.89	
1203	D156	18		-0.11	
1264	D156	18		-0.11	
1276	D156	19		1.29	
1277	D156	18		-0.11	
1279	D156	19		1.29	
1284	D156	17		-1.51	
1288		----		----	
1318	D156	17		-1.51	
1402	D156	17		-1.51	
1428	D156	18		-0.11	
1528	D156	16		-2.91	
1531		----		----	
1538	D156	18		-0.11	
1543		----		----	
1610	D156	17		-1.51	
1634		----		----	
1635		----		----	
1636	D156	18		-0.11	
1651	D156	20.5		3.39	
1715	D6045	20		2.69	
1720	D156	17		-1.51	
1724	D156	22	C,DG(0.01)	5.49	First reported 25
1730		----		----	
1811	D156	18		-0.11	
1833	D156	16		-2.91	
1842		----		----	
1854	D156	19		1.29	
1913	D156	20		2.69	
2129	D156	18		-0.11	
2130	D6045	18		-0.11	
2133		----		----	
2136	D156	17		-1.51	
					<u>Only ASTM D156 data:</u>
	normality	not OK			not OK
	n	43			34
	outliers	4			4
	mean (n)	18.1			18.0
	st.dev. (n)	1.03			1.10
	R(calc.)	2.9			3.1
	R(D156:07a)	2.0			2.0













Determination of Density @ 15°C on sample #1017; results in kg/m³

lab	method	value	mark	z(targ)	remarks
225		-----		-----	
353	D4052	795.0		-0.01	
463	D4052	795.0		-0.01	
495	D4052	795.0		-0.01	
496	D4052	794.92		-0.45	
851	D4052	795.1	C	0.55	First reported 795.71
1017	D4052	795.0		-0.01	
1021	D4052	795.01		0.05	
1026	D4052	795.1		0.55	
1032	D4052	795.04		0.22	
1034	D4052	794.9		-0.57	
1038	D4052	794.9		-0.57	
1039	D4052	794.9		-0.57	
1047	D4052	794.9		-0.57	
1049	D4053	794.82		-1.01	
1059	D4052	795.0		-0.01	
1062	D4052	795.0		-0.01	
1064	D4052	794.9		-0.57	
1079	D4052	795.0		-0.01	
1081	ISO12185	795.1		0.55	
1097	ISO12185	795.0		-0.01	
1105	D4052	795.1		0.55	
1108	D4052	795.07		0.39	
1109	D4052	795.19		1.06	
1126	D4052	794.99		-0.06	
1131	D4052	795.1		0.55	
1140	D4052	795.1		0.55	
1150		-----		-----	
1155	D1298	795.20		1.11	
1161	ISO12185	794.90		-0.57	
1167	D4052	795.0		-0.01	
1194	ISO12185	781.3	G(0.01)	-76.73	
1201	D4052	794.9		-0.57	
1203	ISO12185	795.0		-0.01	
1264	D4052	795.0		-0.01	
1276	D4052	794.97		-0.17	
1277	D1298	794.7	G(0.01)	-1.69	
1279	D1298	794.6	DG(0.05)	-2.25	
1284	D4052	795.08		0.44	
1288	D4052	794.9		-0.57	
1318	D4052	795.00		-0.01	
1402	D4052	794.9		-0.57	
1428	ISO12185	795.0		-0.01	
1528	D4052	794.95		-0.29	
1531		-----		-----	
1538	D4052	795.1		0.55	
1543		-----		-----	
1610	IP365	794.9		-0.57	
1634	D4052	795.0		-0.01	
1635		-----		-----	
1636	D4052	794.9		-0.57	
1651	D4052	795.02		0.11	
1715	ISO12185	794.8		-1.13	
1720	D4052	795.2		1.11	
1724	D4052	794.6	DG(0.05)	-2.25	
1730	D4052	795.00		-0.01	
1811	D4052	795.0		-0.01	
1833	D4052	795.0		-0.01	
1842		-----		-----	
1854	D4052	795.0		-0.01	
1913	D4052	794.9		-0.57	
2129	D4052	795.1		0.55	
2130	D4052	795.2		1.11	
2133	D4052	794.63	G(0.05)	-2.08	
2136	D4052	794.99		-0.06	
	normality	not OK			
	n	54			
	outliers	5			
	mean (n)	795.00			
	st.dev. (n)	0.092			
	R(calc.)	0.26			
	R(D4052:09)	0.50			





Determination of Distillation ASTM D86 (automated + manual) on sample #1017; results in °C

lab	method	IBP	mark	10%	mark	50%	mark	90%	mark	FBP	mark	Remark
225		----		----		----		----		----		
353	D86A	151.8		171.5		195.3		231.5		260.5		
463	D86A	153.1		173.3		194.4		229.8		261.1		
495	D86A	152.2		170.3		193.1		228.2		254.3		
496		----		----		----		----		----		
851	D86A	151.5		172.0		194.6		229.4		258.4		
1017	D86A	152.0		169.8		192.8		226.4		254.4		
1021	D86A	151.8		171.8		194.7		231.5		258.8		
1026	ISO3405A	147.1		173.0		195.0		230.0		258.5		
1032	D86A	152.2		171.1		194.7		230.1		258.2		
1034	D86A	154.0		171.8		194.3		229.8		259.8		
1038	D86A	154.3		170.4		193.7		229.6		258.3		
1039	D86A	151.8		172.9		194.5		229.3		256.9		
1047	D86A	153.7		171.5		195.3		231.4		258.9		
1049	D86A	149.9		172.0		193.7		228.5		256.6		
1059	D86A	151.2		171.8		195.4		229.8		259.1		
1062	D86A	153.7		172.2		195.3		229.3		258.0		
1064	D86A	155.9		171.6		194.6		230.4		258.9		
1079	D86A	153.4		172.4		194.3		228.8		256.7		
1081	D86A	150.9		171.7		194.7		230.9		256.2		
1097	ISO3405A	152.4		171.1		194.5		229.5		257.8		
1105	D86A	149.1		170.9		194.0		226.7		249.5		G(0.01)
1108	D86A	152.5		171.1		193.0		227.3		258.5		
1109	D86A	152.4		170.5		194.2		229.8		258.4		
1126	D2887A	148.2		170.6		196.8		229.0		256.6		
1131	D86A	151.7		172.7		194.7		229.7		258.0		
1140	D86A	154.0		172.5		195.7		232.1		259.8		
1150		----		----		----		----		----		
1155	ISO3405M	152.36		170.37		192.39		228.42		258.45		
1161	ISO3405A	149.65		171.25		196.00		230.05		258.65		
1167	D86A	150.0		170.1		194.3		227.4		257.1		
1194	D86	132.9	G(0.01)	168.0	G(0.05)	195.7		228.2		256.3		
1201	D86A	152.7		171.4		194.4		230.3		257.9		
1203	ISO3405A	150.1		171.9		195.8		229.9		258.8		
1264	D86A	153		172		195		229		260.5		
1276	D86A	149.4		169.4		192.3		227.1		257.2		
1277	D86A	150.2		169.0		193.3		228.1		258.3		
1279	D86M	152.0		171.5		194.5		229.0		259.5		
1284	D86A	153.3		171.6		194.0		228.3		256.3		
1288	D86A	151.4		171.1		194.2		229.3		257.2		
1318	D86A	151.8		171.4		194.0		228.7		259.3		
1402	D86A	155.5		171.7		194.0		227.6		258.2		
1428	ISO3405A	152.9		172.3		194.6		229.0		262.0		
1528	ISO3405A	151.1		172.0		194.6		229.5		258.1		
1531	D86A	155.7		171.6		195.1		229.7		259.4		
1538	D86A	156.5		171.2		194.0		229.9		255.2		
1543		----		----		----		----		----		
1610	IP123A	154.8		171.8		194.8		230.8		258.8		
1634	D86A	151.3		172.1		194.3		230.1		258.9		
1635		----		----		----		----		----		
1636	D86A	149.2		169.0		195.1		229.9		257.2		
1651	D86	152.05		172.93		195.0		229.45		257.05		
1715	D86A	154.2		171.0		193.9		229.1		258.7		
1720	D86M	147.5		171.5		195.5		231.0		260.5		
1724	D86A	151.5		171.4		195.3		230.8		258.4		
1730		----		----		----		----		----		
1811	D86A	146.3		170.6		194.3		229.2		256.5		
1833	D86	152.8		171.2		194.8		229.6		259.1		
1842		----		----		----		----		----		
1854	D86M	151.5		170.5		192.5		226.5		256.5		
1913	D86M	151.2		171.7		194		228.6		258.4		
2129	D86A	147.5		169.4		193.0		227.5		252.7		CG(0.05) Fr 253.3
2130	D86A	152.8		171.3		193.5		227.5		259.2		
2133	D86A	151.0		171.3		193.7		227.5		257.1		
2136	D86A	150.7		172.6		194.5		228.8		258.1		
	normality	OK		not OK		OK		OK		OK		
	n	57		57		58		58		56		
	outliers	1		1		0		0		2		
	mean (n)	151.80		171.38		194.41		229.22		258.13		
	st.dev. (n)	2.168		0.960		0.909		1.290		1.512		
	R(calc.)	6.07		2.69		2.55		3.61		4.23		
	R(D86:09e1-A)	8.35		3.77		2.97		3.44		7.10		



Determination of Existent Gum on sample #1017; results in mg/100 mL

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	IP540	1.2		0.06	
463	IP540	1.4		0.23	
495	D381	0.5		-0.54	
496		----		----	
851	IP540	1		-0.11	
1017	D381	0.6		-0.46	
1021	IP540	2.0		0.75	
1026	ISO6246	2		0.75	
1032	D381	0.3		-0.71	
1034	D381	1.2		0.06	
1038	IP540	0.6		-0.46	
1039	D381	3	DG(0.05)	1.61	
1047	ISO6246	1.7		0.49	
1049	D381	1.2		0.06	
1059	D381	1		-0.11	
1062	D381	1.5		0.32	
1064	D381	0.8		-0.28	
1079	D381	0.2		-0.80	
1081	D381	2		0.75	
1097	IP540	0.9		-0.20	
1105	D381	0.70		-0.37	
1108	D381	1		-0.11	
1109	IP540	1.0		-0.11	
1126		----		----	
1131	IP540	1.6		0.40	
1140	IP540	1		-0.11	
1150		----		----	
1155	ISO6246	0.70		-0.37	
1161		----		----	
1167		----		----	
1194		----		----	
1201	D381	1.9		0.66	
1203	ISO6246	3.8	G(0.05)	2.29	
1264	D381	2.5		1.18	
1276	D381	1.6		0.40	
1277	D381	NIL		----	
1279		----		----	
1284	D381	1.5		0.32	
1288	D381	1.5		0.32	
1318		----		----	
1402	D381	<1		<-0.11	
1428	ISO6246	0.8	C	-0.28	First reported 3.8
1528	IP540	0.8		-0.28	
1531		----		----	
1538	D381	2.7		1.35	
1543		----		----	
1610	IP540	0.4		-0.63	
1634	D381	1		-0.11	
1635		----		----	
1636	D381	1.3		0.15	
1651	D381	0.6		-0.46	
1715		----		----	
1720	D381	0.8		-0.28	
1724	D381	1.8		0.58	
1730		----		----	
1811	D381	3.2	DG(0.05)	1.78	
1833	D381	0.4		-0.63	
1842		----		----	
1854	D381	1.1		-0.03	
1913	IP540	0.50		-0.54	
2129	D381	1.2		0.06	
2130	D381	1.0		-0.11	
2133	D381	1.0		-0.11	
2136	D381	0.4		-0.63	
	normality	not OK			
	n	45			
	outliers	3			
	mean (n)	1.13			
	st.dev. (n)	0.579			
	R(calc.)	1.62			
	R(D381:09)	3.26			



Determination of Flash Point on sample #1017; results in °C



lab	method	Value	mark	z(targ)	remarks
225		----		----	
353	IP170-F	42.875		-0.39	
463	IP170-AE	42.5		-0.72	
495	IP170-AE	44.5		1.03	
496	D3828-MF	42.0		-1.15	
851	IP170-MF	41.5	C	-1.59	First reported 41.2
1017	D93-AE	45.0		1.47	
1021		----		----	
1026	IP170-	44.5		1.03	
1032	IP170-AE	43.5		0.16	
1034	IP170-AE	43.5		0.16	
1038	IP170-MF	43.0		-0.28	
1039	IP170-AE	43.0		-0.28	
1047	D56-MF	43.0		-0.28	
1049	IP170-E	43		-0.28	
1059	IP170-AE	42.0		-1.15	
1062	IP170-AE	42.5		-0.72	
1064	IP170-E	43.5		0.16	
1079	IP170-AE	44.0		0.60	
1081	IP170-	43.0		-0.28	
1097	ISO13736-F	42.6		-0.63	
1105	IP170-AF	42.8		-0.45	
1108	D56-AE	42.0		-1.15	
1109	IP170-AF	42.5		-0.72	
1126	ISO2719-E	45.0		1.47	
1131	D56-AE	45.8		2.17	
1140	IP170-A	43.5		0.16	
1150		----		----	
1155		----		----	
1161	ISO271-AE	46.0	G(0.05)	2.35	
1167	IP170-	42.5		-0.72	
1194		----		----	
1201	IP170-AE	45.5		1.91	
1203	D56-F	44.5		1.03	
1264	IP170-AE	43		-0.28	
1276	IP170-MF	41.5		-1.59	
1277	IP170-A	44.5		1.03	
1279	IP170-AE	44.0		0.60	
1284	D56-F	42.0		-1.15	
1288	D3828-AF	43.0		-0.28	
1318	IP170-AE	46.5	G(0.05)	2.78	
1402	IP170-AF	43.0		-0.28	
1428	D56F	43.4		0.07	
1528	D53-AE	44.5		1.03	
1531		----		----	
1538	D56-	44.3		0.86	
1543		----		----	
1610	IP170-	43.8		0.42	
1634	IP170-AE	45.0		1.47	
1635		----		----	
1636		----		----	
1651	IP170-AE	42.80		-0.45	
1715	D56-AE	42		-1.15	
1720	D3828-AF	44.0		0.60	
1724	IP170-	42.5		-0.72	
1730		----		----	
1811	IP170-F	42.3		-0.89	
1833	IP170-MF	43		-0.28	
1842		----		----	
1854	IP170-MF	44		0.60	
1913	IP170-AF	43.0		-0.28	
2129	IP170-F	42.5		-0.72	
2130	IP170-AF	43.5		0.16	
2133	D56-E	47.0	G(0.05)	3.22	
2136	D56-AE	44.0		0.60	
					<u>After excluding ASTM D93/ISO2719:</u>
	normality	not OK			not OK
	n	51			49
	outliers	3			2
	mean (n)	43.32			43.25
	st.dev. (n)	1.027			0.987
	R(calc.)	2.87			2.76
	R(IP170:09)	3.20			3.20
					Compare R(IP170:99) = 1.50

ONLY ASTM D56

normality OK
 n 10
 outliers 0
 mean (n) 43.55
 st.dev. (n) 1.298
 R(calc.) 3.63
 R(D56:05) 4.30



ONLY IP170/ISO13736

normality not OK
 n 36
 outliers 1
 mean (n) 43.19
 st.dev. (n) 0.901
 R(calc.) 2.52
 R(IP170:09) 3.20



Determination of Freezing Point on sample #1017; results in °C

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	IP16-M	-54.77		0.85	
463	D2386-M	-54.5		1.15	
495	D2386-M	-56.5		-1.09	
496	D2386-M	-55.5		0.03	
851	D2386-M	-55.7		-0.19	
1017		----		----	
1021	D2386-M	-55.64		-0.12	
1026	D2386-M	-56.5		-1.09	
1032	IP529-A	-55.5		0.03	
1034	D2386-A	-55.1		0.48	
1038	D5972-A	-56.3		-0.86	
1039	D2386-A	-55.5		0.03	
1047	D2386-A	-55.4		0.14	
1049	D7153-A	-55.3		0.26	
1059	D2386-M	-55.0		0.59	
1062	IP529-A	-55.1		0.48	
1064	D7153-A	-55.2		0.37	
1079	D5972-A	-55.7		-0.19	
1081	D7153-A	-55.2		0.37	
1097	IP529-A	-55.2		0.37	
1105	D2386-A	-55.3		0.26	
1108	D5972-A	-55.9		-0.42	
1109	D5972-A	-56.7		-1.31	
1126		----		----	
1131	D2386-	-56.0		-0.53	
1140	D5972-A	-56.0		-0.53	
1150		----		----	
1155	ISO3013-M	-56.5		-1.09	
1161		----		----	
1167		----		----	
1194		----		----	
1201	D2386-M	-54.5		1.15	
1203	D2386-M	-58	G(0.01)	-2.77	
1264	D5972-A	-55.2		0.37	
1276	D2386-M	-50	G(0.01)	6.19	
1277	D2386-A	-54.9		0.70	
1279	D2386-M	-55.0		0.59	
1284	D7153-A	-55.1		0.48	
1288	D2386-A	-55.4		0.14	
1318	D7153-A	-55.1		0.48	
1402	D2386-A	-55.1		0.48	
1428	D7153-A	-55.2		0.37	
1528	D5901-A	-55.76		-0.26	
1531		----		----	
1538	D597-M	-57.1		-1.76	
1543		----		----	
1610	IP16-M	-55.5		0.03	
1634	D2386-	-56.0		-0.53	
1635		----		----	
1636	D2386-M	-55.0		0.59	
1651	D2386-	-56.03		-0.56	
1715	D5972-A	-55.9		-0.42	
1720	D5972-A	-55.8		-0.30	
1724	D2386-	-55.5		0.03	
1730	D5901-A	-55.09		0.49	
1811	D5972-A	-56.2		-0.75	
1833	IP528-A	-55.1		0.48	
1842		----		----	
1854	D2386-M	-55.5		0.03	
1913	D7153-A	-55.6		-0.08	
2129	D2386-A	-55.0		0.59	
2130	D2386-M	-56.0		-0.53	
2133	D2386-A	-55.6		-0.08	
2136	D7153-A	-55.3		0.26	
	normality	OK			
	n	52			
	outliers	2			
	mean (n)	-55.53			
	st.dev. (n)	0.553			
	R(calc.)	1.55			
	R(D2386:06)	2.50			





-- empty page --

Determination of JFTOT; Tube Rating, Delta P in mmHg, Pumped Vol. in mL, Temp. in °C

lab	method	tube	mark	Delta P	mark	volume	mark	temp	mark	remarks
225		----		----		----		----		
353		----		----		----		----		
463		----		----		----		----		
495	D3241	1		0		485		260		
496		----		----		----		----		
851	D3241	<1		0.1		500		260		
1017		----		----		----		----		
1021	D3241	0		0		450		260		
1026	D3241	3		1.0		----		260		
1032	D3241	1		0.0		455		260		
1034	D3241	1		0		450		260		
1038		----		----		----		----		
1039	D3241	<1		<1.0		450		260		
1047	D3241	0		2		475		260		
1049	D3241	1		0		510		260		
1059		----		----		----		----		
1062	D3241	0		0		500		260		
1064	D3241	0		0		450		260		
1079	D3241	<1		0		455		260		
1081	D3241	<2		1		450		260		
1097	D3241	1		0.0		480		260		
1105	D3241	1		1		465		260		
1108		----		----		----		----		
1109	D3241	<1		0		440		260		
1126		----		----		----		----		
1131		----		----		----		----		
1140	D3241	1		<1		510		260		
1150		----		----		----		----		
1155		----		----		----		----		
1161		----		----		----		----		
1167		----		----		----		----		
1194		----		----		----		----		
1201	D3241	<1		1		450		260		
1203	D3241	1		0		450		260		
1264	D3241	<1		0.6		500		260		
1276	D3241	<1		0		450		260		
1277	D3241	1		2		450		260		
1279	D3241	<1		0.0		450		260		
1284	D3241	<1.0		0.1		510		260		
1288		----		----		----		----		
1318	D3241	<1		0.0		450		260		
1402	D3241	1		0		460		260		
1428	D3241	<1		0		450		260		
1528	D3241	1		0		450		260		
1531		----		----		----		----		
1538	D3241	<1		1		450		260		
1543		----		----		----		----		
1610	IP323	<1		<1		455		260		
1634		----		----		----		----		
1635		----		----		----		----		
1636		----		----		----		----		
1651	D3241	1		0		450		260		
1715		----		----		----		----		
1720		----		----		----		----		
1724	D3241	<1		0		450		260		
1730		----		----		----		----		
1811		----		----		----		----		
1833		----		----		----		----		
1842		----		----		----		----		
1854	D3241	1		----		600	G(0.01)	260		
1913	D3241	<1		0.0		450		260		
2129	D3241	1		0		450		260		
2130	D3241	1		0		450		260		
2133	D3241	<1		1		475		260		
2136	D3241	1		0		450		260		
	normality	not OK		not OK		not OK		n.a.		
	n	21		32		36		38		
	outliers	0		2		1		0		
	mean (n)	0.90		0.34		463.2		260		
	st.dev. (n)	0.625		0.588		21.32		0		
	R(calc.)	1.7		1.65		59.7		0		
	R(D3241:09)	n.a.		n.a.		n.a.		n.a.		



Determination of Mercaptan Sulphur on sample #1017; results in % M/M

lab	method	value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463		----		----	
495	D3227	0.0014		-1.07	
496		----		----	
851	D3227	0.00158		0.27	
1017		----		----	
1021		----		----	
1026	D3227	0.0007	G(0.01)	-6.30	
1032		----		----	
1034	D3227	0.0014		-1.07	
1038	D3227	0.0015		-0.32	
1039	D3227	0.0018		1.92	
1047	D3227	0.00164		0.72	
1049		----		----	
1059	D3227	0.0016		0.42	
1062	D3227	0.0016		0.42	
1064	D3227	0.00152		-0.17	
1079	D3227	0.0016		0.42	
1081	D3227	0.0009	G(0.01)	-4.81	
1097	ISO3012	0.0014		-1.07	
1105	D3227	0.00146		-0.62	
1108	D3227	0.0014		-1.07	
1109	D3227	0.00151		-0.25	
1126		----		----	
1131	D3227	0.0016		0.42	
1140	D3227	0.0015	C	-0.32	First reported 15
1150		----		----	
1155	D3227	0.001463		-0.60	
1161	ISO3012	0.00154		-0.02	
1167	D3227	0.00312	G(0.01)	11.78	
1194		----		----	
1201	D3227	<0.0003		----	False negative?
1203	UOP163	0.00149		-0.40	
1264	D3227	0.0020	C	3.41	First reported 0.0022
1276	D3227	0.00172		1.32	
1277		----		----	
1279		----		----	
1284		----		----	
1288		----		----	
1318		----		----	
1402	D3227	0.0015		-0.32	
1428	ISO3012	0.0017		1.17	
1528	D3227	0.00152		-0.17	
1531		----		----	
1538	D3227	0.00147		-0.55	
1543		----		----	
1610	IP342	0.0013		-1.82	
1634		----		----	
1635		----		----	
1636		----		----	
1651		----		-10.11	
1715		----		----	
1720		----		----	
1724	D3227	0.00146		-0.62	
1730		----		----	
1811		----		----	
1833	D3227	0.0018		1.92	
1842		----		----	
1854	D3227	0.00018	G(0.05)	-10.18	
1913	D3227	0.00152		-0.17	
2129	D3227	0.0014		-1.07	
2130	D3227	0.00155		0.05	
2133	D3227	0.0005	G(0.01)	-7.79	
2136	D3227	0.00144		-0.77	
	normality	not OK			
	n	32			
	outliers	5			
	mean (n)	0.00154			
	st.dev. (n)	0.000142			
	R(calc.)	0.00040			
	R(D3227:04a)	0.00038			





Determination of MSEP on sample #1017;

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	D3948	85		-0.61	
463	D3948	75		-2.76	
495	D3948	84		-0.83	
496	D3948	91		0.67	
851	D3948	95		1.53	
1017		----		----	
1021	D3948	78		-2.12	
1026	D3948	82		-1.26	
1032	D3948	75		-2.76	
1034	D3948	88		0.03	
1038	D3948	78		-2.12	
1039	D3948	88		0.03	
1047	D3948	89		0.24	
1049	D3948	89		0.24	
1059		----		----	
1062	D3948	88		0.03	
1064	D3948	83		-1.04	
1079	D3948	92		0.89	
1081	D3948	86		-0.40	
1097	D3948	97		1.96	
1105	D3948	85		-0.61	
1108	D3948	92		0.89	
1109	D3948	89		0.24	
1126		----		----	
1131	D3948	97		1.96	
1140	D3948	76		-2.55	
1150		----		----	
1155		----		----	
1161		----		----	
1167		----		----	
1194		----		----	
1201	D3948	88		0.03	
1203		----		----	
1264		----		----	
1276	D3948	100		2.61	
1277	D3948	91		0.67	
1279	D3948	94		1.32	
1284	D3948	85		-0.61	
1288	D3948	94		1.32	
1318	D3948	72	G(0.05)	-3.41	
1402	D3948	90		0.46	
1428		----		----	
1528	D3948	94		1.32	
1531		----		----	
1538	d3227	84		-0.83	
1543		----		----	
1610	D3948	80	C	-1.69	First reported 73
1634		----		----	
1635		----		----	
1636		----		----	
1651	D3948	95		1.53	
1715		----		----	
1720	D3948	85		-0.61	
1724	D3948	86		-0.40	
1730		----		----	
1811	D3948	97		1.96	
1833	D3948	92		0.89	
1842		----		----	
1854	D3948	86		-0.40	
1913	D3948	85		-0.61	
2129	D3948	94		1.32	
2130	D3948	83		-1.04	
2133	D3948	69	G(0.05)	-4.05	
2136	D3948	93		1.10	
	normality	OK			
	n	43			
	outliers	2			
	mean (n)	87.86			
	st.dev. (n)	6.262			
	R(calc.)	17.53			
	R(D3948:08)	13.03			





Determination of Naphthalenes on sample #1017; results in %V/V

lab	method	value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463		----		----	
495	D1840-B	0.95		-1.38	
496	D1840-A	0.97		-0.78	
851	D1840-A	0.992		-0.13	
1017		----		----	
1021	D1840-B	0.50	G(0.05)	-14.82	
1026		----		----	
1032		----		----	
1034	D1840-A	0.949		-1.41	
1038	D1840-B	0.99		-0.19	
1039	D1840-B	0.98		-0.49	
1047	D1840-B	1.06		1.90	
1049	D1840-A	0.99		-0.19	
1059		----		----	
1062	D1840-A	0.99		-0.19	
1064	D1840-A	0.986		-0.31	
1079	D1840-A	0.996		-0.01	
1081	D1840-B	0.97		-0.78	
1097	D1840-A	0.99		-0.19	
1105	D1840-A	1.0282		0.95	
1108	D1840-B	0.978		-0.55	
1109	D1840-A	1.023		0.80	
1126		----		----	
1131	D1840-B	0.97		-0.78	
1140	D1840-	1.03		1.01	
1150		----		----	
1155		----		----	
1161		----		----	
1167		----		----	
1194	D1840-A	1.79	G(0.01)	23.70	
1201	D1840-A	0.99		-0.19	
1203	D1840-A	1.05	C	1.60	
1264	D1840-A	1.04		1.30	
1276	D1840-A	1.038		1.25	
1277		----		----	
1279		----		----	
1284		----		----	
1288	D1840-B	0.99		-0.19	
1318	D1840-B	1.016		0.59	
1402		----		----	
1428		----		----	
1528	D1840-B	1.12	G(0.01)	3.69	
1531		----		----	
1538	D1840-b	0.975		-0.64	
1543		----		----	
1610	D1840-A	0.975		-0.64	
1634		----		----	
1635		----		----	
1636	D1840-B	0.986		-0.31	
1651		----		----	
1715		----		----	
1720	D1840-A	0.98		-0.49	
1724		----		----	
1730		----		----	
1811	D1840-A	1.054		1.72	
1833		----		----	
1842		----		----	
1854		----		----	
1913	D1840-B	0.976		-0.61	
2129	D1840-B	1.203	G(0.01)	6.17	
2130	D1840-B	0.989		-0.22	
2133	D1840-A	2.338	G(0.01)	40.06	
2136	D1840-A	0.98		-0.49	
	normality	not OK			
	n	32			
	outliers	5			
	mean (n)	0.996			
	st.dev. (n)	0.0293			
	R(calc.)	0.082			
	R(D1840:07-B)	0.094			Compare R(D1840:07-A) = 0.060





Determination of Smoke Point on sample #1017; results in mm

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	IP57	24.311		0.31	
463		----		----	
495	D1322	23.8		-0.17	
496	D1322	25.0		0.95	
851	D1322	26.0		1.88	
1017		----		----	
1021	D1322	23.69		-0.27	
1026	D1322	25.0		0.95	
1032	D1322	24		0.01	
1034	D1322	24.5		0.48	
1038	D1322	26.0		1.88	
1039	D1322	23.5		-0.45	
1047	D1322	23		-0.92	
1049	D1322	23		-0.92	
1059	D1322	21.0		-2.79	
1062	D1322	24.5		0.48	
1064	D1322	25.2		1.13	
1079	D1322	22		-1.85	
1081	D1322	21.0		-2.79	
1097	D1322	25.16		1.10	
1105	D1322	25.2		1.13	
1108		----		----	
1109	D1322	23.35		-0.59	
1126		----		----	
1131	D1322	20.8		-2.97	
1140	D1322	23		-0.92	
1150		----		----	
1155	D1322	25.333		1.26	
1161	ISO3014	24.4		0.39	
1167		----		----	
1194	D1322	20.0		-3.72	
1201	D1322	20.5		-3.25	
1203	D1322	24.9		0.85	
1264	D1322	25		0.95	
1276	D1322	25.0		0.95	
1277	D1322	26.9		2.72	
1279		----		----	
1284	D1322	21.5		-2.32	
1288	D1322	24.0		0.01	
1318	D1322	23.992		0.01	
1402	D1322	25.0		0.95	
1428	D1322	25		0.95	
1528	D1322	23.9		-0.08	
1531		----		----	
1538	D1322	22.3		-1.57	
1543		----		----	
1610	IP57	22.5		-1.39	
1634		----		----	
1635		----		----	
1636	D1322	24.39		0.38	
1651		----		----	
1715	D1322	23.5		-0.45	
1720	D1322	23.0		-0.92	
1724	D1322	25		0.95	
1730		----		----	
1811	D1322	24.5		0.48	
1833	D1322	26.4		2.25	
1842		----		----	
1854		----		----	
1913	D1322	23.2		-0.73	
2129	D1322	27.0		2.81	
2130	D1322	25.0		0.95	
2133	D1322	26.00		1.88	
2136	D1322	24.0		0.01	
	normality	OK			
	n	49			
	outliers	0			
	mean (n)	23.98			
	st.dev. (n)	1.638			
	R(calc.)	4.59			
	R(D1322:08)	3.00			





Determination of Specific Energy on sample #1017; results in MJ/kg

lab	method	Value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463		----		----	
495	D3338	43.33		3.26	
496		----		----	
851	D3338	43.247		-1.79	
1017		----		----	
1021		----		----	
1026		----		----	
1032	D3338	43.315	C	2.35	First reported 43.108
1034	D3338	43.2644		-0.73	
1038		----		----	
1039	D3338	43.251		-1.55	
1047		----		----	
1049	D3338	43.266	C	-0.63	First reported 43.472
1059	D3338	43.254		-1.36	
1062	D3338	43.243		-2.03	
1064	D3338	43.2991		1.38	
1079	D3338	43.26		-1.00	
1081	D3338	43.296		1.19	
1097	D3338	43.248		-1.73	
1105	D3338	43.291		0.89	
1108		----		----	
1109	D3338	43.26		-1.00	
1126		----		----	
1131	D4521	43.32	ex	2.65	Result excluded, test method is not compatible, different calculation
1140	D3338	43.265		-0.69	
1150		----		----	
1155	ISO3648	43.340	ex	3.87	Result excluded, test method is not compatible, different calculation
1161		----		----	
1167		----		----	
1194	D3338	43.079	G(0.01)	-12.02	
1201	D3338	43.314		2.29	
1203	D3338	43.25		-1.61	
1264		----		----	
1276	D3338	43.3		1.44	
1277		----		----	
1279		----		----	
1284	D3338	43.251		-1.55	
1288	D3338	43.233		-2.64	
1318		----		----	
1402	D3338	43.251		-1.55	
1428		----		----	
1528	D3338	43.29263		0.99	
1531		----		----	
1538		----		----	
1543		----		----	
1610	D3338	43.271		-0.33	
1634		----		----	
1635		----		----	
1636	D3338	43.3056		1.78	
1651		----		----	
1715		----		----	
1720	D3338	43.33		3.26	
1724	D3338	43.253		-1.42	
1730		----		----	
1811	D3338	43.2494		-1.64	
1833	D3338	43.261		-0.94	
1842		----		----	
1854	D3338	43.323		2.84	
1913	D3338	43.26		-1.00	
2129	D3338	43.282		0.34	
2130	D3338	43.2917		0.93	
2133	D3338	43.288		0.71	
2136	D3338	43.302		1.56	
	normality	not OK			
	n	34			
	outliers	1			
	mean (n)	43.2764			
	st.dev. (n)	0.02763			
	R(calc.)	0.0774			
	R(D3338:09)	0.0460			






Determination of Total Acidity on sample #1017; results in mg KOH/g

lab	method	value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463	D3242	0.00375		1.00	
495		----		----	
496		----		----	
851	D3242	0.0012		-2.23	
1017	D3242	0.000	ex	-3.75	Result excluded not a real result
1021	D3242	0.003		0.05	
1026	D3242	0.005		2.58	
1032	D3242	0.0010		-2.49	
1034	D3242	0.0041		1.44	
1038	D3242	0.0029		-0.08	
1039	D3242	0.003		0.05	
1047	D3242	0.0026		-0.46	
1049	D3242	0.00269		-0.35	
1059	D3242	0.002		-1.22	
1062	D3242	0.0160	G(0.01)	16.51	
1064	D3242	0.00221		-0.95	
1079	D3242	0.00258		-0.49	
1081	D3242	0.0034		0.55	
1097	D3242	0.003		0.05	
1105	D3242	0.0008		-2.74	
1108	D3242	0.0214	G(0.01)	23.36	
1109	D3242	0.0041		1.44	
1126		----		----	
1131	D3242	0.0031		0.17	
1140	D3242	0.003		0.05	
1150		----		----	
1155		----		----	
1161		----		----	
1167		----		----	
1194		----		----	
1201	D3242	0.070	G(0.01)	84.92	
1203	D3242	0.0029		-0.08	
1264	D3242	0.0051	C	2.71	First reported 0.0086
1276	D3242	0.0032		0.30	
1277		----		----	
1279	D3242	0.002		-1.22	
1284	D3242	0.015	G(0.01)	15.25	
1288		----		----	
1318		----		----	
1402	D3242	0.004		1.31	
1428		----		----	
1528	D3242	0.0044		1.82	
1531		----		----	
1538	D3242	0.0014		-1.98	
1543		----		----	
1610	IP354	0.003		0.05	
1634		----		----	
1635		----		----	
1636	D3242	0.00163		-1.69	
1651	D3242	0.0013		-2.11	
1715		----		----	
1720		----		----	
1724	D3242	0.004		1.31	
1730		----		----	
1811		----		----	
1833	D3242	0.0067		4.73	
1842		----		----	
1854	D3242	0.0081		6.51	
1913	D3242	0.0022		-0.97	
2129	D3242	0.00200		-1.22	
2130	IP354	0.0019		-1.35	
2133	D3242	0.0009		-2.61	
2136	D3242	0.0015		-1.85	
	normality	OK			
	n	37			
	outliers	5			
	mean (n)	0.00296			
	st.dev. (n)	0.001556			
	R(calc.)	0.00436			
	R(D3242:08)	0.00221			





Determination of Total Sulphur on sample #1017; results in mg/kg

lab	method	value	mark	z(targ)	remarks
225		----		----	
353	IP336	1577	ex	-2.50	Result excluded, see paragraph 4.1
463	D4294	1720		0.09	
495	D5453	1678		-0.67	
496		----		----	
851	D4294	1770.5		1.01	
1017		----		----	
1021	D2622	1734		0.35	
1026	ISO20884	1720		0.09	
1032	D5453	1788		1.32	
1034		----		----	
1038	D2622	1505	ex	-3.81	Result excluded, see paragraph 4.1
1039	D2622	1715	C	0.00	First reported 0.1715
1047	ISO8754	1655		-1.09	
1049	D5453	1459	ex	-4.64	Result excluded, see paragraph 4.1
1059	ISO14596	1700		-0.27	
1062	D4927	1673		-0.76	
1064	D5453	1465.9	C,ex	-4.51	First reported 0.1466, result excluded see paragraph 4.1.
1079	D5453	1699		-0.29	
1081	ISO20846	740	G(0.01)	-17.67	
1097	D5453	1824		1.98	
1105	D4294	1670		-0.81	
1108	D4294	1697		-0.33	
1109	D4294	1662		-0.96	
1126	ISO20846	1574.8	ex	-2.54	Result excluded, see paragraph 4.1
1131		----		----	
1140	IP336	1640		-1.36	
1150		----		----	
1155	D1266	1707.94		-0.13	
1161	ISO20846	1771.6		1.03	
1167	ISO20846	1582.4	ex	-2.40	Result excluded, see paragraph 4.1
1194	D4294	1847		2.39	
1201	D5453	1420	ex	-5.35	Result excluded, see paragraph 4.1
1203	ISO14596	1702		-0.23	
1264	D5453	1543	ex	-3.12	Result excluded, see paragraph 4.1
1276	D5453	1490	ex	-4.08	Result excluded, see paragraph 4.1
1277		----		----	
1279		----		----	
1284		----		----	
1288	D5453	1743		0.51	
1318	D4294	1660		-1.00	
1402	IP490	1685		-0.54	
1428	ISO8754	1604		-2.01	
1528	D4294	1650		-1.18	
1531		----		----	
1538		----		----	
1543		----		----	
1610	IP336	1634		-1.47	
1634		----		----	
1635		----		----	
1636	D4294	1774		1.07	
1651	D5453	1693.2		-0.39	
1715		----		----	
1720		----		----	
1724	ip336	1751	C	0.65	First reported 0.1751
1730		----		----	
1811	D4294	1765		0.91	
1833	D4294	1782		1.22	
1842		----		----	
1854	D5453	1735		0.36	
1913	D4294	1691.8		-0.42	
2129	IP496	1726		0.20	
2130	D5453	1750		0.64	
2133	D4294	1720		0.09	
2136		----		----	
		<u>Only group 2</u>			<u>All data</u>
	normality	OK			not OK
	n	36			45
	outliers	1			1
	mean (n)	1714.9			1674.6
	st.dev. (n)	53.96			98.11
	R(calc.)	151.1			274.7
	R(D5453:09)	154.5			151.8



Determination of Viscosity @ -20°C on sample #1017; results in cSt

lab	method	value	mark	z(targ)	remarks
225		----		----	
353		----		----	
463		----		----	
495	D445	3.6188		-0.58	
496	D445-A	3.630		-0.13	
851	D445-M	3.634		0.03	
1017	D445-M	3.5997		-1.36	
1021	D445-M	3.6215		-0.47	
1026		----		----	
1032	D445	3.621		-0.49	
1034	D445-M	3.6508		0.71	
1038	D445-M	3.617		-0.66	
1039	D445-M	3.624		-0.37	
1047	D445-M	3.652		0.76	
1049	D445-M	3.6605		1.11	
1059	D445-M	3.668		1.41	
1062	D445-M	3.633		-0.01	
1064	D445-A	3.6400		0.28	
1079	D445-M	3.63496		0.07	
1081	D445-M	3.603		-1.22	
1097	ISO3104-M	3.651		0.72	
1105	D445-M	3.5763	C	-2.31	First reported 3.5363
1108		----		----	
1109	D445-M	3.6459		0.52	
1126		----		----	
1131	D445-A	3.6560		0.93	
1140	D445-A	3.625		-0.33	
1150		----		----	
1155	D445-M	3.6159		-0.70	
1161		----		----	
1167		----		----	
1194		----		----	
1201	D445-M	3.634		0.03	
1203	ISO3104-M	3.619		-0.57	
1264	D7042-A	3.6318		-0.06	
1276		----		----	
1277	D445-M	3.96	G(0.01)	13.26	
1279		----		----	
1284	D445	3.6337		0.02	
1288		----		----	
1318		----		----	
1402	D445-M	3.642		0.36	
1428	ISO3104-A	3.642		0.36	
1528	D445-M	3.64562		0.50	
1531		----		----	
1538	D445-M	3.6849		2.10	
1543		----		----	
1610	IP71-M	3.638		0.20	
1634		----		----	
1635		----		----	
1636		----		----	
1651	D445	3.6723		1.59	
1715		----		----	
1720	D445-M	3.564		-2.81	
1724	D445	3.705	C	2.91	First reported 3.735
1730		----		----	
1811	D445-A	3.6219		-0.46	
1833	D445-M	3.62		-0.53	
1842		----		----	
1854	D445-M	3.909	G(0.01)	11.19	
1913	D445-A	3.6420		0.36	
2129	D445-M	3.633		-0.01	
2130	D445-M	3.582		-2.08	
2133	D445-A	3.6386		0.22	
2136	D445-A	3.6319		-0.05	
	normality	OK			
	n	41			
	outliers	2			
	mean (n)	3.6332			
	st.dev. (n)	0.02640			
	R(calc.)	0.0739			
	R(D445:09)	0.0690			



Determination of Particle Size Distribution on sample #1018; results per mL

lab	method	>4 µm	mark	>6 µm	mark	>14µm	mark	>21µm	mark	>25µm	mark	>30µm	mark
225		----		----		----		----		----		----	
353		----		----		----		----		----		----	
495	IP564	19507		2948		94		19		8		3	
851	IP564	17112.0		2716.3		116.1		23.8		7.3		3.5	
1023		----		----		----		----		----		----	
1038		----		----		----		----		----		----	
1039	IP565	13066		1912		103		36		21		13	
1047	ISO4407	2373	ex	1008	ex	206	ex	134	ex	106	ex	54	ex
1062	IP564	10628	DG(1)	1703		69		13		5		2	
1079	IP564	18684.4		2617.0		155.5		50.5		29.1		16.1	
1081	IP564	2638	ex	47	ex	1	ex	0	ex	0	ex	0	ex
1095	IP564	21430		4077		129		18		6		3	
1105		----		----		----		----		----		----	
1108	IP564	22504		2970		108		33		20		11	
1109	IP565	20619.9		3232.1		128.1		30.8		15.9		7.0	
1131	IP565	18671		4501		171		26		10		3	
1146	IP564	19840.3		2931.8		141.9		42.5		19.2		6.4	
1201	IP564	9691	DG(1)	565	G(1)	10	G(1)	1	G(1)	<1	false -	<1	false -
1279	IP565	18672.7		5788.6	DG(1)	250.2	G(1)	60.9		30.7		12.8	
1528	IP565	16213.5		3295.1		86		16.8		6.8		2.9	
1724	IP565	18392.3		2920.6		131.8		31.7		15.6		8.5	
1811	IP565	20369.1		3775.9		132.2		30.1		11.2		3.7	
1833	IP565	14561.7		2330.5		99.4		22.6		10.5		4.5	
1913	IP565	21575.1		5564.9	DG(1)	113.9		30.3		15.9		8.0	
2130	IP564	22650.6		3584.2		69.6		12.0		4.2		1.3	
	normality	OK		OK		OK		OK		OK		OK	
	n	16		15		16		16		17		17	
	outliers	2		3		2		1		0		0	
	mean (n)	18992		3034		115		29.2		13.9		6.5	
	st.dev. (n)	2714.5		757.6		28.7		13.07		8.06		4.46	
	R(calc.)	7601		2121		80		36.6		22.6		12.5	
*)	R(IP564:10)	3575		959		63		37.1		18.6		10.7	
	R(IP565:10)	2091		704		71		24.8		14.5		8.9	

*) IP564 is used for calculation of the z-scores


Lab 1047: results are cumulative
 Lab 1081: analyzed wrong sample

G(1) = Grubbs outlier G(0.01); DG(1) = Grubbs outlier DG(0.01)

7

z-scores particle size

lab	method	>4 µm	>6 µm	>14µm	>21µm	>25µm	>30µm
225		----	----	----	----	----	----
353		----	----	----	----	----	----
495		0.40	-0.25	-0.95	-0.77	-0.89	-0.91
851		-1.47	-0.93	0.03	-0.41	-0.99	-0.77
1023		----	----	----	----	----	----
1038		----	----	----	----	----	----
1039		-4.64	-3.28	-0.55	0.51	1.07	1.72
1047		-13.02	-5.91	4.00	7.91	13.85	12.47
1062		-6.55	-3.89	-2.06	-1.23	-1.34	-1.17
1079		-0.24	-1.22	1.77	1.61	2.28	2.53
1081		-12.81	-8.72	-5.07	-2.21	-2.09	-1.69
1095		1.91	3.04	0.60	-0.85	-1.19	-0.91
1105		----	----	----	----	----	----
1108		2.75	-0.19	-0.33	0.28	0.92	1.19
1109		1.28	0.58	0.56	0.12	0.30	0.14
1131		-0.25	4.28	2.45	-0.24	-0.59	-0.91
1146		0.66	-0.30	1.17	1.00	0.80	-0.01
1201		-7.29	-7.21	-4.67	-2.13	----	----
1279		-0.25	8.04	5.96	2.39	2.53	1.66
1528		-2.18	0.76	-1.31	-0.94	-1.07	-0.93
1724		-0.47	-0.33	0.72	0.19	0.25	0.54
1811		1.08	2.16	0.74	0.07	-0.41	-0.72
1833		-3.47	-2.05	-0.71	-0.50	-0.51	-0.51
1913		2.02	7.39	-0.07	0.08	0.30	0.41
2130		2.87	1.61	-2.03	-1.30	-1.46	-1.35



APPENDIX 2

Z-scores of individual participants for distillation of sample #1017

lab	IBP	10%	50%	90%	FBP
225	----	----	----	----	----
353	0.00	0.09	0.84	1.86	0.93
463	0.44	1.43	-0.01	0.47	1.17
495	0.13	-0.80	-1.23	-0.83	-1.51
496	----	----	----	----	----
851	-0.10	0.46	0.18	0.15	0.10
1017	0.07	-1.17	-1.52	-2.29	-1.47
1021	0.00	0.31	0.27	1.86	0.26
1026	-1.58	1.20	0.56	0.64	0.14
1032	0.13	-0.21	0.27	0.72	0.03
1034	0.74	0.31	-0.10	0.47	0.66
1038	0.84	-0.73	-0.67	0.31	0.07
1039	0.00	1.13	0.09	0.07	-0.49
1047	0.64	0.09	0.84	1.78	0.30
1049	-0.64	0.46	-0.67	-0.58	-0.61
1059	-0.20	0.31	0.93	0.47	0.38
1062	0.64	0.61	0.84	0.07	-0.05
1064	1.37	0.16	0.18	0.96	0.30
1079	0.54	0.76	-0.10	-0.34	-0.57
1081	-0.30	0.24	0.27	1.37	-0.76
1097	0.20	-0.21	0.09	0.23	-0.13
1105	-0.91	-0.36	-0.39	-2.05	-3.41
1108	0.23	-0.21	-1.33	-1.56	0.14
1109	0.20	-0.65	-0.20	0.47	0.10
1126	-1.21	-0.58	2.25	-0.18	-0.61
1131	-0.03	0.98	0.27	0.39	-0.05
1140	0.74	0.83	1.22	2.35	0.66
1150	----	----	----	----	----
1155	0.19	-0.75	-1.90	-0.65	0.12
1161	-0.72	-0.10	1.50	0.68	0.20
1167	-0.60	-0.95	-0.10	-1.48	-0.41
1194	-6.34	-2.51	1.22	-0.83	-0.72
1201	0.30	0.01	-0.01	0.88	-0.09
1203	-0.57	0.39	1.31	0.56	0.26
1264	0.40	0.46	0.56	-0.18	0.93
1276	-0.81	-1.47	-1.99	-1.72	-0.37
1277	-0.54	-1.77	-1.04	-0.91	0.07
1279	0.07	0.09	0.09	-0.18	0.54
1284	0.50	0.16	-0.39	-0.75	-0.72
1288	-0.14	-0.21	-0.20	0.07	-0.37
1318	0.00	0.01	-0.39	-0.42	0.46
1402	1.24	0.24	-0.39	-1.32	0.03
1428	0.37	0.68	0.18	-0.18	1.52
1528	-0.24	0.46	0.18	0.23	-0.01
1531	1.31	0.16	0.65	0.39	0.50
1538	1.58	-0.13	-0.39	0.56	-1.16
1543	----	----	----	----	----
1610	1.01	0.31	0.37	1.29	0.26
1634	-0.17	0.53	-0.10	0.72	0.30
1635	----	----	----	----	----
1636	-0.87	-1.77	0.65	0.56	-0.37
1651	0.08	1.15	0.56	0.19	-0.43
1715	0.80	-0.28	-0.48	-0.10	0.22
1720	-1.44	0.09	1.03	1.45	0.93
1724	-0.10	0.01	0.84	1.29	0.10
1730	----	----	----	----	----
1811	-1.85	-0.58	-0.10	-0.01	-0.64
1833	0.33	-0.13	0.37	0.31	0.38
1842	----	----	----	----	----
1854	-0.10	-0.65	-1.80	-2.21	-0.64
1913	-0.20	0.24	-0.39	-0.50	0.10
2129	-1.44	-1.47	-1.33	-1.40	-2.14
2130	0.33	-0.06	-0.86	-1.40	0.42
2133	-0.27	-0.06	-0.67	-1.40	-0.41
2136	-0.37	0.91	0.09	-0.34	-0.01

APPENDIX 3**List of participants**

Number of laboratories	Country
2 laboratories in	AUSTRALIA
5 laboratories in	BELGIUM
1 laboratory in	BULGARIA
1 laboratory in	CÔTE D'IVOIRE
1 laboratory in	CROATIA
1 laboratory in	CYPRUS
1 laboratory in	CZECH REPUBLIC
1 laboratory in	DENMARK
1 laboratory in	FRANCE
6 laboratories in	GERMANY
4 laboratories in	GREECE
1 laboratory in	HONG KONG
2 laboratories in	HUNGARY
1 laboratory in	IRELAND
1 laboratory in	LITHUANIA
1 laboratory in	NORWAY
3 laboratories in	P.R. of CHINA
2 laboratories in	POLAND
1 laboratory in	PORTUGAL
1 laboratory in	QATAR
1 laboratory in	REPUBLIC OF MACEDONIA
1 laboratory in	ROMANIA
1 laboratory in	SERBIA
1 laboratory in	SLOVENIA
2 laboratories in	SUDAN
1 laboratory in	SWEDEN
6 laboratories in	THE NETHERLANDS
6 laboratories in	TURKEY
2 laboratories in	U.A.E.
6 laboratories in	UNITED KINGDOM
1 laboratory in	ZAMBIA

APPENDIX 4

Abbreviations:

C	= final result after checking of first reported suspect result
U	= reported in wrong unit
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
ex	= excluded from calculations
E	= error in calculations
n.a.	= not applicable
W	= withdrawn on request participant
fr.	= first reported
U	= reported in a deviating unit
SDS	= Safety Data Sheet

Literature:

- 1 i.i.s. Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, January 2010
- 2 Defence Standard 91-91, Issue 6 Publication date 8 April '08.
- 3 Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS), Issue 22–28 June 2007, Bulletin No. 21.
- 4 ASTM E178-02
- 5 ASTM E1301-03
- 6 ISO 13528-05
- 7 ISO 5725, parts 1-6, 1994
- 8 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 9 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 10 IP 367/84
- 11 DIN 38402 T41/42
- 12 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 13 J.N. Miller, Analyst, 118, 455, (1993)
- 14 Analytical Methods Committee Technical Brief, No4 January 2001
- 15 The Royal Society of Chemistry 2002, Analyst 2002, 127 page1359-1364, P.J. Lowthian and M. Thompson. (see <http://www.rsc.org/suppdata/an/b2/b205600n/>)