

**Results of Proficiency Test
Dissolved Gas Analysis
November 2010**

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: Ing. M. Jafar
Corrector: Dr. R.G. Visser & ing. R. Starink
Report: iis10L03DGA

January 2011

CONTENTS

1	INTRODUCTION.....	3
2	SET UP.....	3
2.1	QUALITY SYSTEM.....	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALITY STATEMENT.....	3
2.4	SAMPLES.....	4
2.5	STABILITY OF THE SAMPLES.....	5
2.6	ANALYSES.....	5
3	RESULTS.....	5
3.1	STATISTICS.....	5
3.2	GRAPHICS.....	6
3.3	Z-SCORES.....	6
4	EVALUATION.....	7
4.1	EVALUATION PER TEST.....	7
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	9
4.3	COMPARISON OF THE PROFICIENCY TEST OF NOVEMBER 2010 WITH PREVIOUS PTS.....	9
4.4	DISCUSSION.....	10

Appendices:

1.	Data and statistical results.....	12
2.	Number of participants per country.....	22
3.	Abbreviations and literature.....	23

1 INTRODUCTION

Since 2001, the Institute for Interlaboratory Studies organized a proficiency test for Transformer Oil every year. Since then several requests were received from laboratories to organize also a proficiency test for Dissolved Gas Analysis. The analytical test results of this DGA test are very important to evaluate the condition of a transformer oil. The first proficiency study for DGA was organized by iis in December 2007.

During the annual proficiency testing program 2010/2011, it was decided to continue the PT for Dissolved Gas Analysis.

In this international Interlaboratory study, 24 laboratories from 14 different countries have participated. See appendix 2 for a list of participants in alphabetical country order. In this report the results of the DGA proficiency test are presented and discussed.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkensisse, The Netherlands, was the organizer of this proficiency test.

In total one batch of 26 syringes (of 50 mL) was prepared (lot BG46) on September 8, 2010. Each syringe was uniquely numbered and provided by Morgan Schaffer, Quebec, Canada (True North). Each syringe was uniquely numbered and one syringe was sent to each participating laboratory. Participants were requested to report rounded and unrounded results. The unrounded results were preferably used for statistical evaluation.

2.1 QUALITY SYSTEM

The Institute for Interlaboratory Studies in Spijkensisse, the Netherlands, has implemented a quality system based on ISO guide 43, ILAC-G13:2007 and ISO17043:2010. This ensures 100% confidentiality of participant's data. Also customer's satisfaction is measured on regular basis by the distribution of questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2).

2.3 CONFIDENTIALITY STATEMENT

All data present in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

In this proficiency test only one sample was used. The 50 mL gas tight syringes with sample material were prepared by Morgan Schaffer (Quebec, Canada) in accordance with principles outlined in ASTM Method D3612-01, Annex A2 (2001) and IEC 60567, clause 6.2 (1992). In total one batch of 26 syringes was prepared (lot BG46) on September 8, 2010. Each syringe was uniquely numbered and a certificate of analysis was provided by Morgan Schaffer. These certificates were removed after receipt by iis prior to the forwarding of the samples to the participating laboratories. The values, given on the Morgan Schaffer certificates are listed in below table:

	Syringe S/N	Ethane in µl/L	Ethylene in µl/L	CO ₂ in µl/L	Oxygen in µl/L	Nitrogen in µl/L
sample #1088-1	4736	2.8	6.7	1500	23300	60700
sample #1088-2	4754	2.8	6.7	1500	23300	60700
sample #1088-3	4758	2.8	6.7	1500	23300	60700
sample #1088-4	4768	2.8	6.7	1500	23300	60700
sample #1088-5	4774	2.8	6.7	1500	23300	60700
sample #1088-6	4779	2.8	6.7	1500	23300	60700
sample #1088-7	4805	2.8	6.7	1500	23300	60700
sample #1088-8	4816	2.8	6.7	1500	23300	60700
sample #1088-9	4817	2.8	6.7	1500	23300	60700
sample #1088-10	4829	2.8	6.7	1500	23300	60700
sample #1088-11	5169	2.8	6.7	1500	23300	60700
sample #1088-12	5234	2.8	6.7	1500	23300	60700
sample #1088-13	5317	2.8	6.7	1500	23300	60700
sample #1088-14	5601	2.8	6.7	1500	23300	60700
sample #1088-15	5674	2.8	6.7	1500	23300	60700
sample #1088-16	5712	2.8	6.7	1500	23300	60700
sample #1088-17	5805	2.8	6.7	1500	23300	60700
sample #1088-18	5887	2.8	6.7	1500	23300	60700
sample #1088-19	5989	2.8	6.7	1500	23300	60700
sample #1088-20	6047	2.8	6.7	1500	23300	60700
sample #1088-21	6087	2.8	6.7	1500	23300	60700
sample #1088-22	6273	2.8	6.7	1500	23300	60700
sample #1088-23	6419	2.8	6.7	1500	23300	60700
sample #1088-24	6769	2.8	6.7	1500	23300	60700
sample #1088-25	6769	2.8	6.7	1500	23300	60700
sample #1088-26	6769	2.8	6.7	1500	23300	60700

Table 1: homogeneity test of sub samples #1088

From above data, the homogeneity of the prepared syringes was judged to be sufficiently to be used for the proficiency test.

To each of the participating laboratories one syringe of 50 mL (labelled #1088) was sent on October 20, 2010.

2.5 STABILITY OF THE SAMPLES

Morgan Schaffer declare that bulk storage prior to shipping has a shelf life of at least 6 months. This was assumed to be sufficient for the proficiency testing purposes.

2.6 ANALYSES

The participants were asked to determine on sample #1088: Hydrogen, Oxygen, Nitrogen, Carbon Monoxide, Carbon Dioxide, Methane, Ethane, Ethene, Ethyn, Propane and Propene. Also some method details were requested to be reported.

To get comparable results a detailed report form, on which the units were prescribed, was sent together with each set of samples. Also a letter of instructions and a SDS were added to the package as well as the Morgan Shaffer procedure how to deal with small gas bubbles in the syringe that may be present after transport.

3 RESULTS

During four weeks after sample despatch, the results of the individual laboratories were gathered. The original results are tabulated per determination in the appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline the available results were screened for suspect data. A result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the results. Additional or corrected data are put under 'Remarks' in the result tables in appendix 1. Results that came in after deadline were not taken into account in the screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded results. Results reported as '<...' or '>...' were not used in the statistical evaluation.

First the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test. After removal of outliers this check was repeated. In case a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

In accordance with ISO 5725 (1986 and 1994) the original results per determination were submitted subsequently to Dixon and Grubbs outlier tests. Outliers are marked by D(0.01) for the Dixon test and by G(0.01) or DG(0.01) for the Grubbs test. Stragglers are marked by D(0.05) for the Dixon test and by G(0.05) or DG(0.05) for the Grubbs test. Both outliers and stragglers were not included in the calculations of the averages and the standard deviations.

Finally the reproducibilities were calculated from the standard deviations by multiplying these with a factor of 2.8.

3.2 GRAPHICS

In order to visualise the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the results from a sample are plotted. The corresponding laboratory numbers are under the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected standard. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms (see appendix 3; nr.13 and 14).

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against a target standard deviation, the z-scores were calculated using the IEC 60567 reproducibilities standard deviations. This results in an evaluation independent of the spread of this interlaboratory study. The target standard deviation was calculated from the literature reproducibility by division with 2.8.

The standard uncertainty (u_x) was calculated from the (target) standard deviation in accordance with ISO13528, paragraph 5.6:

$$u_x = 1.23 * (\text{st.dev} (n)) / \sqrt{n}$$

In ISO13528 is stated that if $u_x \geq 0.3 * \text{standard deviation}$ for proficiency testing, the uncertainty of the assigned value is not negligible and need to be included in the interpretation of the results of the proficiency test. Therefore in these cases (Hydrogen and Ethane) z'-scores were calculated in stead of the usual z-scores.

The z-scores were calculated in accordance with:

$$z(\text{target}) = (\text{result} - \text{average of PT}) / \text{target standard deviation}$$

The $z'_{(\text{target})}$ were calculated in accordance with ISO13528 paragraph 7.6:

$$z'_{(\text{target})} = (\text{result} - \text{mean of PT}) / \sqrt{((\text{target standard deviation})^2 + (u_x)^2)}$$

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. The evaluation of $z'_{(\text{target})}$ is not different as for common z-scores and both are evaluated as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this proficiency test no problems were encountered during execution. All participants did report one or more test results. In total 24 participating laboratories reported 218 numerical results. Observed were 10 outlying results, which is 4.6% of the numerical results. In proficiency studies outlier percentages of 3% - 7.5% are quite normal.

4.1 EVALUATION PER TEST

In this section the results are discussed per component. The methods, which are used by the various laboratories, are taken into account for explaining the observed differences when possible and applicable. These methods are also in the tables together with the original data. The abbreviations, used in these tables, are listed in appendix 3.

All original data sets proved to have a normal distribution.

All eight test results reported by laboratory 398 were very low. Three of the eight test results appeared to be statistical outliers and at least two false negative test results were reported. As the eight test results are not independent, it was decided not to use any of the eight test results for the statistical evaluation.

Hydrogen: The determination of this component was problematic. Only one statistical outlier was observed. However, the calculated reproducibility is, after rejection of the statistical outlier, not at all in agreement with the requirements of IEC 60657:2005.

Oxygen: The determination of this component was problematic. Only one statistical outlier was observed. However, the calculated reproducibility is, after rejection of the statistical outlier, not in agreement with the requirements of IEC 60657:2005.

- Nitrogen: The determination of this component was problematic. No statistical outlier was observed. However, the calculated reproducibility is not in agreement with the requirements of IEC 60657:2005.
- Carbon monoxide: The determination of this component was problematic. Two statistical outliers were observed. Also, the calculated reproducibility is, after rejection of the statistical outliers, not in agreement with the requirements of IEC 60657:2005.
- Carbon dioxide: The determination of this component was problematic. Two statistical outliers were observed. Also, the calculated reproducibility is, after rejection of the statistical outliers, not in agreement with the requirements of IEC 60657:2005.
- Methane: The determination of this component was problematic. No statistical outlier was observed. However, the calculated reproducibility is not in agreement with the requirements of IEC 60657:2005.
- Ethane: The determination of this component was problematic. Only one statistical outlier was observed. However, the calculated reproducibility is, after rejection of the statistical outlier, not in agreement with the requirements of IEC 60657:2005.
- Ethene: The determination of this component was problematic. Two statistical outliers were observed. Also, the calculated reproducibility is, after rejection of the statistical outliers, not in agreement with the requirements of IEC 60657:2005.
- Ethyn: The determination of this component was problematic. Only one statistical outlier was observed. However, the calculated reproducibility is, after rejection of the statistical outlier, not in agreement with the requirements of IEC 60657:2005.
- Propane & Propene: To few analytical test results were received to draw any significant conclusions.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant standard and the reproducibility as found for the group of participating laboratories. The average results per component, calculated reproducibilities and reproducibilities, derived from literature standards (in casu IEC 60657:2005) are compared in the next table.

Parameter	unit	n	average	2.8 * sd	R(lit)
Hydrogen H ₂	µl/L	22	20.8	16.1	4.2
Oxygen O ₂	mg/L	19	23.1	13.6	4.6
Nitrogen N ₂	mg/L	20	61.0	34.7	12.2
Carbon Monoxide CO	µl/L	22	185.7	85.5	37.1
Carbon Dioxide CO ₂	µl/L	22	1467	469.6	293.5
Methane CH ₄	µl/L	23	9.8	4.9	2.0
Ethane C ₂ H ₆	µl/L	22	2.5	1.8	0.5
Ethene C ₂ H ₄	µl/L	21	6.6	3.4	1.3
Ethyn C ₂ H ₂	µl/L	22	2.7	1.7	0.5
Propane C ₃ H ₈	µl/L	4	n/a	n/a	n/a
Propene C ₃ H ₆	µl/L	5	n/a	n/a	n/a
Sum of Propane C ₃ H ₈ and Propene C ₃ H ₆	µl/L	6	4.6	4.6	n/a

Table 2: Performance of the group on sample #1088

Without further statistical calculations it can be concluded from the overview given in table 2 that there is not a compliance of the performance of the group of participating laboratories with the relevant standard IEC 60657:2010.

The problematic components have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF NOVEMBER 2010 WITH PREVIOUS PTS

	<i>November 2010</i>	<i>November 2009</i>	<i>January 2009</i>	<i>December 2007</i>
Number of reporting labs	24	18	14	13
Number of results reported	218	182	140	129
Statistical outliers	10	9	8	7
Percentage outliers	4.6%	5.0%	5.7%	5.4%

Table 3: Comparison of statistical summary parameters with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective standards. The conclusions are given the following table:

Determination	November 2010	November 2009	January 2009	December 2007
Hydrogen H ₂	--	--	--	--
Oxygen O ₂	-	-	-	--
Nitrogen N ₂	-	--	--	--
Carbon Monoxide CO	-	--	--	--
Carbon Dioxide CO ₂	-	-	-	--
Methane CH ₄	-	--	-	--
Ethane C ₂ H ₆	--	--	-	--
Ethene C ₂ H ₄	-	--	-	--
Ethyn C ₂ H ₂	--	--	-	--
Propane C ₃ H ₈	n.e.	n.e	-	--
Propene C ₃ H ₆	n.e.	n.e	-	--

Table 4: Comparison determinations against IEC 60657:2005

The performance of the determinations against the requirements of the respective standards is listed in the above table. The following performance categories were used:

- ++: group performed much better than the standard
- + : group performed better than the standard
- +/-: group performance equals the standard
- : group performed worse than the standard
- : group performed much worse than the standard
- n.e.: not evaluated

4.4 DISCUSSION

The consensus values as determined in this PT are compared with the average values from the homogeneity testing by Morgan Schaffer in the following table. From this comparison it is clear that all consensus values as determined in this PT are very well in line with the values as determined by Morgan Schaffer after the preparation of the syringes.

Parameter	Average values by Morgan Schaffer in µL/L	Consensus values from participants results in µL/L	Absolute differences in µL/L
Hydrogen H ₂	16	21	+5
Oxygen O ₂	23300	23100	-200
Nitrogen N ₂	60700	61000	+300
Carbon Monoxide CO	207	186	-21
Carbon Dioxide CO ₂	1500	1467	-33
Methane CH ₄	9.9	9.8	-0.1
Ethane C ₂ H ₆	2.8	2.5	-0.3
Ethene C ₂ H ₄	6.7	6.6	-0.1
Ethyn C ₂ H ₂	2.9	2.7	-0.2

Table 5: comparison of consensus values with values determined by Morgan Schaffer

No correlation could be found between the testing date and the reported results, see table 6:

laboratory	<i>actual testing date</i>	laboratory	<i>actual testing date</i>
398	unknown	1516	16 November 2010
445	unknown	1529	unknown
614	11 November 2010	1660	19 November 2010
963	13 – 14 November 2010	1702	9 November 2010
1072	15 November 2010	1719	10 November 2010
1152	10 November 2010	1801	9 November 2010
1178	12 November 2010	1923	9 November 2010
1304	18 November 2010	1924	5 November 2010
1374	10 November 2010	1925	8 November 2010
1430	16 November 2010	1943	9 November 2010
1435	4 November 2010	2125	4 November 2010
1513	9 November 2010	7003	14 November 2010

Table 6: Actual testing dates

However, a correlation could be found between the methods used by the laboratories and the reported results. Five laboratories (1072, 1178, 1513, 1516 and 1702) used the Toepler method (IEC 60567 clause 7.2), eleven laboratories (398, 963, 1152, 1374, 1430, 1435, 1529, 1660, 1943, 2125 and 7003) used the head-space method (IEC 60567 clause 7.5), three laboratories (1923, 1924 and 1925) used the Kelman method, two laboratories used partial extraction (1304 and 1801) one laboratory (614) used the ToGas method, one laboratory (1719) used the ASTM D3612-B stripper column extraction and one laboratory (445) did not report the extraction method that was used.

It is remarkable to note that the headspace results are the lowest observed for all components except for Ethane and Ethene. The spread of the headspace results is larger than the spread of the results of the other test methods for all components except for Hydrogen, Ethene and Ethyn.



The target reproducibilities as required by IEC 60567 obviously appear to be very hard to meet, although the observed reproducibilities are decreasing during the subsequent annual PTs. Still, it is clear that the reproducibility requirements of IEC 60567 are quite strict as they are all smaller than the reproducibilities estimated using the Horwitz equation.

APPENDIX 1

Determination of Hydrogen on sample #1088; results in µl/L

lab	method	value	mark	z'(targ)	remarks
398	IEC60567	8.7	ex	-5.71	
445	IEC60567	17.9		-1.37	
614	IEC60567	24.114		1.56	
963	D3612C	12.281		-4.02	
1072	IEC60567	31		4.81	
1152	D3612	14.8387		-2.82	
1178	IEC60567	20.79		-0.01	
1304	in house	25		1.98	
1374	D3612 (mod)	21.2		0.18	
1430		23		1.03	
1435	IEC60567	16.7		-1.94	
1513	IEC60567	24.6		1.79	
1516	IEC60567	20.6		-0.10	
1529	IEC60567	20.9		0.04	
1660	IEC60567	16.9		-1.84	
1702	IEC60567	26		2.45	
1719	D3612B	32		5.28	
1801	IEC60567	17.7	C	-1.47	First reported 202.9
1923	IEC60567	14		-3.21	
1924	in house	14		-3.21	
1925	IEC60567	14		-3.21	
1943	D3612	19.693		-0.53	
2125	IEC60567	30.58		4.61	
7003	D3612	62.61	C,G(0.01)	19.71	First reported 42.03



		<u>Only head space results:</u>	<u>All other results:</u>
normality	OK	OK	OK
n	22	9	13
outliers	1	1	0
mean (n)	20.809	19.566	21.670
st.dev. (n)	5.7668	5.3421	6.1001
R(calc.)	16.147	14.958	17.080
R(IEC60567)	4.162	3.913	4.334
U(mean)	1.512		
R'(IEC60567)	5.937	Compare R(Horwitz) = 5.904	



Determination of Oxygen on sample #1088; results in µl/L

lab	method	value	mark	z(targ)	Remarks
398	IEC60567	7825	ex	-9.25	
445	IEC60567	21533.5		-0.94	
614	IEC60567	23215.722		0.08	
963	D3612C	11168.26	G(0.05)	-7.23	
1072	IEC60567	25568		1.51	
1152	D3612	20276.6625		-1.70	
1178	IEC60567	21362.72		-1.04	
1304	in house	26158		1.86	
1374	D3612 (mod)	25280.4		1.33	
1430		13120		-6.04	
1435	IEC60567	18661		-2.68	
1513	IEC60567	27900		2.92	
1516	IEC60567	25680.0		1.57	
1529	IEC60567	27400		2.62	
1660	IEC60567	14061		-5.47	
1702	IEC60567	32500		5.71	
1719	D3612B	22926		-0.10	
1801	IEC60567	24983.2	C	1.15	First reported 28707.9
1923		----		----	
1924		----		----	
1925		----		----	
1943	D3612	23822.175		0.45	
2125	IEC60567	26932		2.33	
7003	D3612	17247.66		-3.54	



		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	19	9	10
outliers	1	1	0
mean (n)	23085.69	20755.66	25182.72
st.dev. (n)	4854.761	5387.672	3319.705
R(calc.)	13593.33	15085.48	9295.17
R(IEC60567)	4617.14	4151.13	5036.54
U(mean)	1369.92		
R'(IEC60567)	n/a	Compare R(Horwitz) = 2279.67	



Determination of Nitrogen on sample #1088; results in µ/L

lab	method	value	mark	z(targ)	Remarks
398	IEC60567	26875	ex	-7.83	
445	IEC60567	57510.5		-0.81	
614	IEC60567	55484.202		-1.27	
963	D3612C	34652.06		-6.05	
1072	IEC60567	65873		1.11	
1152	D3612	46007.9889		-3.44	
1178	IEC60567	67377.71		1.46	
1304	in house	67917		1.58	
1374	D3612 (mod)	66818.1		1.33	
1430		41838		-4.40	
1435	IEC60567	49717		-2.59	
1513	IEC60567	74400		3.07	
1516	IEC60567	76040.0		3.45	
1529	IEC60567	72000		2.52	
1660	IEC60567	52797		-1.89	
1702	IEC60567	74614		3.12	
1719	D3612B	79271		4.19	
1801	IEC60567	60374.4	C	-0.15	First reported 695873.5
1923		----		----	
1924		----		----	
1925		----		----	
1943	D3612	61555.975		0.12	
2125	IEC60567	67526		1.49	
7003	D3612	48647.53		-2.84	



		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	20	10	10
outliers	0	0	0
mean (n)	61021.07	54155.96	67886.18
st.dev. (n)	12378.388	12304.812	8205.657
R(calc.)	34659.49	34453.47	22975.84
R(IEC60567)	12204.21	10831.19	13577.24
U(mean)	3404.51		
R'(IEC60567)	n/a	Compare R(Horwitz) = 5205.61	



Determination of Carbon monoxide on sample #1088; results in µl/L



lab	method	value	mark	z(targ)	remarks
398	IEC60567	68.4	D(0.05)	-8.84	
445	IEC60567	193.7		0.60	
614	IEC60567	188.999		0.25	
963	D3612C	132	C	-4.05	First reported 122.9795
1072	IEC60567	199		1.00	
1152	D3612	145.8844		-3.00	
1178	IEC60567	177.51		-0.62	
1304	in house	230		3.34	
1374	D3612 (mod)	224.7		2.94	
1430		170		-1.19	
1435	IEC60567	157		-2.17	
1513	IEC60567	194.4		0.65	
1516	IEC60567	160.0		-1.94	
1529	IEC60567	229		3.26	
1660	IEC60567	167.7		-1.36	
1702	IEC60567	254		5.15	
1719	D3612B	300	D(0.05)	8.61	
1801	IEC60567	171.8	C	-1.05	First reported 99.4
1923	IEC60567	170		-1.19	
1924	in house	161		-1.86	
1925	IEC60567	178		-0.58	
1943	D3612	196.303		0.80	
2125	IEC60567	218.1		2.44	
7003	D3612	167.01		-1.41	

		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	22	10	12
outliers	2	1	1
mean (n)	185.732	180.770	189.867
st.dev. (n)	30.5209	34.2011	27.9335
R(calc.)	85.458	95.763	78.214
R(IEC60567)	37.146	36.154	37.973
U(mean)	8.003		
R'(IEC60567)	n/a	Compare R(Horwitz) = 37.902	



Determination of Carbon dioxide on sample #1088; results in µl/L



lab	method	value	mark	z(targ)	remarks
398	IEC60567	525	G(0.05)	-8.99	
445	IEC60567	1417.7		-0.47	
614	IEC60567	1310.284		-1.50	
963	D3612C	1261.783		-1.96	
1072	IEC60567	1283		-1.76	
1152	D3612	1360.0235		-1.02	
1178	IEC60567	1450.27		-0.16	
1304	in house	1584		1.11	
1374	D3612 (mod)	1628.3		1.54	
1430		1325		-1.36	
1435	IEC60567	1319		-1.42	
1513	IEC60567	1460		-0.07	
1516	IEC60567	1359.2		-1.03	
1529	IEC60567	1730		2.51	
1660	IEC60567	1525		0.55	
1702	IEC60567	1878		3.92	
1719	D3612B	1724		2.45	
1801	IEC60567	1328.4	C	-1.33	First reported 1974.8
1923	IEC60567	1478		0.10	
1924	in house	1412		-0.53	
1925	IEC60567	1512		0.43	
1943	D3612	1626.573		1.52	
2125	IEC60567	1308.5		-1.52	
7003	D3612	614.76	C,G(0.01)	-8.13	First reported 809.01
	normality	OK		OK	<u>Only head space results:</u> OK
	n	22		9	<u>All other test results:</u> OK
	outliers	2		2	13
	mean (n)	1467.318		1453.7977	0
	st.dev. (n)	167.7010		174.3537	1476.681
	R(calc.)	469.563		488.190	169.4643
	R(IEC60567)	293.464		290.760	474.500
	U(mean)	43.977			295.336
	R'(IEC60567)	n/a		Compare R(Horwitz) = 219.378	



Determination of Methane on sample #1088; results in µ/L

lab	method	value	mark	z(targ)	remarks
398	IEC60567	<1		< -12.58	False negative?
445	IEC60567	8.8		-1.48	
614	IEC60567	8.444		-1.98	
963	D3612C	6.4225		-4.86	
1072	IEC60567	11		1.66	
1152	D3612	11.1206		1.83	
1178	IEC60567	9.70		-0.19	
1304	in house	10.9		1.51	
1374	D3612 (mod)	11.8		2.79	
1430		6		-5.46	
1435	IEC60567	8.6		-1.76	
1513	IEC60567	9.3		-0.76	
1516	IEC60567	10.1		0.37	
1529	IEC60567	13.2		4.79	
1660	IEC60567	9.4		-0.62	
1702	IEC60567	13		4.50	
1719	D3612B	9		-1.19	
1801	IEC60567	9.4	C	-0.62	First reported 14661.7
1923	IEC60567	10		0.23	
1924	in house	9		-1.19	
1925	IEC60567	10		0.23	
1943	D3612	10.053		0.31	
2125	IEC60567	11.73		2.69	
7003	D3612	9.28		-0.79	



		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	23	10	13
outliers	0	0	0
mean (n)	9.837	9.7606	9.896
st.dev. (n)	1.7370	2.3285	1.2069
R(calc.)	4.864	6.520	3.379
R(IEC60567)	1.967	1.952	1.979
U(mean)	0.4455		
R'(IEC60567)	n/a	Compare R(Horwitz) = 3.124	



Determination of Ethane on sample #1088; results in µ/L

lab	method	value	mark	z'(targ)	remarks
398	IEC60567	2.2	ex	-1.39	
445	IEC60567	2.5		-0.17	
614	IEC60567	2.593		0.21	
963	D3612C	1.9315		-2.49	
1072	IEC60567	3		1.87	
1152	D3612	2.2446		-1.21	
1178	IEC60567	2.23		-1.27	
1304	in house	2.8		1.05	
1374	D3612 (mod)	2.5		-0.17	
1430		5	G(0.05)	10.02	
1435	IEC60567	2.6		0.24	
1513	IEC60567	2.4		-0.58	
1516	IEC60567	2.7		0.65	
1529	IEC60567	3.28		3.01	
1660	IEC60567	2.9		1.46	
1702	IEC60567	3		1.87	
1719	D3612B	2		-2.21	
1801	IEC60567	2.1	C	-1.80	First reported 58.0
1923	IEC60567	2		-1.39	
1924	in house	1		-0.17	
1925	IEC60567	2		0.21	
1943	D3612	4.155		-2.49	
2125	IEC60567	3.18	C	1.87	First reported 3.41
7003	D3612	2.80		-1.21	



		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	22	9	13
outliers	1	1	0
mean (n)	2.542	2.843	2.333
st.dev. (n)	0.6287	0.6516	0.5414
R(calc.)	1.760	1.824	1.516
R(IEC60567)	0.508	0.569	0.467
U(mean)	0.165		
R'(IEC60567)	0.687	Compare R(Horwitz) = 0.989	



Determination of Ethene (ethylene) on sample #1088; results in µ/L

lab	method	value	mark	z(targ)	remarks
398	IEC60567	5.0	ex	-3.41	
445	IEC60567	7.3		1.46	
614	IEC60567	5.083		-3.24	
963	D3612C	5.1065		-3.19	
1072	IEC60567	7		0.82	
1152	D3612	5.4444		-2.47	
1178	IEC60567	5.55		-2.25	
1304	in house	6.1		-1.08	
1374	D3612 (mod)	7.1		1.04	
1430		2	G(0.05)	-9.76	
1435	IEC60567	5.8		-1.72	
1513	IEC60567	5.3		-2.78	
1516	IEC60567	5.1		-3.20	
1529	IEC60567	7.48		1.84	
1660	IEC60567	7.8		2.52	
1702	IEC60567	6		-1.29	
1719	D3612B	12	G(0.05)	11.41	
1801	IEC60567	5.3	C	-2.78	First reported 22.8
1923	IEC60567	7.5		1.88	
1924	in house	8		2.94	
1925	IEC60567	9		5.06	
1943	D3612	8.528		4.06	
2125	IEC60567	7.66		2.22	
7003	D3612	6.68		0.15	



		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	21	9	12
outliers	2	1	1
mean (n)	6.611	6.844	6.436
st.dev. (n)	1.2316	1.1716	1.2969
R(calc.)	3.449	3.2803	3.631
R(IEC60567)	1.322	1.369	1.287
U(mean)	0.331		
R'(IEC60567)	n/a	Compare R(Horwitz) = 2.229	



Determination of Ethyn (acetylene) on sample #1088; results in µl/L

lab	method	value	mark	z(targ)	remarks
398	IEC60567	1.4	ex	-6.86	
445	IEC60567	3.6		4.36	
614	IEC60567	2.114		-3.22	
963	D3612C	2.073		-3.43	
1072	IEC60567	3		1.30	
1152	D3612	1.7374		-5.14	
1178	IEC60567	2.54		-1.05	
1304	in house	3.1		1.81	
1374	D3612 (mod)	3.1		1.81	
1430		0.5	G(0.05)	-11.45	
1435	IEC60567	2.3		-2.27	
1513	IEC60567	3.1		1.81	
1516	IEC60567	3.5		3.85	
1529	IEC60567	3.25		2.57	
1660	IEC60567	2.5		-1.25	
1702	IEC60567	2	C	-3.80	First reported 1
1719	D3612B	4		6.40	
1801	IEC60567	3.2	C	2.32	First reported 35.2
1923	IEC60567	2.5		-1.25	
1924	in house	2.0		-3.80	
1925	IEC60567	2.5		-1.25	
1943	D3612	3.263		2.64	
2125	IEC60567	2.65		-0.49	
7003	D3612	2.37		-1.91	

		<u>Only head space results:</u>	<u>All other test results:</u>
normality	OK	OK	OK
n	22	9	13
outliers	1	1	0
mean (n)	2.745	2.583	2.858
st.dev. (n)	0.6014	0.5346	0.6393
R(calc.)	1.684	1.4970	1.790
R(IEC60567)	0.549	0.517	0.572
U(mean)	0.158		
R'(IEC60567)	n/a	Compare R(Horwitz) = 1.056	



Determination of Propane and Propene on sample #1088; results in µl/L

lab	method	Propane	Propene	Sum of Propane & Propene	remarks
398	IEC60567	<1	<1	<1	False negative?
445		----	----	----	
614		----	----	----	
963	D3612C	0	4.769	4.769	
1072		----	----	----	
1152		----	----	----	
1178	IEC60567	3.42	----	3.42	
1304		----	----	----	
1374		----	----	----	
1430		----	----	----	
1435		----	----	----	
1513		----	----	----	
1516	IEC60567	1.9	3.9	5.8	
1529	IEC60567	<0.1	5.2	5.2	
1660	IEC60567	1.9	0	1.9	
1702		----	----	----	
1719		----	----	----	
1801		----	----	----	
1923		----	----	----	
1924		----	----	----	
1925		----	----	----	
1943	D3612	----	6.398	6.398	
2125		----	----	----	
7003		----	----	----	
	normality	n.a.	n.a.	OK	
	n	4	5	6	
	outliers	0	0	0	
	mean (n)	n.a.	n.a.	4.58	
	st.dev. (n)	n.a.	n.a.	1.658	
	R(calc.)	n.a.	n.a.	4.64	
	R(IEC60567)	n.a.	n.a.	0.92	

APPENDIX 2

Number of participants per country

3 labs in AUSTRALIA
2 labs in BELGIUM
3 labs in BULGARIA
1 lab in CROATIA
1 lab in IRAN
2 labs in ITALY
1 lab in LATVIA
2 labs in MALAYSIA
1 lab in NEW ZEALAND
2 labs in PORTUGAL
1 lab in SAUDI ARABIA
3 labs in SPAIN
1 lab in THE NETHERLANDS
1 lab in UNITED KINGDOM

APPENDIX 3

Abbreviations:

C	= final result after checking of first reported suspect result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
ex	= excluded from calculations
n/a	= not applicable
W	= withdrawn on request participant
U	= reported in wrong unit
E	= error in calculations
SDS	= Safety Data Sheet

Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics and Evaluation, January 2010 (version 3.2.).
- 2 prNEN 12766-2:2000.
- 3 ASTM E178-89
- 4 ASTM E1301-89
- 5 ISO 5725-86
- 6 ISO 5725, parts 1-6, 1994
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367/84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, First reported Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee Technical Brief, No4 January 2001
- 14 The Royal Society of Chemistry 2002, Analyst 2002, 127 page1359-1364, P.J. Lowthian and M. Thompson. (see <http://www.rsc.org/suppdata/an/b2/b205600n/>)