

Results of Proficiency Test  
Aviation gasoline  
April 2012

Organised by: Institute for Interlaboratory Studies  
Spijkensisse, the Netherlands

Authors: ing. R. J. Starink  
Correctors: dr. R.G. Visser & ing. L. Sweere  
Report: iis12B02

June 2012

**CONTENTS**

1	INTRODUCTION .....	3
2	SET UP .....	3
2.1	QUALITY SYSTEM .....	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALY STATEMENT .....	3
2.4	SAMPLES .....	4
2.5	ANALYSES .....	4
3	RESULTS.....	5
3.1	STATISTICS .....	5
3.2	GRAPHICS .....	6
3.3	Z-SCORES .....	6
4	EVALUATION .....	7
4.1	EVALUATION PER TEST .....	7
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	9
4.3	COMPARISON OF THE PROFICIENCY TEST OF APRIL 2012 WITH PREVIOUS PT .....	10

## Appendices:

1.	Data and statistical results .....	11
2.	Number of participants per country .....	27
3.	Abbreviations and literature.....	28

## **1 INTRODUCTION**

On request of several participants, the Institute for Interlaboratory Studies decided to organise a new proficiency test for the analysis of Aviation Gasoline during the annual proficiency testing program 2011/2012. In this interlaboratory study 19 laboratories in 17 different countries have participated. See appendix 2 for the number of participants per country. In this report, the results of the Aviation Gasoline proficiency test are presented and discussed. This report is also electronically available through the iis internet site [www.iisnl.com](http://www.iisnl.com).

## **2 SET UP**

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organiser of this proficiency test. It was decided to evaluate the Aviation Gasoline according to the test scope of ASTM D910:11. The analyses for fit-for-use and homogeneity testing were subcontracted. In this proficiency test, the participants received one sample of Aviation Gasoline (Avgas 100LL). Participants were requested to report rounded and unrounded results. The unrounded results were preferably used for statistical evaluation.

### **2.1 QUALITY SYSTEM**

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, has implemented a quality system based on ISO17043:2010 and ILAC-G13:2007. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Also customer's satisfaction is measured on a regular basis by sending out questionnaires.

### **2.2 PROTOCOL**

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organization, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2), which can be downloaded from [www.iisnl.com](http://www.iisnl.com).

### **2.3 CONFIDENTIALITY STATEMENT**

All data present in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

## 2.4 SAMPLES

The necessary sample was obtained from an European supplier. After homogenisation, the material was transferred into 74 brown glass bottles of 1 litre (#12050). The homogeneity of the subsamples #12050 was checked by determination of Density @15°C in accordance with ASTM D4052:11 on 8 stratified randomly selected samples.

	Density @ 15°C in kg/m <sup>3</sup>
Sample #12050-1	710.77
Sample #12050-2	710.80
Sample #12050-3	710.79
Sample #12050-4	710.79
Sample #12050-5	710.78
Sample #12050-6	710.74
Sample #12050-7	710.78
Sample #12050-8	710.80

table 1: homogeneity test results of subsamples #12050

From the above test results, the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method which is in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	Density @ 15 °C in kg/m <sup>3</sup>
r (sample #12050)	0.02
reference test	D4052:11
0.3*R (reference test)	0.97

table 2: evaluation of repeatability of the subsamples #12050

The calculated repeatability is less than 0.3 times the reproducibility of the corresponding reference method. Therefore, homogeneity of the subsamples #12050 was assumed.

To the participants one 1L bottle with sample #12050 was sent on 28 March, 2012.

## 2.5 ANALYSIS

The participants were requested to determine on sample #12050: Colour, Copper Strip Corrosion 2 hrs/100 °C, Density, Distillation, Existent Gum, Freezing Point, Heat of Combustion (Net), Lead as Pb, Lead as TEL, Lead participate, MON, Potential Gum, Sulphur and Water reaction interface (volume change).

To get comparable results a detailed report form, on which the units were prescribed as well as some of the required standards and a letter of instructions were prepared and made available for download on the iis website.

A SDS and a form to confirm receipt of the samples were added to the sample package.

### 3 RESULTS

During four weeks after sample despatch, the results of the individual laboratories were gathered. The original data are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder fax was sent to the laboratories that had not reported results at that moment. Shortly after the deadline, the available results were screened for suspect data. A result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the results. Additional or corrected results are used for data analysis and original results are placed under 'Remarks' in the result tables in appendix 1.

#### 3.1 STATISTICS

Statistical calculations were performed as described in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of January 2010 (iis-protocol, version 3.2).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded results. Results reported as '<...>' or '>...>' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test. After removal of outliers, this check was repeated. Not all data sets proved to have a normal distribution, in which cases the statistical evaluation of the results should be used with due care.

In accordance to ISO 5725 (1986 and 1994) the original results per determination were submitted subsequently to Dixon and Grubbs outlier tests. Outliers are marked by D(0.01) for the Dixon test, by G(0.01) or DG(0.01) for the Grubbs test. Stragglers are marked by D(0.05) for the Dixon test, by G(0.05) or DG(0.05) for the Grubbs test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. When the uncertainty passed the evaluation, no remarks are made in the report. However, when the uncertainty failed the evaluation it is mentioned in the report and it will have consequences for the evaluation of the test results.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

### 3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported analysis results are plotted. The corresponding laboratory numbers are under the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected standard. Outliers and other data, which were excluded from the calculations, are represented as a "x". Accepted data are represented as a triangle. Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms (see appendix 3; nr.13 and 14).

### 3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM reproducibilities, the z-scores were calculated using a target standard deviation. The target standard deviation was calculated from the literature reproducibility by division with 2.8.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate the fit-for-useness of the reported test result.

The z-scores were calculated in accordance with:

$$Z_{(\text{target})} = (\text{result} - \text{average of PT}) / \text{target standard deviation}$$

The  $Z_{(\text{target})}$  scores are listed in the result tables in appendix 1.

Absolute values for  $z < 2$  are very common and absolute values for  $z > 3$  are very rare. Therefore the usual interpretation of z-scores is as follows:

$ z  < 1$	good
$1 <  z  < 2$	satisfactory
$2 <  z  < 3$	questionable
$3 <  z $	unsatisfactory

## 4 EVALUATION

No problems were encountered during the execution of this proficiency test. Three laboratories reported the test results after the final reporting date. One laboratory reported no test results. And not all laboratories were able to perform all analyses requested. Finally, 18 laboratories did report 222 numerical test results. Observed were 7 outlying test results, which is 3.2%. In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

### 4.1 EVALUATION PER TEST

In this section the results are discussed per test. For the Density determination a not normal distribution was found. For many other tests the number of reported test results was too small to determine whether the data set was normally distributed.

Aromatics: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of D6379:11.

Colour: Five laboratories reported a test result of "Blue". No further conclusions were drawn.

Copper Corrosion: No conclusions were drawn. All participants agreed on result 1.

Density @15°C: This determination was not problematic. Only one statistical outlier was observed and the calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of D4052:11 and with the much more strict requirements of the previous version D4052:02e1.

Distillation: This determination is not problematic. No statistical outliers were observed and all calculated reproducibilities, except for 10% evaporated, are in good agreement with the requirements of D86:11a.

Existent Gum: This determination was not problematic. Only one statistical outlier was observed and the calculated reproducibility, after rejection of the statistical outlier, is in good agreement with the requirements of ASTM D381:09.

Freezing Point: In this determination no problems have been observed. All reporting participants agreed on a result below -58°C. This value is the specification for freezing point according to the test scope D910:11.

Heat of Combustion: This determination was very problematic. One statistical outlier was observed and one result was excluded for statistical evaluation as the test method used ASTM D4529 is not comparable with ASTM D3338. The calculated reproducibility after rejection of the suspect data is not at all in agreement with the requirements of D3338:09. The small number of test results may (partly) explain the large spread.

Lead as Pb: This determination was problematic for two laboratories. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of D3341:05.

Lead as Tel: This determination was not problematic. No statistical outliers were observed and the calculated reproducibility is in agreement with the requirements of D3341:05. IIS converted Lead as Pb test results to Lead as Tel test results for laboratories with codes 120, 496, 1016 and 1521, in order to create a more significant evaluation.

Lead precipitate: No significant conclusions were drawn. The four reporting laboratories agreed on a value "less 1 mg/100ml".

MON: This determination was very problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not at all in agreement with the requirements of D2700:10.

Potential Gum: No significant conclusions were drawn. One statistical outlier was observed. The other seven laboratories agreed on a value "less 1 mg/100ml".

Sulphur: This determination may be problematic for two laboratories. Only two laboratories reported a numerical result between 0.036 %M/M and 0.19 %M/M. However, nine other laboratories reported that no sulphur was detectable: <0.0005 %M/M.

Water reaction: This determination may be not problematic. All participants reported 1 or below 1. Actually, the rating numbers 1, 1b, 2, 3 and 4 should be used for appearance according to D1094:07.



## 4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant standard and the reproducibility as found for the group of participating laboratories. The average values, calculated reproducibilities and reproducibilities derived from literature standards (in casu ASTM standards) are compared in the next table.

Parameter	unit	n	mean	2.8 * sd	R (lit)
Aromatics	%V/V	6	13.4	2.2	1.6
Colour	-----	5	blue	n.e	n.e
Copper Corrosion 2 hrs/100 °C	-----	16	1	n.e	n.e
Density @15°C	kg/m <sup>3</sup>	16	710.9	0.4	3.3
Distillation @ 760 mm Hg	-				
- Initial Boiling Point	°C	17	36.7	3.8	5.2
- 10% evaporated	°C	17	63.8	3.9	3.2
- 40% evaporated	°C	17	97.2	2.0	unknown
- 50% evaporated	°C	17	103.8	1.5	1.9
- 90% evaporated	°C	17	125.9	2.0	3.1
- Final Boiling Point	°C	17	152.2	2.5	6.8
Existent Gum	mg/100ml	10	0.5	0.7	(2.2)
Freezing Point	°C	10	<-58	n.e	2.5
Heat of Combustion (Net)	MJ/kg	6	43.758	0.115	0.046
Lead as Pb	g/l	8	0.55	0.01	0.03
Lead as Tel	ml/l	8	0.52	0.01	0.03
Lead participate content	mg/100ml	4	<1	n.e	(1)
MON (lean mixture)	-----	6	102.2	3.3	0.9
Potential Gum	mg/100ml	4	0.8	0.7	(3.0)
Sulphur	%M/M	9	<0.0003	n.e.	n.e.
Water reaction interface	ml	9	n.e	n.e	n.e

table 3: performance evaluation sample #12050

\*( ) = near or below the detection limit

Without further statistical calculations, it can be concluded that for several tests there is not a good compliance of the group of participants with the relevant standards. The problematic tests have been discussed in paragraph 4.1.

### 4.3 COMPARISON OF THE PROFICIENCY TEST OF APRIL 2012 WITH PREVIOUS PT

	April 2012	May 2011
Number of reporting labs	18	13
Number of results reported	222	163
Statistical outliers	7	2
Percentage outliers	3.2%	1.2%

table 4: Comparison with previous proficiency tests

In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective standards. The conclusions are given the following table:

Parameter	April 2012	May 2011
Aromatics	--	n.e.
Density @15°C	++	++
Distillation @ 760 mm Hg	+	++
Existent Gum	(++)	++
Heat of Combustion (Net)	--	--
Lead as Pb	++	--
Lead as Tel	++	--
MON (lean mixture)	--	n.e.
Potential Gum	(++)	n.e.
Sulphur	n.e.	--

table 5: comparison determinations against the standard requirements


The performance of the determinations against the requirements of the respective standards is listed in the above table. The following performance categories were used:

- ++: group performed much better than the standard
- + : group performed better than the standard
- +/-: group performance equals the standard
- : group performed worse than the standard
- : group performed much worse than the standard
- n.e.: not evaluated

**APPENDIX 1**

Determination of Aromatics on sample #12050; results in %V/V

lab	method	value	mark	z(targ)	remarks
120		----		----	
225		----		----	
273		----		----	
340	D6379	12.3		-1.99	
353		----		----	
445		----		----	
447	D1319	13.0		-0.75	
463	D1319	13.0		-0.75	
496		----		----	
606		----		----	
631		----		----	
823		----		----	
962		----		----	
1016		----		----	
1017		----		----	
1080	in house	14.23		1.45	
1293		----		----	
1521	EN15553	13.88		0.82	
1833	D2854	14.1		1.21	
	normality	OK			
	n	6			
	outliers	0			
	mean (n)	13.42			
	st.dev. (n)	0.766			
	R(calc.)	2.15			
	R(D6379:11)	1.57			



## Determination of Colour on sample #12050

lab	method	value	mark	z(targ)	remarks
120	Visual	blue		----	
225		----		----	
273		----		----	
340		----		----	
353		----		----	
445	Visual	blue		----	
447		----		----	
463		----		----	
496		----		----	
606		----		----	
631		----		----	
823		----		----	
962	D2392	blue		----	
1016		----		----	
1017		----		----	
1080	D1500	0.0		----	
1293		----		----	
1521	IP17	3.1 blue		----	
1833	D2392	blue		----	
	normality	n.e.			
	n				
	outliers	n.e.			
	mean (n)	blue			
	st.dev. (n)	n.e.			
	R(calc.)	n.e.			
	R(D2392:11)	n.e.			

## Determination of Copper Corrosion on sample #12050




lab	method	value	mark	z(targ)	remarks
120	D130	1a		----	
225	D130	1a		----	
273	D130	1a		----	
340	D130	1a		----	
353	D130	1a		----	
445	D130	1a		----	
447	D130	1a		----	
463	D130	1a		----	
496	D130	1a		----	
606		----		----	
631	D130	1a		----	
823	D130	1a		----	
962	D130	1a		----	
1016	D130	1a		----	
1017		----		----	
1080	D130	1b		----	
1293		----		----	
1521	D130	1		----	
1833	D130	1a		----	
	normality	n.a.			
	n	16			
	outliers	n.a.			
	mean (n)	1 (1a)			
	st.dev. (n)	n.a.			
	R(calc.)	n.a.			
	R(D130:10)	n.a.			

Determination of Density @ 15°C on sample #12050; results in kg/m<sup>3</sup>

lab	method	value	mark	z(targ)	remarks
120	D4052	710.9		0.01	
225	D4052	710.7		-0.17	
273	D4052	711.1		0.18	
340	D4052	710.91		0.01	
353	D4052	710.8		-0.08	
445	D4052	710.8		-0.08	
447	D4052	710.8		-0.08	
463	D4052	711.10		0.18	
496	D4052	710.89		0.00	
606	D4052	710.3	G(0.01)	-0.51	
631		-----		-----	
823	D4052	711.0		0.09	
962	D4052	711.1		0.18	
1016	D4052	710.9		0.01	
1017		-----		-----	
1080	D4052	710.8		-0.08	
1293	ISO12185	710.89		0.00	
1521	D4052	710.8		-0.08	
1833	D4052	710.8		-0.08	

normality OK  
 n 16  
 outliers 1  
 mean (n) 710.89  
 st.dev. (n) 0.124  
 R(calc.) 0.35  
 R(D4052:11) 3.25

Compare R(D4052:02e1) = 0.50



Distillation @ 760 mm Hg (automated + manual) on sample #12050.


lab	method	IBP	mark	10%	mark	40%	mark	50%	mark	90%	mark	FBP	mark
120	D86-A	35.31		63.08		97.03		103.61		125.36		153.50	
225	D86-M	39.0		62.0		96.0		103.0		125.0		151.0	
273	D86-A	36.1		64.1		96.6		103.2		125.7		152.3	
340	D86-A	36.5		64.5		97.9		104.2		126.6		152.8	
353	IP123-A	35.4		62.1		96.8		103.4		126.1		152.7	
445	D86-A	38.1		62.7		96.9		103.7		125.9	C	152.7	
447	D86-A	35.2		62.4		97.5		104.0		126.1		151.6	
463	D86-A	36.9		66.7		97.1		104.3		126.1		151.6	
496	D86-A	38.8		65.8		98.7		105.0		127.8		152.6	
606	D86-A	34.9		63.2		98.0		104.3		126.7		153.3	
631	D86-M	38.0		64.0		97.0		104.0		126.0		153	C
823	D86-A	38.1		63.3		97.8		103.9		125.2		150.2	
962	D86-M	36.0		64.0		97.0		103.5		125.0		153.0	
1016	D86-A	35.9		64.1		97.3		103.8		125.6		151.3	
1017		----		----		----		----		----		----	
1080	D86-A	36.7		62.3		96.3		103.2		125.2		151.3	
1293		----		----		----		----		----		----	
1521	D86-A	35.3		64.1		97.8		104.1		126.2		152.3	
1833	D86-	37.9		66.1		96.4		103.0		125.4		152.5	
	normality	OK		OK		OK		OK		OK		OK	
	n	17		17		17		17		17		17	
	outliers	0		0		0		0		0		0	
	mean (n)	36.71		63.79		97.18		103.78		125.88		152.22	
	st.dev. (n)	1.355		1.397		0.699		0.534		0.717		0.904	
	R(calc.)	3.79		3.91		1.96		1.50		2.01		2.53	
	R(D86:11a)	5.23		3.20		unknown		1.88		3.11		6.78	

Lab 445 first reported 115.7

Lab 631 first reported 157.0

## Z-scores



lab	IBP	10%	40%	50%	90%	FBP
120	-0.75	-0.62	----	-0.25	-0.47	0.53
225	1.22	-1.57	----	-1.16	-0.79	-0.50
273	-0.33	0.27	----	-0.86	-0.16	0.03
340	-0.11	0.62	----	0.63	0.65	0.24
353	-0.70	-1.48	----	-0.56	0.20	0.20
445	0.74	-0.96	----	-0.11	0.02	0.20
447	-0.81	-1.22	----	0.33	0.20	-0.26
463	0.10	2.54	----	0.78	0.20	-0.26
496	1.12	1.76	----	1.82	1.73	0.16
606	-0.97	-0.52	----	0.78	0.74	0.45
631	0.69	0.18	----	0.33	0.11	0.32
823	0.74	-0.43	----	0.18	-0.61	-0.83
962	-0.38	0.18	----	-0.41	-0.79	0.32
1016	-0.43	0.27	----	0.03	-0.25	-0.38
1017	----	----	----	----	----	----
1080	-0.01	-1.31	----	-0.86	-0.61	-0.38
1293	----	----	----	----	----	----
1521	-0.76	0.27	----	0.48	0.29	0.03
1833	0.64	2.02	----	-1.16	-0.43	0.12





Determination of Existent Gum on sample #12050; results in mg/100ml

lab	method	value	mark	z(targ)	remarks
120	D381	0.53		0.05	
225	D381	2.0	G(0.01)	1.96	
273					
340	D381	0.4		-0.12	
353	D381	<1			
445	D381	0.9		0.53	
447	D381	1		0.66	
463	D381	0.4		-0.12	
496	D381	0.3		-0.25	
606	D381	0.2		-0.38	
631	D381	0.4		-0.12	
823	D381	<1.0			
962	D381	0.40		-0.12	
1016	D381	<1			
1017					
1080	ISO6246	<1			
1293					
1521	D381	<1			
1833	D381	0.4		-0.12	
normality		not OK			
n		10			
outliers		1			
mean (n)		0.49			
st.dev. (n)		0.256			
R(calc.)		0.72			
R(D381:09)		2.15			





## Determination of Freezing Point on sample #12050; results in °C

lab	method	value	mark	z(targ)	remarks
120	D2386	<-75		----	
225	D2386	<-70		----	
273		----		----	
340	D2386	<-70		----	
353		----		----	
445		----		----	
447	D2386	<-58		----	
463	D2386	<-65.0		----	
496	D2386	<-77.5		----	
606		----		----	
631	D2386	<-58.0		----	
823	D2386	<-60.0		----	
962		----		----	
1016	D2386	<-58.0		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D2386	<-60		----	
1833		----		----	
	normality	n.a.			
	n	10			
	outliers	n.a.			
	mean (n)	<-58			
	st.dev. (n)	n.a.			
	R(calc.)	n.a.			
	R(D2386:06)	2.5			



Determination of Heat of Combustion (Net) on sample #12050; results in MJ/kg

lab	method	value	mark	z(targ)	remarks
120	D3338	43.737		-1.29	
225	D3338	42.79	G(0.01)	-58.93	
273		----		----	
340	D3338	43.8012		2.62	
353		----		----	
445		----		----	
447	D3338	43.769		0.66	
463	D3338	43.76		0.11	
496		----		----	
606		----		----	
631	D3338	43.7925		2.09	
823	D3338	43.689		-4.21	
962		----		----	
1016		----		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D4529	43.908	ex	9.12	Result excluded as test method is not comparable with D3338
1833		----		----	
normality		OK			
n		6			
outliers		1			
mean (n)		43.7581			
st.dev. (n)		0.04095			
R(calc.)		0.1147			
R(D3338:09)		0.0460			



Determination of Lead as Pb on sample #12050; results in g/l



lab	method	value	mark	z(targ)	remarks
120	D5059	0.55		0.40	
225		-----		-----	
273	D3341	0.54		-0.59	
340		-----		-----	
353	IP270	0.5478		0.18	
445	IP270	0.5389		-0.70	
447	IP362	0.50	G(0.01)	-4.56	
463	D3341	0.5472		0.12	
496	D3341	0.55315		0.71	
606		-----		-----	
631		-----		-----	
823		-----		-----	
962		-----		-----	
1016	D5059	0.544		-0.20	
1017		-----		-----	
1080		-----		-----	
1293		-----		-----	
1521	D5059	0.5468		0.08	
1833	EN237	41.8	G(0.01)	4090.13	Probably reported in a different unit?
normality	OK				
n	8				
outliers	2				
mean (n)	0.5460				
st.dev. (n)	0.00482				
R(calc.)	0.0135				
R(D3341:05)	0.0282				



Determination of Lead as TEL on sample #12050; results in ml/l

lab	method	value	mark	z(targ)	remarks
120		0.52		0.36	converted by iis from the reported Lead as Pb result
225		-----		-----	
273	D3341	0.511		-0.56	
340		-----		-----	
353	IP270	0.5182		0.18	
445	IP270	0.5098		-0.68	
447		-----		-----	
463	D3341	0.5176		0.11	
496		0.5233		0.70	converted by iis from the reported Lead as Pb result
606		-----		-----	
631		-----		-----	
823		-----		-----	
962		-----		-----	
1016		0.515		-0.15	converted by iis from the reported Lead as Pb result
1017		-----		-----	
1080		-----		-----	
1293		-----		-----	
1521		0.517		0.05	converted by iis from the reported Lead as Pb result
1833		-----		-----	

normality OK  
 n 8  
 outliers 0  
 mean (n) 0.5165  
 st.dev. (n) 0.00448  
 R(calc.) 0.0125  
 R(D3341:05) 0.0274




## Determination of Lead precipitate content on sample #12050; results in mg/100ml

lab	method	value	mark	z(targ)	remarks
120		----		----	
225		----		----	
273		----		----	
340		----		----	
353		----		----	
445		----		----	
447	D873	<1		----	
463		----		----	
496		----		----	
606		----		----	
631		----		----	
823	D873	0.3		----	
962	D873	<1		----	
1016		----		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D873	<1		----	
1833		----		----	
	normality	n.e.			
	n	4			
	outliers	n.e.			
	mean (n)	<1			
	st.dev. (n)	n.e.			
	R(calc.)	n.e.			
	R(D873:02)	(1)			


Determination of MON (lean mixture) on sample #12050

lab	method	value	mark	z(targ)	remarks
120		----		----	
225		----		----	
273		----		----	
340	D2700	101.5		-2.26	
353		----		----	
445	D2700	100.6		-5.06	
447	D2700	102.8		1.79	
463	D2700	101.9		-1.01	
496	D2700	102.55		1.01	
606		----		----	
631		----		----	
823		----		----	
962		----		----	
1016		----		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D2700	104.0		5.52	
1833	D2700	128.0	G(0.01)	80.19	
normality		OK			
n		6			
outliers		1			
mean (n)		102.2			
st.dev. (n)		1.17			
R(calc.)		3.3			
R(D2700:10)		0.9			



Determination of Potential Gum on sample #12050; results in mg/100ml

lab	method	value	mark	z(targ)	remarks
120		----		----	
225		----		----	
273		----		----	
340	D873	3.0	G(0.05)	----	
353		----		----	
445		----		----	
447	D873	1		----	
463	D873	<1		----	
496		----		----	
606		----		----	
631	D873	0.4		----	
823	D873	0.8		----	
962	D873	0.80		----	
1016	D873	<1		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D873	<1		----	
1833		----		----	
normality		unknown			
n		4			
outliers		1			
mean (n)		0.75			
st.dev. (n)		0.252			
R(calc.)		0.71			
R(D873:02)		(3.00)			





## Determination of Sulphur content on sample #12050; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	D7039	0.0002		----	
225		----		----	
273		----		----	
340	D2622	<0.0001		----	
353		----		----	
445	D5453	<0.0001		----	
447	IP490	0.00005		----	
463	D4294	0.0586		----	False positive result?
496	D2622	<0.0005		----	
606		----		----	
631		----		----	
823		----		----	
962	D2622	0.0516		----	False positive result?
1016	D2622	<0.0003		----	
1017		----		----	
1080	D5453	<0.0001		----	
1293		----		----	
1521	D2622	<0.0003		----	
1833	D5453	0.000012		----	
	normality	not OK			
	n	9			
	outliers	0			
	mean (n)	<0.0003			
	st.dev. (n)	n.a.			
	R(calc.)	n.a.			
	R(D2622:10)	n.a.			Application range >0.0003 %M/M

## Determination of Water reaction interface on sample #12050; results in ml

lab	method	value	mark	z(targ)	remarks
120	D1094	<0.5		----	
225		----		----	
273	D1094	0		----	
340	D1094	0.0		----	
353		----		----	
445	D1094	1		----	
447	D1094	<0.5		----	
463	D1094	1.0		----	
496		----		----	
606		----		----	
631	D1094	0.5		----	
823	D1094	<0.5		----	
962		----		----	
1016		----		----	
1017		----		----	
1080		----		----	
1293		----		----	
1521	D1094	0.5		----	
1833		----		----	
	normality	n.a.			
	n	9			
	outliers	n.a.			
	mean (n)	n.a.			
	st.dev. (n)	n.a.			
	R(calc.)	n.a.			
	R(D1094:07)	n.a.			

## **APPENDIX 2**

### **Number of participants per country**

1 lab in BELGIUM  
1 lab in CÔTE D'IVOIRE  
1 lab in FRANCE  
1 lab in GERMANY  
1 lab in HUNGARY  
1 lab in IRELAND  
1 lab in KOREA  
1 lab in MALAYSIA  
1 lab in PHILIPPINES  
1 lab in POLAND  
1 lab in SAUDI ARABIA  
1 lab in SOUTH AFRICA  
1 lab in SWEDEN  
1 lab in THE NETHERLANDS  
2 labs in TURKEY  
1 lab in U.S.A.  
2 labs in UNITED KINGDOM

### APPENDIX 3

#### Abbreviations:

C	= final result after checking of first reported suspect result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
ex	= excluded from calculations
E	= error in calculations
n.e.	= not evaluated
W	= withdrawn
fr.	= first reported
U	= reported in different unit
SDS	= Safety Data Sheet

#### Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, January 2010
- 2 ASTM E178-02
- 3 ASTM E1301-03
- 4 ISO 5725-86
- 5 ISO 5725, parts 1-6, 1994
- 6 ISO13528-05
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367/84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee Technical Brief, No4 January 2001
- 14 The Royal Society of Chemistry 2002, Analyst 2002, 127 page1359-1364, P.J. Lowthian and M. Thompson. (see <http://www.rsc.org/suppdata/an/b2/b205600n/>)
- 15 H. Verplaetse and M. Lacourt, Accred Qual Assur (2006) 11:521-522