

**Results of Proficiency Test
Biodiesel 100% FAME (B100)
April 2014**

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: ing. L. Dijkstra
Correctors: dr. R.G. Visser & ing. L. Sweere
Report no.: iis14G02

July 2014

CONTENTS

1	INTRODUCTION.....	3
2	SET UP.....	3
2.1	QUALITY SYSTEM.....	3
2.2	PROTOCOL	3
2.3	CONFIDENTIALITY STATEMENT.....	3
2.4	SAMPLES.....	4
2.5	STABILITY OF THE SAMPLES	5
2.6	ANALYSES.....	5
3	RESULTS.....	6
3.1	STATISTICS.....	7
3.2	GRAPHICS.....	7
3.3	Z-SCORES.....	8
4	EVALUATION.....	8
4.1	EVALUATION PER TEST	9
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	13
4.3	COMPARISON OF THE PROFICIENCY TEST OF APRIL 2014 WITH PREVIOUS PTS	14

Appendices:

1	Data and statistical results	16
2	Number of participants per country	75
3	Abbreviations and literature	76

1 INTRODUCTION

Since 2001, a proficiency test for Fatty Acid Methyl Esters (FAME) used as Biodiesel B100 is organised every year by the Institute for Interlaboratory Studies (iis).

In this interlaboratory study 73 laboratories from 33 different countries have participated.

See appendix 2 for a list of number of participants per country. In this report the results of the 2014 Biodiesel B100 proficiency test are presented and discussed.

2 SET UP

In this proficiency test Biodiesel B100, a sample Rapeseed methyl ester was used. Sample analyses for fit-for-use and homogeneity testing were subcontracted. In this proficiency test, the participants received, depending on the registration, from one up to four different samples of Biodiesel B100, see table below.

Samples	Amount in mL	Purpose	Spiked
#14045	1000	For regular analysis	-
#14046	100	Analysis of Phosphorus, Potassium, Sodium and Calcium & Magnesium	Phosphorus, Sodium Calcium & Magnesium
#14047	1000	Total Contamination test	Quartz material
#14048	500	Cold Soak Test	-

table 1: four different Biodiesel B100 samples used in iis14G02

The test scope was set up according to both EN14214+A1:2014 and ASTM D6751:12 specifications. Participants were requested to report the analytical results as “rounded and unrounded results” and to use the indicated units on the report form(s). The unrounded results were preferably used for statistical evaluation.

2.1 QUALITY SYSTEM

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, has implemented a quality system based on ISO/IEC17043:2010 (R007). This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Also customer's satisfaction is measured on regular basis by the distribution of questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report ‘iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation’ of April 2014 (iis-protocol, version 3.3).

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

The necessary bulk material of Biodiesel B100 was obtained from a European producer. After fit-for-use testing and homogenisation in a precleaned metal drum, the B100 was transferred to 81 brown glass bottles of 1 litre and labelled #14045.

The homogeneity of the subsamples #14045 was checked by the determination of Density in accordance with ASTM D4052 on 8 stratified randomly selected samples:

	Density at 15°C in kg/m ³
sample 1 #14045-1	883.40
sample 2 #14045-2	883.41
sample 3 #14045-3	883.41
sample 4 #14045-4	883.41
sample 5 #14045-5	883.43
sample 6 #14045-6	883.46
sample 7 #14045-7	883.35
sample 8 #14045-8	883.38

table 2: homogeneity test of subsamples #14045

	Density at 15°C in kg/m ³
r (sample #14045)	0.09
reference test	ISO12185:96
0.3*R _(reference test)	0.15

table 3: repeatability of subsamples #14045

For Subsample #14046, 3.0 kg Biodiesel #13182 used in iis13G05M that was positive on Calcium & Magnesium, Phosphorus and Sodium was diluted with the Rapeseed methyl ester to 6.7 kg. The concentrations after dilution are approx: Calcium & Magnesium (10.6 mg/kg), Phosphorus (6.0 mg/kg) and Sodium (3.7 mg/kg). The material was subsequently divided over 66 glass bottles of 0.1L and labelled #14046. The homogeneity of the subsamples #14046 was checked by determination of Phosphorus on 8 stratified randomly selected samples:

	Phosphorus in mg/kg
sample 1 #14046-1	6.6
sample 2 #14046-2	6.6
sample 3 #14046-3	6.6
sample 4 #14046-4	6.7
sample 5 #14046-5	6.5
sample 6 #14046-6	6.7
sample 7 #14046-7	6.6
sample 8 #14046-8	6.6

table 4: homogeneity test of subsamples #14046

	Phosphorus in mg/kg
r (sample #14046)	0.2
reference test	EN14106:03
0.3*R(reference test)	0.4

table 5: repeatability of subsamples #14046

For Total Contamination, out of the same batch of Biodiesel B100, another 50 amber glass bottles of 1 litre with inner and outer caps were filled.

Each sample bottle was spiked (approx 15 mg/kg) with a fresh prepared and well shaken particulate quartz material BCR-070 (\varnothing 1.2 – 20.0 μm) in oil suspension.

Therefore, an amount of the quartz suspension was weighed in the bottle. This bottle was filled up to 850 mL and subsequently labelled #14047.

After homogenization, a random sample was taken to check the Total Contamination.

For the “Cold Soak Test” determination 23 bottles of 1 litre with regular Biodiesel B100 were filled and labelled #14048. The homogeneity of the subsamples was checked by the determination of density in accordance with ISO12185.

The calculated repeatabilities for samples #14045 and #14046 were less than 0.3 times the corresponding reproducibilities of the respective reference method. Therefore, homogeneity of the subsamples was assumed.

Depending on the registration of the participant, one 1 litre bottle labelled #14045, and/or and/or one 0.1 litre bottle labelled #14046, and/or 1 litre bottle labelled #14047, and/or one 0.5 litre bottle labelled #14048, were dispatched to each of the participating laboratories on April 02, 2014.

2.5 STABILITY OF THE SAMPLES

The stability of the Biodiesel B100, packed in the brown glass bottles, was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYSES

The tests methods to be used by the participating laboratories should be in accordance with the requirements of EN14214:12+A1:2014 and/or ASTM D6751:12, e.g.:

Parameter	EN14214:12+A1:2014	Parameter	ASTM D6751:12
Acid Value	EN14104	Acid Number	ASTM D664
Carbon Residue on 10% dist.*	ISO10370*		
CFPP	EN116		
		Cloud Point	ASTM D2500
Copper Strip Corrosion	ISO2160	Copper Strip Corrosion	ASTM D130
Density @ 15°C	ISO12185		
Flash Point (Recc)	ISO3679	Flash Point	ASTM D93
Flash Point (PMcc)	ISO2719		
Iodine Value	EN14111		
Kin. Visc. @ 40°C	ISO3104	Kin. Visc. @ 40°C	ASTM D445
Oxidation Stability	EN14112	Oxidation Stability	EN15751
Sulphated Ash	ISO3987	Sulphated Ash	ASTM D874
Sulphur	ISO20846	Sulphur	ASTM D5453
Water	ISO12937	Water and Sediment	ASTM D2709
Calcium + Magnesium	EN14538	Calcium + Magnesium	EN14538
Phosphorus	EN14107	Phosphorus	ASTM D4951
Polyunsaturated esters	EN15779		
Potassium + Sodium	EN14108/14109	Potassium + Sodium	EN14538
Methanol	EN14110	Methanol	EN14110
mono-, di-, tri-Glycerides	EN14105		
Free + Total Glycerol	EN14105	Free + Total Glycerol	ASTM D6584
Total ester content	EN14103		
Linolenic Acid	EN14103		
Total Contamination	EN12662		

table 6: requirements and test methods acc. to specifications EN14214:12+A1:2014 and ASTM D6751:12

* = not applicable for B100 according to EN14214:12+A1:2014

3 RESULTS

During four weeks after sample despatch, the results of the individual laboratories were gathered. The original data are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder fax was sent to the laboratories that had not reported results at that moment. Shortly after the deadline, the available results were screened for suspect data. A result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the results. Additional or corrected results are used for data analysis and original results are placed under 'Remarks' in the result tables in appendix 1.

3.1 STATISTICS

Statistical calculations were performed as described in the report ‘iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation’ (iis-protocol, April 2014 version 3.3). For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded results. Results reported as ‘<...’ or ‘>...’ were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either ‘unknown’, ‘OK’, ‘suspect’ or ‘not OK’. After removal of outliers, this check was repeated. Not all data sets proved to have a normal distribution, in which cases the statistical evaluation of the results should be used with due care.

In accordance to ISO 5725 (1986 and 1994) the original results per determination were submitted subsequently to Dixon, Grubbs and Rosner outlier tests. Outliers are marked by D(0.01) for the Dixon test, by G(0.01) or DG(0.01) for the Grubbs test and by R(0.01) for the Rosner General ESD test (see appendix 3, no.15). Stragglers are marked by D(0.05) for the Dixon test, by G(0.05) or DG(0.05) for the Grubbs test and by R(0.05). Both outliers and stragglers were not included in the calculations of the averages and the standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. When the uncertainty passed the evaluation, no remarks are made in the report. However, when the uncertainty failed the evaluation it is mentioned in the report and it will have consequences for the evaluation of the test results.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported analysis results are plotted. The corresponding laboratory numbers are under the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected standard. Outliers and other data, which were excluded from the calculations, are represented as a “x”. Accepted data are represented as a triangle. Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms (see appendix 3; Nos.13 and 14). Also a normal Gauss curve was projected over the Kernel Density Graph.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the spread of this interlaboratory study. The target standard deviation was calculated from the literature reproducibility by division with 2.8.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated in accordance with:

$$z_{(\text{target})} = (\text{result} - \text{average of PT}) / \text{target standard deviation}$$

The $z_{(\text{target})}$ scores are listed in the result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. Therefore the usual interpretation of z-scores maybe as follows:

- $|z| < 1$ good
- $1 < |z| < 2$ satisfactory
- $2 < |z| < 3$ questionable
- $3 < |z|$ unsatisfactory

4 EVALUATION

In this proficiency test some problems were encountered during the execution.

For the regular Biodiesel PT: 12 participants reported test results after the final reporting date and 5 participants did not report any test results at all.

For the Total Contamination PT: 11 participants reported the test results after the final reporting date and 3 participants did not report any test results at all.

For the Metals in Biodiesel PT: 7 participants reported the test results after the final reporting date and 6 participants did not report any test results at all.

For the Cold Soak Test PT: 3 participants reported the test results after the final reporting date and 5 participants did not report any test results at all.

Finally, 68 participants reported in total 1093 numerical results. Observed were 54 outlying results, which is 5.2%. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

4.1 EVALUATION PER TEST

In this section, the results are discussed per sample and per test. The specified test methods and requirements acc. to EN14214:12+A1:2014 and ASTM D6751:12 were taken into account for explaining the observed differences when possible and applicable. These methods are also in the tables together with the reported data. The abbreviations, used in these tables, are listed in appendix 3.

In the iis PT reports, ASTM methods are referred to with a number (e.g. D2086) and an added designation for the year that the method was adopted or revised (e.g. D2086-08). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D2086-08 (2013)). In the results tables of Appendix 1 only the method number and year of adoption or revision will be used.

The majority of the data sets proved to have a normal distribution. For some other tests the number of reported test results was too small to determine whether the data set was normally distributed. In these cases the results of the statistical evaluations should be used with care.

For Biodiesel B100 sample #14045

<u>Acid Value (EN)</u>	This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of EN14104:03.
<u>Acid Number (ASTM)</u>	This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D664:11a (method B).
<u>Carbon Residue on 10% distillation residue</u>	This determination was very problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not at all in agreement with the requirements of ISO10370:93.
<u>Carbon Residue on 100%</u>	The consensus value found was near or below the applicable lower limit of D4530:11 (<0.1% M/M). Therefore no significant conclusions were drawn.
<u>CFPP</u>	This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN116:97.
<u>Cloud Point</u>	This determination was problematic for a number of laboratories. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D2500:11.
<u>Copper Corrosion D130 / ISO2160</u>	No problems have been observed. All participants agreed on a result of 1.

<u>Density @15°C</u>	This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.
<u>Flash Point conform EN spec.</u>	This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of ISO3679:04.
<u>Flash Point PMcc ISO2719 / D93</u>	This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of ASTM D93-C:13e.
<u>Iodine Number</u>	This determination was problematic for a number of laboratories. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers, is in agreement with the requirements of EN14111:03.
<u>Kin.Visco. @ 40°C ISO3104 / D445</u>	The determination of this component may be problematic for a number of laboratories, depending on the test method used by the laboratory. Five statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO3104:96. However, the calculated reproducibility is not in agreement with the requirements of ASTM D445:12.
<u>Oxidation Stability</u>	This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of EN14112:03.
<u>Sulphur ISO20846</u>	This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in full agreement with the requirements of ISO20846:11.
<u>Sulphur ASTM D5453</u>	This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D5453:12.
<u>Sulphated Ash</u>	All reported results were near or below the applicable lower limit of ASTM D874:13a and/or ISO3987:10 (0.005% M/M). Therefore no significant conclusions were drawn.

Water

This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12937:00. Remarkably, eight laboratories made probably a unit error in the reported test result.

Methanol

This determination was problematic for a number of laboratories. Five statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of EN14110:03 which are the same for procedure A (automatic headspace system) and B (manual procedure). When the EN14110-A and EN14110-B methods were evaluated separately, the calculated reproducibilities are respectively not and in full agreement with the requirements of EN14110:03.

mono-Glycerides

This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN14105:11 and also in good agreement with the requirements of ASTM D6584:13.

di-Glycerides

This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of EN14105:11 and also in good agreement with ASTM D6584:13.

tri-Glycerides

This determination was problematic for number of laboratories. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN14105:11 and also in good agreement with ASTM D6584:13.

Free Glycerol

This determination was not problematic. Two statistical outliers were observed and one result was excluded. However, the calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of EN14105:11.

Total Glycerol

This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN14105:11.

Total Ester content

This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of EN14103:11.

Five laboratories reported to have used the 2003 version and one laboratory the 2009 version of EN14103 instead of the 2011 version.

The 2011 version of EN14103 is mentioned in EN14214:2012+A1:2014 as the test method to be used.

[Linolenic Acid Methyl Ester](#)

This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in full agreement with the requirements of EN14103:11.

Three laboratories reported to have used the 2003 version and one laboratory the 2009 version of EN14103 instead of the 2011 version. The 2011 version of EN14103 is mentioned in EN14214:2012+A1:2014 as the test method to be used.

[Polyunsaturated](#)

All reported results, were near or below the lower application limit of EN15779:09 (0.3 – 3.0 %M/M). Therefore no significant conclusions were drawn.

For Biodiesel B100 sample #14046

[Calcium and Magnesium](#)

This determination was not problematic. Two statistical outliers and one false negative test result were observed. However, the calculated reproducibility, after rejection the suspect data is in agreement with the requirements of EN141538:06.

The samples were spiked with Calcium and Magnesium. The average recovery of 103% Calcium and Magnesium (theoretical increment of 10.6 mg/kg) can be considered as good.

[Phosphorus](#)

This determination was very problematic. One false negative test result but no statistical outliers were observed. The calculated reproducibility, after rejection of the suspect test results, is not at all in agreement with the requirements of EN14107:03.

The samples were spiked with Phosphorus. The average recovery of 112% Phosphorus (theoretical increment of 6.0 mg/kg) can be considered as satisfactory.

[Potassium](#)

All reported results except one were near or below the lower application limit of EN14109:03 (0.5 mg/kg). Therefore no significant conclusions were drawn.

[Sodium](#)

This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of EN14108:03.

The samples were spiked with Sodium. The average recovery of 109% Sodium (theoretical increment of 3.7 mg/kg) can be considered as satisfactory.

For Biodiesel B100 sample #14047

<u>Total Contamination</u>	<p>This determination was very problematic. No statistical outliers were observed. However, the calculated reproducibility is not at all in agreement with the requirements of EN12662:14.</p> <p>When the EN12662:14 data were evaluated separately, the calculated reproducibility is in full agreement with the requirements of EN12662:14</p> <p>The samples were spiked with particulate quartz material BCR. Therefore the minimum contamination concentration to be found was known (added amount = 15.31 mg/kg). The laboratories should be able to find at least 6.80 mg/kg [15.31 mg/kg_(added amount) – 8.51 mg/kg_(R EN12662:14)]. No laboratories reported lower amounts than 6.8 mg/kg.</p>
----------------------------	---

For Biodiesel B100 sample #14048

<u>Cold Soak test</u>	Two test results were excluded because these laboratories probably reported a test result for filter blocking tendency instead of a cold soak test result in seconds. The other five reported test results vary over a large range: 63 – 720 s. Therefore no significant conclusions were drawn. Important aspects such as conditioning of test samples that might have been cooled to temperatures below 20 °C (68 °F) (§ 8.4 - 8.6), the choice and the preparation of the glass and following the correct filtration procedure (§11) are important aspects to attain comparable results.
<u>Filter Blocking Tendency</u>	The five reported test results vary from 3.73 – 10.05. Therefore no significant conclusions were drawn.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant standard and the reproducibility as found for the group of laboratories that participated. The reproducibilities derived from literature standards (in casu the ASTM, EN, ISO and IP standards) and the calculated reproducibilities of the samples (see appendix 1) are compared in the next table.

Parameter	unit	n	average	R (Calc.)	R (lit)
Acid Value (EN14104)	mg KOH/g	42	0.158	0.048	0.060
Acid Number (D664-B)	mg KOH/g	28	0.147	0.061	0.082
Carbon Residue on 10% dist.	%M/M	14	0.17	0.14	0.07
Carbon Residue on 100% FAME	%M/M	7	0.02	0.08	(0.14)
Cold Filter Plugging Point	°C	53	-14.5	2.3	3.3
Cloud Point	°C	53	-4.3	2.1	3.0
Copper Strip Corrosion		52	1	n.a.	n.a.
Density @ 15°C	kg/m ³	63	883.3	0.3	0.5
Flash Point EN spec.	°C	26	174.9	15.2	15.0
Flash Point (PMcc) ISO/ASTM	°C	36	167.1	15.3	14.7
Iodine Value	g I ₂ /100g	41	113.1	4.5	5.0
Kin. Viscosity @ 40°C	mm ² /s	52	4.552	0.040	0.045

Oxidation Stability EN14112	hours	51	8.38	1.41	2.41
Sulphur (ISO20846)	mg/kg	32	2.04	1.34	1.35
Sulphur (D5453)	mg/kg	21	2.06	0.74	1.00
Sulphated Ash	%M/M	14	0.0011	0.0023	(0.0006)
Water	%M/M	60	0.033	0.007	0.013
Methanol	%M/M	34	0.021	0.007	0.008
mono-Glycerides	%M/M	35	0.59	0.12	0.18
di-Glycerides	%M/M	35	0.13	0.05	0.05
tri-Glycerides	%M/M	33	0.10	0.06	0.08
Free Glycerol	%M/M	23	0.0026	0.0060	0.0066
Total Glycerol	%M/M	36	0.181	0.034	0.046
Total Ester Content	%M/M	45	98.51	2.69	4.16
Linolenic Acid Methyl Ester	%M/M	43	9.71	0.68	0.66
Polyunsat. Methyl esters	%M/M	16	0.22	0.43	(0.27)

table 7: comparison of the observed and target reproducibilities of Biodiesel B100 sample #14045

Parameter	unit	n	average	R (Calc.)	R (lit)
Calcium & Magnesium	mg/kg	26	10.9	3.1	2.8
Phosphorus	mg/kg	26	6.7	3.1	1.3
Potassium	mg/kg	10	0.2	0.5	(1.0)
Sodium	mg/kg	28	4.1	2.9	2.4

table 8: comparison of the observed and target reproducibilities of Biodiesel B100 sample #14046

Parameter	unit	n	average	R (Calc.)	R (lit)
Total Contamination	mg/kg	39	27.1	17.4	8.5

table 9: comparison of the observed and target reproducibilities of Biodiesel B100 sample #14047

Parameter	unit	n	average	R (Calc.)	R (lit)
Cold Soak Filter Test	s	4	95.0	63.0	(10.9)
Filter Blocking Tendency		5	6.70	6.50	(2.78)

table 10: comparison of the observed and target reproducibilities of Biodiesel B100 sample #14048

* Values between brackets were below the application range of the respective reference test method, therefore results should be used with care.

Without further statistical calculations it can be concluded that for many tests there is not a good compliance of the group of participating laboratories with the relevant standards. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF APRIL 2014 WITH PREVIOUS PTS

	April 2014	October 2013	April 2013	October 2012
Type of FAME	Rapeseed	Fat of Offal	Rapeseed	Used cooking oil
Number of reporting labs	68	58	75	46
Number of results reported	1093	768	1010	728
Number of statistical outliers	54	44	67	40
Percentage statistical outliers	5.2%	5.7%	6.6%	5.5%

table 11: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective standards. The conclusions are given the following table:

Determination	April 2014	October 2013	April 2013	October 2012
Acid Value (EN14104)	+	--	+	+/-
Acid Number (D664-B)	+	--	+	+
Carbon Residue on 10% dist.	(--)	(--)	--	--
Carbon Residue on 100% FAME	(++)	n.e.	++	n.e.
Cold Filter Plugging Point	+	--	+	-
Cloud Point	+	+	+	+
Density @15°C	+	++	++	++
Flash Point EN spec.	+/-	++	-	++
Flash Point PMcc ISO/ASTM	+/-	+/-	-	+/-
Iodine Value	+	--	+/-	-
Kin. Viscosity @ 40°C	+/-	--	-	-
Oxidation Stability	+	--	+	++
Sulphated Ash	(--)	(-)	(--)	(--)
Sulphur (EN spec.)	+/-	--	-	-
Sulphur (D5453)	+	+	+	-
Water	+	+	++	+
Methanol	+/-	n.e.	--	--
mono-Glycerides	+	(-)	+	-
di-Glycerides	+/-	(-)	+	+/-
tri-Glycerides	+	(+)	+	--
Free Glycerol	+/-	+	+	++
Total Glycerol	+	--	+/-	+/-
Total Ester content	+	--	++	-
Linolenic Acid Methyl Ester	+/-	--	-	++
Polyunsat. Methyl esters	(-)	--	(--)	-
Total Contamination	--	--	--	(--)
Calcium and Magnesium	-	(--)	(--)	(++)
Phosphorus	--	--	--	(--)
Potassium	(++)	(++)	+/-	(++)
Sodium	-	-	+/-	+
Cold Soak Filter Test	(--)	n.e.	(--)	n.e.
Filter Blocking Tendency	(--)	n.e.	(--)	n.e.

table 12 : comparison of group performances against the standard requirements

* Signs between brackets are for assigned values below the application range of the respective reference test method and therefore should be used with due care

The performance of the determinations against the requirements of the respective standards is listed in the above table. The following performance categories were used:



- ++: group performed much better than the standard
- + : group performed better than the standard
- +/-: group performance equals the standard
- : group performed worse than the standard
- : group performed much worse than the standard
- n.e.: not evaluated

APPENDIX 1

Determination of Acid Value conform EN spec. on sample #14045; results in mg KOH/g

lab	method	value	mark	z(targ)	remarks
120	EN14104	0.129		-1.34	
150	EN14104	0.203		2.12	
171	EN14104	0.1863		1.34	
311	EN14104	0.15		-0.36	
312	EN14104	0.141		-0.78	
323	EN14104	0.15		-0.36	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14104	0.155		-0.12	
344	EN14104	0.1261		-1.47	
345	EN14104	0.16		0.11	
360	EN14104	0.161		0.16	
391	EN14104	0.143		-0.68	
398	EN14104	0.148		-0.45	
445	EN14104	0.1374		-0.95	
447		----		----	
494	EN14104	0.136		-1.01	
496	EN14104	0.150		-0.36	
529		----		----	
540	EN14104	0.17		0.58	
554		----		----	
603		----		----	
657	EN14104	0.13		-1.29	
824	EN14104	0.15		-0.36	
862	EN14104	0.166		0.39	
863	EN14104	0.151		-0.31	
1016	EN14104	0.138		-0.92	
1033		----		----	
1059	EN14104	0.17		0.58	
1067	EN14104	0.16		0.11	
1095	EN14104	0.16		0.11	
1131	EN14104	0.186		1.32	
1134	EN14104	0.1516		-0.28	
1161		----		----	
1199		----		----	
1201	EN14104	0.52	R(0.01)	16.91	
1213		----		----	
1227		----		----	
1231		----		----	
1240	EN14104	0.178		0.95	
1268		----		----	
1286		----		----	
1290	EN14104	0.17		0.58	
1292	EN14104	0.167		0.44	
1299	EN14104	0.15		-0.36	
1316		----		----	
1397	EN14104	0.16		0.11	
1402		----		----	
1429	EN14104	0.17		0.58	
1443		----		----	
1457	EN14104	0.150		-0.36	
1459	EN14104	0.18		1.04	
1494		----		----	
1510	EN14104	0.174		0.76	
1528	EN14104	0.183		1.18	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14104	0.151		-0.31	
1654		----		----	
1684		----		----	
1706	EN14104	0.164		0.30	
1721	EN14104	0.152		-0.26	
1739	EN14104	0.149		-0.40	
1744		----		----	
1769		----		----	
1807	EN14104	0.14		-0.82	
1948	EN14104	0.1751		0.81	



normality	OK
n	42
outliers	1
mean (n)	0.1577
st.dev. (n)	0.01710
R(calc.)	0.0479
R(EN14104:03)	0.0600



Determination of Acid Number conform ASTM spec. on sample #14045; results in mg KOH/g

lab	method	value	mark	z(targ)	remarks
120	D664-B	0.119		-0.97	
150	D664-B	0.16		0.43	
171	D664-B	0.0371	R(0.01)	-3.75	
311		----		----	
312		----		----	
323		----		----	
334	D664-B	0.09		-1.95	
335		----		----	
337		----		----	
338		----		----	
343		----		----	
344		----		----	
345		----		----	
360	D664-B	0.149		0.05	
391		----		----	
398	D664-B	0.158		0.36	
445	D664	0.121		-0.90	
447		----		----	
494		----		----	
496		----		----	
529		----		----	
540		----		----	
554		----		----	
603	D664-B	0.132		-0.53	
657	D664-B	0.14		-0.25	
824		----		----	
862	D664-B	0.152		0.15	
863	D664-B	0.137		-0.36	
1016		----		----	
1033	D974	0.15		0.09	
1059		----		----	
1067	D664-B	0.16		0.43	
1095		----		----	
1131	D664-A	0.134		-0.46	
1134		----		----	
1161	D664	0.140		-0.25	
1199		----		----	
1201	D664-B	0.177		1.01	
1213	D664	0.15		0.09	
1227	D974	0.159		0.39	
1231	D664	0.1485		0.04	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292		----		----	
1299	D664-B	0.190		1.45	
1316	D664-B	0.16		0.43	
1397		----		----	
1402	D664-A	0.15		0.09	
1429		----		----	
1443		----		----	
1457		----		----	
1459		----		----	
1494	D664-B	0.14600		-0.05	
1510		----		----	
1528	D664-B	0.201		1.82	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635	D664-B	0.15		0.09	
1643	D664	0.138		-0.32	
1650	D664-B	0.123		-0.83	
1654		----		----	
1684		----		----	
1706		----		----	
1721		----		----	
1739		----		----	
1744	D664-B	0.133		-0.49	
1769	D664-B	0.161		0.46	
1807		----		----	
1948		----		----	

normality	suspect
n	28
outliers	1
mean (n)	0.1474
st.dev. (n)	0.02182
R(calc.)	0.0611
R(D664B:11a)	0.0823





Determination of Carbon Residue on 10% distillation residue on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	ISO10370	0.200		1.27	
150		----		----	
171	ISO10370	0.225		2.20	
311		----		----	
312		----		----	
323	ISO10370	<0.10		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	ISO10370	0.203		1.38	
344		----		----	
345	ISO10370	0.230		2.40	
360	ISO10370	0.130		-1.38	
391		----		----	
398		----		----	
445		----		----	
447		----		----	
494	ISO10370	0.098		-2.60	
496		----		----	
529		----		----	
540	ISO10370	0.110		-2.13	
554		----		----	
603		----		----	
657	ISO10370	0.150		-0.62	
824		----		----	
862		----		----	
863		----		----	
1016	ISO10370	0.564	G(0.01)	15.01	false positive result?
1033		----		----	
1059		----		----	
1067		----		----	
1095		----	W	----	result withdrawn, first reported: 0.67
1131		----		----	
1134		----		----	
1161	ISO10370	0.148		-0.70	
1199		----		----	
1201		----		----	
1213		----		----	
1227		----		----	
1231		----		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292	ISO10370	0.198		1.18	
1299		----		----	
1316		----		----	
1397		----		----	
1402	ISO10370	0.110		-2.13	
1429		----		----	
1443		----		----	
1457	ISO10370	0.218		1.95	
1459		----		----	
1494		----		----	
1510		----		----	
1528		----		----	
1539	ISO10370	0.110		-2.13	
1582		----		----	
1588		----		----	
1634		----		----	
1635	ISO10370	0.009	G(0.05)	-5.94	
1643		----		----	
1650		----		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721		----		----	
1739		----		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948	ISO10370	0.202		1.32	

normality OK
n 14
outliers 2
mean (n) 0.1665
st.dev. (n) 0.04884
R(calc.) 0.1368
R(ISO10370:93) 0.0742

application range: 0.10 – 30.0 %M/M





Determination of Carbon Residue on 100% FAME on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	remarks
120		----		----	
150		----		----	
171	D4530	0.2319	G(0.01)	----	
311		----		----	
312		----		----	
323		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343		----		----	
344		----		----	
345		----		----	
360	D4530	0.02		----	
391		----		----	
398		----		----	
445		----		----	
447		----		----	
494		----		----	
496		----		----	
529		----		----	
540		----		----	
554		----		----	
603		----		----	
657	D4530	0.01		----	
824		----		----	
862	D4530	0.003		----	
863		----		----	
1016		----		----	
1033		----		----	
1059		----		----	
1067		----		----	
1095		----		----	
1131		----		----	
1134		----		----	
1161		----		----	
1199		----		----	
1201		----		----	
1213		----		----	
1227	D4530	0.01		----	
1231	D4530	0.08		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292		----		----	
1299		----		----	
1316		----		----	
1397	D4530	<0.01		----	
1402	D4530	0		----	
1429		----		----	
1443		----		----	
1457	D4530	0.009		----	
1459		----		----	
1494		----		----	
1510		----		----	
1528		----		----	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650		----		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721		----		----	
1739		----		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948		----		----	

normality	unknown
n	7
outliers	1
mean (n)	0.019
st.dev. (n)	0.0277
R(calc.)	0.078
R(D4530:11)	(0.141)



application range: 0.10 – 30.0 %M/M



Determination of Cold Filter Plugging Point on sample #14045; results in °C

lab	method	value	mark	z(targ)	remarks
120	EN116	2	C,R(0.01)	14.17	first reported: -4
150	EN116	-14.0		0.46	
171	EN116	-15		-0.40	
311	EN116	-14		0.46	
312	EN116	-14		0.46	
323	EN116	-15		-0.40	
334	EN116	-14		0.46	
335	EN116	-14		0.46	
337	EN116	-15		-0.40	
338	EN116	-15		-0.40	
343	EN116	-15		-0.40	
344	EN116	-14		0.46	
345	EN116	-15		-0.40	
360	EN116	-15		-0.40	
391	EN116	-15		-0.40	
398	EN116	-14		0.46	
445	IP309	-14		0.46	
447	IP309	-14		0.46	
494	EN116	-15		-0.40	
496	EN116	-14.5		0.03	
529	----	----		----	
540	EN116	-16		-1.25	
554	----	----		----	
603	----	----		----	
657	EN116	-15		-0.40	
824	EN116	-14		0.46	
862	EN116	-14		0.46	
863	IP309	-15		-0.40	
1016	EN116	-15		-0.40	
1033	IP309	-15		-0.40	
1059	EN116	-14		0.46	
1067	EN116	-14		0.46	
1095	EN116	-16		-1.25	
1131	EN116	-13		1.32	
1134	IP309	-14		0.46	
1161	EN116	-12		2.17	
1199	----	----		----	
1201	EN116	-14		0.46	
1213	----	----		----	
1227	EN116	-15		-0.40	
1231	D6371	-15		-0.40	
1240	EN116	-14.0		0.46	
1268	----	----		----	
1286	EN116	-16.0		-1.25	
1290	EN116	-13		1.32	
1292	----	----		----	
1299	EN116	-16		-1.25	
1316	EN116	-14.0		0.46	
1397	EN116	-15		-0.40	
1402	EN116	-15		-0.40	
1429	EN116	-14		0.46	
1443	EN116	-14		0.46	
1457	EN116	-15		-0.40	
1459	EN116	-14		0.46	
1494	----	----		----	
1510	IP309	-16		-1.25	
1528	EN116	-15		-0.40	
1539	EN116	-14		0.46	
1582	----	----		----	
1588	----	----		----	
1634	----	----		----	
1635	EN116	-15		-0.40	
1643	----	----		----	
1650	EN116	-14		0.46	
1654	----	----		----	
1684	----	----		----	
1706	----	----		----	
1721	EN116	-14		0.46	
1739	EN116	-16		-1.25	
1744	----	----		----	
1769	----	----		----	
1807	----	----		----	
1948	EN116	-8	C,R(0.01)	5.60	first reported: -5



normality	OK
n	53
outliers	2
mean (n)	-14.54
st.dev. (n)	0.820
R(calc.)	2.29
R(EN116:97)	3.27



Determination of Cloud Point on sample #14045; results in °C

lab	method	value	mark	z(targ)	remarks
120	D2500	-5		-0.66	
150	D2500/EN23015	-5		-0.66	
171	D2500/EN23015	-7	R(0.05)	-2.53	
311	D2500/EN23015	-4		0.27	
312	EN23015	-5		-0.66	
323	EN23015	-5		-0.66	
334	EN23015	-5		-0.66	
335	D2500/EN23015	-5		-0.66	
337	D2500/EN23015	-4		0.27	
338	EN23015	-4.0		0.27	
343	D2500/EN23015	-5		-0.66	
344	D2500	-3.9		0.37	
345	D5771	-4.8		-0.47	
360	EN23015	-4		0.27	
391	D2500/EN23015	-5		-0.66	
398	EN23015	-4		0.27	
445	EN23015	-1	R(0.01)	3.07	
447	D2500	-4		0.27	
494	EN23015	-5		-0.66	
496	EN23015	-4.0		0.27	
529	D2500	-5		-0.66	
540	D2500	-5		-0.66	
554		----		----	
603	D2500	-2		2.14	
657	D2500	-1	R(0.01)	3.07	
824	D2500	-5		-0.66	
862	D2500	-4		0.27	
863	D2500	-5		-0.66	
1016	D2500/EN23015	-3.4		0.83	
1033	IP219	-5.2		-0.85	
1059	EN23015	-4		0.27	
1067	D2500/EN23015	-5		-0.66	
1095	D2500	-4		0.27	
1131	EN23015	-4		0.27	
1134	IP219	-5		-0.66	
1161	EN23015	-5		-0.66	
1199		----		----	
1201	D2500/EN23015	-3		1.21	
1213	D2500	-3	C	1.21	first reported: 1
1227	D2500	-3.9		0.37	
1231	D2500	-3		1.21	
1240	EN23015	-4.0		0.27	
1268		----		----	
1286		----		----	
1290	D2500	-5.26		-0.90	
1292		----		----	
1299	D2500/EN23015	-3.7		0.55	
1316	EN23015	-4.0		0.27	
1397	D2500/EN23015	-3		1.21	
1402	D2500/EN23015	-4		0.27	
1429		----		----	
1443		----		----	
1457	D2500/EN23015	-4.8		-0.47	
1459	EN23015	-5		-0.66	
1494		----		----	
1510	D2500	-4		0.27	
1528	D2500/EN23015	-5		-0.66	
1539	ISO3015	0	R(0.01)	4.01	
1582		----		----	
1588		----		----	
1634		----		----	
1635	EN23015	-4		0.27	
1643	D2500	-3		1.21	
1650	D5771	-4.0		0.27	
1654		----		----	
1684		----		----	
1706		----		----	
1721	D2500/EN23015	-4.5		-0.19	
1739	EN23015	-5		-0.66	
1744	D2500	-4		0.27	
1769		----		----	
1807		----		----	
1948	D2500/EN23015	-4		0.27	

normality	OK
n	53
outliers	4
mean (n)	-4.29
st.dev. (n)	0.741
R(calc.)	2.07
R(D2500:11)	3.00



Determination of Copper Strip Corrosion 3 hrs/50°C on sample #14045



lab	method	value	mark	z(targ)	remarks
120	D130	1A		----	
150	D130/ISO2160	1A		----	
171	D130/ISO2160	1A		----	
311	D130/ISO2160	1A		----	
312	D130	1A		----	
323	ISO2160	1A		----	
334		----		----	
335	D130/ISO2160	1		----	
337		----		----	
338		----		----	
343	D130/ISO2160	1A		----	
344	D130	1A		----	
345	ISO2160	1A		----	
360	D130	1A		----	
391	D130/ISO2160	1A		----	
398	ISO2160	1A		----	
445	ISO2160	1A		----	
447	D130	1A		----	
494	D130/ISO2160	1A		----	
496	ISO2160	1A		----	
529	D130	1A		----	
540	D130	1A		----	
554		----		----	
603	D130	1A		----	
657	D130	1A		----	
824	D130	1A		----	
862	D130	1A		----	
863	D130	1A		----	
1016	D130/ISO2160	1A		----	
1033	IP154	1B		----	
1059	ISO2160	1A		----	
1067	D130/ISO2160	1A		----	
1095	ISO2160	1A		----	
1131	ISO2160	1A		----	
1134	D130/ISO2160	1A		----	
1161	ISO2160	1		----	
1199		----		----	
1201	D130/ISO2160	1A		----	
1213	D130	1A		----	
1227	D130	1A		----	
1231	D130	1A		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292	D130/ISO2160	1A		----	
1299	D130/ISO2160	1A		----	
1316	D130	1A		----	
1397	D130/ISO2160	1		----	
1402	D130/ISO2160	1A		----	
1429	D130	1A		----	
1443		----		----	
1457	D130/ISO2160	1A		----	
1459		----		----	
1494		----		----	
1510	D130	1A		----	
1528	D130/ISO2160	1A		----	
1539	ISO2160	1A		----	
1582		----		----	
1588		----		----	
1634	D130/ISO2160	1A		----	
1635	ISO2160	1A		----	
1643		----		----	
1650	ISO2160	1A		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721	D130/ISO2160	1		----	
1739	ISO2160	1A		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948	D130/ISO2160	1A		----	

normality	n.a
n	52
outliers	n.a
mean (n)	1(1A)
st.dev. (n)	n.a
R(calc.)	n.a
R(D130:12)	n.a

Determination of Density @ 15°C conform EN spec. on sample #14045; results in kg/m³

lab	method	value	mark	z(targ)	remarks
120	ISO12185	883.3		0.00	
150	ISO12185	883.6		1.68	
171	ISO12185	883.32		0.12	
311	ISO12185	883.4		0.56	
312	ISO12185	883.3		0.00	
323	ISO12185	883.4		0.56	
334	ISO12185	883.0	C	-1.68	first reported:882.5
335	ISO12185	883.2		-0.56	
337	ISO12185	883.4		0.56	
338	ISO12185	883.3		0.00	
343	ISO12185	883.35		0.28	
344	ISO12185	883.4		0.56	
345	ISO12185	883.3		0.00	
360	ISO12185	883.5		1.12	
391	ISO12185	883.1		-1.12	
398	ISO12185	883.3		0.00	
445	ISO12185	883.3		0.00	
447	D4052	883.3		0.00	
494	ISO12185	883.3		0.00	
496	ISO12185	883.31		0.06	
529	D4052	883.4		0.56	
540	ISO12185	883.3		0.00	
554		-----		-----	
603	ISO12185	883.4		0.56	
657	ISO12185	883.0		-1.68	
824	ISO12185	883.3		0.00	
862	ISO12185	883.4		0.56	
863	ISO12185	883.40		0.56	
1016	ISO12185	883.4		0.56	
1033	IP365	883.4		0.56	
1059	ISO12185	883.2		-0.56	
1067	ISO12185	883.3		0.00	
1095	ISO12185	883.2		-0.56	
1131	ISO12185	883.24		-0.33	
1134	ISO12185	883.3		0.00	
1161	ISO12185	883.2		-0.56	
1199		-----		-----	
1201	ISO12185	883.35		0.28	
1213	D4052	883.32		0.12	
1227	D4052	883.4		0.56	
1231	D4052	883.05	C	-1.40	probably unit error, reported 0.88305
1240	ISO12185	883.3		0.00	
1268		-----		-----	
1286	ISO12185	883.319		0.11	
1290	ISO12185	883.3		0.00	
1292	ISO12185	881.5	R(0.01)	-10.08	
1299	ISO12185	883.3		0.00	
1316	ISO12185	883.4		0.56	
1397	ISO12185	883.4		0.56	
1402	ISO12185	883.2		-0.56	
1429	ISO12185	883.4		0.56	
1443	ISO12185	883.40		0.56	
1457	ISO12185	883.3		0.00	
1459	ISO12185	883.30		0.00	
1494	D4052	883.470		0.96	
1510	IP365	883.3		0.00	
1528	ISO12185	883.25		-0.28	
1539	ISO12185	883.3		0.00	
1582		-----		-----	
1588		-----		-----	
1634	ISO12185	883.308		0.05	
1635		-----		-----	
1643	ISO12185	883.0	C	-1.68	first reported: 0.8830
1650	ISO12185	883.37		0.40	
1654		-----		-----	
1684	ISO3675	884.1	R(0.01)	4.48	
1706	ISO12185	883.3		0.00	
1721	ISO12185	883.2		-0.56	
1739	ISO3675	882.97		-1.84	
1744	D2500	883.30		0.00	
1769	D4052	883.14	C	-0.89	first reported:0.88314
1807		-----		-----	
1948	ISO12185	883.4		0.56	



normality	Suspect
n	63
outliers	2
mean (n)	883.299
st.dev. (n)	0.1207
R(calc.)	0.338
R(ISO12185:96)	0.500



Determination of Flash Point conform EN spec. on sample #14045; results in °C

lab	method	value	mark	z(targ)	remarks
120		----		----	
150		----		----	
171	ISO3679	178		0.58	
311	ISO3679	168.5		-1.19	
312	ISO3679	179.0		0.77	
323		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	ISO3679	164.0		-2.03	
344		----		----	
345	ISO3679	169.5		-1.00	
360	ISO3679	179.0		0.77	
391		----		----	
398	ISO3679	169.0		-1.10	
445	IP523	185.0		1.89	
447	IP523	177.8		0.55	
494		----		----	
496	ISO3679	179.9		0.94	
529		----		----	
540		----		----	
554		----		----	
603		----		----	
657		----		----	
824		----		----	
862	ISO3679	177.9		0.57	
863		----		----	
1016	ISO3679	166.0		-1.66	result was reported under flash point (PMcc)
1033		----		----	
1059		----		----	
1067		----		----	
1095		----		----	
1131		----		----	
1134	ISO3679	173.5		-0.26	
1161	ISO3679	180		0.96	
1199		----		----	
1201		----		----	
1213		----		----	
1227		----		----	
1231		----		----	
1240	ISO3679	170.0		-0.91	
1268		----		----	
1286	ISO3679	175.3		0.08	
1290		----		----	
1292	ISO3679	170.1		-0.89	
1299		----		----	
1316	ISO3679	177		0.40	
1397		----		----	
1402	ISO3679	177		0.40	
1429	ISO3679	179		0.77	
1443	ISO3679	181.0		1.14	
1457	ISO3679	175.0		0.02	
1459		----		----	
1494		----		----	
1510		----		----	
1528		----		----	
1539	ISO3679	174.5		-0.07	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650		----		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721	ISO3679	166		-1.66	
1739	ISO3679	180.2		0.99	
1744		----		----	
1769		----		----	
1807		----		----	
1948	ISO3679	174.5		-0.07	



normality	OK
n	26
outliers	0
mean (n)	174.87
st.dev. (n)	5.427
R(calc.)	15.19
R(ISO3679:04)	15.00



Determination of Flash Point (PMcc) conform ISO/ASTM spec. on sample #14045; results in °C

lab	method	value	mark	z(targ)	remarks
120	D93	167		-0.02	
150	D93/ISO2719	>130		----	
171	D93/ISO2719	176.5		1.79	
311		----		----	
312		----		----	
323	ISO2719	171.0		0.74	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	D93/ISO2719	167.5		0.07	
344	D93	160.0		-1.36	
345		----		----	
360	D93	163.5		-0.69	
391	D93/ISO2719	177		1.88	
398	ISO2719	154.0		-2.50	
445	ISO2719	164.0		-0.59	
447		----		----	
494	D93/ISO2719	160.5		-1.26	
496	ISO2719	165.9		-0.23	
529		----		----	
540	D93	171.0		0.74	
554		----		----	
603	D93	163.0		-0.78	
657	D93	172		0.93	
824	ISO2719	166.0		-0.21	
862	D93	170.3		0.61	
863		----		----	
1016		----		----	
1033	IP34	168.0		0.17	
1059	D93/ISO2719	170.5		0.64	
1067		----		----	
1095	ISO2719	167.0		-0.02	
1131		----		----	
1134	ISO2719	165.7		-0.27	
1161		----		----	
1199		----		----	
1201	D93/ISO2719	170		0.55	
1213		----		----	
1227	D93	164		-0.59	
1231	D93	172.0		0.93	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292		----		----	
1299	D93/ISO2719	>120.0		----	
1316	D93/ISO2719	163.0		-0.78	
1397	D93/ISO2719	164.5		-0.50	
1402	D93/ISO2719	161.5		-1.07	
1429		----		----	
1443	ISO2719	164.5		-0.50	
1457	D93/ISO2719	161.0		-1.17	
1459	ISO2719	171.5		0.83	
1494		----		----	
1510	D93	174		1.31	
1528	D93/ISO2719	167.0		-0.02	
1539	ISO2719	164.5		-0.50	
1582		----		----	
1588		----		----	
1634	D93/ISO2719	167.0		-0.02	
1635	ISO2719	182		2.83	
1643	D93	168.35		0.23	
1650	ISO2719	161.0		-1.17	
1654		----		----	
1684		----		----	
1706		----		----	
1721		----		----	
1739		----		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948		----		----	



normality	OK
n	36
outliers	0
mean (n)	167.12
st.dev. (n)	5.474
R(calc.)	15.33
R(D93-C:13e)	14.70



Determination of Iodine Value conform EN spec. on sample #14045; results in g I₂/100g

lab	method	value	mark	z(targ)	remarks
120		----		----	
150		----		----	
171	EN14111	111.11		-1.13	
311	EN14111	114		0.49	
312	EN14111	114.0		0.49	
323	EN14111	114		0.49	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14111	112.5		-0.35	
344	EN14111	116.1		1.67	
345	EN14111	113		-0.07	
360	EN14111	113.6		0.27	
391	EN14111	112		-0.63	
398	EN14111	112.0		-0.63	
445	EN14111	115.6		1.39	
447		----		----	
494	EN14111	115.4		1.27	
496	EN14111	113		-0.07	
529		----		----	
540	EN14111	111		-1.19	
554		----		----	
603		----		----	
657	EN14111	113		-0.07	
824	EN14111	113.7		0.32	
862	EN14111	114.3		0.66	
863		----		----	
1016	EN14111	114.5		0.77	
1033		----		----	
1059	EN14111	113		-0.07	
1067	EN14111	109.9		-1.81	
1095		----		----	
1131	EN14111	114.49		0.77	
1134	EN14111	114.75		0.91	
1161	EN14111	112.7		-0.24	
1199		----		----	
1201	EN14111	131	R(0.01)	10.01	
1213	EN14111	113.3		0.10	
1227	EN14111	109.8		-1.86	
1231	EN14111	92.5	R(0.01)	-11.55	
1240	EN16300	113.0		-0.07	
1268		----		----	
1286		----		----	
1290	EN14111	112.63		-0.28	
1292	EN14111	110.06		-1.72	
1299	EN14111	110		-1.75	
1316		----		----	
1397	EN16300	112.4		-0.41	
1402	EN14111	102	R(0.01)	-6.23	
1429	EN14111	114.9		0.99	
1443		----		----	
1457	EN14111	114.6		0.83	
1459	EN16300	113.4		0.15	
1494	EN14111	113.700		0.32	
1510		----		----	
1528	EN14111	103	C,R(0.01)	-5.67	first reported: 90.9
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14111	110.5		-1.47	
1654		----		----	
1684	EN14111	112.8		-0.18	
1706		----		----	
1721	EN14111	112.8		-0.18	
1739	EN14111	112.6		-0.29	
1744	EN14111	114.56		0.80	
1769		----		----	
1807	EN14111	115		1.05	
1948	EN14111	114.374		0.70	

normality	OK
n	41
outliers	4
mean (n)	113.12
st.dev. (n)	1.614
R(calc.)	4.52
R(EN14111:03)	5.00





Determination of Kinematic Viscosity @ 40°C on sample #14045; results in mm²/s

lab	method	value	mark	z(targ)	remarks
120	D445	4.481		1.82	
150	D445/ISO3104	4.4648		0.81	
171	D445/ISO3104	4.444		-0.49	
311	D445/ISO3104	4.417		-2.18	
312	ISO3104	4.453		0.07	
323	ISO3104	4.434		-1.12	
334	----			----	
335	D445/ISO3104	4.482		1.89	
337	----			----	
338	----			----	
343	D445/ISO3104	4.4392		-0.79	
344	----			----	
345	ISO3104	4.478		1.64	
360	ISO3104	4.4528		0.06	
391	D445/ISO3104	4.450		-0.12	
398	ISO3104	4.4178		-2.13	
445	ISO3104	4.447		-0.31	
447	D445	4.447		-0.31	
494	D445/ISO3104	4.513	C,R(0.05)	3.83	first reported:4.523
496	ISO3104	4.4570		0.32	
529	----			----	
540	ISO3104	4.461		0.57	
554	----			----	
603	D445	4.462		0.63	
657	D445	4.472	C	1.26	first reported:4.555
824	ISO3104	4.448		-0.24	
862	D445	4.4501		-0.11	
863	D445	4.4493		-0.16	
1016	D445/ISO3104	4.4433		-0.54	
1033	IP71	4.471		1.20	
1059	D445/ISO3104	4.433		-1.18	
1067	D445/ISO3104	4.448		-0.24	
1095	ISO3104	4.444		-0.49	
1131	D445	4.4368		-0.94	
1134	ISO3104	4.459		0.45	
1161	ISO3104	4.461		0.57	
1199	----			----	
1201	D445/ISO3104	4.450		-0.12	
1213	D445	4.459		0.45	
1227	D445	4.4508		-0.07	
1231	D445	4.4495		-0.15	
1240	ISO3104	4.265	C,R(0.01)	-11.71	first reported:4.383
1268	----			----	
1286	----			----	
1290	D7042	4.5033	R(0.05)	3.22	
1292	D445/ISO3104	3.79285	R(0.01)	-41.28	
1299	D445/ISO3104	4.439		-0.81	
1316	ISO3104	4.456		0.26	
1397	D445/ISO3104	4.468		1.01	
1402	D445/ISO3104	4.446		-0.37	
1429	D445	4.4542		0.14	
1443	ISO3104	4.4480		-0.24	
1457	D445/ISO3104	4.459		0.45	
1459	D7042	4.485		2.07	
1494	D445	4.42500		-1.68	
1510	D445	4.449		-0.18	
1528	D445/ISO3104	4.45149		-0.02	
1539	ISO3104	4.450		-0.12	
1582	----			----	
1588	----			----	
1634	----			----	
1635	----			----	
1643	D445	4.445		-0.43	
1650	ISO3104	4.4550		0.20	
1654	----			----	
1684	ISO3104	4.529	R(0.01)	4.83	
1706	----			----	
1721	D445/ISO3104	4.468		1.01	
1739	ISO3104	4.4469		-0.31	
1744	D445	4.4391		-0.80	
1769	D445	4.4510		-0.06	
1807	----			----	
1948	D445/ISO3104	4.450		-0.12	

normality OK
 n 52
 outliers 5
 mean (n) 4.4519
 st.dev. (n) 0.01428
 R(calc.) 0.0400
 R(ISO3104:94) 0.0447



Compare R(445:12) = 0.0338



Determination of Oxidation Stability on sample #14045; results in hours

lab	method	value	mark	z(targ)	Remarks
120	EN14112	9		0.72	
150	EN14112	8.37		-0.01	
171	EN14112	8.3		-0.09	
311	EN14112	8.1		-0.33	
312	EN14112	8.46		0.09	
323	EN14112	9.0		0.72	
334	EN14112	8.5		0.14	
335		----		----	
337		----		----	
338		----		----	
343	EN14112	8.9		0.60	
344	EN14112	8.305		-0.09	
345	EN14112	9.5		1.30	
360	EN14112	7.74		-0.74	
391	EN14112	8.3		-0.09	
398	EN14112	9.03		0.76	
445	EN15751	8.76		0.44	
447		----		----	
494	EN14112	8.15		-0.27	
496	EN14112	8.23		-0.17	
529		----		----	
540	EN14112	8.5		0.14	
554		----		----	
603	EN14112	8.23		-0.17	
657	EN14112	8.3		-0.09	
824	EN14112	8.2		-0.21	
862	EN14112	9.10		0.84	
863	EN14112	7.52		-1.00	
1016	EN14112	7.73		-0.76	
1033	EN14112	8.47		0.10	
1059	EN14112	8.6		0.26	
1067	EN14112	8.5		0.14	
1095	EN14112	8.2		-0.21	
1131	EN14112	8.66		0.33	
1134	EN14112	8.32		-0.07	
1161	EN14112	7.5		-1.02	
1199	EN14112	8.42		0.05	
1201	EN14112	8.64		0.30	
1213		----		----	
1227	EN14112	7.96		-0.49	
1231		----		----	
1240	EN15751	9.04		0.77	
1268		----		----	
1286		----		----	
1290	EN14112	8.34		-0.05	
1292	EN14112	8.66		0.33	
1299	EN15751	9.1		0.84	
1316	EN14112	8.08		-0.35	
1397	EN14112	9.1		0.84	
1402	EN14112	8.2		-0.21	
1429	EN14112	8.04		-0.40	
1443	EN14112	6.825		-1.81	
1457	EN14112	8.47		0.10	
1459	EN15751	8.9		0.60	
1494		----		----	
1510	EN14112	7.45		-1.08	
1528	EN14112	8.44		0.07	
1539	EN14112	8.4		0.02	
1582		----		----	
1588		----		----	
1634		----		----	
1635	EN14112	8.45		0.08	
1643		----		----	
1650	EN14112	7.45		-1.08	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14112	8.67		0.34	
1739	EN14112	8.27		-0.13	
1744		----		----	
1769		----		----	
1807		----		----	
1948	EN14112	14.01	C,R(0.01)	6.54	first reported:25.09

normality	suspect
n	51
outliers	1
mean (n)	8.380
st.dev. (n)	0.5049
R(calc.)	1.414
R(EN14112:03)	2.409





Determination of Sulphur (EN spec.) on sample #14045; results in mg/kg

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	ISO20846	0.94		-2.29	
311	ISO20846	<3		----	
312		----		----	
323	ISO20846	<3.0		----	
334		----		----	
335	ISO20846	2.1		0.12	
337	ISO20846	1.22		-1.71	
338	ISO20846	1.3		-1.54	
343	ISO20846	2.7		1.37	
344		----		----	
345	ISO20846	1.26		-1.62	
360	ISO20846	2.0		-0.09	
391	ISO20846	2.0		-0.09	
398	ISO20846	2.56		1.07	
445	IP490	2.22		0.37	
447		----		----	
494	ISO20846	1.8		-0.50	
496	ISO20846	2.01		-0.07	
529		----		----	
540	ISO20846	3.0		1.99	
554		----		----	
603		----		----	
657	ISO20846	1.9		-0.30	
824	ISO20846	2.1		0.12	
862		----		----	
863		----		----	
1016	ISO20846	2.219		0.37	
1033		----		----	
1059	ISO20884	2.0		-0.09	
1067	ISO20846	1.8		-0.50	
1095	ISO20846	2.2		0.33	
1131	ISO20846	2.4		0.74	
1134	ISO20846	2.15		0.22	
1161	ISO20846	2.24		0.41	
1199	ISO20884	<5.0		----	
1201		----		----	
1213		----		----	
1227		----		----	
1231		----		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----	W R(0.01)	----	result withdrawn, first reported: 17.72
1292	in house	5.695		7.58	
1299	ISO20846	1.3		-1.54	
1316		----		----	
1397		----		----	
1402	ISO20846	1.7		-0.71	
1429		----		----	
1443	ISO20846	3.04		2.07	
1457	ISO20846	1.9		-0.30	
1459	in house	<2		----	
1494		----		----	
1510		----		----	
1528	ISO20846	1.88		-0.34	
1539	ISO20846	2.2		0.33	
1582		----		----	
1588		----		----	
1634		----		----	
1635	ISO20846	2.2		0.33	
1643		----		----	
1650	ISO20846	2.34		0.62	
1654		----		----	
1684		----		----	
1706		----		----	
1721	ISO20846	2.3		0.53	
1739		----		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948	ISO20846	2.38		0.70	

normality OK
 n 32
 outliers 1
 mean (n) 2.042
 st.dev. (n) 0.4792
 R(calc.) 1.342
 R(ISO20846:11) 1.349

application range 3 – 500 mg/kg





Determination of Sulphur (ASTM D5453) on sample #14045; results in mg/kg

lab	method	value	mark	z(targ)	remarks
120	D5453	2.1	C	0.13	first reported: 3.1
150	D5453	1.92		-0.38	
171	D5453	0.94	R(0.05)	-3.14	
311		----		----	
312	D5453	1.9		-0.44	
323		----		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343		----		----	
344	D5453	1.973		-0.23	
345		----		----	
360	D5453	2.0		-0.16	
391		----		----	
398		----		----	
445		----		----	
447		----		----	
494	D5453	1.8		-0.72	
496		----		----	
529		----		----	
540	D5453	2.6		1.53	
554		----		----	
603		----		----	
657	D5453	1.9		-0.44	
824	D5453	2.1		0.13	
862	D5453	2.16		0.29	
863	D5453	2.25		0.55	
1016		----		----	
1033		----		----	
1059		----		----	
1067		----		----	
1095	D5453	2.2		0.41	
1131		----		----	
1134		----		----	
1161	D5453	2.20		0.41	
1199		----		----	
1201		----		----	
1213	D5453	1.50		-1.56	
1227	D5453	1.8		-0.72	
1231		----		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292		----		----	
1299		----		----	
1316		----		----	
1397	D5453	1.8		-0.72	
1402	D5453	1.7		-1.00	
1429	D5453	2.46		1.14	
1443		----		----	
1457		----		----	
1459		----		----	
1494		----		----	
1510		----		----	
1528	D5453	2.26		0.58	
1539	D5453	2.2		0.41	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	D5453	2.34		0.80	
1654		----		----	
1684		----		----	
1706		----		----	
1721		----		----	
1739		----		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948		----		----	

normality OK
n 21
outliers 1
mean (n) 2.055
st.dev. (n) 0.2651
R(calc.) 0.742
R(D5453:12) 0.995

Application range: 1- 8000 mg/kg





Determination of Sulphated Ash on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	D874	<0.005		----	
150	D874/ISO3987	<0.001		----	
171	D874/ISO3987	0.000		----	
311		----		----	
312		----		----	
323	ISO3987	<0.005		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	D874/ISO3987	<0.005		----	
344	ISO3987	<0.005		----	
345	ISO3987	<0.005		----	
360	D874	0.0009		----	
391		----		----	
398	ISO3987	<0.005		----	
445	ISO3987	<0.01		----	
447		----		----	
494	D874/ISO3987	0.0002		----	
496	ISO3987	<0.001		----	
529		----		----	
540	ISO3987	<0.005		----	
554		----		----	
603	D874	<0.005		----	
657	D874	<0.005		----	
824		----		----	
862	D874	0.002		----	
863	D874	<0.001		----	
1016	D874	<0.001		----	
1033		----		----	
1059	ISO3987	0.00012		----	
1067		----		----	
1095		----		----	
1131		----		----	
1134	ISO3987	0.00071		----	
1161	ISO3987	0.00266		----	
1199		----		----	
1201	D874/ISO3987	<0.001		----	
1213	D874	<0.001		----	
1227		----		----	
1231	D874	0.00178		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292	D874/ISO3987	0.0015		----	
1299		----		----	
1316		----		----	
1397		----		----	
1402	D874/ISO3987	0	ex	----	result excluded, zero is not a real value
1429		----		----	
1443		----	W	----	result withdrawn, first reported:0.0146
1457	D874/ISO3987	0.00096		----	
1459	ISO3987	0.0010		----	
1494		----		----	
1510		----		----	
1528	D874/ISO3987	0.0019		----	
1539	ISO3987	<0.005		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	ISO3987	0.0005		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721	D874/ISO3987	<0.005		----	
1739	ISO3987	<0.01		----	
1744		----		----	
1769		----		----	
1807	ISO3987	<0.005		----	
1948	D874/ISO3987	0.0005		----	

normality OK
 n 14
 outliers 1 + 1excl.
 mean (n) 0.00105
 st.dev. (n) 0.000804
 R(calc.) 0.00225
 R(D874:13a) (0.00056)



Compare R(ISO3987:10) = 0.00013
 Applicable lower limit of 0.005 %M/M



Determination of Water on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	ISO12937	0.0321		-0.22	
150	ISO12937	0.036		0.65	
171	ISO12937	0.03356	C	0.11	probably unit error, reported 335.6
311	ISO12937	0.031		-0.47	
312	ISO12937	0.0330		-0.02	
323	ISO12937	0.0310	C	-0.47	first reported: 310
334	ISO12937	0.032		-0.24	
335	ISO12937	0.033		-0.02	
337	ISO12937	0.0317		-0.31	
338	ISO12937	0.033144	C	0.01	probably unit error, reported 331.44
343	ISO12937	0.0315		-0.36	
344	ISO12937	0.03366		0.13	
345	ISO12937	0.032		-0.24	
360	ISO12937	0.0338		0.16	
391	ISO12937	0.035	C	0.43	probably unit error, reported 350
398	ISO12937	0.0224	R(0.01)	-2.39	
445	ISO12937	0.0339		0.18	
447	IP438	0.031705	C	-0.31	probably unit error, reported 317.05
494	ISO12937	0.0408	C	1.73	first reported: 408
496	ISO12937	0.0325	C	-0.13	first reported: 352.5
529	E1064	0.033		-0.02	
540	ISO12937	0.031		-0.47	
554		-----		-----	
603		-----		-----	
657	ISO12937	0.034		0.20	
824	ISO12937	0.048	R(0.01)	3.34	
862	ISO12937	0.03125	C	-0.41	probably unit error, reported 312.5
863	ISO12937	0.032		-0.24	
1016	ISO12937	0.0371	C	0.90	first reported: 371.2
1033	IP438	0.033		-0.02	
1059	ISO12937	0.032	C	-0.24	probably unit error, reported 320
1067	ISO12937	0.033		-0.02	
1095	ISO12937	0.032		-0.24	
1131	ISO12937	0.0443	R(0.01)	2.51	
1134	ISO12937	0.0330115	C	-0.02	probably unit error, reported 330.115
1161	ISO12937	0.0324		-0.15	
1199	ISO12937	0.0329		-0.04	
1201	ISO12937	0.035		0.43	
1213	D6304	0.0377		1.03	
1227	D6304	0.0381		1.12	
1231	ISO12937	0.044	R(0.01)	2.44	
1240	ISO12937	0.0314		-0.38	
1268		-----		-----	
1286		-----		-----	
1290	ISO12937	0.031		-0.47	
1292	ISO12937	0.03144		-0.37	
1299	ISO12937	0.0322		-0.20	
1316	E1064	0.0349	C	0.41	first reported: 349
1397	ISO12937	0.029		-0.92	
1402	ISO12937	0.033947	C	0.19	first reported: 339.47
1429	ISO12937	0.0422		2.04	
1443	ISO12937	0.0358		0.61	
1457	ISO12937	0.0334	C	0.07	first reported: 334
1459	ISO12937	0.029		-0.92	
1494	E203	0.031550		-0.34	
1510	IP438	0.03196	C	-0.25	probably unit error, reported 319.6
1528	ISO12937	0.0315		-0.36	
1539	ISO12937	0.032		-0.24	
1582		-----		-----	
1588		-----		-----	
1634	ISO12937	0.032		-0.24	
1635	ISO12937	0.03	C	-0.69	first reported: 326
1643	ISO12937	0.0346		0.34	
1650	ISO12937	0.0308		-0.51	
1654		-----		-----	
1684	ISO12937	0.03846		1.20	
1706		-----		-----	
1721	ISO12937	0.0316		-0.33	
1739	ISO12937	0.0312		-0.42	
1744	E203	0.0327		-0.09	
1769	ISO12937	0.0304		-0.60	
1807		-----		-----	
1948	ISO12937	0.0334	C	0.07	first reported: 333.8



normality	not OK
n	60
outliers	4
mean (n)	0.03309
st.dev. (n)	0.002529
R(calc.)	0.00708
R(ISO12937:00)	0.01251



Determination of Methanol on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	EN14110	<0.01		<-2.94	false negative result?
150		----		----	
171	EN14110-A	0.03		3.46	
311	EN14110-B	0.02		-0.24	
312	EN14110-B	0.0202		-0.16	
323	EN14110-B	0.02		-0.24	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14110-A	0.02		-0.24	
344	EN14110-A	0.02053		-0.04	
345	EN14110-B	0.022	C	0.50	first reported: 0.04
360	EN14110-B	0.0179		-1.02	
391		----		----	
398		----		----	
445		----		----	
447		----		----	
494	EN14110-A	0.02		-0.24	
496	EN14110-B	0.031	R(0.01)	3.83	
529		----		----	
540	EN14110-A	0.02		-0.24	
554		----		----	
603		----		----	
657	EN14110-A	0.019		-0.61	
824	EN14110-B	0.02		-0.24	
862	EN14110-A	0.022		0.50	
863	EN14110-A	0.0193		-0.50	
1016	EN14110-B	0.017		-1.35	
1033		----		----	
1059	EN14110-B	0.02		-0.24	
1067	EN14110-B	0.02		-0.24	
1095		----		----	
1131		----		----	
1134	EN14110-A	0.0205		-0.05	
1161	EN14110-A	0.02		-0.24	
1199		----		----	
1201		----		----	
1213		----		----	
1227		----		----	
1231	EN14110	0.02		-0.24	
1240	EN14110-A	0.020		-0.24	
1268		----		----	
1286	EN14110-B	0.0258		1.91	
1290		----		----	
1292	EN14110-B	0.141	R(0.01)	44.56	
1299	EN14110-B	0.02		-0.24	
1316	EN14110-B	0.036	R(0.01)	5.69	
1397	EN14110	0.023		0.87	
1402		----		----	
1429	EN14110-	0.0208		0.06	
1443	EN14110-B	0.016	R(0.01)	-1.72	
1457	EN14110-B	0.018		-0.98	
1459		----		----	
1494	EN14110-B	0.02687	C	2.31	first reported: 0.0277
1510		----		----	
1528	EN14110-AMod.	0.01	R(0.01)	-3.94	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14110Mod.	0.0194		-0.46	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14110-B	0.020	C	-0.24	first reported: 0.014
1739	EN14110	0.020		-0.24	
1744	EN14110-B	0.019		-0.61	
1769	EN14110-B	0.0196		-0.39	
1807	EN14110	0.02		-0.24	
1948	EN14110	0.021		0.13	

		<u>Method A</u>	<u>Method B</u>
normality	not OK	not OK	not OK
n	34	11	19
outliers	5	0	3
mean (n)	0.0206	0.0210	0.0215
st.dev. (n)	0.00250	0.00307	0.00501
R(calc.)	0.0070	0.0086	0.0076
R(EN14110:03) A+B	0.0076	0.0076	0.0074





Determination of mono-Glycerides on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	EN14105	0.57		-0.33	
311	EN14105	0.58		-0.17	
312	EN14105	0.627		0.58	
323	EN14105	0.55		-0.65	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14105	0.582		-0.14	
344	EN14105	0.5804		-0.16	
345	EN14105	0.58		-0.17	
360		----		----	
391	EN14105	0.56		-0.49	
398	EN14105	0.621		0.49	
445		----		----	
447		----		----	
494	EN14105	0.605		0.23	
496	EN14105	0.5946		0.07	
529		----		----	
540	EN14105	0.56		-0.49	
554		----		----	
603		----		----	
657	EN14105	0.58		-0.17	
824	EN14105	0.58		-0.17	
862	EN14105	0.586		-0.07	
863		----		----	
1016	EN14105	0.546		-0.71	
1033		----		----	
1059	EN14105	0.642		0.82	
1067	EN14105	0.583		-0.12	
1095		----		----	
1131		----		----	
1134	EN14105	0.530		-0.96	
1161	EN14105	0.61		0.31	
1199		----		----	
1201		----		----	
1213	D6584	0.680		1.43	
1227		----		----	
1231		----		----	
1240	EN14105	0.641		0.81	
1268		----		----	
1286		----		----	
1290		----		----	
1292	EN14105	0.029	R(0.01)	-8.95	
1299	EN14105	0.57		-0.33	
1316		----		----	
1397	EN14105	0.61		0.31	
1402		----		----	
1429	EN14105	0.84	R(0.01)	3.98	
1443	EN14105	0.631		0.65	
1457	EN14105	0.606		0.25	
1459		----		----	
1494	D6584	0.5239		-1.06	
1510	EN14105	0.379	R(0.01)	-3.37	
1528	EN14105	0.59	C	-0.01	first reported: 0.424
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14105	0.685		1.51	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14105	0.60		0.15	
1739	EN14105	0.59		-0.01	
1744	D6584	0.5848		-0.09	
1769	D6584	0.490		-1.60	
1807	EN14105	0.54		-0.80	
1948	EN14105	0.658		1.08	

normality OK
 n 35
 outliers 3
 mean (n) 0.5905
 st.dev. (n) 0.04191
 R(calc.) 0.1173
 R(EN14105:11) 0.1756

Compare R(D6584:13) = 0.3318
 Application limit EN14105:11 = 0.10 %M/M





Determination of di-Glycerides on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	EN14105	0.13		0.11	
311	EN14105	0.14		0.64	
312	EN14105	0.103		-1.31	
323	EN14105	0.13		0.11	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14105	0.137		0.48	
344	EN14105	0.1288		0.05	
345	EN14105	0.11		-0.94	
360		----		----	
391	EN14105	0.16		1.70	
398	EN14105	0.141		0.69	
445		----		----	
447		----		----	
494	EN14105	0.15		1.17	
496	EN14105	0.1416		0.73	
529		----		----	
540	EN14105	0.10		-1.47	
554		----		----	
603		----		----	
657	EN14105	0.10		-1.47	
824	EN14105	0.14		0.64	
862	EN14105	0.080		-2.53	
863		----		----	
1016	EN14105	0.092		-1.89	
1033		----		----	
1059	EN14105	0.104		-1.26	
1067	EN14105	0.134		0.32	
1095		----		----	
1131		----		----	
1134	EN14105	0.136		0.43	
1161	EN14105	0.14		0.64	
1199		----		----	
1201		----		----	
1213	D6584	0.110		-0.94	
1227		----		----	
1231		----		----	
1240	EN14105	0.152		1.28	
1268		----		----	
1286		----		----	
1290		----		----	
1292	EN14105	0.006	R(0.01)	-6.44	
1299	EN14105	0.14		0.64	
1316		----		----	
1397	EN14105	0.14		0.64	
1402		----		----	
1429	EN14105	0.27	R(0.01)	7.51	
1443	EN14105	0.130		0.11	
1457	EN14105	0.118		-0.52	
1459		----		----	
1494	D6584	0.1266	C	-0.07	first reported: 0.2502
1510	EN14105	0.053	R(0.05)	-3.95	
1528	EN14105	0.13	C	0.11	first reported: 0.075
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14105	0.151		1.22	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14105	0.13		0.11	
1739	EN14105	0.135		0.38	
1744	D6584	0.1350		0.38	
1769	D6584	0.138		0.54	
1807	EN14105	0.11	C	-0.94	first reported: 0.07
1948	EN14105	0.132		0.22	

normality OK
 n 35
 outliers 3
 mean (n) 0.1279
 st.dev. (n) 0.01843
 R(calc.) 0.0516
 R(EN14105:11) 0.0530

Compare R(D6584:13) = 0.1158
 Application limit EN14105:11 = 0.10 %M/M





Determination of tri-Glyceriden on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	EN14105	<0.10		----	
311	EN14105	0.09		-0.22	
312	EN14105	0.104		0.26	
323	EN14105	0.10		0.12	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14105	0.097		0.02	
344	EN14105	0.0996		0.11	
345	EN14105	0.13		1.14	
360		----		----	
391	EN14105	0.08		-0.56	
398	EN14105	0.089		-0.25	
445		----		----	
447		----		----	
494	EN14105	0.07		-0.89	
496	EN14105	0.0922		-0.14	
529		----		----	
540	EN14105	0.08		-0.56	
554		----		----	
603		----		----	
657	EN14105	0.09		-0.22	
824	EN14105	0.086		-0.35	
862	EN14105	0.095		-0.05	
863		----		----	
1016	EN14105	0.09	C	-0.22	first reported: 0.013
1033		----		----	
1059	EN14105	0.182	R(0.05)	2.90	
1067	EN14105	0.108		0.39	
1095		----		----	
1131		----		----	
1134	EN14105	0.102		0.19	
1161	EN14105	0.065		-1.06	
1199		----		----	
1201		----		----	
1213	D6584	0.136		1.34	
1227		----		----	
1231		----		----	
1240	EN14105	0.101		0.16	
1268		----		----	
1286		----		----	
1290		----		----	
1292	EN14105	0.005	R(0.05)	-3.10	
1299	EN14105	0.10		0.12	
1316		----		----	
1397	EN14105	0.10		0.12	
1402		----		----	
1429	EN14105	0.25	R(0.01)	5.20	
1443	EN14105	0.092		-0.15	
1457	EN14105	0.093		-0.12	
1459		----		----	
1494	D6584	0.1562		2.03	
1510	EN14105	0.005	R(0.05)	-3.10	
1528	EN14105	0.072		-0.83	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14105	0.111		0.49	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14105	0.09		-0.22	
1739	EN14105	0.095		-0.05	
1744	D6584	0.0644		-1.08	
1769	D6584	0.080		-0.56	
1807	EN14105	0.15		1.82	
1948	EN14105	0.073	C	-0.79	first reported: 0.016

normality	not OK
n	33
outliers	4
mean (n)	0.0964
st.dev. (n)	0.02142
R(calc.)	0.0600
R(EN14105:11)	0.0827

Compare R(6584:13) = 0.2165
Application limit EN14105:11 = 0.10 %M/M





Determination of Free Glycerol on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	EN14105	0.001		-0.66	
311	EN14105	0.003		0.19	
312	EN14105	<0.001		----	
323	EN14105	<0.01		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14105	<0.005		----	
344	EN14105	<0.05		----	
345	EN14105	0.001		-0.66	
360		----		----	
391	EN14105	<0.01		----	
398	EN14105	<0.01		----	
445		----		----	
447		----		----	
494	EN14105	<0.001		----	
496	EN14105	0.0012		-0.58	
529		----		----	
540	EN14105	<0.001		----	
554		----		----	
603		----		----	
657	EN14105	0.002		-0.24	
824	EN14105	0.001		-0.66	
862	EN14105	0.007		1.90	
863		----		----	
1016	EN14105	0.00098		-0.67	
1033		----		----	
1059	EN14105	0.005		1.04	
1067	EN14105	<0.010		----	
1095		----		----	
1131		----		----	
1134	EN14105	0.006		1.47	
1161	EN14105	0.008		2.32	
1199		----		----	
1201		----		----	
1213	D6584	0	ex	-1.09	result excluded, zero is not a real value
1227		----		----	
1231		----		----	
1240	EN14105	0.0007		-0.79	
1268		----		----	
1286		----		----	
1290	in house	0.0005		-0.87	
1292	EN14105	0.0637	R(0.01)	26.07	
1299	EN14105	<0.01	C	----	first reported: 0.01
1316		----		----	
1397	EN14105	<0.005		----	
1402		----		----	
1429	EN14105	0.138	R(0.01)	57.75	
1443	EN14105	0.0021		-0.19	
1457	EN14105	0.002		-0.24	
1459		----		----	
1494	D6584	0.0024		-0.06	
1510	EN14105	0.002		-0.24	
1528	EN14105	0.0008		-0.75	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14105	0.0045		0.83	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14105	0.004		0.62	
1739	EN14105	0.0013		-0.53	
1744	D6584	0.0012		-0.58	
1769	D6584	0.001		-0.66	
1807	EN14105	<0.001		----	
1948	EN14105	<0.001		----	

normality	suspect
n	23
outliers	2 +1 excl
mean (n)	0.00255
st.dev. (n)	0.002160
R(calc.)	0.00605
R(EN14105:11)	0.00657

Compare R(6584:13) = 0.0050
Application limit EN14105:11 = 0.001 %M/M





Determination of Total Glycerol on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171	EN14105	0.174		-0.43	
311	EN14105	0.179		-0.12	
312	EN14105	0.1855		0.27	
323	EN14105	0.17		-0.67	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14105	0.181		0.00	
344	EN14105	0.1773		-0.23	
345	EN14105	0.181		0.00	
360		----		----	
391	EN14105	0.175		-0.37	
398	EN14105	0.189		0.49	
445		----		----	
447		----		----	
494	EN14105	0.183		0.12	
496	EN14105	0.1812		0.01	
529		----		----	
540	EN14105	0.17		-0.67	
554		----		----	
603		----		----	
657	EN14105	0.173		-0.49	
824	EN14105	0.178		-0.18	
862	EN14105	0.1715		-0.58	
863		----		----	
1016	EN14105	0.1549		-1.59	
1033		----		----	
1059	EN14105	0.203		1.34	
1067	EN14105	0.179		-0.12	
1095		----		----	
1131		----		----	
1134	EN14105	0.172		-0.55	
1161	EN14105	0.19		0.55	
1199		----		----	
1201		----		----	
1213	D6584	0.207		1.59	
1227		----		----	
1231		----		----	
1240	EN14105	0.1967		0.96	
1268		----		----	
1286		----		----	
1290	in house	0.1858		0.29	
1292	EN14105	0.0741	R(0.01)	-6.52	
1299	EN14105	0.18		-0.06	
1316		----		----	
1397	EN14105	0.185		0.24	
1402		----		----	
1429	EN14105	0.42	R(0.01)	14.57	
1443	EN14105	0.1910		0.61	
1457	EN14105	0.183		0.12	
1459		----		----	
1494	D6584	0.1710	C	-0.61	first reported: 0.1912
1510	EN14105	0.107	R(0.01)	-4.51	
1528	EN14105	0.18	C	-0.06	first reported: 0.127
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14105	0.2127		1.93	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN14105	0.185		0.24	
1739	EN14105	0.181		0.00	
1744	D6584	0.1795		-0.09	
1769	D6584	0.157		-1.46	
1807	EN14105	0.164		-1.04	
1948	EN14105	0.1897		0.53	

normality	suspect
n	36
outliers	3
mean (n)	0.1810
st.dev. (n)	0.01202
R(calc.)	0.0337
R(EN14105:11)	0.0459

Compare R(6584:13) = 0.0833
Application limit EN14105:11 = 0.10 %M/M





Determination of Total Ester content on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171		----		----	
311	EN14103:11	99.9		0.93	
312	EN14103:11	97.19		-0.89	
323	EN14103:11	98.4		-0.08	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14103:11	99.2		0.46	
344	EN14103:11	98.634		0.08	
345	EN14103:03	98.7		0.13	
360	EN14103:11	97.994		-0.35	
391	EN14103:11	98.0		-0.34	
398	EN14103:11	97.74		-0.52	
445		----		----	
447	EN14103	99.4		0.60	
494	EN14103:11	97.9		-0.41	
496	EN14103:11	97.84		-0.45	
529		----		----	
540		----		----	
554		----		----	
603		----		----	
657	EN14103:11	98.9		0.26	
824	EN14103:11	98.5		-0.01	
862	EN14103:11	98.317		-0.13	
863	EN14103:11	99.79		0.86	
1016		----		----	
1033		----		----	
1059	EN14103:11	99.7		0.80	
1067	EN14103:11	99.9		0.93	
1095	EN14103:03	99.6		0.73	
1131	EN14103:11	97.84		-0.45	
1134	EN14103:11	97.86		-0.44	
1161	EN14103:11	98.3		-0.14	
1199		----		----	
1201		----		----	
1213	EN14103:03	97.76		-0.51	
1227		----		----	
1231	EN14103:11	97.2		-0.88	
1240	EN14103:11	99.64		0.76	
1268		----		----	
1286		----		----	
1290		----		----	
1292	EN14103	96.36		-1.45	
1299	EN14103:11	99.0		0.33	
1316	EN14103:11	98.02		-0.33	
1397	EN14103	100.0		1.00	
1402	EN14103:11	99.830		0.89	
1429	EN14103:11	>99		----	
1443	EN14103:11	97.77		-0.50	
1457	EN14103:11	99.01		0.34	
1459	EN14103:11	97.0		-1.02	
1494	EN14103:11	98.3065		-0.14	
1510	EN14103:11	99.81		0.87	
1528	EN14103:11	98.15		-0.24	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14103:11	98.78		0.18	
1654		----		----	
1684	EN14103	97.7		-0.55	
1706	EN14103:09	100.066		1.05	
1721	EN14103:11	97.8		-0.48	
1739	EN14103	97.81		-0.47	
1744	EN14103:03	98.75		0.16	
1769	EN14103:03	98.18		-0.22	
1807	EN14103:11	96.75		-1.19	
1948	EN14103:11	99.73		0.82	

normality	OK
n	45
outliers	0
mean (n)	98.512
st.dev. (n)	0.9616
R(calc.)	2.692
R(EN14103:11)	4.160

Compare R(EN14103:03) = 3.1





Determination of Linolenic Acid Methyl Ester content on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171		----		----	
311	EN14103:11	9.96		1.08	
312	EN14103:11	9.5		-0.87	
323	EN14103:11	9.9		0.82	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN14103:11	9.8		0.40	
344	EN14103:11	9.683		-0.09	
345	EN14103:03	9.8		0.40	
360	EN14103:11	9.437		-1.13	
391	EN14103:11	9.6		-0.44	
398	EN14103:11	9.82		0.48	
445		----		----	
447	EN14103	9.7		-0.02	
494	EN14103:11	9.6		-0.44	
496	EN14103:11	9.80		0.40	
529		----		----	
540		----		----	
554		----		----	
603		----		----	
657	EN14103:11	9.8		0.40	
824	EN14103:11	10.0		1.24	
862	EN14103:11	9.506		-0.84	
863	EN14103:11	9.83		0.53	
1016	EN14103:11	9.53		-0.74	
1033		----		----	
1059	EN14103:11	9.8		0.40	
1067	EN14103:11	9.9		0.82	
1095	EN14103:03	10.0		1.24	
1131	EN14103:11	9.24		-1.96	
1134	EN14103:11	9.34		-1.54	
1161	EN14103:11	9.7		-0.02	
1199		----		----	
1201		----		----	
1213	EN14103:03	9.47		-0.99	
1227		----		----	
1231	EN14103:11	9.335		-1.56	
1240	EN14103:11	9.65		-0.23	
1268		----		----	
1286		----		----	
1290		----		----	
1292	EN14103	10.63	G(0.05)	3.90	
1299	EN14103:11	9.4		-1.29	
1316	EN14103:11	9.68		-0.11	
1397	EN14103	9.9		0.82	
1402	EN14103:11	9.177		-2.23	
1429	EN14103:11	9.8		0.40	
1443	EN14103:11	9.78		0.32	
1457	EN14103:11	9.68		-0.11	
1459	EN14103:11	10.0		1.24	
1494		----		----	
1510	EN14103:11	10.409		2.97	
1528	EN14103:11	9.7		-0.02	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650	EN14103:11	9.77		0.27	
1654		----		----	
1684	EN14103	9.40		-1.29	
1706	EN14103:09	10.004		1.26	
1721	EN14103:11	9.7		-0.02	
1739	EN14103	9.52		-0.78	
1744		----		----	
1769		----		----	
1807	EN14103:11	10.0		1.24	
1948	EN14103:11	9.70		-0.02	

normality OK
 n 43
 outliers 1
 mean (n) 9.705
 st.dev. (n) 0.2419
 R(calc.) 0.677
 R(EN14103:11) 0.664

Compare R(EN14103:03) = 3.038





Determination of Polyunsaturated Methyl Esters content on sample #14045; results in %M/M

lab	method	value	mark	z(targ)	Remarks
120		----		----	
150		----		----	
171		----		----	
311	EN15779	<0.7		----	
312	EN15779	0.59		----	false positive result?
323	EN15779	<0.6		----	
334		----		----	
335		----		----	
337		----		----	
338		----		----	
343	EN15779	<0.30		----	
344	EN15779	0.4073		----	
345	EN15779	0.21		----	
360	EN15779	0.197		----	
391		----		----	
398	EN15779	<0.1		----	
445		----		----	
447		----		----	
494		----		----	
496	EN15779	0.02		----	
529		----		----	
540		----		----	
554		----		----	
603		----		----	
657	EN15779	0.25		----	
824	EN15779	<0.6		----	
862		----		----	
863		----		----	
1016	EN15779	0.11		----	
1033		----		----	
1059	EN15779	0.1		----	
1067	EN15779	0.16		----	
1095		----		----	
1131		----		----	
1134	EN15779	0.06		----	
1161	EN15779	0.23		----	
1199		----		----	
1201		----		----	
1213		----		----	
1227		----		----	
1231		----		----	
1240		----		----	
1268		----		----	
1286		----		----	
1290		----		----	
1292		----		----	
1299	EN15779	0.2		----	
1316		----		----	
1397		----		----	
1402		----		----	
1429		----		----	
1443		----		----	
1457	EN15779	0.284		----	
1459		----		----	
1494		----		----	
1510		----		----	
1528	in house	0.07		----	
1539		----		----	
1582		----		----	
1588		----		----	
1634		----		----	
1635		----		----	
1643		----		----	
1650		----		----	
1654		----		----	
1684		----		----	
1706		----		----	
1721	EN15779	0.15		----	
1739	EN15779	<0.6		----	
1744		----		----	
1769		----		----	
1807		----		----	
1948	EN15779	0.44		----	



normality not OK
n 16
outliers 0
mean (n) 0.217
st.dev. (n) 0.1526
R(calc.) 0.427
R(EN15779:09) (0.270)

Application range 0.3 -3.0 %M/M





Determination of sum of Calcium and Magnesium on sample #14046; results in mg/kg

lab	method	value	mark	z(targ)	Remarks
120	EN14538	8.8	C	-2.09	first reported: 2.5
171	EN14538	10.1		-0.80	
311	EN14538	10.2		-0.70	
312	EN14538	11.8		0.90	
323	EN14538	11.2		0.30	
343	EN14538	11.2		0.30	
345	EN14538	11.6		0.70	
360	EN14538	10.25		-0.65	
391		----		----	
398		----		----	
445	EN14538	12.40		1.49	
447		----		----	
494	EN14538	10.32		-0.58	
496	EN14538	11.84		0.94	
540	EN14538	10.8		-0.10	
603		----		----	
657		----		----	
824	EN14538	11.4		0.50	
1067	EN14538	10.66		-0.24	
1134		----		----	
1161	EN14538	<0.5		<-10.36	false negative result?
1201	D5185mod	9.8		-1.10	
1240	EN14538	12.47		1.56	
1268		----		----	
1290	EN14538	9.162		-1.73	
1299	EN14538	16.7	R(0.01)	5.78	
1316	D7111	10.75		-0.15	
1402	EN14538	11.15		0.25	
1429	EN14538	11.8		0.90	
1457	EN14538	11.78		0.88	
1510	EN14538	8.7		-2.19	
1528	D4628	10.19		-0.71	
1643	D5185	16.631	R(0.01)	5.71	
1650	EN14538	10.45		-0.45	
1721	EN14538	13.0		2.09	
1739	EN14538	11.59		0.69	
normality					
n		OK			<u>Spike</u>
outliers		n			
mean (n)		26			
st.dev. (n)		2			
R(calc.)		10.90			
R(EN14538:06)		1.096			
		3.07			
		2.81			
application range: 1- 10 mg/kg					





Determination of Phosphorus on sample #14046; results in mg/kg

lab	method	value	mark	z(targ)	Remarks
120		----		----	
171	EN14107	5.1		-3.45	
311	EN14107	6.6		-0.26	
312	EN14107	7.0		0.60	
323	EN14107	6.8		0.17	
343	EN14107	5.3		-3.02	
345	EN14107	6.4		-0.68	
360	EN14107	7.07		0.74	
391		----		----	
398		----		----	
445	EN14107	6.70		-0.04	
447		----		----	
494	EN14107	8.55		3.89	
496	EN14107	7.31		1.26	
540	EN14107	5.7		-2.17	
603		----		----	
657		----		----	
824	EN14107	7.2		1.02	
1067	EN14107	4.92		-3.83	
1134		----		----	
1161	EN14107	<1		<-12.18	false negative result?
1201	D5185mod	6.0		-1.53	
1240	EN16294	7.40		1.45	
1268		----		----	
1290	EN14107	5.466		-2.67	
1299	EN14107	7.9		2.51	
1316	D7111	6.55		-0.36	
1402	EN14107	8.190		3.13	
1429	EN14107	7.7		2.09	
1457	EN14107	7.30		1.23	
1510	EN14107	5.3		-3.02	
1528	D3231	5.4	C	-2.81	first reported: 1.81
1643	D5185	9.014		4.88	
1650	EN14107	6.10		-1.32	
1721	EN14107	<1		<-12.18	false negative result?
1739	EN14107	7.76		2.21	
normality		OK		Spike	
n		26			
outliers		0			
mean (n)		6.72		6.01	recovery: <111.8%
st.dev. (n)		1.115			
R(calc.)		3.12			
R(EN14107:03)		1.32			application range: EN14107:03 = 4 -20 mg/kg





Determination of Potassium on sample #14046; results in mg/kg

lab	method	value	mark	z(targ)	remarks
120	EN14109	<0.10	C	----	first reported: 0.83
171	EN14109	<0.1		----	
311	EN14109	<1		----	
312	EN14109	<0.5		----	
323	EN14538	<1.0		----	
343	EN14109	<0.5		----	
345	EN14538	0.17		----	
360	EN14538	0.2		----	
391		----		----	
398		----		----	
445	EN14538	<1		----	
447		----		----	
494	EN14538	0.22		----	
496	EN14538	<1.0		----	
540	EN14538	<0.5		----	
603		----		----	
657	EN14109	<0.5		----	
824	EN14109	<1		----	
1067	EN14538	0.6		----	
1134		----		----	
1161	EN14109	<1		----	
1201	D5185mod	4.3	G(0.01)	----	false positive result?
1240	EN14538	<1.0		----	
1268		----		----	
1290	EN14538	0.0984		----	
1299	EN14109	<1.0	C	----	first reported: 1.0
1316	D7111	<0.50		----	
1402	EN14538	0.14		----	
1429	EN14107	<1.0		----	
1457	EN14538	0.37		----	
1510	EN14538	0.07		----	
1528	EN14109	0.09		----	
1643		----		----	
1650	EN14109	<0.5		----	
1721	EN14109	<1		----	
1739	EN14538	0.08		----	
normality					
not OK					
n					
10					
outliers					
1					
mean (n)					
0.20					
st.dev. (n)					
0.166					
R(calc.)					
0.46					
R(EN14109:03)					
(1.03)					
application range: >0.5 mg/kg					





Determination of Sodium on sample #14046; results in mg/kg

lab	method	value	mark	z(targ)	Remarks
120	EN14108	2.05		-2.34	
171	EN14108	2.2		-2.17	
311	EN14108	4.3		0.25	
312	EN14108	4.2		0.14	
323	EN14538	4.0		-0.09	
343	EN14108	4.8		0.83	
345	EN14538	5.05		1.12	
360	EN14538	3.75		-0.38	
391		----		----	
398		----		----	
445	EN14538	4.50		0.49	
447		----		----	
494	EN14538	3.33		-0.86	
496	EN14538	4.88		0.92	
540	EN14538	4.2		0.14	
603		----		----	
657	EN14108	3.89		-0.22	
824	EN14108	5.0		1.06	
1067	EN14358	3.87		-0.24	
1134		----		----	
1161	EN14108	0.2139	R(0.05)	-4.46	
1201	D5185mod	6.2		2.45	
1240	EN14538	3.56		-0.60	
1268		----		----	
1290	EN14538	3.294		-0.91	
1299	EN14108	6.8		3.14	
1316	D7111	2.62		-1.68	
1402	EN14538	3.92		-0.18	
1429		----		----	
1457	EN14538	3.72		-0.41	
1510	EN14358	3.0		-1.24	
1528	EN14108	4.85		0.89	
1643	D5185	4.038		-0.05	
1650	EN14108	3.92		-0.18	
1721	EN14108	4.0		-0.09	
1739	EN14538	4.28		0.23	
normality		suspect			<u>Spike</u>
n		28			
outliers		1			
mean (n)		4.08		3.74	recovery: < 109 %
st.dev. (n)		1.028			
R(calc.)		2.88			
R(EN14108:03)		2.43			Application range: $\geq 1 \text{ mg/kg}$




Determination of Total Contamination on sample #14047; results in mg/kg

lab	method	value	mark	z(targ)	remarks
120		----		----	
171	EN12662:08	12.1		-4.93	
311	EN12662:14	17.3		-3.22	
312	EN12662:14	25.2		-0.62	
323	EN12662:14	24.2		-0.95	
334	EN12662:08	>30		----	
335	EN12662:14	22.8		-1.41	
337	EN12662:08	19.2		-2.59	
343	EN12662:98	33.4		2.08	
345	EN12662:98	29.6		0.83	
351	EN12662:14	26.3		-0.26	
360	EN12662:08	19.4		-2.53	
391	EN12662:14	26.2		-0.29	
398	EN12662:98	27.17		0.03	
445	EN12662:98	37.51		3.43	
540	EN12662:98	29.0		0.63	
603	EN12662	27.5		0.14	
657	EN12662:08	28.9		0.60	
862		----		----	
1016	EN12662:14	22.97		-1.35	
1033	IP440/EN12662:08	17.5		-3.15	
1059	EN12662	29.4		0.76	
1067	EN12662:14	23.5		-1.18	
1095	EN12662:98	27.0		-0.03	
1134	EN12662:08	22.77		-1.42	
1161	EN12662	30.0		0.96	
1199		----		----	
1201	EN12662:98	21.63		-1.79	
1240	EN12662:14	27.2		0.04	
1290	EN12662:08	28.32		0.41	
1299	EN12662:08	31.6		1.49	
1402	EN12662:99	36.5		3.10	
1429	EN12662:98	39.98		4.24	
1457	EN12662:14	27.53		0.15	
1494	EN12662:08	28.50		0.47	
1528	EN12662:08	31.02		1.30	
1539	EN12662:09	28.3		0.40	
1582		----		----	
1650	EN12662:08	45.02		5.90	
1721	EN12662:09	30.4		1.09	recovery: < 177 %
1739	EN12662:98	26.1		-0.32	
1744	EN12662:08	23.85		-1.06	
1769	EN12662:08	25.82		-0.41	
1948	EN12662	25.34	C Only EN12662:14 data	-0.57	first reported: 10.80
	normality	suspect	not OK		<u>Spike</u>
n		39	10		
outliers		0	0		
mean (n)		27.08	24.32	15.31	
st.dev. (n)		6.207	2.999		
R(calc.)		17.38	8.40		
R(EN12662:14)		8.51	8.51		Application range: 12 – 30 mg/kg



Determination of Cold Soak Filter Test on sample #14048; results in s

lab	method	value	mark	z(targ)	filter blocked after (ml)	remarks
120	D7501	63		----	----	
171	D7501	96		----	----	
311		----		----	----	
343		----		----	----	
398		----		----	50	
494	D7501	108		----	----	
657	D7501	>720		----	294	
1033	IP PM-EA	1.47	ex	----	280	
1067		----		----	----	
1134		----		----	----	
1201	D7501	10.05	ex	----	30	
1287		----		----	----	
1402		----		----	30	
1429		----		----	----	
1539	D7501	113		----	300	
1582		----		----	----	
normality						
n		unknown				
		4				
outliers						
mean (n)		0 + 2 excl				
st.dev. (n)		95.0				
R(calc.)		22.49				
R(D7501:12a)		63.0				
		(10.9)				



Determination of Filter Blocking Tendency on sample #14048

lab	method	value	mark	z(targ)	remarks
120		----		----	
171		----		----	
311		----		----	
343	IP387-B	3.73		----	
398	IP387-B	6.08		----	
494	IP387-B	7.57		----	
657		----		----	
1033	IP387	6.08		----	
1067		----		----	
1134		----		----	
1201		----		----	
1287		----		----	
1402	IP387-B	10.05		----	
1429		----		----	
1539		----		----	
1582		----		----	
normality		unknown			
n		5			
outliers		0			
mean (n)		6.702			
st.dev. (n)		2.3221			
R(calc.)		6.502			
R(IP387B:13)		(2.782)			



APPENDIX 2**Number of participants per country**

1 lab in ARGENTINA
4 labs in BELGIUM
1 lab in BOSNIA and HERZEGOVINA
2 labs in BRAZIL
3 labs in BULGARIA
3 labs in COLOMBIA
1 lab in CROATIA
1 lab in CZECH REPUBLIC
7 labs in FRANCE
4 labs in GERMANY
2 labs in HONG KONG
2 labs in HUNGARY
2 lab in ITALY
1 lab in LITHUANIA
1 lab in MALAYSIA
4 labs in NETHERLANDS
1 lab in NORWAY
1 lab in PHILIPPINES
4 labs in PORTUGAL
1 lab in SERBIA
1 lab in SLOVENIA
7 labs in SPAIN
4 labs in TURKEY
4 labs in UNITED KINGDOM
3 labs in UNITED STATES OF AMERICA

APPENDIX 3

Abbreviations:

C	= final result after checking of first reported suspect result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner outlier test
R(0.05)	= straggler in Rosner outlier test
ex	= excluded from calculations
S	= scope of the reported method is not applicable
U	= reported in different unit
n.a.	= not applicable
n.e.	= not evaluated
SDS	= Safety Data Sheet

Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, April 2014
- 2 ASTM E178-02
- 3 ASTM E1301-03
- 4 ISO13528-05
- 5 ISO 5725-86
- 6 ISO 5725, parts 1-6, 1994
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367/84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee Technical Brief, No4 January 2001
- 14 The Royal Society of Chemistry 2002, Analyst 2002, 127 page1359-1364, P.J. Lowthian and M. Thompson. (see <http://www.rsc.org/suppdata/an/b2/b205600n/>)
- 15 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, *Technometrics*, 25(2), pp. 165-172, (1983).