

Results of Proficiency Test
Jet Fuel A1
March 2017

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: ing. A.S. Noordman – de Neef
Correctors: dr. R.G. Visser, ing. R.J. Starink
Report: iis17J01

June 2017

--- empty page ---

CONTENTS

1	INTRODUCTION	4
2	SET UP	4
2.1	ACCREDITATION	4
2.2	PROTOCOL	4
2.3	CONFIDENTIALITY STATEMENT	5
2.4	SAMPLES	5
2.5	STABILITY OF THE SAMPLES	7
2.6	ANALYSES	7
3	RESULTS	7
3.1	STATISTICS	8
3.2	GRAPHICS	8
3.3	Z-SCORES	9
4	EVALUATION	9
4.1	EVALUATION PER SAMPLE AND PER TEST	10
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	15
4.3	COMPARISON OF THE PROFICIENCY TEST OF MARCH 2017 WITH PREVIOUS PTS	16
Appendices:		
1.	Data, statistical results and graphic results	18
2.	Z-scores for Distillation and Particle Size distribution	66
3.	Equipment used in Particle Size distribution	70
4.	Number of participants per country	71
5.	Abbreviations and literature	72

1 INTRODUCTION

Since 1995, the Institute for Interlaboratory Studies organises every year proficiency tests (PT) for Jet Fuel A1. In the annual proficiency testing program of 2016/2017, it was decided to continue proficiency tests on Jet Fuel A1 and Jet Fuel Particle Size in accordance with the latest applicable version (October 2016) of the "Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS)", sometimes referred to as the "Joint Fuelling System Check List for Jet A-1".

In total 113 laboratories from 54 different countries registered for participation in the interlaboratory study for Jet Fuel A1 and Particle Size Distribution. See appendix 4 for the number of participants per country. For Jet Fuel A1 (main round) 112 participants from 54 countries did register and for Particle Size Distribution 47 participants from 31 countries. In this report, the results of the two proficiency tests are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organiser of this proficiency test. Sample analyses for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC 17025 accredited laboratory. For the main round Jet Fuel A1, it was decided to send two identical samples (both 1L bottles were labelled #17030) for the analyses according to the latest version of "Joint Fuelling System Check List for Jet A-1". For the Particle Size Distribution round, it was decided to send one 0.5L sample (labelled #17031). The participants were requested to report the test results using the indicated units and to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC 17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4). This protocol can be downloaded from the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

2.4.1 JET FUEL A1 (MAIN)

The necessary bulk material, approximately 400 litres, was obtained from a local refinery and homogenised in a mixing vessel. From this batch 290 amber glass bottles of one litre were filled, later closed with inner and outer caps and labelled #17030. The remainder of the batch was stored. The homogeneity of the subsamples #17030 was checked by the determination of Density at 15°C in accordance with ASTM D4052 on ten stratified randomly selected samples.

	Density at 15°C in kg/m ³
Sample #17030-1	796.20
Sample #17030-2	796.20
Sample #17030-3	796.20
Sample #17030-4	796.19
Sample #17030-5	796.20
Sample #17030-6	796.19
Sample #17030-7	796.19
Sample #17030-8	796.20
Sample #17030-9	796.19
Sample #17030-10	796.19

Table 1: homogeneity test of the test results of subsamples #17030

From the above test results, the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table:

	Density at 15°C in kg/m ³
r (observed)	0.01
reference test method	ASTM D4052:16
0.3 x R (ref. test method)	0.15

Table 2: evaluation of repeatability of subsamples #17030

The calculated repeatability is less than 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of all subsamples #17030 was assumed.

2.4.2 JET FUEL PARTICLE SIZE DISTRIBUTION DETERMINATION (PS)

The bulk material for Particle Size Distribution Determination was obtained from a local refinery. Approximately 60 litres bulk material was homogenized. From this material 80 amber glass bottles of 0.5 litres were filled, later closed with inner and outer caps and labelled #17031. Each bottle was spiked with 0.83 grams of Lube oil, enriched with 10 mg/kg Arizona Dust A3. The homogeneity of the subsamples #17031 was checked by the determination of Particle Size Distribution in accordance with IP565 on eight stratified randomly selected samples.

	> 4 µm (c) counts/ml	> 6 µm (c) counts/ml	> 14 µm (c) counts/ml	> 21 µm (c) counts/ml	> 25 µm (c) counts/ml	> 30 µm (c) counts/ml
Sample #17031-1	19927	9217	505	72	20	6
Sample #17031-2	20360	9439	518	74	22	6
Sample #17031-3	20498	9494	509	66	16	4
Sample #17031-4	20273	9421	519	74	21	6
Sample #17031-5	20288	9368	492	66	16	5
Sample #17031-6	20441	9517	540	83	25	7
Sample #17031-7	20453	9494	532	84	22	6
Sample #17031-8	20153	9350	513	73	21	8

Table 3: homogeneity test results of subsamples #17031

From the above test results, the calculated repeatabilities were calculated and compared with the repeatabilities (r) and/or with 0.3 times the reproducibilities (R) of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table:

	> 4 µm (c) counts/ml	> 6 µm (c) counts/ml	> 14 µm (c) counts/ml	> 21 µm (c) counts/ml	> 25 µm (c) counts/ml	> 30 µm (c) counts/ml
r (observed)	527	278	42	19	9	3
reference test method	IP565:13	IP565:13	IP565:13	IP565:13	IP565:13	IP565:13
0.3 x R (ref. test method)	666	593	76	17	6	2
r (ref. test method)	1753	1531	194	41	13	5

Table 4: evaluation of repeatabilities of subsamples #17031

The calculated repeatabilities (r) for the particle sizes > 4 µm (c), > 6 µm (c), > 14 µm (c), > 21 µm (c), > 25 µm (c) and > 30 µm (c) are in agreement with the corresponding target repeatabilities (r) of the reference test method. The calculated repeatabilities (r) for the particle sizes > 4 µm (c), > 6 µm (c) and > 14 µm (c) are also in agreement with 3 times the target reproducibility (R) of the respective reference test method. Therefore, homogeneity of the subsamples of #17031 was assumed.

Depending on their registration to each of the participating laboratories 2 x 1 litre bottle of Jet Fuel A1 labelled #17030 and/or a 0.5 litre bottle of Jet Fuel PS labelled #17031 was/were sent on March 1, 2017. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of Jet Fuel A1, packed in the brown glass bottles was checked. The type of bottle was chosen in accordance with ASTM D4306:15. The material has been found sufficiently stable for the period of the proficiency test.

2.6 ANALYSES

The participants were requested to determine on sample #17030: Total Acidity, Aromatics by FIA, Aromatics by HPLC (in %M/M and %V/V), Colour Saybolt (ASTM D156 and ASTM 6045), Density at 15°C, Distillation (IBP, 10%, 50%, 90% recovered and FBP), Existent Gum (unwashed), Flash Point, Freezing Point, JFTOT, Kinematic Viscosity at -20°C, Mercaptan Sulphur, MSEP, Naphthalenes, Smoke Point, Specific Energy (on Sulphur free basis) and Total Sulphur.

The participants were requested to determine only Particle Size Distribution on sample #17031.

The analyses should be performed according to the "Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS), version October 2016", also referred to as the "Joint Fuelling System Check List" or simply "Check List".

It was explicitly requested to treat the samples as if they were routine samples. Therefore, each laboratory is advised to perform only those analyses that normally are done in daily routine (but the laboratories are allowed to do all analyses). Furthermore, it was requested to report the test results using the indicated units on the report form and not to round the test results more, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical calculations.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalysis). Additional or corrected test results are

used for data analysis and original test results are placed under 'Remarks' in the test result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...>' or '>...>' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the results of the statistical evaluation should be used with due care.

According to ISO 5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. When the uncertainty passed the evaluation, no remarks are made in the report. However, when the uncertainty failed the evaluation it is mentioned in the report and it will have consequences for the evaluation of the test results.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualise the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also a normal Gauss curve was projected over the Kernel Density Graph for reference.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM or IP reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation of this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other targets values were used. In some cases a reproducibility based on former iis proficiency tests could be used.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The $z_{(\text{target})}$ scores are listed in the test result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. The usual interpretation of z-scores is as follows:

	$ z < 1$	good
1 <	$ z < 2$	satisfactory
2 <	$ z < 3$	questionable
3 <	$ z $	unsatisfactory

4 EVALUATION

In this interlaboratory study problems with sample dispatch were encountered due to several reasons. Some laboratories in Afghanistan, Azerbaijan, Egypt, Georgia, Mauritius and Saudi-Arabia received the sample(s) rather late or not at all. Another reason for not reporting test results was brake down of equipment (especialty in the PT on particle size determination). For the main round Jet Fuel A1, five participants reported the test results after the final reporting date and another five participants did not report any test results at all. For the PT on Particle Size Distribution determination one participant reported the test results after the final reporting date and another twelve! participants did not report any test results at all. Not all laboratories were able to report all analyses requested.

Finally 108 participants reported in total 2091 numerical test results. Observed were 63 outlying test results, which is 3.0% of the reported numerical test results. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

4.1 EVALUATION PER SAMPLE AND PER TEST

In this section, the reported test results are discussed per sample and per test. The test methods, which were used by the various laboratories, were taken into account for explaining the observed differences where possible and applicable. These test methods are also in the tables together with the reported test results. The abbreviations, used in these tables, are listed in appendix 5.

In the iis PT reports, ASTM test methods are referred to with a number and if appropriate an indication of sub test method (e.g. D1840-B) and an added designation for the year that the test method was adopted or revised (e.g. D1840-B:07). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D1840-B:07(2013)). In the test results tables of Appendix 1 only the test method number and year of adoption or revision e.g. D1840-B:07 will be used.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care.

Since the Joint Fuelling System Check List for Jet-A1 is continuously updated, the participants are advised to monitor the updates. The latest version at the time of this Round Robin is “DEF STAN 91-091/Issue 9, publication date: October 2016” and ASTM D1655:16c. One must keep in mind that ISO test methods are not mentioned in the “Checklist”.

Jet Fuel A1: sample #17030

Acidity, Total: This determination was problematic at the low level 0.0016 mg KOH/g. Four statistical outliers were observed. Three participants reported a test method (ASTM D664 and ASTM D974) other than ASTM D3242 or IP354 mentioned in the check list for Jet Fuel A1. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirement of ASTM D3242:11. The large variation may (partly) be explained by rounding of the test results to three digits by about 50% of the reporting participants.

Aromatics by FIA: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with ASTM D1319:15.

Aromatics by HPLC: The determination in %M/M was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with ASTM D6379:11.

The determination in %V/V may not be problematic. Two statistical outliers were observed. Regretfully, no precision data for the determination in %V/V is mentioned in ASTM D6379:11. However, the calculated reproducibility after rejection of the statistical outliers was comparable to or smaller than the calculated reproducibility in %V/V of the proficiency tests iis16J01 and iis16J02 of 2016.

Colour Saybolt General:

Over the years the Colour Saybolt determination is problematic. The requirements of both Automated (ASTM D6045) and Manual (ASTM D156) appeared to be very strict in comparison to the reproducibilities observed over the years.

Colour Saybolt: The determination was problematic for the automatic test method ASTM D6045. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirement of ASTM D6045:12. In last year's proficiency test (iis16J02) it was suggested that some of the variation may be caused by the fact that the majority of the laboratories reported to have used a different cell than the suggested 100 mm cell in ASTM D6045. The reported test results of seven participants that used a 100 mm cell is almost in line with the requirement of ASTM D6045:12. The calculated reproducibility of labs that reported to use a 50 mm cell is larger than the calculated reproducibility of labs that reported to use a 100 mm cell.

The manual test method ASTM D156 was also problematic. Five statistical outliers were observed. The calculated reproducibility is not in agreement with the requirement of ASTM D156:15.

Density: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility is in good agreement with the requirement of ASTM D4052:16.

Distillation: This determination was not problematic. In total only one statistical outlier was observed over the five distillation parameters. All calculated reproducibilities, after rejection of the statistical outlier, are in agreement with the automated mode requirements of ASTM D86-A:16a.

When compared to the manual mode requirements of ASTM D86-M:16a only the calculated reproducibilities for 10% rec, 50% rec, 90% rec are in agreement.

Existent Gum: This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in good agreement with ASTM D381:12.

Flash Point: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility is in agreement with IP170:14.

In the Joint Fuelling System Checklist both IP170/IP523/ISO13736 and ASTM D56 or ASTM D3828 are mentioned as test methods. Still several participants (four in total) reported test methods which are not equivalent to IP170, like D93 and ISO2719.

Freezing Point: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirement of ASTM D2386:15e1.

JFTOT: The test results for visual tube rating (VTR) are reported mostly as <1 or 1. No statistical analysis can be done on this parameter as this is a qualitative parameter. The AFQRJOS mentioned a limit for VTR<3 and all of the participants agreed on this.

Interferometric (ITR) and ellipsometric (ETR) tube ratings are new parameters for tube rating listed in the latest AFQRJOS of Oct 2016. In total 7 participants reported a test value for ITR and 3 participants a test value for VTR. The limits for ITR and for ETR are both 85 nm maximum. Therefore it is surprisingly that two participants reported >85 nm for ITR. Further it is observed that the reported values for ITR or ETR are quite different from each other so that no conclusions could be drawn.

For Delta P the majority of the participants reported a value ≤ 1 mmHg. Three participants reported a different value than ≤ 1 mmHg.

Kin. Viscosity at -20°C: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirement of ASTM D445:17a.

Mercaptan Sulphur: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirement of ASTM D3227:16.

MSEP: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirement of ASTM D3948:14.

Naphthalenes: This determination may be problematic, depending on the test method used. The calculated reproducibility after rejection of the three statistical outliers is in full agreement with the requirement of ASTM D1840-B:07(2013), but not with the requirement of ASTM D1840-A:07(2013),.

When the test results from the reported procedures A and B are evaluated separately, only the calculated reproducibility of procedure B test results after rejection of the statistical-outliers is in agreement with the respective requirements of ASTM D1840:07(2013).

Smoke Point: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirement of ASTM D1322:15e1-Manual.

When the test results from the reported Automated and Manual modes are evaluated separately, only the calculated reproducibility of the Manual mode is in agreement with the respective requirements of ASTM D1322:15e1.

One participant reported to use test method IP57. Please note that test method IP57 is not longer part of the Joint Fuelling System Checklist.

Specific Energy: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in full agreement with the requirement of ASTM D3338:09e2 (2014).

Sulphur, Total: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirement of ASTM D5453:16e1.

Jet Fuel PS sample #17031: Particle Size Distribution Determination:

The Joint Fuelling System Check List for Jet-A1 mentions test methods IP564, IP565 and IP577 as the reference test methods to determine the Particle Size Distribution in Jet Fuel A1. Over the last years iis has observed and concluded that these methods are biased and not as interchangeable as suggested by the checklist. Although no equipment suppliers are mentioned in the test methods, the brand of the automatic particle counter (APC) defines the test method. The automatic particle counter in test method IP564 is Parker Hannifin, in test method IP565 it is Stanhope-Seta and in test method IP577 it is Pamas. All Particle Size Distribution test data in counts per ml were collected in one test result table via the Data Entry Tool. For the statistical evaluation iis has separated the test data by test method IP564 or by IP565. The test results of IP577 were placed in the table of IP565 as iis had concluded in an earlier PT that the IP577 test data is most similar to IP565 (see iis16J01). However for the statistical evaluations the test data from IP577 were excluded.

In this PT again the brand of the APC was asked (see appendix 3). All participants mentioned the expected combination of test method and brand of APC, except one participant. This participant had mentioned test method IP564 along with the brand Pamas. It was decided to use the reported APC brand as leading and to place the data in the test results table of IP565.

The Particle Size Distribution Determination could also be reported in scale numbers according to ISO4406 scale. In the PTs of 2016 (iis16J01 and iis16J02) it was concluded that the bias effect observed between test methods IP564 and IP565 as reported in counts per ml is also present in the test results reported in scale number. Therefore, it was decided to separate the test results reported in scale number as well as based on reported test method IP564 or IP565. And the test results in scale number of IP577 were again placed in the IP565 test result tables.

In case a statistical outlier was observed in one of the particle size categories the test result in the corresponding category in counts/ml or ISO 4406 scale was excluded.

IP564: The determination according to IP564 was very problematic. In total fourteen statistical outliers were observed for the six particle size categories and one another test result was excluded. The calculated reproducibilities after rejection of the suspect data are not at all in agreement with the requirements of IP564:13 for almost all of the particle size categories. The test results varied over a large range, especially at the lower particle size categories: $\geq 4\mu\text{m}$ (c), $\geq 6\mu\text{m}$ (c) and $\geq 14\mu\text{m}$ (c).

The determination according to ISO4406 scale numbers may not be problematic. Four statistical outliers were observed for the six particle size categories and eleven other test results were excluded. However, the calculated reproducibilities after rejection of the suspect data for $\geq 4\mu\text{m}$ (c), $\geq 6\mu\text{m}$ (c) and for $\geq 14\mu\text{m}$ (c) categories are in agreement with the indicative requirements of IP564:13 annex C. No precision data were given in IP564:13 annex C for the other three categories.

IP565: The determination according to IP565 was not problematic for the determinations $\geq 6\mu\text{m}$ (c) and $\geq 30\mu\text{m}$ (c) categories, but it was problematic for $\geq 4\mu\text{m}$ (c), $\geq 14\mu\text{m}$ (c), $\geq 21\mu\text{m}$ (c) and $\geq 25\mu\text{m}$ (c). In total ten statistical outliers were observed and fourteen other test results excluded. The calculated reproducibilities, after rejection of the suspect data, are in agreement with the requirements of IP565:13 for $\geq 6\mu\text{m}$ (c) and $\geq 30\mu\text{m}$ (c) particle size categories but not for the other four categories.

As noticed in previous PTs it is observed again that the variation in the test results is much smaller with method IP565 than with IP564. A possible explanation could be that the calibration with Parker-Hannifin (IP564) is based on ISO4406 scale numbers and with Stanhope-Seta based (IP565) on counts/ml.

The determination according to ISO4406 scale numbers may not be problematic. Three statistical outliers were observed for the six particle size categories and fifteen other test results were excluded. However, the calculated reproducibilities after rejection of the suspect data for $\geq 4\mu\text{m}$ (c) and $\geq 6\mu\text{m}$ (c) categories are in agreement with the indicative requirements of IP565:13 annex C but not for $\geq 14\mu\text{m}$ (c) category. No precision data were given in IP565:13 annex C for the other categories.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant reference test method and the reproducibility as found for the group of laboratories that participated. The reproducibilities derived from literature reference test methods (in casu ASTM test methods) and the calculated reproducibilities of samples #17030 and #17031 are compared in the next tables.

Parameter	unit	n	average	2.8 * sd	R (lit.)
Acidity, Total	mg KOH/g	63	0.0016	0.0022	0.0016
Aromatics by FIA	%V/V	52	16.5	1.7	2.7
Aromatics by HPLC	%M/M	18	19.8	1.2	2.0
Aromatics by HPLC	%V/V	21	17.5	1.2	n.a.
Colour Saybolt (automated)		39	18.0	2.1	1.2
Colour Saybolt (manual)		50	17.7	2.9	2
Density at 15°C	kg/m ³	104	796.2	0.3	0.5
Initial Boiling Point	°C	104	147.3	7.2	8.1
Temp at 10% recovered	°C	104	169.0	2.8	3.7
Temp at 50% recovered	°C	104	199.0	2.9	3.0
Temp at 90% recovered	°C	103	244.1	4.0	3.7
Final Boiling Point	°C	104	271.1	5.1	7.1
Existent Gum (unwashed)	mg/100mL	62	0.77	1.35	3.16
Flash Point	°C	99	40.6	2.7	3.2
Freezing Point	°C	84	-53.8	2.0	2.5
JFTOT – Visual Tube Rating		58	≤1	n.a.	n.a.
JFTOT – Delta P	mmHg	57	≤1	n.a.	n.a.
Kinematic Viscosity at -20°C	mm ² /s	67	3.957	0.080	0.075
Mercaptan Sulphur as S	%M/M	59	0.0004	0.0003	0.0003
MSEP	rating	76	91.4	11.1	10.9
Naphthalenes	%V/V	52	0.53	0.06	0.06
Smoke Point	mm	76	24.5	2.8	3.8
Specific Energy (Net)	MJ/kg	56	43.337	0.046	0.046
Sulphur, Total	mg/kg	88	593	72	75

Table 5: comparison of the observed and target reproducibilities of sample #17030

Parameter – IP564	unit	n	average	2.8 * sd	R (lit.)
Particle Size ≥ 4 µm (c)	counts/ml	10	15602	7679	2968
Particle Size ≥ 6 µm (c)	counts/ml	10	5748	4259	1740
Particle Size ≥ 14 µm (c)	counts/ml	10	142	178	75
Particle Size ≥ 21 µm (c)	counts/ml	9	11.0	15.0	13.9
Particle Size ≥ 25 µm (c)	counts/ml	9	4.0	6.0	5.4
Particle Size ≥ 30 µm (c)	counts/ml	9	1.4	2.5	2.3
Particle Size ≥ 4 µm (c)	ISO4406 scale number	9	21.0	0.0	1.0
Particle Size ≥ 6 µm (c)	ISO4406 scale number	9	19.8	1.2	1.4
Particle Size ≥ 14 µm (c)	ISO4406 scale number	9	14.1	2.2	2.2
Particle Size ≥ 21 µm (c)	ISO4406 scale number	7	10.7	2.1	n.a.
Particle Size ≥ 25 µm (c)	ISO4406 scale number	7	9.4	1.5	n.a.
Particle Size ≥ 30 µm (c)	ISO4406 scale number	6	7.7	2.3	n.a.

Table 6: comparison of the observed and target reproducibilities of sample #17031 according to IP564

Parameter – IP565	unit	n	average	2.8 * sd	R (lit.)
Particle Size ≥ 4 µm (c)	counts/ml	18	21977	3924	2388
Particle Size ≥ 6 µm (c)	counts/ml	18	8679	1794	1829
Particle Size ≥ 14 µm (c)	counts/ml	20	328	354	168
Particle Size ≥ 21 µm (c)	counts/ml	17	27.8	42.7	23.8
Particle Size ≥ 25 µm (c)	counts/ml	17	6.9	10.4	8.1
Particle Size ≥ 30 µm (c)	counts/ml	17	1.6	2.5	2.7
Particle Size ≥ 4 µm (c)	ISO4406 scale number	16	21.9	1.0	1.0
Particle Size ≥ 6 µm (c)	ISO4406 scale number	17	20.0	0.0	1.0
Particle Size ≥ 14 µm (c)	ISO4406 scale number	18	15.3	1.9	1.4
Particle Size ≥ 21 µm (c)	ISO4406 scale number	13	11.7	2.4	n.a.
Particle Size ≥ 25 µm (c)	ISO4406 scale number	13	9.7	2.4	n.a.
Particle Size ≥ 30 µm (c)	ISO4406 scale number	13	7.3	2.1	n.a.

Table 7: comparison of the observed and target reproducibilities of sample #17031 according to IP565

Without further statistical calculations, it can be concluded that for many tests there is a good compliance of the group of participants with the relevant reference test methods. The tests that are problematic have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF MARCH 2017 WITH PREVIOUS PTS

	March 2017	September 2016	March 2016	September 2015	March 2015	September 2014
Number of reporting labs	108	137	103	129	102	132
Number of test results reported	2091	2710	1809	2695	1803	2729
Statistical outliers	63	49	40	74	44	62
Percentage outliers	3.0%	1.8%	2.2%	2.7%	2.4%	2.3%

Table 8: Comparison with previous proficiency tests

In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective reference test methods. The conclusions are given the following table:

Parameter	March 2017	September 2016	March 2016	September 2015	March 2015	September 2014
Acidity, Total	-	+/-	-	--	-	--
Aromatics by FIA	+	+	+/-	++	+	++
Aromatics by HPLC	+	+/-	-	+	+/-	-
Colour Saybolt (automated)	-	-	--	-	--	-
Colour Saybolt (manual)	-	--	--	-	-	--
Density at 15°C	+	+	+	++	++	++
Distillation	+	+	+	+	+	+
Existent Gum	++	++	++	++	++	++
Flash Point	+	+	+/-	+	+	+
Freezing Point	+	+/-	+	+	+	+
Kinematic Viscosity at -20°C	+/-	-	+	+/-	-	+/-
Mercaptan Sulphur	+/-	+/-	+/-	+	+/-	++
MSEP	+/-	+/-	+/-	+/-	+	+

Parameter	March 2017	September 2016	March 2016	September 2015	March 2015	September 2014
Naphthalenes	+/-	-	+/-	-	+/-	+
Smoke Point	+	+	+	+/-	+	+
Specific Energy (Net)	+/-	+/-	+/-	-	-	-
Sulphur, Total	+/-	+/-	+/-	++	++	+
BOCLE	n.e.	-	n.e.	-	n.e.	+
Particle Size Distribution IP564	--	+	--	--	--	-
ISO scale numbers IP564	+	-	n.e.	n.e.	n.e.	n.e.
Particle Size Distribution IP565	-	+	-	-	-	--
ISO scale numbers IP565	+	-	n.e.	n.e.	n.e.	n.e.
FAME	n.e.	--	n.e.	-	n.e.	--

Table 9: comparison determinations against the requirements of the reference test methods

The performance of the determinations against the requirements of the respective reference test methods is listed in the above table. The following performance categories were used:



- ++: group performed much better than the reference test method
- + : group performed better than the reference test method
- +/-: group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method

APPENDIX 1**Determination of Acidity, Total on sample #17030; results in mg KOH/g**

lab	method	value	mark	z(targ)	remarks
120	D3242	0.002		0.63	
131		----		----	
132	D3242	0.001		-1.08	
140	D3242	0.001		-1.08	
150	D3242	0.002		0.63	
158		----		----	
159	D3242	0.004		4.04	
169	D3242	0.001		-1.08	
171	D3242	0.003		2.33	
175	D3242	0.0005		-1.93	
177	D3242	0.001		-1.08	
194		----		----	
228		----		----	
230	D3242	0.0011		-0.91	
237	D3242	0.0005434		-1.86	
238		----		----	
253	D3242	0.002	C	0.63	first reported: 0.005
273	D974	0.0062	R(0.01)	7.80	
317		----		----	
323	D3242	0.002		0.63	
333	D3242	0.001		-1.08	
334	D3242	0.002		0.63	
335	D3242	0.000	R(0.01)	-2.79	
336		----		----	
353		----		----	
360	D3242	0.0019		0.46	
372	D3242	0.002		0.63	
391		----		----	
398		----		----	
399		----		----	
445	IP354	0.0606	R(0.01)	100.65	
447	D3242	0.002		0.63	
463	D3242	0.0022		0.97	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631		----		----	
663	D3242	0.0040		4.04	
671		----		----	
824	D3242	0.002	C	0.63	first reported: 0.005
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D3242	0.0011	C	-0.91	first reported: 0.005
998		----		----	
1006		0.0012		-0.74	
1023		<0.07	C	----	first reported: >0.07
1039	D3242	0.001		-1.08	
1049	D3242	0.00148		-0.26	
1059	D3242	0.001		-1.08	
1062	D3242	0.0010		-1.08	
1064	D3242	0.0007		-1.59	
1080		----		----	
1082		----		----	
1097	D3242	0.0023		1.14	
1105	D3242	0.00080		-1.42	
1109	D3242	0.0032		2.68	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161	D664	0.0019	C	0.46	first reported: 0.009
1191	D3242	0.0003		-2.27	
1205		----		----	
1212	D3242	0.0014		-0.40	
1237		----		----	
1284	D3242	0.002		0.63	
1297	D664	<0.05		----	
1299	D3242	0.002		0.63	

lab	method	value	mark	z(targ)	remarks
1318	D3242	0.0019		0.46	
1347	D3242	0.0016		-0.06	
1348	D3242	0.003		2.33	
1372	D3242	0.002219		1.00	
1376	D3242	0.0018		0.29	
1379		----		----	
1397	D3242	0.001		-1.08	
1399	D3242	0.0027495		1.91	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D3242	0.0012		-0.74	
1498		----		----	
1528	D3242	0.0029		2.16	
1531		----		----	
1538		----		----	
1585	D3242	0.0013		-0.57	
1586	D3242	0.002		0.63	
1587	D3242	0.00095	C	-1.16	first reported: 0.0095
1610	IP354	0.005	R(0.01)	5.75	
1631	D3242	0.0010		-1.08	
1634	D3242	0.0011		-0.91	
1710	D3242	0.0022		0.97	
1720	D3242	0.001		-1.08	
1724	D3242	0.001		-1.08	
1740		----		----	
1755	D3242	0.001		-1.08	
1757	D3242	0.0014		-0.40	
1776	D3242	0.0022		0.97	
1810	D3242	0.0016		-0.06	
1811	D3242	0.0019		0.46	
1883		----		----	
2129	D3242	0.0012		-0.74	
2130	IP354	0.001		-1.08	
6049	D3242	0.0014		-0.40	
6075	D3242	0.0016		-0.06	
6101		----		----	
6103	D3242	0.001	C	-1.08	first reported: 0.0047
6108		----		----	

normality not OK
n 63
outliers 4
mean (n) 0.00163
st.dev. (n) 0.000781
R(calc.) 0.00219
R(D3242:11) 0.00164





Determination of Aromatics by FIA (without oxygenate correction) on sample #17030; results in %V/V

lab	method	value	mark	z(targ)	remarks
120	D1319	17.2		0.75	
131		----		----	
132	D1319	16.5		0.04	
140	D1319	17.2		0.75	
150	D1319	16.3		-0.17	
158		----		----	
159	D1319	17.6		1.16	
169	D1319	17.1		0.65	
171	D1319	15.8		-0.68	
175	D1319	17.6		1.16	
177	D1319	15.6		-0.88	
194	D1319	16.44		-0.02	
228		----		----	
230	D1319	16.07		-0.40	
237	D1319	16.3		-0.17	
238		----		----	
253	D1319	16.7		0.24	
273		----		----	
317		----		----	
323	D1319	15.8		-0.68	
333		----		----	
334	D1319	14.1	R(0.05)	-2.41	
335	D1319	16.4		-0.06	
336		----		----	
353		----		----	
360	D1319	16.2		-0.27	
372	D1319	16.9		0.45	
391	D1319	16.6		0.14	
398		----		----	
399		----		----	
445	D1319	16.17		-0.30	
447	D1319	16.4		-0.06	
463	D1319	15.35		-1.14	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631	D1319	14.29	R(0.05)	-2.22	
663	D1319	16.41		-0.05	
671		----		----	
824	D1319	17.0		0.55	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D1319	16.6		0.14	
998		----		----	
1006		----		----	
1023		----		----	
1039		----		----	
1049		----		----	
1059	D1319	15.5		-0.98	
1062	D1319	17.4		0.96	
1064	D1319	16.88		0.43	
1080		----		----	
1082		----		----	
1097	D1319	16.21		-0.26	
1105	D1319	17.47		1.03	
1109	D1319	15.82		-0.66	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161		----		----	
1191		----		----	
1205		----		----	
1212	D1319	16.34		-0.13	
1237		----		----	
1284		----		----	
1297		----		----	
1299	D1319	16.5		0.04	
1318		----		----	

lab	method	value	mark	z(targ)	remarks
1347	D1319	16.96		0.51	
1348	D1319	16.9		0.45	
1372	D1319	16.84		0.38	
1376	D1319	16.29		-0.18	
1379		----		----	
1397		----		----	
1399	D1319	17.1		0.65	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D1319	16.6		0.14	
1498		----		----	
1528	EN15553	15.73		-0.75	
1531		----		----	
1538	D1319	16.4		-0.06	
1585		----		----	
1586	D1319	16.9		0.45	
1587		----		----	
1610	IP156	16.2		-0.27	
1631	D1319	16.16		-0.31	
1634		----		----	
1710	D1319	15.6		-0.88	
1720		----		----	
1724	D1319	16.7		0.24	
1740		----		----	
1755		----		----	
1757		----		----	
1776		----		----	
1810		----		----	
1811	D1319	15.84		-0.64	
1883		----		----	
2129	D1319	16.9		0.45	
2130	IP156	15.0		-1.49	
6049	D1319	16.45		-0.01	
6075	D1319	16.35		-0.12	
6101	D1319	17.11		0.66	
6103	D1319	15.7	C	-0.78	first reported: 13.4
6108		----		----	

normality OK
 n 52
 outliers 2
 mean (n) 16.463
 st.dev. (n) 0.5914
 R(calc.) 1.656
 R(D1319:15) 2.744





Determination of Aromatics by HPLC on sample #17030; results in %M/M

lab	method	value	mark	z(targ)	remarks
120		----		----	
131		----		----	
132		----		----	
140		----		----	
150		----		----	
158		----		----	
159		----		----	
169		----		----	
171	D6379	23.4	D(0.01)	5.02	False positive test results?
175		----		----	
177		----		----	
194		----		----	
228		----		----	
230		----		----	
237		----		----	
238		----		----	
253		----		----	
273		----		----	
317		----		----	
323		----		----	
333		----		----	
334		----		----	
335		----		----	
336		----		----	
353		----		----	
360	D6379	20.50		1.03	
372	D6379	19.6		-0.21	
391		----		----	
398		----		----	
399		----		----	
445		----		----	
447		----		----	
463	D6379	20.00		0.34	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631		----		----	
663		----		----	
671		----		----	
824	IP391	19.2		-0.76	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974		----		----	
998		----		----	
1006		----		----	
1023		----		----	
1039		----		----	
1049	D6379	19.469		-0.39	
1059		----		----	
1062		----		----	
1064		----		----	
1080	D6379	19.47		-0.39	
1082		----		----	
1097		----		----	
1105	D6379	19.65		-0.14	
1109	D6591	20.22		0.64	
1121		----		----	
1126	EN12916	20.2		0.62	
1146		----		----	
1150		----		----	
1155		----		----	
1161	EN12916	19.4		-0.49	
1191		----		----	
1205		----		----	
1212		----		----	
1237		----		----	
1284		----		----	
1297	EN12916	19.87		0.16	
1299	IP436	20.5		1.03	
1318	D6379	19.47		-0.39	

lab	method	value	mark	z(targ)	remarks
1347		----		----	
1348		----		----	
1372		----		----	
1376		----		----	
1379		----		----	
1397		----		----	
1399		----		----	
1412		----		----	
1429		----		----	
1491		----		----	
1496		----		----	
1498		----		----	
1528	D6379	19.99		0.33	
1531		----		----	
1538		----		----	
1585	D6379	19.01		-1.02	
1586		----		----	
1587		----		----	
1610		----		----	
1631		----		----	
1634		----		----	
1710		----		----	
1720		----		----	
1724		----		----	
1740		----		----	
1755		----		----	
1757		----		----	
1776		----		----	
1810	D6379	19.6		-0.21	
1811	D6379	19.7		-0.07	
1883		----		----	
2129	IP391	19.70	C	-0.07	first reported: 17.75
2130		----		----	
6049		----		----	
6075		----		----	
6101		----		----	
6103		----		----	
6108		----		----	

normality OK
 n 18
 outliers 1
 mean (n) 19.753
 st.dev. (n) 0.4166
 R(calc.) 1.167
 R(D6379:11) 2.033





Determination of Aromatics by HPLC on sample #17030; results in %V/V

lab	method	value	mark	z(targ)	remarks
120		----		----	
131		----		----	
132		----		----	
140		----		----	
150		----		----	
158		----		----	
159		----		----	
169		----		----	
171	D6379	21.2	R(0.01)	----	
175		----		----	
177		----		----	
194		----		----	
228		----		----	
230		----		----	
237		----		----	
238		----		----	
253		----		----	
273		----		----	
317		----		----	
323		----		----	
333	D6379	17.6		----	
334		----		----	
335		----		----	
336		----		----	
353		----		----	
360		----		----	
372	D6379	17.6		----	
391		----		----	
398		----		----	
399		----		----	
445		----		----	
447		----		----	
463	D6379	17.92		----	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631		----		----	
663		----		----	
671		----		----	
824	IP391	17.1		----	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974		----		----	
998		----		----	
1006		----		----	
1023		----		----	
1039	D6379	15.2	R(0.01)	----	
1049	D6379	17.67186		----	
1059		----		----	
1062		----		----	
1064		----		----	
1080		----		----	
1082	D6379	17.4		----	
1097		----		----	
1105	D6379	17.68		----	
1109		----		----	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161	EN12916	17.0		----	
1191	D6379	17.23		----	
1205		----		----	
1212		----		----	
1237		----		----	
1284	D6379	18.1		----	
1297	EN12916	17.38		----	
1299	IP436	17.9		----	
1318	D6379	17.55		----	

lab	method	value	mark	z(targ)	remarks
1347		----		----	
1348		----		----	
1372		----		----	
1376		----		----	
1379		----		----	
1397	D6379	16.697	C	----	first reported: 20.663
1399	IP436	17.3		----	
1412		----		----	
1429		----		----	
1491		----		----	
1496		----		----	
1498		----		----	
1528	D6379	17.93		----	
1531		----		----	
1538		----		----	
1585	D6379	17.23		----	
1586		----		----	
1587		----		----	
1610		----		----	
1631		----		----	
1634		----		----	
1710		----		----	
1720	D6379	18.3		----	
1724	D6379	17.4		----	
1740		----		----	
1755		----		----	
1757		----		----	
1776	D6379	17.0		----	
1810		----		----	
1811		----		----	
1883		----		----	
2129	IP391	16.62	C	----	first reported: 14.97
2130		----		----	
6049		----		----	
6075		----		----	
6101		----		----	
6103		----		----	
6108		----		----	

normality OK
 n 21
 outliers 2
 mean (n) 17.458
 st.dev. (n) 0.4397
 R(calc.) 1.231
 R(lit.) unknown
 compare R(iis16J01) 1.443
 compare R(iis16J02) 2.333





Determination of Colour Saybolt (Automated) on sample #17030; cell size in mm;

lab	method	cell	value	mark	z(targ)	remarks
120	D6045	50	18		0.09	
131		----	----		----	
132		----	----		----	
140		----	----		----	
150	D6045	----	17		-2.17	
158		----	----		----	
159	D6045	100	19		2.35	
169	D6045	50	16.4		-3.52	
171	D6045	----	22	R(0.01)	9.12	
175		----	----		----	
177		----	----		----	
194		----	----		----	
228		----	----		----	
230		----	----		----	
237		----	----		----	
238		----	----		----	
253		----	----		----	
273		----	----		----	
317		----	----		----	
323	D6045	50	17		-2.17	
333	D6045	----	18		0.09	
334		----	----		----	
335		----	----		----	
336		----	----		----	
353	D6045	50	17		-2.17	
360		----	----		----	
372	D6045	50	19		2.35	
391		----	----		----	
398		----	----		----	
399		----	----		----	
445	D6045	50	17		-2.17	
447	D6045	100	19		2.35	
463		----	----		----	
468		----	----		----	
473	D6045	----	18		0.09	
603		----	----		----	
604		----	----		----	
608		----	----		----	
631	D6045	50	18		0.09	
663		----	----		----	
671		----	----		----	
824	D6045	----	18		0.09	
825		----	----		----	
904		----	----		----	
962		----	----		----	
963		----	----		----	
974		----	----		----	
998		----	----		----	
1006		----	----		----	
1023	D6045	50	18		0.09	
1039		----	----		----	
1049	D6045	50	18		0.09	
1059		----	----		----	
1062	D6045	50	18	C	0.09	first reported: 28 and cell 100
1064	D6045	50	18.0		0.09	
1080		----	----		----	
1082	D6045	----	18		0.09	
1097		----	----		----	
1105	D6045	50	18		0.09	
1109	D6045	100	18		0.09	
1121		----	----		----	
1126		----	----		----	
1146		----	----		----	
1150		----	----		----	
1155		----	----		----	
1161	D6045	----	17		-2.17	
1191	D6045	100	19		2.35	
1205		----	----		----	
1212		----	----		----	
1237		----	----		----	
1284	D6045	50	19		2.35	
1297		----	----		----	
1299	D6045	50	18		0.09	
1318	D6045	100	18		0.09	

lab	method	cell	value	mark	z(targ)	remarks
1347		----	----		----	
1348		----	----		----	
1372	D6045	100	18		0.09	
1376	D6045	----	19		2.35	
1379		----	----		----	
1397		----	----		----	
1399	D6045	----	20	R(0.05)	4.61	
1412		----	----		----	
1429	D6045	50	19		2.35	
1491		----	----		----	
1496		----	----		----	
1498		----	----		----	
1528	D6045	50	17		-2.17	
1531	D6045	----	17		-2.17	
1538		----	----		----	
1585		----	----		----	
1586	D6045	100	18		0.09	
1587	D6045	50	18		0.09	
1610		----	----		----	
1631	D6045	----	18		0.09	
1634		----	----		----	
1710		----	----		----	
1720	D6045	50	19.0		2.35	
1724		----	----		----	
1740		----	----		----	
1755		----	----		----	
1757		----	----		----	
1776		----	----		----	
1810	D6045	50	17		-2.17	
1811	D6045	50	19		2.35	
1883		----	----		----	
2129	D6045	50	18.0		0.09	
2130	D6045	50	18		0.09	
6049		----	----		----	
6075		----	----		----	
6101		----	----		----	
6103	D6045	----	17		-2.17	
6108		----	----		----	

		<u>Test results 50mm Cell only</u>	<u>Test results 100mm Cell only</u>
normality	OK	OK	OK
n	39	22	7
outliers	2	0	0
mean (n)	17.96	17.93	18.43
st.dev. (n)	0.734	0.770	0.535
R(calc.)	2.06	2.16	1.50
R(D6045:12)	1.24	1.24	1.24





Determination of Colour Saybolt (Manual) on sample #17030;

lab	method	value	mark	z(targ)	remarks
120	D156	18		0.48	
131		----		----	
132	D156	18		0.48	
140	D156	17		-0.92	
150		----		----	
158		----		----	
159		----		----	
169		----		----	
171	D156	19		1.88	
175		----		----	
177	D156	18		0.48	
194		----		----	
228		----		----	
230	D156	16		-2.32	
237	D156	19		1.88	
238	D156	20		3.28	
253	D156	19		1.88	
273	D156	18		0.48	
317	D156	17		-0.92	
323	D156	17		-0.92	
333		----		----	
334	D156	16		-2.32	
335	D156	17		-0.92	
336	D156	18		0.48	
353		----		----	
360	D156	18		0.48	
372	D156	18		0.48	
391	D156	16		-2.32	
398	D156	15		-3.72	
399	D156	17		-0.92	
445		----		----	
447	D156	18		0.48	
463		----		----	
468	D156	17	C	-0.92	first reported: 9
473		----		----	
603		----		----	
604	D156	18		0.48	
608	D156	13	R(0.05)	-6.52	
631		----		----	
663	D156	18		0.48	
671	D156	18		0.48	
824		----		----	
825	D156	18		0.48	
904	D156	18		0.48	
962		----		----	
963		----		----	
974	D156	19		1.88	
998	D156	13	R(0.05)	-6.52	
1006		----		----	
1023		----		----	
1039	D156	18		0.48	
1049		----		----	
1059	D156	19		1.88	
1062		----		----	
1064	D156	18.0		0.48	
1080	D156	19		1.88	
1082		----		----	
1097	NF M 07-003	18		0.48	
1105		----		----	
1109	D156	18		0.48	
1121	D156	15		-3.72	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161		----		----	
1191		----		----	
1205		----		----	
1212	D156	18		0.48	
1237		----		----	
1284		----		----	
1297		----		----	
1299		----		----	
1318		----		----	

lab	method	value	mark	z(targ)	remarks
1347	D156	17		-0.92	
1348	D156	19		1.88	
1372		----		----	
1376		----		----	
1379		----		----	
1397		----		----	
1399		----		----	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D156	21	R(0.05)	4.68	
1498		----		----	
1528	D156	18		0.48	
1531		----		----	
1538	D156	18		0.48	
1585	D156	17		-0.92	
1586		----		----	
1587		----		----	
1610	D156	26	R(0.01)	11.68	
1631		----		----	
1634	D156	17		-0.92	
1710	D156	18		0.48	
1720		----		----	
1724	D156	17		-0.92	
1740		----		----	
1755	D156	18		0.48	
1757		----		----	
1776	D156	16		-2.32	
1810		----		----	
1811		----		----	
1883	D156	17		-0.92	
2129	D156	18		0.48	
2130		----		----	
6049	D156	18		0.48	
6075	D156	17		-0.92	
6101	D156	22	C,R(0.05)	6.08	first reported: 24
6103		----		----	
6108		----		----	

normality OK
 n 50
 outliers 5
 mean (n) 17.66
 st.dev. (n) 1.022
 R(calc.) 2.86
 R(D156:15) 2





Determination of Density at 15°C on sample #17030; results in kg/m³

lab	method	value	mark	z(targ)	remarks
120	D4052	796.1		-0.55	
131		----		----	
132	D4052	796.2		0.01	
140	D4052	796.2		0.01	
150	D4052	796.2		0.01	
158		----		----	
159	D4052	796.3		0.57	
169	D4052	796.2		0.01	
171	D4052	796.1		-0.55	
175	D4052	796.2		0.01	
177	D4052	796.0		-1.11	
194	D4052	796.1		-0.55	
228	D4052	796.3		0.57	
230	D4052	796.20		0.01	
237	D4052	796.27		0.40	
238	D4052	795.6	R(0.01)	-3.35	
253	D4052	796.0		-1.11	
273	D4052	796.2		0.01	
317	D4052	796.2		0.01	
323	D4052	796.0		-1.11	
333	D4052	796.2		0.01	
334	D4052	796.3		0.57	
335	D4052	796.3		0.57	
336	D4052	796.3		0.57	
353	IP365	796.1		-0.55	
360	D4052	796.2		0.01	
372	D4052	796.2		0.01	
391	D4052	796.2		0.01	
398	D4052	796.2		0.01	
399	D4052	796.3		0.57	
445	D4052	796.2		0.01	
447	D4052	796.2		0.01	
463	D4052	796.28		0.45	
468	D4052	796.1		-0.55	
473	D4052	796.0		-1.11	
603	D4052	796.5		1.69	
604	D4052	796.23		0.17	
608	D4052	796.3		0.57	
631	D4052	796.17		-0.16	
663	D4052	796.195		-0.02	
671	D4052	796.2		0.01	
824	D4052	796.2		0.01	
825	D4052	796.2		0.01	
904	D4052	796.3		0.57	
962		----		----	
963		----		----	
974	D4052	796.2		0.01	
998	D4052	796.3		0.57	
1006	D4052	796.4		1.13	
1023	D4052	796.2	C	0.01	first reported: 0.7962 kg/m ³
1039	ISO12185	796.4		1.13	
1049	D4052	796.04		-0.89	
1059	D4052	796.2		0.01	
1062	D4052	796.2	C	0.01	first reported: 736.2
1064	D4052	796.10		-0.55	
1080	D4052	796.1		-0.55	
1082	D4052	796.1		-0.55	
1097	ISO12185	796.24		0.23	
1105	D4052	796.05		-0.83	
1109	D4052	796.2		0.01	
1121	D4052	796.3		0.57	
1126	D4052	796.17		-0.16	
1146	D4052	796.17		-0.16	
1150	ISO12185	796.19		-0.05	
1155	ISO3675	796.4		1.13	
1161	ISO12185	796.20		0.01	
1191	D4052	796.2		0.01	
1205	ISO12185	796.21		0.06	
1212	D4052	796.2		0.01	
1237	ISO12185	796.0		-1.11	
1284	D4052	796.27		0.40	
1297	D4052	796.21	C	0.06	first reported: 792.1
1299	D4052	796.3		0.57	
1318	D4052	796.18	C	-0.11	reported 796.18 kg/L

lab	method	value	mark	z(targ)	remarks
1347	D4052	796.26		0.34	
1348	D4052	796.2		0.01	
1372	D4052	796.74	R(0.01)	3.03	
1376	D4052	796.2		0.01	
1379		----		----	
1397	D4052	796.2		0.01	
1399	D4052	796.1	C	-0.55	first reported: 797.288
1412	D4052	796.2		0.01	
1429	D4052	796.2		0.01	
1491	D4052	796.19		-0.05	
1496	D1298	796.3		0.57	
1498	D4052	796.2		0.01	
1528	D4052	796.16		-0.22	
1531	D4052	796.4		1.13	
1538	D1298	796.3		0.57	
1585	D4052	796.20		0.01	
1586	D4052	796.3		0.57	
1587	D4052	796.16		-0.22	
1610	IP365	796.2		0.01	
1631	D4052	796.1		-0.55	
1634	D4052	796.208		0.05	
1710	D4052	796.2		0.01	
1720	D4052	796.3	C	0.57	first reported: 791.1
1724	D4052	796.2		0.01	
1740	D4052	796.2		0.01	
1755	D4052	796.12		-0.44	
1757	D7777	796.1		-0.55	
1776	D4052	796.2		0.01	
1810	D4052	796.1		-0.55	
1811	D4052	796.2		0.01	
1883	D1298	795.9		-1.67	
2129	D4052	796.2		0.01	
2130	IP365	796.2		0.01	
6049	D4052	796.2		0.01	
6075	ISO12185	796.05		-0.83	
6101	D1298	796.2	C	0.01	first reported: 795.3
6103	ISO12185	796.45		1.41	
6108		----		----	

normality suspect
n 104
outliers 2
mean (n) 796.199
st.dev. (n) 0.0975
R(calc.) 0.273
R(D4052:16) 0.50



Determination of Distillation ASTM D86 on sample #17030; results in °C






lab	method	IBP	mark	10% rec	mark	50% rec	mark	90% rec	mark	FBP	mark
120	D86-A	147.6		167.4		198.0		243.9		269.3	
131	D86-A	149.4		169.0		199.3		245.2		269.5	
132	D86-A	149.6		169.6		199.0		244.2		269.4	
140	D86-A	144.0		168.5		199.0		243.1		269.4	
150	D86-A	147.0		168.9		198.8		245.0		269.7	
158		----		----		----		----		----	
159	D86-A	152.4		169.6		199.9		245.9		273.5	
169	D86-A	147.7		169.4		198.7		243.4		269.5	
171	D86-A	149.2		168.9		196.6		242.7		268.3	
175	D86-A	146.4		169.0		199.6		246.8		272.7	
177	D86-A	141.9		168.5		199.9		245.5		273.5	
194	D86-A	144.7		168.5		198.4		242.6		271.3	
228	D86	145.0		167.0		196.0		240.0	R(0.05)	270.0	
230	D86-A	149.0		169.0		197.8		245.1		270.1	
237	D86-M	149.0		168.0		198.0		243.5		270.0	
238	D86-M	146.0		168.0		196.0		240.5		269.5	
253	D86-M	144.0		168.5		198.0		242.0		270.5	
273	D86-A	149.8		168.2		199.4		244.3		273.6	
317	D86-A	143.3		168.3		199.1		243.8		269.2	
323	D86-A	147.8		168.5		198.9		243.7		269.0	
333	D86-A	149.9		169.3		198.9		243.3		272.5	
334	D86-A	150.5		168.1		198.1		246.7		272.9	
335	D86-A	147.3		168.9		199.0		246.4		269.8	
336	D86-A	150.1		169.5		199.6		244.7		270.9	
353	IP123-A	150.5		169.3		199.6		246.3		273.7	
360	D86-A	148.1		170.4		200.1		246.0		270.7	
372	D86-A	147.6		169.5		199.1		244.7		271.0	
391	D86-A	148.0		168.4		199.7		244.6		273.4	
398	D86-A	152.0		170.0		199.3		244.0		270.8	
399	D86-A	148.9		170.1		199.6		243.2		271.0	
445	IP123-A	143.3		167.7		198.6		244.4		270.7	
447	D86-A	140.7		167.7		198.0		245.2		270.1	
463	D86-A	150.6		169.8		199.0		243.5		273.2	
468		----		----		----		----		----	
473		148.4		168.8		198.9		244.2		272.3	
603	D86-A	151.3		170.4		200.2		244.7		273.0	
604	D86-A	143.8		166.8		196.8		241.8		267.8	
608	D86-M	149.0		170.0		201.0		245.0		272.0	
631	D86-A	148.4		168.7		198.0		245.6		269.4	
663	D86-A	143.15		169.25		199.80		246.20		272.50	
671		----		----		----		----		----	
824	D86-A	149.9		169.8		199.5		244.8		274.0	
825	D86-A	148.3		169.4		199.2		244.3		271.0	
904	D86-A	146.0		166.8		197.1		241.9		268.8	
962		----		----		----		----		----	
963		----		----		----		----		----	
974	D86-A	147.1		167.4		197.3		242.0		269.9	
998	D86-M	147.5		169.5		198.0		243.5		274.0	
1006	D86-A	151.3		170.7		199.9	C	244.9	C	272.4	C
1023	D86-A	150.6		170.9		199.9		244.3		273.6	
1039	D2887	146.7		168.7		200.3		243.4		272.9	
1049	D86-A	149.3		169.8		199.3		244.0		273.0	
1059	D86-A	145.6		169.0		198.2		242.7		270.1	
1062	D86-A	145.2		169.7		200.0		244.4		270.5	
1064	D86-A	148.9		169.9		200.4		246.7		273.3	
1080		----		----		----		----		----	
1082	D86-A	146.7		170.5		199.8		244.8		272.2	
1097	ISO3405-A	149.9		168.9		199.4		245.0		271.7	
1105	D86-A	146.9		168.6		198.5		240.5		269.3	
1109	D86-A	146.4		168.0		198.2		243.5		270.8	
1121	D86-M	142.5		166.5		197.0		241.5		268.0	
1126		146.6		170.0		202.1		244.8		273.6	
1146	ISO3405-A	145.9		169.0		199.2		244.4		270.7	
1150	ISO3405-A	150.0		170.4		199.1		243.4		277.2	
1155	ISO3405-A	148.8		169.1		198.2		242.6		270.2	
1161	ISO3405-A	143.6		169.4		199.9		244.1		271.3	
1191	D86-A	143.8		170.7		199.5		243.1		271.5	
1205	D86-A	151.2		170.4		200.4		244.7		272.1	
1212	D86-A	147.1		168.5		199.2		245.8		272.5	
1237	ISO3405-M	152.4		168.4		198.3		243.3		271.2	
1284	D86-A	147.7		168.4		197.9		243.5		272.1	
1297	D86-A	146.0		168.8		198.3		241.7		269.8	
1299	D86-A	147.8		169.1		199.3		245.1		272.8	
1318	D86-A	144.6		169.0		198.8		243.4		270.4	

lab	method	IBP	mark	10% rec	mark	50% rec	mark	90% rec	mark	FBP	mark
1347	D86-M	150.8		169.8		199.28		242.8		271.75	
1348	D86-A	146.1		170		198		242.5		269	
1372	D86-M	146		167		198		243		271	
1376	D86-A	148.2		169.4		198.9		243.7		272.4	
1379		----		----		----		----		----	
1397	D86-A	145.6		168.8		198.5		242.2		270.2	
1399	D86-A	145.0		167.5		198.1		245.3		268.9	
1412	D86-M	149.5		168.5		198.0		242.5		266.0	
1429	D86-A	144.8		167.6		197.6		242.8		268.1	
1491	D86-A	148.3		169.6		199.6		245.6		273.5	
1496	D86-A	147.6		171.6		200.6		246.7		272.1	
1498	D86-A	145.8		169.1		199.9		246.3		273.8	
1528	D86-A	150.7		169.9		200.0		245.6		272.4	
1531	D86-A	147.5		168.5		200.0		245.4		273.1	
1538	D86-A	146.9		168.0		197.7		241.5		269.3	
1585	D86-A	147.5		169.5		199.3		242.0		272.4	
1586	D86-A	146.7		169.1		198.8		246.2		271.4	
1587	D86-A	146.1		169.3		199.5		243.2		270.6	
1610	D86-A	146.1		168.2		198.3		244.2		269.8	
1631	D86-A	149.7		169.8		199.5		244.4		273.3	
1634	D86-A	145.2		169.8		199.1		245.5		271.2	
1710	D86-A	148.9		170.1		199.7		245.3		271.6	
1720	D86-A	141.2		169.0		199.7		243.4		273.9	
1724	D86-A	146.4		168.8		199.4		244		271.3	
1740	D86-A	150.1		170.2		199.1		243.9		269.7	
1755	D86-A	145.5		169.0		199.0		245.5		269.3	
1757	D86-A	146.6		168.8		198.9		246.0		268.8	
1776	D86-A	141.4		168.4		198.9		245.1		270.4	
1810	D86-A	146.1		167.6		197.0		242.0		267.8	
1811	D86-A	145.9		167.8		197.4		242.6		269.1	
1883	D86-M	150		166		198		243		270	
2129	D86-A	144.9		169.0		198.4		243.7		269.9	
2130	IP123-A	150.2		170.3		199.7		244.8		272.0	
6049	D86-A	149.6		168.5		198.0		244.3		269.9	
6075	D86-A	145.2		167.7		198.8		244.1		273.6	
6101	D86-A	149.0		168.9		199.3		244.2		271.8	
6103	ISO3405-A	147.65		169.1		198.95		245.5		270.45	
6108		----		----		----		----		----	
	normality	OK		OK		OK		OK		OK	
	n	104		104		104		103		104	
	outliers	0		0		0		1		0	
	mean (n)	147.33		168.95		198.89		244.10		271.12	
	st.dev. (n)	2.559		1.015		1.024		1.416		1.824	
	R(calc.)	7.16		2.84		2.87		3.97		5.11	
	R(D86:16a-A)	8.10		3.72		3.0		3.66		7.1	
Comp.	R(D86:16a-M)	4.81		3.17		3.05		3.72		4.23	

-A is mode automated

-M is mode manual

Lab 1006 first reported: 278.6; 279.0; 281.7 respectively for T at 50% rec. T at 90% rec. and FBP.





-- Empty page --

Determination of Existent Gum (unwashed) on sample #17030; results in mg/100mL

lab	method	value	mark	z(targ)	remarks
120	IP540	1.0		0.21	
131		----		----	
132	D381	3	R(0.01)	1.98	
140	D381	1		0.21	
150	D381	2		1.09	
158		----		----	
159	D381	1.0		0.21	
169	D381	0.2		-0.50	
171	D381	<1		----	
175	D381	<1		----	
177	D381	2		1.09	
194	D381	<1		----	
228		----		----	
230	IP540	0.6		-0.15	
237	D381	1.0		0.21	
238		----		----	
253	IP540	<1		----	
273	D381	<1		----	
317	D381	<1		----	
323	D381	2.5	R(0.05)	1.54	
333		----		----	
334	D381	2.0		1.09	
335	D381	<0.5		----	
336		----		----	
353	IP540	0.2		-0.50	
360	D381	0.8		0.03	
372	IP540	1		0.21	
391		----		----	
398		----		----	
399	D381	1.6		0.74	
445	IP540	<0.5		----	
447	IP540	1		0.21	
463	IP540	0.6		-0.15	
468	IP540	1.2		0.38	
473	D381	0.034		-0.65	
603	D381	0.4		-0.33	
604		----		----	
608	D381	<0.1		----	
631	IP540	0.5		-0.24	
663	D381	1.10		0.30	
671	D381	0.6		-0.15	
824	D381	0.4		-0.33	
825	D381	<1		----	
904		----		----	
962		----		----	
963		----		----	
974	IP540	1		0.21	
998	D381	0.9		0.12	
1006	D381	<1.0		----	
1023		----		----	
1039	IP540	<1		----	
1049	D381	0.2		-0.50	
1059	D381Mod.	1		0.21	
1062	D381	0		-0.68	
1064	D381	0.0		-0.68	
1080	ISO6246	1		0.21	
1082	IP540	0.4		-0.33	
1097	IP540	0.3		-0.41	
1105	D381	0.6		-0.15	
1109	IP540	0.6		-0.15	
1121		----		----	
1126		----		----	
1146		----		----	
1150	ISO6246	0.8		0.03	
1155	ISO6246	0.5		-0.24	
1161		----		----	
1191	IP540	0.94		0.15	
1205		----		----	
1212	D381	0.4		-0.33	
1237		----		----	
1284	D381	0.6		-0.15	
1297		----		----	
1299	D381	0.4		-0.33	
1318	IP540	0.1		-0.59	

lab	method	value	mark	z(targ)	remarks
1347	D381	< 1		----	
1348	D381	< 1		----	
1372	IP540	1.0		0.21	
1376	D381	1.2		0.38	
1379		----		----	
1397		----		----	
1399		----		----	
1412	D381	0.85		0.07	
1429		----		----	
1491	D381	0.6		-0.15	
1496	D381	0.5		-0.24	
1498		----		----	
1528	IP540	0.4		-0.33	
1531		----		----	
1538	D381	1		0.21	
1585		----		----	
1586	D381	3	R(0.01)	1.98	
1587	IP540	0.6		-0.15	
1610	IP540	< 1		----	
1631	IP540	1.0		0.21	
1634	D381	0.15		-0.55	
1710	IP540	1.0		0.21	
1720		----		----	
1724	D381	1.2		0.38	
1740		----		----	
1755	D381	0.69		-0.07	
1757		----		----	
1776	IP540	2		1.09	
1810	D381	0.4		-0.33	
1811	D381	0.4		-0.33	
1883	D381	0.6		-0.15	
2129	D381	0.4		-0.33	
2130	IP540	1		0.21	
6049	IP540	3.3	R(0.01)	2.25	
6075	ISO6246	1		0.21	
6101	D381	1.2		0.38	
6103	ISO6246	0.35		-0.37	
6108		----		----	

normality OK
 n 62
 outliers 4
 mean (n) 0.766
 st.dev. (n) 0.4805
 R(calc.) 1.345
 R(D381:12) 3.155





Determination of Flash Point on sample #17030; results in °C

lab	method	value	mark	z(targ)	remarks
120		----		----	
131	D56	42.0		1.22	
132	D56	41.0		0.35	
140	D56	41.0		0.35	
150	D56	42.0		1.22	
158		----		----	
159	D56	39	C	-1.40	first reported: 45
169	D56	40.8		0.17	
171	D56	40.5		-0.09	
175	D56	42.0		1.22	
177	D56	41.5		0.79	
194		----		----	
228	IP523	40.0		-0.53	
230	IP170	40.0		-0.53	
237	IP170	39.0		-1.40	
238	IP170	41.0		0.35	
253	IP170	40.5		-0.09	
273	IP170	40.5		-0.09	
317	IP170	40.5		-0.09	
323	IP170	41.0		0.35	
333	IP170	40.0		-0.53	
334	IP170	40.0		-0.53	
335	IP170	40.5		-0.09	
336		----		----	
353	IP170	39.875		-0.63	
360	D56	40.0		-0.53	
372	IP170	40.5		-0.09	
391	IP170	40		-0.53	
398	IP170	41.0		0.35	
399	IP170	40		-0.53	
445	IP170	40.2		-0.35	
447	IP170	41.0		0.35	
463	IP170	40.43		-0.15	
468	IP170	39.0		-1.40	
473	IP170	40.5		-0.09	
603	IP170	41.5		0.79	
604	IP170	42.0		1.22	
608	IP170	40.0		-0.53	
631	D56	42.5		1.66	
663	D56	39.3		-1.14	
671	IP170	40		-0.53	
824	IP170	39.5		-0.96	
825	IP170	40.0		-0.53	
904		----		----	
962		----		----	
963		----		----	
974	IP170	40.0		-0.53	
998	IP170	41.5		0.79	
1006	D56	40.0		-0.53	
1023	ISO13736	40.5		-0.09	
1039	IP170	41.0		0.35	
1049	ISO13736	40.5		-0.09	
1059	IP170	42.0		1.22	
1062	IP170	39.5		-0.96	
1064	IP170	41.0		0.35	
1080		----		----	
1082	IP170	41.5		0.79	
1097	ISO13736	40.9		0.26	
1105	IP170	38.9		-1.49	
1109	IP170	41.0		0.35	
1121	IP170	40.5		-0.09	
1126	ISO2719	42.5		1.66	
1146	IP170	41.0		0.35	
1150	D56	40.6		0.00	
1155	ISO13736	40.5		-0.09	
1161		----		----	
1191	IP170	42.0		1.22	
1205	D93	42.0		1.22	
1212	IP170	40.5		-0.09	
1237		----		----	
1284	IP170	39.5		-0.96	
1297	D56	41		0.35	
1299	IP170	41.5		0.79	
1318	IP170	40.0		-0.53	

lab	method	value	mark	z(target)	remarks
1347	IP170	41.0		0.35	
1348	IP170	41.0		0.35	
1372	IP170	40.0		-0.53	
1376	D56	41.5		0.79	
1379		----		----	
1397	D56	39.1		-1.31	
1399	IP170	43.5		2.54	
1412	D93A	41.0		0.35	
1429	D56	39.0		-1.40	
1491	IP170	40.0		-0.53	
1496	IP170	39.0		-1.40	
1498	D56	41.5		0.79	
1528	D56	41.0		0.35	
1531	D93	43		2.10	
1538	D56	41.0		0.35	
1585	IP170	40.5		-0.09	
1586	IP170	40.5		-0.09	
1587	IP170	39.5		-0.96	
1610	IP170	41.5		0.79	
1631	IP170	40.5		-0.09	
1634	IP170	39.7		-0.79	
1710	D56	40.5		-0.09	
1720	D3828	38.0		-2.28	
1724	IP170	40.5		-0.09	
1740	IP170	40.5		-0.09	
1755	D56	40.8		0.17	
1757	D56	40.9		0.26	
1776	IP170	41.0		0.35	
1810	D56	45.5	R(0.01)	4.29	
1811	D56	40.7		0.09	
1883	D3828	41		0.35	
2129	IP170	39.5		-0.96	
2130	IP170	41.0		0.35	
6049	IP170	41.0		0.35	
6075	IP170	38.5		-1.84	
6101	IP170	40.7		0.09	
6103	ISO13736	41.0		0.35	
6108		----		----	

normality OK
n 99
outliers 1
mean (n) 40.600
st.dev. (n) 0.9540
R(calc.) 2.671
R(IP170:14) 3.2





Determination of Freezing Point on sample #17030; results in °C

lab	method	value	mark	z(targ)	remarks
120	D2386	-54.5		-0.83	
131		----		----	
132	D2386	-54.5		-0.83	
140	D5972	-53.9		-0.16	
150	D7153	-54.0		-0.27	
158		----		----	
159	D2386	-54.5		-0.83	
169	D2386	-55.0		-1.39	
171	D2386	-55.0		-1.39	
175	D2386	-53		0.85	
177	D2386	-55.0		-1.39	
194		----		----	
228		----		----	
230	D2386	-53.30		0.51	
237	D2386	-52.0		1.97	
238		----		----	
253	D2386	-54.5		-0.83	
273		----		----	
317	D5972	-53.8		-0.05	
323	D2386	-54.0		-0.27	
333	IP529	-53.5		0.29	
334	IP529	-53.3		0.51	
335	IP529	-53.7		0.07	
336		----		----	
353	IP16	-53.57		0.21	
360	D2386	-53.7		0.07	
372	D2386	-53.5		0.29	
391	D2386	-53.5		0.29	
398	D2386	-52.0		1.97	
399	D7153	-54.1		-0.38	
445	IP529	-53.5		0.29	
447	IP16	-53.0		0.85	
463	D2386	-54.5		-0.83	
468	D2386	-53.0		0.85	
473	D2386	-53.0		0.85	
603	D2386	-53.0		0.85	
604		----		----	
608	D2386	-55.0		-1.39	
631	D2386	-56.5	R(0.05)	-3.07	
663	D2386	-57	R(0.01)	-3.63	
671	D2386	-53.5		0.29	
824	D2386	-53.0		0.85	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D2386	-55.5		-1.95	
998	D2386	-53.5		0.29	
1006		----		----	
1023	D7153	-53.4	C	0.40	first reported: 53.4
1039	IP529	-53.6		0.18	
1049	D7153	-53.8		-0.05	
1059	D2386	-52.5		1.41	
1062	D2386	-53.6		0.18	
1064	D7153	-53.80		-0.05	
1080		----		----	
1082	IP529	-53.9		-0.16	
1097	IP529	-53.0		0.85	
1105	D7153	-53.7		0.07	
1109	D5972	-54.8		-1.17	
1121	D2386	-53.5	C	0.29	first reported: -58.5
1126		----		----	
1146		----		----	
1150	D2386	-53		0.85	
1155	ISO3013	-55.5		-1.95	
1161		----		----	
1191	IP529	-54.2		-0.49	
1205		----		----	
1212	D2386	-54.5		-0.83	
1237		----		----	
1284	D7153	-53.8		-0.05	
1297	D5972	-54.1		-0.38	
1299	D2386	-54.5		-0.83	
1318	D7153	-53.2		0.63	

lab	method	value	mark	z(target)	remarks
1347	D2386	-54.0		-0.27	
1348	D2386	-53.0		0.85	
1372	D7153	-53.7		0.07	
1376	D7153	-54.0		-0.27	
1379		----		----	
1397	D7153	-53.9		-0.16	
1399		----		----	
1412		----		----	
1429		----		----	
1491	D2386	-54.0		-0.27	
1496	D5972	-52.9		0.96	
1498		----		----	
1528	D7153	-53.7		0.07	
1531		----		----	
1538	D5972	-54.6		-0.94	
1585	D2386	-53.75		0.01	
1586	D2386	-53.0		0.85	
1587	IP529	-53.7		0.07	
1610	D5972	-54.4		-0.72	
1631	D7153	-53.7		0.07	
1634	D2386	-54.0		-0.27	
1710	D7153	-53.6		0.18	
1720	D7153	-54.1		-0.38	
1724	D5972	-53.7		0.07	
1740	D2386	-54.5		-0.83	
1755	D2386	-53.40		0.40	
1757	D2386	-53.0	C	0.85	first reported: 53.0
1776	IP529	-53.2		0.63	
1810	D5972	-54.7		-1.05	
1811	D2386	-53.5		0.29	
1883	D2386	-55		-1.39	
2129	D2386	-53.0		0.85	
2130	IP16	-53.1		0.74	
6049	D2386	-53		0.85	
6075	IP529	-53.0		0.85	
6101	D7153	-53.8		-0.05	
6103	D7153	-54.1		-0.38	
6108		----		----	

normality OK
n 84
outliers 2
mean (n) -53.760
st.dev. (n) 0.7055
R(calc.) 1.975
R(D2386:15e1) 2.5



Determination of JFTOT at 260°C; Visual Tube Rating (VTR), Interferiometric Tube Rating (ITR) in nm, Elipsometric Tube Rating (ETR) in nm, Delta P in mmHg, Heater tube control temp. in °C on sample #17030

lab	method	VTR	mark	method	ITR	mark	method	ETR	mark	method	Delta P	mark	Temp
120	D3241 - A1	1			----			----		D3241	0		260
131		----			----			----			----		----
132	D3241 - A1	<1			----			----		D3241	0		260
140	D3241 - A1	1		D3241 - A2	337.6			----		D3241	1		260
150	D3241 - A1	1		D3241 - A2	18			----		D3241	0		260
158		----			----			----			----		----
159	D3241 - A1	1		D3241 - A2	13.8			----		D3241	0.5		260
169	D3241 - A1	1			----		D3241 - A3	0.1		D3241	0		260
171	D3241 - A1	<1			----			----		D3241	0		260
175	D3241 - A1	<1			----			----		D3241	0		260
177	D3241 - A1	1		D3241 - A2	10.7		D3241 - A3	0.003150		D3241	0		260
194	D3241 - A1	0			----			----		D3241	1		260
228		----			----			----			----		----
230	D3241	<1			----			----		D3241	0.5		260
237	D3241 - A1	1			----			----		D3241	0		260
238		----			----			----			----		----
253	D3241 - A1	<1			----			----		D3241	0		260
273		----			----			----			----		----
317		----			----			----			----		----
323	D3241 - A1	<1			----			----		D3241	<1		260
333		----		D3241 - A2	8			----		D3241	0		260
334	D3241 - A1	0			----			----		D3241	0		260
335	D3241 - A1	<1		D3241 - A2	93.5			----		D3241	0		260
336		----			----			----			----		----
353		----			----			----			----		----
360	D3241 - A1	0			----			----		D3241	1		260
372	D3241 - A1	1			----			----		D3241	0		260
391	D3241 - A1	0			----			----		D3241	0		260
398	D3241 - A1	1			----			----		D3241	3.6	f+, R(1)	260
399		----			----			----			----		----
445	IP323 - B	1			----			----		IP323	<1		260
447		----			----			----			----		----
463		----			----			----			----		----
468		----			----			----			----		----
473		----			----			----			----		----
603		----			----			----			----		----
604		----			----			----			----		----
608		----			----			----			----		----
631	D3241 - A1	<1			----			----		D3241	<1.0		260
663		----			----			----			----		----
671	D3241 - A1	1			----			----		D3241	0		260
824	D3241 - A1	1			----			----		D3241	0		260
825		----			----			----			----		----
904		----			----			----			----		----
962		----			----			----			----		----
963		----			----			----			----		----
974	D3241 - A1	<1			----			----		D3241	0		260
998		----			----			----			----		----
1006		----			----			----			----		----
1023		----			----			----			----		----
1039	D3241 - A1	<1			----			----		D3241	0.2		260
1049	D3241 - A1	1			----			----		D3241	0.0		260
1059		----			----			----			----		----
1062	D3241 - A1	0			----			----		D3241	0		260
1064	D3241 - A1	1			----			----		D3241	0.0		260
1080		----			----			----			----		----
1082	D3241 - A1	<1			----			----		D3241	0		260
1097	D3241 - A1	1			----			----		D3241	0.0		260
1105	D3241 - A1	0			----			----		D3241	0		260
1109	D3241 - A1	1			----			----		D3241	0.0		260
1121	D3241 - A1	<2			----			----		D3241	1		260
1126		----			----			----			----		----
1146	D3241 - A1	1			----			----		D3241	0.1		260
1150		----			----			----			----		----
1155		----			----			----			----		----
1161		----			----			----			----		----
1191	D3241 - A1	1			----			----		D3241	0		260
1205		----			----			----			----		----
1212	D3241 - A1	<1			----			----		D3241	<2		260
1237	D3241 - A1	<1			----			----		D3241	0.0		260
1284		----			----			----			----		----
1297		----			----			----			----		----

lab	method	VTR	mark	method	ITR	mark	method	ETR	mark	method	Delta P	mark	Temp
1299	D3241 - A1	1			----			----		D3241	0		260
1318	D3241 - A1	<1			----			----		D3241	0.0		260
1347	D3241 - A1	1			----			----		D3241	0		260
1348	D3241 - A1	1			----			----		D3241	0		260
1372	D3241	1			----			----		D3241	2	R(0.01)	260
1376	D3241 - A1	1			----			----		D3241	0.0		260
1379		----			----			----			----		----
1397		----			----			----			----		----
1399		----			----			----			----		----
1412	D3241	1			----			----		D3241	1		260
1429		----			----			----			----		----
1491		----			----			----			----		----
1496		----			----			----			----		----
1498		----			----			----			----		----
1528	D3241 - A1	1			----			----		D3241	1		260
1531		----			----			----			----		----
1538	D3241 - A1	<1		D3241 - A2	<1		D3241 - A3	3.32		D3241	0.1		260
1585		----			----			----			----		----
1586	D3241 - A1	<1			----			----		D3241	0		260
1587	D3241 - A1	1			----			----		D3241	0.1		260
1610	IP323 - B	< 1			----			----		IP323	< 1		260
1631	D3241 - A1	1			----			----		D3241	0		260
1634	D3241 - A1	<1			----			----		D3241	0		260
1710	D3241 - A1	1			----			----		D3241	0		260
1720		----			----			----			----		----
1724	D3241 - A1	1			----			----		D3241	0.3		260
1740		----			----			----			----		----
1755		----			----			----			----		----
1757		----			----			----			----		----
1776		----			----			----			----		----
1810		----			----			----			----		----
1811		----			----			----			----		----
1883		----			----			----			----		----
2129	IP323 - B	1			----			----		IP323	0		260
2130	IP323 - B	1			----			----		IP323	0		260
6049	D3241 - A1	< 1			----			----		D3241	zero		260
6075	D3241 - A1	1			----			----		D3241	0		260.0
6101	D3241 - A1	<1			----			----		D3241	0.2		260
6103		----			----			----			----		----
6108		----			----			----			----		----
normality		n.a.			n.a.			n.a.			n.a.		
n		58			7			3			57		
outliers		n.a.			n.a.			n.a.			2		
mean (n)		≤ 1			n.a.			n.a.			≤ 1		
st.dev. (n)		n.a.			n.a.			n.a.			n.a.		
R(calc.)		n.a.			n.a.			n.a.			n.a.		
R(D3241:16a)		unknown			unknown			unknown			unknown		



Lab 399: false positive test result?

Determination of Kinematic Viscosity at -20°C on sample #17030; results in mm²/s

lab	method	value	mark	z(targ)	remarks
120	D445	3.960		0.11	
131		----		----	
132	D445	3.904		-1.97	
140	D445	3.965		0.30	
150	D445	3.954		-0.11	
158		----		----	
159	D445	3.938		-0.71	
169	D445	3.9190		-1.42	
171	D445	3.965		0.30	
175	D445	3.949		-0.30	
177	D445	3.984		1.01	
194	D445	3.924		-1.23	
228		----		----	
230	D445	3.9661		0.34	
237	D445	3.979		0.82	
238		----		----	
253	D445	3.9871		1.12	
273		----		----	
317		----		----	
323	D445	3.937		-0.74	
333	D445	3.984		1.01	
334		----		----	
335	D445	3.947		-0.37	
336		----		----	
353		----		----	
360	D445	3.9753		0.68	
372	D445	3.956		-0.04	
391		----		----	
398		----		----	
399		----		----	
445	IP71	3.9647		0.29	
447	D445	3.959		0.07	
463		----		----	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631	D445	3.9741		0.64	
663	D445	3.9724		0.57	
671		----		----	
824	D445	3.950	C	-0.26	first reported: 1.866
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D445	4.0112		2.02	
998		----		----	
1006		----		----	
1023		----		----	
1039	ISO3104	3.995		1.42	
1049	D445	3.966		0.34	
1059	D445	3.940		-0.63	
1062	D445	3.935		-0.82	
1064	D445	3.9535		-0.13	
1080		----		----	
1082	D445	3.9379		-0.71	
1097	ISO3104	3.9741		0.64	
1105	D445	4.0327		2.82	
1109	D445	3.9349		-0.82	
1121		----		----	
1126		----		----	
1146		----		----	
1150	ISO3104	3.97847		0.80	
1155	ISO3104	3.9493		-0.29	
1161		----		----	
1191	D445	3.9185		-1.43	
1205		----		----	
1212	D445	3.9443		-0.47	
1237		----		----	
1284		----		----	
1297		----		----	
1299	D445	3.979		0.82	
1318	D7042	3.9281		-1.08	

lab	method	value	mark	z(targ)	remarks
1347	D445	3.976		0.71	
1348	D445	3.88		-2.87	
1372	D445	3.91395		-1.60	
1376	D445	3.9647		0.29	
1379		----		----	
1397	D7042	3.945		-0.45	
1399	D445	3.985		1.04	
1412	D445	3.986		1.08	
1429		----		----	
1491		----		----	
1496	D445	3.923		-1.27	
1498		----		----	
1528	D445	3.9994		1.58	
1531		----		----	
1538	D445	3.9650		0.30	
1585	D445	3.9682		0.42	
1586	D445	3.983		0.97	
1587	D445	3.91809		-1.45	
1610	D7042	3.986		1.08	
1631	D445	3.977		0.74	
1634	IP71	3.944		-0.48	
1710	D445	3.970		0.48	
1720		----		----	
1724	D445	3.958		0.04	
1740		----		----	
1755	D445	3.94105		-0.59	
1757		----		----	
1776	D445	3.970		0.48	
1810	D445	3.888		-2.57	
1811	D445	3.9466		-0.39	
1883		----		----	
2129	D445	3.966		0.34	
2130	IP71	3.917		-1.49	
6049	D445	3.994		1.38	
6075	D445	3.9965		1.47	
6101	D445	3.935		-0.82	
6103	ISO3104	3.931		-0.97	
6108		----		----	

normality OK
n 67
outliers 0
mean (n) 3.9570
st.dev. (n) 0.02857
R(calc.) 0.0800
R(D445:17a) 0.0752





Determination of Mercaptan Sulphur as S on sample #17030; results in %M/M

lab	method	value	mark	z(targ)	remarks
120	D3227	0.0004		-0.31	
131		----		----	
132	D3227	0.0007		2.25	
140	D3227	0.0003		-1.16	
150	D3227	0.0005		0.55	
158		----		----	
159	D3227	0.0005		0.55	
169	D3227	0.0005		0.55	
171		0.0005		0.55	
175	D3227	0.0004		-0.31	
177	D3227	0.0005		0.55	
194		----		----	
228		----		----	
230	D3227	0.00069		2.17	
237		----		----	
238		----		----	
253		----		----	
273		----		----	
317		----		----	
323	D3227	0.0004		-0.31	
333	D3227	0.0004		-0.31	
334	D3227	0.0004		-0.31	
335	D3227	0.0006		1.40	
336		----		----	
353		----		----	
360	D3227	0.00038		-0.48	
372	D3227	0.0005		0.55	
391		----		----	
398		----		----	
399		----		----	
445	D3227	0.0003		-1.16	
447	D3227	0.00048		0.38	
463		----		----	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631		----		----	
663	D3227	0.00036		-0.65	
671		----		----	
824	D3227	0.00035		-0.73	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974		----		----	
998		----		----	
1006	D3227	0.0005		0.55	
1023		----		----	
1039		----		----	
1049	D3227	0.000386		-0.43	
1059	D3227	0.0004		-0.31	
1062	D3227	0.0004		-0.31	
1064	D3227	0.00037		-0.56	
1080		----		----	
1082	ISO3012	0.000335		-0.86	
1097	ISO3012	0.00045		0.12	
1105	D3227	0.000414		-0.19	
1109	D3227	0.00036		-0.65	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155	ISO3012	0.00039		-0.39	
1161	ISO3012	0.0005		0.55	
1191	ISO3012	0.00033		-0.90	
1205		----		----	
1212	D3227	0.00049		0.46	
1237		----		----	
1284	D3227	0.00038		-0.48	
1297	D3227	0.00055		0.97	
1299	D3227	0.0003		-1.16	
1318		----		----	

lab	method	value	mark	z(targ)	remarks
1347	D3227	0.00066		1.91	
1348	D3227	0.00047		0.29	
1372	D3227	0.00048		0.38	
1376	D3227	0.00042		-0.14	
1379		----		----	
1397	D3227	0.0006		1.40	
1399	D3227	0.000392		-0.38	
1412	D3227	0.0005		0.55	
1429	D3227	0.00038		-0.48	
1491		----		----	
1496	D3227	0.00035		-0.73	
1498		----		----	
1528	D3227	0.00038		-0.48	
1531		----		----	
1538		----		----	
1585	D3227	0.00037		-0.56	
1586	D3227	0.0003		-1.16	
1587		----		----	
1610	IP342	< 0.0003		----	
1631	D3227	0.00046		0.20	
1634	D3227	0.0003		-1.16	
1710	D3227	0.0004		-0.31	
1720		----		----	
1724	D3227	0.00039		-0.39	
1740		----		----	
1755		----		----	
1757		----		----	
1776	D3227	0.0004		-0.31	
1810	D3227	0.00035		-0.73	
1811	D3227	0.000507		0.61	
1883		----		----	
2129	D3227	0.0004	C	-0.31	first reported: 0.003
2130	IP342	0.0004		-0.31	
6049	D3227	0.0005		0.55	
6075	D3227	0.0006		1.40	
6101		----		----	
6103		----		----	
6108		----		----	

normality OK
n 59
outliers 0
mean (n) 0.000436
st.dev. (n) 0.0000952
R(calc.) 0.000266
R(D3227:16) 0.000328





Determination of MSEP on sample #17030;

lab	method	value	mark	z(targ)	remarks
120	D3948	90		-0.37	
131		----		----	
132	D3948	90		-0.37	
140	D3948	93		0.40	
150		----		----	
158		----		----	
159	D3948	90		-0.37	
169	D3948	89		-0.63	
171	D3948	88		-0.89	
175	D3948	89		-0.63	
177	D3948	90		-0.37	
194		----		----	
228		----		----	
230	D3948	94		0.66	
237	D3948	96.0		1.18	
238		----		----	
253	D3948	95		0.92	
273		----		----	
317	D3948	93		0.40	
323	D3948	86		-1.41	
333	D3948	85		-1.66	
334	D3948	86		-1.41	
335	D3948	95		0.92	
336	D3948	85		-1.66	
353	D3948	86		-1.41	
360	D3948	93		0.40	
372	D3948	93		0.40	
391		----		----	
398	D3948	94		0.66	
399		----		----	
445	D3948	91		-0.11	
447	D3948	93		0.40	
463	D3948	86	C	-1.41	first reported: 73
468		----		----	
473	D3948	89		-0.63	
603	D3948	91		-0.11	
604		----		----	
608	D3948	97		1.43	
631	D3948	97		1.43	
663	D3948	93		0.40	
671	D3948	98		1.69	
824	D3948	95		0.92	
825		----		----	
904	D3948	97		1.43	
962		----		----	
963		----		----	
974	D3948	96		1.18	
998		----		----	
1006		----		----	
1023	D3948	81		-2.70	
1039	D3948	90		-0.37	
1049	D3948	94		0.66	
1059	D3948	89		-0.63	
1062	D3948	94		0.66	
1064	D3948	94		0.66	
1080		----		----	
1082		----		----	
1097	D3948	83		-2.18	
1105	D3948	95		0.92	
1109	D3948	90		-0.37	
1121	D3948	89		-0.63	
1126		----		----	
1146		----		----	
1150	D3948	88		-0.89	
1155		----		----	
1161		----		----	
1191	D3948	94		0.66	
1205		----		----	
1212	D3948	92		0.14	
1237		----		----	
1284		----		----	
1297		----		----	
1299	D3948	95		0.92	
1318	D3948	87		-1.15	

lab	method	value	mark	z(targ)	remarks
1347	D3948	91		-0.11	
1348	D3948	87		-1.15	
1372	D3948	97		1.43	
1376	D3948	94		0.66	
1379		----		----	
1397	D3948	96		1.18	
1399	D3948	89		-0.63	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D3948	95		0.92	
1498		----		----	
1528	D3948	97.5		1.56	
1531		----		----	
1538	D3948	97		1.43	
1585	D3948	88.7		-0.71	
1586	D3948	84		-1.92	
1587	D3948	88		-0.89	
1610	D3948	93		0.40	
1631	D3948	87		-1.15	
1634	D3948	91		-0.11	
1710	D3948	92		0.14	
1720	D3948	88		-0.89	
1724	D3948	94		0.66	
1740		----		----	
1755	D3948	95		0.92	
1757	D3948	95		0.92	
1776		----		----	
1810		----		----	
1811	D3948	90		-0.37	
1883	D3948	94		0.66	
2129	D3948	95		0.92	
2130	D3948	98		1.69	
6049	D3948	88		-0.89	
6075	D3948	91.5		0.01	
6101	D3948	85		-1.66	
6103	D3948	91		-0.11	
6108		----		----	

normality OK
n 76
outliers 0
mean (n) 91.44
st.dev. (n) 3.976
R(calc.) 11.13
R(D3948:14) 10.85





Determination of Naphthalenes on sample #17030; results in %V/V

lab	method	value	mark	z(targ)	remarks
120	D1840 - meth.B	0.54		0.24	
131		----		----	
132	D1840 - meth.B	0.53		-0.19	
140	D1840 - meth.A	0.53		-0.19	
150	D1840 - meth.A	0.61	R(0.05)	3.28	
158		----		----	
159	D1840 - meth.B	0.54		0.24	
169	D1840 - meth.B	0.53		-0.19	
171	D1840 - meth.A	0.53		-0.19	
175	D1840 - meth.B	0.54		0.24	
177	D1840 - meth.A	0.58		1.98	
194		----		----	
228		----		----	
230		----		----	
237	D1840 - meth.B	0.53		-0.19	
238		----		----	
253	D1840 - meth.B	0.53		-0.19	
273		----		----	
317		----		----	
323	D1840 - meth.A	0.54		0.24	
333	D1840 - meth.B	0.53		-0.19	
334		----		----	
335	D1840 - meth.A	0.58		1.98	
336		----		----	
353		----		----	
360		----		----	
372	D1840 - meth.B	0.53		-0.19	
391	D1840 - meth.B	0.50		-1.49	
398		----		----	
399		----		----	
445	D1840 - meth.B	0.57		1.54	
447	D1840 - meth.B	0.52		-0.63	
463	D1840 - meth.B	0.559		1.07	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631	D1840 - meth.A	0.531		-0.15	
663	D1840 - meth.A	0.537		0.11	
671		----		----	
824	D1840 - meth.B	0.51		-1.06	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D1840 - meth.A	0.52		-0.63	
998		----		----	
1006		----		----	
1023		----		----	
1039	D1840 - meth.B	0.54		0.24	
1049	D1840 - meth.A	0.527		-0.32	
1059	D1840 - meth.B	0.54	C	0.24	first reported: 0.43
1062	D1840 - meth.A	0.52		-0.63	
1064	D1840 - meth.A	0.522		-0.54	
1080		----		----	
1082	D1840 - meth.A	0.54		0.24	
1097	D1840 - meth.A	0.532		-0.11	
1105	D1840 - meth.A	0.5875		2.30	
1109	D1840 - meth.B	0.490		-1.93	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161		----		----	
1191	D1840 - meth.B	0.56		1.11	
1205		----		----	
1212	D1840 - meth.B	0.506		-1.23	
1237		----		----	
1284		----		----	
1297		----		----	
1299	D1840 - meth.B	0.53		-0.19	
1318	D1840 - meth.B	0.52		-0.63	

lab	method	value	mark	z(targ)	remarks
1347	D1840 - meth.B	0.522		-0.54	
1348	D1840 - meth.B	0.51		-1.06	
1372		----		----	
1376	D1840 - meth.A	0.535		0.02	
1379		----		----	
1397	D6379	0.491		-1.88	
1399		----		----	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D1840 - meth.B	0.5325		-0.08	
1498		----		----	
1528		----		----	
1531		----		----	
1538	D1840 - meth.B	0.54		0.24	
1585	D1840 - meth.B	0.545		0.46	
1586		----		----	
1587		----		----	
1610	D1840 - meth.B	0.36	R(0.01)	-7.57	
1631		----		----	
1634	D1840 - meth.B	0.52		-0.63	
1710	D1840 - meth.A	0.56		1.11	
1720	D1840 - meth.B	0.65	C,R(0.01)	5.01	first reported: 0.76
1724		----		----	
1740		----		----	
1755		----		----	
1757		----		----	
1776	D1840 - meth.A	0.54		0.24	
1810	D1840 - meth.A	0.53		-0.19	
1811	D1840 - meth.A	0.5119		-0.98	
1883		----		----	
2129	D1840 - meth.A	0.5453		0.47	
2130	D1840 - meth.B	0.58	C	1.98	first reported: 1.01
6049	D1840 - meth.A	0.54		0.24	
6075	D1840 - meth.B	0.536		0.07	
6101		----		----	
6103	D1840 - meth.B	0.5311		-0.15	
6108		----		----	

		D1840 method A only	D1840 method B only
normality	OK	suspect	OK
n	52	21	30
outliers	3	1	2
mean (n)	0.53445	0.53994	0.53205
st.dev. (n)	0.020751	0.020550	0.019319
R(calc.)	0.05810	0.05754	0.05409
R(D1840-B:07)	0.06455	-	0.06437
Compare R(D1840-A:07)	0.04588	0.04604	-





Determination of Smoke Point on sample #17030; results in mm

lab	method	value	mark	z(targ)	remarks
120	D1322_A	25.5		0.76	
131		----		----	
132	D1322_M	24.1		-0.27	
140	D1322_A	24.9		0.32	
150	D1322_A	24.2		-0.20	
158		----		----	
159	D1322_A	25.1		0.47	
169	D1322_A	25.0		0.39	
171	D1322_A	24.5		0.02	
175	D1322_M	22.2		-1.68	
177	D1322_A	24.7		0.17	
194	D1322_M	23.7		-0.57	
228		----		----	
230	D1322_M	24.9		0.32	
237	D1322_A	24.1		-0.27	
238		----		----	
253	D1322_M	24		-0.35	
273	D1322	24		-0.35	
317		----		----	
323	D1322_M	22.0		-1.83	
333	D1322_A	24.6		0.10	
334	D1322_A	24.8		0.24	
335	D1322_M	25.6		0.84	
336		----		----	
353	IP57_M	25.92		1.07	
360	D1322_M	24.5		0.02	
372	D1322_M	23.2		-0.94	
391	D1322_M	24		-0.35	
398		----		----	
399		----		----	
445	IP598_M	27.0		1.87	
447	D1322_M	24		-0.35	
463	D1322_M	23.50		-0.72	
468	D1322_M	27.0		1.87	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631	D1322_A	24.8		0.24	
663		----		----	
671		----		----	
824	D1322_A	24.2		-0.20	
825		----		----	
904	D1322	25		0.39	
962		----		----	
963		----		----	
974	D1322_M	25		0.39	
998		----		----	
1006		----		----	
1023		----		----	
1039	D1322_A	24.5		0.02	
1049	D1322_A	25.0		0.39	
1059	D1322_M	21.2		-2.42	
1062	D1322_M	22		-1.83	
1064	D1322_A	24.6		0.10	
1080		----		----	
1082	D1322_A	24.4		-0.05	
1097	D1322_M	24.33		-0.10	
1105	_A	24.3		-0.13	
1109	D1322_M	23.76		-0.52	
1121	D1322_M	23.5		-0.72	
1126		----		----	
1146		----		----	
1150		----		----	
1155	D1322	24.33		-0.10	
1161	ISO3014_M	24.2		-0.20	
1191	D1322_A	24.2		-0.20	
1205		----		----	
1212	D1322_M	24.8		0.24	
1237		----		----	
1284	D1322_A	24.5		0.02	
1297	D1322	24.5		0.02	
1299	D1322_A	24.1		-0.27	
1318	D1322_A	25.0		0.39	

lab	method	value	mark	z(targ)	remarks
1347	D1322_M	24		-0.35	
1348	D1322_M	25		0.39	
1372	D1322_A	24.3		-0.13	
1376	D1322_M	24.33	C	-0.10	first reported: 20.08
1379		----		----	
1397	D1322	24.0		-0.35	
1399	D1322_A	26		1.13	
1412	D1322_M	23.5		-0.72	
1429		----		----	
1491		----		----	
1496	D1322_A	24.4		-0.05	
1498		----		----	
1528	D1322_M	25.0		0.39	
1531		----		----	
1538	D1322_M	24.0		-0.35	
1585	D1322_M	26.25		1.32	
1586	D1322_M	24.0		-0.35	
1587	D1322_A	25.1		0.47	
1610	IP598	25.0		0.39	
1631	D1322_M	25		0.39	
1634	D1322_A	24.4		-0.05	
1710	D1322_M	25.5		0.76	
1720	_A	24.5		0.02	
1724	D1322	25		0.39	
1740		----		----	
1755		----		----	
1757		----		----	
1776	D1322_A	24.8		0.24	
1810	D1322_M	24.5		0.02	
1811	D1322_A	24.2		-0.20	
1883		----		----	
2129	D1322_M	24.5		0.02	
2130	IP598_A	24.4		-0.05	
6049	D1322_M	23.0		-1.09	
6075	D1322	24.3		-0.13	
6101	D1322_M	26.9		1.80	
6103	D1322_A	24.6		0.10	
6108		----		----	

		<u>D1322 Manual only</u>	<u>D1322 Automated only</u>
normality	not OK	OK	not OK
n	76	34	28
outliers	0	0	0
mean (n)	24.47	24.20	24.66
st.dev. (n)	0.984	1.279	0.441
R(calc.)	2.76	3.58	1.23
R(D1322:15e1-M)	3.79	3.76	-
Compare R(D1322:15e1-A)	0.90	-	0.90

A = automated and _M = manual





Determination of Specific Energy (Net, on Sulphur free basis) on sample #17030; results in MJ/kg

lab	method	value	mark	z(targ)	remarks
120	D3338	43.336		-0.05	
131		----		----	
132	D3338	43.345		0.50	
140	D3338	43.330		-0.41	
150	D3338	43.349		0.74	
158		----		----	
159	D3338	43.310	C	-1.63	first reported: 43.064
169	D3338	43.315		-1.33	
171	D3338	43.350		0.80	
175		----		----	
177	D3338	43.342		0.32	
194		----		----	
228		----		----	
230	D3338	43.351		0.86	
237	D3338	43.343		0.38	
238		----		----	
253	D3338	43.3372		0.02	
273		----		----	
317		----		----	
323	D3338	43.356		1.17	
333		----		----	
334		----		----	
335	D3338	43.328		-0.54	
336		----		----	
353		----		----	
360	D3338	43.356		1.17	
372	D3338	43.341		0.25	
391		----		----	
398		----		----	
399		----		----	
445	D3338	43.346		0.56	
447	D3338	43.344		0.44	
463	D3338	43.36		1.41	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608		----		----	
631	D3338	43.361		1.47	
663	D3338	43.351		0.86	
671		----		----	
824	D3338	43.339		0.13	
825		----		----	
904		----		----	
962		----		----	
963		----		----	
974	D3338	43.334		-0.17	
998		----		----	
1006		----		----	
1023		----		----	
1039	D3338	43.359		1.35	
1049	D3338	43.3087		-1.71	
1059	D3338	43.336		-0.05	
1062	D3338	43.313		-1.45	
1064	D3338	43.3471		0.63	
1080		----		----	
1082	D3338	43.3125		-1.48	
1097	D3338	43.332		-0.29	
1105	D3338	43.32		-1.02	
1109	D3338	43.33		-0.41	
1121		----		----	
1126		----		----	
1146		----		----	
1150		----		----	
1155		----		----	
1161	D3338	43.338		0.07	
1191	D3338	43.315		-1.33	
1205		----		----	
1212	D3338	43.350		0.80	
1237		----		----	
1284	D3338	43.311		-1.57	
1297		----		----	
1299	D3338	43.33		-0.41	
1318	D3338	43.324		-0.78	

lab	method	value	mark	z(targ)	remarks
1347	D3338	43.321		-0.96	
1348	D3338	43.32		-1.02	
1372	D3338	43.3269		-0.60	
1376	D3338	43.3275		-0.57	
1379		----		----	
1397	D3338	43.29		-2.85	
1399		----		----	
1412		----		----	
1429		----		----	
1491		----		----	
1496	D3338	43.3564		1.19	
1498		----		----	
1528	D3338	43.3430		0.38	
1531		----		----	
1538		----		----	
1585	D3338	43.350		0.80	
1586	D3338	43.34		0.19	
1587		----		----	
1610	D3338	43.324		-0.78	
1631		43.354		1.05	
1634		----		----	
1710	D3338	43.364		1.65	
1720		----		----	
1724		----		----	
1740		----		----	
1755		----		----	
1757		----		----	
1776	D3338	43.34		0.19	
1810		----		----	
1811	D3338	43.3480		0.68	
1883		----		----	
2129	D3338	43.336		-0.05	
2130	D3338	43.37		2.02	
6049	D3338	43.32		-1.02	
6075	D3338	43.3454		0.52	
6101	D3338	43.33499	C	-0.11	first reported: 43.2449
6103	D4809	43.158	C,R(0.01)	-10.88	first reported: 44.056
6108		----		----	

normality OK
n 56
outliers 1
mean (n) 43.3368
st.dev. (n) 0.01648
R(calc.) 0.0461
R(D3338:09e2) 0.0460





Determination of Sulphur, Total on sample #17030; results in mg/kg

lab	method	value	mark	z(targ)	remarks
120	D4294	585		-0.29	
131		----		----	
132	D2622	600		0.26	
140		----		----	
150	D5453	573		-0.74	
158		----		----	
159	D4294	591		-0.07	
169	D4294	572		-0.78	
171	D5453	558		-1.30	
175		----		----	
177	D4294	553	C	-1.49	first reported: 0.0553 mg/kg
194	D2622	606		0.49	
228		----		----	
230	D4294	603	C	0.38	reported: 0.0603 mg/kg
237	D5453	566.0		-1.00	
238	D4294	573.0		-0.74	
253	D4294	590		-0.11	
273	D5453	600		0.26	
317		----		----	
323	D5453	587		-0.22	
333	D4294	599.0	C	0.23	first reported: 0.06 mg/kg
334	D5453	536		-2.12	
335	D4294	605	C	0.45	first reported: 0.0605 mg/kg
336	ISO8754	593		0.00	
353	IP336	628	C	1.31	first reported: 0.0628 mg/kg
360	D5453	580		-0.48	
372	D5453	596		0.12	
391	D5453	592		-0.03	
398	D4294	600		0.26	
399	D4294	605		0.45	
445	D5453	633.3		1.51	
447	IP336	580	C	-0.48	first reported: 0.058 mg/kg
463	D4294	610		0.64	
468		----		----	
473		----		----	
603		----		----	
604		----		----	
608	D5453	581.3		-0.43	
631	D4294	594		0.04	
663	D5453	578.6		-0.53	
671		----		----	
824	D5453	590		-0.11	
825	IP336	593		0.00	
904		----		----	
962		----		----	
963		----		----	
974	D4294	582		-0.41	
998	D4294	624		1.16	
1006	D2622	606		0.49	
1023	ISO14596	580		-0.48	
1039	D2622	598		0.19	
1049	D5453	603.1		0.38	
1059	ISO14596	520		-2.72	
1062	D5453	602		0.34	
1064	D5453	566.55		-0.98	
1080	D5453	597		0.15	
1082	D4294	650		2.13	
1097	D5453	563.14		-1.11	
1105	D4294	590.55		-0.09	
1109	D2622	610.3		0.65	
1121	IP336	606		0.49	
1126	ISO20846	595.1		0.08	
1146	D4294	530		-2.34	
1150	ISO20884	547.56		-1.69	
1155	D1266	631.90		1.45	
1161	ISO20846	565		-1.04	
1191	D4294	600		0.26	
1205	ISO14596	603.7		0.40	
1212	D4294	634		1.53	
1237	ISO8754	640		1.76	
1284	D2622	583		-0.37	
1297	D4294	596.2		0.12	
1299	D2622	590		-0.11	
1318	D5453	552.3		-1.51	

lab	method	value	mark	z(targ)	remarks
1347	D5453	597.5		0.17	
1348	D5453	539		-2.01	
1372	D4294	615		0.82	
1376	D5453	593.9		0.04	
1379		----		----	
1397	D2622	604		0.41	
1399	D4294	554		-1.45	
1412	D5453	600		0.26	
1429		----		----	
1491		----		----	
1496	D4294	580.5		-0.46	
1498	D5453	684.0	R(0.05)	3.40	
1528	D2622	605.8		0.48	
1531		----		----	
1538	D5453	567		-0.97	
1585	D4294	557.5		-1.32	
1586	D5453	618		0.94	
1587	D4294	588.4		-0.17	
1610	IP336	690	R(0.05)	3.62	
1631		579.4		-0.50	
1634	D5453	626	C	1.23	first reported: 18.2
1710	D4294	597		0.15	
1720	D5453	600.5		0.28	
1724	D5453	586		-0.26	
1740	D4294	625		1.20	
1755	D2622	619.1062		0.98	
1757		----		----	
1776	D5453	564		-1.08	
1810	ISO8754	624		1.16	
1811	D5453	619.2		0.98	
1883		----		----	
2129	IP496	623.0		1.12	
2130	D5453	646	C	1.98	first reported: 137
6049	D4294	613		0.75	
6075	D5453	600	C	0.26	reported: 0.06 mg/kg
6101	D4294	613		0.75	
6103		----		----	
6108		----		----	

normality OK
n 88
outliers 2
mean (n) 592.891
st.dev. (n) 25.6748
R(calc.) 71.890
R(D5453:16e1) 75.119




Determination of Particle Size Distribution on sample #17031 acc. to IP564, results in counts/ml

lab	method	≥4µm (c)	m	≥6µm (c)	m	≥14µm (c)	m	≥21µm (c)	m	≥25µm (c)	m	≥30µm (c)	m
140		----		----		----		----		----		----	
150		----		----		----		----		----		----	
171		----		----		----		----		----		----	
230	IP564	10585.6		3693.5		178.4		75.9	G1	68.2	G1	64.4	G1
237		----		----		----		----		----		----	
253	IP564	14770		5866		343	D5	17		4		1	
323		----		----		----		----		----		----	
333		----		----		----		----		----		----	
334		----		----		----		----		----		----	
335		----		----		----		----		----		----	
360		----		----		----		----		----		----	
372	IP564	15106.6		5012.0		106.0		11.7		5.2		2.0	
445		----		----		----		----		----		----	
447	IP564	12431.2		3386.1		39.4		5.4		2.3		0.8	
824		----		----		----		----		----		----	
904		----		----		----		----		----		----	
962		----		----		----		----		----		----	
963		----		----		----		----		----		----	
974		----		----		----		----		----		----	
1023		----		----		----		----		----		----	
1039		----		----		----		----		----		----	
1059	IP564	14312.7		5056.4		150.0		38.9	DG5	30.4	G1	27.8	G1
1062		----		----		----		----		----		----	
1064	IP564	17452.47		5752.97		133.50		9.23		2.17		0.50	
1095		----		----		----		----		----		----	
1097	IP564	19146.8		7667.7		201.9		16.7		6.1		1.9	
1105	IP564	8119.1	D5	2749.3	D5	68.5		7.4		2.6		1.2	
1109		----		----		----		----		----		----	
1299		----		----		----		----		----		----	
1376		----		----		----		----		----		----	
1498	IP564	4226.3	C,D5	968.9	C,D5	19.6	ex,C	4.6	C	2.5	C	1.0	C
1528		----		----		----		----		----		----	
1585		----		----		----		----		----		----	
1587	IP564-B	16053.1		6029.3		149.8		18.9		8.4		3.3	
1610		----		----		----		----		----		----	
1631		----		----		----		----		----		----	
1634		----		----		----		----		----		----	
1710		----		----		----		----		----		----	
1720		----		----		----		----		----		----	
1724		----		----		----		----		----		----	
1776	IP564	18920.5		7627.3		129.6		7.9		2.8		0.6	
1810		----		----		----		----		----		----	
1811		----		----		----		----		----		----	
2130	IP564	17238		7389		260		33	DG5	17	G1	5	G5
6075		----		----		----		----		----		----	
6101		----		----		----		----		----		----	
6108		----		----		----		----		----		----	
normality		OK		OK		OK		OK		suspect		not OK	
n		10		10		10		9		9		9	
outliers		2		2		1+1ex		3		3		3	
mean (n)		15602		5748		141.7		10.98		4.01		1.37	
st.dev. (n)		2742.7		1521.1		63.61		5.355		2.153		0.893	
R(calc.)		7679		4259		178.1		14.99		6.03		2.50	
R(IP564:13)		2968		1740		75.1		13.92		5.37		2.26	

m=mark
D5=D(0.05)
DG5=DG(0.05)
G(1)=G(0.01)
G(5)=G(0.05)

Lab 1498 reported the test results in counts/ml in the table of scale number. Test result at ≥14µm is excluded as corresponding test result in ISO4406 scale number is a statistical outlier.
Lab 1587 reported to use IP564 Annex B



Determination of Particle Size Distribution by IP564 on sample #17031 acc. to ISO4406 scale, results in scale number


lab	method	≥4µm (c)	m	≥6µm (c)	m	≥14µm (c)	m	≥21µm (c)	m	≥25µm (c)	m	≥30µm (c)	m
140		----		----		----		----		----		----	
150		----		----		----		----		----		----	
171		----		----		----		----		----		----	
230	ISO4406	21		19		15		13	ex	13	ex	13	D5
237		----		----		----		----		----		----	
253	ISO4406	21		20		15	ex	12		9		7	
323		----		----		----		----		----		----	
333		----		----		----		----		----		----	
334		----		----		----		----		----		----	
335		----		----		----		----		----		----	
360		----		----		----		----		----		----	
372	ISO4406	21		20		14		11		10		8	
445		----		----		----		----		----		----	
447	ISO4406	21		19		13		10		9		7	
824		----		----		----		----		----		----	
904		----		----		----		----		----		----	
962		----		----		----		----		----		----	
963		----		----		----		----		----		----	
974		----		----		----		----		----		----	
1023		----		----		----		----		----		----	
1039		----		----		----		----		----		----	
1059	ISO4406	21		20		14		13	ex	12	ex	12	ex
1062		----		----		----		----		----		----	
1064		----		----		----		----		----		----	
1095		----		----		----		----		----		----	
1097	ISO4406	21		20		15		11		10		8	
1105	ISO4406	20	ex	19	ex	13		10		9		7	
1109		----		----		----		----		----		----	
1299		----		----		----		----		----		----	
1376		----		----		----		----		----		----	
1498	ISO4406	19	C,G1	17	C,G1	11	C,G5	----	C	----	C	----	C
1528		----		----		----		----		----		----	
1585		----		----		----		----		----		----	
1587	ISO4406	21		20		14		11		10		9	
1610		----		----		----		----		----		----	
1631		----		----		----		----		----		----	
1634		----		----		----		----		----		----	
1710		----		----		----		----		----		----	
1720		----		----		----		----		----		----	
1724		----		----		----		----		----		----	
1776	ISO4406	21		20		14		10		9		>6	
1810		----		----		----		----		----		----	
1811		----		----		----		----		----		----	
2130	ISO4406	21		20		15		12	ex	11	ex	10	ex
6075		----		----		----		----		----		----	
6101		----		----		----		----		----		----	
6108		----		----		----		----		----		----	
normality		n.a.		OK		OK		OK		OK		OK	
n		9		9		9		7		7		6	
outliers		1+1ex		1+1ex		1+1ex		0+3ex		0+3ex		1+2ex	
mean (n)		21.00		19.78		14.11		10.71		9.43		7.67	
st.dev. (n)		0.000		0.441		0.782		0.756		0.535		0.816	
R(calc.)		0.00		1.23		2.19		2.12		1.50		2.29	
R(IP564:13)		1.02		1.43		2.22		unknown		unknown		unknown	

m=mark
D5=D(0.05)
G(1)=G(0.01)
G(5)=G(0.05)

Test results of labs 230, 253, 1059, 1105 and 2130 are excluded as corresponding test results in counts are statistical outliers.

Lab 1498 reported the test results in counts/ml in the table of scale number.

Lab 1498 first reported: 48.6; 10.8; 0.4; 0.1; 0.1; 0.1



Determination of Particle Size Distribution on sample #17031 acc. to IP565, results in counts/ml

lab	method	≥4µm (c)	m	≥6µm (c)	m	≥14µm (c)	m	≥21µm (c)	m	≥25µm (c)	m	≥30µm (c)	m
140		----		----		----		----		----		----	
150	IP565	23374		9554		583		152	R1	58	G1	21	R1
171		----		----		----		----		----		----	
230		----		----		----		----		----		----	
237	IP565	18128.6	D5	6237.8	DG5	122.3		3.7		0.6		0.1	
253		----		----		----		----		----		----	
323		----		----		----		----		----		----	
333	IP565	23103		9400		248		35		9		2	
334	IP565	20497		7679		250		17		4		1	
335	IP565	22145.0		8019.0		455.8		58.5		14.0		2.5	
360	IP565	22979.4		9041.0		269.4		18.0		4.7		0.8	
372		----		----		----		----		----		----	
445		----		----		----		----		----		----	
447		----		----		----		----		----		----	
824	IP565	22393		8548		282		23		5		1	
904		----		----		----		----		----		----	
962		----		----		----		----		----		----	
963		----		----		----		----		----		----	
974	IP565	21780		8601		458		16.8		3.7		1.2	
1023	IP577	25667.53	ex,C	11359.57	ex,C	721.03	ex,C	148.30	ex,C	82.60	ex,C	-----	W
1039	IP565	23548		9188		382		43		10		2	
1059		----		----		----		----		----		----	
1062	IP565	19512		7322		117		2.9	ex	0.4	ex	0.3	ex
1064		----		----		----		----		----		----	
1095	IP565	22226		8878		452		58		13		3	
1097		----		----		----		----		----		----	
1105		----		----		----		----		----		----	
1109	IP565	22635.1		8852.7		191.5		5.3		2.0		0.5	
1299	IP577	19631.6	ex,C	8361.4	ex	295.2	ex	33.5	ex	12.3	ex	7.3	ex
1376		----		----		----		----		----		----	
1498		----		----		----		----		----		----	
1528	IP565	23079.4		9543.9		568.7		108.0	R1	43.9	G1	18.6	R1
1585	IP565	21751.5		8030.3		285.1		27.2		8.5		2.1	
1587		----		----		----		----		----		----	
1610		----		----		----		----		----		----	
1631	IP565	20274.8		8342.0		326.5		30.8		8.8		2.5	
1634		----		----		----		----		----		----	
1710	IP565	22491.2		8479.6		267.5		17.8		4.1		1.5	
1720	IP565	19828.1		6483.3	DG5	308.4		32.0		10.4		3.1	
1724	IP565	19904.9		8422		335		27.3		5.5		0.6	
1776		----		----		----		----		----		----	
1810	IP565	-----	W	-----	W	-----	W	-----	W	-----	W	-----	W
1811	IP565	24062.1		9357.0		343.4		31.9		7.0		2.2	
2130		----		----		----		----		----		----	
6075	IP565	31389.8	D1	8965.9		316.8		26.7		6.8		1.3	
6101		----		----		----		----		----		----	
6108		----		----		----		----		----		----	
normality		OK		OK		OK		OK		OK		OK	
n		18		18		20		17		17		17	
outliers		2+2ex		2+2ex		0+2ex		2+3ex		2+3ex		2+2ex	
mean (n)		21977		8679		328.1		27.76		6.89		1.61	
st.dev. (n)		1401.3		640.6		126.55		15.261		3.714		0.896	
R(calc.)		3924		1794		354.3		42.73		10.40		2.51	
R(IP565:13)		2388		1829		168.1		23.83		8.08		2.70	

m=mark

D1=D(0.01) / D5=D(0.05)

DG5=DG(0.05) / G(1)=G(0.01) / R1=R(0.01)


Lab 1023 reported to use Pamas and test method IP564; Pamas is in accordance with test method IP577, therefore test method IP577 is given by iis.

Labs 1023 and 1299: test results are excluded as the reported equipment is Pamas, see also §4.1.

Lab 1023 reported the test results in counts/ml in the table of scale number. Lab 1023 first reported respectively: 2566753; 1135957; 72103; 14830; 14830; 14830 counts/ml.

Lab 1063: the test results at ≥ 21, 25 and 30 µm are excluded as corresponding test results in ISO4406 scale number are statistical outliers.

Lab 1810 first reported: 6417; 2685; 83; 7; 3; 1.4



Determination of Particle Size Distribution by IP565 on sample #17031 acc. to ISO4406 scale, results in scale number

lab	method	≥4µm (c)	m	≥6µm (c)	m	≥14µm (c)	m	≥21µm (c)	m	≥25µm (c)	m	≥30µm (c)	m
140		----		----		----		----		----		----	
150	ISO4406	22		20		16		14	ex	13	ex	12	ex
171		----		----		----		----		----		----	
230		----		----		----		----		----		----	
237	ISO4406	21	ex	20	ex	14		≥9		≥6		≥4	
253		----		----		----		----		----		----	
323		----		----		----		----		----		----	
333	ISO4406	22		20		15		12		10		8	
334	ISO4406	22		20		15		11		9		7	
335	ISO4406	22		20		16		13		11		8	
360	ISO4406	22		20		15		11		9		7	
372		----		----		----		----		----		----	
445		----		----		----		----		----		----	
447		----		----		----		----		----		----	
824	ISO4406	22		20		15		12		10		7	
904		----		----		----		----		----		----	
962		----		----		----		----		----		----	
963		----		----		----		----		----		----	
974	ISO4406	22		20		16		11		9		7	
1023	ISO4406	----	C	----	C	----	C	----	C	----	C	----	C
1039	ISO4406	22		20		16		13		11		8	
1059		----		----		----		----		----		----	
1062	ISO4406	21		20		14		9	D5	6	G5	5	C,D5
1064		----		----		----		----		----		----	
1095	ISO4406	22		20		16		----		----		----	
1097		----		----		----		----		----		----	
1105		----		----		----		----		----		----	
1109		22		20		15		10		8		6	
1299	ISO4406	21	ex	20	ex	15	ex	12	ex	11	ex	10	ex
1376		----		----		----		----		----		----	
1498		----		----		----		----		----		----	
1528	ISO4406	22		20		16		14	ex	13	ex	11	ex
1585	ISO4406	22		20		15		12		10		8	
1587		----		----		----		----		----		----	
1610		----		----		----		----		----		----	
1631	ISO4406	22		20		16		12		10		8	
1634		----		----		----		----		----		----	
1710	ISO4406	22		20		15		11		9		8	
1720		----		----		----		----		----		----	
1724	ISO4406	21		20		16		12		10		6	
1776		----		----		----		----		----		----	
1810		----		----		----		----		----		----	
1811		----		----		----		----		----		----	
2130		----		----		----		----		----		----	
6075		22	ex	20		15		12		10		7	
6101		----		----		----		----		----		----	
6108		----		----		----		----		----		----	
normality		not OK		n.a.		OK		OK		OK		OK	
n		16		17		18		13		13		13	
outliers		0+3ex		0+2ex		0+1ex		1+3ex		1+3ex		1+3ex	
mean (n)		21.88		20.00		15.33		11.69		9.69		7.31	
st.dev. (n)		0.342		0.000		0.686		0.855		0.855		0.751	
R(calc.)		0.96		0.00		1.92		2.39		2.39		2.10	
R(IP565:13)		1.03		1.03		1.37		unknown		unknown		unknown	











m=mark

D5=D(0.05)

Lab 1299: test results are excluded as the reported equipment is Pamas (test method IP577), see also §4.1.

Lab 1023 reported the test results of counts/ml in the table of scale number.

Lab 1062 first reported: 0



APPENDIX 2

z-scores distillation ASTM D86.

lab	method	IBP	10% rec.	50% rec.	90%rec.	FBP
120	D86-A	0.09	-1.17	-0.83	-0.16	-0.72
131	D86-A	0.71	0.04	0.38	0.84	-0.64
132	D86-A	0.78	0.49	0.10	0.07	-0.68
140	D86-A	-1.15	-0.34	0.10	-0.77	-0.68
150	D86-A	-0.11	-0.04	-0.09	0.69	-0.56
158		----	----	----	----	----
159	D86-A	1.75	0.49	0.94	1.37	0.94
169	D86-A	0.13	0.34	-0.18	-0.54	-0.64
171	D86-A	0.65	-0.04	-2.14	-1.07	-1.11
175	D86-A	-0.32	0.04	0.66	2.06	0.62
177	D86-A	-1.88	-0.34	0.94	1.07	0.94
194	D86-A	-0.91	-0.34	-0.46	-1.15	0.07
228	D86	-0.81	-1.47	-2.70	-3.14	-0.44
230	D86-A	0.58	0.04	-1.02	0.76	-0.40
237	D86-M	0.58	-0.72	-0.83	-0.46	-0.44
238	D86-M	-0.46	-0.72	-2.70	-2.76	-0.64
253	D86-M	-1.15	-0.34	-0.83	-1.61	-0.24
273	D86-A	0.85	-0.57	0.47	0.15	0.98
317	D86-A	-1.39	-0.49	0.19	-0.23	-0.76
323	D86-A	0.16	-0.34	0.01	-0.31	-0.84
333	D86-A	0.89	0.26	0.01	-0.61	0.54
334	D86-A	1.10	-0.64	-0.74	1.99	0.70
335	D86-A	-0.01	-0.04	0.10	1.76	-0.52
336	D86-A	0.96	0.41	0.66	0.46	-0.09
353	IP123-A	1.10	0.26	0.66	1.68	1.02
360	D86-A	0.27	1.09	1.13	1.45	-0.17
372	D86-A	0.09	0.41	0.19	0.46	-0.05
391	D86-A	0.23	-0.42	0.75	0.38	0.90
398	D86-A	1.61	0.79	0.38	-0.08	-0.13
399	D86-A	0.54	0.87	0.66	-0.69	-0.05
445	IP123-A	-1.39	-0.94	-0.27	0.23	-0.17
447	D86-A	-2.29	-0.94	-0.83	0.84	-0.40
463	D86-A	1.13	0.64	0.10	-0.46	0.82
468		----	----	----	----	----
473		0.37	-0.11	0.01	0.07	0.47
603	D86-A	1.37	1.09	1.22	0.46	0.74
604	D86-A	-1.22	-1.62	-1.95	-1.76	-1.31
608	D86-M	0.58	0.79	1.97	0.69	0.35
631	D86-A	0.37	-0.19	-0.83	1.14	-0.68
663	D86-A	-1.44	0.22	0.85	1.60	0.54
671		----	----	----	----	----
824	D86-A	0.89	0.64	0.57	0.53	1.14
825	D86-A	0.33	0.34	0.29	0.15	-0.05
904	D86-A	-0.46	-1.62	-1.67	-1.68	-0.91
962		----	----	----	----	----
963		----	----	----	----	----
974	D86-A	-0.08	-1.17	-1.49	-1.61	-0.48
998	D86-M	0.06	0.41	-0.83	-0.46	1.14
1006	D86-A	1.37	1.32	0.94	0.61	0.51
1023	D86-A	1.13	1.47	0.94	0.15	0.98
1039	D2887	-0.22	-0.19	1.31	-0.54	0.70
1049	D86-A	0.68	0.64	0.38	-0.08	0.74
1059	D86-A	-0.60	0.04	-0.65	-1.07	-0.40
1062	D86-A	-0.74	0.56	1.03	0.23	-0.24
1064	D86-A	0.54	0.71	1.41	1.99	0.86
1080		----	----	----	----	----
1082	D86-A	-0.22	1.17	0.85	0.53	0.43
1097	ISO3405-A	0.89	-0.04	0.47	0.69	0.23
1105	D86-A	-0.15	-0.26	-0.37	-2.76	-0.72
1109	D86-A	-0.32	-0.72	-0.65	-0.46	-0.13
1121	D86-M	-1.67	-1.85	-1.77	-1.99	-1.23
1126		-0.25	0.79	2.99	0.53	0.98
1146	ISO3405-A	-0.49	0.04	0.29	0.23	-0.17
1150	ISO3405-A	0.92	1.09	0.19	-0.54	2.40
1155	ISO3405-A	0.51	0.11	-0.65	-1.15	-0.36
1161	ISO3405-A	-1.29	0.34	0.94	0.00	0.07
1191	D86-A	-1.22	1.32	0.57	-0.77	0.15
1205	D86-A	1.34	1.09	1.41	0.46	0.39
1212	D86-A	-0.08	-0.34	0.29	1.30	0.54
1237	ISO3405-M	1.75	-0.42	-0.55	-0.61	0.03
1284	D86-A	0.13	-0.42	-0.93	-0.46	0.39

lab	method	IBP	10% rec.	50% rec.	90%rec.	FBP
1297	D86-A	-0.46	-0.11	-0.55	-1.84	-0.52
1299	D86-A	0.16	0.11	0.38	0.76	0.66
1318	D86-A	-0.94	0.04	-0.09	-0.54	-0.28
1347	D86-M	1.20	0.64	0.36	-1.00	0.25
1348	D86-A	-0.43	0.79	-0.83	-1.23	-0.84
1372	D86-M	-0.46	-1.47	-0.83	-0.84	-0.05
1376	D86-A	0.30	0.34	0.01	-0.31	0.51
1379		----	----	----	----	----
1397	D86-A	-0.60	-0.11	-0.37	-1.46	-0.36
1399	D86-A	-0.81	-1.09	-0.74	0.92	-0.88
1412	D86-M	0.75	-0.34	-0.83	-1.23	-2.02
1429	D86-A	-0.87	-1.02	-1.21	-1.00	-1.19
1491	D86-A	0.33	0.49	0.66	1.14	0.94
1496	D86-A	0.09	2.00	1.59	1.99	0.39
1498	D86-A	-0.53	0.11	0.94	1.68	1.06
1528	D86-A	1.16	0.71	1.03	1.14	0.51
1531	D86-A	0.06	-0.34	1.03	0.99	0.78
1538	D86-A	-0.15	-0.72	-1.11	-1.99	-0.72
1585	D86-A	0.06	0.41	0.38	-1.61	0.51
1586	D86-A	-0.22	0.11	-0.09	1.60	0.11
1587	D86-A	-0.43	0.26	0.57	-0.69	-0.20
1610	D86-A	-0.43	-0.57	-0.55	0.07	-0.52
1631	D86-A	0.82	0.64	0.57	0.23	0.86
1634	D86-A	-0.74	0.64	0.19	1.07	0.03
1710	D86-A	0.54	0.87	0.75	0.92	0.19
1720	D86-A	-2.12	0.04	0.75	-0.54	1.10
1724	D86-A	-0.32	-0.11	0.47	-0.08	0.07
1740	D86-A	0.96	0.94	0.19	-0.16	-0.56
1755	D86-A	-0.63	0.04	0.10	1.07	-0.72
1757	D86-A	-0.25	-0.11	0.01	1.45	-0.91
1776	D86-A	-2.05	-0.42	0.01	0.76	-0.28
1810	D86-A	-0.43	-1.02	-1.77	-1.61	-1.31
1811	D86-A	-0.49	-0.87	-1.39	-1.15	-0.80
1883	D86-M	0.92	-2.22	-0.83	-0.84	-0.44
2129	D86-A	-0.84	0.04	-0.46	-0.31	-0.48
2130	IP123-A	0.99	1.02	0.75	0.53	0.35
6049	D86-A	0.78	-0.34	-0.83	0.15	-0.48
6075	D86-A	-0.74	-0.94	-0.09	0.00	0.98
6101	D86-A	0.58	-0.04	0.38	0.07	0.27
6103	ISO3405-A	0.11	0.11	0.05	1.07	-0.26
6108		----	----	----	----	----

z-scores Particle Size Distribution on sample #17031 acc. to IP564 and IP565, results in counts/ml

		IP 564						IP565						
lab	method	≥4 µm	≥6 µm	≥14 µm	≥21 µm	≥25 µm	≥30 µm	method	≥4 µm	≥6 µm	≥14 µm	≥21 µm	≥25 µm	≥30 µm
140		----	----	----	----	----	----							
150		----	----	----	----	----	----	IP565	1.64	1.34	4.24	14.60	17.70	20.11
171		----	----	----	----	----	----							
230	IP564	-4.73	-3.31	1.37	13.05	33.49	78.03							
237		----	----	----	----	----	----	IP565	-4.51	-3.74	-3.43	-2.83	-2.18	-1.57
253	IP564	-0.78	0.19	7.50	1.21	0.00	-0.45							
323		----	----	----	----	----	----							
333		----	----	----	----	----	----	IP565	1.32	1.10	-1.33	0.85	0.73	0.40
334		----	----	----	----	----	----	IP565	-1.73	-1.53	-1.30	-1.27	-1.00	-0.63
335		----	----	----	----	----	----	IP565	0.20	-1.01	2.13	3.61	2.46	0.92
360		----	----	----	----	----	----	IP565	1.18	0.55	-0.98	-1.15	-0.76	-0.84
372	IP564	-0.47	-1.18	-1.33	0.14	0.62	0.78							
445		----	----	----	----	----	----							
447	IP564	-2.99	-3.80	-3.81	-1.12	-0.89	-0.70							
824		----	----	----	----	----	----	IP565	0.49	-0.20	-0.77	-0.56	-0.65	-0.63
904		----	----	----	----	----	----							
962		----	----	----	----	----	----							
963		----	----	----	----	----	----							
974		----	----	----	----	----	----	IP565	-0.23	-0.12	2.16	-1.29	-1.10	-0.43
1023		----	----	----	----	----	----	IP577 *)	4.33	4.10	6.54	14.16	26.22	----
1039		----	----	----	----	----	----	IP565	1.85	0.78	0.90	1.79	1.08	0.40
1059	IP564	-1.22	-1.11	0.31	5.61	13.77	32.72							
1062		----	----	----	----	----	----	IP565	-2.89	-2.08	-3.52	-2.92	-2.25	-1.36
1064	IP564	1.75	0.01	-0.31	-0.35	-0.96	-1.07							
1095		----	----	----	----	----	----	IP565	0.29	0.30	2.06	3.55	2.12	1.44
1097	IP564	3.34	3.09	2.24	1.15	1.09	0.66							
1105	IP564	-7.06	-4.83	-2.73	-0.72	-0.73	-0.21							
1109		----	----	----	----	----	----	IP565	0.77	0.27	-2.28	-2.64	-1.69	-1.15
1299		----	----	----	----	----	----	IP577	-2.75	-0.49	-0.55	0.67	1.87	5.90
1376		----	----	----	----	----	----							
1498	IP564	-10.73	-7.69	-4.55	-1.28	-0.79	-0.45							
1528		----	----	----	----	----	----	IP565	1.30	1.32	4.01	9.43	12.82	17.62
1585		----	----	----	----	----	----	IP565	-0.26	-0.99	-0.72	-0.07	0.56	0.51
1587	IP564-B	0.43	0.45	0.30	1.59	2.29	2.39							
1610		----	----	----	----	----	----							
1631		----	----	----	----	----	----	IP565	-1.99	-0.52	-0.03	0.36	0.66	0.92
1634		----	----	----	----	----	----							
1710		----	----	----	----	----	----	IP565	0.61	-0.31	-1.01	-1.17	-0.97	-0.12
1720		----	----	----	----	----	----	IP565	-2.52	-3.36	-0.33	0.50	1.22	1.54
1724		----	----	----	----	----	----	IP565	-2.43	-0.39	0.11	-0.05	-0.48	-1.05
1776	IP564	3.13	3.02	-0.45	-0.62	-0.63	-0.95							
1810		----	----	----	----	----	----	IP565	----	----	----	----	----	----
1811		----	----	----	----	----	----	IP565	2.45	1.04	0.25	0.49	0.04	0.61
2130	IP564	1.54	2.64	4.41	4.43	6.78	4.50							
6075		----	----	----	----	----	----	IP565	11.04	0.44	-0.19	-0.13	-0.03	-0.32
6101		----	----	----	----	----	----							
6108		----	----	----	----	----	----							

*) Lab 1023 reported to use Pamas and test method IP564; Pamas is in accordance with test method IP577

z-scores Particle Size Distribution on sample #17031 by IP564 and IP565 acc. to ISO4406 scale, results in scale number

		IP 564						IP565						
lab	method	≥4 µm	≥6 µm	≥14 µm	≥21 µm	≥25 µm	≥30 µm	method	≥4 µm	≥6 µm	≥14 µm	≥21 µm	≥25 µm	≥30 µm
140		----	----	----	----	----	----							
150		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
171		----	----	----	----	----	----							
230	ISO4406	0.00	-1.52	1.12	----	----	----							
237		----	----	----	----	----	----	ISO4406	-2.39	0.00	-2.72	----	----	----
253	ISO4406	0.00	0.44	1.12	----	----	----							
323		----	----	----	----	----	----							
333		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
334		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
335		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
360		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
372	ISO4406	0.00	0.44	-0.14	----	----	----							
445		----	----	----	----	----	----							
447	ISO4406	0.00	-1.52	-1.40	----	----	----							
824		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
904		----	----	----	----	----	----							
962		----	----	----	----	----	----							
963		----	----	----	----	----	----							
974		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
1023		----	----	----	----	----	----							
1039		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
1059	ISO4406	0.00	0.44	-0.14	----	----	----							
1062		----	----	----	----	----	----	ISO4406	-2.39	0.00	-2.72	----	----	----
1064		----	----	----	----	----	----							
1095		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
1097	ISO4406	0.00	0.44	1.12	----	----	----							
1105	ISO4406	-2.75	-1.52	-1.40	----	----	----							
1109		----	----	----	----	----	----		0.34	0.00	-0.68	----	----	----
1299		----	----	----	----	----	----	ISO4406	-2.39	0.00	-0.68	----	----	----
1376		----	----	----	----	----	----							
1498	ISO4406	-5.49	-5.44	-3.92	----	----	----							
1528		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
1585		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
1587	ISO4406	0.00	0.44	-0.14	----	----	----							
1610		----	----	----	----	----	----							
1631		----	----	----	----	----	----	ISO4406	0.34	0.00	1.36	----	----	----
1634		----	----	----	----	----	----							
1710		----	----	----	----	----	----	ISO4406	0.34	0.00	-0.68	----	----	----
1720		----	----	----	----	----	----							
1724		----	----	----	----	----	----	ISO4406	-2.39	0.00	1.36	----	----	----
1776	ISO4406	0.00	0.44	-0.14	----	----	----							
1810		----	----	----	----	----	----							
1811		----	----	----	----	----	----							
2130	ISO4406	0.00	0.44	1.12	----	----	----							
6075		----	----	----	----	----	----		0.34	0.00	-0.68	----	----	----
6101		----	----	----	----	----	----							
6108		----	----	----	----	----	----							

APPENDIX 3

Equipment used in Particle Size distribution

lab	Equipment	Test Method based on equipment	Test Method reported	Calibration method reported	Remark
140					
150	Stanhope-Seta	IP565	IP565		
171					
230	Parker Hannifin	IP564	IP564	ISO11171	
237	Stanhope-Seta	IP565	IP565	ISO11171	
253	Parker Hannifin	IP564	IP564		
323					
333	Stanhope-Seta	IP565	IP565	ISO11171	
334	Stanhope-Seta	IP565	IP565		
335	Stanhope-Seta	IP565	IP565	ISO11171	
360	Stanhope-Seta	IP565	IP565	ISO11171	
372	Parker Hannifin	IP564	IP564	ISO11171	
445					
447	Parker Hannifin	IP564	IP564	ISO11171	
824	Stanhope-Seta	IP565	IP565		
904					
962					
963					
974	Stanhope-Seta	IP565	IP565	ISO11171	
1023	Pamas	IP577	IP564	ISO11171	test method not in accordance with equipment
1039	Stanhope-Seta	IP565	IP565		
1059	Parker Hannifin	IP564	IP564	ISO11171	
1062	Stanhope-Seta	IP565	IP565	ISO11171	
1064	Parker Hannifin	IP564	IP564		
1095	Stanhope-Seta	IP565	IP565		
1097	Parker Hannifin	IP564	IP564	ISO11171	
1105	Parker Hannifin	IP564	IP564	ISO11171	
1109	Stanhope-Seta	IP565	IP565	ISO11171	
1299	Pamas	IP577	IP577	ISO11171	
1376					
1498	Parker Hannifin	IP564	IP564	ISO11171	
1528	Stanhope-Seta	IP565	IP565	ISO11171	
1585	Stanhope-Seta	IP565	IP565	ISO11171	
1587	Parker Hannifin	IP564	IP564	ISO11171	
1610					
1631	Stanhope-Seta	IP565	IP565	ISO11171	
1634					
1710	Stanhope-Seta	IP565	IP565	ISO11171	
1720	Stanhope-Seta	IP565	IP565	ISO11171	
1724	Stanhope-Seta	IP565	IP565	ISO11171	
1776	Parker Hannifin	IP564	IP564		
1810	Stanhope-Seta	IP565	IP565	ISO11171	
1811	Stanhope-Seta	IP565	IP565	ISO11171	
2130	Parker Hannifin	IP564	IP564	ISO11171	
6075	Stanhope-Seta	IP565	IP565	ISO11171	
6101					
6108					

APPENDIX 4**Number of participants in iis17J01+ iis17J01PS per country**

2 labs in AFGHANISTAN
1 lab in AUSTRALIA
1 lab in AZERBAIJAN
4 labs in BELGIUM
4 labs in BULGARIA
1 lab in CHILE
1 lab in CHINA, People's Republic
1 lab in CROATIA
1 lab in CYPRUS
2 labs in CZECH REPUBLIC
1 lab in DJIBOUTI
1 lab in EGYPT
1 lab in ESTONIA
2 labs in FINLAND
5 labs in FRANCE
1 lab in GEORGIA
1 lab in GERMANY
3 labs in GREECE
1 lab in GUAM
2 labs in HUNGARY
1 lab in IRELAND
3 labs in ITALY
3 labs in LEBANON
3 labs in MALAYSIA
1 lab in MALTA
1 lab in MARTINIQUE
1 lab in MAURITIUS
1 lab in MOROCCO
5 labs in NETHERLANDS
2 labs in NIGERIA
3 labs in NORWAY
1 lab in PHILIPPINES
1 lab in POLAND
3 labs in PORTUGAL
1 lab in QATAR
1 lab in ROMANIA
1 lab in RUSSIAN FEDERATION
2 labs in SAUDI ARABIA
1 lab in SERBIA
2 labs in SLOVENIA
2 labs in SOUTH AFRICA
2 labs in SOUTH KOREA
1 lab in SPAIN
1 lab in SUDAN
4 labs in SWEDEN
1 lab in TAIWAN
1 lab in THAILAND
1 lab in TOGO
4 labs in TURKEY
2 labs in UNITED ARAB EMIRATES
7 labs in UNITED KINGDOM
12 labs in UNITED STATES OF AMERICA
1 lab in URUGUAY
1 lab in VIETNAM

APPENDIX 5

Abbreviations:

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= probably an error in calculations
U	= test result probably reported in a different unit
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
SDS	= Safety Data Sheet

Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, March 2017
- 2 Defence Standard 91-091, Issue 9, Publication date 3rd of October 2016.
- 3 Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS), Issue 29 Oct 2016, Bulletin No. 96.
- 4 Joel Schmitgal and Jill Bramer, Field Evaluation of Particle Counter Technology for Aviation Fuel Contamination Detection, US Army TARDEC, Technical Report 23966, (June 2013)
- 5 ASTM E178:02
- 6 ASTM E1301:03
- 7 ISO 13528:05
- 8 ISO 5725, parts 1-6, 1994
- 9 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 10 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 11 IP 367:84
- 12 DIN 38402 T41/42
- 13 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 14 J.N. Miller, Analyst, 118, 455, (1993)
- 15 Analytical Methods Committee Technical Brief, No. 4 January 2001
- 16 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 17 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, *Technometrics*, 25(2), 165-172, (1983).