

Results of Proficiency Test
Fuel Oil
June 2018

Organised by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: A. Lewinska, MSc.
Correctors: ing. A.S. Noordman – de Neef & M. Meijer
Report: iis18F01

October 2018

CONTENTS

1	INTRODUCTION	3
2	SET UP	3
2.1	ACCREDITATION	3
2.2	PROTOCOL	3
2.3	CONFIDENTIALITY STATEMENT	3
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES	5
2.6	ANALYSES	5
3	RESULTS	6
3.1	STATISTICS	6
3.2	GRAPHICS	7
3.3	Z-SCORES	7
4	EVALUATION	8
4.1	EVALUATION PER SAMPLE AND PER TEST	8
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	13
4.3	COMPARISON OF THE PROFICIENCY TEST OF JUNE 2018 WITH PREVIOUS PTs	14

Appendices:

1.	Data, statistical results and graphic results	16
2.	Number of participants per country	86
3.	Abbreviations and literature	87

1 INTRODUCTION

Since 1994 the Institute for Interlaboratory Studies (iis) organizes a proficiency test for Fuel Oil every year and twice per year since 2016. During the annual proficiency testing program of 2017/2018, it was decided to continue twice per year with the round robin for the analyses of Fuel Oil in accordance with the latest applicable version of the specifications ISO 8217 and ASTM D396.

In this round robin with regular Fuel Oil 151 laboratories in 59 different countries registered for participation. In the round robin for Metals in Fuel Oil 98 laboratories in 48 different countries registered for participation. See appendix 2 for the number of participants per country for both rounds. In this report, the test results of the June 2018 interlaboratory study on Fuel Oil are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET-UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyses for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC 17025 accredited laboratory. Depending on the registration it was decided to send one bottle of 1L Fuel Oil (labelled #18105) and/or one bottle of 0.1L Fuel Oil (labelled #18106) specifically prepared for metal determinations.

Participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC 17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by means of questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

Approximately 200 litre Fuel Oil was obtained from a supplier in Germany. After heating to 60°C and homogenisation of this batch 188 amber glass bottles of 1L were filled and labelled #18105. The homogeneity of these subsamples was checked by determination of Density at 15°C in accordance with ISO12185 on 12 stratified randomly selected samples.

	Density at 15°C in kg/m ³
Sample #18105-1	987.0
Sample #18105-2	987.1
Sample #18105-3	987.0
Sample #18105-4	987.0
Sample #18105-5	987.1
Sample #18105-6	987.0
Sample #18105-7	987.0
Sample #18105-8	987.0
Sample #18105-9	986.9
Sample #18105-10	986.9
Sample #18105-11	987.0
Sample #18105-12	987.0

Table 1: homogeneity test results of subsamples #18105

From the above test results the repeatability was calculated and compared with 0.3 times the corresponding reproducibility of the reference test method in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	Density at 15°C in kg/m ³
r (observed)	0.17
reference test method	ISO12185:96
0.3 * R (ref. test method)	0.45

Table 2: evaluation of the repeatability of test results of subsamples #18105

The calculated repeatability for Density was in agreement with 0.3 times the corresponding reproducibility of the reference test method. Therefore, homogeneity of the subsamples of #18105 was assumed.

For sample #18106 another batch of Fuel Oil obtained from a local supplier was used. After homogenisation 128 plastic PE bottles of 0.1L were filled and labelled #18106. The homogeneity of the subsamples was checked by determination of Nickel and Vanadium in accordance with IP501 on 7 stratified randomly selected samples.

	Nickel in mg/kg	Vanadium in mg/kg
Sample #18106-1	39	112
Sample #18106-2	37	109
Sample #18106-3	37	107
Sample #18106-4	36	111
Sample #18106-5	36	112
Sample #18106-6	38	118
Sample #18106-7	39	118

Table 3: homogeneity test results of subsamples #18106

From above test results the repeatabilities were calculated and compared with 0.3 times the corresponding reproducibilities of the reference test methods in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	Nickel in mg/kg	Vanadium in mg/kg
r (observed)	3.6	11.8
reference test method	IP470:05	IP470:05
0.3 * R (ref. test method)	5.1	10.4

Table 4: evaluation of the repeatabilities of subsamples #18106

The calculated repeatability for Nickel was in agreement with 0.3 times the reproducibility of the reference test method and the calculated repeatability for Vanadium was close to 3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples of #18106 was assumed.

Depending on the registration of the participant; one bottle of 1L Fuel Oil, labelled #18105 and/or one bottle of 0.1L Fuel Oil, labelled #18106 were sent to the participating laboratories on May 30, 2018. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of Fuel Oil, packed in the amber glass and PE bottles was checked. The material has been found sufficiently stable for the period of the proficiency test.

2.6 ANALYSES

The participants were requested to determine the following parameters:

On sample #18105: Acid Number (total), API Gravity, Ash Content, Asphaltenes, Calculated Carbon Aromaticity Index (CCAI), Carbon Residue – Micro Method, Conradson Carbon Residue, Density at 15°C, Flash Point PMcc, Heat of Combustion (Gross and Net), Kinematic Viscosity (at 50°C and 100°C), Viscosity Stabinger (at 50°C and 100°C), Nitrogen, Pour Point (Lower, Upper and Automated), Sediment by Extraction, Total Sediment (Hot filtration, Accelerated and Potential), Total Sulphur, Water by Distillation, Water and Sediment, Distillation (IBP, 5% - 50% recovered and FBP) and Total Carbon, Hydrogen and Nitrogen (CHN-analyzer).

On sample #18106: Aluminium, Silicon, Sum of Aluminium and Silicon, Iron, Nickel, Sodium, Vanadium, Calcium, Zinc and Phosphorus content.

It was explicitly requested to treat the samples as if they were routine samples and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalysis). Additional or corrected test results are used for data analysis and original test results are placed under 'Remarks' in the test result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organization, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4).

For the statistical evaluation, the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the results of the statistical evaluation should be used with due care.

According to ISO 5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1 was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle. Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve was projected over the Kernel Density Graph for reference.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM or IP reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation of this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other targets values were used. In some cases a reproducibility based on former iis proficiency tests could be used.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z(\text{target}) = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The z(target) scores are listed in the result tables of appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this proficiency test several problems were encountered with the dispatch of the samples. Four participants did not report any test results for sample #18105 and five participants reported the test results after the final reporting date. Eighteen participants did not report any test results for sample #18106 and six participants reported the test results after the final reporting date. Not all laboratories were able to report all analyses requested. Finally, 149 participants reported in total 2631 numerical test results. Observed were 88 statistically outlying test results, which is 3.3 %. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

Not all data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

4.1 EVALUATION PER SAMPLE AND PER TEST

In this section, the reported test results are discussed per sample and per test. The test methods, which are used by the various laboratories, were taken into account for explaining the observed differences when possible and applicable. These test methods are also listed in the tables together with the reported test results. The abbreviations, used in these tables, are listed in appendix 3.

In the iis PT reports, ASTM test methods are referred to with a number (e.g. D1298) and an added designation for the year that the test method was adopted or revised (e.g. D1298:12b). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D1298:12b(2017)). In the test results tables of appendix 1 only the test method number and year of adoption or revision (e.g. D1298:12b) are used.

Sample #18105:

Acid Number: This determination was problematic depending on used requirements of the type of end point and titration volume. Six statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D664-A:17a mode BEP at both titration volumes (60mL and 125mL) but in agreement with inflection point at both titration volumes.

API Gravity: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D1298:12b(2017).

- Ash Content:** This determination was problematic at an ash content of 0.02 %M/M. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with requirements of ISO6245:01 and ASTM D482:13.
When evaluated separately over method ISO6245:01 and ASTM D482:13 the calculated reproducibility after rejection of the statistical outliers was still not in agreement for both methods.
- Asphaltenes:** This determination was problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of IP143:04.
- Calculated Carbon Aromaticity Index:** This determination was not problematic. Six statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO8217:17.
- Carbon Residue Micro Method:** This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO10370:14.
- Conradson Carbon Residue:** This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of ASTM D189:06(2014).
- Density at 15°C:** This determination was not problematic. Nine statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.
- Flash Point PMcc:** This determination was problematic. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ISO2719-B:16.
- HOC Gross:** This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D240:17.
- HOC Net:** This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D240:17.
- Kin. Visc. at 50°C:** This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO3104:94.
- Kin. Visc. at 100°C:** This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ISO3104:94.

Vis Stab.at 50°C: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D7042:16e3.

Vis Stab.at 100°C: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D7042:16e3.

Nitrogen: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of ASTM D5762:18. When the test results of ASTM D5762 volumetric and gravimetric test methods were evaluated separately, the calculated reproducibility over the volumetric test results was much smaller than the calculated reproducibility over the gravimetric test results. The calculated reproducibility over the volumetric test results is in agreement with the requirements of ASTM D5762:18.

Pour Point Lower: This determination was not problematic. No statistical outliers were observed, but one test result was excluded. The calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ISO3016:94.

Pour Point Upper: This determination was problematic. One statistical outlier was observed and one other test result was excluded. The calculated reproducibility after rejection of the suspect data is not in agreement with ISO3016:94. Rounding to 3 degrees acc. ISO3016:94 may (partly) explain the large variation.

Pour Point Automated: This determination was problematic. No statistical outliers were observed but two test results were excluded. However, the calculated reproducibility after rejection of the suspect data is not in agreement with ASTM D5950:14.
The large variation may (partly) be explained by possible problems with the detector sensitivity or by not following the test method properly (see Note 8 in ASTM D5950:14).

Sediment by Extraction: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D473:07e1(2017).

Sediment (Hot Filtration): This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of IP375:11.

Total Sediment (Accelerated): This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with IP390:11. IP390:11 is identical to ISO10307-2:09 and technically equivalent to ASTM D4870.

Total Sediment (Potential): This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with IP390:11. IP390:11 is identical to ISO10307-2:09 and technically equivalent to ASTM D4870.

Total Sulphur: This determination may be problematic dependent on the requirements of the test method used. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ISO 8754:03, but not with the more strict requirements of ASTM D4294:16e1.

Water by Distillation: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ISO3733:99 or ASTM D95:13e1.

Water and Sediment: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D1796:11(2016).

Vacuum Distillation: This determination may be problematic for 5%, 10% and 50% recovered and FBP. In total six statistical outliers were observed and five other test results were excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement for IBP, 20%, 30% and 40% recovered with the requirements of ASTM D1160:18. The calculated reproducibilities for 5%, 10% and 50% recovered and FBP are not in agreement with the requirements of ASTM D1160:18.

CHN-Analyzer: This determination was not problematic for Carbon, Hydrogen and Nitrogen. In total one statistical outlier was observed. The calculated reproducibilities after rejection of the statistical outlier are in agreement with the requirements of ASTM D5291-ABC:16.

Sample #18106:

Aluminum: This determination was problematic. Six statistical outliers were observed and one other test result was excluded. The calculated reproducibility after rejection of the suspect data is not in agreement with the reproducibility of IP470:05 and IP501:05.

Silicon: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the reproducibility of IP470:05 and with the more strict requirements of IP501:05.

Total Al/Si: This determination was problematic. Two statistical outliers were observed and five other test results were excluded. The calculated reproducibility after rejection of the suspect data is not in agreement with the reproducibility of IP470:05 and with the more strict requirements of IP501:05.

- Iron: This determination may not be problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in full agreement with the reproducibility of IP470:05 but not in agreement with the more strict requirements of IP501:05.
- Nickel: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the reproducibilities of IP470:05 and IP501:05.
- Sodium: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the reproducibility of IP470:05 and with the more strict requirements of IP501:05.
- Vanadium: This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the reproducibilities of IP470:05 and IP501:05.
- Calcium: This determination was problematic. Two statistical outliers were observed and one other test result was excluded. The calculated reproducibility after rejection of the suspect data is not in agreement with the requirements of IP470:05 and with the more strict requirements of IP501:05.
- Zinc: This determination was problematic. Four statistical outliers were observed and one other test result was excluded. The calculated reproducibility after rejection of the suspect data is not in agreement with the reproducibilities of IP470:05 or IP501:05.
- Phosphorus: This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the reproducibility of IP500:03 or IP501:05.

Finally, it should be noted that proper attention for homogenization is crucial for a material such as Fuel Oil. Due to the nature of the material it is very susceptible to problems when not handled correctly. Practically most test methods for the determination of metals in Fuel Oil have similar statements regarding homogenization. Recommended is the use of a quality control fuel oil with known amounts of metals like Al, Fe, Si and V. This control standard may be of use to detect deviations in metals with respect to the preparation steps.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant reference test method and the reproducibility as found for the group of participating laboratories. The average results, calculated reproducibilities and reproducibilities derived from literature reference test methods (in casu ASTM and IP reference test methods), are compared in the next tables:

Parameters	unit	n	average	2.8 * sd	R (lit)
Acid Number (Total)	mg KOH/g	72	0.20	0.11	0.09
API Gravity		75	11.7	0.3	0.5
Ash Content	%M/M	100	0.020	0.008	0.005
Asphaltenes	%M/M	56	3.27	1.13	0.65
Calc. Carbon Aromaticity Index		52	851.3	1.2	2.4
Carbon Residue, Micro Method	%M/M	88	9.66	0.82	1.11
Conradson Carbon Residue	%M/M	32	10.00	1.53	1.60
Density at 15°C	kg/m ³	123	987.3	1.3	1.5
Flash Point PMcc	°C	118	121.7	7.4	6
Heat of Combustion, Gross	MJ/kg	58	42.99	0.37	0.40
Heat of Combustion, Net	MJ/kg	47	40.68	0.41	0.40
Kinematic Viscosity at 50°C	mm ² /s	113	286.5	16.1	21.2
Kinematic Viscosity at 100°C	mm ² /s	85	27.1	1.1	1.4
Viscosity Stabinger at 50°C	mm ² /s	12	287.5	11.4	29.6
Viscosity Stabinger at 100°C	mm ² /s	12	27.0	0.7	1.9
Nitrogen	mg/kg	28	2999	1044	798
Pour Point, Lower	°C	47	4.2	7.0	6.6
Pour Point, Upper	°C	87	6.7	8.4	6.6
Pour Point (automated), Δ3°C	°C	28	3.4	11.4	6.1
Sediment by Extraction	%M/M	75	0.016	0.021	0.037
Total Sediment (Hot Filtration)	%M/M	52	0.014	0.012	0.035
Total Sediment (Accelerated)	%M/M	54	0.014	0.011	0.034
Total Sediment (Potential)	%M/M	54	0.014	0.013	0.034
Total Sulphur	%M/M	130	0.95	0.08	0.09
Water by Distillation	%V/V	74	0.04	0.09	0.2
Water and Sediment	%V/V	39	0.05	0.07	0.1
Distillation at 10mmHg calculated to 760 mmHg					
Initial Boiling Point	°C	15	214.0	35.4	49
5% recovered	°C	16	273.6	28.6	24.2
10% recovered	°C	17	317.6	23.0	20.1
20% recovered	°C	17	371.6	14.0	17.1
30% recovered	°C	17	406.6	12.7	15.8
40% recovered	°C	17	445.1	11.8	17.4
50% recovered	°C	16	498.8	25.5	18.5
Final Boiling Point	°C	12	520.6	40.7	27
CHN analyser					
Total Carbon	%M/M	22	87.8	2.1	2.5
Total Hydrogen	%M/M	22	10.6	0.5	0.8
Total Nitrogen	%M/M	18	0.34	0.16	0.45

Table 5: comparison of the observed and target reproducibilities of the sample #18105

Parameters	unit	n	average	2.8 * sd	R (lit)
Aluminium as Al	mg/kg	61	4.5	2.8	2.2
Silicon as Si	mg/kg	60	6.8	6.6	5.0
Total Aluminium + Silicon	mg/kg	53	11.2	7.0	5.5
Iron as Fe	mg/kg	69	37.3	16.1	16.5
Nickel as Ni	mg/kg	73	35.5	20.0	16.6
Sodium as Na	mg/kg	76	17.5	8.6	7.3
Vanadium as V	mg/kg	74	121.4	23.0	35.9
Calcium as Ca	mg/kg	68	4.6	4.7	4.1
Zinc as Zn	mg/kg	56	1.33	1.3	0.7
Phosphorus as P	mg/kg	36	0.6	1.0	0.9

Table 6: comparison of the observed and target reproducibilities of the sample #18106

Without further statistical calculations it can be concluded that for a number tests there is a good compliance of the group of participating laboratories with the relevant reference test method. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF JUNE 2018 WITH PREVIOUS PTS

	June 2018	December 2017	June 2017	December 2016	January 2016
Number of reporting labs	149	143	169	83	226
Number of test results	2631	3081	3119	1936	4787
Statistical outliers	88	133	102	72	115
Percentage outliers	3.3%	4.3%	3.3%	3.7%	2.4%

Table 7: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective reference test methods. The conclusions are given the following table:

Determination	June 2018	December 2017	June 2017	December 2016	January 2016
Acid Number (Total)	-	++	+	++	+
API Gravity	+	++	++	++	++
Ash Content	-	--	--	--	-
Asphaltenes	-	+/-	-	--	+/-
Calc. Carbon Aromaticity Index	++	+	+	+	+
Carbon Residue, Micro Method	+	+	+/-	+	+
Conradson Carbon Residue	+/-	+	+	+/-	+
Density at 15°C	+	+/-	+/-	+	+/-
Flash Point PMcc	-	-	+/-	+/-	-
Heat of Combustion Gross	+/-	+/-	+	++	-
Heat of Combustion Net	+/-	+/-	+	++	-
Kinematic Viscosity at 50°C	+	+	+/-	++	+
Kinematic Viscosity at 100°C	+	-	+	-	+/-
Viscosity Stabinger at 50°C	++	++	+	+	n.e.
Viscosity Stabinger at 100°C	++	++	++	++	n.e.
Nitrogen	-	-	-	-	-

Determination	June 2018	December 2017	June 2017	December 2016	January 2016
Pour Point Lower	+/-	-	-	-	-
Pour Point Upper	-	-	-	-	-
Pour Point (automated), $\Delta 3^{\circ}\text{C}$	-	-	-	--	-
Sediment by Extraction	+	+	+	++	++
Total Sediment (Hot Filtration)	++	++	++	++	++
Total Sediment (Accelerated)	++	++	++	++	++
Total Sediment (Potential)	++	++	++	++	++
Total Sulphur	+/-	+/-	+	+/-	+
Water by Distillation	++	++	++	++	++
Water and Sediment	+	+	+/-	++	+
Distillation at 10mmHg to AET	+/-	+	+/-	-	+/-
Total Carbon	+	+	+	+	+
Total Hydrogen	+	+	+	+	+
Total Nitrogen	++	-	++	+/-	-
Aluminium as Al	-	+	+/-	n.e.	n.e.
Silicon as Si	-	+	+	n.e.	n.e.
Total Aluminium + Silicon	-	+	+/-	n.e.	n.e.
Iron as Fe	+/-	++	+	+	+
Nickel as Ni	-	+	++	+	+
Sodium as Na	-	+	+	+/-	+/-
Vanadium as V	+	++	++	+	+
Calcium as Ca	-	+/-	+	+/-	-
Zinc as Zn	-	+/-	+	+	-
Phosphorus as P	+/-	-	+	+	+

Table 8: comparison determinations against the reference test methods

The performance of the determinations against the requirements of the respective reference test methods is listed in the above table. The following performance categories were used:

- ++: group performed much better than the reference test method
- + : group performed better than the reference test method
- +/-: group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e.: not evaluated



APPENDIX 1

Determination of Acid Number (Total) on sample #18105; results in mg KOH/g

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913	D664-A	0.18		-0.75
92		----		----	922	D664-A	0.189		-0.47
120	D664-A	0.392	R(1)	5.85	962		----		----
131	D664-A	0.034	R(1)	-5.30	963	D664-A	0.21		0.18
132	D664-A	0.26		1.74	971	ISO6619	0.19		-0.44
140	D664-B	0.1905		-0.43	974	D664-A	0.20		-0.13
150	D664-A	0.23		0.80	982		----		----
154		----		----	1006		----		----
158	D664-A	0.251		1.46	1011	D664-A	0.210		0.18
159	D664-A	0.225		0.65	1019		----		----
168		----		----	1059	ISO6619	0.17		-1.07
169		----		----	1095	D664-A	0.193		-0.35
171	D664-A	0.14		-2.00	1107	D664-A	0.181		-0.72
175		----		----	1109	D664-A	0.16		-1.38
194		----		----	1121	D664-A	0.21		0.18
212	D664-A	0.21		0.18	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	D664-A	0.592	R(1)	12.09
225		----		----	1161	D664-A	0.202		-0.07
228		----		----	1167		----		----
237	D664-A	0.189		-0.47	1177		----		----
238		----		----	1233	D664-A	0.21		0.18
253		----		----	1259	D664-A	0.228		0.74
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP177	0.18		-0.75
323	D664-A	0.16		-1.38	1299	D664-B	0.19		-0.44
331	D664Mod.	0.17		-1.07	1345	D664-A	0.185		-0.60
333		----		----	1356	D664-A	0.02	R(1)	-5.74
334		----		----	1367	D664-A	0.16		-1.38
337		----		----	1402	IP177	0.20		-0.13
339		----		----	1404	D664-A	0.105		-3.09
342	D664-A	0.25		1.43	1428	D664-A	0.24		1.12
343		----		----	1459		----		----
349		----		----	1510		----		----
356	D664-A	0.21		0.18	1556	D664-A	0.145		-1.84
371		----		----	1569	D664-A	0.16		-1.38
391	D664-A	0.287		2.58	1586	D664-A	0.20		-0.13
398		----		----	1613	D664-A	0.16		-1.38
399		----		----	1631		----		----
440		----		----	1643	D664-A	0.196		-0.25
444		----		----	1650		----		----
463	D664-A	0.213		0.27	1677	D664-A	0.18		-0.75
511		----		----	1720	D664-A	0.14		-2.00
529		----		----	1724	D664-A	0.257		1.65
541		----		----	1728		----		----
557	D664-A	0.17873		-0.79	1740	D664-A	0.18		-0.75
562		----		----	1796	D664-A	0.1892		-0.47
575	D664-A	0.32		3.61	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605	D664-A	0.27		2.05	1833	D664-A	0.274		2.18
631	D664-A	0.289		2.64	1849		----		----
633	D664-A	0.22		0.49	1906		----		----
663	D664-A	0.215		0.34	1936		----		----
671		----		----	1937		----		----
750	D664-A	0.16		-1.38	1938		----		----
753		----		----	1950	D664-A	0.183		-0.66
759		----		----	1956		----		----
781	D664-A	0.21		0.18	1964		----		----
785		----		----	1967	D664-A	0.247		1.33
824	D664-A	0.214		0.31	1995	D664-A	0.8	C,R(1)	18.57
825		----		----	6016		----		----
850	D664-A	0.20		-0.13	6026	D664-A	0.2084		0.13
851		----		----	6028	D664-A	0.65	C,R(1)	13.89
855	D664-A	0.22		0.49	6039		----		----
858	D664-A	0.19		-0.44	6044	D664-A	0.162		-1.31
862	D664-A	0.23		0.80	6054		----		----
863	D664-A	0.20		-0.13	6057	D664-A	0.17		-1.07
864	D664-A	0.2		-0.13	6075		----		----
865	D664-A	0.172		-1.00	6092		----		----
866	D664-A	0.215		0.34	6142	D664-A	0.235		0.96
870	D664-A	0.21		0.18	6180	D664-A	0.282		2.43
875	D664-A	0.24		1.12	6192		----		----
886		----		----					

normality	OK
n	72
outliers	6
mean (n)	0.2042
st.dev. (n)	0.03892
R(calc.)	0.1090
st.dev. (D664-A:17a BEP-125mL)	0.03209
R(D664-A:17a BEP-125mL)	0.0898
Compare	
R(D664-A:17a Inflection-125mL)	0.1698
R(D664-A:17a BEP-60mL)	0.0958
R(D664-A:17a Inflection-60mL)	0.1093

Lab 1995 first reported 0.4
 Lab 6028 first reported 0.4





Determination of API Gravity on sample #18105

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62	D4052	11.7		-0.13	912	D287	11.60		-0.69
90	D4052	11.74		0.09	913		----		----
92	D4052	11.86		0.76	922	D1298	11.8		0.43
120	ISO12185	11.7		-0.13	962		----		----
131	D4052	11.73		0.03	963	ISO12185	11.7		-0.13
132		----		----	971	ISO12185	11.75		0.15
140	D4052	11.7		-0.13	974	D4052	11.72		-0.02
150	D287	11.9		0.99	982	D1298	11.72		-0.02
154	D4052	11.7		-0.13	1006		----		----
158	D4052	11.71		-0.08	1011		----		----
159	D4052	11.8		0.43	1019		----		----
168	D287	12.0		1.55	1059		----		----
169	D1298	11.9		0.99	1095		----		----
171	D4052	11.7		-0.13	1107		----		----
175	D4052	11.75		0.15	1109	D287	11.69		-0.19
194		----		----	1121	D4052	11.47		-1.42
212	ISO12185	11.7	C	-0.13	1126		----		----
221		----		----	1134	D4052	11.73		0.03
224	D1298	11.89	C	0.93	1135	D1298	11.72		-0.02
225	D4052	11.8		0.43	1161	D1298	11.9		0.99
228		----		----	1167		----		----
237	D1298	11.59		-0.75	1177		----		----
238	D4052	11.5		-1.25	1233		----		----
253	D4052	11.89		0.93	1259	Calculation	11.52		-1.14
254		----		----	1266	D1298	11.4		-1.81
273		----		----	1269		----		----
313		----		----	1275		----		----
323	D1298	11.70		-0.13	1299		----		----
331	ISO12185	11.7		-0.13	1345	D4052	11.75		0.15
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402	D4052	11.69		-0.19
339		----		----	1404		----		----
342		----		----	1428		----		----
343		----		----	1459		----		----
349		----		----	1510		----		----
356	D4052	11.74		0.09	1556		----		----
371	D1298	11.73		0.03	1569		----		----
391	ISO12185	11.73		0.03	1586	D1298	11.78		0.31
398		----		----	1613	D4052	11.68		-0.25
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	D1298	11.76		0.20	1677		----		----
511		----		----	1720		----		----
529		----		----	1724	D4052	11.78		0.31
541	D4052	11.40		-1.81	1728	D287	11.832		0.61
557	D4052	11.68666		-0.21	1740		----		----
562	D1298	12.4714	R(0.01)	4.19	1796	D1250	11.742		0.10
575	D1298	11.95		1.27	1807		----		----
603		----		----	1811		----		----
604	D4052	11.66		-0.36	1832		----		----
605	D4052	11.64		-0.47	1833	ISO12185	11.69		-0.19
631	D1298	11.37		-1.98	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671	D287	11.69		-0.19	1937		----		----
750	D1298	11.7		-0.13	1938		----		----
753	ISO12185	11.67		-0.30	1950	D1298	11.74		0.09
759	D1298	11.765		0.23	1956		----		----
781	ISO12185	11.65		-0.41	1964		----		----
785		----		----	1967	D1298	11.75		0.15
824	D4052	11.78		0.31	1995		----		----
825	D4052	11.76		0.20	6016		----		----
850		----		----	6026	D1298	11.71		-0.08
851		----		----	6028	D1298	11.79		0.37
855	D1298	11.80		0.43	6039		----		----
858	D1298	11.8		0.43	6044	D1298	11.70		-0.13
862	ISO12185	11.77		0.26	6054		----		----
863	ISO12185	11.76		0.20	6057	D1298	11.78		0.31
864	ISO12185	11.77		0.26	6075		----		----
865	D1250	11.72		-0.02	6092		----		----
866	D1298	11.8		0.43	6142		----		----
870	D1298	11.72		-0.02	6180		----		----
875	D1298	11.68		-0.25	6192		----		----
886		----		----					

normality	not OK
n	75
outliers	1
mean (n)	11.724
st.dev. (n)	0.1122
R(calc.)	0.314
st.dev.(D1298:12b)	0.1786
R(D1298:12b)	0.5

Lab 212 first reported 13.4
 Lab 224 first reported 12.21





Determination of Ash Content on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D482	0.0169		-1.43
90	D482	0.0253		3.27	913		----		----
92	D482	0.017		-1.38	922	D482	0.0205		0.58
120	ISO6245	0.020		0.30	962		----		----
131	D482	0.019		-0.26	963	ISO6245	0.021		0.86
132	ISO6245	0.022		1.42	971	ISO6245	0.0204		0.53
140	ISO6245	0.006	R(0.01)	-7.54	974	D482	0.020		0.30
150	D482	0.021	C	0.86	982	D482	0.020		0.30
154	D482	0.019		-0.26	1006		----		----
158		----		----	1011		----		----
159	D482	0.020		0.30	1019	ISO6245	0.0142		-2.95
168	D482	0.016		-1.94	1059	ISO6245	0.015		-2.50
169	D482	0.0200		0.30	1095	ISO6245	0.018		-0.82
171		----		----	1107		----		----
175		----		----	1109	D482	0.0176		-1.04
194	D482	0.020		0.30	1121	ISO6245	0.016		-1.94
212	ISO6245	0.019		-0.26	1126		----		----
221	D482	0.020		0.30	1134	IP4	0.0168		-1.49
224	D482	0.025		3.10	1135	ISO6245	0.0216		1.20
225	D482	0.024		2.54	1161	ISO6245	0.017		-1.38
228	D482	0.01296	C	-3.64	1167	ISO6245	0.01984		0.21
237		----		----	1177		----		----
238		----		----	1233	ISO6245	0.034	R(0.01)	8.14
253		----		----	1259	ISO6245	0.01844		-0.57
254	D482	0.018		-0.82	1266	ISO6245	0.018		-0.82
273	D482	0.022		1.42	1269		----		----
313		----		----	1275	IP4	0.0200057		0.31
323	ISO6245	0.018		-0.82	1299	D482	0.018		-0.82
331		----		----	1345	D482	0.0185		-0.54
333		----		----	1356	ISO6245	0.018		-0.82
334		----		----	1367	IP4	0.018	C	-0.82
337		----		----	1402	IP4	0.020		0.30
339		----		----	1404		----		----
342	ISO6245	0.019		-0.26	1428	ISO6245	0.018		-0.82
343	ISO6245	0.022		1.42	1459		----		----
349		----		----	1510		----		----
356	ISO6245	0.019		-0.26	1556	ISO6245	0.0193		-0.09
371	D482	0.0187		-0.43	1569	ISO6245	0.0195		0.02
391		----		----	1586	ISO6245	0.016		-1.94
398	D482	0.015		-2.50	1613	D482	0.020		0.30
399		----		----	1631	ISO6245	0.020		0.30
440		----		----	1643	D482	0.01745		-1.13
444		----		----	1650	ISO6245	0.0195		0.02
463	ISO6245	0.0192		-0.15	1677	D482	0.0175		-1.10
511		----		----	1720		----		----
529		----		----	1724		----	W	----
541	D482	0.0178		-0.93	1728	D482	0.0232		2.09
557	D482	0.0129247		-3.66	1740	ISO6245	0.0217		1.25
562	D482	0.0205		0.58	1796	ISO6245	0.01448		-2.79
575		----		----	1807	ISO6245	0.0260		3.66
603	D482	0.020		0.30	1811	ISO6245	0.013		-3.62
604		----		----	1832	ISO6245	0.0220		1.42
605	D482	0.018		-0.82	1833	ISO6245	0.024		2.54
631	D482	0.0219		1.37	1849	ISO6245	0.02	C	0.30
633	D482	0.0233		2.15	1906		----		----
663	D482	0.0219		1.37	1936		----		----
671	D482	0.016		-1.94	1937		----		----
750	D482	0.021		0.86	1938		----		----
753		----		----	1950	ISO6245	0.018		-0.82
759		----		----	1956	ISO6245	0.026	C	3.66
781	ISO6245	0.023		1.98	1964		----		----
785	ISO6245	0.025		3.10	1967	D482	0.0163		-1.77
824	ISO6245	0.02055		0.61	1995	D482	0.02		0.30
825	D482	0.019		-0.26	6016		----		----
850	ISO6245	0.020		0.30	6026	D482	0.0175		-1.10
851	ISO6245	0.020	C	0.30	6028		----		----
855	D482	0.019		-0.26	6039		----		----
858	D482	0.019		-0.26	6044	ISO6245	0.021		0.86
862	D482	0.0197		0.13	6054	D482	0.020049		0.33
863	D482	0.022		1.42	6057	ISO6245	0.025		3.10
864	D482	0.016		-1.94	6075	ISO6245	0.0184		-0.59
865	D482	0.022		1.42	6092	ISO6245	0.015		-2.50
866	ISO6245	0.0196		0.08	6142		----		----
870	D482	0.0201		0.36	6180		----		----
875	D482	0.024		2.54	6192		----		----
886		----		----					

		<u>Only ISO6245 results:</u>	<u>Only D482 results:</u>
normality	OK	OK	OK
n	100	45	51
outliers	2	2	0
mean (n)	0.0195	0.0196	0.0194
st.dev. (n)	0.00280	0.00300	0.00271
R(calc.)	0.0078	0.0084	0.0076
st.dev.(ISO6245:01)	0.00179	0.00179	0.00179
R(ISO6245:01)	0.005	0.005	0.005
Compare			
R(D482:13)	0.005		

Lab 150 first reported 0.012
 Lab 228 first reported 0.029
 Lab 851 first reported 0.012
 Lab 1367 first reported 0.031
 Lab 1724 first reported 0.028
 Lab 1849 first reported 0.03
 Lab 1956 first reported 0.008





Determination of Asphaltenes on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90	IP143	3.16		-0.45	913		----		----
92		----		----	922		----		----
120	D6560	2.923		-1.47	962		----		----
131	D6560	0.145	R(0.01)	-13.38	963	IP143	3.10		-0.71
132		----		----	971	IP143	3.24		-0.11
140	IP143	1.755	R(0.05)	-6.47	974	IP143	3.28		0.06
150		----		----	982		----		----
154	D6560	3.1		-0.71	1006		----		----
158		----		----	1011	IP143	3.4		0.58
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095	IP143	2.5		-3.28
171	IP143	3.0		-1.14	1107	IP143	3.14		-0.54
175		----		----	1109		----		----
194		----		----	1121	IP143	3.75		2.08
212		----		----	1126		----		----
221	D6560	2.9		-1.56	1134	IP143	2.900		-1.56
224		----		----	1135	IP143	3.5		1.01
225	D6560	2.8		-1.99	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259	IP143	4.72	R(0.05)	6.24
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP143	4.3.02		----
323	IP143	3.8		2.29	1299		----		----
331		----		----	1345	IP143	2.19		-4.61
333		----		----	1356	D6560	3.3		0.15
334	IP143	3.8	C	2.29	1367		----		----
337		----		----	1402	IP143	3.6		1.44
339		----		----	1404		3.45		0.79
342	IP143	4.011		3.20	1428		----		----
343	IP143	4.1	C	3.58	1459		----		----
349		----		----	1510		----		----
356		----		----	1556	IP143	6.1	R(0.01)	12.16
371		----		----	1569	IP143	3.5		1.01
391		----		----	1586	IP143	3.2		-0.28
398		----		----	1613	IP143	3.67		1.74
399		----		----	1631		----		----
440		----		----	1643	D6560	3.798		2.29
444		----		----	1650		----		----
463	IP143	3.38		0.49	1677	IP143	3.1056		-0.68
511		----		----	1720		----		----
529		----		----	1724		----		----
541	IP143	3.32		0.24	1728	D6560	3.62		1.52
557	IP143	3.5875926		1.38	1740	IP143	3.2		-0.28
562		----		----	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605	IP143	4.2		4.01	1833	IP143	2.96		-1.31
631	IP143	3.0914		-0.74	1849		----		----
633		----		----	1906		----		----
663	IP143	3.10		-0.71	1936		----		----
671		----		----	1937		----		----
750	IP143	3.4		0.58	1938		----		----
753		----		----	1950	IP143	3.34		0.32
759		----		----	1956		----		----
781	IP143	3.37		0.45	1964		----		----
785		----		----	1967	IP143	3.045		-0.94
824		----		----	1995		----		----
825		----		----	6016		----		----
850	IP143	2.90		-1.56	6026	IP143	3.1187		-0.63
851		----		----	6028		----		----
855	IP143	2.9		-1.56	6039		----		----
858	IP143	2.8		-1.99	6044		----		----
862	IP143	3.05		-0.92	6054	D6560	2.90343479		-1.55
863	IP143	3.0		-1.14	6057	IP143	3.320		0.24
864		----		----	6075		3.95		2.94
865	IP143	2.84		-1.82	6092	IP143	3.2		-0.28
866	IP143	2.80		-1.99	6142		----		----
870	IP143	2.782		-2.07	6180	IP143	3.86		2.55
875	IP143	3.58		1.35	6192		----		----
886		----		----					

normality	OK
n	56
outliers	4
mean (n)	3.265
st.dev. (n)	0.4040
R(calc.)	1.131
st.dev.(IP143:04)	0.2332
R(IP143:04)	0.653

Lab 334 first reported 4.8
 Lab 343 first reported 4.668





Determination of Calculated Carbon Aromaticity Index on sample #18105

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	ISO8217	852		0.88
90		----		----	913		----		----
92		----		----	922	ISO8217	850.4		-1.03
120	ISO8217	851		-0.32	962		----		----
131	ISO8217	851		-0.32	963	ISO8217	851		-0.32
132	ISO8217	853	R(0.05)	2.07	971	ISO8217	851.8		0.64
140		----		----	974	ISO8217	851		-0.32
150	ISO8217	853	R(0.05)	2.07	982	ISO8217	851.073		-0.23
154	ISO8217	851		-0.32	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171	ISO8217	851		-0.32	1107		----		----
175		----		----	1109		----		----
194		----		----	1121	ISO8217	851.9		0.76
212	ISO8217	852		0.88	1126		----		----
221	ISO8217	851		-0.32	1134	ISO8217	851.4		0.16
224		----		----	1135		----		----
225	ISO8217	851.487		0.26	1161		----		----
228	ISO8217	851		-0.32	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254	ISO8217	852		0.88	1266	ISO8217	853.4	R(0.05)	2.54
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331		----		----	1345	ISO8217	851		-0.32
333		----		----	1356		----		----
334		----		----	1367	ISO8217	852		0.88
337		----		----	1402	ISO8217	852		0.88
339		----		----	1404		----		----
342	ISO8217	852		0.88	1428		----		----
343		----		----	1459		----		----
349		----		----	1510		----		----
356	ISO8217	851		-0.32	1556	ISO8217	852		0.88
371	ISO8217	852		0.88	1569		----		----
391	ISO8217	851.6		0.40	1586	ISO8217	851.0		-0.32
398		----		----	1613	ISO8217	850.9		-0.43
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	ISO8217	851.0		-0.32	1677		----		----
511		----		----	1720		----		----
529		----		----	1724	ISO8217	851.0		-0.32
541	ISO8217	853.7	E,R(0.05)	2.90	1728		----		----
557	ISO8217	851.9310		0.79	1740	ISO8217	851		-0.32
562		----		----	1796	ISO8217	851		-0.32
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605	ISO8217	851.6		0.40	1833	ISO8217	853.56	R(0.05)	2.73
631	ISO8217	853.67	R(0.05)	2.87	1849		----		----
633	ISO8217	851.46		0.23	1906		----		----
663	ISO8217	851.9		0.76	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753	ISO8217	851		-0.32	1950	ISO8217	851		-0.32
759		----		----	1956		----		----
781	ISO8217	851		-0.32	1964		----		----
785		----		----	1967	ISO8217	851		-0.32
824	ISO8217	851		-0.32	1995		----		----
825	ISO8217	850.6		-0.79	6016		----		----
850	ISO8217	851.2		-0.08	6026	ISO8217	851		-0.32
851		----		----	6028		----		----
855	ISO8217	851.1		-0.20	6039		----		----
858	ISO8217	851		-0.32	6044		----		----
862	ISO8217	850.9		-0.43	6054		----		----
863	ISO8217	851.1		-0.20	6057	ISO8217	851		-0.32
864	ISO8217	850.9		-0.43	6075		----		----
865	ISO8217	851.2		-0.08	6092		----		----
866	ISO8217	851.2		-0.08	6142		----		----
870	ISO8217	851.12		-0.17	6180		----		----
875		----		----	6192		----		----
886		----		----					

normality	OK
n	52
outliers	6
mean (n)	851.265
st.dev. (n)	0.4320
R(calc.)	1.210
st.dev.(ISO8217:17)	0.8393
R(ISO8217:17)	2.35

Lab 541 probably a calc. error, iis calculated 851.35





Determination of Carbon Residue - micro method on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	ISO10370	9.755		0.23
90		----		----	913	D4530	10.31		1.63
92	D4530	9.09		-1.44	922	D4530	10.1		1.10
120	ISO10370	9.74		0.19	962		----		----
131	D4530	9.61		-0.13	963	ISO10370	9.64		-0.06
132	ISO10370	9.57		-0.23	971	ISO10370	9.60		-0.16
140	ISO10370	9.61		-0.13	974	D4530	9.59		-0.18
150	D4530	9.67		0.02	982		----		----
154	D4530	9.70		0.09	1006	D4530	9.42		-0.61
158		----		----	1011	ISO10370	9.52		-0.36
159	D4530	9.78		0.29	1019		----		----
168		----		----	1059	ISO10370	9.49		-0.44
169		----		----	1095	ISO10370	9.33		-0.84
171	ISO10370	9.81		0.37	1107	D4530	9.32		-0.86
175		----		----	1109	D4530	10.01		0.87
194	D4530	9.81		0.37	1121	ISO10370	9.59		-0.18
212	ISO10370	10.22		1.40	1126		----		----
221		----		----	1134	D4530	9.673		0.03
224		----		----	1135	ISO10370	9.27		-0.99
225	D4530	10.2		1.35	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233	ISO10370	9.315		-0.88
253		----		----	1259	ISO10370	10.313		1.64
254		----		----	1266		----		----
273	D4530	9.43		-0.59	1269		----		----
313		----		----	1275	IP398	9.7007		0.10
323	ISO10370	11.6	R(0.01)	4.88	1299	D4530	9.30		-0.91
331	ISO10370	9.8		0.35	1345		----		----
333		----		----	1356	ISO10370	9.73		0.17
334	ISO10370	9.91		0.62	1367	ISO10370	9.61		-0.13
337		----		----	1402	IP398	9.51		-0.39
339		----		----	1404	ISO10370	9.60		-0.16
342	ISO10370	9.59		-0.18	1428		----		----
343	ISO10370	9.34		-0.81	1459		----		----
349		----		----	1510		----		----
356	ISO10370	9.67		0.02	1556	ISO10370	9.64		-0.06
371		----		----	1569	ISO10370	9.82		0.40
391	ISO10370	9.93		0.67	1586	ISO10370	9.45		-0.54
398		----		----	1613		----	W	----
399	D4530	8.33	R(0.01)	-3.36	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	ISO10370	9.523		-0.35	1677	D4530	9.646		-0.04
511		----		----	1720	D4530	10.4	C	1.86
529		----		----	1724	D4530	9.53		-0.33
541		----		----	1728	D4530	9.45		-0.54
557		----		----	1740	ISO10370	9.50		-0.41
562	D4530	9.67		0.02	1796		----		----
575		----		----	1807	ISO10370	9.37		-0.74
603	D4530	10.2		1.35	1811	ISO10370	9.99		0.82
604		----		----	1832		----		----
605	D4530	9.58		-0.21	1833		----		----
631	D4530	9.436		-0.57	1849		----		----
633		----		----	1906		----		----
663	D4530	9.086		-1.45	1936		----		----
671	D4530	9.35		-0.79	1937		----		----
750	ISO10370	10.06		1.00	1938		----		----
753		----		----	1950	ISO10370	9.51		-0.39
759		----		----	1956	ISO10370	9.24		-1.07
781	ISO10370	9.44		-0.56	1964		----		----
785	ISO10370	9.80		0.35	1967	D4530	9.794		0.33
824	ISO10370	9.54		-0.31	1995	D4530	8.8		-2.17
825	D4530	9.68		0.04	6016		----		----
850	ISO10370	9.660		-0.01	6026	D4530	9.9433		0.71
851	ISO10370	8.96		-1.77	6028	ISO10370	10	C	0.85
855	ISO10370	9.83		0.42	6039		----		----
858	D4530	9.84		0.45	6044	ISO10370	9.51		-0.39
862	D4530	9.92		0.65	6054	D4530	9.69		0.07
863	D4530	9.60		-0.16	6057	ISO10370	9.74		0.19
864	D4530	9.83		0.42	6075	ISO10370	9.927		0.66
865	D4530	9.58		-0.21	6092		----		----
866	ISO10370	9.63		-0.08	6142	ISO10370	9.70		0.09
870	D4530	9.922		0.65	6180	ISO10370	9.665		0.01
875	D4530	9.51		-0.39	6192		----		----
886	D4530	10.2		1.35					

normality	OK
n	88
outliers	2
mean (n)	9.6629
st.dev. (n)	0.29379
R(calc.)	0.8226
st.dev.(ISO10370:14)	0.39712
R(ISO10370:14)	1.1119

Lab 1613 first reported 6.85
 Lab 1720 first reported 11.64
 Lab 6028 first reported 13.7





Determination of Conradson Carbon Residue on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D189	10.11		0.20
90		----		----	913		----		----
92		----		----	922	D189	10.4		0.71
120		----		----	962		----		----
131		----		----	963	D189	9.70		-0.52
132		----		----	971	ISO6615	9.68		-0.55
140		----		----	974	D189	9.68		-0.55
150		----		----	982	D189	11.02		1.79
154		----		----	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171		----		----	1107		----		----
175		----		----	1109		----		----
194		----		----	1121		----		----
212	D189	10.25		0.44	1126		----		----
221		----		----	1134		----		----
224	D189	11.27		2.23	1135		----		----
225		----		----	1161		----		----
228	D189	9.685		-0.54	1167		----		----
237	D189	9.314		-1.19	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254	D189	9.6		-0.69	1266		----		----
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331		----		----	1345	D189	10.63		1.11
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402	IP13	10.4		0.71
339		----		----	1404		----		----
342		----		----	1428		----		----
343		----		----	1459		----		----
349		----		----	1510		----		----
356	D189	9.3		-1.22	1556		----		----
371	D189	9.91		-0.15	1569		----		----
391		----		----	1586	ISO6615	10.14		0.25
398	D4530	9.60		-0.69	1613		----	W	----
399		----		----	1631		----		----
440		----		----	1643	D189	10.42		0.74
444		----		----	1650	D189	10.676		1.19
463		----		----	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541	D189	9.034		-1.68	1728	D189	9.77		-0.40
557		----		----	1740	D189	9.55		-0.78
562		----		----	1796		----		----
575	D189	11.08	C	1.90	1807		----		----
603		----		----	1811		----		----
604		----		----	1832	ISO6615	10.090		0.16
605		----		----	1833		----		----
631		----		----	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750	D189	10.00		0.01	1938		----		----
753		----		----	1950	D189	9.73		-0.47
759		----		----	1956		----		----
781	D189	9.51		-0.85	1964		----		----
785		----		----	1967	D189	9.495		-0.88
824		----		----	1995		----		----
825		----		----	6016		----		----
850		----		----	6026	D189	10.5379		0.95
851		----		----	6028		----		----
855		----		----	6039		----		----
858		----		----	6044	D189	9.81		-0.33
862		----		----	6054		----		----
863		----		----	6057		----		----
864		----		----	6075		----		----
865	ISO6615	9.59		-0.71	6092		----		----
866		----		----	6142		----		----
870	D189	9.894		-0.18	6180		----		----
875		----		----	6192		----		----
886		----		----					

normality	OK
n	32
outliers	0
mean (n)	9.9961
st.dev. (n)	0.54524
R(calc.)	1.5267
st.dev.(D189:06)	0.57192
R(D189:06)	1.6014

Lab 575 first reported 11.92
 Lab 1613 first reported 7.35





Determination of Density at 15°C on sample #18105; results in kg/m³

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62	D4052	987.6		0.53	912	D1298	988.2		1.65
90	D4052	987.5		0.34	913	D1298	987.2		-0.22
92	D4052	986.2		-2.08	922	D1298	986.3		-1.90
120	ISO12185	987.5		0.34	962		----		----
131	D4052	987.5		0.34	963	ISO12185	987.3		-0.03
132	D4052	989.2	R(0.05)	3.52	971	ISO12185	987.21		-0.20
140	D4052	987.0		-0.59	974	D4052	987.4		0.16
150		----		----	982	D1298	987.4		0.16
154	D4052	987.5		0.34	1006	D1298	987.2	C	-0.22
158	D4052	988.0	C	1.28	1011	ISO12185	988.0		1.28
159	D4052	987.1		-0.40	1019	ISO3838	988.84		2.84
168		----		----	1059		----		----
169		----		----	1095	ISO12185	987.0		-0.59
171	ISO12185	987.0		-0.59	1107	D4052	987.5		0.34
175	D4052	987.2		-0.22	1109	D4052	987.3		-0.03
194	D4052	988.1		1.46	1121	ISO12185	988.9		2.96
212	ISO12185	987.9		1.09	1126	ISO12185	987.51		0.36
221	D4052	986.8		-0.96	1134	D4052	987.25		-0.12
224	D1298	986.2	C	-2.08	1135	ISO12185	987.4		0.16
225	D4052	987.0		-0.59	1161	ISO12185	986.2		-2.08
228	D1298	987.0		-0.59	1167	ISO12185	987.7		0.72
237	D4052	987.9		1.09	1177		----		----
238	D4052	989.30	C,R(0.05)	3.70	1233	ISO12185	986.9		-0.78
253	D4052	986.2		-2.08	1259	ISO12185	988.7		2.58
254	D1298	987.4		0.16	1266	ISO3675	989.8	R(0.01)	4.64
273		----		----	1269		----		----
313	ISO12185	987.4		0.16	1275	IP365	987.7		0.72
323	ISO12185	987.5		0.34	1299	D4052	987.4		0.16
331	ISO12185	987.5		0.34	1345	ISO12185	987.2		-0.22
333	ISO12185	987.5		0.34	1356	ISO12185	987.3	C	-0.03
334	ISO12185	987.5		0.34	1367	IP365	987.1		-0.40
337	ISO12185	987.7		0.72	1402	IP365	987.6		0.53
339		----		----	1404	ISO12185	987.4		0.16
342	D4052	987.49		0.32	1428	ISO12185	986.5	C	-1.52
343	D4052	986.4		-1.71	1459	ISO12185	987.2		-0.22
349		----		----	1510		----		----
356	ISO12185	987.3		-0.03	1556	ISO12185	987.1		-0.40
371	D4052	987.33		0.02	1569	ISO12185	986.5		-1.52
391	ISO12185	987.3		-0.03	1586	ISO12185	987.0		-0.59
398	ISO12185	987.5		0.34	1613	D4052	987.3		-0.03
399	D7042	986.4		-1.71	1631	ISO12185	987.2		-0.22
440	D4052	986.9		-0.78	1643	D4052	987.4		0.16
444		----		----	1650	ISO12185	987.1		-0.40
463	ISO12185	987.10		-0.40	1677	D4052	987.3		-0.03
511		----		----	1720	D4052	985.4	C,R(0.05)	-3.58
529		----		----	1724	D4052	987.0		-0.59
541	ISO12185	987.2	C	-0.22	1728	D4052	986.62		-1.30
557	D4052	987.6533		0.63	1740	D4052	987.0		-0.59
562	D1298	982.3	R(0.01)	-9.36	1796	ISO12185	987.27		-0.09
575		----		----	1807	ISO3675	987.1		-0.40
603	D1298	987.7		0.72	1811	ISO12185	987.5		0.34
604	D4052	988.4		2.02	1832	ISO12185	987.0		-0.59
605	D4052	987.6		0.53	1833	ISO12185	987.2		-0.22
631	D1298	989.78	R(0.01)	4.60	1849	ISO12185	987.1		-0.40
633	D1298	987.50		0.34	1906		----		----
663	D4052	987.9		1.09	1936	ISO12185	987.3		-0.03
671	D1298	987.5		0.34	1937	ISO12185	987.4		0.16
750	ISO12185	987.5		0.34	1938	ISO12185	987.3		-0.03
753	ISO12185	987.4		0.16	1950	ISO12185	987.3		-0.03
759	D1298	987.1		-0.40	1956	ISO3675	987.4		0.16
781	ISO12185	987.5		0.34	1964		----		----
785	D1298	987.8		0.90	1967	D1298	987.2		-0.22
824	ISO12185	987.0		-0.59	1995	D4052	985.48	R(0.05)	-3.43
825	ISO12185	987.1		-0.40	6016		----		----
850	ISO3675	987.3		-0.03	6026	D1298	987.5		0.34
851	ISO12185	987.9		1.09	6028	ISO12185	986.9		-0.78
855	ISO12185	987.29		-0.05	6039		----		----
858	ISO12185	987.4		0.16	6044	ISO12185	987.19		-0.24
862	ISO12185	987.1		-0.40	6054	D4052	986.999		-0.59
863	ISO12185	987.1		-0.40	6057	ISO12185	987.0		-0.59
864	ISO12185	987.1		-0.40	6075	ISO12185	881.2	R(0.01)	-198.08
865	D1298	987.4		0.16	6092	ISO12185	987.4		0.16
866	D1298	987.3		-0.03	6142	ISO12185	988.0		1.28
870	D1298	987.4		0.16	6180	ISO12185	987.2		-0.22
875	D1298	987.6		0.53	6192	ISO3675	987.2		-0.22
886	D4052	989.3	R(0.05)	3.70					

normality	not OK
n	123
outliers	9
mean (n)	987.317
st.dev. (n)	0.4643
R(calc.)	1.300
st.dev.(ISO12185:96)	0.5357
R(ISO12185:96)	1.5

Lab 158 reported 0.988 kg/m³
 Lab 224 first reported 983.96
 Lab 238 first reported 989.03
 Lab 541 first reported 989.5
 Lab 1006 first reported 0.9872 kg/m³
 Lab 1356 reported 0.9873 kg/m³
 Lab 1428 first reported 989.1
 Lab 1720 first reported 985.7





Determination of Flash Point PMcc on sample #18105; results in °C

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D93-B	120		-0.78
90	D93-B	>110		----	913	D93-B	122		0.15
92	D93-B	119.0		-1.25	922	D93-B	123		0.62
120	ISO2719-B	>110		----	962		----		----
131		----		----	963	ISO2719-B	121.5		-0.08
132	ISO2719-A	121.0		-0.31	971	ISO2719-B	120.5		-0.55
140		----		----	974	D93-B	120.5		-0.55
150	D93-A	>110.0		----	982	D93-B	116.1		-2.60
154	D93-B	123.0		0.62	1006	D93-B	122.5		0.39
158	D93-B	128.0		2.95	1011		----		----
159	D93-B	>110		----	1019		----		----
168	D93-B	118		-1.71	1059	ISO2719-B	121.0		-0.31
169	D93-B	123.9		1.04	1095	D93-B	121.0		-0.31
171	ISO2719-A	126.5		2.25	1107	D93-B	121.0		-0.31
175	D93-B	120		-0.78	1109	D93-B	>100		----
194	D93-B	120.5		-0.55	1121	ISO2719-B	123.5		0.85
212	ISO2719-B	122.0		0.15	1126	ISO2719-A	126		2.02
221	D93-B	123		0.62	1134	D93-B	123.0		0.62
224	D93-B	118.2		-1.62	1135	ISO2719-B	123.0		0.62
225	D93-B	122.0		0.15	1161		----		----
228	D93-B	122		0.15	1167	ISO2719-B	125.5		1.79
237	D93-B	120.0		-0.78	1177		----		----
238	D93-B	123		0.62	1233	ISO2719-A	121		-0.31
253	D93-B	120		-0.78	1259	ISO2719-A	130.5		4.12
254	D93-B	118		-1.71	1266	ISO2719-B	120.0		-0.78
273	D93-B	110	R(0.01)	-5.45	1269	D93-B	119.5		-1.01
313	ISO2719-B	122.0		0.15	1275	IP34-B	120.0		-0.78
323	ISO2719-B	119.0		-1.25	1299	D93-B	116.5	C	-2.41
331	D93-B	121.5		-0.08	1345	D93-B	120.0		-0.78
333	ISO2719-B	122.0		0.15	1356		145	R(0.01)	10.89
334	ISO2719-B	121.0		-0.31	1367	D93-B	122		0.15
337		----		----	1402	IP34-B	121.0		-0.31
339		----		----	1404	ISO2719-A	120.0		-0.78
342	ISO2719-B	119.5		-1.01	1428	ISO2719-B	127.5		2.72
343	D93-B	124		1.09	1459	ISO2719-A	120.0		-0.78
349		----		----	1510		----		----
356	ISO2719-B	126.0		2.02	1556	ISO2719-B	123.5		0.85
371	D93-B	118.0		-1.71	1569		----		----
391		----		----	1586	ISO2719-A	120.0		-0.78
398	ISO2719-B	127	C	2.49	1613	D93-B	123.0		0.62
399	D93-B	135.0	R(0.01)	6.22	1631	ISO2719-A	121		-0.31
440	IP34-B	119.9	C	-0.83	1643	D93-B	124.1		1.13
444		----		----	1650	ISO2719-A	118.7		-1.39
463	D93-B	121.0		-0.31	1677	D93-B	122.5		0.39
511		----		----	1720		----		----
529		----		----	1724	D93-B	119		-1.25
541	D93-B	121.00		-0.31	1728	D93-B	121		-0.31
557	D93-B	123.5		0.85	1740	ISO2719-B	121		-0.31
562	D93-B	118.8		-1.34	1796	D93-B	118.2		-1.62
575	D93-B	126	C	2.02	1807	ISO2719-A	122.0		0.15
603	D93-B	121		-0.31	1811	ISO2719-A	124.5		1.32
604	D93-B	122		0.15	1832	ISO2719-B	121.0	C	-0.31
605	D93-B	118.0		-1.71	1833	ISO2719-A	119		-1.25
631	D93-B	116.0		-2.65	1849	ISO2719-B	121.5		-0.08
633	D93-B	124.6		1.37	1906		----		----
663	D93-B	118.35		-1.55	1936		----		----
671	D93-B	>110.0		----	1937	ISO2719-B	123		0.62
750	D93-B	121.0		-0.31	1938		----		----
753	D93-B	123.0		0.62	1950	ISO2719-B	122.3		0.29
759	ISO2719-B	123.0		0.62	1956	ISO2719-A	125		1.55
781	ISO2719-B	120.0		-0.78	1964		----		----
785	ISO2719-B	123.0		0.62	1967	D93-B	122.5		0.39
824	ISO2719-B	118.0		-1.71	1995	D93-A	125.2		1.65
825	D93-B	120.0		-0.78	6016		----		----
850	ISO2719-B	123.0		0.62	6026	D93-B	120.0		-0.78
851	ISO2719-B	121.0		-0.31	6028	ISO2719-A	122.0		0.15
855	D93-B	122		0.15	6039		----		----
858	D93-B	122.0		0.15	6044	ISO2719-A	128		2.95
862	D93-B	121.0		-0.31	6054	D93-B	118.0		-1.71
863	D93-B	121.0		-0.31	6057	ISO2719-B	123.0		0.62
864	D93-B	121.0		-0.31	6075	ISO2719-A	126.0		2.02
865	D93-B	124		1.09	6092	D93-B	115		-3.11
866	ISO2719-B	121.0		-0.31	6142	ISO2719-B	120.5		-0.55
870	D93-A	122		0.15	6180	ISO2719-A	126.0		2.02
875	D93-B	122.0		0.15	6192	ISO2719-A	124		1.09
886	D93-A	122.5		0.39					

normality	OK
n	118
outliers	3
mean (n)	121.672
st.dev. (n)	2.6417
R(calc.)	7.397
st.dev.(ISO2719-B:16)	2.1429
R(ISO2719-B:16)	6

Lab 398 first reported 137
 Lab 440 first reported 113.8
 Lab 575 first reported 110
 Lab 1299 first reported 106.5
 Lab 1832 first reported 114.0





Determination of Heat of Combustion Gross on sample #18105; results in MJ/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		43.02		0.21
90	D240	42.823		-1.17	913		----		----
92	D240	42.998		0.06	922	D240	42.9612		-0.20
120	D4868	42.98		-0.07	962		----		----
131		----		----	963		----		----
132		42.940		-0.35	971	D240	42.90		-0.63
140		----		----	974	D4868	42.99		0.00
150	D240	43.273	C	1.98	982	D4868	43		0.07
154		----		----	1006		----		----
158		----		----	1011	D240	----		----
159		----		----	1019	D4809	43.078		0.62
168		----		----	1059	D240	43.208		1.53
169		----		----	1095		----		----
171	D240	42.555		-3.05	1107		----		----
175		----		----	1109	D4868	43.028		0.27
194		----		----	1121		----		----
212		----		----	1126		----		----
221	D4868	42.60		-2.73	1134		----		----
224		----		----	1135	D240	42.955		-0.25
225		42.99		0.00	1161	D240	43.170		1.26
228		----		----	1167	DIN51900-2	42.695		-2.07
237		----		----	1177	DIN51900-1/2	42.928		-0.43
238		----		----	1233		----		----
253		----		----	1259	D4868	42.97		-0.14
254	D4868	42.988		-0.01	1266		42.94		-0.35
273		----		----	1269	DIN51900-3	42.983		-0.05
313		----		----	1275		----		----
323	D240	43.140		1.05	1299		----		----
331		----		----	1345		43.00		0.07
333		----		----	1356		----		----
334		----		----	1367		41.906	C,R(0.01)	-7.59
337		----		----	1402	D240	42.925		-0.46
339		----		----	1404		----		----
342	D4868	43.000	C	0.07	1428		----		----
343		42.932		-0.41	1459		----		----
349		----		----	1510		----		----
356	D4868	43.00		0.07	1556	D4868	43.011		0.15
371	D4868	42.977		-0.09	1569	D240	----		----
391		----		----	1586	D240	43.121		0.92
398		----		----	1613	D240	43.020		0.21
399		----		----	1631		----		----
440		----		----	1643	D240	42.865		-0.88
444		----		----	1650		----		----
463	D4868	42.996		0.04	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541		----		----	1728	D4868	43.01094		0.15
557		43.0001		0.07	1740	D240	43.040		0.35
562		----		----	1796		----		----
575		----		----	1807		43.027		0.26
603		----		----	1811		42.90		-0.63
604		----		----	1832		----		----
605		----		----	1833		----		----
631	D240	43.0514		0.43	1849		----		----
633		----		----	1906	D4809	43.606	R(0.01)	4.31
663		----		----	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753		----		----	1950	D4868	43.00		0.07
759		----		----	1956		42.864	C	-0.88
781	D4868	43.00		0.07	1964		----		----
785		----		----	1967		----		----
824	D240	42.765		-1.58	1995		----		----
825		42.9545		-0.25	6016		----		----
850	GB/T384	43.114		0.87	6026	D4868	43.000		0.07
851		----		----	6028	D240	----		----
855		43.118	C	0.90	6039		----		----
858	D4868	43.02		0.21	6044		----		----
862	D240	43.065		0.52	6054		----		----
863	D240	43.065		0.52	6057		----		----
864	D4868	43.286		2.07	6075	D4868	10.802	R(0.01)	-225.32
865	D240	43.175		1.29	6092		----		----
866	D240	43.084		0.66	6142		----		----
870		----		----	6180		42.9546		-0.25
875	D4868	42.97		-0.14	6192		----		----
886		----		----					

normality	not OK
n	58
outliers	3
mean (n)	42.9901
st.dev. (n)	0.13054
R(calc.)	0.3655
st.dev.(D240:17)	0.14286
R(D240:17)	0.40

Lab 150 first reported 45.273
 Lab 342 first reported 43.000
 Lab 855 first reported 40.663
 Lab 1367 first reported 41.906.15
 Lab 1956 first reported 42.364





Determination of Heat of Combustion Net on sample #18105; results in MJ/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		40.68		0.03
90	D240	----		----	913		----		----
92	D240	----		----	922	D240	40.9355		1.82
120	D4868	40.64		-0.25	962		----		----
131		----		----	963		----		----
132		----		----	971	D240	40.64		-0.25
140		----		----	974	D4868	40.65		-0.18
150	D240	41.021		2.42	982	D4868	40.65		-0.18
154		----		----	1006		----		----
158		----		----	1011	D240	----		----
159		----		----	1019	D4809	40.859		1.28
168		----		----	1059	D240	40.833		1.10
169		----		----	1095		----		----
171	D240	----		----	1107		----		----
175		----		----	1109	D4868	40.687		0.08
194		----		----	1121		----		----
212		----		----	1126		----		----
221	D4868	40.29		-2.70	1134		----		----
224		----		----	1135	D240	40.684		0.06
225		40.65		-0.18	1161	D240	----		----
228		----		----	1167	DIN51900-2	----		----
237		----		----	1177	DIN51900-1/2	40.657		-0.13
238		----		----	1233		----		----
253		----		----	1259	D4868	40.63		-0.32
254	D4868	40.648		-0.19	1266		40.61		-0.46
273		----		----	1269	DIN51900-3	----		----
313		----		----	1275		----		----
323	D240	40.890		1.50	1299		----		----
331		----		----	1345		40.66		-0.11
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402	D240	40.735		0.42
339		----		----	1404		----		----
342	D4868	40.660	C	-0.11	1428		----		----
343		40.459		-1.52	1459		----		----
349		----		----	1510		----		----
356	D4868	40.66		-0.11	1556	D4868	40.671		-0.03
371	D4868	40.638		-0.26	1569	D240	----		----
391		----		----	1586	D240	----		----
398		----		----	1613	D240	----		----
399		----		----	1631		----		----
440		----		----	1643	D240	40.637		-0.27
444		----		----	1650		----		----
463	D4868	40.655		-0.14	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541		----		----	1728	D4868	40.6683		-0.05
557		40.6615		-0.10	1740	D240	40.725		0.35
562		----		----	1796		----		----
575		----		----	1807		40.751		0.53
603		----		----	1811		40.65		-0.18
604		----		----	1832		----		----
605		----		----	1833		----		----
631	D4868	40.6200		-0.39	1849		----		----
633		----		----	1906	D4809	41.397	R(0.01)	5.05
663		----		----	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753		----		----	1950	D4868	40.66		-0.11
759		----		----	1956		40.662		-0.09
781	D4868	40.66		-0.11	1964		----		----
785		----		----	1967		----		----
824	D240	40.534		-0.99	1995		----		----
825		----		----	6016		----		----
850	GB/T384	40.638		-0.26	6026	D4868	40.660		-0.11
851		----		----	6028	D240	----		----
855	GB/T384	40.663	C	-0.09	6039		----		----
858	D4868	40.68		0.03	6044		----		----
862	D240	40.775		0.70	6054		----		----
863	D240	----		----	6057		----		----
864	D4868	40.949		1.91	6075	D4868	10.441	R(0.01)	-211.64
865	D240	40.685		0.07	6092		----		----
866	ISO8217	40.93		1.78	6142		----		----
870		----		----	6180		40.117		-3.91
875	D4868	40.63		-0.32	6192		----		----
886		----		----					

normality	not OK
n	47
outliers	2
mean (n)	40.6755
st.dev. (n)	0.14780
R(calc.)	0.4138
st.dev.(D240:17)	0.14286
R(D240:17)	0.40

Lab 342 first reported 40660
 Lab 855 first reported 43.118





Determination of Kinematic Viscosity at 50°C on sample #18105; results in mm²/s

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62	D445	291.1		0.61	912	D445	292.2		0.76
90	D445	289.13		0.35	913	D445	280.1		-0.84
92	D445	289.423		0.39	922	D445	280.5		-0.79
120	ISO3104	281.9		-0.60	962		----		----
131	D445	290.2		0.49	963	ISO3104	290.1		0.48
132	ISO3104	276.8		-1.28	971	ISO3104	280.5		-0.79
140	D445	285.3		-0.15	974	D445	288.5		0.27
150	D445	288.0		0.20	982	D445	293.1		0.88
154	D445	285.5		-0.13	1006	D445	292.0		0.73
158		----		----	1011	ISO3104	288.7		0.30
159	D445	286.8		0.05	1019	ISO3104	288.2		0.23
168	D445	282.6		-0.51	1059		----		----
169	D445	291.1940		0.63	1095	ISO3104	282.3		-0.55
171	ISO3104	283.4		-0.40	1107	ISO3104	277.6		-1.17
175	D445	287.48		0.14	1109	D445	287.60		0.15
194	D445	283.9		-0.34	1121	ISO3104	312.7	R(0.01)	3.47
212	ISO3104	286.4		-0.01	1126	ISO3104	261.87	R(0.01)	-3.25
221	D445	293.2		0.89	1134	IP71	282.4535		-0.53
224		----		----	1135	ISO3104	297.1		1.41
225	D445	285.1		-0.18	1161	ISO3104	292.7		0.83
228	D445	294.18		1.02	1167		----		----
237	D445	287.62		0.15	1177		----		----
238		----		----	1233	ISO3104	290.6		0.55
253	D445	283.10		-0.44	1259	ISO3104	283.1		-0.44
254	D445	279.65		-0.90	1266	ISO3104	296.0		1.26
273		----		----	1269		----		----
313	ISO3104	276.9		-1.26	1275	IP71	289.401		0.39
323	ISO3104	282.5		-0.52	1299	D445	274.9		-1.53
331	ISO3104	288.2		0.23	1345	D445	290.99		0.60
333	ISO3104	284.7		-0.23	1356	ISO3104	292.230		0.76
334	ISO3104	282.8		-0.48	1367	IP71	278.9		-1.00
337		----		----	1402	IP71	279.8		-0.88
339		----		----	1404	ISO3104	287.45		0.13
342	ISO3104	279.4		-0.93	1428	ISO3104	289.4		0.39
343	ISO3104	279.0		-0.98	1459		----		----
349		----		----	1510		----		----
356	ISO3104	283.1		-0.44	1556	ISO3104	273.88		-1.66
371	D445	279.29		-0.95	1569		----		----
391	ISO3104	277		-1.25	1586	ISO3104	283.8		-0.35
398		----		----	1613	D445	294.38		1.05
399	D445	279.12		-0.97	1631		----		----
440		----		----	1643	D445	281.07		-0.71
444		----		----	1650	ISO3104	287.11		0.09
463	ISO3104	286.73		0.04	1677	D445	283.94		-0.33
511		----		----	1720		----		----
529		----		----	1724	D445	284.7		-0.23
541	D445	280.32		-0.81	1728	D445	295.60		1.21
557	D445	277.685578		-1.16	1740	ISO3104	283.9		-0.34
562		----		----	1796	ISO3104	281.39		-0.67
575	D445	288.905		0.32	1807		----		----
603	D445	289.1		0.35	1811		----		----
604	D445	291.64		0.69	1832	ISO3104	280.679		-0.76
605	D445	285.2		-0.17	1833		----		----
631	D445	287.82		0.18	1849		----		----
633	D445	285.41		-0.14	1906		----		----
663	D445	283.17		-0.43	1936		----		----
671	D445	296.74		1.36	1937		----		----
750	ISO3104	283.1		-0.44	1938		----		----
753	D445	282.3		-0.55	1950	ISO3104	285.8		-0.09
759	ISO3104	282.7		-0.50	1956	ISO3104	302.8		2.16
781	ISO3104	284.5		-0.26	1964		----		----
785	ISO3104	282.8		-0.48	1967	D445	283.29		-0.42
824	ISO3104	297.2		1.42	1995		----		----
825	D445	298.51		1.59	6016		----		----
850	ISO3104	286.00		-0.06	6026	D445	285.7464		-0.09
851	ISO3104	288.3		0.24	6028	ISO3104	297.0	C	1.39
855	D445	290.3		0.51	6039		----		----
858	D445	291.6		0.68	6044	ISO3104	282.05		-0.58
862	D445	290.26		0.50	6054	D445	289.62		0.42
863	D445	283.83		-0.35	6057	ISO3104	279.6		-0.91
864	D445	290.6		0.55	6075	ISO3104	289.2		0.36
865	D445	289.9		0.46	6092	D445	283.92		-0.33
866	ISO3104	284.68		-0.23	6142	ISO3104	296.1738		1.28
870	D445	291.7		0.69	6180	ISO3104	300.26		1.82
875	D445	282.9		-0.47	6192		----		----
886	D445	289		0.34					

normality	OK
n	113
outliers	2
mean (n)	286.4532
st.dev. (n)	5.73736
R(calc.)	16.0646
st.dev.(ISO3104:94)	7.57055
R(ISO3104:94)	21.20

Lab 6028 first reported 328.8





Determination of Kinematic Viscosity at 100°C on sample #18105; results in mm²/s

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D445	26.12		-1.95
90		----		----	913		----		----
92		----		----	922		----		----
120		----		----	962		----		----
131		----		----	963	ISO3104	27.51		0.82
132		----		----	971	ISO3104	27.15		0.10
140	D445	27.17		0.14	974	D445	27.20		0.20
150	D445	26.34		-1.51	982	D445	26.82		-0.56
154		----		----	1006		----		----
158		----		----	1011	ISO3104	27.18		0.16
159		----		----	1019		----		----
168	D445	26.64		-0.92	1059		----		----
169	D445	26.905		-0.39	1095	ISO3104	27.94		1.68
171	ISO3104	27.14		0.08	1107	ISO3104	26.86		-0.48
175		----		----	1109	D445	27.115		0.03
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134	IP71	27.685	C	1.17
224		----		----	1135	ISO3104	26.94		-0.32
225	D445	27.44		0.68	1161	ISO3104	27.03		-0.14
228		----		----	1167	ISO3104	27.51		0.82
237	D445	27.05		-0.10	1177		----		----
238		----		----	1233		----		----
253	D445	28.34		2.47	1259	ISO3104	27.13		0.06
254		----		----	1266		----		----
273		----		----	1269		----		----
313	ISO3104	26.94		-0.32	1275	IP71	26.299		-1.60
323	ISO3104	39.90	R(0.01)	25.53	1299	D445	26.72		-0.76
331	ISO3104	27.17		0.14	1345		----		----
333	ISO3104	26.73		-0.74	1356		----		----
334	ISO3104	26.99		-0.22	1367	IP71	26.62		-0.96
337		----		----	1402	IP71	27.26		0.32
339		----		----	1404	ISO3104	27.04		-0.12
342		----		----	1428	ISO3104	26.65		-0.90
343	ISO3104	26.98		-0.24	1459		----		----
349		----		----	1510		----		----
356	ISO3104	27.02		-0.16	1556	ISO3104	26.748		-0.70
371	D445	26.952		-0.29	1569	D445	26.89		-0.42
391	ISO3104	26.94		-0.32	1586	ISO3104	26.84		-0.52
398		----		----	1613	D445	27.18		0.16
399		----		----	1631	ISO3104	26.759		-0.68
440		----		----	1643	D445	26.79		-0.62
444		----		----	1650	ISO3104	27.455		0.71
463	ISO3104	27.135		0.07	1677	D445	27.416		0.63
511		----		----	1720		----		----
529		----		----	1724	D445	26.798		-0.60
541	D445	27.515		0.83	1728	D445	27.630		1.06
557	D445	26.3238		-1.55	1740	ISO3104	26.72		-0.76
562		----		----	1796	ISO3104	27.402		0.60
575	D445	27.09		-0.02	1807	ISO3104	27.01		-0.18
603		----		----	1811	ISO3104	27.29		0.38
604		----		----	1832		----		----
605	D445	27.51		0.82	1833	ISO3104	26.87		-0.46
631	D445	26.734		-0.73	1849	ISO3104	26.80		-0.60
633		----		----	1906		----		----
663	D445	27.239		0.28	1936	ISO3104	27.02		-0.16
671		----		----	1937	ISO3104	26.91		-0.38
750	ISO3104	27.10		0.00	1938	ISO3104	27.059		-0.08
753	D445	27.14		0.08	1950	ISO3104	27.07		-0.06
759		----		----	1956	ISO3104	27.4		0.60
781	ISO3104	27.10		0.00	1964		----		----
785	ISO3104	27.09		-0.02	1967		----		----
824	ISO3104	27.01		-0.18	1995		----		----
825	D445	27.121		0.04	6016		----		----
850	ISO3104	27.094		-0.01	6026	D445	27.8697	C	1.54
851		----		----	6028	ISO3104	27.36		0.52
855	D445	26.98		-0.24	6039		----		----
858	D445	26.97		-0.26	6044		----		----
862	D445	27.601		1.00	6054		----		----
863	D445	27.302		0.40	6057	ISO3104	27.65		1.10
864		----		----	6075		----		----
865		----		----	6092		----		----
866	ISO3104	26.991		-0.22	6142	ISO3104	27.9161		1.63
870	D445	27.202		0.20	6180	ISO3104	27.181		0.16
875	D445	26.87		-0.46	6192	ISO3104	27.98		1.76
886	D445	26.8		-0.60					

normality	suspect
n	85
outliers	1
mean (n)	27.0995
st.dev. (n)	0.38004
R(calc.)	1.0641
st.dev.(ISO3104:94)	0.50142
R(ISO3104:94)	1.4040



Lab 1134 first reported 29.685
 Lab 6026 first reported 28.8697



Determination of Viscosity Stabinger at 50°C on sample #18105; results in mm²/s

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913		----		----
92		----		----	922		----		----
120		----		----	962		----		----
131		----		----	963		----		----
132		----		----	971		----		----
140		----		----	974		----		----
150		----		----	982		----		----
154		----		----	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171	D7042	288.0		0.05	1107		----		----
175		----		----	1109		----		----
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	D7042	286.9		-0.05
225		----		----	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331		----		----	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402		----		----
339		----		----	1404		----		----
342		----		----	1428		----		----
343		----		----	1459	D7042	285.09		-0.23
349		----		----	1510		----		----
356		----		----	1556		----		----
371		----		----	1569		----		----
391		----		----	1586	D7042	287.7		0.02
398	D7042	290.2		0.26	1613		----		----
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	D7042	280.85		-0.63	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541		----		----	1728	D7042	296.24		0.83
557		----		----	1740	D7042	283.8		-0.35
562		----		----	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833		----		----
631		----		----	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753		----		----	1950		----		----
759		----		----	1956		----		----
781		----		----	1964		----		----
785		----		----	1967		----		----
824		----		----	1995	D7042	291.77		0.41
825		----		----	6016		----		----
850		----		----	6026		----		----
851		----		----	6028		----		----
855		----		----	6039		----		----
858		----		----	6044	D7042	284.65		-0.27
862	D7042	285.14		-0.22	6054		----		----
863		----		----	6057	D7042	289.3		0.17
864		----		----	6075		----		----
865		----		----	6092		----		----
866		----		----	6142		----		----
870		----		----	6180		----		----
875		----		----	6192		----		----
886		----		----					



normality	OK
n	12
outliers	0
mean (n)	287.4700
st.dev. (n)	4.08102
R(calc.)	11.4269
st.dev.(D7042:16e3)	10.56452
R(D7042:16e3)	29.5807



Determination of Viscosity Stabinger at 100°C on sample #18105; results in mm²/s

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913		----		----
92		----		----	922		----		----
120		----		----	962		----		----
131		----		----	963		----		----
132		----		----	971		----		----
140		----		----	974		----		----
150		----		----	982		----		----
154		----		----	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171	D7042	26.74		-0.44	1107		----		----
175		----		----	1109		----		----
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	D7042	27.35		0.44
225		----		----	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331		----		----	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402		----		----
339		----		----	1404		----		----
342		----		----	1428		----		----
343		----		----	1459	D7042	26.83		-0.31
349		----		----	1510		----		----
356		----		----	1556		----		----
371		----		----	1569		----		----
391		----		----	1586	D7042	26.90		-0.21
398	D7042	26.94		-0.15	1613		----		----
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	D7042	27.065		0.03	1677		----		----
511		----		----	1720	D7042	25.17	D(0.01)	-2.70
529		----		----	1724		----		----
541		----		----	1728	D7042	27.689		0.93
557		----		----	1740	D7042	26.85		-0.28
562		----		----	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833		----		----
631		----		----	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753		----		----	1950		----		----
759		----		----	1956		----		----
781		----		----	1964		----		----
785		----		----	1967		----		----
824		----		----	1995	D7042	26.981		-0.09
825		----		----	6016		----		----
850		----		----	6026		----		----
851		----		----	6028		----		----
855		----		----	6039		----		----
858		----		----	6044	D7042	27.05		0.01
862	D7042	26.950		-0.14	6054		----		----
863		----		----	6057	D7042	27.19		0.21
864		----		----	6075		----		----
865		----		----	6092		----		----
866		----		----	6142		----		----
870		----		----	6180		----		----
875		----		----	6192		----		----
886		----		----					



normality	not OK
n	12
outliers	1
mean (n)	27.0446
st.dev. (n)	0.26140
R(calc.)	0.7319
st.dev.(D7042:16e3)	0.69408
R(D7042:16e3)	1.9434



Determination of Nitrogen on sample #18105; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913		----		----
92		----		----	922		----		----
120		----		----	962		----		----
131	D4629	3459.33		1.62	963		----		----
132		----		----	971	D5762 Gravimetric	3049		0.18
140	D5762 Gravimetric	2000		-3.51	974		----		----
150	D5762 Volumetric	2900		-0.35	982		----		----
154		----		----	1006		----		----
158		----		----	1011		----		----
159	D4629	2800		-0.70	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171	D5762 Volumetric	3100		0.36	1107		----		----
175		----		----	1109	D4629	3747		2.63
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134	D5762 Gravimetric	2923.704		-0.26
224		----		----	1135	D5762 Volumetric	2690.4		-1.08
225		----		----	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259	D4629	2400		-2.10
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	In house	3013		0.05
323	D5762 Gravimetric	3500		1.76	1299		----		----
331		----		----	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402	D5762 Volumetric	3200		0.71
339		----		----	1404	D5762 Volumetric	3325		1.15
342		----		----	1428		----		----
343		----		----	1459		----		----
349		----		----	1510		----		----
356		----		----	1556	D5762 Volumetric	3262		0.92
371		----		----	1569	D4629	3098		0.35
391		----		----	1586	D5762 Volumetric	3267		0.94
398		----		----	1613		----		----
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463	D5762 Gravimetric	3005		0.02	1677	D4629	2995		-0.01
511		----		----	1720		----		----
529		----		----	1724		----		----
541		----		----	1728		----		----
557		----		----	1740		----		----
562		----		----	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833		----		----
631		----		----	1849		----		----
633		----		----	1906		----		----
663	D5762 Gravimetric	2580		-1.47	1936		----		----
671		----		----	1937		----		----
750	D5762 Gravimetric	2700		-1.05	1938		----		----
753		----		----	1950		----		----
759		----		----	1956		----		----
781	D3228	3057		0.20	1964		----		----
785		----		----	1967		----		----
824	D5762 Gravimetric	3070		0.25	1995		----		----
825	D5762 Gravimetric	2910		-0.31	6016		----		----
850		----		----	6026	D5762 Volumetric	3039.65		0.14
851		----		----	6028		----		----
855		----		----	6039		----		----
858		----		----	6044		----		----
862	D5762 Gravimetric	2570		-1.50	6054		----		----
863		----		----	6057	D5762 Gravimetric	2703		-1.04
864		----		----	6075		----		----
865		----		----	6092		----		----
866		----		----	6142	D5762 Gravimetric	3600		2.11
870		----		----	6180		----		----
875		----		----	6192		----		----
886		----		----					

normality	OK	D5762	D5762
n	28	<u>Volumetric only</u>	<u>Gravimetric only</u>
outliers	0	8	12
mean (n)	2998.72	3098.01	2884.23
st.dev. (n)	372.963	215.851	427.457
R(calc.)	1044.30	604.38	1196.88
st.dev.(D5762:18)	284.878	284.878	284.878
R(D5762:18)	797.66	824.07	767.20





Determination of Pour Point (Lower) on sample #18105; results in °C

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90	D97	3		-0.51	913		----		----
92	D97	0		-1.78	922	D97	6		0.77
120		----		----	962		----		----
131		----		----	963	ISO3016	3		-0.51
132		----		----	971		----		----
140		----		----	974		----		----
150		----		----	982	D97	3		-0.51
154		----		----	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171		----		----	1107		----		----
175		----		----	1109		----		----
194		----		----	1121	ISO3016	6		0.77
212		----		----	1126		----		----
221	D97	9		2.04	1134	D97	6		0.77
224	D97	2.0		-0.93	1135	ISO3016	3		-0.51
225		----		----	1161		----		----
228		----		----	1167	ISO3016	6		0.77
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259	ISO3016	9		2.04
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP15	6.0		0.77
323		----		----	1299		----		----
331		----		----	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402		----		----
339		----		----	1404		----		----
342		----		----	1428	ISO3016	3		-0.51
343		----		----	1459		----		----
349		----		----	1510		----		----
356		----		----	1556		----		----
371		----		----	1569		----		----
391		----		----	1586		----		----
398		----		----	1613	D97	6.0		0.77
399	D97	6		0.77	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463		----		----	1677		----		----
511		----		----	1720		----		----
529	D97	6		0.77	1724	D97	3		-0.51
541		----		----	1728	D97	3		-0.51
557	D97	0		-1.78	1740	ISO3016	6		0.77
562	D97	9		2.04	1796	ISO3016	3		-0.51
575		----		----	1807		----		----
603	D97	6		0.77	1811		----		----
604	D97	3		-0.51	1832	ISO3016	3		-0.51
605	D97	3		-0.51	1833		----		----
631	D97	6		0.77	1849	ISO3016	6		0.77
633		----		----	1906		----		----
663		----		----	1936		----		----
671	D97	3		-0.51	1937		----		----
750		----		----	1938		----		----
753		----		----	1950	ISO3016	3		-0.51
759	ISO3016	6		0.77	1956		----		----
781	ISO3016	6		0.77	1964		----		----
785	ISO3016	9		2.04	1967	D97	3		-0.51
824	ISO3016	0		-1.78	1995		----		----
825	D97	3		-0.51	6016		----		----
850	ISO3016	3		-0.51	6026	D97	3		-0.51
851		----		----	6028	D97	6		0.77
855	ISO3016	3		-0.51	6039		----		----
858		----		----	6044		----		----
862	D97	9	ex	2.04	6054		----		----
863	D97	0		-1.78	6057	ISO3016	0		-1.78
864		----		----	6075		----		----
865	D97	3		-0.51	6092		----		----
866	ISO3016	0		-1.78	6142		----		----
870	D97	6		0.77	6180		----		----
875	D97	6		0.77	6192		----		----
886		----		----					

normality	OK
n	47
outliers	0 (+1ex)
mean (n)	4.19
st.dev. (n)	2.490
R(calc.)	6.97
st.dev.(ISO3016:94)	2.354
R(ISO3016:94)	6.59

Lab 862 test result excluded, PP lower > PP upper which is in principle not possible





Determination of Pour Point (Upper) on sample #18105; results in °C

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D97	12		2.25
90	D97	6		-0.30	913		----		----
92	D97	3		-1.58	922	D97	9		0.97
120		----		----	962		----		----
131		----		----	963	ISO3016	3		-1.58
132	ISO3016	3		-1.58	971	ISO3016	6		-0.30
140		----		----	974	D97	3		-1.58
150	D97	0		-2.85	982	D97	6		-0.30
154	D97	3		-1.58	1006	D97	3		-1.58
158		----		----	1011	D97	6		-0.30
159	D97	12		2.25	1019	ISO3016	9		0.97
168	D97	9		0.97	1059	ISO3016	3		-1.58
169	D97	6		-0.30	1095		----		----
171		----		----	1107		----		----
175		----		----	1109	D97	6		-0.30
194	D97	9		0.97	1121	ISO3016	9		0.97
212	ISO3016	9		0.97	1126		----		----
221	D97	12		2.25	1134	D97	6		-0.30
224	D97	5.0		-0.73	1135	ISO3016	3		-1.58
225		----		----	1161	ISO3016	9		0.97
228	D97	0		-2.85	1167	ISO3016	9		0.97
237	D97	3		-1.58	1177		----		----
238	D97	9		0.97	1233	ISO3016	6		-0.30
253	D97	3		-1.58	1259	ISO3016	12		2.25
254	D97	3		-1.58	1266		----		----
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331		----		----	1345	D97	9		0.97
333		----		----	1356	ISO3016	6		-0.30
334		----		----	1367	ISO3016	3		-1.58
337		----		----	1402	D97	9		0.97
339		----		----	1404		----		----
342	ISO3016	9	C	0.97	1428	ISO3016	6		-0.30
343		----		----	1459		----		----
349		----		----	1510		----		----
356	ISO3016	6		-0.30	1556	ISO3016	6		-0.30
371	D97	6		-0.30	1569		----		----
391		----		----	1586	ISO3016	0		-2.85
398	ISO3016	-9	R(0.01)	-6.68	1613	D97	9.0		0.97
399	D97	9		0.97	1631	ISO3016	9		0.97
440		----		----	1643	D97	9		0.97
444		----		----	1650		----		----
463		----		----	1677	D97	6		-0.30
511		----		----	1720	D97	12		2.25
529	D97	9		0.97	1724	D97	9		0.97
541		----		----	1728	D97	6		-0.30
557	D97	3		-1.58	1740	ISO3016	9		0.97
562	D97	9		0.97	1796	ISO3016	6		-0.30
575		----		----	1807	D97	12		2.25
603		----		----	1811		----		----
604	D97	6		-0.30	1832	ISO3016	6		-0.30
605	D97	6		-0.30	1833		----		----
631		----		----	1849	ISO3016	9		0.97
633		----		----	1906		----		----
663	D97	0		-2.85	1936		----		----
671	D97	9		0.97	1937		----		----
750	D97	9		0.97	1938		----		----
753		----		----	1950	ISO3016	6		-0.30
759	ISO3016	9		0.97	1956		----		----
781	ISO3016	9		0.97	1964		----		----
785	ISO3016	12		2.25	1967	D97	6		-0.30
824	ISO3016	6		-0.30	1995		----		----
825	D97	6		-0.30	6016		----		----
850	ISO3016	6		-0.30	6026	D97	6		-0.30
851	ISO3016	6		-0.30	6028	D97	9		0.97
855	D97	6		-0.30	6039		----		----
858	D97	9		0.97	6044		----		----
862	D97	0	ex	-2.85	6054		----		----
863	D97	3		-1.58	6057	ISO3016	3		-1.58
864	D97	9		0.97	6075		----		----
865	D97	6		-0.30	6092		----		----
866	ISO3016	3		-1.58	6142		----		----
870	D97	9		0.97	6180	ISO3016	9		0.97
875	D97	9		0.97	6192		----		----
886		----		----					

normality	OK
n	87
outliers	1 (+1ex)
mean (n)	6.71
st.dev. (n)	3.004
R(calc.)	8.41
st.dev.(ISO3016:94)	2.354
R(ISO3016:94)	6.59

Lab 342 first reported -3
 Lab 862 test result excluded, PP lower > PP upper which is in principle not possible





Determination of Pour Point (Automated), 3°C interval on sample #18105; results in °C

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913	D6749	9		2.57
92		----		----	922		----		----
120	D5950	0		-1.56	962		----		----
131		----		----	963		----		----
132		----		----	971		----		----
140		----		----	974		----		----
150		----		----	982		----		----
154		----		----	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171		----		----	1107	D5950	0		-1.56
175	D5950	3		-0.18	1109		----		----
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	D5950	0		-1.56
225		----		----	1161	D6749	9		2.57
228		----		----	1167	D6749	6		1.20
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254		----		----	1266	D5950	0.0		-1.56
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299	D97	6	ex	1.20
331	D5950	9		2.57	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402		----		----
339		----		----	1404		----		----
342		----		----	1428	D6749	2		-0.64
343		----		----	1459	In house	0.0		-1.56
349		----		----	1510		----		----
356		----		----	1556		----		----
371		----		----	1569	D5950	3		-0.18
391		----		----	1586	D5950	6		1.20
398		----		----	1613		----		----
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650	D5950	6.0		1.20
463	D6892	6		1.20	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541	D5950	-3		-2.93	1728		----		----
557		----		----	1740		----		----
562		----		----	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833	D5950	6		1.20
631		----		----	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750		----		----	1938		----		----
753		----		----	1950		----		----
759		----		----	1956		9		2.57
781	D5950	6		1.20	1964		----		----
785		----		----	1967		----		----
824		----		----	1995	D5950	-3		-2.93
825		----		----	6016		----		----
850		----		----	6026		----		----
851		----		----	6028	D5950	9		2.57
855		----		----	6039		----		----
858		----		----	6044	D6749	-3	C	-2.93
862	D5950	9		2.57	6054	D5950	0.0		-1.56
863		----		----	6057	D5950	3		-0.18
864		----		----	6075	NF T60-105	0		-1.56
865		----		----	6092		----		----
866		----		----	6142	ISO3016	-3	ex	-2.93
870		----		----	6180		----		----
875		----		----	6192	D5950	0		-1.56
886	D5950	3		-0.18					

normality	OK
n	28
outliers	0 (+2ex)
mean (n)	3.39
st.dev. (n)	4.067
R(calc.)	11.39
st.dev.(D5950:14)	2.179
R(D5950:14)	6.1

Lab 6044 first reported -9
 Lab 1295 and Lab 6142 test results excluded, manual method





Determination of Sediment by Extraction on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D473	0.033		1.28
90	D473	0.013		-0.23	913		----		----
92	D473	0.024		0.60	922	D473	0.016		-0.01
120	D473	0.025		0.67	962		----		----
131	D473	0.01		-0.46	963	D473	0.018		0.15
132	D473	0.01		-0.46	971	ISO3735	0.016		-0.01
140	D473	0.01		-0.46	974	D473	0.017		0.07
150	D473	0.01		-0.46	982	D473	0.016		-0.01
154	D473	0.02		0.30	1006		----		----
158	D473	0.005		-0.84	1011		----		----
159	D473	0.03		1.05	1019		----		----
168	D473	0.03		1.05	1059		----		----
169		----		----	1095		----		----
171		----		----	1107		----		----
175		----		----	1109	D473	0.014		-0.16
194		----		----	1121	ISO3735	0.024		0.60
212	D473	0.01		-0.46	1126		----		----
221	D473	0.016		-0.01	1134	IP53	0.008		-0.61
224	D473	0.0193		0.24	1135	ISO3735	0.01		-0.46
225	D473	0.010		-0.46	1161	ISO3735	0.02		0.30
228	D473	0.019		0.22	1167		----		----
237		----		----	1177		----		----
238	D473	0.009		-0.53	1233		----		----
253		----		----	1259		----		----
254	D473	0.02		0.30	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP53	0.0056		-0.79
323		----		----	1299		----		----
331		----		----	1345	D473	0.015		-0.08
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402	IP53	0.01		-0.46
339		----		----	1404		----		----
342		----		----	1428	D473	0.02		0.30
343		----		----	1459		----		----
349		----		----	1510		----		----
356		----		----	1556	ISO3735	0.005		-0.84
371	D473	0.014		-0.16	1569		----		----
391		----		----	1586	ISO3735	0.01		-0.46
398		----		----	1613	D473	0.015		-0.08
399		----		----	1631		----		----
440		----		----	1643	D473	0.02897		0.97
444		----		----	1650	D473	0.011		-0.38
463	D473	0.006		-0.76	1677	D473	0.0301		1.06
511		----		----	1720		----		----
529		----		----	1724		----		----
541	D473	0.018		0.15	1728	D473	0.020		0.30
557	D473	0.00695908		-0.69	1740	ISO3735	0.015		-0.08
562	D473	0.035		1.43	1796	D473	0.013		-0.23
575	D473	0.020		0.30	1807	D473	0.034		1.35
603	D473	0.01		-0.46	1811		----		----
604		----		----	1832	ISO3735	0.0176		0.12
605	D473	0.03		1.05	1833	ISO3735	0.011		-0.38
631	D473	0.003		-0.99	1849		----		----
633		----		----	1906		----		----
663	D473	0.016		-0.01	1936		----		----
671	D473	0.046	C,R(0.05)	2.26	1937		----		----
750	D473	0.015		-0.08	1938		----		----
753		----		----	1950	D473	0.014		-0.16
759	D473	0.014		-0.16	1956		----		----
781	D473	0.018		0.15	1964		----		----
785	D473	0.019		0.22	1967	D473	0.0231		0.53
824	D473	0.02		0.30	1995		----		----
825		----		----	6016		----		----
850	D473	0.018		0.15	6026		----		----
851		----		----	6028		----		----
855	D473	0.01		-0.46	6039		----		----
858	D473	0.01		-0.46	6044		----		----
862	D473	0.010		-0.46	6054	D473	0.009		-0.53
863	D473	0.018		0.15	6057	D473	0.015		-0.08
864	D473	0.01		-0.46	6075		----		----
865	D473	0.013		-0.23	6092		----		----
866	D473	0.029		0.98	6142		----		----
870	D473	0.012		-0.31	6180	D473	0.011		-0.38
875	D473	0.0148		-0.10	6192		----		----
886		----		----					

normality	OK
n	75
outliers	1
mean (n)	0.0161
st.dev. (n)	0.00738
R(calc.)	0.0207
st.dev.(D473:07e1)	0.01325
R(D473:07e1)	0.0371



Lab 671 first reported 0.072



Determination of Total Sediment (Hot filtration) of sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D4870	0.018		0.28
90		----		----	913		----		----
92		----		----	922	ISO10307-1	0.015		0.05
120	D4870	0.03	R(0.05)	1.24	962		----		----
131	D4870	0.01		-0.35	963	IP375	0.015		0.05
132	ISO10307-1	0.01		-0.35	971	ISO10307-1	0.012		-0.19
140	IP375	<0.01		----	974	IP375	0.014		-0.03
150		----		----	982		----		----
154	D4870	0.02		0.44	1006		----		----
158		----		----	1011	ISO10307-1	0.02		0.44
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171		----		----	1107	IP375	0.014		-0.03
175		----		----	1109		----		----
194	ISO10307-1	0.01		-0.35	1121	IP375	0.011		-0.27
212	ISO10307-1	0.01		-0.35	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	ISO10307-1	<0.01		----
225		----		----	1161		----		----
228		----		----	1167	ISO10307-1	0.02		0.44
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP375	0.0167		0.18
323		----		----	1299	ISO10307-1	<0.01		----
331		----		----	1345	IP375	0.0191		0.37
333		----		----	1356	ISO10307-1	0.01		-0.35
334	IP375	0.03	R(0.05)	1.24	1367	IP375	<0.01		----
337		----		----	1402	IP375	0.02		0.44
339		----		----	1404		----		----
342	ISO10307-1	0.00	R(0.05)	-1.14	1428	ISO10307-1	0.019		0.36
343	ISO10307-1	0.01		-0.35	1459		----		----
349		----		----	1510		----		----
356	IP375	0.01		-0.35	1556		----		----
371	ISO10307-1	0.014		-0.03	1569		----		----
391		----		----	1586	ISO10307-1	0.007		-0.59
398		----		----	1613	IP375	0.016		0.13
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463		----		----	1677	IP375	0.0154		0.08
511		----		----	1720		----		----
529		----		----	1724		----		----
541	D4870	0.0106		-0.30	1728		----		----
557		----		----	1740	ISO10307-1	0.011		-0.27
562		----		----	1796	ISO10307-1	0.01575		0.11
575		----		----	1807	D4870	0.01		-0.35
603		----		----	1811	IP375	0.02		0.44
604		----		----	1832		----		----
605		----		----	1833		----		----
631	D4870	0.0299	R(0.05)	1.23	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750	IP375	0.01		-0.35	1938		----		----
753		----		----	1950	IP375	0.015		0.05
759		----		----	1956		----		----
781	IP375	0.022		0.60	1964		----		----
785	IP375	0.01		-0.35	1967	IP375	0.0212		0.54
824	ISO10307-1	0.02		0.44	1995	IP375	0.02		0.44
825		----		----	6016		----		----
850	ISO10307-1	0.012		-0.19	6026	IP375	0.0167		0.18
851	ISO10307-1	<0.01		----	6028		----		----
855	D4870	0.01		-0.35	6039		----		----
858	D4870	0.01		-0.35	6044	ISO10307-1	0.012		-0.19
862	ISO10307-1	0.014		-0.03	6054		----		----
863	D4870	0.018		0.28	6057	ISO10307-1	0.01		-0.35
864	D4870	0.014		-0.03	6075		----		----
865	D4870	0.013		-0.11	6092	IP375	0.02		0.44
866	ISO10307-1	0.022		0.60	6142		----		----
870	IP375	0.014		-0.03	6180	IP375	0.012		-0.19
875	IP375	0.01		-0.35	6192		----		----
886		----		----					



normality	OK
n	52
outliers	4
mean (n)	0.0144
st.dev. (n)	0.00417
R(calc.)	0.0117
st.dev.(IP375:11)	0.01261
R(IP375:11)	0.0353



Determination of Total Sediment (Accelerated) of sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D4870	0.017		0.28
90		----		----	913		----		----
92		----		----	922	ISO10307-2	0.014		0.04
120	D4870	0.04	R(0.01)	2.17	962		----		----
131	D4870	0.01		-0.29	963	IP390	0.015		0.12
132	ISO10307-2	0.01		-0.29	971	ISO10307-2	0.013		-0.04
140		----		----	974	IP390	0.015		0.12
150		----		----	982		----		----
154	D4870	0.01		-0.29	1006		----		----
158		----		----	1011		----		----
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095	ISO10307-2	0.01		-0.29
171		----		----	1107	IP390	0.012		-0.12
175		----		----	1109		----		----
194		----		----	1121	IP390	0.004		-0.78
212	ISO10307-2	0.01		-0.29	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	ISO10307-2	0.006		-0.62
225		----		----	1161	ISO10307-2	0.02		0.53
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233	ISO10307-2	0.01		-0.29
253		----		----	1259		----		----
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299		----		----
331	ISO10307-2	0.012		-0.12	1345	ISO10307-2	0.020		0.53
333		----		----	1356	ISO10307-2	0.01		-0.29
334	IP390	0.02		0.53	1367	IP390	<0.01		----
337		----		----	1402	IP390	0.02		0.53
339		----		----	1404	ISO10307-2	0.02		0.53
342	ISO10307-2	0.01		-0.29	1428	ISO10307-2	0.021		0.61
343	IP390	0.02		0.53	1459		----		----
349		----		----	1510		----		----
356	IP390	0.01		-0.29	1556	ISO10307-2	0.01		-0.29
371	ISO10307-2	0.014		0.04	1569		----		----
391		----		----	1586	ISO10307-2	0.0065		-0.58
398		----		----	1613		----		----
399		----		----	1631	ISO10307-2	0.013		-0.04
440		----		----	1643		----		----
444		----		----	1650		----		----
463		----		----	1677	IP390	0.0115		-0.17
511		----		----	1720		----		----
529		----		----	1724	IP390	0.011		-0.21
541		----		----	1728		----		----
557	D4870	0.0139071		0.03	1740	ISO10307-2	0.012		-0.12
562		----		----	1796	ISO10307-2	0.01775		0.35
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833	ISO10307-2	0.011		-0.21
631	D4870	0.0326	R(0.01)	1.56	1849	ISO10307-2	0.02		0.53
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750	IP390	0.01		-0.29	1938		----		----
753		----		----	1950	IP390	0.015		0.12
759		----		----	1956		----		----
781	IP390	0.030	R(0.05)	1.35	1964		----		----
785	IP390	0.01		-0.29	1967	IP390	0.0150		0.12
824	ISO10307-2	0.017		0.28	1995		----		----
825		----		----	6016		----		----
850	ISO10307-2	0.014		0.04	6026	IP390	0.0176		0.33
851		----		----	6028		----		----
855	D4870	0.01		-0.29	6039		----		----
858	D4870	0.01		-0.29	6044	ISO10307-2	0.015		0.12
862	ISO10307-2	0.015		0.12	6054		----		----
863	D4870	0.016		0.20	6057	ISO10307-2	0.01		-0.29
864	D4870	0.014		0.04	6075		----		----
865	IP390	0.013		-0.04	6092		----		----
866	ISO10307-2	0.019		0.45	6142		----		----
870	IP390	0.014		0.04	6180	IP390	0.016		0.20
875		----		----	6192	ISO10307-2	0.038	R(0.01)	2.00
886		----		----					



normality	OK
n	54
outliers	4
mean (n)	0.0135
st.dev. (n)	0.00406
R(calc.)	0.0114
st.dev.(IP390:11)	0.01221
R(IP390:11)	0.0342



Determination of Total Sediment (Potential) of sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D4870	0.013		-0.04
90		----		----	913		----		----
92		----		----	922	ISO10307-2	0.015		0.12
120	D4870	0.007		-0.53	962		----		----
131	D4870	0.01		-0.29	963	IP390	0.013		-0.04
132		----		----	971	ISO10307-2	0.013		-0.04
140	IP390	0.02		0.53	974	IP390	0.014		0.04
150		----		----	982		----		----
154		----		----	1006		----		----
158		----		----	1011	ISO10307-2	0.01		-0.29
159		----		----	1019		----		----
168		----		----	1059		----		----
169		----		----	1095		----		----
171		----		----	1107	IP390	0.010		-0.29
175		----		----	1109		----		----
194		----		----	1121	IP390	0.000		-1.11
212	ISO10307-2	0.01		-0.29	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	ISO10307-2	<0.01		----
225		----		----	1161	ISO10307-2	0.02		0.53
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259		----		----
254		----		----	1266		----		----
273		----		----	1269		----		----
313		----		----	1275	IP390	0.0127		-0.07
323	IP390	0.02		0.53	1299	ISO10307-2	0.01		-0.29
331	ISO10307-2	0.014		0.04	1345	ISO10307-2	0.019		0.45
333		----		----	1356	ISO10307-2	0.01		-0.29
334	IP390	0.02		0.53	1367	IP390	<0.01		----
337		----		----	1402	IP390	0.01		-0.29
339		----		----	1404	ISO10307-2	0.02		0.53
342	ISO10307-2	0.01		-0.29	1428	ISO10307-2	0.020		0.53
343	ISO10307-2	0.01		-0.29	1459		----		----
349		----		----	1510		----		----
356	IP390	0.02		0.53	1556		----		----
371	IP390	0.015		0.12	1569		----		----
391		----		----	1586	ISO10307-2	0.014		0.04
398		----		----	1613		----		----
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650		----		----
463		----		----	1677	IP390	0.0045		-0.74
511		----		----	1720		----		----
529		----		----	1724	IP390	0.01		-0.29
541		----		----	1728		----		----
557	D4870	0.0119541		-0.13	1740	ISO10307-2	0.013		-0.04
562		----		----	1796	ISO10307-2	0.01743		0.32
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605		----		----	1833	ISO10307-2	0.0198		0.51
631	D4870	0.0418	R(0.01)	2.31	1849		----		----
633		----		----	1906		----		----
663		----		----	1936		----		----
671		----		----	1937		----		----
750	IP390	0.01		-0.29	1938		----		----
753		----		----	1950	IP390	0.016		0.20
759		----		----	1956		----		----
781	IP390	0.018		0.37	1964		----		----
785	IP390	0.01		-0.29	1967	IP390	0.0216		0.66
824	ISO10307-2	0.01		-0.29	1995		----		----
825		----		----	6016		----		----
850	ISO10307-2	0.011		-0.21	6026	IP390	0.0219		0.68
851		----		----	6028		----		----
855	D4870	0.01		-0.29	6039		----		----
858	D4870	0.01		-0.29	6044	ISO10307-2	0.012		-0.13
862	ISO10307-2	0.013		-0.04	6054		----		----
863	D4870	0.016		0.20	6057	ISO10307-2	0.01		-0.29
864	D4870	0.013		-0.04	6075		----		----
865	ISO10307-2	0.014		0.04	6092		----		----
866	ISO10307-2	0.020		0.53	6142		----		----
870	IP390	0.012		-0.13	6180	IP390	0.015		0.12
875		----		----	6192	ISO10307-2	0.011		-0.21
886		----		----					

normality	OK
n	54
outliers	1
mean (n)	0.0135
st.dev. (n)	0.00460
R(calc.)	0.0129
st.dev.(IP390:11)	0.01222
R(IP390:11)	0.0342





Determination of Total Sulphur on sample #18105; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62	D4294	0.9184		-1.11	912	D4294	0.904		-1.56
90	D4294	0.9755		0.67	913		----		----
92	D4294	1.03		2.37	922	D4294	0.937		-0.53
120	ISO8754	0.966		0.37	962		----		----
131	D4294	0.9995		1.42	963	ISO8754	0.9093		-1.40
132		----		----	971	ISO8754	0.952		-0.06
140	D4294	0.99354		1.23	974	D4294	0.947		-0.22
150	D4294	0.977		0.72	982	D4294	0.939		-0.47
154	D4294	0.979		0.78	1006	D2622	0.999		1.41
158	D4294	0.98281		0.90	1011	ISO8754	0.96		0.19
159	D4294	0.903		-1.59	1019	D1552	0.918		-1.12
168	D4294	0.996		1.31	1059	ISO14596Mod	0.89		-2.00
169	D4294	1.00889		1.71	1095	ISO8754	0.98		0.81
171	D4294	1.00		1.44	1107	ISO8754	0.925		-0.91
175	D4294	0.976		0.69	1109	D4294	0.913		-1.28
194	D4294	1.00		1.44	1121	ISO8754	0.984		0.94
212	ISO8754	0.916		-1.19	1126	In house	0.95		-0.13
221	D4294	0.8963	C	-1.80	1134	IP336	0.9314		-0.71
224	D4294	0.933		-0.66	1135	ISO8754	0.933		-0.66
225	D4294	0.97		0.50	1161	ISO8754	0.9419		-0.38
228	D4294	0.999		1.41	1167	ISO8754	0.9580		0.12
237	D4294	0.95		-0.13	1177	DIN10304-1	0.949		-0.16
238	D4294	0.938		-0.50	1233	ISO8754	0.91		-1.37
253	D4294	1.0		1.44	1259	ISO8754	0.962		0.25
254	D4294	0.9631		0.28	1266	ISO8754	0.97		0.50
273		----		----	1269	In house	1.01		1.75
313	ISO8754	0.97		0.50	1275	IP336	0.933		-0.66
323	ISO8754	0.97		0.50	1299	D2622	0.94		-0.44
331		----		----	1345	D4294	0.945		-0.28
333	ISO8754	0.921		-1.03	1356	ISO8754	0.940		-0.44
334		----		----	1367	ISO8754	0.93	C	-0.75
337	D2622	1.03		2.37	1402	IP336	0.9990		1.41
339	In house	0.93		-0.75	1404	ISO8754	0.955		0.03
342	D4294	0.982		0.87	1428	ISO8754	0.98		0.81
343	IP336	0.99		1.12	1459	ISO8754	0.9438		-0.32
349		----		----	1510		----		----
356	ISO8754	0.93		-0.75	1556	ISO8754	0.936		-0.56
371	D4294	0.997		1.34	1569	ISO8754	0.975		0.66
391	ISO8754	0.949		-0.16	1586	ISO8754	0.946		-0.25
398	ISO8754	0.941		-0.41	1613	D4294	0.919		-1.09
399		----		----	1631	ISO8754	0.93		-0.75
440		----		----	1643	D1552	0.961		0.22
444		----		----	1650	ISO8754	0.957		0.09
463	ISO8754	0.935		-0.59	1677	D4294	0.946		-0.25
511		----		----	1720	D4294	0.897		-1.78
529		----		----	1724	IP336	0.93		-0.75
541	D4294	1.002		1.50	1728	D4294	0.941		-0.41
557	D4294	0.98478		0.96	1740	ISO8754	0.95		-0.13
562		----		----	1796	D4294	0.9392		-0.46
575	D4294	0.9612		0.22	1807	ISO8754	0.92		-1.06
603		----		----	1811	ISO8754	0.979		0.78
604	D4294	0.911		-1.34	1832	ISO8754	0.948		-0.19
605	D4294	0.931		-0.72	1833	ISO8754	0.96		0.19
631	D4294	0.93785		-0.50	1849	ISO8754	1.00		1.44
633		----		----	1906	D5623	0.822	R(0.01)	-4.12
663	D4294	0.9190		-1.09	1936	ISO8754	0.952		-0.06
671	D4294	1.003		1.53	1937	ISO8754	0.947		-0.22
750	D4294	0.932		-0.69	1938	ISO8754	0.940		-0.44
753	ISO8754	0.957		0.09	1950	ISO8754	0.928		-0.81
759	ISO8754	0.945		-0.28	1956	ISO8754	1.01		1.75
781	ISO8754	0.920		-1.06	1964		----		----
785	ISO8754	0.94		-0.44	1967	D4294	0.965		0.34
824	ISO8754	0.967		0.41	1995	D4294	0.923		-0.97
825	D4294	0.973		0.59	6016		----		----
850	D4294	0.937		-0.53	6026	D4294	0.956		0.06
851	ISO8754	0.9572		0.10	6028	ISO8754	0.98	C	0.81
855	D4294	0.948		-0.19	6039	ISO8754	0.96		0.19
858	D4294	0.939		-0.47	6044	D4294	0.961		0.22
862	D4294	0.951		-0.09	6054	D4294	0.94		-0.44
863	D4294	0.950		-0.13	6057	ISO8754	0.96		0.19
864	D4294	0.937		-0.53	6075	ISO8754	0.938		-0.50
865	D4294	0.955		0.03	6092	D4294	0.97		0.50
866	ISO8754	0.950		-0.13	6142	ISO8754	0.9279		-0.82
870	D4294	0.9427		-0.35	6180	ISO8754	0.977		0.72
875	D4294	0.947		-0.22	6192	ISO8754	1.06	C,R(0.05)	3.31
886		----		----					

normality	OK
n	130
outliers	2
mean (n)	0.9540
st.dev. (n)	0.02889
R(calc.)	0.0809
st.dev.(ISO8754:03)	0.03202
R(ISO8754:03)	0.0896
Compare	
R(D4294:16e1)	0.0705



Lab 221 first reported 2.1897
 Lab 1367 first reported 0.093
 Lab 6028 first reported 2.1
 Lab 6192 first reported 1.36



Determination of Water by distillation on sample #18105; results in %V/V

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912	D95	0.05		0.13
90	D95	<0.1		----	913	D95	0.05		0.13
92	D95	<0.05		----	922	D95	<0.05		----
120	ISO3733	<0.05		----	962		----		----
131	D95	0.05		0.13	963	ISO3733	0.05		0.13
132	ISO3733	<0.025		----	971	ISO3733	0.05		0.13
140	D95	0.25	R(0.01)	2.93	974	D95	0.05		0.13
150	D95	<0.05		----	982	D95	<0.05		----
154	D95	0.1		0.83	1006	D95	0.025		-0.22
158		----		----	1011	ISO3733	<0.10		----
159	D95	0.0		-0.57	1019	ISO3733	<0,1		----
168		----		----	1059	ISO3733	<0,05		----
169		----		----	1095	D95	<0.1		----
171		----		----	1107		----		----
175	D95	0.0		-0.57	1109	D95	<0.05		----
194	D95	0.05		0.13	1121	ISO3733	<0.05		----
212	ISO3733	0.00		-0.57	1126	ISO3733	0.00		-0.57
221	D95	<0.05		----	1134	IP74	0.00		-0.57
224		----		----	1135	ISO3733	0.05		0.13
225	D95	0.05		0.13	1161	EN1428	0.1		0.83
228	D95	<0.05		----	1167	EN1428	<0.1		----
237	D95	0.05		0.13	1177		----		----
238		----		----	1233		----		----
253	D95	0.07		0.41	1259	ISO3733	<0.05		----
254	D95	0.05		0.13	1266	D95	0.05		0.13
273	D95	<0.05		----	1269		----		----
313	ISO3733	0.05		0.13	1275		0.00		-0.57
323	ISO3733	<0.05		----	1299	D95	<0.1		----
331	D95	<0.05		----	1345	D95	0.05		0.13
333	ISO3733	0.10		0.83	1356	ISO3733	<0.05		----
334		----		----	1367	ISO3733	<0.05		----
337		----		----	1402	IP74	0.00		-0.57
339		----		----	1404		----		----
342	ISO3733	<0,10		----	1428	D95	0.05		0.13
343	D95	<0,1		----	1459	In house	0		-0.57
349	D95	<0,1		----	1510		----		----
356	ISO3733	0.05		0.13	1556	D6304-C	0.03		-0.15
371	D95	0.048		0.10	1569	D95	<0.10		----
391	ISO3733	0.05		0.13	1586	D95	0		-0.57
398	ISO3733	0.05		0.13	1613	D95	0.05		0.13
399		----		----	1631	D95	<0.1		----
440		----		----	1643	D95	0.03		-0.15
444		----		----	1650	D95	< 0.05		----
463	ISO3733	0.059		0.25	1677	ISO3733	0.05		0.13
511		----		----	1720		----		----
529		----		----	1724	D95	<0,05		----
541	D95	<0.05		----	1728	D95	0.04		-0.01
557	D95	0.0506329		0.13	1740	ISO3733	0.05		0.13
562	D95	0.00		-0.57	1796	ISO3733	0.05		0.13
575	D95	<0.05		----	1807	ISO3733	0		-0.57
603	D95	0.05		0.13	1811		----		----
604		----		----	1832	ISO3733	0.046		0.07
605	D95	0.00		-0.57	1833	D95	<0.1		----
631	D95	0.05		0.13	1849	EN1428	<0,1		----
633	D95	0		-0.57	1906	D6304-C	0.045		0.06
663	D95	<0.1		----	1936	EN1428	<0,1		----
671	D95	0		-0.57	1937	EN1428	0.1		0.83
750	D95	0.1		0.83	1938	EN1428	0.05		0.13
753		----		----	1950	ISO3733	<0.05		----
759	D95	0.05		0.13	1956		----		----
781	ISO3733	0.05		0.13	1964		----		----
785	D95	0.05		0.13	1967	D95	<0.05		----
824	ISO3733	0.00		-0.57	1995	D95	0.05		0.13
825	D95	0.05		0.13	6016		----		----
850	ISO3733	<0.05		----	6026	D95	<0.05		----
851	ISO3733	<0.05		----	6028	D95	0.0		-0.57
855	D95	<0.05		----	6039		----		----
858	D95	<0.05		----	6044	D6304-A	0.132		1.27
862	ISO3733	0.00		-0.57	6054	D95	<0.05		----
863	D95	0		-0.57	6057	ISO3733	0.05		0.13
864	D95	<0.05		----	6075	ISO3733	0.00		-0.57
865	D95	0.03		-0.15	6092	D95	0.05		0.13
866	ISO3733	<0.05		----	6142	D95	0.032		-0.13
870	D95	0		-0.57	6180	ISO3733	0.10		0.83
875	D95	0.10		0.83	6192	ISO3733	0.05		0.13
886	D95	0.05		0.13					



normality	OK
n	74
outliers	1
mean (n)	0.0410
st.dev. (n)	0.03120
R(calc.)	0.0874
st.dev.(ISO3733:99)	0.07143
R(ISO3733:99)	0.2
Compare	
R(D95:13e1)	0.2



Determination of Water and Sediment on sample #18105; results in %V/V

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
62		----		----	912		----		----
90		----		----	913		----		----
92	D1796	0.05		0.08	922	D1796	<0.05		----
120	D1796	0.05		0.08	962		----		----
131	D1796	0.10		1.48	963	D1796	0.05		0.08
132	D1796	0.05		0.08	971	ISO3734	0.05		0.08
140	D1796	0		-1.32	974	D1796	0.05		0.08
150	D1796	0.05		0.08	982		----		----
154	D1796	0.075		0.78	1006	D1796	0.050		0.08
158		----		----	1011		----		----
159	D1796	0.050		0.08	1019		----		----
168		----		----	1059	ISO3734	<0,05		----
169	D1796	0.050		0.08	1095		----		----
171		----		----	1107		----		----
175		----		----	1109	D1796	0.05		0.08
194		----		----	1121		----		----
212		----		----	1126		----		----
221		----		----	1134		----		----
224		----		----	1135	D1796	0.025		-0.62
225		----		----	1161		----		----
228		----		----	1167		----		----
237		----		----	1177		----		----
238		----		----	1233		----		----
253		----		----	1259	ISO3734	0.05		0.08
254		----		----	1266	D1796	0.10		1.48
273		----		----	1269		----		----
313		----		----	1275		----		----
323		----		----	1299	D1796	0.05		0.08
331		----		----	1345		----		----
333		----		----	1356		----		----
334		----		----	1367		----		----
337		----		----	1402		----		----
339		----		----	1404		----		----
342		----		----	1428	D1796	0.05		0.08
343	D1796	0.000		-1.32	1459		----		----
349		----		----	1510		----		----
356		----		----	1556		----		----
371		----		----	1569		----		----
391		----		----	1586	D1796	0.05		0.08
398		----		----	1613	D1796	0.05		0.08
399		----		----	1631		----		----
440		----		----	1643		----		----
444		----		----	1650	D1796	0.05		0.08
463		----		----	1677		----		----
511		----		----	1720		----		----
529		----		----	1724		----		----
541		----		----	1728	D1796	0.060		0.36
557	D1796	0.05		0.08	1740	D1796	0.075		0.78
562	D1796	0.05		0.08	1796		----		----
575		----		----	1807		----		----
603		----		----	1811		----		----
604		----		----	1832		----		----
605	D1796	0.00		-1.32	1833		----		----
631	D1796	0.05		0.08	1849		----		----
633	D1796	0		-1.32	1906		----		----
663	D1796	0.05		0.08	1936		----		----
671	D1796	0		-1.32	1937		----		----
750	D1796	0.075		0.78	1938		----		----
753		----		----	1950		----		----
759		----		----	1956		----		----
781	D1796	0.025		-0.62	1964		----		----
785		----		----	1967	D1796	0.05		0.08
824	D1796	0.05		0.08	1995		----		----
825	D1796	0.05		0.08	6016		----		----
850		----		----	6026		----		----
851		----		----	6028		0.0		-1.32
855		----		----	6039		----		----
858		----		----	6044		----		----
862		----		----	6054		----		----
863		----		----	6057		----		----
864		----		----	6075		----		----
865		----		----	6092		----		----
866		----		----	6142		----		----
870		----		----	6180		----		----
875		----		----	6192		----		----
886	D1796	0.1		1.48					

normality	OK
n	39
outliers	0
mean (n)	0.0471
st.dev. (n)	0.02585
R(calc.)	0.0724
st.dev.(D1796:11)	0.03571
R(D1796:11)	0.1



Vacuum Distillation according to ASTM D1160 (as AET) on sample #18105, results in °C

lab	method	IBP	5%rec	10%rec	20%rec	30%rec	40%rec	50%rec	FBP
62		----	----	----	----	----	----	----	----
90		----	----	----	----	----	----	----	----
92		----	----	----	----	----	----	----	----
120		----	----	----	----	----	----	----	----
131		----	----	----	----	----	----	----	----
132		----	----	----	----	----	----	----	----
140		----	----	----	----	----	----	----	----
150	D1160	226	265	308	378	402	446	492	588
154		----	----	----	----	----	----	----	----
158		----	----	----	----	----	----	----	----
159		----	----	----	----	----	----	----	----
168		----	----	----	----	----	----	----	----
169		----	----	----	----	----	----	----	----
171		----	----	----	----	----	----	----	----
175		----	----	----	----	----	----	----	----
194		----	----	----	----	----	----	----	----
212		----	----	----	----	----	----	----	----
221		----	----	----	----	----	----	----	----
224		----	----	----	----	----	----	----	----
225		----	----	----	----	----	----	----	----
228		----	----	----	----	----	----	----	----
237		----	----	----	----	----	----	----	----
238		----	----	----	----	----	----	----	----
253		----	----	----	----	----	----	----	----
254		----	----	----	----	----	----	----	----
273		----	----	----	----	----	----	----	----
313		----	----	----	----	----	----	----	----
323	D1160	219	282	320	372	408	448	500	----
331		----	----	----	----	----	----	----	----
333		----	----	----	----	----	----	----	----
334		----	----	----	----	----	----	----	----
337		----	----	----	----	----	----	----	----
339		----	----	----	----	----	----	----	----
342		----	----	----	----	----	----	----	----
343		----	----	----	----	----	----	----	----
349		----	----	----	----	----	----	----	----
356		----	----	----	----	----	----	----	----
371		----	----	----	----	----	----	----	----
391		----	----	----	----	----	----	----	----
398		----	----	----	----	----	----	----	----
399		----	----	----	----	----	----	----	----
440		----	----	----	----	----	----	----	----
444		----	----	----	----	----	----	----	----
463		----	----	----	----	----	----	----	----
511		----	----	----	----	----	----	----	----
529		----	----	----	----	----	----	----	----
541		----	----	----	----	----	----	----	----
557		----	----	----	----	----	----	----	----
562		----	----	----	----	----	----	----	----
575		----	----	----	----	----	----	----	----
603		----	----	----	----	----	----	----	----
604		----	----	----	----	----	----	----	----
605		----	----	----	----	----	----	----	----
631		----	----	----	----	----	----	----	----
633		----	----	----	----	----	----	----	----
663		----	----	----	----	----	----	----	----
671		----	----	----	----	----	----	----	----
750	D1160	214	275	320	375	411	447	500	523
753		----	----	----	----	----	----	----	----
759		----	----	----	----	----	----	----	----
781	D1160	206	269	314	371	405	441	496	531
785		----	----	----	----	----	----	----	----
824		----	----	----	----	----	----	----	----
825		----	----	----	----	----	----	----	----
850		----	----	----	----	----	----	----	----
851		----	----	----	----	----	----	----	----
855		----	----	----	----	----	----	----	----
858		----	----	----	----	----	----	----	----
862		----	----	----	----	----	----	----	----
863		----	----	----	----	----	----	----	----
864		----	----	----	----	----	----	----	----
865	D1160	----	----	331.1	379.1	417.8	450.1	----	----
866		----	----	----	----	----	----	----	----
870		----	----	----	----	----	----	----	----
875	D1160	200	266	312	369	405	443	494	----
886		----	----	----	----	----	----	----	----
912		----	----	----	----	----	----	----	----

lab	method	IBP	5%rec	10%rec	20%rec	30%rec	40%rec	50%rec	FBP
913		----	----	----	----	----	----	----	----
922		----	----	----	----	----	----	----	----
962		----	----	----	----	----	----	----	----
963	D1160	214.4	274.4	314.9	370.1	407.0	442.1	499.1	524.5
971	D1160	218.0	280.6	319.4	374.1	410.3	449.4	502.0	526.6
974		----	----	----	----	----	----	----	----
982		----	----	----	----	----	----	----	----
1006		----	----	----	----	----	----	----	----
1011		----	----	----	----	----	----	----	----
1019		----	----	----	----	----	----	----	----
1059		----	----	----	----	----	----	----	----
1095		----	----	----	----	----	----	----	----
1107		218.7	281.7	320.0	373.2	410.0	448.0	500.2	515.2
1109		----	----	----	----	----	----	----	----
1121		----	----	----	----	----	----	----	----
1126		----	----	----	----	----	----	----	----
1134		----	----	----	----	----	----	----	----
1135		----	----	----	----	----	----	----	----
1161		----	----	----	----	----	----	----	----
1167		----	----	----	----	----	----	----	----
1177		----	----	----	----	----	----	----	----
1233		----	----	----	----	----	----	----	----
1259		----	----	----	----	----	----	----	----
1266		----	----	----	----	----	----	----	----
1269		----	----	----	----	----	----	----	----
1275		----	----	----	----	----	----	----	----
1299		----	----	----	----	----	----	----	----
1345		----	----	----	----	----	----	----	----
1356		----	----	----	----	----	----	----	----
1367		----	----	----	----	----	----	----	----
1402	D1160	<u>80.1</u>	289.5	328.6	369.9	406.6	446.5	502.3	530.8
1404		----	----	----	----	----	----	----	----
1428		----	----	----	----	----	----	----	----
1459		----	----	----	----	----	----	----	----
1510		----	----	----	----	----	----	----	----
1556		----	----	----	----	----	----	----	----
1569		----	----	----	----	----	----	----	----
1586		----	----	----	----	----	----	----	----
1613	D1160	244.0	285.5	327.5	378.6	411.2	448.8	505.2	512.8
1631		----	----	----	----	----	----	----	----
1643		----	----	----	----	----	----	----	----
1650		----	----	----	----	----	----	----	----
1677		210	278	319	374	407	445	491	502
1720		----	----	----	----	----	----	----	----
1724		----	----	----	----	----	----	----	----
1728		----	----	----	----	----	----	----	----
1740		----	----	----	----	----	----	----	----
1796	D1160	218	265	323	371 C	404	450	513	517
1807		----	----	----	----	----	----	----	----
1811		----	----	----	----	----	----	----	----
1832		----	----	----	----	----	----	----	----
1833		----	----	----	----	----	----	----	----
1849		----	----	----	----	----	----	----	----
1906		----	----	----	----	----	----	----	----
1936		----	----	----	----	----	----	----	----
1937		----	----	----	----	----	----	----	----
1938		----	----	----	----	----	----	----	----
1950	D1160	197	263	304	359	400	436	481	508
1956		----	----	----	----	----	----	----	----
1964		----	----	----	----	----	----	----	----
1967		194	250	302	366	400	438	484	502
1995		----	----	----	----	----	----	----	----
6016		----	----	----	----	----	----	----	----
6026	D1160	<u>221</u>	<u>303</u>	<u>343 C</u>	<u>392 C</u>	<u>434 C</u>	<u>465 C</u>	<u>516</u>	<u>523</u>
6028		----	----	----	----	----	----	----	----
6039		----	----	----	----	----	----	----	----
6044		----	----	----	----	----	----	----	----
6054		----	----	----	----	----	----	----	----
6057		223.6	281.4	322.8	369.8	403.0	441.2	514.6	<u>566.1</u>
6075		----	----	----	----	----	----	----	----
6092		----	----	----	----	----	----	----	----
6142		----	----	----	----	----	----	----	----
6180	D1160	206.5	271.2	313.1	366.8	405.0	445.8	506.1	553.8
6192		----	----	----	----	----	----	----	----





	IBP	5%rec	10%rec	20%rec	30%rec	40%rec	50%rec	FBP
Normality	suspect	OK	OK	suspect	OK	OK	OK	suspect
n	15	16	17	17	17	17	16	12
outliers	1(+1ex)	0(+1ex)	0(+1ex)	1	1	1	0(+1ex)	2(+1ex)
mean (n)	213.95	273.58	317.61	371.56	406.64	445.05	498.78	520.56
st.dev. (n)	12.640	10.212	8.186	5.009	4.536	4.209	9.117	14.538
R(calc.)	35.39	28.59	22.92	14.03	12.70	11.79	25.53	40.71
st.dev.(D1160:18)	17.500	8.684	7.162	6.082	5.660	6.236	6.662	9.643
R(D1160:18)	49	24.2	20.1	17.1	15.8	17.4	18.5	27





Lab 1796 first reported 351 at 20% recovered

Lab 6026 first reported respectively; 350; 401; 440; 471 for 10%; 20%; 30% and 40% recovered

The reported results underlined and bold are statistical outliers

The reported results underlined and italic are excluded (see §4.1) because the other related test results were statistical outliers





z-scores of Vacuum Distillation according to ASTM D1160 on sample #18105

lab	method	IBP	5%rec	10%rec	20%rec	30%rec	40%rec	50%rec	FBP
62		----	----	----	----	----	----	----	----
90		----	----	----	----	----	----	----	----
92		----	----	----	----	----	----	----	----
120		----	----	----	----	----	----	----	----
131		----	----	----	----	----	----	----	----
132		----	----	----	----	----	----	----	----
140		----	----	----	----	----	----	----	----
150	D1160	0.69	-0.99	-1.34	1.05	-0.82	0.15	-1.03	6.99
154		----	----	----	----	----	----	----	----
158		----	----	----	----	----	----	----	----
159		----	----	----	----	----	----	----	----
168		----	----	----	----	----	----	----	----
169		----	----	----	----	----	----	----	----
171		----	----	----	----	----	----	----	----
175		----	----	----	----	----	----	----	----
194		----	----	----	----	----	----	----	----
212		----	----	----	----	----	----	----	----
221		----	----	----	----	----	----	----	----
224		----	----	----	----	----	----	----	----
225		----	----	----	----	----	----	----	----
228		----	----	----	----	----	----	----	----
237		----	----	----	----	----	----	----	----
238		----	----	----	----	----	----	----	----
253		----	----	----	----	----	----	----	----
254		----	----	----	----	----	----	----	----
273		----	----	----	----	----	----	----	----
313		----	----	----	----	----	----	----	----
323	D1160	0.29	0.97	0.33	0.07	0.24	0.47	0.18	----
331		----	----	----	----	----	----	----	----
333		----	----	----	----	----	----	----	----
334		----	----	----	----	----	----	----	----
337		----	----	----	----	----	----	----	----
339		----	----	----	----	----	----	----	----
342		----	----	----	----	----	----	----	----
343		----	----	----	----	----	----	----	----
349		----	----	----	----	----	----	----	----
356		----	----	----	----	----	----	----	----
371		----	----	----	----	----	----	----	----
391		----	----	----	----	----	----	----	----
398		----	----	----	----	----	----	----	----
399		----	----	----	----	----	----	----	----
440		----	----	----	----	----	----	----	----
444		----	----	----	----	----	----	----	----
463		----	----	----	----	----	----	----	----
511		----	----	----	----	----	----	----	----
529		----	----	----	----	----	----	----	----
541		----	----	----	----	----	----	----	----
557		----	----	----	----	----	----	----	----
562		----	----	----	----	----	----	----	----
575		----	----	----	----	----	----	----	----
603		----	----	----	----	----	----	----	----
604		----	----	----	----	----	----	----	----
605		----	----	----	----	----	----	----	----
631		----	----	----	----	----	----	----	----
633		----	----	----	----	----	----	----	----
663		----	----	----	----	----	----	----	----
671		----	----	----	----	----	----	----	----
750	D1160	0.00	0.16	0.33	0.56	0.77	0.31	0.18	0.25
753		----	----	----	----	----	----	----	----
759		----	----	----	----	----	----	----	----
781	D1160	-0.45	-0.53	-0.50	-0.09	-0.29	-0.65	-0.42	1.08
785		----	----	----	----	----	----	----	----
824		----	----	----	----	----	----	----	----
825		----	----	----	----	----	----	----	----
850		----	----	----	----	----	----	----	----
851		----	----	----	----	----	----	----	----
855		----	----	----	----	----	----	----	----
858		----	----	----	----	----	----	----	----
862		----	----	----	----	----	----	----	----
863		----	----	----	----	----	----	----	----
864		----	----	----	----	----	----	----	----
865	D1160	----	----	1.88	1.23	1.97	0.81	----	----
866		----	----	----	----	----	----	----	----
870		----	----	----	----	----	----	----	----
875	D1160	-0.80	-0.88	-0.78	-0.42	-0.29	-0.33	-0.72	----
886		----	----	----	----	----	----	----	----

lab	method	IBP	5%rec	10%rec	20%rec	30%rec	40%rec	50%rec	FBP
912		----	----	----	----	----	----	----	----
913		----	----	----	----	----	----	----	----
922		----	----	----	----	----	----	----	----
962		----	----	----	----	----	----	----	----
963	D1160	0.03	0.09	-0.38	-0.24	0.06	-0.47	0.05	0.41
971	D1160	0.23	0.81	0.25	0.41	0.65	0.70	0.49	0.63
974		----	----	----	----	----	----	----	----
982		----	----	----	----	----	----	----	----
1006		----	----	----	----	----	----	----	----
1011		----	----	----	----	----	----	----	----
1019		----	----	----	----	----	----	----	----
1059		----	----	----	----	----	----	----	----
1095		----	----	----	----	----	----	----	----
1107		0.27	0.94	0.33	0.27	0.59	0.47	0.21	-0.56
1109		----	----	----	----	----	----	----	----
1121		----	----	----	----	----	----	----	----
1126		----	----	----	----	----	----	----	----
1134		----	----	----	----	----	----	----	----
1135		----	----	----	----	----	----	----	----
1161		----	----	----	----	----	----	----	----
1167		----	----	----	----	----	----	----	----
1177		----	----	----	----	----	----	----	----
1233		----	----	----	----	----	----	----	----
1259		----	----	----	----	----	----	----	----
1266		----	----	----	----	----	----	----	----
1269		----	----	----	----	----	----	----	----
1275		----	----	----	----	----	----	----	----
1299		----	----	----	----	----	----	----	----
1345		----	----	----	----	----	----	----	----
1356		----	----	----	----	----	----	----	----
1367		----	----	----	----	----	----	----	----
1402	D1160	-7.65	1.84	1.53	-0.27	-0.01	0.23	0.53	1.06
1404		----	----	----	----	----	----	----	----
1428		----	----	----	----	----	----	----	----
1459		----	----	----	----	----	----	----	----
1510		----	----	----	----	----	----	----	----
1556		----	----	----	----	----	----	----	----
1569		----	----	----	----	----	----	----	----
1586		----	----	----	----	----	----	----	----
1613	D1160	1.72	1.38	1.38	1.15	0.81	0.60	0.97	-0.80
1631		----	----	----	----	----	----	----	----
1643		----	----	----	----	----	----	----	----
1650		----	----	----	----	----	----	----	----
1677		-0.23	0.51	0.19	0.40	0.06	-0.01	-1.18	-1.92
1720		----	----	----	----	----	----	----	----
1724		----	----	----	----	----	----	----	----
1728		----	----	----	----	----	----	----	----
1740		----	----	----	----	----	----	----	----
1796	D1160	0.23	-0.99	0.75	-0.09	-0.47	0.79	2.15	-0.37
1807		----	----	----	----	----	----	----	----
1811		----	----	----	----	----	----	----	----
1832		----	----	----	----	----	----	----	----
1833		----	----	----	----	----	----	----	----
1849		----	----	----	----	----	----	----	----
1906		----	----	----	----	----	----	----	----
1936		----	----	----	----	----	----	----	----
1937		----	----	----	----	----	----	----	----
1938		----	----	----	----	----	----	----	----
1950	D1160	-0.97	-1.22	-1.90	-2.05	-1.17	-1.45	-2.69	-1.30
1956		----	----	----	----	----	----	----	----
1964		----	----	----	----	----	----	----	----
1967		-1.14	-2.73	-2.18	-0.91	-1.17	-1.13	-2.24	-1.92
1995		----	----	----	----	----	----	----	----
6016		----	----	----	----	----	----	----	----
6026	D1160	0.40	3.40	3.55	3.34	4.83	3.20	2.60	0.25
6028		----	----	----	----	----	----	----	----
6039		----	----	----	----	----	----	----	----
6044		----	----	----	----	----	----	----	----
6054		----	----	----	----	----	----	----	----
6057		0.55	0.90	0.72	-0.29	-0.64	-0.62	2.39	4.72
6075		----	----	----	----	----	----	----	----
6092		----	----	----	----	----	----	----	----
6142		----	----	----	----	----	----	----	----
6180	D1160	-0.43	-0.28	-0.63	-0.78	-0.29	0.12	1.11	3.45
6192		----	----	----	----	----	----	----	----




Determination of Total Carbon, Hydrogen and Nitrogen on sample #18105; results in %M/M

lab	method	Total C	m	z(targ)	Total H	m	z(targ)	Total N	m	z(targ)
62		----		----	----		----	----		----
90		----		----	----		----	----		----
92		----		----	----		----	----		----
120		----		----	----		----	----		----
131		----		----	----		----	----		----
132		----		----	----		----	----		----
140		----		----	----		----	----		----
150	D5291 - C	88.28		0.52	10.61		0.13	<0.75		----
154		----		----	----		----	----		----
158		----		----	----		----	----		----
159		----		----	----		----	----		----
168		----		----	----		----	----		----
169		----		----	----		----	----		----
171		----		----	----		----	----		----
175		----		----	----		----	----		----
194		----		----	----		----	----		----
212		----		----	----		----	----		----
221		----		----	----		----	----		----
224		----		----	----		----	----		----
225		----		----	----		----	----		----
228		----		----	----		----	----		----
237		----		----	----		----	----		----
238		----		----	----		----	----		----
253		----		----	----		----	----		----
254		----		----	----		----	----		----
273		----		----	----		----	----		----
313		----		----	----		----	----		----
323	D5291 - C	87.8		-0.03	10.6		0.09	0.3		-0.27
331		----		----	----		----	----		----
333	D5291 - D	89.1		1.45	10.6		0.09	----		----
334		----		----	----		----	----		----
337		----		----	----		----	----		----
339		87.8		-0.03	10.56		-0.06	0.30		-0.27
342		----		----	----		----	----		----
343	D5291 - A	87.85		0.02	10.8	C	0.83	0.3905		0.30
349		----		----	----		----	----		----
356		----		----	----		----	----		----
371		----		----	----		----	----		----
391		----		----	----		----	----		----
398		----		----	----		----	----		----
399		----		----	----		----	----		----
440		----		----	----		----	----		----
444		----		----	----		----	----		----
463		----		----	----		----	----		----
511		----		----	----		----	----		----
529		----		----	----		----	----		----
541		----		----	----		----	----		----
557		----		----	----		----	----		----
562		----		----	----		----	----		----
575		----		----	----		----	----		----
603		----		----	----		----	----		----
604		----		----	----		----	----		----
605		----		----	----		----	----		----
631		----		----	----		----	----		----
633		----		----	----		----	----		----
663		----		----	----		----	----		----
671		----		----	----		----	----		----
750		----		----	----		----	----		----
753		----		----	----		----	----		----
759		----		----	----		----	----		----
781		----		----	----		----	----		----
785		----		----	----		----	----		----
824	D5291 - D	87.96		0.15	10.51		-0.25	----		----
825		----		----	----		----	----		----
850		----		----	----		----	----		----
851		----		----	----		----	----		----
855		----		----	----		----	----		----
858		----		----	----		----	----		----
862	D5291 - D	87.9		0.08	10.8		0.83	0.3		-0.27
863		----		----	----		----	----		----
864		----		----	----		----	----		----
865		----		----	----		----	----		----
866		----		----	----		----	----		----
870		----		----	----		----	----		----
875		----		----	----		----	----		----
886		----		----	----		----	----		----

lab	method	Total C	m	z(targ)	Total H	m	z(targ)	Total N	m	z(targ)
912		----		----	----		----	----		----
913		----		----	----		----	----		----
922		----		----	----		----	----		----
962		----		----	----		----	----		----
963		----		----	----		----	----		----
971		----		----	----		----	----		----
974		----		----	----		----	----		----
982		----		----	----		----	----		----
1006		----		----	----		----	----		----
1011	D5291 - A	87.25		-0.66	10.64		0.24	0.27		-0.46
1019	D5291 - A	88.08		0.29	10.47		-0.39	0.349		0.03
1059		----		----	----		----	----		----
1095		----		----	----		----	----		----
1107		----		----	----		----	----		----
1109		----		----	----		----	----		----
1121		----		----	----		----	----		----
1126		----		----	----		----	----		----
1134		----		----	----		----	----		----
1135	D5291 - A	87.29		-0.61	10.73		0.57	0.28		-0.40
1161		----		----	----		----	----		----
1167		----		----	----		----	----		----
1177	D5291 - D	89.22		1.59	10.70		0.46	0.461		0.74
1233		----		----	----		----	----		----
1259		----		----	----		----	----		----
1266		----		----	----		----	----		----
1269		----		----	----		----	----		----
1275	D5291 - D	87.64		-0.22	10.60		0.09	0.43		0.54
1299		----		----	----		----	----		----
1345		----		----	----		----	----		----
1356		----		----	----		----	----		----
1367		----		----	----		----	----		----
1402	D5291 - C	88.12		0.33	10.30		-1.03	----		----
1404		----		----	----		----	----		----
1428		----		----	----		----	----		----
1459		----		----	----		----	----		----
1510		----		----	----		----	----		----
1556		----		----	----		----	----		----
1569	D5291 - A	88.47		0.73	10.33		-0.92	0.36		0.10
1586		----		----	----		----	----		----
1613		----		----	----		----	----		----
1631		----		----	----		----	----		----
1643	D5291 - A	86.438		-1.59	10.457		-0.44	0.306		-0.24
1650		----		----	----		----	----		----
1677		----		----	----		----	----		----
1720		----		----	----		----	----		----
1724		----		----	----		----	----		----
1728		----		----	----		----	----		----
1740	D5291 - A	87.4		-0.49	10.9		1.21	0.406		0.39
1796		----		----	----		----	----		----
1807	D5291 - A	87.2		-0.72	10.7		0.46	<0,75		----
1811		----		----	----		----	----		----
1832		----		----	----		----	----		----
1833		----		----	----		----	----		----
1849		----		----	----		----	----		----
1906		88.26		0.49	10.41		-0.62	0.419		0.47
1936		----		----	----		----	----		----
1937		----		----	----		----	----		----
1938		----		----	----		----	----		----
1950		----		----	----		----	----		----
1956		88.6		0.88	10.37		-0.77	0.37		0.17
1964		87.72125		-0.12	10.49789		-0.29	0.32333		-0.13
1967		----		----	----		----	----		----
1995		----		----	----		----	----		----
6016	D5291 - A	86.06		-2.02	10.25		-1.21	0.3214		-0.14
6026		----		----	----		----	----		----
6028		----		----	----		----	----		----
6039	D5291 - C	87.79		-0.04	10.84		0.98	0.33		-0.08
6044		----		----	----		----	----		----
6054		----		----	----		----	----		----
6057		----		----	----		----	----		----
6075		----		----	----		----	----		----
6092		----		----	----		----	----		----
6142		----		----	----		----	----		----
6180	D5291 - D	----		----	12.998	R(1)	9.01	0.267		-0.48
6192		----		----	----		----	----		----

	Total C	Total H	Total N
Normality	OK	OK	OK
n	22	22	18
outliers	0	1	0
mean (n)	87.8286	10.5761	0.3435
st.dev. (n)	0.73793	0.18068	0.05819
R(calc.)	2.0662	0.5059	0.1629
st.dev.(D5291-ABC:16)	0.87627	0.26876	0.15914
R(D5291-ABC:16)	2.4536	0.7525	0.4456

Lab 343 first reported 11.65



- empty page -

Determination of Aluminium as Al, Silicon as Si and total Al+Si on sample #18106; results in mg/kg

lab	method	Al	mark	z(targ)	Si	mark	z(targ)	Sum Al+Si	mark	z(targ)
62		----		----	----		----	----		----
90	D5184	4.6		0.17	4.9		-1.07	9.5		-0.88
92		----		----	----		----	----		----
120	IP501	1.9925	ex	-3.22	25.64	R(0.01)	10.47	27.6325	R(0.01)	8.40
131	IP501	5.441		1.26	6.462		-0.20	11.903		0.35
132		----		----	----		----	----		----
140	IP501	4.22		-0.33	3.06		-2.10	7.28		-2.02
150	IP501	7		3.29	<10		----	<15		----
154	IP501	6		1.99	<10		----	<15		----
159	IP501	4		-0.61	10	C	1.77	14	C	1.42
168	D5184	<5		----	<10		----	<15		----
171	IP501	3.47		-1.30	9.26		1.35	12.73		0.77
175		----		----	----		----	----		----
194	IP470	6	C	1.99	----		----	----		----
212	IP470	0.2	R(0.01)	-5.56	4.5		-1.30	4.7	ex	-3.34
221	IP470	4		-0.61	12		2.88	16		2.44
225		----		----	----		----	----		----
237		----		----	----		----	----		----
254	IP501	5.461		1.29	11.037		2.34	16.50		2.70
273		----		----	----		----	----		----
323	IP501	<5		----	<10		----	<15		----
331	IP501	<5		----	<5		----	<15		----
333	IP501	6		1.99	<10		----	15		1.93
334	IP501	4		-0.61	6		-0.46	----		----
342	IP501	4.5649		0.12	6.7773		-0.03	11.3422		0.06
343	IP501	7		3.29	<10		----	15		1.93
356	IP501	12	R(0.01)	9.80	11		2.32	23	ex	6.03
357	IP501	4.8		0.43	10.4		1.99	15.2		2.03
371	IP470	3.5		-1.26	4.9		-1.07	8.4		-1.45
391	IP501	4		-0.61	5		-1.02	9		-1.14
398		----		----	----		----	----		----
399		----		----	----		----	----		----
444		----		----	----		----	----		----
463	IP470	4.9		0.56	4.55		-1.27	9.45		-0.91
511		----		----	----		----	----		----
529		----		----	----		----	----		----
541	IP470	<5		----	<10		----	<15		----
557	IP470	5.57200		1.43	5.48628		-0.75	11.05828		-0.09
605	IP501	3		-1.91	5		-1.02	8		-1.65
631		----		----	----		----	----		----
663	IP501	4.1		-0.48	4.9		-1.07	9		-1.14
750	IP501	3.8		-0.87	5.6		-0.68	9.4		-0.93
781	IP501	4.4		-0.09	6.7		-0.07	11.1		-0.06
785	IP470	4.8		0.43	8		0.65	12.8		0.81
824	IP501	3.7		-1.00	5.9		-0.52	9.6		-0.83
850	IP501	4.1		-0.48	4.8		-1.13	8.9		-1.19
851	IP501	<5	C	----	<10	C	----	<15	C	----
855	IP501	4.2		-0.35	6.0		-0.46	10.2		-0.53
862	IP501	4.4		-0.09	7.2		0.21	11.6		0.19
863	IP501	4		-0.61	8.2		0.76	12.2		0.50
864		----		----	----		----	----		----
865	IP501	5.8		1.73	11.3		2.49	17.1		3.01
875	IP501	<5		----	<10		----	<15		----
912	IP501	7		3.29	7		0.09	14		1.42
913		----		----	----		----	----		----
922	IP501	<5		----	<10		----	<15		----
963	IP501	3		-1.91	3		-2.13	6		-2.68
971	IP501	4.3		-0.22	6.4		-0.24	10.7		-0.27
974	IP501	4		-0.61	5		-1.02	9		-1.14
1011	ISO10478	4		-0.61	5		-1.02	9		-1.14
1059	In house	10	R(0.01)	7.20	<6		----	----		----
1107	IP501	3.9		-0.74	6.0		-0.46	9.9		-0.68
1109	IP470	3.5		-1.26	5.4		-0.80	8.9		-1.19
1121	IP501	5.4		1.21	4.5		-1.30	9.9		-0.68
1126	IP501	3.9		-0.74	7.9		0.60	----		----
1134	IP501	3.18		-1.68	4.34		-1.39	7.52		-1.90
1135	IP501	5.1		0.82	8.0		0.65	13.1		0.96
1233	IP501	9	R(0.01)	5.89	9		1.21	18	ex	3.47
1259		----		----	----		----	----		----
1275	IP501	4.0365		-0.56	13.985		3.98	18.0215	R(0.01)	3.48
1299		----		----	----		----	----		----
1345		----		----	----		----	----		----
1356	IP501	3		-1.91	6		-0.46	9		-1.14
1367	IP501	4		-0.61	10		1.77	14		1.42
1402	IP501	6		1.99	7		0.09	13		0.91
1404	IP470	3		-1.91	8		0.65	11		-0.12

lab	method	Al	mark	z(targ)	Si	mark	z(targ)	Sum Al+Si	mark	z(targ)
1510		----		----	----		----	----		----
1556	IP501	4.1		-0.48	6.2		-0.35	10.4		-0.42
1586	IP501	<5		----	<10		----	<15		----
1613	IP470	4.6		0.17	4.0		-1.58	8.6		-1.34
1643		----		----	----		----	----		----
1677	IP501	3.43		-1.35	4.89		-1.08	8.32		-1.49
1720	D5708	4.4		-0.09	----		----	----		----
1724	IP501	6.4		2.51	7.7		0.48	14.1		1.47
1740	IP501	4		-0.61	7		0.09	11		-0.12
1833	IP501	4.95		0.62	6.93		0.06	11.88		0.33
1854	IP501	4.4		-0.09	7.0		0.09	11.4		0.09
1950	IP470	3.9		-0.74	7		0.09	10.9		-0.17
1967	IP470	4		-0.61	7		0.09	11		-0.12
1995	IP501	3.8	C	-0.87	9.5		1.49	13.3	C	1.06
6016		----		----	----		----	----		----
6026	IP470	4.0640		-0.53	8.464		0.91	12.528		0.67
6028	D5184	1.7	R(0.01)	-3.60	1.7		-2.86	3.4	ex	-4.01
6044		----		----	----		----	----		----
6054	IP501	4.50675		0.05	6.76267		-0.04	11.26942		0.02
6057	IP501	5		0.69	8		0.65	13		0.91
6075	In house	15.1	R(0.01)	13.83	6.16		-0.37	21.3	ex	5.16
6092	IP501	3		-1.91	6		-0.46	----		----
normality		OK			OK			OK		
n		61			60			53		
outliers		6 (+1ex)			1			1 (+5ex)		
mean (n)		4.47			6.83			11.23		
st.dev. (n)		1.012			2.346			2.508		
R(calc.)		2.83			6.57			7.02		
st.dev.(IP470:05)		0.769			1.796			1.953		
R(IP470:05)		2.15			5.03			5.47		
Compare										
R(IP501:05)		1.51			2.27			2.72		

Lab 120 test results are excluded because of the statistical outliers in the Si and sum Al+Si determination


Lab 159 first reported 14 for Si and 18 for sum Al+Si

Lab 194 first reported 16

Lab 851 first reported 6.564 for Al, 14.13 for Si and 20.694 for sum Al+Si

Lab 1995 first reported 9.7 for Al and 19.2 for sum Al+Si

Labs: 212, 356, 1059, 1233, 6028 and 6075 test results are excluded because of a statistical outlier in the Al determination




Determination of Iron as Fe, Nickel as Ni, Sodium as Na on sample #18106; results in mg/kg

lab	method	Fe	mark	z(targ)	Ni	mark	z(targ)	Na	mark	z(targ)
62		----		----	----		----	----		----
90		----		----	----		----	20		0.97
92		----		----	----		----	----		----
120	IP501	31.215		-1.02	28.93		-1.11	6.351	R(0.05)	-4.30
131	IP501	38.11		0.15	33.78		-0.30	15.64		-0.71
132		----		----	----		----	----		----
140		----		----	----		----	13.75		-1.44
150	IP501	40		0.47	38		0.41	20		0.97
154	IP501	30		-1.23	25		-1.78	18		0.20
159	IP501	34		-0.55	28		-1.27	17		-0.18
168		----		----	----		----	13		-1.73
171	IP501	39.9		0.45	36.02		0.08	20.94		1.34
175		----		----	----		----	----		----
194		----		----	----		----	16		-0.57
212	IP470	39		0.30	36		0.08	22		1.75
221		----		----	----		----	16		-0.57
225		----		----	----		----	----		----
237		----		----	----		----	----		----
254	IP501	39.292		0.35	33.924		-0.27	16.807		-0.26
273		----		----	----		----	----		----
323	IP501	41		0.63	38		0.41	20		0.97
331	IP501	21		-2.75	17		-3.12	17		-0.18
333	IP501	39		0.30	38		0.41	20		0.97
334	IP501	42		0.80	38		0.41	17		-0.18
342	IP501	46.6626		1.59	40.7075		0.87	16.8915		-0.23
343	D5708	30		-1.23	45		1.59	8.89	C	-3.32
356	IP501	34		-0.55	25		-1.78	20		0.97
357	IP501	48.3		1.87	55.7		3.39	26.4		3.44
371	IP470	37.2		-0.01	39.3		0.63	21.0		1.36
391	IP501	33		-0.72	30		-0.93	13		-1.73
398		----		----	----		----	----		----
399		----		----	----		----	----		----
444		----		----	----		----	----		----
463	IP470	39.7		0.41	37.9		0.40	18.45		0.38
511		----		----	----		----	----		----
529		----		----	----		----	----		----
541	IP470	35		-0.38	27		-1.44	18		0.20
557	IP470	39.7346		0.42	41.2629		0.96	18.0419		0.22
605	IP501	40		0.47	38		0.41	12		-2.12
631		----		----	----		----	----		----
663	IP501	40.3		0.52	38.7		0.53	14.7		-1.07
750	IP501	41.1		0.65	39.3		0.63	18.7		0.47
781	IP501	43.5		1.06	44.8		1.56	19.6		0.82
785	IP470	43.8		1.11	38.0		0.41	20.0		0.97
824	IP501	39.1		0.31	38.3		0.46	18.2		0.28
850	IP501	35.6		-0.28	30.0		-0.93	17.2		-0.11
851		----		----	----		----	24		2.52
855	IP501	35.5		-0.30	26.8		-1.47	18.6		0.43
862	IP501	40.7		0.58	33.2		-0.39	19.9		0.93
863	IP501	39.8		0.43	36.7		0.19	17.7		0.09
864		----		----	----		----	----		----
865	IP501	32.9		-0.74	25.5		-1.69	16.8		-0.26
875	IP501	46.2		1.51	41.3		0.97	20.9		1.32
912	IP501	36		-0.21	39		0.58	----		----
913		----		----	----		----	----		----
922	IP501	47		1.65	35		-0.09	19		0.59
963	IP501	36		-0.21	43		1.26	16		-0.57
971	IP501	38.6		0.23	38.8		0.55	18.6		0.43
974	IP501	39		0.30	40		0.75	16		-0.57
1011		----		----	36		0.08	14		-1.34
1059	In house	37		-0.04	42		1.09	----		----
1107	IP501	39.4		0.36	38.9		0.57	18.0		0.20
1109	IP470	37.2		-0.01	32.4		-0.53	17.2		-0.11
1121	IP501	37		-0.04	32.8		-0.46	20.7		1.24
1126	IP501	39.5		0.38	38.1		0.43	19.3		0.70
1134	IP501	26.17		-1.88	26.24		-1.57	14.38		-1.20
1135	IP501	42.0		0.80	37.3		0.30	13.6		-1.50
1233	IP501	31		-1.06	29		-1.10	19		0.59
1259		----		----	----		----	----		----
1275	IP501	38.23		0.17	36.21		0.11	18.35		0.34
1299		----		----	----		----	----		----
1345	IP470	38.31		0.18	24.66		-1.83	19.12		0.63
1356	IP501	20		-2.92	18		-2.95	16		-0.57
1367	IP501	37		-0.04	34		-0.26	16		-0.57
1402	IP501	36		-0.21	30		-0.93	18		0.20
1404	IP470	35		-0.38	33		-0.43	19		0.59

lab	method	Fe	mark	z(targ)	Ni	mark	z(targ)	Na	mark	z(targ)
1510		----		----	----		----	----		----
1556		----		----	38.0		0.41	18.8		0.51
1586	IP501	44		1.14	41		0.92	19		0.59
1613	IP470	44.7		1.26	48.3		2.15	18.6		0.43
1643	D5185	25.29		-2.03	39.28		0.63	10.34		-2.76
1677	IP501	37.0		-0.04	37.4		0.31	16.8		-0.26
1720	D5708	33.7		-0.60	36.0		0.08	21.6	C	1.59
1724	IP501	35.6		-0.28	30.6		-0.83	14.99		-0.96
1740	IP501	39		0.30	36		0.08	17		-0.18
1833	IP501	34.85		-0.41	29.57		-1.01	15.26		-0.86
1854	IP501	45.5		1.40	36.9		0.23	19.3		0.70
1950	IP470	41.3		0.69	36.9		0.23	19.4		0.74
1967	IP470	39		0.30	40		0.75	20		0.97
1995	IP501	9.1	C,R(0.01)	-4.77	20.3		-2.57	13.8		-1.42
6016		----		----	----		----	----		----
6026	IP470	39.9399		0.45	43.8593		1.40	20.2670		1.08
6028	D5185	31.4		-0.99	43.9		1.41	11.3		-2.39
6044	IP501	23.32	C	-2.36	40.65		0.86	10.05		-2.87
6054	IP501	39.8344		0.44	37.4956		0.33	18.5089		0.40
6057	IP501	40		0.47	32		-0.60	19		0.59
6075		----		----	53.9		3.09	14.0		-1.34
6092	IP501	29		-1.40	25		-1.78	16		-0.57
normality		suspect			OK					
n		69			73			76		
outliers		1			0			1		
mean (n)		37.25			35.54			17.48		
st.dev. (n)		5.755			7.158			3.076		
R(calc.)		16.12			20.04			8.61		
st.dev.(IP470:05)		5.906			5.938			2.590		
R(IP470:05)		16.54			16.63			7.25		
Compare										
R(IP501:05)		6.86			11.98			5.15		

Lab 343 first reported 7 for Na
 Lab 1720 first reported 32.5 for Na
 Lab 1995 first reported 20 for Fe
 Lab 6044 first reported 19.62 for Fe




Determination of Vanadium as V, Calcium as Ca and Zinc as Zn on sample #18106; results in mg/kg

lab	method	V	mark	z(targ)	Ca	mark	z(targ)	Zn	mark	z(targ)
62		----		----	----		----	----		----
90	D5863-B	136		1.14	----		----	----		----
92		----		----	----		----	----		----
120	IP501	128.1		0.52	0	ex	-3.17	1.5335		0.79
131	IP501	113.0		-0.66	3.342		-0.87	----		----
132		----		----	----		----	----		----
140	IP501	122		0.05	3.12		-1.02	1.49		0.62
150	IP501	131		0.75	4		-0.42	1	C	-1.29
154	IP501	114		-0.58	4		-0.42	1		-1.29
159	IP501	108		-1.05	5		0.27	----		----
168	D5863-B	80	R(0.01)	-3.23	----		----	----		----
171	IP501	126.3		0.38	5.94		0.91	1.2		-0.51
175	D5863-B	134		0.98	----		----	----		----
194	IP470	123		0.12	----		----	1		-1.29
212	IP470	99		-1.75	4		-0.42	1		-1.29
221	IP470	124		0.20	7		1.64	----		----
225		----		----	----		----	----		----
237		----		----	----		----	----		----
254	IP501	115.617		-0.45	20.557	C,R(0.01)	10.95	1.688		1.39
273		----		----	----		----	----		----
323	IP501	131		0.75	5		0.27	1		-1.29
331	IP501	114		-0.58	<5		----	<5		----
333	IP501	132		0.83	5		0.27	1		-1.29
334	IP501	126		0.36	5		0.27	1		-1.29
342	IP501	123.2090		0.14	5.3425		0.50	1.6104		1.09
343	D5708	139		1.37	5		0.27	<1		----
356	IP501	126		0.36	8		2.33	1.5		0.66
357	IP501	125.6		0.33	7.2		1.78	4.9	R(0.01)	13.90
371	IP470	131.1		0.76	7.4		1.92	1.1		-0.90
391	IP501	115		-0.50	4		-0.42	1		-1.29
398		----		----	----		----	----		----
399		----		----	----		----	----		----
444		----		----	----		----	----		----
463	IP470	124.8		0.26	4.25		-0.25	1.65		1.24
511		----		----	----		----	----		----
529		----		----	----		----	----		----
541	IP470	112		-0.73	4		-0.42	2		2.60
557	IP470	125.298		0.30	4.70778		0.07	1.5077		0.69
605	IP501	128		0.51	3		-1.11	0	ex	-5.19
631		----		----	----		----	----		----
663	IP501	121.4		0.00	6.1		1.02	1.4		0.27
750	IP501	132.0		0.83	4.7		0.06	1.5		0.66
781	IP501	125.5		0.32	3.2		-0.97	1.2		-0.51
785	IP470	111.3		-0.79	3.98		-0.43	0.97		-1.41
824	IP501	127.5		0.47	4.4		-0.15	1.4		0.27
850	IP501	108.0		-1.05	5.2		0.40	1.2		-0.51
851	IP501	121	C	-0.03	5	C	0.27	<1	C	----
855	IP501	115.6		-0.45	6.1		1.02	1.4		0.27
862	IP501	109.4		-0.94	5.1		0.34	1.4		0.27
863	IP501	119.6		-0.14	6		0.95	1.3		-0.12
864		----		----	----		----	----		----
865	IP501	115.8		-0.44	6.4		1.23	1.8		1.83
875	IP501	123		0.12	----		----	1.3		-0.12
912	IP501	115		-0.50	5		0.27	<1		----
913		----		----	----		----	----		----
922	IP501	131		0.75	7		1.64	22	C,R(0.01)	80.50
963	IP501	132		0.83	3		-1.11	2		2.60
971	IP501	126.6		0.40	3.6		-0.70	1.4		0.27
974	IP501	130		0.67	4		-0.42	1		-1.29
1011	D5863-B	109		-0.97	----		----	----		----
1059	In house	126		0.36	12	R(0.01)	5.07	3		6.50
1107	IP501	124.8		0.26	4.3		-0.21	1.4		0.27
1109	IP470	115.6		-0.45	----		----	----		----
1121	IP501	117.5		-0.30	9.5		3.36	1.4		0.27
1126	IP501	124.7		0.26	4.4		-0.15	1.6		1.05
1134	IP501	91.61	R(0.05)	-2.32	3.49		-0.77	0.97		-1.41
1135	IP501	121.7		0.02	3.9		-0.49	<1		----
1233	IP501	131		0.75	10		3.70	4	R(0.01)	10.39
1259		----		----	----		----	----		----
1275	IP501	123.4		0.16	4.326		-0.20	1.481		0.58
1299		----		----	----		----	----		----
1345	IP470	98.50		-1.79	4.73	C	0.08	----		----
1356	IP501	118		-0.27	1		-2.48	<1		----
1367	IP501	110		-0.89	5		0.27	2		2.60
1402	IP501	123		0.12	5		0.27	1		-1.29
1404	IP470	117		-0.34	4		-0.42	0.1		-4.80

lab	method	V	mark	z(targ)	Ca	mark	z(targ)	Zn	mark	z(targ)
1510		----		----	----		----	----		----
1556	IP501	125.6		0.33	4.7		0.06	0.6		-2.85
1586	IP501	131		0.75	3		-1.11	1		-1.29
1613		----	W	----	4.5		-0.08	1.4		0.27
1643	D5185	121.2		-0.02	1.943		-1.83	----		----
1677	IP501	123.7		0.18	4.36		-0.17	1.25		-0.32
1720	D5708	117.9		-0.27	4.1	C	-0.35	----		----
1724	IP501	111.2		-0.80	7.06		1.68	1.9		2.21
1740	IP501	121		-0.03	4		-0.42	1.5		0.66
1833	IP501	112.4		-0.70	3.57		-0.72	1.31		-0.08
1854	IP501	117.3		-0.32	5.0		0.27	1.4		0.27
1950	IP470	123.3		0.15	4.5		-0.08	1.3		-0.12
1967	IP470	127		0.44	4		-0.42	1		-1.29
1995	IP501	26	C,R(0.01)	-7.44	1.3	C	-2.27	0.03	C	-5.07
6016		----		----	----		----	----		----
6026	IP470	113.4852		-0.62	3.8746		-0.51	2.0658		2.86
6028	D5185	154.3	R(0.05)	2.56	0.85		-2.58	4.8	C,R(0.01)	13.51
6044	IP501	127		0.44	1.4		-2.21	1.8		1.83
6054		----		----	4.72469		0.08	----		----
6057	IP501	120		-0.11	4		-0.42	1.5		0.66
6075	D5863-B	128.1		0.52	----		----	----		----
6092	IP501	112		-0.73	4		-0.42	1		-1.29
normality		OK			suspect			not OK		
n		74			68			56		
outliers		4			2 (+1ex)			4 (+1ex)		
mean (n)		121.41			4.61			1.33		
st.dev. (n)		8.228			1.680			0.460		
R(calc.)		23.04			4.71			1.29		
st.dev.(IP470:05)		12.829			1.456			0.257		
R(IP470:05)		35.92			4.08			0.72		
Compare										
R(IP501:05)		29.91			1.74			0.62		

Lab 150 first reported 5 for Zn
 Lab 254 first reported 11.184 for Ca
 Lab 851 first reported 22.66 for V, 5.838 for Ca and 0.696 for Zn
 Lab 922 first reported 28 for Zn
 Lab 1345 first reported 10.73 for Ca
 Lab 1613 first reported 202 for V
 Lab 1720 first reported 10 for Ca
 Lab 1995 first reported 41 for V, 19 for Ca and 15.2 for Zn
 Lab 6028 first reported 3.8 for Zn
 Labs 120 and 605 excluded as zero is not a real test result




Determination of Phosphorus as P on sample #18106; results in mg/kg

lab	method	value	mark	z(targ)	remarks
62		----		----	
90		----		----	
92		----		----	
120	IP501	0.378		-0.58	
131		----		----	
132		----		----	
140	IP501	0		-1.70	
150	IP501	<1		----	
154	IP501	<1		----	
159		----		----	
168		----		----	
171	IP501	<1		----	
175		----		----	
194		----		----	
212	IP500	1		1.28	
221		----		----	
225		----		----	
237		----		----	
254	IP501	0.769		0.59	
273		----		----	
323	IP501	<1		----	
331	IP501	<5		----	
333	IP501	<1		----	
334	IP501	<1		----	
342	IP501	0.2304		-1.02	
343	IP501	<1		----	
356	IP501	0.7		0.39	
357	IP501	7.4	R(0.01)	20.39	
371		----		----	
391	IP501	<1		----	
398		----		----	
399		----		----	
444		----		----	
463	IP500	0.75		0.54	
511		----		----	
529		----		----	
541	IP501	<1		----	
557	IP501	0.381814		-0.56	
605	IP501	0		-1.70	
631		----		----	
663	IP501	0.4		-0.51	
750	IP501	<1		----	
781	IP501	0.4		-0.51	
785		----		----	
824	IP501	0.6		0.09	
850	IP501	0.2		-1.11	
851	IP501	<1	C	----	first reported 1.829
855	IP501	0.5		-0.21	
862	IP501	1.0		1.28	
863	IP501	0.7		0.39	
864		----		----	
865	IP501	0.4		-0.51	
875	IP501	0.7		0.39	
912	IP501	<1		----	
913		----		----	
922		----		----	
963	IP501	2	R(0.05)	4.27	
971	IP501	<1		----	
974	IP501	<1		----	
1011		----		----	
1059	In house	6	R(0.01)	16.21	
1107	IP501	0.4		-0.51	
1109		----		----	
1121	IP501	0.7		0.39	
1126		----		----	
1134	IP501	0.23		-1.02	
1135	IP501	<1		----	
1233	IP501	3	R(0.01)	7.25	
1259		----		----	
1275	IP501	0.345		-0.67	
1299		----		----	
1345		----		----	
1356	IP501	<1		----	
1367	IP501	1		1.28	
1402	IP501	1		1.28	
1404		----		----	

lab	method	value	mark	z(targ)	remarks
1510		----		----	
1556		----		----	
1586	IP501	1		1.28	
1613		----		----	
1643	D5185	0.589		0.05	
1677	IP501	0.56		-0.03	
1720		----		----	
1724	IP501	0.8		0.68	
1740	IP501	0.5		-0.21	
1833	IP501	<1		----	
1854	IP501	0.5		-0.21	
1950		----		----	
1967	IP501	0.7		0.39	
1995	IP501	0.02	C	-1.64	first reported 9.4
6016		----		----	
6026		----		----	
6028	D5185	1.0		1.28	
6044	IP501	0		-1.70	
6054	IP501	1.59561		3.06	
6057	IP501	<1		----	
6075		----		----	
6092	IP501	0.5		-0.21	
normality		OK			
n		36			
outliers		4			
mean (n)		0.57			
st.dev. (n)		0.349			
R(calc.)		0.98			
st.dev.(IP501:05)		0.335			
R(IP501:05)		0.94			
Compare R(IP500:03)		0.97			

Range 0.2 – 40 mg/kg



APPENDIX 2**Number of participants per country****Main round**

1 lab in ARGENTINA
 1 lab in AUSTRALIA
 5 labs in BELGIUM
 1 lab in BRAZIL
 3 labs in CANADA
 1 lab in CHILE
 9 labs in CHINA, People's Republic
 1 lab in COLOMBIA
 2 labs in COTE D'IVOIRE
 1 lab in CROATIA
 1 lab in CZECH REPUBLIC
 1 lab in DJIBOUTI
 1 lab in EGYPT
 2 labs in ESTONIA
 8 labs in FRANCE
 1 lab in GEORGIA
 2 labs in GERMANY
 4 labs in GREECE
 1 lab in GUAM
 1 lab in GUINEA REPUBLIC
 1 lab in HONG KONG
 2 labs in INDIA
 1 lab in IRAN, Islamic Republic of
 1 lab in IRELAND
 3 labs in ITALY
 1 lab in JORDAN
 1 lab in KAZAKHSTAN
 1 lab in KENYA
 2 labs in LATVIA
 1 lab in MACEDONIA
 3 labs in MALAYSIA
 2 labs in MALTA
 1 lab in MARTINIQUE
 1 lab in MEXICO
 1 lab in MOROCCO
 2 labs in NETHERLANDS
 2 labs in NIGERIA
 1 lab in PAKISTAN
 1 lab in PERU
 2 labs in PHILIPPINES
 2 labs in PORTUGAL
 1 lab in ROMANIA
 10 labs in RUSSIAN FEDERATION
 2 labs in SAUDI ARABIA
 1 lab in SENEGAL
 2 labs in SLOVENIA
 1 lab in SOUTH AFRICA
 2 labs in SOUTH KOREA
 8 labs in SPAIN
 1 lab in SUDAN
 4 labs in SWEDEN
 2 labs in TAIWAN
 1 lab in THAILAND
 1 lab in TOGO
 1 lab in TUNISIA
 10 labs in TURKEY
 2 labs in UNITED ARAB EMIRATES
 10 labs in UNITED KINGDOM
 13 labs in UNITED STATES OF AMERICA

Metals in Fuel Oil

1 lab in ARGENTINA
 1 lab in AUSTRALIA
 3 labs in BELGIUM
 1 lab in BRAZIL
 3 labs in CANADA
 6 labs in CHINA, People's Republic
 2 labs in COTE D'IVOIRE
 1 lab in CROATIA
 1 lab in EGYPT
 1 lab in ESTONIA
 1 lab in FINLAND
 4 labs in FRANCE
 1 lab in GEORGIA
 1 lab in GERMANY
 3 labs in GREECE
 1 lab in HONG KONG
 2 labs in INDIA
 3 labs in ITALY
 1 lab in JORDAN
 1 lab in KAZAKHSTAN
 1 lab in KENYA
 1 lab in LATVIA
 1 lab in MALAYSIA
 2 labs in MALTA
 1 lab in MARTINIQUE
 1 lab in MEXICO
 1 lab in MOROCCO
 1 lab in NETHERLANDS
 1 lab in NIGERIA
 1 lab in PAKISTAN
 1 lab in PERU
 1 lab in PHILIPPINES
 1 lab in PORTUGAL
 7 labs in RUSSIAN FEDERATION
 1 lab in SAUDI ARABIA
 1 lab in SENEGAL
 1 lab in SLOVENIA
 1 lab in SOUTH AFRICA
 1 lab in SOUTH KOREA
 4 labs in SPAIN
 1 lab in SUDAN
 3 labs in SWEDEN
 1 lab in THAILAND
 1 lab in TUNISIA
 2 labs in TURKEY
 2 labs in UNITED ARAB EMIRATES
 9 labs in UNITED KINGDOM
 11 labs in UNITED STATES OF AMERICA

APPENDIX 3**Abbreviations:**

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
R(1)	= outlier in Rosner's outlier test
R(5)	= straggler in Rosner's outlier test
E	= probably an error in calculations
U	= test result reported probably in a different unit
W	= test result withdrawn on request of participant
ex	= test result excluded from the statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
SDS	= Material Safety Data Sheet

Literature:

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, March 2017
- 2 ASTM E178:08
- 3 ASTM E1301:03
- 4 ISO 5725:86
- 5 ISO 5725, parts 1-6, 1994
- 6 ISO13528:05
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367:84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee Technical Brief, No 4 January 2001
- 14 P.J. Lowthian and M.Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 15 H. Verplaetse and M. Lacourt, Accred Qual Assur 11, 521-522, 2006
- 16 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, *Technometrics*, 25(2), 165-172, (1983)