

**Results of Proficiency Test
Gasoil B10 (10% FAME)
June 2018**

Organised by: Institute for Interlaboratory Studies (iis)
Spijkenisse, the Netherlands

Authors: ing. C.M. Nijssen-Wester
Correctors: ing. A.S. Noordman-de Neef & ing. G.A. Oosterlaken-Buijs
Report: iis18G03

September 2018

CONTENTS

1	INTRODUCTION.....	3
2	SET UP	3
2.1	ACCREDITATION	3
2.2	PROTOCOL	3
2.3	CONFIDENTIALITY STATEMENT	4
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES	5
2.6	ANALYSES	5
3	RESULTS.....	6
3.1	STATISTICS.....	6
3.2	GRAPHICS.....	7
3.3	Z-SCORES.....	7
4	EVALUATION.....	8
4.1	EVALUATION PER SAMPLE AND PER TEST	8
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	11
4.3	COMPARISON OF THE PROFICIENCY TEST OF JUNE 2018 WITH PREVIOUS PTS	12

Appendices:

1.	Data and statistical results	14
2.	z-scores Distillation	71
3.	Number of participants per country	72
4.	Abbreviations and literature	73

1 INTRODUCTION

Since 2005, the Institute for Interlaboratory Studies organizes a proficiency test for Gasoil B10 containing 6-10% FAME, in accordance with the latest applicable version of the EN590 (0-7% FAME) and ASTM D7467 (6-20% FAME) specifications. During the annual proficiency testing program 2017/2018, it was decided to continue the round robin for the analysis of Gasoil B10.

In this interlaboratory study, a total of 66 laboratories in 35 different countries registered for participation. See appendix 3 for the number of participants per country. In this report, the results of the 2018 Gasoil B10 proficiency test are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organiser of this proficiency test. The sample analyses for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC 17025 accredited laboratory. In this proficiency test, the participants received, depending on their registration, 1 litre plus a 0.5 litre bottle Gasoil B10 (both labelled #18090) and/or a 1 litre bottle with Gasoil B10 (\pm 850 mL filled, labelled #18091) for Total Contamination only.

Participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC 17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4). This protocol can be downloaded via the FAQ page of the iis website www.iisnl.com.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

The necessary sample material of about 180 litres of regular EN590 Diesel (containing no FAME) was purchased from a local petrol station. To this batch 20 litre Biodiesel B100 was added to reach a final FAME concentration of approx. 10%V/V. From this batch, after homogenisation, 90 amber glass bottles of 1 litre and 90 amber glass bottles of 0.5 litre (both labelled #18090) were filled. The homogeneity of the subsamples #18090 was checked by determination of Density at 15°C in accordance with ASTM D4052 on 8 stratified randomly selected samples.

	<i>Density at 15 °C in kg/m³</i>
sample #18090-1	836.53
sample #18090-2	836.51
sample #18090-3	836.52
sample #18090-4	836.52
sample #18090-5	836.52
sample #18090-6	836.52
sample #18090-7	836.52
sample #18090-8	836.52

Table 1: homogeneity test results of subsamples #18090

From the above test results the repeatability was calculated and compared with 0.3 times the corresponding reproducibility of the reference test method in agreement with the procedure of ISO 13528, Annex B2 in the next table:

	<i>Density at 15 °C in kg/m³</i>
r (observed)	0.01
reference test method	ISO12185:96
0.3 * R (reference test method)	0.15

Table 2: evaluation of repeatability of the subsamples #18090

The calculated repeatability was less than 0.3 times the reproducibility of the corresponding reference test method. Therefore, homogeneity of the subsamples #18090 was assumed.

For the preparation of the Total Contamination PT subsamples, the remaining batch of 65 L was used. To each bottle (labelled #18091) Arizona Dust material in an oil suspension was added to give a total contamination of approx 15 mg/kg. To do this, a defined volume of the fresh prepared and well shaken dust suspension was added to an empty bottle by means of a calibrated pipette. The addition was checked by weighing the bottle before and after addition. In total 55 bottles were prepared and subsequently filled up to 850 mL with Gasoil B10. After homogenization, a random sample was taken to verify the actual Total Contamination content (14 mg/kg).

To the participants, depending on their registration, a 1 litre and a 0.5 litre bottle of sample #18090 and/or a 1 litre (\pm 850 mL filled) of sample #18091 were sent on May 16, 2018. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of Gasoil B10, packed in the amber glass bottles, was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYSES

The participants were requested to determine on sample #18090: Acid number (total), Aromatics by FIA, Ash Content, Cetane Indices as per D976 and ISO4264, Cloud Point, Cold Filter Plugging Point (CFPP), Carbon Residue on 10% distillation residue, Ramsbottom Carbon Residue on 10% distillation residue, Copper Corrosion, Density at 15°C, Distillation (IBP, 10%rec., 50%rec., 90%rec., 95%rec., FBP, Volume at 250°C and 350°C), FAME, Flash Point PMcc, Kinematic Viscosity at 40°C, Lubricity by HFRR at 60°C, Oxidation Stability as per ISO12205 and EN15751, Polycyclic Aromatic Hydrocarbons (MAH, DAH, T+AH and Total AH), Pour Point (manual and automated), Sulphur and Water.

The participants were requested to determine on sample #18091 Total Contamination only.

It was explicitly requested to treat the samples as if they were routine samples. Therefore, each laboratory is advised to perform only those analyses that normally are done in daily routine (but the laboratories are allowed to do all analyses). Furthermore, it was requested to report the test results using the indicated units on the report form and not to round the test results more, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical calculations.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the appropriate reference test methods that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment.

Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyses). Additional or corrected test results are used for data analysis and original test results are placed under 'Remarks' in the test result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organisation of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of March 2017 (iis-protocol, version 3.4).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

According to ISO 5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirements based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1 was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report. Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also a normal Gauss curve was projected over the Kernel Density Graph for reference.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM, ISO or EN reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation of this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used. In some cases, a reproducibility based on former IIS proficiency tests could be used.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$Z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The $Z_{(\text{target})}$ scores are listed in the test result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. The usual interpretation of z-scores is as follows:

	$ z < 1$	good
1 <	$ z < 2$	satisfactory
2 <	$ z < 3$	questionable
3 <	$ z $	unsatisfactory

4 EVALUATION

In this proficiency test, no problems were encountered during the dispatch of the samples. All laboratories reported test results, nine laboratories reported test results after the final reporting date and not all laboratories were able to perform all analyses requested. Finally, 66 laboratories did report 1343 numerical test results. Observed were 40 outlying test results, which is 3.0%. In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

4.1 EVALUATION PER SAMPLE AND PER TEST

In this section, the reported test results are discussed per sample and per test. The test methods, which are used by the various laboratories, are taken into account for explaining the observed differences when possible and applicable. These methods are also in the tables together with the reported test results. The abbreviations, used in these tables, are listed in appendix 4.

In the iis PT reports, ASTM methods are referred to with a number (e.g. D976) and an added designation for the year that the method was adopted or revised (e.g. D976:06). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D976:06 (2016)). In the results tables of appendix 1 only the method number and year of adoption or revision (e.g. D976:06) will be used.

Sample #18090

Acid number: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D664-B:17a.

Aromatics (FIA): No z-scores were calculated as the precision and bias of ASTM D1319 with biodiesel blends is not known and is currently under investigation, see paragraph X1.11.1 of ASTM D7467:18.

Ash: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ISO6245:01.

C.I. D976: This determination was not problematic. One statistical outlier was observed and six calculation errors were found. However, the calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ASTM D976:06(2016).

- C.I. ISO4264: This determination was not problematic. One statistical outlier was observed and four calculation errors were found. However, the calculated reproducibility after rejection of the statistical outlier is in good agreement with the requirements of ISO4264:07.
- Cloud Point: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN23015:94.
- CFPP: This determination was problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of EN116:15.
- CR on 10% res.: The consensus value of the group was below the application range (0.1% - 30% M/M) of ISO10370:14. Therefore, no z-scores were calculated.
- Ramsbottom CR: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of ASTM D524:15.
- Copper Corr.: No problems were observed. All reporting participants agreed on a test result of 1 or 1A/1B.
- Density at 15°C: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.
- Distillation: This determination was not problematic. In total eight statistical outliers were observed over eight parameters. However, the calculated reproducibilities after rejection of the statistical outliers are in agreement with the requirements of ISO3405:11 (automated) for all parameters, except for Temperature at 10% recovered, which was not in agreement.
- FAME: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN14078-B:14.
- Flash Point: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of both ISO2719-A:16 and D93-A:16a, but not in agreement with the (more strict) requirements of EN590-Annex A:13.
- Kin. Visc. 40°C: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO3104:94+corr.1997 and with EN590-Annex A:13.

Lubricity by HFRR: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of ISO12156-1-B:16.

Ox. Stab. ISO12205: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12205:95.

Ox. Stab. EN15751: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN15751:14.

PAH: This determination was not problematic. One statistical outlier was observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of EN12916:16.

MAH: This determination was not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of EN12916:16.

DAH: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of EN12916:16.

Tri+-AH: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of EN12916:16.

Total AH: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of EN12916:16.

Pour Point (M): This determination was not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ISO3016:94.

Pour Point (A): This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D5950:14.

Sulphur: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ISO20846:11 and ASTM D5453:16e1.

Water: This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in good agreement with the requirements of ISO12937:00.

Sample #18091

Total Contamination: This determination was problematic. One statistical outlier was observed and four other test results were excluded. The samples were spiked with a freshly prepared suspension of Arizona Dust. Therefore, the minimum Total Contamination to be found was known. The laboratories should be able to find at least 8.4 mg/kg [15.0 mg/kg_(added amount) – 6.6 mg/kg_(R EN12662)]. The calculated reproducibility after rejection of the suspect data is not in agreement with the requirements of EN12664:14.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant reference test method and the reproducibility as found for the group of participating laboratories. The assigned values, calculated reproducibilities and reproducibilities, derived from literature reference test methods (in casu ASTM, ISO, EN reference test methods) are compared in the next table.

<i>Parameters</i>	<i>unit</i>	<i>n</i>	<i>average</i>	<i>2.8 * sd</i>	<i>R (lit)</i>
Acid Number (total)	mgKOH/g	28	0.031	0.046	0.044
Aromatics by FIA	%V/V	13	22.9	9.6	n.a.
Ash content	%M/M	17	0.0007	0.0016	0.005
Cetane Index D976		29	54.0	0.5	2
Cetane Index ISO4264		45	54.6	0.9	2
Cloud Point	°C	53	-7.0	2.5	4
Cold Filter Plugging Point (CFPP)	°C	44	-25.8	5.2	4.6
CR micro method on 10% residue	%M/M	26	0.033	0.045	(0.025)
Ramsbottom CR on 10% residue	%M/M	8	0.088	0.050	0.034
Copper Corrosion 3hrs at 50°C	rating	46	1	n.a.	n.a.
Density at 15°C	kg/m ³	62	836.5	0.2	0.5
Initial Boiling Point	°C	59	177.0	9.8	9.7
10% recovery	°C	60	218.7	5.5	4.8
50% recovery	°C	60	278.4	3.2	3.0
90% recovery	°C	60	334.3	3.5	5.0
95% recovery	°C	61	346.8	7.1	8.7
Final Boiling Point	°C	58	355.7	5.0	7.1
Volume at 250°C	%V/V	48	27.9	1.9	2.7
Volume at 350°C	%V/V	52	95.8	1.8	2.7
Fatty Acid Methyl Ester (FAME)	%V/V	39	8.2	0.5	0.6
Flash Point PMcc	°C	61	67.8	4.2	4.8
Kinematic Viscosity at 40°C	mm ² /s	53	2.897	0.026	0.032
Lubricity by HFRR at 60°C	µm	33	194	91	90
Oxidation Stability ISO12205	g/m ³	18	2.84	5.48	7.74
Oxidation Stability EN15751	hours	24	34.8	5.4	7.0

<i>Parameters</i>	<i>unit</i>	<i>n</i>	<i>average</i>	<i>2.8 * sd</i>	<i>R (lit)</i>
Polycyclic Aromatic Hydrocarbons	%M/M	22	2.51	0.89	0.93
Mono-Aromatic Hydrocarbons	%M/M	21	20.1	3.2	3.2
Di-Aromatic Hydrocarbons	%M/M	22	2.29	0.83	0.92
Tri+-Aromatic Hydrocarbons	%M/M	20	0.22	0.40	0.57
Total Aromatic Hydrocarbons	%M/M	21	22.7	3.9	5.0
Pour Point (manual)	°C	27	-26.3	4.8	6.6
Pour Point (automated)	°C	26	-24.9	3.8	6.1
Sulphur	mg/kg	51	7.3	1.5	1.9
Water content	mg/kg	53	77.6	35.3	60.6
Total Contamination (#18091)	mg/kg	24	15.6	10.5	6.7

Table 3: summary of test results samples #18090 and #18091

Between brackets: evaluation with care as consensus value was below application range of the reference test method

Without further statistical calculations, it can be concluded that for several tests there is a good compliance of the group of participants with the relevant reference test methods. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE INTERLABORATORY STUDY OF JUNE 2018 WITH PREVIOUS PTS

	<i>June 2018</i>	<i>June 2017</i>	<i>June 2016</i>	<i>May 2015</i>	<i>May 2014</i>
Number of reporting labs	66	68	76	73	67
Number of results reported	1343	1444	1522	1371	1317
Statistical outliers	40	33	51	32	33
Percentage outliers	3.0%	2.3%	3.4%	2.3%	2.5%

Table 4: comparison with previous proficiency tests.

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective reference test methods. The conclusions are given the following table:

	<i>June 2018</i>	<i>June 2017</i>	<i>June 2016</i>	<i>May 2015</i>	<i>May 2014</i>
Acid number (total)	+/-	+	++	+	+
Aromatics by FIA	n.e.	n.e.	n.e.	n.e.	n.e.
Ash content	++	++	++	+	++
Cetane Index D976	++	++	++	++	++
Cetane Index ISO4264	++	++	n.e.	n.e.	n.e.
Cloud Point	+	++	++	+	++
Cold Filter Plugging Point (CFPP)	-	+	++	+	+
CR micro method on 10% res.	(-)	(--)	(--)	(-)	(-)
Ramsbottom CR on 10% res.	-	++	+	+/-	n.e.
Density at 15°C	++	+	++	+	+
Distillation	+	+	-	+	+
Fatty Acid Methyl Ester (FAME)	+	-	+/-	-	-
Flash Point PMcc	+	+	+	+	+
Kinematic Viscosity at 40°C	+	+	+	+/-	+
Lubricity by HFRR at 60°C	+/-	+/-	+	-	++
Oxidation Stability ISO12205	+	+	-	+	+

	<i>June 2018</i>	<i>June 2017</i>	<i>June 2016</i>	<i>May 2015</i>	<i>May 2014</i>
Oxidation Stability EN15751	+	--	+	-	+/-
Poly Aromatic Hydrocarbons	+/-	+	-	+	-
Pour Point (manual)	+	+/-	+	+	--
Pour Point (automated)	+	+	++	-	+
Sulphur	+	+	+	+	+/-
Water content	+	++	++	++	++
Total Contamination	-	+/-	+	--	--

Table 5: comparison determinations against the reference test method



Between brackets: consensus value is below application range of the reference test method

The performance of the determinations against the requirements of the respective reference test methods is listed in the above table. The following performance categories were used:

- ++: group performed much better than the reference test method
- + : group performed better than the reference test method
- +/-: group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e.: not evaluated



APPENDIX 1**Determination of Acid Number, total on sample #18090; result in mgKOH/g**

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D664-B	0.020	C	-0.69	first reported: 20
140	D664-B	0.1055	R(0.05)	4.75	
150	D664-B	0.03		-0.05	
171	D664-A	<0.10		----	
175		----		----	
194		----		----	
230	D664-B	0.0026		-1.79	
237	D664-B	<0.1		----	
238		----		----	
312	D974	0.02		-0.69	
323	D664-B	0.09	R(0.05)	3.77	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445		----		----	
447	D664-B	0.04		0.59	
463	D664-B	0.043		0.78	
494	D664-B	0.027		-0.24	
496	D664-B	0.022		-0.56	
511		----		----	
529		----		----	
541	D974	<0.05		----	
551	D974	0.02		-0.69	
603		----		----	
633	D664-B	0.01		-1.32	
663	D664-B	0.030		-0.05	
1017		----		----	
1026	D974	0.03		-0.05	
1033		----		----	
1059		----		----	
1065	D664-B	0.030		-0.05	
1134	D664-B	0.03		-0.05	
1146		----		----	
1161	D664-B	0.03		-0.05	
1194		----		----	
1227	D974	0.0079		-1.46	
1299	D664-B	0.03		-0.05	
1389	D664-B	0.04		0.59	
1397	D664-B	0.07		2.49	
1406		----		----	
1407		----		----	
1441		----		----	
1459	In house	0.07		2.49	
1510		----		----	
1544	D664-B	0.0391		0.53	
1556	D664-B	0.02		-0.69	
1569	D664-B	0.04		0.59	
1631		----		----	
1634		----		----	
1635	D664-B	0.04		0.59	
1706		----		----	
1724	D664-B	0.02		-0.69	
1728	D974	0.0156		-0.97	
1807	D664-B	0.06		1.86	
1811		----		----	
1984		----		----	
2130	D974	0.0252		-0.36	
6016		----		----	
6075		----		----	
	normality	suspect			
	n	28			
	outliers	2			
	mean (n)	0.0308			
	st.dev. (n)	0.01630			
	R(calc.)	0.0456			
	st.dev.(D664-B:17a)	0.01571			
	R(D664-B:17a)	0.0440			





Determination of Aromatics by FIA on sample #18090; result in %V/V

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D1319	24.4		----	
140	D1319	27.7		----	
150	D1319	18.4		----	
171	D1319	25.1		----	
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312		----		----	
323	D1319	15.5		----	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	D1319	24.8		----	
444		----		----	
445	D1319	26.05		----	
447	D1319	23.1		----	
463		----		----	
494		----		----	
496	D1319	20.78		----	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026		----		----	
1033		----		----	
1059		----		----	
1065	D1319	21.3		----	
1134		----		----	
1146		----		----	
1161		----		----	
1194		----		----	
1227		----		----	
1299		----		----	
1389		----		----	
1397	D1319	20.9		----	
1406		----		----	
1407		----		----	
1441	D1319	24.58		----	
1459		----		----	
1510		----		----	
1544	D1319	25.55		----	
1556		----		----	
1569		----		----	
1631		----		----	
1634		----		----	
1635		----		----	
1706		----		----	
1724		----		----	
1728		----		----	
1807		----		----	
1811		----		----	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	13			
	outliers	0			
	mean (n)	22.935			
	st.dev. (n)	3.4107			
	R(calc.)	9.550			
	st.dev.(D1319:15)	n.a.			
	R(D1319:15)	n.a.			compare R(D1319:15) for Gasoil without FAME = 3.7





Determination of Ash Content on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D482	0.0		-0.40	
140	D482	0		-0.40	
150	D482	<0.001		----	
171	D482	<0.001		----	
175		----		----	
194	D482	<0.001		----	
230	ISO6245	0.0005		-0.12	
237	D482	<0.001		----	
238		----		----	
312		----		----	
323	ISO6245	< 0.001		----	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353	IP4	0.0034	G(0.01)	1.51	
381		----		----	
444	D482	<0.001		----	
445	IP4	<0.001		----	
447	D482	<0.001		----	
463	D482	0.00098		0.15	
494	ISO6245	< 0,001		----	
496	ISO6245	0.0000		-0.40	
511	D482	0.0016		0.50	
529	D482	0.000803		0.05	
541	ISO6245	<0.001		----	
551	D482	<0,001		----	
603	D482	0.00059		-0.07	
633	D482	0.0017		0.56	
663	D482	0.0000		-0.40	
1017		----		----	
1026	ISO6245	<0.001		----	
1033		----		----	
1059	ISO6245	<0.001		----	
1065		----		----	
1134	IP4	<0.001		----	
1146	D482	0.0006		-0.06	
1161	ISO6245	0.000883		0.10	
1194		----		----	
1227		----		----	
1299		----		----	
1389	D482	<0.001		----	
1397		----		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459		----		----	
1510	IP4	<0.001		----	
1544	ISO6245	0.00055		-0.09	
1556	ISO6245	0.001		0.16	
1569	ISO6245	<0.001		----	
1631	ISO6245	<0.001		----	
1634		----		----	
1635	ISO6245	<0.001		----	
1706		----		----	
1724	D482	0.0001		-0.34	
1728	D482	0.0011		0.22	
1807	ISO6245	0.0016		0.50	
1811		----		----	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	17			
	outliers	1			
	mean (n)	0.00071			
	st.dev. (n)	0.000581			
	R(calc.)	0.00163			
	st.dev.(ISO6245:01)	0.001786			application range: 0.001 - 0.180%
	R(ISO6245:01)	0.005			compare R(D482:13) = 0.005





Determination of Cetane Index, two variables D976 on sample #18090

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D976	53.73	E	-0.43	iis calculated: 53.91
140	D976	54.1		0.09	
150	D976	54.0		-0.05	
171	D976	54.0		-0.05	
175		----		----	
194	D976	53.90	E	-0.19	iis calculated: 54.09
230	D976	54.1		0.09	
237		----		----	
238		----		----	
312		----		----	
323	D976	54.2		0.23	
334		----		----	
335		----		----	
336	D976	54.2		0.23	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445	D976	53.9		-0.19	
447	D976	54.1		0.09	
463	D976	54.33		0.41	
494		----		----	
496		----		----	
511	D976	53.9		-0.19	
529	D976	54.3169		0.39	
541	D976	54.2		0.23	
551	D976	53.9		-0.19	
603		----		----	
633	D976	53.8		-0.33	
663	D976	53.97		-0.09	
1017		----		----	
1026		----		----	
1033	D976	54.0		-0.05	
1059		----		----	
1065		----		----	
1134	D976	53.8985		-0.19	
1146	D976	54.1		0.09	
1161		----		----	
1194		----		----	
1227	D976	53.983	E	-0.08	iis calculated: 54.28
1299	D976	54.0		-0.05	
1389	D976	54.2		0.23	
1397		----		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459		----		----	
1510		25.8	E,R(0.01)	-39.53	iis calculated: 54.07
1544	D976	54.4		0.51	
1556		----		----	
1569		----		----	
1631		----		----	
1634		----		----	
1635		----		----	
1706		----		----	
1724	D976	53.97	E	-0.09	iis calculated: 53.83
1728	D976	53.6056		-0.60	
1807	D976	53.9		-0.19	
1811	D976	54.3	E	0.37	iis calculated: 54.03
1984		----		----	
2130	D976	54.07		0.05	
6016		----		----	
6075		----		----	
	normality	OK			
	n	29			
	outliers	1			
	mean (n)	54.037			
	st.dev. (n)	0.1847			
	R(calc.)	0.517			
	st.dev.(D976:06)	0.7143			
	R(D976:06)	2			





Determination of Cetane Index, four variables ISO4264 on sample #18090

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D4737-A	54.26		-0.43	
140	D4737-A	54.6		0.05	
150	ISO4264	54.4		-0.23	
171	D4737-A	54.4		-0.23	
175		----		----	
194	D4737-A	54.58		0.02	
230	ISO4264	54.3		-0.37	
237		----		----	
238		----		----	
312		----		----	
323	ISO4264	54.8		0.33	
334		----		----	
335	ISO4264	54.8		0.33	
336	ISO4264	54.6		0.05	
338	ISO4264	54.5		-0.09	
353	IP380	54.399		-0.24	
381	ISO4264	54.6	E	0.05	iis calculated: 54.49 (ISO6246)
444	ISO4264	54.1		-0.65	
445	IP380	54.1		-0.65	
447	D4737-A	54.5		-0.09	
463	ISO4264	54.94		0.52	
494	ISO4264	54.3		-0.37	
496	ISO4264	54.39		-0.25	
511	D4737	54.1		-0.65	
529	D4737	54.840		0.38	
541	D4737-A	54.69		0.17	
551	D4737-A	54.2		-0.51	
603		----		----	
633		----		----	
663	D4737-A	54.59		0.03	
1017	ISO4264	54.75		0.26	
1026	ISO4264	54.8		0.33	
1033	IP380	54.5		-0.09	
1059	ISO4264	54.6		0.05	
1065		----		----	
1134	ISO4264	54.3483		-0.31	
1146		----		----	
1161	ISO4264	54.73		0.23	
1194	D4737-A	58.6	R(0.01)	5.65	
1227		----		----	
1299	D4737-B	54.4	E	-0.23	iis calculated: 53.73 (D4737-B) and 54.43 (D4737-A)
1389	ISO4264	55.0		0.61	
1397	ISO4264	54.8		0.33	
1406	ISO4264	55.44		1.22	
1407	ISO4264	55.1		0.75	
1441	D4737-A	54.3		-0.37	
1459	D4737-A	54.7		0.19	
1510		----		----	
1544		----		----	
1556	ISO4264	54.6		0.05	
1569	ISO4264	54.4	E	-0.23	iis calculated: 54.51 (ISO6246)
1631	ISO4264	55.3		1.03	
1634	ISO4264	54.72		0.21	
1635		----		----	
1706		----		----	
1724	D4737-A	54.48	E	-0.12	iis calculated: 54.31 (D4737-A)
1728	ISO4264	53.9679		-0.84	
1807	ISO4264	54.2		-0.51	
1811		----		----	
1984	ISO4264	54.8		0.33	
2130	IP380	54.61		0.06	
6016		----		----	
6075		----		----	
	normality	OK			
	n	45			
	outliers	1			
	mean (n)	54.567			
	st.dev. (n)	0.3094			
	R(calc.)	0.866			
	st.dev.(ISO4264:07)	0.7143			
	R(ISO4264:07)	2			





Determination of Cloud Point on sample #18090; result in °C

lab	method	value	mark	z(targ)	remarks
62	D5773	-5.8		0.81	
120	D5773	-7		-0.03	
140	D2500	-6		0.67	
150	D5771	-7.0		-0.03	
171	D2500	-3	R(0.01)	2.77	
175	D2500	-6		0.67	
194	D2500	-7		-0.03	
230	D2500	-7		-0.03	
237	D2500	-7		-0.03	
238	D2500	-7		-0.03	
312	D2500	-9		-1.43	
323	EN23015	-8		-0.73	
334	EN23015	-6		0.67	
335	EN23015	-8		-0.73	
336	EN23015	-8		-0.73	
338	EN23015	-7		-0.03	
353	IP219	-6		0.67	
381	ISO3015	-7		-0.03	
444		----		----	
445	IP219	-6		0.67	
447	D2500	-9		-1.43	
463	D2500	-8		-0.73	
494	EN23015	-8		-0.73	
496	EN23015	-7.0		-0.03	
511		----		----	
529	D2500	-6		0.67	
541	D5771	-7.3		-0.24	
551		----		----	
603		----		----	
633		----		----	
663	D2500	-7		-0.03	
1017	D2500	-7		-0.03	
1026	ISO3015	-6		0.67	
1033	D7689	-6.7		0.18	
1059	EN23015	-8		-0.73	
1065	D5771	-5.8		0.81	
1134	IP219	-6		0.67	
1146	D2500	-5.0		1.37	
1161	EN23015	-6		0.67	
1194		----		----	
1227	D2500	-7.6		-0.45	
1299	D2500	-7		-0.03	
1389	D2500	-7		-0.03	
1397	EN23015	-7		-0.03	
1406	ISO3015	-6.0		0.67	
1407		----		----	
1441	D2500	-11	R(0.01)	-2.83	
1459	ISO3015	-8.0		-0.73	
1510	D2500	-7		-0.03	
1544	EN23015	-7.0		-0.03	
1556	ISO3015	-7.5		-0.38	
1569	EN23015	-8	C	-0.73	first reported: -28
1631	EN23015	-5		1.37	
1634		----		----	
1635	D7689	-8		-0.73	
1706		----		----	
1724	D2500	-7		-0.03	
1728	D2500	-7		-0.03	
1807	EN23015	-7		-0.03	
1811	EN23015	-6		0.67	
1984	EN23015	-8		-0.73	
2130	EN23015	-8.0		-0.73	
6016	D2500	-6.4		0.39	
6075	EN23015	-7		-0.03	
	normality	OK			
	n	53			
	outliers	2			
	mean (n)	-6.96			
	st.dev. (n)	0.898			
	R(calc.)	2.52			
	st.dev.(EN23015:94)	1.429			
	R(EN23015:94)	4			compare R(D2500:17a) = 5





Determination of Cold Filter Plugging Point (CFPP) on sample #18090; result in °C

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D6371	-20	R(0.05)	3.59	
140	D6371	-21		2.98	
150	EN116	-27		-0.71	
171	D6371	-25		0.52	
175		----		----	
194		----		----	
230	IP309	-25	C	0.52	first reported: -19
237	D6371	-25		0.52	
238		----		----	
312	EN116	-26		-0.10	
323	EN116	-26		-0.10	
334	EN116	-25		0.52	
335	EN116	-27		-0.71	
336	EN116	-27		-0.71	
338	EN116	-24		1.13	
353	IP309	-25		0.52	
381	EN116	-23		1.75	
444		----		----	
445	IP309	-26		-0.10	
447	IP309	-26		-0.10	
463	EN116	-26		-0.10	
494	EN116	-26		-0.10	
496	EN116	-29.5		-2.25	
511		----		----	
529		----		----	
541	D6371	-26		-0.10	
551		----		----	
603		----		----	
633		----		----	
663	EN116	-27		-0.71	
1017	EN116	-27		-0.71	
1026	EN16329	-25		0.52	
1033	IP309	-28		-1.33	
1059	EN116	-27		-0.71	
1065	D6371	-26		-0.10	
1134	EN116	-28		-1.33	
1146		----		----	
1161	EN116	-21		2.98	
1194	EN116	-19	R(0.05)	4.21	
1227	EN116	-24		1.13	
1299	IP309	-21		2.98	
1389	IP309	-25		0.52	
1397	EN116	-28		-1.33	
1406		----		----	
1407	EN116	-25		0.52	
1441		----		----	
1459	EN116	-28.0		-1.33	
1510	IP309	-19	R(0.05)	4.21	
1544	EN116	-27.5		-1.02	
1556	EN116	-25		0.52	
1569		----		----	
1631	EN116	-27		-0.71	
1634	EN116	-28		-1.33	
1635	EN116	-27		-0.71	
1706	EN116	-27		-0.71	
1724	IP309	-26		-0.10	
1728	D6371	-26		-0.10	
1807	EN116	-25		0.52	
1811	EN116	-20	R(0.05)	3.59	
1984	EN116	-26		-0.10	
2130	EN116	-27.0		-0.71	
6016		----		----	
6075		----		----	
	normality	not OK			
	n	44			
	outliers	4			
	mean (n)	-25.84			
	st.dev. (n)	1.839			
	R(calc.)	5.15			
	st.dev.(EN116:15)	1.625			
	R(EN116:15)	4.55			





Determination of Carbon Residue, micro method on 10% dist. res. on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D4530	0.074		----	
140		----		----	
150		----		----	
171	D4530	<0.1		----	
175		----		----	
194	D4530	0.05		----	
230	ISO10370	0.059		----	
237	D4530	0.004		----	
238		----		----	
312		----		----	
323	ISO10370	< 0.10		----	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353	IP398	0.053		----	
381		----		----	
444		----		----	
445	IP398	<0.1		----	
447	IP398	<0.10		----	
463	ISO10370	0.03		----	
494	ISO10370	0.059		----	
496	ISO10370	0.0215		----	
511	D189	0.026		----	
529		----		----	
541		----		----	
551	D4530	<0.1		----	
603	D4530	0.021		----	
633		----		----	
663	D4530	0.044		----	
1017	ISO10370	0.0286		----	
1026	ISO10370	0.02		----	
1033		----		----	
1059	ISO10370	0.05		----	
1065		----		----	
1134		----		----	
1146		----		----	
1161	ISO10370	0.02		----	
1194		----		----	
1227	D4530	0.023		----	
1299	D4530	0.03		----	
1389	D4530	0.04		----	
1397	ISO10370	0.021		----	
1406	D4530	0.026		----	
1407	ISO10370	0.032		----	
1441		----		----	
1459		----		----	
1510	IP398	<0.10		----	
1544	ISO10370	0.0345		----	
1556	ISO10370	0.02		----	
1569	ISO10370	0.024		----	
1631	ISO10370	<0.10		----	
1634		----		----	
1635	ISO10370	0.03		----	
1706		----		----	
1724	D4530	<0,1		----	
1728	ISO10370	0.0255		----	
1807		----		----	
1811		----		----	
1984		----		----	
2130	ISO10370	<0.01		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	26			
	outliers	0			
	mean (n)	0.0333			
	st.dev. (n)	0.01591			
	R(calc.)	0.0446			
	st.dev.(ISO10370:14)	(0.00906)			application range: 0.1 – 30%M/M
	R(ISO10370:14)	(0.0254)			compare R(EN590:13-Annex A) = 0.0388



Determination of Ramsbottom Carbon Residue on 10% dist. res. on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D524	0.091		0.25	
140	D524	0.06		-2.34	
150	D524	0.09		0.17	
171	D524	0.09		0.17	
175	D524	0.078		-0.83	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312		----		----	
323		----		----	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445	IP14	0.10		1.00	
447		----		----	
463		----		----	
494		----		----	
496		----		----	
511	D524	0.075		-1.09	
529		----		----	
541		----		----	
551	D524	0.12		2.67	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026		----		----	
1033		----		----	
1059		----		----	
1065		----		----	
1134		----		----	
1146		----		----	
1161		----		----	
1194		----		----	
1227		----		----	
1299		----		----	
1389		----		----	
1397		----		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459		----		----	
1510		----		----	
1544		----		----	
1556		----		----	
1569		----		----	
1631		----		----	
1634		----		----	
1635		----		----	
1706		----		----	
1724		----		----	
1728		----		----	
1807		----		----	
1811		----		----	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	unknown			
	n	8			
	outliers	0			
	mean (n)	0.0880			
	st.dev. (n)	0.01788			
	R(calc.)	0.0501			
	st.dev.(D524:15)	0.01198			
	R(D524:15)	0.0335			





Determination of Copper Strip 3hrs at 50 °C on sample #18090

lab	method	value	mark	z(targ)	remarks
62	D130	1b		----	
120	D130	1A		----	
140	D130	1a		----	
150	D130	1a		----	
171	D130	1a		----	
175		----		----	
194		----		----	
230	D130	1a		----	
237	D130	1		----	
238	D130	1a		----	
312		----		----	
323	D130	1A		----	
334	D130	1a		----	
335		----		----	
336	D130	1		----	
338		----		----	
353	IP154	1a		----	
381		----		----	
444		----		----	
445	IP154	1a		----	
447	D130	1a		----	
463	ISO2160	1A		----	
494	ISO2160	1a		----	
496	ISO2160	1a		----	
511	D130	1A		----	
529	D130	1A		----	
541	D130	1A		----	
551	D130	1A		----	
603	D130	1A		----	
633	D130	1a		----	
663	D130	1a		----	
1017	D130	1A		----	
1026	ISO2160	1A		----	
1033	IP154	1b		----	
1059	ISO2160	1a		----	
1065		----		----	
1134	D130	1a		----	
1146		----		----	
1161	ISO2160	1a		----	
1194		----		----	
1227	D130	1A		----	
1299	D130	1A		----	
1389	D130	1A		----	
1397	ISO2160	1		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459		----		----	
1510	D130	1a		----	
1544	ISO2160	1a		----	
1556	ISO2160	class 1a		----	
1569	ISO2160	1a		----	
1631	ISO2160	1		----	
1634	ISO2160	1a		----	
1635	D130	1A		----	
1706		----		----	
1724	D130	1a		----	
1728	D130	1A		----	
1807	D130	1A		----	
1811		----		----	
1984		----		----	
2130	IP154	1a		----	
6016		----		----	
6075	ISO2160	1a		----	
	n	46			
	mean	1 (1A/1B)			

- empty page -

Determination of Density at 15 °C on sample #18090; result in kg/m³

lab	method	value	mark	z(targ)	remarks
62	D4052	836.54		0.00	
120	D4052	836.4		-0.79	
140	D4052	836.6		0.33	
150	ISO12185	836.7		0.89	
171	D4052	836.6		0.33	
175	D4052	836.55		0.05	
194	D4052	836.6		0.33	
230	ISO12185	836.56		0.11	
237	D4052	836.5		-0.23	
238	D4052	836.37		-0.95	
312	D4052	836.5		-0.23	
323	ISO12185	836.4		-0.79	
334	ISO12185	836.5		-0.23	
335	ISO12185	836.6		0.33	
336	ISO12185	836.3		-1.35	
338	ISO12185	836.6		0.33	
353	IP365	836.5		-0.23	
381	ISO12185	836.5		-0.23	
444	D4052	836.7		0.89	
445	IP365	836.5		-0.23	
447	D4052	836.5		-0.23	
463	ISO12185	836.54		0.00	
494	ISO12185	836.6		0.33	
496	ISO12185	836.62		0.45	
511	D4052	836.5		-0.23	
529	D4052	836.54		0.00	
541	ISO12185	836.5		-0.23	
551	D4052	836.6		0.33	
603	D4052	836.58		0.22	
633	D1298	837.12	R(0.01)	3.25	
663	D4052	836.55		0.05	
1017	D4052	836.6		0.33	
1026	D4052	836.5		-0.23	
1033	IP365	836.6		0.33	
1059	ISO12185	836.5		-0.23	
1065	D4052	836.6		0.33	
1134	IP365	836.7		0.89	
1146	D4052	836.56		0.11	
1161	ISO12185	836.49		-0.28	
1194	ISO12185	836.6	C	0.33	reported: 0.8366 kg/m ³ (probably a unit error?)
1227	D4052	836.6	C	0.33	first reported: 0.8366
1299	D4052	836.5		-0.23	
1389	D4052	836.6		0.33	
1397	ISO12185	836.5		-0.23	
1406	ISO12185	836.56		0.11	
1407	ISO12185	836.5		-0.23	
1441	D4052	836.54		0.00	
1459	ISO12185	836.50		-0.23	
1510	IP365	836.6	C	0.33	reported: 0.8366 kg/m ³ (probably a unit error?)
1544	ISO12185	836.60		0.33	
1556	ISO12185	836.6		0.33	
1569	ISO12185	836.5		-0.23	
1631	ISO12185	836.5		-0.23	
1634	ISO12185	836.538		-0.01	
1635	ISO12185	836.5		-0.23	
1706	ISO12185	836.4		-0.79	
1724	D1298	836.9	R(0.01)	2.01	
1728	D4052	836.53		-0.06	
1807	ISO12185	836.5		-0.23	
1811	ISO12185	836.6		0.33	
1984	ISO12185	836.4		-0.79	
2130	ISO12185	836.6		0.33	
6016	D4052	836.50		-0.23	
6075	ISO12185	836.60		0.33	
	normality	suspect			
	n	62			
	outliers	2			
	mean (n)	836.540			
	st.dev. (n)	0.0762			
	R(calc.)	0.213			
	st.dev.(ISO12185:96)	0.1786			
	R(ISO12185:96)	0.5			















Determination of Distillation on sample #18090; result in °C

lab method	IBP	mark	10%rec	mark	50%rec	mark	90%rec	mark	95%rec	mark	FBP	mark
62	----		----		----		----		----		----	
120 D86-automated	174.9		217.0		277.3		335.7		350.6		355.4	
140 D86-automated	179.9		219.1		278.5		334.4		346.6		354.3	
150	178.2		218.6		278.1		333.2		344.8		353.2	
171 D86-automated	175.2		217.8		278.3		335.1		348.8		353.8	
175 D86-automated	176.4		219.1		279.2		337.2		351.6		354.8	
194 D86-automated	179.0		219.0		278.6		335.3		347.7		356.6	
230	177.0		216.5		278.5		333.5		345.5		359.5	
237 D86-manual	175.0		213.0		276.0		333.0		345.0		356.0	
238 D86-manual	178		222		281		334		344		355	
312 D86-automated	179.4		220.0		279.6		334.5		347.4		354.2	
323 ISO3405-automated	181.6		219.9		278.7		335.3		349.7		355.6	
334	169.6		214.6		276.5		333.3		339.6		352.7	
335	----		----		----		----		----		----	
336 D86-automated	176.1		219.0		278.4		334.0		346.7		356.8	
338 ISO3405-automated	176.3		217.6	C	278.8		335.1		347.1		356.7	
353 D86-automated	179.2	C	218.1		277.7		337.7		352.6		358.0	
381	169.1		217.3		278.7		335.4		348.1		355.9	
444 D86-automated	172.4		218.4		276.5		332.2		342.8		356.4	
445 IP123-automated	171.3		216.6		277.2		333.2		346.0		353.8	
447 D86-automated	169.3		218		278.6		334.6		347.5		357.4	
463 ISO3405-automated	181.2		220.2		279.7		336.8		349.9		358.9	
494 ISO3405-automated	171.5		217.8		277.5		334.5		347.8		356.6	
496 D86-automated	168.8		219.3		277.5		332.7		344.5		355.3	
511 D86-manual	180.5		215.5		277.5		333		345.5		355	
529 D86	175.1		219.4		279.6		335.7		348.4		356.7	
541 ISO3405-automated	178.9		219.3		278.9		334.4		347.3		356.1	
551 D86-automated	174.4		217.4		277.5		332.6		343.9		353.2	
603 D86-automated	177.5		218.5		279.5		333.9		345.4		356.6	
633 D86-automated	176.9		219.6		278.0		334.4		345.0		346.3	R(0.01)
663 D86-automated	181.25		220.55		277.90		333.00		345.20		355.15	
1017 ISO3405-automated	180.48		220.45		278.78		334.98		347.99		357.38	
1026 ISO3405-automated	176.2		220.1		279.0		335.1		348		356.2	
1033 IP123-automated	176.4		218.5		278.4		334.4		346.8		355.7	
1059 ISO3405-automated	178.2		218.3		279.1		334.7		347.9		356.9	
1065 D7345	174.5	C	212.5	C	277.2		333.7		346.4		353.9	
1134 D86-automated	176.1		218.6		277.8		333.7		342.2		351.8	
1146 D86-automated	175.8		217.5		278.8		336.1		349.4		358.0	
1161 ISO3405-automated	178.2		218.8		279.3		335.4		347.9		355.1	
1194	----		----		----		----		----		----	
1227 D86-automated	180.8		220.7		279.6		335.4		348.9		355.8	
1299 D86-automated	177.7		218.1		278.0		334.1		346.8		355.4	
1389 D86-automated	178.3		222.5		279.6		335.2		347.8		357.6	
1397	179.9		220.6		279.0		334.9		348.0		357.2	
1406 ISO3405-manual	189.5	R(0.05)	222.5		281.5		336.5		353.0		357.0	
1407 ISO3405-automated	185.6		222.1		279.6		333.7		345.1		357.7	
1441 D86-automated	175.4		218.3		277.0		332.5		345.0		353.7	
1459 ISO3405-automated	179.6		219.0		279.0		332.6		344.8		352.8	
1510 D86-automated	177.8		218.7		278.5		332.1		344.9		351.2	
1544 ISO3405-automated	176.05		221.65		280.05		336.70		350.80		357.40	
1556 ISO3405-automated	182.1		218.7		278.9		335.2		348.0		356.7	
1569 ISO3405-automated	173.8		220.3		277.3		333.9	C	352.1		----	
1631 ISO3405-automated	----		----		----		----		346.5		----	
1634 D86-automated	179.7		220.6		278.7		333.9		345.8		357.7	
1635	178.0		216.0		276.0		333.0		343.0		358.0	
1706 ISO3405-automated	182.45		220.05		279.15		333.7		346.2		353.95	
1724 D86-automated	177		219.1		277.8		333.9		347.4		355.6	
1728 ISO3405-manual	175.0		217.0		276.0		334.5		347.0		355.5	
1807 D86-automated	177.5		217.3		277.3		333.2		344.7		356.9	
1811 D86-automated	177.8		218.6		278.3		333.2		344.3		355.1	
1984 ISO3405-automated	180.4		220.2		278.5		334.9		347.8		356.8	
2130 IP123-automated	180.0		219.6		278.8		333.9		346.4		355.8	
6016 D86-automated	171.3		216.9		276.9		333.1		344.8		351.7	
6075 ISO3405	177.2		217.8		278.5		333.0		345.6		356.0	
normality	OK		suspect		OK		OK		OK		OK	
n	59		60		60		60		61		58	
outliers	1		0		0		0		0		1	
mean (n)	177.00		218.67		278.37		334.31		346.82		355.69	
st.dev. (n)	3.514		1.980		1.124		1.265		2.518		1.791	
R(calc.)	9.84		5.54		3.15		3.54		7.05		5.02	
st.dev.(ISO3405-A:11)	3.477		1.718		1.061		1.79		3.122		2.536	
R(ISO3405-A:11)	9.74		4.81		2.97		5.01		8.74		7.1	
comp R(ISO3405-M:11)	6.67		4.50		3.72		3.82		4.64		3.84	





Lab 338 first reported for T. at 10% rec.: 203.8
 Lab 353 first reported for IBP: 206.7

Lab 1065 first reported for IBP: 166.2 and for T. at 10% rec.: 208.6
 Lab 1569 first reported for T. at 90% rec.: 345.4





Determination of Distillation on sample #18090; result in %V/V

Lab method	vol at 250°C	mark	z(targ)	vol at 350°C	mark	z(targ)	% residue	remark
62	----		----	----		----	----	
120 D86-automated	28.2		0.32	94.9		-0.96	1.2	C f.r. 97.5
140 D86-automated	27.8		-0.09	95.8		-0.03	1.4	
150	28.0		0.11	96.9		1.11	1.4	
171 D86-automated	----		----	----		----	0.9	
175 D86-automated	----		----	----		----	1.3	
194 D86-automated	----		----	----		----	1.4	
230	27.8		-0.09	96.6		0.80	1.3	
237 D86-manual	68.0	R(0.01)	41.60	30.0	R(0.01)	-68.27	96.0	
238 D86-manual	25	R(0.01)	-3.00	96.5		0.70	1.5	
312 D86-automated	27.0		-0.92	95.6		-0.24	2.0	
323 ISO3405-automated	27.5		-0.40	95.1		-0.76	1.4	
334	29.1		1.25	96.3		0.49	1.4	
335	----		----	----		----	----	
336 D86-automated	27.9		0.01	95.8		-0.03	1.7	
338 ISO3405-automated	27.6		-0.30	95.8		-0.03	0.2	
353 D86-automated	27.3		-0.61	94.4		-1.48	1.2	
381	27.53		-0.37	95.47		-0.37	1.6	
444 D86-automated	29.1		1.25	97.0		1.22	1.4	
445 IP123-automated	28.7		0.84	96.0		0.18	1.6	
447 D86-automated	27.6		-0.30	95.5		-0.34	1.4	
463 ISO3405-automated	27.0		-0.92	95.0		-0.86	1.6	
494 ISO3405-automated	28.1		0.22	95.4		-0.44	0	
496 D86-automated	28.7		0.84	96.4		0.59	----	
511 D86-manual	31	R(0.01)	3.22	96.5		0.70	1.4	
529 D86	27.4		-0.51	95.6		-0.24	1.4	
541 ISO3405-automated	27.7		-0.20	95.6		-0.24	0.5	
551 D86-automated	29		1.15	97		1.22	1.2	
603 D86-automated	30.5		2.71	96.3		0.49	1.4	
633 D86-automated	----		----	----		----	2.0	
663 D86-automated	28.20		0.32	96.45		0.64	1.55	
1017 ISO3405-automated	27.8		-0.09	95.5		-0.34	1.4	
1026 ISO3405-automated	27.1		-0.82	95.5		-0.34	1.7	
1033 IP123-automated	27.8		-0.09	95.9		0.07	1.4	
1059 ISO3405-automated	27.9		0.01	95.5		-0.34	1.4	
1065 D7345	----		----	----		----	----	
1134 D86-automated	28.3		0.42	96.0		0.18	1.4	
1146 D86-automated	28		0.11	95		-0.86	1.2	
1161 ISO3405-automated	27.5		-0.40	95.2		-0.65	1.2	
1194	----		----	----		----	----	
1227 D86-automated	26.8		-1.13	95.2		-0.65	1.0	
1299 D86-automated	28.1		0.22	95.9		0.07	1.4	
1389 D86-automated	26.6		-1.34	95.6		-0.24	1.4	
1397	27.8		-0.09	95.4		-0.44	1.1	
1406 ISO3405-manual	25.5	R(0.01)	-2.48	94.5		-1.38	1.0	
1407 ISO3405-automated	27.2		-0.72	96.2		0.39	0.7	
1441 D86-automated	----		----	----		----	1.4	
1459 ISO3405-automated	27.5		-0.40	96.3		0.49	1.4	
1510 D86-automated	28.0		0.11	96.2		0.39	1.5	
1544 ISO3405-automated	26.80		-1.13	94.70		-1.17	1.60	
1556 ISO3405-automated	27.6		-0.30	95.5		-0.34	1.4	
1569 ISO3405-automated	----		----	----		----	1.4	
1631 ISO3405-automated	27.5		-0.40	95.8		-0.03	----	
1634 D86-automated	28.0		0.11	96.5		0.70	1.0	
1635	34	R(0.01)	6.34	97		1.22	1.5	
1706 ISO3405-automated	28.0		0.11	95.95		0.13	1.8	
1724 D86-automated	28.2		0.32	95.6		-0.24	1.3	
1728 ISO3405-manual	27.5		-0.40	95.5		-0.34	1.4	
1807 D86-automated	28.7		0.84	96.3		0.49	1.2	
1811 D86-automated	28.1		0.22	96.6		0.80	1.2	
1984 ISO3405-automated	28.2		0.32	95.9		0.07	1.4	
2130 IP123-automated	27.8		-0.09	95.8		-0.03	1.0	
6016 D86-automated	----		----	----		----	----	
6075 ISO3405	28.2		0.32	96.1		0.28	1.1	
normality	not OK			OK				
n	48			52				
outliers	5			1				
mean (n)	27.89			95.83				
st.dev. (n)	0.689			0.625				
R(calc.)	1.93			1.75				
st.dev.(ISO3405-A:11)	0.964			0.964				
R(ISO3405-A:11)	2.7			2.7				
comp R(ISO3405-M:11)	5.13			5.34				





Determination of Fatty Acid Methyl Esters (FAME) content on sample #18090; result in %V/V

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D7371	8.21		0.20	
140	D7371	7.7		-2.16	
150	D7371	8.54		1.72	
171		----		----	
175		----		----	
194		----		----	
230	EN14078 - B	7.97		-0.91	
237		----		----	
238		----		----	
312	EN14078 - B	8.15		-0.08	
323	EN14078 - B	8.2		0.15	
334	EN14078 - B	8.0		-0.77	
335		----		----	
336	EN14078 - B	8.2		0.15	
338	EN14078 - B	8.29		0.57	
353	EN14078 - B	8.235		0.31	
381	EN14078 - B	8.2		0.15	
444		----		----	
445	EN14078 - B	8.26		0.43	
447	EN14078 - B	8.0		-0.77	
463	EN14078 - B	8.453		1.32	
494	EN14078 - B	8.11		-0.27	
496	EN14078 - B	8.28		0.52	
511	D7371	8.126		-0.19	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017	EN14078 - B	8.33		0.75	
1026	EN14078 - B	8.1		-0.31	
1033		----		----	
1059	EN14078 - B	8.3		0.61	
1065		----		----	
1134		----		----	
1146	D7371	9.0356	R(0.01)	4.01	
1161	EN14078 - B	8.20		0.15	
1194		8.7		2.46	
1227		----		----	
1299	EN14078 - B	8.1		-0.31	
1389	EN14078 - B	8.1		-0.31	
1397	EN14078 - B	8.1		-0.31	
1406	EN14078 - B	8.15		-0.08	
1407	EN14078 - B	8.32		0.70	
1441	EN14078 - B	9.05	C,R(0.01)	4.08	first reported: 9.5
1459	EN14078 - B	8.0		-0.77	
1510		----		----	
1544	EN14078 - B	8.19		0.10	
1556	EN14078 - B	8.192		0.11	
1569	EN14078 - B	7.95		-1.00	
1631		----		----	
1634	EN14078 - B	8.145		-0.10	
1635	EN14078 - B	7.2	R(0.01)	-4.47	
1706	EN14078 - B	8.0		-0.77	
1724	EN14078 - B	8.3		0.61	
1728	EN14078 - B	8.48		1.44	
1807	EN14078 - B	7.9		-1.24	
1811	EN14078 - A	8.2		0.15	
1984	EN14078 - B	7.79		-1.74	
2130	EN14078 - B	8.06		-0.50	
6016		----		----	
6075		----		----	
	normality	suspect			
	n	39			
	outliers	3			
	mean (n)	8.167			
	st.dev. (n)	0.1921			
	R(calc.)	0.538			
	st.dev.(EN14078-B:14)	0.2166			
	R(EN14078-B:14)	0.606			compare R(D7371:14) = 1.101





Determination of Flash Point PMcc on sample #18090; result in °C

lab	method	value	mark	z(targ)	remarks
62	D93-C	68.0		0.11	
120	D93-A	66.7		-0.64	
140	D93-A	68		0.11	
150	D93-A	67.0		-0.47	
171	ISO2719-A	67.8	C	0.00	first reported: 59
175	D93-A	68		0.11	
194	D93-A	66.0		-1.05	
230	ISO2719-A	66.0		-1.05	
237	D93-A	68.0		0.11	
238	D93-A	66.0		-1.05	
312	D93-A	67.5		-0.18	
323	ISO2719-A	68.0		0.11	
334	ISO2719-A	68.5		0.40	
335	ISO2719-A	69.0		0.69	
336	ISO2719-A	69.5		0.99	
338	ISO2719-A	71.5		2.15	
353	IP34-A	69.0		0.69	
381	ISO2719-A	66.0		-1.05	
444	D93-A	68		0.11	
445	D93-A	69.0		0.69	
447	D93-A	69.0		0.69	
463	D93-A	68.5		0.40	
494	ISO2719-A	68.0		0.11	
496	ISO2719-A	68.2		0.23	
511	D93	66.5		-0.76	
529	D93	67		-0.47	
541	ISO2719-A	66.0		-1.05	
551	D93-A	69.0		0.69	
603	D93-A	69.0		0.69	
633		----		----	
663	D93-A	68.35		0.32	
1017	D93-A	69.0		0.69	
1026	ISO2719-A	67.0		-0.47	
1033	IP34-A	69.0		0.69	
1059	ISO2719-A	67.5		-0.18	
1065	D93-A	71		1.86	
1134	D93-A	67.0		-0.47	
1146	D93-A	66.2		-0.93	
1161	ISO2719-A	63.5		-2.50	
1194		----		----	
1227	D93-A	65.7		-1.22	
1299	D93-A	69.5		0.99	
1389	D93-A	65.0		-1.63	
1397	ISO2719-A	68.0		0.11	
1406	ISO2719-A	67.0		-0.47	
1407	ISO2719-A	67.5		-0.18	
1441	D93-A	66.5		-0.76	
1459	ISO2719-A	65.5		-1.34	
1510	IP34-A	67.0		-0.47	
1544	ISO2719-A	68.00		0.11	
1556	ISO2719-A	68.5		0.40	
1569	D93-A	68		0.11	
1631	ISO2719-A	67.0		-0.47	
1634	D93-A	68.2		0.23	
1635	ISO2719-A	70.0		1.28	
1706		----		----	
1724	D93-A	68		0.11	
1728	D93-A	68.0		0.11	
1807	ISO2719-A	72.0		2.44	
1811	D93-A	67.5		-0.18	
1984	ISO2719-A	68.0		0.11	
2130	ISO2719-A	69.0		0.69	
6016	D93-A	67.0		-0.47	
6075	ISO2719-A	68.0		0.11	
	normality	suspect			
	n	61			
	outliers	0			
	mean (n)	67.806			
	st.dev. (n)	1.4869			
	R(calc.)	4.163			
	st.dev.(ISO2719-A:16)	1.7194			compare R(D93-A:16a) =4.81
	R(ISO2719-A:16)	4.814			compare R(EN590-Annex A:13) = 3.5





Determination of Kinematic Viscosity at 40°C on sample #18090; result in mm²/s

lab	method	value	mark	z(targ)	remarks
62	D445	2.8996		0.25	
120	D445	2.8951		-0.14	
140	D445	2.89		-0.59	
150	D445	2.897		0.03	
171	ISO3104	2.901		0.38	
175	D445	2.8983		0.14	
194		----		----	
230	ISO3104	2.8911		-0.49	
237	D445	2.9237		2.37	
238		----		----	
312	D445	2.9127		1.40	
323	ISO3104	2.883		-1.20	
334	ISO3104	2.894		-0.24	
335		----		----	
336		----		----	
338	ISO3104	2.898		0.11	
353	IP71	2.8942		-0.22	
381	D445	2.888		-0.76	
444		----		----	
445	IP71	2.8936		-0.27	
447	D445	2.899		0.20	
463	ISO3104	2.8925		-0.37	
494	ISO3104	2.889		-0.67	
496	ISO3104	2.8940		-0.24	
511	D445	2.889		-0.67	
529	D445	2.9091		1.09	
541	ISO3104	2.8885		-0.72	
551	D445	2.901	C	0.38	first reported: 2.938
603	D445	2.886		-0.94	
633		----		----	
663	D445	2.88930		-0.65	
1017		----		----	
1026	ISO3104	2.900		0.29	
1033	IP71	3.67	R(0.01)	67.76	
1059	ISO3104	2.883		-1.20	
1065	D445	2.9157		1.66	
1134	IP71	2.9023		0.49	
1146	D445	2.8914		-0.46	
1161	ISO3104	2.896		-0.06	
1194		----		----	
1227	D445	2.8881		-0.75	
1299	D445	2.903		0.55	
1389	D445	2.914		1.52	
1397	D7042	2.884		-1.11	
1406	ISO3104	2.884		-1.11	
1407	In house	2.889		-0.67	
1441	D445	2.900		0.29	
1459	D7042	2.9077		0.96	
1510	IP71	2.886		-0.94	
1544	ISO3104	2.9005		0.33	
1556	ISO3104	2.8956	C	-0.10	first reported: 3.284
1569	ISO3104	2.901		0.38	
1631	ISO3104	2.896		-0.06	
1634	ISO3104	2.897		0.03	
1635	ISO3104	2.899		0.20	
1706	ISO3104	2.901		0.38	
1724	D445	2.894		-0.24	
1728	D445	2.9082		1.01	
1807	ISO3104	2.886		-0.94	
1811	ISO3104	2.9030		0.55	
1984	ISO3104	2.887		-0.85	
2130	ISO3104	2.916	C	1.69	first reported: 2.961
6016		----		----	
6075	ISO3104	2.854	R(0.01)	-3.74	
	normality	OK			
	n	53			
	outliers	2			
	mean (n)	2.8967			
	st.dev. (n)	0.00922			
	R(calc.)	0.0258			
	st.dev.(ISO3104:94)	0.01141			
	R(ISO3104:94)	0.0320			compare R(EN590-Annex A:13) = 0.0521





Determination of Lubricity by HFRR at 60°C on sample #18090; result in µm

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D6079	252		1.81	
140	D6079	231		1.16	
150	D6079	250		1.75	
171	ISO12156-1 meth A	208		0.44	
175		----		----	
194		----		----	
230		----		----	
237	D6079	240		1.44	
238		----		----	
312	ISO12156-1 meth A	270		2.37	
323	ISO12156-1 meth A	150		-1.36	
334	ISO12156-1 meth B	200		0.19	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445	ISO12156-1 meth A	155		-1.21	
447	IP450	181		-0.40	
463	ISO12156-1 meth A	209		0.47	
494	ISO12156-1 meth A	181		-0.40	
496	ISO12156-1 meth A	184		-0.30	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	ISO12156-1 (2006)	151		-1.33	
1033	ISO12156-1 meth A	181		-0.40	
1059	ISO12156-1 meth B	170		-0.74	
1065	IP450	168		-0.80	
1134		----		----	
1146		----		----	
1161	ISO12156-1 meth A	180.0		-0.43	
1194		----		----	
1227		----		----	
1299	ISO12156-1 (2006)	167		-0.83	
1389	ISO12156-1 meth A	157		-1.14	
1397	ISO12156-1 meth B	171		-0.71	
1406		----		----	
1407		----		----	
1441	D6079	223		0.91	
1459		----		----	
1510		----		----	
1544	ISO12156-1 (2006)	175.5		-0.57	
1556	ISO12156-1 meth A	201		0.22	
1569	ISO12156-1 (2006)	187		-0.21	
1631	ISO12156-1 meth B	186		-0.24	
1634		----		----	
1635	ISO12156-1 meth A	180		-0.43	
1706		----		----	
1724	IP450	154		-1.24	
1728		----		----	
1807	ISO12156-1	180		-0.43	
1811	ISO12156-1 (2006)	203		0.29	
1984	ISO12156-1 meth A	175		-0.58	
2130	ISO12156-1 meth A	238		1.38	
6016		----		----	
6075	ISO2156-A	236.0		1.31	
	normality	OK			
	n	33			
	outliers	0			
	mean (n)	193.77			ISO12156-1 method A = digital camera and method B = visual
	st.dev. (n)	32.444			
	R(calc.)	90.84			compare R(ISO12156-1-A:16) = 80 (digital camera)
	st.dev.(ISO12156-1-B:16)	32.143			compare R(D6079:11) = 80 (digital camera)
	R(ISO12156-1-B:16)	90			compare R(IP450:00) = 102 (visual)





Determination of Oxidation Stability ISO12205 on sample #18090; result in g/m³

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D2274	3.1		0.10	
140		----		----	
150	D2274	1		-0.66	
171	ISO12205	3		0.06	
175		----		----	
194		----		----	
230	ISO12205	0.86		-0.71	
237		----		----	
238		----		----	
312	ISO12205	1		-0.66	
323	ISO12205	5		0.78	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445	IP388	1		-0.66	
447	ISO12205	2.0		-0.30	
463	ISO12205	1.43		-0.51	
494		----		----	
496	ISO12205	0.86		-0.71	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	ISO12205	<1		----	
1033	D2274	2.72		-0.04	
1059	ISO12205	1		-0.66	
1065		----		----	
1134		----		----	
1146		----		----	
1161	ISO12205	2.9		0.02	
1194		----		----	
1227		----		----	
1299	D2274	7.4		1.65	
1389	D2274	5		0.78	
1397		----		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459		----		----	
1510		----		----	
1544		----		----	
1556		----		----	
1569		----		----	
1631	ISO12205	4		0.42	
1634		----		----	
1635		----		----	
1706		----		----	
1724	D2274	2.86		0.01	
1728		----		----	
1807	ISO12205	<1		----	
1811		----		----	
1984	ISO12205	5.9		1.11	
2130	D2274	15	R(0.01)	4.40	
6016		----		----	
6075	ISO12205	10.5	R(0.05)	2.77	
	normality	OK			
	n	18			
	outliers	2			
	mean (n)	2.835			
	st.dev. (n)	1.9557			
	R(calc.)	5.476			
	st.dev.(ISO12205:95)	2.7624			
	R(ISO12205:95)	7.735			





Determination of Oxidation Stability induction period EN15751 on sample #18090; result in hours

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	EN15751	16	R(0.01)	-7.51	
140		----		----	
150		----		----	
171		----		----	
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN15751	32.3		-0.98	
323		----		----	
334	EN15751	36.5		0.70	
335		----		----	
336	EN15751	36.4		0.66	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445	EN15751	37.22		0.99	
447	EN15751	35.8		0.42	
463		----		----	
494	EN15751	33.7		-0.42	
496	EN15751	32.39		-0.95	
511		----		----	
529		----		----	
541		----		----	
551	EN15751	34		-0.30	
603		----		----	
633		----		----	
663		----		----	
1017	EN15751	36.99		0.90	
1026	EN15751	33.79		-0.39	
1033	EN15751	34.35		-0.16	
1059	EN15751	34.4		-0.14	
1065	EN15751	37.6		1.14	
1134		----		----	
1146		----		----	
1161		----		----	
1194		----		----	
1227	EN15751	36.26		0.60	
1299	EN15751	35.1		0.14	
1389		----		----	
1397		----		----	
1406		----		----	
1407		----		----	
1441		----		----	
1459	EN15751	37.88		1.25	
1510	EN14112	33.9		-0.34	
1544	EN15751	35.88		0.45	
1556	EN15751	34.52		-0.09	
1569	EN15751	32.8		-0.78	
1631		----		----	
1634		----		----	
1635	EN15751	36.63		0.75	
1706		----		----	
1724	EN15751	32.37		-0.95	
1728		----		----	
1807	EN15751	31.8		-1.18	
1811	EN15751	24	R(0.01)	-4.31	
1984	EN15751	31.5		-1.30	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	24			
	outliers	2			
	mean (n)	34.753			
	st.dev. (n)	1.9459			
	R(calc.)	5.448			
	st.dev.(EN15751:14)	2.4961			
	R(EN15751:14)	6.989			





Determination of Polycyclic Aromatic Hydrocarbons on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120		----		----	
140		----		----	
150		----		----	
171	EN12916	1.3	ex	-3.64	excluded, DAH value corrected without correcting PAH
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN12916	2.84		1.01	
323	EN12916	2.8		0.89	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	EN12916	2.1		-1.22	
444		----		----	
445	IP391	2.47		-0.11	
447	IP391	2.4		-0.32	
463	EN12916	2.54		0.10	
494	EN12916	2.89		1.16	
496	EN12916	2.71		0.62	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	EN12916	2.2		-0.92	
1033		----		----	
1059	EN12916	2.5		-0.02	
1065		----		----	
1134		----		----	
1146		----		----	
1161	EN12916	2.1		-1.22	
1194		1.5	R(0.05)	-3.03	
1227	EN12916	2.9		1.19	
1299	EN12916	2.5		-0.02	
1389		2.643		0.41	
1397	EN12916	2.38		-0.38	
1406		----		----	
1407		----		----	
1441		----		----	
1459		2.8		0.89	
1510		----		----	
1544		----		----	
1556	EN12916	2.566		0.18	
1569	EN12916	2.1		-1.22	
1631	EN12916	3.0		1.49	
1634		----		----	
1635	EN12916	1.8		-2.13	
1706		----		----	
1724		2.69		0.56	
1728		----		----	
1807		----		----	
1811	IP391	2.20		-0.92	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	22			
	outliers	1 (+1ex)			
	mean (n)	2.506			
	st.dev. (n)	0.3191			
	R(calc.)	0.894			
	st.dev.(EN12916:16)	0.3316			
	R(EN12916:16)	0.929			





Determination of Mono-Aromatic Hydrocarbons on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120		----		----	
140		----		----	
150		----		----	
171	EN12916	20.0	C	-0.08	first reported: 19.3
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN12916	20.71		0.54	
323	EN12916	21.8		1.48	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	EN12916	20.7		0.53	
444		----		----	
445	IP391	20.33		0.21	
447	IP391	19.8		-0.25	
463	EN12916	19.51		-0.50	
494	EN12916	20.37		0.24	
496	EN12916	20.33		0.21	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	EN12916	14.3	R(0.01)	-5.01	
1033		----		----	
1059	EN12916	19.9		-0.16	
1065		----		----	
1134		----		----	
1146		----		----	
1161	EN12916	19.5		-0.51	
1194		----		----	
1227	EN12916	20.0		-0.08	
1299	EN12916	20.2		0.10	
1389		20.328		0.21	
1397	EN12916	19.98	C	-0.09	first reported: 15.2
1406		----		----	
1407		----		----	
1441		----		----	
1459		23.1	C	2.60	first reported: 26.3
1510		----		----	
1544		----		----	
1556	EN12916	21.22		0.98	
1569	EN12916	18.20		-1.63	
1631		----		----	
1634		----		----	
1635	EN12916	18.2		-1.63	
1706		----		----	
1724		19.7		-0.34	
1728		----		----	
1807		----		----	
1811	IP391	17.990		-1.82	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	suspect			
	n	21			
	outliers	1			
	mean (n)	20.089			
	st.dev. (n)	1.1580			
	R(calc.)	3.242			
	st.dev.(EN12916:16)	1.1560			
	R(EN12916:16)	3.237			





Determination of Di-Aromatic Hydrocarbons on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120		----		----	
140		----		----	
150		----		----	
171	EN12916	2.4	C	0.34	first reported: 1.3
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN12916	2.47		0.56	
323	EN12916	2.7		1.26	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	EN12916	2.1		-0.57	
444		----		----	
445	IP391	2.18		-0.33	
447	IP391	2.3		0.04	
463	EN12916	2.40		0.34	
494	EN12916	2.66		1.14	
496	EN12916	2.54		0.77	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	EN12916	1.6		-2.10	
1033		----		----	
1059	EN12916	2.5		0.65	
1065		----		----	
1134		----		----	
1146		----		----	
1161	EN12916	2.0		-0.88	
1194		----		----	
1227	EN12916	2.6		0.95	
1299	EN12916	2.3		0.04	
1389		2.313		0.08	
1397	EN12916	2.16		-0.39	
1406		----		----	
1407		----		----	
1441		----		----	
1459		2.5		0.65	
1510		----		----	
1544		----		----	
1556	EN12916	2.513		0.69	
1569	EN12916	1.9		-1.18	
1631		----		----	
1634		----		----	
1635	EN12916	1.7		-1.79	
1706		----		----	
1724		2.41		0.37	
1728		----		----	
1807		----		----	
1811	IP391	2.090		-0.60	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	22			
	outliers	0			
	mean (n)	2.288			
	st.dev. (n)	0.2959			
	R(calc.)	0.829			
	st.dev.(EN12916:16)	0.3277			
	R(EN12916:16)	0.918			





Determination of Tri+-Aromatic Hydrocarbons on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120		----		----	
140		----		----	
150		----		----	
171	EN12916	0.0		-1.07	
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN12916	0.40		0.91	
323	EN12916	0.1		-0.57	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	EN12916	<0,1		----	
444		----		----	
445	IP391	0.28		0.32	
447	IP391	0.1		-0.57	
463	EN12916	0.14		-0.37	
494	EN12916	0.23		0.07	
496	EN12916	0.17		-0.23	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	EN12916	0.6		1.90	
1033		----		----	
1059	EN12916	<0,1		----	
1065		----		----	
1134		----		----	
1146		----		----	
1161	EN12916	0.1		-0.57	
1194		----		----	
1227	EN12916	0.3		0.42	
1299	EN12916	0.2		-0.08	
1389		0.329		0.56	
1397	EN12916	0.22		0.02	
1406		----		----	
1407		----		----	
1441		----		----	
1459		0.4		0.91	
1510		----		----	
1544		----		----	
1556	EN12916	0.053		-0.80	
1569	EN12916	0.20		-0.08	
1631		----		----	
1634		----		----	
1635	EN12916	0.1		-0.57	
1706		----		----	
1724		0.28		0.32	
1728		----		----	
1807		----		----	
1811	IP391	0.111		-0.52	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	suspect			
	n	20			
	outliers	0			
	mean (n)	0.216			
	st.dev. (n)	0.1432			
	R(calc.)	0.401			
	st.dev.(EN12916:16)	0.2023			
	R(EN12916:16)	0.566			





Determination of Total Aromatic Hydrocarbons on sample #18090; result in %M/M

lab	method	value	mark	z(targ)	remarks
62		----		----	
120		----		----	
140		----		----	
150		----		----	
171	EN12916	22.6	C	-0.03	first reported: 2036
175		----		----	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	EN12916	23.55		0.50	
323	EN12916	24.6		1.09	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353		----		----	
381	EN12916	22.8		0.08	
444		----		----	
445	IP391	22.81		0.08	
447	IP391	22.2		-0.26	
463	EN12916	22.05		-0.34	
494	EN12916	23.26		0.34	
496	EN12916	23.04		0.21	
511		----		----	
529		----		----	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017		----		----	
1026	EN12916	16.5	R(0.05)	-3.46	
1033		----		----	
1059	EN12916	22.4		-0.15	
1065		----		----	
1134		----		----	
1146		----		----	
1161	EN12916	21.6		-0.59	
1194		23.7		0.58	
1227	EN12916	22.9		0.14	
1299	EN12916	22.7		0.02	
1389		22.971		0.17	
1397	EN12916	17.58	E, R(0.05)	-2.85	calculation error, iis calculated: 22.36
1406		----		----	
1407		----		----	
1441		----		----	
1459		25.9	C	1.82	first reported: 29.1
1510		----		----	
1544		----		----	
1556	EN12916	23.78		0.63	
1569	EN12916	20.40		-1.27	
1631		----		----	
1634		----		----	
1635	EN12916	20.0		-1.49	
1706		----		----	
1724		22.39		-0.15	
1728		----		----	
1807		----		----	
1811	IP391	20.190		-1.39	
1984		----		----	
2130		----		----	
6016		----		----	
6075		----		----	
	normality	OK			
	n	21			
	outliers	2			
	mean (n)	22.659			
	st.dev. (n)	1.3825			
	R(calc.)	3.871			
	st.dev.(EN12916:16)	1.7826			
	R(EN12916:16)	4.991			





Determination of Pour Point, manual on sample #18090; results in °C

lab	method	value	mark	z(targ)	remarks
62	D97	-27		-0.32	
120	D97	-27		-0.32	
140	D97	-24		0.96	
150		----		----	
171	D97	-30		-1.59	
175		----		----	
194	D97	<-24		----	
230	ISO3016	-27		-0.32	
237	D97	<-21		----	
238	D97	-30		-1.59	
312		----		----	
323	ISO3016	-27		-0.32	
334		----		----	
335		----		----	
336		----		----	
338		----		----	
353	IP15	-27		-0.32	
381		----		----	
444		----		----	
445	IP15	-26		0.11	
447	D97	-27		-0.32	
463		----		----	
494		----		----	
496	D97	-27.0		-0.32	
511		----		----	
529	D97	-24		0.96	
541		----		----	
551		----		----	
603		----		----	
633		----		----	
663	D97	-27		-0.32	
1017		----		----	
1026	D97	-24		0.96	
1033	IP15	<-42		<-6.69	possible false negative test result?
1059	ISO3016	-24		0.96	
1065		----		----	
1134	IP15	-24		0.96	
1146	D97	-26.8		-0.23	
1161	ISO3016	-24		0.96	
1194		----		----	
1227		----		----	
1299	D97	-24		0.96	
1389	D97	<-21		----	
1397		----		----	
1406	ISO3016	-26.0		0.11	
1407	ISO3016	-27		-0.32	
1441		----		----	
1459		----		----	
1510	D97	<-21		----	
1544	ISO3016	-24.0		0.96	
1556	ISO3016	-27		-0.32	
1569		----		----	
1631		----		----	
1634		----		----	
1635	ISO3016	-6	R(0.01)	8.60	
1706		----		----	
1724	D97	-27		-0.32	
1728	D97	-27		-0.32	
1807		----		----	
1811		----		----	
1984	ISO3016	-27		-0.32	
2130		----		----	
6016	D97	-27		-0.32	
6075		----		----	
	normality	OK			
	n	27			
	outliers	1			
	mean (n)	-26.25			
	st.dev. (n)	1.720			
	R(calc.)	4.82			
	st.dev.(ISO3016:94)	2.354			
	R(ISO3016:94)	6.59			





Determination of Pour Point, automated (3°C interval) on sample #18090; results in °C

lab	method	value	mark	z(targ)	remarks
62		----		----	
120	D5950	-24		0.39	
140	D5949	-24		0.39	
150		----		----	
171	D5950	-27		-0.99	
175	D5950	-24		0.39	
194		----		----	
230		----		----	
237		----		----	
238		----		----	
312	D5950	-27		-0.99	
323	D5950	-24		0.39	
334	D5950	-27		-0.99	
335		----		----	
336	D5950	-27		-0.99	
338		----		----	
353		----		----	
381		----		----	
444		----		----	
445		----		----	
447		----		----	
463	D6892	-24		0.39	
494	D6892	-24		0.39	
496	D6892	-24.0		0.39	
511		----		----	
529		----		----	
541	D5950	-27		-0.99	
551		----		----	
603		----		----	
633		----		----	
663		----		----	
1017	D5950	-24		0.39	
1026	D5950	-24		0.39	
1033		----		----	
1059		----		----	
1065	D5950	-24.0		0.39	
1134		----		----	
1146		----		----	
1161	D6749	-24		0.39	
1194		----		----	
1227	D97	-24		0.39	
1299		----		----	
1389		----		----	
1397	D5950	-24		0.39	
1406		----		----	
1407		----		----	
1441	D5950	-24		0.39	
1459	In house	-27.0		-0.99	
1510	D5950	-24		0.39	
1544		----		----	
1556		----		----	
1569	D5950	-24		0.39	
1631		----		----	
1634		----		----	
1635		----		----	
1706	D5950	-25		-0.07	
1724		----		----	
1728		----		----	
1807		----		----	
1811	D5950	-24		0.39	
1984		----		----	
2130	D5950	-24.0		0.39	
6016		----		----	
6075	NFT 60-105	-27		-0.99	
	normality	OK			
	n	26			
	outliers	0			
	mean (n)	-24.85			
	st.dev. (n)	1.347			
	R(calc.)	3.77			
	st.dev.(D5950:14)	2.179			
	R(D5950:14)	6.1			





Determination of Sulphur Content on sample #18090; result in mg/kg

lab	method	value	mark	z(targ)	remarks
62	D5453	6.7		-0.91	
120	D5453	8.59		1.81	
140	D2622	7.45	C	0.17	first reported: 3.45
150	D5453	6.6		-1.06	
171	D4294	7.02	C	-0.45	first reported: 121
175	D5453	7.5		0.24	
194	D2622	6.81		-0.75	
230		----		----	
237	D5453	7.4		0.10	
238		----		----	
312	D5453	7.2		-0.19	
323	ISO20846	7.3		-0.05	
334	ISO20846	6.8		-0.77	
335	ISO20846	7.3		-0.05	
336	ISO20846	7.3		-0.05	
338		----		----	
353	IP490	7.4		0.10	
381	D5453	7.7		0.53	
444	IP490	7.11		-0.32	
445	IP490	8.08		1.08	
447	D5453	8.0		0.96	
463	ISO20846	7.06		-0.39	
494	ISO20846	7.18		-0.22	
496	ISO20846	6.30		-1.49	
511		----		----	
529		----		----	
541	ISO20846	7.53		0.28	
551	D5453	7.8		0.67	
603		----		----	
633		----		----	
663	D5453	8.06		1.05	
1017	ISO20846	6.84		-0.71	
1026	ISO20846	6.4		-1.35	
1033		----		----	
1059	ISO20846	7.4		0.10	
1065	D5453	7.2		-0.19	
1134	IP490	7.84		0.73	
1146		----		----	
1161	ISO20846	6.84		-0.71	
1194		----		----	
1227	D5453	7		-0.48	
1299	ISO20884	7.3		-0.05	
1389	ISO20846	7.6		0.39	
1397	ISO20846	6.2		-1.63	
1406	ISO20846	7.4		0.10	
1407		----		----	
1441	D7039	8.0		0.96	
1459	ISO8754	7.1	C	-0.34	first reported 10.7
1510		----		----	
1544	ISO20846	7.96		0.90	
1556	ISO20884	7.2		-0.19	
1569	ISO20846	7.3		-0.05	
1631	ISO20846	7.7		0.53	
1634	ISO20846	8.195		1.24	
1635	ISO20846	6.5		-1.20	
1706	ISO20846	8.0		0.96	
1724	D5453	7.15		-0.26	
1728	D5453	7.1		-0.34	
1807	ISO20846	6.3		-1.49	
1811	ISO20846	7.8		0.67	
1984	ISO20846	7.18		-0.22	
2130	IP490	8.3		1.39	
6016		----		----	
6075	ISO20846	7.99		0.95	
	normality	OK			
	n	51			
	outliers	0			
	mean (n)	7.33			
	st.dev. (n)	0.551			
	R(calc.)	1.54			
	st.dev.(ISO20846:11)	0.693			
	R(ISO20846:11)	1.94			compare R(D5453:16e1) = 2.58



Determination of Water content, KF on sample #18090; result in mg/kg

lab	method	value	mark	z(targ)	remarks
62	D6304-A	69		-0.40	
120	E1064	105		1.27	
140	D6304-A	139.5	R(0.01)	2.86	
150	D6304-A	152	C,R(0.01)	3.44	first reported: 193
171	D6304-A	104		1.22	
175		----		----	
194		----		----	
230	ISO3733	<500		----	
237	D6304-A	66.53		-0.51	
238		----		----	
312	ISO12937	80		0.11	
323	ISO12937	80		0.11	
334	ISO12937	90		0.58	
335	ISO12937	75.6		-0.09	
336	ISO12937	90		0.58	
338	ISO12937	65.78		-0.54	
353	IP438	87.2		0.45	
381	ISO12937	60		-0.81	
444	IP438	139	R(0.01)	2.84	
445	D6304-A	63		-0.67	
447	IP438	69		-0.40	
463	D6304-A	62.5		-0.70	
494	ISO12937	72.2		-0.25	
496	D6304-A	81		0.16	
511	D6304	89.042		0.53	
529	D6304	76.593		-0.04	
541	ISO12937	74.5		-0.14	
551	D6304-A	76.6		-0.04	
603		----		----	
633	D6304-C	72.6		-0.23	
663	D6304-A	78.8		0.06	
1017	ISO12937	164.2	R(0.01)	4.01	
1026	D6304-A	85		0.34	
1033	IP438	85.0		0.34	
1059	ISO12937	80		0.11	
1065	D6304-C	61		-0.77	
1134	IP438	68		-0.44	
1146	D6304-C	47		-1.41	
1161	ISO12937	65.317		-0.57	
1194	ISO12937	68.96		-0.40	
1227	D6304	73.30		-0.20	
1299	ISO12937	80		0.11	
1389	ISO12937	65		-0.58	
1397	ISO12937	71		-0.30	
1406	ISO12937	62		-0.72	
1407	ISO12937	84.9		0.34	
1441		----		----	
1459	ISO12937	68		-0.44	
1510	IP438	87		0.44	
1544	ISO12937	77.86		0.01	
1556	ISO12937	83		0.25	
1569	In house	76.5		-0.05	
1631	ISO12937	80.0		0.11	
1634	ISO12937	61.3		-0.75	
1635	ISO12937	95		0.81	
1706		----		----	
1724	D6304-A	79		0.07	
1728	E203	90		0.58	
1807	ISO12937	88		0.48	
1811	ISO12937	83.8		0.29	
1984	ISO12937	78.4		0.04	
2130	IP438	90		0.58	
6016	D6304-A	70		-0.35	
6075	ISO12937	117.3		1.84	
	normality	suspect			
	n	53			
	outliers	4			
	mean (n)	77.558			
	st.dev. (n)	12.6167			
	R(calc.)	35.327			
	st.dev.(ISO12937:00)	21.6299			
	R(ISO12937:00)	60.564			




Determination of Total Contamination on sample #18091; result in mg/kg

lab	method	value	mark	z(targ)	Incomplete filtration	Volume used	filtration stopped after
120	D7321:14	10.23	ex	-2.24	NO	850	104.4
140	EN12662:2014	23.9		3.50	NO	300	----
150	EN12662:2014	13.0		-1.08	NO	300	----
171	EN12662:2014	3.0	ex	-5.28	NO	300	----
230	D6217	7.60	ex	-3.35	NO	1000	9
312	EN12662:2014	15.7		0.05	NO	300	----
323	EN12662:2014	11.0		-1.92	----	----	----
334	EN12662:2014	15.5		-0.03	NO	300	----
335		----		----	----	----	----
336		----		----	----	----	----
353	EN12662:2014	23.3		3.24	NO	300	----
445	EN12662:1998	15.1		-0.20	NO	300	Complete
447	IP440	16.8		0.52	YES	----	----
463	EN12662:2014	12.54		-1.27	NO	300	----
494	EN12662:2014	13.25		-0.97	----	----	----
496	EN12662:2014	11.42		-1.74	NO	----	----
663	EN12662:2014	10.29		-2.22	NO	322	12.20
1017		----		----	----	----	----
1026	EN12662:2014	20		1.86	NO	850	----
1033		----		----	----	----	----
1059	EN12662:2014	16.0		0.18	NO	300	----
1064	EN12662:2014	16.71		0.48	NO	309	----
1134	EN12662:2014	12.9		-1.12	----	----	----
1299	EN12662:2014	17.2		0.68	NO	350	----
1397	EN12662:2014	12.1		-1.46	NO	----	----
1544	EN12662:2014	16.85	C	0.54	NO	300	25 C
1724	IP440	13.02		-1.07	NO	----	----
1807	EN12662:2014	13.5		-0.87	NO	250	----
1811	EN12662:2014	22.48		2.90	----	300	----
1984	EN12662:2014	15.45		-0.05	NO	300	----
2130	IP440	6.9	ex	-3.64	NO	800	----
2493	EN12662:2014	50.65	R(0.01)	14.72	NO	----	----
6016		----		----	----	----	----
6075	EN12662:2014	15.7		0.05	NO	----	----
normality		OK					
n		24					
outliers		1 (+4ex)	spike:	recovery:			
mean (n)		15.571	15.0	<104%			
st.dev. (n)		3.7370					
R(calc.)		10.463					
st.dev.(EN12662:14)		2.3825					
R(EN12662:14)		6.671					

Lab 120 and 230 excluded for reporting test results in mg/L

Lab 171 and 2130 excluded for being too low in relation to the known concentration (see paragraph 4.1)

Lab 1544 first reported for Total Contamination: 33.55 and for Filtration stopped after: 5.00



APPENDIX 2

z-scores Distillation

lab	IBP	10% rec	50% rec	90% rec	95% rec	FBP
62	----	----	----	----	----	----
120	-0.61	-0.97	-1.01	0.77	1.21	-0.12
140	0.83	0.25	0.12	0.05	-0.07	-0.55
150	0.34	-0.04	-0.25	-0.62	-0.65	-0.98
171	-0.52	-0.51	-0.07	0.44	0.63	-0.75
175	-0.17	0.25	0.78	1.61	1.53	-0.35
194	0.57	0.19	0.22	0.55	0.28	0.36
230	0.00	-1.26	0.12	-0.45	-0.42	1.50
237	-0.58	-3.30	-2.23	-0.73	-0.58	0.12
238	0.29	1.94	2.48	-0.18	-0.90	-0.27
312	0.69	0.77	1.16	0.10	0.18	-0.59
323	1.32	0.72	0.31	0.55	0.92	-0.04
334	-2.13	-2.37	-1.76	-0.57	-2.31	-1.18
335	----	----	----	----	----	----
336	-0.26	0.19	0.03	-0.18	-0.04	0.44
338	-0.20	-0.62	0.41	0.44	0.09	0.40
353	0.63	-0.33	-0.63	1.89	1.85	0.91
381	-2.27	-0.80	0.31	0.61	0.41	0.08
444	-1.32	-0.16	-1.76	-1.18	-1.29	0.28
445	-1.64	-1.20	-1.10	-0.62	-0.26	-0.75
447	-2.22	-0.39	0.22	0.16	0.22	0.67
463	1.21	0.89	1.25	1.39	0.99	1.26
494	-1.58	-0.51	-0.82	0.10	0.31	0.36
496	-2.36	0.37	-0.82	-0.90	-0.74	-0.15
511	1.01	-1.85	-0.82	-0.73	-0.42	-0.27
529	-0.55	0.42	1.16	0.77	0.50	0.40
541	0.55	0.37	0.50	0.05	0.15	0.16
551	-0.75	-0.74	-0.82	-0.96	-0.94	-0.98
603	0.14	-0.10	1.07	-0.23	-0.46	0.36
633	-0.03	0.54	-0.35	0.05	-0.58	-3.70
663	1.22	1.09	-0.44	-0.73	-0.52	-0.21
1017	1.00	1.04	0.39	0.37	0.37	0.67
1026	-0.23	0.83	0.59	0.44	0.38	0.20
1033	-0.17	-0.10	0.03	0.05	-0.01	0.00
1059	0.34	-0.22	0.69	0.22	0.34	0.48
1065	-0.72	-3.59	-1.10	-0.34	-0.14	-0.71
1134	-0.26	-0.04	-0.54	-0.34	-1.48	-1.54
1146	-0.35	-0.68	0.41	1.00	0.82	0.91
1161	0.34	0.08	0.88	0.61	0.34	-0.23
1194	----	----	----	----	----	----
1227	1.09	1.18	1.16	0.61	0.66	0.04
1299	0.20	-0.33	-0.35	-0.12	-0.01	-0.12
1389	0.37	2.23	1.16	0.49	0.31	0.75
1397	0.83	1.12	0.59	0.33	0.38	0.59
1406	3.59	2.23	2.95	1.22	1.98	0.52
1407	2.47	2.00	1.16	-0.34	-0.55	0.79
1441	-0.46	-0.22	-1.29	-1.01	-0.58	-0.79
1459	0.75	0.19	0.59	-0.96	-0.65	-1.14
1510	0.23	0.02	0.12	-1.24	-0.62	-1.77
1544	-0.27	1.73	1.58	1.33	1.27	0.67
1556	1.47	0.02	0.50	0.49	0.38	0.40
1569	-0.92	0.95	-1.01	-0.23	1.69	----
1631	----	----	----	----	-0.10	----
1634	0.78	1.12	0.31	-0.23	-0.33	0.79
1635	0.29	-1.55	-2.23	-0.73	-1.22	0.91
1706	1.57	0.80	0.74	-0.34	-0.20	-0.69
1724	0.00	0.25	-0.54	-0.23	0.18	-0.04
1728	-0.58	-0.97	-2.23	0.10	0.06	-0.08
1807	0.14	-0.80	-1.01	-0.62	-0.68	0.48
1811	0.23	-0.04	-0.07	-0.62	-0.81	-0.23
1984	0.98	0.89	0.12	0.33	0.31	0.44
2130	0.86	0.54	0.41	-0.23	-0.14	0.04
6016	-1.64	-1.03	-1.39	-0.68	-0.65	-1.57
6075	0.06	-0.51	0.12	-0.73	-0.39	0.12

APPENDIX 3**Number of participants per country**

1 lab in ARGENTINA
1 lab in AUSTRIA
2 labs in BELGIUM
1 lab in BRAZIL
1 lab in BULGARIA
1 lab in CANADA
1 lab in CHINA, People's Republic
2 labs in CROATIA
1 lab in CYPRUS
1 lab in CZECH REPUBLIC
6 labs in FRANCE
3 labs in GERMANY
3 labs in GREECE
1 lab in HUNGARY
1 lab in IRELAND
1 lab in ISRAEL
1 lab in KAZAKHSTAN
1 lab in MALAYSIA
1 lab in MARTINIQUE
1 lab in MAURITIUS
1 lab in MEXICO
3 labs in NETHERLANDS
2 labs in NIGERIA
1 lab in PERU
1 lab in PHILIPPINES
1 lab in PORTUGAL
1 lab in ROMANIA
1 lab in SERBIA
1 lab in SLOVENIA
4 labs in SPAIN
2 labs in SWEDEN
1 lab in THAILAND
3 labs in TURKEY
7 labs in UNITED KINGDOM
6 labs in UNITED STATES OF AMERICA

APPENDIX 4

Abbreviations

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= probably an error in calculations
U	= test result probably reported in a different unit
W	= test result withdrawn on request of participant
ex	= test result excluded from the statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
SDS	= Safety Data Sheet

Literature

1. iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, March 2017
2. Horwitz, R. Albert, J. AOAC Int, 79, 3, 589, (1996)
3. ASTM E178:08
4. ASTM E1301:03
5. ISO 5725:86
6. ISO 5725, parts 1-6:94
7. ISO13528:05
8. M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
9. W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
10. DIN 38402 T41/42
11. P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
12. J.N. Miller, Analyst, 118, 455, (1993)
13. Analytical Methods Committee Technical Brief, No. 4 January 2001
14. P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst 2002, 127, 1359-1364 (2002)
15. EN15293:11
16. Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, *Technometrics*, 25(2), 165-172, (1983)