

**Results of Proficiency Test
Total Metals in Polymers
September 2020**

Organized by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: ing. M. Meijer
Correctors: ing. A.S. Noordman-de Neef & ing. C.M. Nijssen-Wester
Report: iis20P06

November 2020

CONTENTS

1	INTRODUCTION	3
2	SET UP	3
2.1	ACCREDITATION	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALITY STATEMENT.....	4
2.4	SAMPLES.....	4
2.5	ANALYZES	5
3	RESULTS	6
3.1	STATISTICS	6
3.2	GRAPHICS	7
3.3	Z-SCORES.....	7
4	EVALUATION	8
4.1	EVALUATION PER SAMPLE AND PER ELEMENT	9
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	9
4.3	COMPARISON OF THE PROFICIENCY TEST OF SEPTEMBER 2020 WITH PREVIOUS PTS.....	10
4.4	EVALUATION OF THE ANALYTICAL DETAILS.....	10
5	DISCUSSION.....	11
6	CONCLUSION	11

Appendices:

1.	Data, statistical and graphic results	12
2.	Other reported test results	20
3.	Analytical details	32
4.	Number of participants per country.....	39
5.	Abbreviations and literature	40

1 INTRODUCTION

World-wide many consumer products with plastic parts are produced and transported. These plastic parts are produced under strict regulations. For instance, in the European Directive 2011/65/EC maximum concentrations are specified for metals in plastic: the content of Lead (Pb), Mercury (Hg), and Hexavalent Chromium (CrVI) may not exceed 0.1%M/M, while the maximum concentration for Cadmium (Cd) may not exceed 0.01%M/M.

Since 1998 a proficiency scheme for the determination of Metals in plastics is organized by the Institute for Interlaboratory Studies every year. Starting with only total Cadmium, over the years the scope was extended with other elements.

In this interlaboratory study 171 laboratories in 39 different countries registered for participation. See appendix 4 for the number of participants per country. In this report the results of this proficiency test are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity were subcontracted to an ISO/IEC17025 accredited laboratory. It was decided to send 2 different polymer samples labelled #20650 and #20651 of approximately 6 grams each.

The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

For the first sample a batch of dark grey PVC rings, artificially fortified to be positive on one or more metals, was selected. After homogenization 200 small bags were filled with approximately 6 grams of the polymer material and labelled #20650. The homogeneity of the subsamples was checked by determination of the totals of Copper and Nickel according to an in house method on 9 stratified randomly selected subsamples.

	Copper in mg/kg	Nickel in mg/kg
sample #20650-1	292.2	198.9
sample #20650-2	295.7	195.4
sample #20650-3	299.7	198.7
sample #20650-4	298.7	208.2
sample #20650-5	295.5	200.7
sample #20650-6	295.4	205.2
sample #20650-7	292.6	202.2
sample #20650-8	284.0	197.9
sample #20650-9	287.2	193.9

Table 1: homogeneity test results of subsamples #20650

From the above test results the repeatabilities were calculated and compared with 0.3 times the corresponding reproducibility estimated from the Horwitz equation in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Copper in mg/kg	Nickel in mg/kg
r (observed)	14.4	12.7
reference method	Horwitz	Horwitz
0.3 * R (reference method)	16.8	12.1

Table 2: evaluation of the repeatabilities of subsamples #20650

The calculated repeatabilities were in agreement with 0.3 times the corresponding target reproducibility estimated from the Horwitz equation. Therefore, homogeneity of the subsamples was assumed.

For the second sample a batch of beige Polypropylene bits, artificially fortified to be positive on one or more metals, was selected. After homogenization 200 small bags were filled with approximately 6 grams of the polymer material and labelled #20651. The homogeneity of the subsamples was checked by determination of total Cadmium using test method EN1122:01 on 4 stratified randomly selected subsamples.

	Cadmium in mg/kg
sample #20651-1	85
sample #20651-2	86
sample #20651-3	85
sample #20651-4	84

Table 3: homogeneity test results of subsamples #20651

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Cadmium in mg/kg
r (observed)	2.3
reference test method	EN1122:01
0.3 * R (reference test method)	8.5

Table 4: evaluation of the repeatability of subsamples #20651

The calculated repeatability was lower than 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

To each of the participating laboratories one sample labelled #20650 and one sample labelled #20651 were sent on August 12, 2020.

2.5 ANALYZES

The participants were requested to determine on both samples the total of: Antimony, Cadmium, Chromium, Hexavalent Chromium, Cobalt, Copper, Lead, Mercury, Nickel, Manganese and Tin.

It was also requested to report if the laboratory was accredited for the requested elements that were determined and to report some analytical details.

It was explicitly requested to treat the samples as if they were routine samples and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis-cts/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis-cts/. The reported test results are tabulated per determination in appendices 1 and 2 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no re-analysis). Additional or corrected test results are used for the data analysis and the original test results are placed under 'Remarks' in the test result tables in appendices 1 or 2. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' (iis-protocol, version 3.5) of June 2018. For statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a dataset does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

According to ISO5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirements based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1 was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis.

The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve was projected over the Kernel Density Graph for reference.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

This target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The $z_{(\text{target})}$ scores are listed in the test result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. The usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this interlaboratory study some problems were encountered with the dispatch of the samples due to the COVID-19 pandemic. Therefore, the reporting time on the data entry portal was extended with one week. After this period still nine participants did not report any test results. Not all participants were able to report all elements requested.

Finally, 162 reporting laboratories submitted 519 numerical test results. Observed were 8 outlying test results, which is 1.5%. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as "not OK" or "suspect". The statistical evaluation of these data sets should be used with due care.

4.1 EVALUATION PER SAMPLE AND PER ELEMENT

In this section the results are discussed per sample and per element. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the reported test results in appendix 1. The abbreviations, used in these tables, are explained in appendix 5.

Test method EN1122 is considered to be the official test method for the determination of Cadmium in polymers. The precision data mentioned in EN1122 were used to evaluate the Cadmium test results. Unfortunately, a suitable reference test method, providing the precision data, is not available for all other elements. For tests that have no available precision data, the calculated reproducibility was compared against the reproducibility estimated from the Horwitz equation.

Sample #20650

Total Copper as Cu: This determination may be problematic. Five statistical outliers were observed. The calculated reproducibility after rejection of statistical outliers is not in agreement with the estimated reproducibility using the Horwitz equation.

Total Nickel as Ni: This determination may be problematic. One statistical outlier was observed. The calculated reproducibility after rejection of statistical outlier is not in agreement with the estimated reproducibility using the Horwitz equation.

Other metals: The majority of participants agreed on a concentration near or below the limit of detection for the other elements. The test results are given in appendix 2.

Sample #20651

Total Cadmium as Cd: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of EN1122:01.

Total Lead as Pb: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of statistical outlier is in full agreement with the estimated reproducibility using the Horwitz equation.

Other metals: The majority of participants agreed on a concentration near or below the limit of detection for the other elements. The test results are given in appendix 2.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the reference test method or as declared by the estimated target reproducibility using the Horwitz equation and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility ($2.8 * \text{standard deviation}$) and the target reproducibility derived from literature reference test methods (in casu EN test method) or estimated using the Horwitz equation are presented in the next table.

Element	unit	n	average	$2.8 * \text{sd}$	R(target)
Total Copper as Cu	mg/kg	96	273.5	61.6	52.7
Total Nickel as Ni	mg/kg	105	187.6	43.5	38.2

Table 5: performance overview on sample #20650

Element	unit	n	average	$2.8 * \text{sd}$	R(target)
Total Cadmium as Cd	mg/kg	156	94.5	18.6	23.6
Total Lead as Pb	mg/kg	154	95.1	20.4	21.5

Table 6: performance overview on sample #20651

Without further statistical calculations, it can be concluded that there is a good compliance of the group of participating laboratories with the reference test method/target for the elements Cadmium and Lead but not a good compliance for Copper and Nickel. See also the discussion in paragraphs 4.1 and 5.

4.3 COMPARISON OF THE PROFICIENCY TEST OF SEPTEMBER 2020 WITH PREVIOUS PTs

	September 2020	August 2019	September 2018	September 2017	September 2016
Number of reporting laboratories	162	177	166	167	142
Number of test results	519	1253	1471	960	1041
Number of statistical outliers	8	42	46	24	24
Percentage of statistical outliers	1.5%	3.4%	3.1%	2.5%	2.3%

Table 7: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared, expressed as relative standard deviation (RSD) of the PTs, see next table.

Element	September 2020	August 2019	September 2018	September 2017	September 2016	2015 - 2004	Target
Sb	--	--	11-14%	10%	--	9%	7-9%
Cd	7%	7-10%	8-12%	8%	9%	7-12%	9%
Cr	--	8-11%	7%	--	10%	7-24%	7-10%
Cr6+	--	9-22%	6%	--	52%	23-76%	7-9%
Co	--	--	--	8%	8%	10-11%	7-10%
Cu	8%	--	6%	--	8%	7%	7-10%
Pb	8%	7%	7%	6%	10%	7-11%	6-9%
Hg	--	--	9-20%	--	8-14%	13-46%	7-13%
Ni	8%	8%	--	10%	9%	--	7-10%
Mn	--	10%	--	--	--	--	7-10%
Sn	--	13%	--	--	--	--	7-10%

Table 8: development of the uncertainties over the years

The RSDs observed in this PT are in line with RSDs observed in previous iis PTs.

Sample #20651 was used before in Proficiency Test iis06P03 as sample #0652. It is observed that the current PT findings of the polymer sample containing Cadmium and Lead show a very similar average concentration level and calculated reproducibility, see next table.

Element	unit	#20651			#0652		
		n	average	R(calc)	n	average	R(calc)
Total Cadmium as Cd	mg/kg	156	94.5	18.6	76	95.7	18.8
Total Lead as Pb	mg/kg	154	95.1	20.4	71	95.9	24.5

Table 9: comparison of sample #20651 with sample #0652

4.4 EVALUATION OF THE ANALYTICAL DETAILS

The participants were asked to provide several analytical details which are listed in appendix 3. Based on the reported answers the following can be summarized:

- A vast majority (90%) mentioned that they are ISO/IEC17025 accredited to determine the reported elements.
- About 72% further cut the samples prior to analysis, 17% used the samples as received and 11% further grinded the samples.
- Almost 90% used 0.5 grams or less of sample intake, mostly 0.1 to 0.2 grams.
- To release the metals from the sample about 80% used Microwave, 13% used Acid Digestion and 7% Other e.g. by XRF which does not actually release the metals from the sample.
- In case of Acid Digestion, Nitric Acid or a mixture of Nitric/Hydrochloric Acid was used.
- The technique to quantify the metals was most often ICP-OES (53%) or ICP-MS (28%) followed by other techniques like AAS, XRF, ICP-AES and UV-VIS.

As the majority of the group follow the same analytical procedures no separate statistical analysis has been performed.

5 DISCUSSION

In this proficiency test for the determination of total metals in polymers two different type of polymers were used: PVC (Polyvinylchloride) and Polypropylene. In the PVC sample two different metals were positively identified by the participants while two other metals were found in the Polypropylene sample. Based on the observed vs. target reproducibilities the metals in the Polypropylene sample could be less difficult to determine than in the PVC sample, but this could also be due to the different type of metals that were found in both samples. See also the discussion in §4.1.

6 CONCLUSION

Each participating laboratory will have to evaluate its performance in this study and decide about any corrective actions if necessary. Therefore, participation on a regular basis in this scheme could be helpful to improve the performance and thus increase of the quality of the analytical results.

APPENDIX 1

Determination of total Copper as Cu on sample #20650; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
110		----		----	2375	EN16711-1	286		0.67
210	CPSC-CH-E1002-08	298.39		1.33	2378	EN16711-1	276.0		0.13
230		----		----	2379		----		----
339	In house	267.8		-0.30	2380	CPSC-CH-E1002-08	286.00		0.67
348		----		----	2381		Out Cap		----
362		----		----	2382	EN16711-1	275.3		0.10
523	EPA3052	280.94		0.40	2384		----		----
551		----		----	2385	In house	239		-1.83
623	In house	242.12		-1.67	2387		----		----
826	EPA3052	271.7		-0.09	2390	CPSC-CH-E1002-08	282.396		0.47
840	EN16711-1	266.560		-0.37	2391	IEC62321-5	270.0		-0.18
841	ISO8124-5	269		-0.24	2392		----		----
1051		----		----	2406		----		----
1082	In house	240		-1.78	2409	In house	220.75	C	-2.80
1126		----		----	2410	IEC62321-5	281		0.40
1527	In house	245.6		-1.48	2413	In house	136.2	R(0.01)	-7.30
1842	not analysed			----	2426	EN16711-1	265.02		-0.45
1910	not analysed			----	2431	In house	288.30		0.79
2108	EN16711-1	239.6		-1.80	2444		----		----
2115	EN16711-1	276.59		0.17	2453		----		----
2118	CPSC-CH-E1002-08	238.541		-1.86	2460		----		----
2120	EN16711-1	211		-3.32	2475	In house	not analysed		----
2129	EN16711-1	264		-0.50	2477		----		----
2131	In house	286.316		0.68	2488		----		----
2132		----		----	2492	In house	not detected		----
2137	KS M6959	275.032		0.08	2500		----		----
2138	CPSC-CH-E1002-08	271.9		-0.08	2504	IEC62321-5	240.076		-1.78
2139	IEC62321-5	273.7		0.01	2511		----		----
2165	not applicable			----	2514	In house	286.0658		0.67
2176	IEC62321-5	not analysed		----	2529		----		----
2182		----		----	2557		----		----
2184	IEC62321-5	not applicable		----	2560	EN16711-1	296.35		1.22
2201	CPSC-CH-E1002-08	279.95		0.34	2561	CPSC-CH-E1002-08	248.11		-1.35
2202	EPA3052	271		-0.13	2563		----		----
2216		----		----	2564		----		----
2218		----		----	2568		----		----
2229	ASTM F963			----	2569	CPSC-CH-E1002-08	284		0.56
2230	EPA3051	239		-1.83	2572	IEC62321-5	283		0.51
2232	IEC62321-5	not applicable		----	2573	EPA3052	227.479		-2.45
2236		----		----	2590	EN16711-1	231.3		-2.24
2250	EN16711-1	293.86		1.08	2591		----		----
2255	EN16711-1	297.2		1.26	2603	GB31604.49/5009.268	285.0510		0.62
2256	EPA3051	268.778		-0.25	2608		----		----
2258	not analysed			----	2621	In house	278.9		0.29
2264		----		----	2624	In house	301.9		1.51
2265		----		----	2630		----		----
2266	In house	284		0.56	2643	EN16711-1	269.67		-0.20
2271	EPA3052	284.6		0.59	2644	EN16711-1	307.2		1.79
2284	EN16711-1	276.42		0.16	2649	CPSC-CH-E1002-08	320.8		2.52
2290	IEC62321-5	279		0.29	2672	IEC62321-5	261.052		-0.66
2293	CPSC-CH-E1002-08	256.09		-0.92	2674		----		----
2294	ISO8124-5	not analysed		----	2678		----		----
2297	IEC62321-5	259.2		-0.76	2734	CPSC-CH-E1002-08	313.45		2.13
2301	EN16711-1	236.9		-1.94	2736		----		----
2310	EN16711-1	304		1.62	2794	IEC62321-3-1	430	R(0.01)	8.32
2311	CPSC-CH-E1002-08	302.78		1.56	2804	EN16711-1	304.199		1.63
2314		----		----	2826		----		----
2316	In house	293.1		1.04	2829		----		----
2320		269		-0.24	2833	IEC62321-3-1	not analysed		----
2339	In house	235		-2.05	2835	IEC62321-5	281.45		0.42
2347	EPA3052	279		0.29	2853		----		----
2350	EPA3052	282.4		0.47	2864	IEC62321-5	253.91		-1.04
2352	In house	275.2		0.09	2870	EN16711-1	308.7		1.87
2353	IEC62321-5	NA		----	2885		----		----
2355	EPA3052	289.8		0.87	2900	In house	not analysed		----
2357	ISO8124-5	282.4		0.47	3100	CPSC-CH-E1002-08	283.532		0.54
2358	EPA3051	N/A		----	3110		----		----
2363	EPA3052	280		0.35	3111		----		----
2365	EN16711-1	268.2		-0.28	3116		----		----
2366	IEC62321-5	268.14		-0.28	3118	EN16711-1	262.62		-0.58
2369	EPA3052	291.2		0.94	3121	In house	259		-0.77
2370	EPA3052	271		-0.13	3122	CPSC-CH-E1002-08	495	R(0.01)	11.78
2372	EPA3052	307.3		1.80	3124	EPA3052	354	R(0.05)	4.28
2374	EN16711-1	270.02		-0.18	3146	In house	278		0.24



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
3153	IEC62321-5	281.1		0.41	3214	EPA3052	280.80		0.39
3154	EN16711-1	297.76		1.29	3216	In house	271.731		-0.09
3160	CPSC-CH-E1002-08	230.66		-2.28	3225		not analysed		-----
3163	In house	287		0.72	3228		-----		-----
3166		-----		-----	3237	EN16711-1	236.98		-1.94
3172		-----		-----	3239		-----		-----
3182	IEC62321-5	not analysed		-----	3246	EN16711-1	282.63		0.49
3183		-----		-----	3248	CPSC-CH-E1002-08	283		0.51
3185	CPSC-CH-E1002-08	279.7		0.33	6146	IEC62321-3-1	356.62	R(0.05)	4.42
3190	IEC62321-5	283		0.51	8005		-----		-----
3197	EPA3052	299.0		1.36	8008		-----		-----
3210	In house	291		0.93					

normality
 n
 outliers
 mean (n)
 st.dev. (n)
 R(calc.)
 st.dev.(Horwitz)
 R(Horwitz)

OK
96
5
273.4716
22.01543
61.6432
18.80357
52.6500

RSD=8%

Lab 2409 first reported 171.800



Determination of total Nickel as Ni on sample #20650; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
110		----		----	2375	EN16711-1	183		-0.33
210	CPSC-CH-E1002-08	209.28	1.59		2378	EN16711-1	189.4		0.14
230		----		----	2379		----		----
339	In house	186.9	-0.05		2380	CPSC-CH-E1002-08	194.340		0.50
348		----		----	2381	Out Cap			----
362		----		----	2382	EN16711-1	180.2		-0.54
523	IEC62321-5	202.72	1.11		2384		----		----
551		----		----	2385	In house	167		-1.51
623	In house	163.79	-1.74		2387		----		----
826	In house	186.1	-0.11		2390	CPSC-CH-E1002-08	153.789		-2.47
840	EN16711-1	185.459	-0.15		2391	IEC62321-5	182.0		-0.41
841	EN16711-1	160	-2.02		2392		----		----
1051		----		----	2406	CPSC-CH-E1002-08	186.9		-0.05
1082	In house	175	-0.92		2409	In house	174.3	C	-0.97
1126		----		----	2410	IEC62321-5	180		-0.55
1527	In house	179.7	-0.58		2413		----		----
1842	not analysed	----		----	2426	EN16711-1	185.81		-0.13
1910	not analysed	----		----	2431	In house	209.47		1.61
2108	EN16711-1	186.0	-0.11		2444		----		----
2115	EN16711-1	206.93	1.42		2453		----		----
2118	CPSC-CH-E1002-08	184.43	-0.23		2460		----		----
2120	EN16711-1	145	-3.12		2475	In house	not analysed		----
2129	EN16711-1	196	0.62		2477		----		----
2131	In house	208.118	1.51		2488		----		----
2132		----		----	2492	In house	160.0		-2.02
2137	KS M6956	183.029	-0.33		2500		----		----
2138	CPSC-CH-E1002-08	168.8	-1.37		2504	IEC62321-5	165.189		-1.64
2139	IEC62321-5	199.5	0.88		2511		195.8		0.60
2165	not applicable	----		----	2514	In house	199.0417		0.84
2176	IEC62321-5	not analysed	----	----	2529		----		----
2182		----		----	2557		----		----
2184	IEC62321-5	not applicable	----	----	2560	EN16711-1	196.78		0.68
2201	CPSC-CH-E1002-08	186.25	-0.10		2561	CPSC-CH-E1002-08	170.00		-1.29
2202	In house	184	-0.26		2563		----		----
2216		----		----	2564		----		----
2218		----		----	2568		----		----
2229		----		----	2569	CPSC-CH-E1002-08	192		0.33
2230	CPSC-CH-E1002-08	175	-0.92		2572	IEC62321-5	187		-0.04
2232	IEC62321-5	not applicable	----	----	2573	EPA3052	200.178		0.92
2236		----		----	2590	EN16711-1	166.1		-1.57
2250	EN16711-1	204.61	1.25		2591		----		----
2255	EN16711-1	199.0	0.84		2603	GB31604.49/5009.268	208.7920		1.56
2256	CPSC-CH-E1002-08	193.638	0.45		2608		----		----
2258	not analysed	----		----	2621	In house	191.2		0.27
2264	CPSC-CH-E1002-08	207.0	1.42		2624	In house	212.3		1.81
2265		----		----	2630		----		----
2266	In house	205	1.28		2643	EN16711-1	160.25		-2.00
2271	EN16711-1	192.5	0.36		2644	EN16711-1	209.1		1.58
2284	EN16711-1	196.11	0.63		2649	CPSC-CH-E1002-08	197.0		0.69
2290	IEC62321-5	192	0.33		2672	IEC62321-5	189.155		0.12
2293	CPSC-CH-E1002-08	183.79	-0.28		2674		----		----
2294	CPSC-CH-E1002-08	not analysed	----	----	2678		----		----
2297	IEC62321-5	188.3	0.05		2734	CPSC-CH-E1002-08	182.79		-0.35
2301	EN16711-1	164.85	-1.66		2736		----		----
2310	EN16711-1	204	1.20		2794	IEC62321-3-1	218		2.23
2311	CPSC-CH-E1002-08	198.626	0.81		2804	EN16711-1	162.624		-1.83
2314		----		----	2826		----		----
2316	In house	220.6	2.42		2829		----		----
2320		173	-1.07		2833	IEC62321-3-1	not analysed		----
2339	In house	145	-3.12		2835	IEC62321-5	181.00		-0.48
2347		178	-0.70		2853		----		----
2350	EPA3052	176.4	-0.82		2864	IEC62321-5	163.00		-1.80
2352	In house	191.1	0.26		2870	EN16711-1	198.99		0.84
2353	IEC62321-5	189.3	0.13		2885		----		----
2355	EPA3052	191.6	0.30		2900	In house	not analysed		----
2357	IEC62321-5	179.8	-0.57		3100	CPSC-CH-E1002-08	189.753		0.16
2358	EPA3051	188.7	0.08		3110		----		----
2363	In house	181	-0.48		3111		----		----
2365	EN16711-1	192.2	0.34		3116		----		----
2366	IEC62321-5	189.30	0.13		3118	EN16711-1	186.56		-0.07
2369	In house	183.2	-0.32		3121	In house	189		0.11
2370	EPA3052	185	-0.19		3122	CPSC-CH-E1002-08	180		-0.55
2372	IEC62321-5	182.0	-0.41		3124	In house	258	R(0.01)	5.16
2374	EN16711-1	172.23	-1.12		3146	In house	205		1.28

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
3153	IEC62321-5	194.4		0.50	3214	EN16711-1	194.15		0.48
3154	EN16711-1	221.22		2.47	3216	In house	189.385		0.13
3160	CPSC-CH-E1002-08	164.35		-1.70	3225		not analysed		-----
3163	In house	216		2.08	3228		-----		-----
3166		-----		-----	3237	EN16711-1	183.92		-0.27
3172	CPSC-CH-E1002-08	193.54		0.44	3239		-----		-----
3182	IEC62321-5	not analysed		-----	3246	EN16711-1	201.58		1.03
3183		-----		-----	3248	CPSC-CH-E1002-08	180		-0.55
3185	CPSC-CH-E1002-08	191.1		0.26	6146	IEC62321-3-1	203.29		1.15
3190	IEC62321-5	<2	f-	<-13.59	8005		-----		-----
3197	IEC62321-5	192.6		0.37	8008		-----		-----
3210	In house	205		1.28					

normality

OK

n

105

outliers

1

mean (n)

187.5544

st.dev. (n)

15.53094

RSD=8%

R(calc.)

43.4866



st.dev.(Horwitz)

13.64921

R(Horwitz)

38.2178

Lab 2409 first reported 106.750
 Lab 3190 possibly a false negative test result?





Determination of Cadmium as Cd on sample #20651; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
110		----		----	2375	EN16711-1	85		-1.12
210	CPSC-CH-E1002-08	89.77		-0.56	2378	EN1122	97.4		0.35
230		----		----	2379	IEC62321-5	97.36		0.34
339	In house	96.62		0.25	2380	CPSC-CH-E1002-08	96.261		0.21
348	In house	89.602		-0.58	2381	EN1122	96.10		0.19
362	EN1122	88.59		-0.70	2382	EN16711-1	94.3		-0.02
523	IEC62321-5	98.91		0.53	2384	IEC62321-5	93.20		-0.15
551		----		----	2385	In house	100		0.65
623	EN1122	67.37	R(0.05)	-3.21	2387	IEC62321-5	105.927		1.36
826	IEC62321-5	98.15		0.44	2390	CPSC-CH-E1002-08	74.901		-2.32
840	EN16711-1	88.564		-0.70	2391	IEC62321-5	86.0		-1.01
841	ISO8124-5	85		-1.12	2392	IEC62321-5	96.15		0.20
1051	EN1122	96.02		0.18	2406	CPSC-CH-E1002-08	85.46		-1.07
1082	In house	85.5		-1.06	2409	EPA3052	105.1	C	1.26
1126	In house	100.5		0.71	2410	EN1122	98		0.42
1527	In house	86.65		-0.93	2413	In house	80.8		-1.62
1842	IEC62321-3-1	92		-0.29	2426	EN16711-1	96.10		0.19
1910	In house	92.962		-0.18	2431	In house	95.79		0.16
2108	EN16711-1	97.80		0.39	2444	IEC62321-5	92.45		-0.24
2115	EN16711-1	87.04		-0.88	2453	EN1122	93.4		-0.13
2118	CPSC-CH-E1002-08	88.832		-0.67	2460	EN1122	87.7		-0.80
2120	EN1122	110		1.84	2475	In house	93.06		-0.17
2129	EN16711-1	96.6		0.25	2477		----		----
2131	In house	95.419		0.11	2488	IEC-5	87.3		-0.85
2132	EN1122	90.82		-0.43	2492	In house	99.3		0.57
2137	KS M6956	90.662		-0.45	2500	EN16711-1	97.83		0.40
2138	CPSC-CH-E1002-08	100.4		0.70	2504	EN1122	99.842		0.64
2139	IEC62321-5	112.1		2.09	2511		92.4		-0.25
2165	IEC62321-5	99.3		0.57	2514	In house	91.3541		-0.37
2176	IEC62321-5	103.0		1.01	2529		----		----
2182	EN1122	98.51		0.48	2557	IEC62321-5	97.9301		0.41
2184	EN1122	95.2		0.09	2560	EN16711-1	93.23		-0.15
2201	CPSC-CH-E1002-08	99.52		0.60	2561	CPSC-CH-E1002-08	93.85		-0.07
2202	IEC62321-5	110		1.84	2563		----		----
2216	CPSC-CH-E1002-08	97.411		0.35	2564	CPSC-CH-E1002-08	98.992		0.53
2218	CPSC-CH-E1002-08	94.494		0.00	2568		----		----
2229	ASTM F963	97.1464		0.32	2569	CPSC-CH-E1002-08	92		-0.29
2230	EN1122	101		0.77	2572	EN1122	95		0.06
2232	IEC62321-5	96.496		0.24	2573	EN1122	93.741		-0.09
2236	In house	90.48		-0.47	2590	EN1122	108.0		1.60
2250	EN16711-1	97.24		0.33	2591	CPSC-CH-E1002-08	91.324		-0.37
2255	EN16711-1	99.8		0.63	2603	GB31604.49/5009.268	99.2450		0.56
2256	CPSC-CH-E1002-08	91.313		-0.38	2608		----		----
2258	EN16711-1	100.97		0.77	2621	IEC62321-5	93.2		-0.15
2264	CPSC-CH-E1002-08	88.038		-0.76	2624	EPA3052	96.5		0.24
2265	EN16711-1	98.1		0.43	2630	IEC62321-5	90.8		-0.44
2266	In house	91		-0.41	2643	EN16711-1	87.05		-0.88
2271	EN1122	93.2		-0.15	2644	EN16711-1	104.6		1.20
2284	EN1122	100.23		0.68	2649	EN1122	93.4		-0.13
2290	EN1122	98		0.42	2672	IEC62321-5	85.104		-1.11
2293	CPSC-CH-E1002-08	84.85		-1.14	2674	EN1122	99.5398		0.60
2294	ISO8124-5	not analysed		----	2678		----		----
2297	EN1122	95.7		0.14	2734	CPSC-CH-E1002-08	91.92		-0.30
2301	EN1122	97.3		0.33	2736		----		----
2310	EN1122	87.7		-0.80	2794	IEC62321-3-1	85		-1.12
2311	CPSC-CH-E1002-08	87.69		-0.80	2804	EN16711-1	90.5444		-0.47
2314	CPSC-CH-E1002-08	88.009		-0.77	2826	IEC62321-5	89.6414		-0.57
2316	IEC62321-5	89.6		-0.58	2829	EN1122	92.16		-0.28
2320		106		1.37	2833	IEC62321-3-1	76		-2.19
2339	In house	72		-2.66	2835	IEC62321-5	96.35		0.22
2347	EN1122	99		0.54	2853		----		----
2350	EN1122	91.44		-0.36	2864	IEC62321-5	96.19		0.20
2352	EN1122	97.3		0.33	2870	EN16711-1	109.11		1.73
2353	IEC62321-5	91.11		-0.40	2885	IEC62321-5	92.96		-0.18
2355	EN1122	99.1		0.55	2900	IEC62321-5	103.2		1.03
2357	IEC62321-5	95.93		0.17	3100	EN1122	98.2575		0.45
2358	EPA3051	93.5		-0.12	3110	ASTM F963	93.65		-0.10
2363	EPA3052	96		0.18	3111	EPA3052	101.3		0.81
2365	EN16711-1	98.9		0.52	3116	CPSC-CH-E1002-08	94.4		-0.01
2366	IEC62321-5	98.733		0.50	3118	EN1122	96.25		0.21
2369	IEC62321-5	96.8		0.28	3121		----		----
2370	EN1122	91.8		-0.32	3122	CPSC-CH-E1002-08	96		0.18
2372	EPA3052	88.7		-0.69	3124	EPA3052	109		1.72
2374	EN1122	93.42		-0.13	3146	In house	98.9		0.52

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
3153	EN1122	95.6		0.13	3214	EN1122	99.80		0.63
3154	EN1122	85.27		-1.09	3216	In house	91.153		-0.39
3160	CPSC-CH-E1002-08	85.47		-1.07	3225	EN1122	95.05		0.07
3163		85		-1.12	3228	IEC62321-5	97.0		0.30
3166		-----		-----	3237	EN16711-1	91.03		-0.41
3172	CPSC-CH-E1002-08	91.03		-0.41	3239	IEC62321-5	102.494		0.95
3182	EN1122	101.600		0.84	3246	EN16711-1	98.60		0.49
3183	In house	104.5965		1.20	3248	EN1122	83		-1.36
3185	EN1122	96.9		0.29	6146	IEC62321-3-1	77.57		-2.00
3190	IEC62321-5	94		-0.06	8005	EN1122	93.8		-0.08
3197	EN1122	97.5		0.36	8008	EN1122	93.56		-0.11
3210	In house	96.5		0.24					
	normality	suspect							
n		156							
outliers		1							
mean (n)		94.4801							
st.dev. (n)		6.62720							
R(calc.)		18.5562							
st.dev.(EN1122:01)		8.43572							
R(EN1122:01)		23.6200							

RSD=7%

Lab 2409 first reported 53.585



Determination of total Lead as Pb on sample #20651; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
110		----		----	2375	EN16711-1	87		-1.06
210	CPSC-CH-E1002-08	95.09		-0.01	2378	EN16711-1	94.9		-0.03
230		----		----	2379	IEC62321-5	99.86		0.61
339		----		----	2380	CPSC-CH-E1002-08	94.034		-0.14
348	CPSC-CH-E1002-08.3	91.254		-0.51	2381	CPSC-CH-E1002-08.1	97.60		0.32
362		----		----	2382	EN16711-1	95.4		0.03
523	IEC62321-5	96.057		0.12	2384	IEC62321-5	92.11		-0.40
551		----		----	2385	In house	101		0.76
623	In house	70.36		-3.23	2387	IEC62321-5	113.335		2.37
826	IEC62321-5	96.43		0.17	2390	CPSC-CH-E1002-08	76.513		-2.43
840	EN16711-1	91.047		-0.53	2391	IEC62321-5	90.6		-0.59
841	ISO8124-5	87		-1.06	2392	IEC62321-5	104.58		1.23
1051	CPSC-CH-E1002-08.3	95.25		0.01	2406	CPSC-CH-E1002-08.3	89.72		-0.71
1082	In house	87.5		-1.00	2409	EPA3052	74.245		-2.73
1126	In house	99.5		0.57	2410	IEC62321-5	98		0.37
1527	In house	115.5		2.65	2413	In house	91.2		-0.51
1842	IEC62321-3-1	88		-0.93	2426	EN16711-1	100.74		0.73
1910	In house	87.959		-0.94	2431	In house	95.99		0.11
2108	CPSC-CH-E1002-08	93.33		-0.24	2444	IEC62321-5	93.56		-0.21
2115	EN16711-1	88.89		-0.82	2453		----		----
2118	CPSC-CH-E1002-08	94.657		-0.06	2460	CPSC-CH-E1002-08.3	91.95		-0.42
2120	CPSC-CH-E1002-08.3	92		-0.41	2475	In house	96.4		0.16
2129	EN16711-1	90		-0.67	2477	CPSC-CH-E1002-08.3	88.99		-0.80
2131	In house	93.8825		-0.16	2488	IEC-5	76.7		-2.41
2132	CPSC-CH-E1002-08	90.46		-0.61	2492	In house	91.5		-0.48
2137	KS M6956	91.527		-0.47	2500	EN16711-1	91.34		-0.50
2138	CPSC-CH-E1002-08	99.1		0.52	2504	IEC62321-5	98.131		0.39
2139	IEC62321-5	115.7		2.68	2511	CPSC-CH-E1002-08.3	94.1		-0.14
2165	IEC62321-5	103.9		1.14	2514	In house	99.0266		0.51
2176	IEC62321-5	106.6		1.49	2529		----		----
2182	CPSC-CH-E1002-08.3	87.62		-0.98	2557	IEC62321-5	103.7066		1.12
2184	IEC62321-5	95.5		0.05	2560	EN16711-1	97.89		0.36
2201	CPSC-CH-E1002-08	97.63		0.32	2561	CPSC-CH-E1002-08	98.92		0.49
2202	IEC62321-5	104		1.15	2563		----		----
2216	CPSC-CH-E1002-08	94.327		-0.11	2564	CPSC-CH-E1002-08	92.090		-0.40
2218	CPSC-CH-E1002-08.1	88.1167		-0.92	2568		----		----
2229	ASTM F963	100.9905		0.76	2569	CPSC-CH-E1002-08	97		0.24
2230	EPA3051	103		1.02	2572	IEC62321-5	94		-0.15
2232	IEC62321-5	98.320		0.41	2573	EPA3052	93.048		-0.27
2236	In house	97.03		0.25	2590	EN16711-1	88.907	C	-0.81
2250	EN16711-1	90.46		-0.61	2591	CPSC-CH-E1002-08.3	90.703		-0.58
2255	EN16711-1	103.05		1.03	2603	GB31604.49/5009.268	106.2470		1.45
2256	CPSC-CH-E1002-08	87.844		-0.95	2608		----		----
2258	EN16711-1	99.989		0.63	2621	IEC62321-5	93.9		-0.16
2264	CPSC-CH-E1002-08	91.341		-0.50	2624	EPA3052	25.4	R(0.01)	-9.09
2265	EN16711-1	86.8		-1.09	2630	IEC62321-5	97.6		0.32
2266	In house	92		-0.41	2643	CPSC-CH-E1002-08.3	88.60		-0.85
2271	EPA3052	92.2		-0.38	2644	EN16711-1	116.7		2.81
2284	EN16711-1	98.22		0.40	2649	CPSC-CH-E1002-08	94.5		-0.08
2290	IEC62321-5	95		-0.02	2672	IEC62321-5	89.144		-0.78
2293	CPSC-CH-E1002-08	88.98		-0.80	2674	CPSC-CH-E1002-08.3	99.3666		0.55
2294	CPSC-CH-E1002-08.3	99		0.50	2678		----		----
2297	IEC62321-5	91.6		-0.46	2734	CPSC-CH-E1002-08	100.34		0.68
2301	EN16711-1	96.4		0.16	2736		----		----
2310	EN16711-1	96.7		0.20	2794	IEC62321-3-1	103		1.02
2311	CPSC-CH-E1002-08.3	100.03		0.64	2804	EN16711-1	96.0717		0.12
2314	CPSC-CH-E1002-08	90.462		-0.61	2826	IEC62321-5	97.7879		0.34
2316	IEC62321-5	97.6		0.32	2829	CPSC-CH-E1002-08	94.67		-0.06
2320		106		1.42	2833	IEC62321-3-1	87		-1.06
2339	In house	79		-2.11	2835	IEC62321-5	94.79		-0.05
2347	IEC62321-5	95		-0.02	2853	CPSC-CH-E1002-08.3	92.55		-0.34
2350	IEC62321-5	95.93		0.10	2864	IEC62321-5	97.17		0.26
2352	CPSC-CH-E1002-08.3	94.8		-0.05	2870	EN16711-1	106.55		1.49
2353	IEC62321-5	90.99		-0.54	2885	IEC62321-5	101.8		0.87
2355	EPA3052	97.4		0.29	2900	IEC62321-5	105.4		1.34
2357	IEC62321-5	96.82		0.22	3100	CPSC-CH-E1002-08.3	96.03135		0.12
2358	EPA3051	93.3		-0.24	3110	ASTM F963	87.7		-0.97
2363	EPA3052	96		0.11	3111	EPA3052	106.5		1.48
2365	EN16711-1	95.2		0.01	3116	CPSC-CH-E1002-08.3	93.9		-0.16
2366	IEC62321-5	95.97		0.11	3118	EN16711-1	93.54		-0.21
2369	IEC62321-5	98.3		0.41	3121		----		----
2370	EPA3052	91.2		-0.51	3122	CPSC-CH-E1002-08	94		-0.15
2372	EPA3052	84.7		-1.36	3124	EPA3052	115		2.59
2374	CPSC-CH-E1002-08.3	91.84		-0.43	3146	In house	93.9		-0.16



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
3153	IEC62321-5	95.5		0.05	3214	EPA3052	100.80		0.74
3154	EN16711-1	107.2		1.57	3216	In house	79.568		-2.03
3160	CPSC-CH-E1002-08.3	87.08		-1.05	3225	CPSC-CH-E1002-08.3	92.88		-0.30
3163		90		-0.67	3228	IEC62321-5	101.0		0.76
3166		-----		-----	3237	EN16711-1	92.40		-0.36
3172	CPSC-CH-E1002-08	90.02		-0.67	3239	IEC62321-5	101.592		0.84
3182	IEC62321-5	95.370		0.03	3246	EN16711-1	103.49		1.09
3183	In house	< 0.1	f-	<-12.39	3248	CPSC-CH-E1002-08	83		-1.58
3185	CPSC-CH-E1002-08.3	96.3		0.15	6146	IEC62321-3-1	98.73		0.47
3190	IEC62321-5	98		0.37	8005		92.3		-0.37
3197	IEC62321-5	107.3		1.58	8008		-----		-----
3210	In house	93.7		-0.19					

normality
n
outliers
mean (n)
st.dev. (n)
R(calc.)
st.dev.(Horwitz)
R(Horwitz)

RSD=8%

Lab 2590 first reported 62.163

Lab 3183 possibly a false negative test result?



APPENDIX 2**Other reported Metals in sample #20650; results in mg/kg**

Lab	Sb	Cd	Cr	Cr6+	Co	Pb
110	----	----	----	----	----	----
210	----	----	----	----	----	----
230	----	----	----	----	----	----
339	<12.5	<1.5	<3	----	<1.5	----
348	----	< 5	< 5	----	----	< 10
362	----	< 5.0	----	----	----	----
523	<2	<2	<2	<8	<2	<2
551	----	----	----	----	----	----
623	not detected	not detected	not detected	not detected	not detected	not detected
826	N.D.	N.D.	N.D.	not detected	N.D.	N.D.
840	<2	<2	<2	<8	<2	<2
841	< 10	< 10	< 10	< 8	< 10	< 10
1051	----	<5	----	----	----	<10
1082	----	0.0335	1.6	----	0.0255	0.175
1126	----	0.958	----	----	----	----
1527	0.071	0.0358	1.72	----	0.186	0.307
1842	not analysed	not detected	not detected	not analysed	not analysed	not detected
1910	not analysed	< 0,34	not analysed	not analysed	not analysed	< 8,2
2108	not detected	not detected	not detected	not detected	not detected	not detected
2115	----	----	1.73	----	----	----
2118	0	0.025	1.632	0	0.031	0
2120	< 33	< 33	19.1	f+	< 33	< 33
2129	----	----	----	----	----	----
2131	0.6495	0	8.4345	----	0.136	0.265
2132	<10.0	<10.0	<10.0	----	----	<10.0
2137	----	----	----	----	----	----
2138	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2139	< 10	< 10	< 10	< 10	< 10	< 10
2165	Not applicable	<10	<10	Not applicable	Not applicable	<10
2176	not analysed	not detected	not detected	not detected	not analysed	not detected
2182	----	----	----	----	----	----
2184	Not applicable	<10	<10	<10	Not applicable	<10
2201	<10	<10	<10	<10	<10	<10
2202	N.D.	N.D.	N.D.	N.D.	N.D.	N.D.
2216	-----	less than 2	less than 5	-----	-----	less than 5
2218	-----	-----	-----	-----	-----	-----
2229	not detected	not detected	not detected	-----	-----	not detected
2230	<10	<10	<10	<10	<10	<10
2232	Not applicable	<10	<10	<10	Not applicable	<10
2236	-----	-----	-----	-----	-----	-----
2250	not detected	not detected	1.7337	-----	not detected	not detected
2255	n.d	n.d	n.d	n.d	n.d	n.d
2256	ND	ND	ND	ND	ND	ND
2258	Not Detected	Not Detected	Not Detected	not analysed	not analysed	Not Detected
2264	-----	less than 3.6	-----	-----	-----	less than 17.8
2265	-----	-----	-----	-----	-----	-----
2266	< detect. limit	< detect. limit	< detect. limit	-----	< detect. limit	< detect. limit
2271	ND	ND	ND	ND	ND	ND
2284	-----	-----	-----	-----	-----	-----
2290	<20	<20	<20	-----	<20	<20
2293	Not detected	Not detected	Not detected	Not detected	Not detected	Not detected
2294	Not analysed	Not analysed	Not analysed	Not analysed	Not analysed	Not detected
2297	<10	<10	<10	<10	<10	<10
2301	<10	<10	<10	<10	<10	<10
2310	Not detected	Not detected	Not detected	Not detected	Not detected	Not detected
2311	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2314	-----	ND	-----	-----	-----	ND
2316	-----	not detected	74.6	f+	not detected	not detected
2320	<1.0	<0.5	1.4	-----	<0.5	<0.5
2339	<5	<1	<5	-----	<5	<5
2347	<10	<2	<8	<8	<5	<2
2350	<10	<1	<5	<8	<5	<5
2352	-----	-----	-----	-----	-----	-----
2353	<10	<10	<10	<10	<10	<10
2355	<10	<5	<2	<8	<10	<2
2357	<10	<10	<10	-----	<10	<10
2358	n.d.	n.d.	n.d.	n.d.	n.d.	n.d.
2363	<10	<2	<2	<8	<5	<2
2365	<10	<10	<10	-----	<10	<10
2366	<10	<5	<10	<8	<10	<10
2369	<5	<2	<2	<8	<5	<2
2370	<4	<2	<4	<8	<6	<2
2372	< 2	< 2	< 2	< 8	< 2	< 2
2374	not detected	not detected	not detected	not analyzed	not detected	not detected

Lab	Sb	Cd	Cr	Cr6+	Co	Pb
2375	<10	<10	<10	<8	<10	<10
2378	----	----	----	----	----	----
2379	----	Not detected	----	Not detected	----	Not detected
2380	ND	ND	ND	ND	ND	ND
2381	Out Cap	<5	Out Cap	Out Cap	Out Cap	<10
2382	<20.0	<5.0	<10.0	NA	<10.0	<5.0
2384	3.59	not detected	not detected	not detected	----	not detected
2385	<5	<0,5	<5	<1,0	<5	<5
2387	----	Not Detected [<2]	Not Detected [<2]	Not Detected [<8]	----	Not Detected [<2]
2390	----	----	----	----	----	----
2391	<3	<3	<3	<2	<3	<3
2392	----	Not detected	----	Not detected	----	Not detected
2406	not detected	not detected	not detected	----	----	not detected
2409	0.077	0.169	6.986	4.775	0.594	0.125
2410	< 5	< 0.5	< 5	< 8	< 5	< 5
2413	----	----	----	----	----	----
2426	<20	<20	<20	----	<20	<20
2431	----	----	----	----	----	----
2444	----	0	----	0	----	0
2453	----	ND	----	----	----	----
2460	----	<6.0	----	----	----	<25
2475	not analysed	<5	not analysed	not analysed	not analysed	<25
2477	----	----	----	----	----	< L.O.Q.
2488	----	<10	<10	----	----	<10
2492	not detected	not detected	not detected	not detected	not detected	not detected
2500	----	not detected	----	----	----	not detected
2504	< 2	<2	<2	<5	<2	<2
2511	<10	<10	<10	<10	<10	<10
2514	----	----	----	----	----	----
2529	----	----	----	----	----	----
2557	----	< 2	< 2	< 8	----	< 2
2560	<10	<10	<10	<10	<10	<10
2561	<0.1	<0.1	4.06	----	<0.1	<0.1
2563	----	----	----	----	----	----
2564	----	not detected	----	----	----	not detected
2568	----	----	----	<50	----	----
2569	ND	ND	ND	----	ND	ND
2572	<20	<20	<20	----	<20	<20
2573	not detected	not detected	not detected	not detected	not detected	not detected
2590	< L.O.Q.	< L.O.Q.	2.3	----	< L.O.Q.	< L.O.Q.
2591	----	not detected	----	----	----	not detected
2603	----	0.0298	1.7392	----	----	0.3192
2608	----	----	----	----	----	----
2621	<5	<5	<5	<5	<5	<5
2624	< 20	< 1	265.3	f+	not determined	not determined
2630	----	not detected	not detected	not detected	----	not detected
2643	----	----	----	----	----	----
2644	----	----	----	----	----	----
2649	<10	<10	<10	<10	<10	<10
2672	< 5	<5	<5	<5	<5	<5
2674	----	< 10	< 10	< 10	----	< 10
2678	----	----	----	----	----	----
2734	not detected	not detected	not detected	not detected	not detected	not detected
2736	----	----	----	----	----	----
2794	not detected	not detected	not detected	not analysed	not detected	not detected
2804	<20	<20	<20	----	<20	<20
2826	----	<20	<20	<20	----	<20
2829	----	----	----	----	----	----
2833	Not detected	Not detected	Not detected	Not applicable	Not detected	Not detected
2835	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2853	----	----	----	----	----	Not detected
2864	not detected	not detected	not detected	not detected	not detected	not detected
2870	----	----	2.96	----	----	----
2885	----	N.D.	----	----	----	N.D.
2900	not analysed	not detected	not analysed	not analysed	not analysed	not detected
3100	<10	<5	<10	<10	<10	<10
3110	<15	<15	<15	----	----	<15
3111	----	0.011	----	----	----	1.197
3116	not detected	not detected	not detected	----	----	not detected
3118	ND	ND	ND	ND	ND	ND
3121	< 0,5	< 0,5	1.77	< 0,25	< 0,5	< 0,5
3122	< 5	< 5	< 5	< 5	< 5	< 5
3124	<0,5	<0,1	----	----	<0,2	<1
3146	not detected	not detected	not detected	----	not detected	not detected
3153	not detected	not detected	not detected	not detected	not detected	not detected
3154	----	----	5.862	----	----	0.22
3160	<16	<16	<16	----	<16	<16

Lab	Sb	Cd	Cr	Cr6+	Co	Pb
3163	not detected	not detected	not detected	not analysed	not detected	not detected
3166	-----	-----	-----	-----	-----	-----
3172	< 10	< 10	< 10	-----	< 10	< 10
3182	not analysed	<5	<5	not detected	not analysed	not detected
3183	-----	< 0.01	-----	-----	-----	< detect. limit
3185	<10	<10	<10	<10	<10	<10
3190	<2	<2	<2	<1	<2	179 f+
3197	<10	<10	<10	<10	<10	<10
3210	<25	<10	<10	-----	<10	<25
3214	<5	<2	<5	<8	<5	<2
3216	0.051	0.041	2.178	Not analysed	0.246	0.147
3225	Not analysed	<25	Not analysed	Not analysed	Not analysed	<15
3228	-----	<10	<10	-----	-----	<10
3237	-----	-----	-----	-----	-----	-----
3239	-----	not detected	not detected	not detected	-----	not detected
3246	not detected	not detected	not detected	not detected	not detected	not detected
3248	<10	<10	<10	<10	<10	<10
6146	Not detected	Not detected	Not detected	Not determined	Not detected	Not detected
8005	not detected	not detected	not detected	-----	-----	not detected
8008	-----	<5	-----	-----	-----	-----

Lab 2120 possibly a false positive test result for Cr

Lab 2316 possibly a false positive test result for Cr

Lab 2624 possibly a false positive test result for Cr

Lab 3190 possibly a false positive test result for Pb

Other reported Metals in sample #20650; results in mg/kg - continued -

Lab	Hg	Mn	Sn
110	-----	-----	-----
210	-----	-----	-----
230	-----	-----	-----
339	<3	5.23	<12.5
348	-----	-----	-----
362	-----	-----	-----
523	<2	4.764	<2
551	-----	-----	-----
623	not detected	not detected	not detected
826	N.D.	N.D.	-----
840	<2	<2	<2
841	< 10	< 10	< 10
1051	-----	-----	-----
1082	-----	4.88	0.030
1126	-----	-----	-----
1527	0.542	-----	-----
1842	not detected	not analysed	not analysed
1910	not analysed	not analysed	not analysed
2108	0.4133	not detected	not detected
2115	0.42	5.10	-----
2118	0.207	4.92	0
2120	< 0,83	-----	-----
2129	-----	-----	-----
2131	0	5.21	78.5965 f+
2132	<10.0	-----	<10.0
2137	-----	-----	-----
2138	Not Detected	Not Detected	Not Detected
2139	< 10	< 10	< 10
2165	<10	Not applicable	Not applicable
2176	not detected	not analysed	not analysed
2182	-----	-----	-----
2184	<10	Not applicable	Not applicable
2201	<10	<10	<10
2202	N.D.	N.D.	N.D.
2216	less than 5	-----	-----
2218	-----	-----	-----
2229	not detected	-----	-----
2230	<10	<10	<10
2232	<10	Not applicable	Not applicable
2236	-----	-----	-----
2250	0.4262	5.119	not detected
2255	n.d	n.d	n.d
2256	ND	ND	ND
2258	Not Detected	not analysed	not analysed
2264	-----	-----	-----
2265	-----	-----	-----
2266	< detect. limit	< detect. limit	11
2271	ND	ND	ND
2284	-----	-----	-----
2290	<20	<20	<20
2293	Not detected	Not Analysed	Not Analysed
2294	Not analysed	Not analysed	Not analysed
2297	<10	<10	<10
2301	<10	<10	<10
2310	Not detected	<10	Not detected
2311	Not Detected	Not Detected	Not Detected
2314	-----	-----	-----
2316	not detected	9.3	-----
2320	0.25	5.0	<1.0
2339	<1	-----	<5
2347	<2	<5	<5
2350	<2	<5	<10
2352	-----	-----	-----
2353	<5	NA	<10
2355	<2	<10	<10
2357	<10	<10	<10
2358	n.d.	N/A	n.d.
2363	<2	<5	<5
2365	<10	<10	<10
2366	<10	<50	<10
2369	<2	<5	<5
2370	<2	<6	<3
2372	< 2	<6	< 2
2374	not detected	not analyzed	not analyzed
2375	<10	<10	<10

Lab	Hg	Mn	Sn
2378	-----	-----	-----
2379	Not detected	-----	-----
2380	ND	ND	ND
2381	Out Cap	Out Cap	Out Cap
2382	<5.0	NA	NA
2384	not detected	-----	-----
2385	0.40	<5	<5
2387	Not Detected [<2]	-----	-----
2390	-----	-----	-----
2391	<3	<3	<3
2392	Not detected	-----	-----
2406	not detected	-----	-----
2409	1.003	4.905	not analyzed
2410	< 2	< 5	< 5
2413	-----	115.3	f+
2426	<10	-----	-----
2431	-----	-----	-----
2444	0	-----	-----
2453	-----	-----	-----
2460	-----	-----	-----
2475	not analysed	not analysed	not analysed
2477	-----	-----	-----
2488	<10	-----	-----
2492	not detected	not detected	not detected
2500	-----	-----	-----
2504	<2	<10	<2
2511	<10	<10	<10
2514	-----	-----	-----
2529	-----	-----	-----
2557	< 0.5	-----	-----
2560	<10	<10	<10
2561	0.44	4.04	<0.1
2563	-----	-----	-----
2564	-----	-----	-----
2568	-----	-----	-----
2569	ND	ND	ND
2572	<20	-----	-----
2573	not detected	not detected	not detected
2590	< L.O.Q.	6.1	< L.O.Q.
2591	-----	-----	-----
2603	-----	-----	-----
2608	-----	-----	-----
2621	<5	<5	<5
2624	< 35	12.1	< 20
2630	not detected	-----	-----
2643	-----	-----	-----
2644	-----	-----	-----
2649	<10	<10	<10
2672	<5	5.479	<5
2674	< 10	-----	-----
2678	-----	-----	-----
2734	not detected	not detected	not detected
2736	-----	-----	-----
2794	not detected	not detected	not detected
2804	<20	-----	-----
2826	<20	-----	-----
2829	-----	-----	-----
2833	Not detected	Not detected	Not analysed
2835	Not Detected	Not Detected	9.88
2853	-----	-----	-----
2864	not detected	not detected	not detected
2870	0.688	5.25	0.75
2885	N.D.	-----	-----
2900	not analysed	not analysed	not analysed
3100	<10	<10	<10
3110	<15	-----	-----
3111	0	-----	-----
3116	not detected	-----	-----
3118	ND	ND	ND
3121	0.37	4.90	< 0,5
3122	< 5	< 10	< 5
3124	0.414	<50	-----
3146	not detected	not detected	not detected
3153	not detected	not detected	not detected
3154	-----	5.801	f+
3160	-----	<16	<16
3163	not detected	not detected	not detected

Lab	Hg	Mn	Sn
3166	----	----	----
3172	< 10	----	< 10
3182	not detected	not analysed	not analysed
3183	----	----	----
3185	<10	<10	<10
3190	<2	<2	<2
3197	<10	<10	<10
3210	0.32	<10	<10
3214	<2	<5	<5
3216	0.587	6.124	0.037
3225	Not analysed	Not analysed	Not analysed
3228	<10	----	----
3237	----	----	----
3239	not detected	----	----
3246	not detected	6.57	not detected
3248	<10	<10	<10
6146	Not detected	Not detected	Not detected
8005	not detected	----	----
8008	----	----	----

Lab 2131 possibly a false positive test result for Sn

Lab 2413 possibly a false positive test result for Sn

Lab 3154 possibly a false positive test result for Sn

Other reported Metals in sample #20651; results in mg/kg

lab	Sb	Cr	Cr6+	Co	Cu	Hg
110	----	----	----	----	----	----
210	----	----	----	----	----	----
230	----	----	----	----	----	----
339	<12.5	<3	----	<1.5	<1.5	<3
348	----	<5	----	----	----	----
362	----	----	----	----	----	----
523	<2	<2	<8	<2	<2	<2
551	----	----	----	----	----	----
623	not detected	not detected	not detected	not detected	not detected	not detected
826	N.D.	N.D.	not detected	N.D.	N.D.	N.D.
840	<2	<2	<8	<2	<2	<2
841	< 10	< 10	< 8	< 10	< 10	< 10
1051	----	----	----	----	----	----
1082	----	1.15	----	0.010	1.1	----
1126	----	----	----	----	----	----
1527	1.322	0.976	----	0.044	1.44	----
1842	not analysed	not detected	not analysed	not analysed	not analysed	not detected
1910	not analysed	not analysed	not analysed	not analysed	not analysed	not analysed
2108	not detected	not detected	not detected	not detected	not detected	not detected
2115	0.85	0.76	----	----	2.91	----
2118	0	0.595	0	0.015	0	0
2120	< 33	< 33	----	< 33	< 33	< 0,83
2129	----	----	----	----	----	----
2131	19.1955	f+	5.758	0.0525	7.922	0
2132	<10.0	<10.0	----	----	----	<10.0
2137	----	----	----	----	----	----
2138	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2139	< 10	< 10	< 10	< 10	< 10	< 10
2165	Not applicable	<10	Not applicable	Not applicable	Not applicable	<10
2176	not analysed	not detected	not detected	not analysed	not analysed	not detected
2182	----	----	----	----	----	----
2184	Not applicable	<10	<10	Not applicable	Not applicable	<10
2201	<10	<10	<10	<10	<10	<10
2202	N.D.	N.D.	N.D.	N.D.	N.D.	N.D.
2216	----	less than 5	----	----	----	less than 5
2218	----	----	----	----	----	----
2229	not detected	not detected	----	----	----	not detected
2230	<10	<10	<10	<10	<10	<10
2232	Not applicable	<10	<10	Not applicable	Not applicable	<10
2236	<10.0	<10.0	----	----	----	<10.0
2250	not detected	not detected	----	not detected	not detected	not detected
2255	n.d	n.d	n.d	n.d	n.d	n.d
2256	ND	ND	ND	ND	ND	ND
2258	not detected	not detected	not analysed	not analysed	not analysed	not detected
2264	----	----	----	----	----	----
2265	----	----	----	----	----	----
2266	< detect. limit	< detect. limit	----	< detect. limit	< detect. limit	< detect. limit
2271	ND	ND	ND	ND	ND	ND
2284	----	----	----	----	----	----
2290	<20	<20	----	<20	<20	<20
2293	Not detected	Not detected	Not detected	Not detected	Not detected	Not detected
2294	Not analysed	Not analysed	Not analysed	Not analysed	Not analysed	Not analysed
2297	<10	<10	<10	<10	<10	<10
2301	<10	<10	<10	<10	<10	<10
2310	Not detected	Not detected	Not detected	Not detected	<10	Not detected
2311	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2314	----	----	----	----	----	----
2316	----	78.7	f+	not detected	6.3	not detected
2320	<1.0	0.70	----	<0.5	1.5	<0.1
2339	<5	<5	----	<5	<5	<1
2347	<10	<8	<8	<5	<5	<2
2350	<10	<5	<8	<5	<5	<2
2352	----	----	----	----	----	----
2353	<10	<10	<10	<10	NA	<5
2355	<10	<2	<8	<10	<10	<2
2357	<10	<10	----	<10	<10	<10
2358	n.d.	n.d.	n.d.	n.d.	N/A	n.d.
2363	<10	<2	<8	<5	<5	<2
2365	<10	<10	----	<10	<10	<10
2366	<10	<10	<8	<10	<50	<10
2369	<10	<2	<8	<5	<5	<2
2370	<4	<4	<8	<6	<6	<2
2372	<2	<2	<28	<2	<2	<2
2374	not detected	not detected	not analyzed	not detected	not detected	not detected
2375	<10	<10	<8	<10	<10	<10
2378	----	----	----	----	----	----

Lab	Sb	Cr	Cr6+	Co	Cu	Hg
2379	----	----	Not detected	----	----	Not detected
2380	ND	ND	ND	ND	ND	ND
2381	Out Cap	Out Cap	Out Cap	Out Cap	Out Cap	Out Cap
2382	<20.0	<10.0	NA	<10.0	<10.0	<5.0
2384	not detected	not detected	not detected	----	----	not detected
2385	<5	<5	<1,0	<5	<5	<0,1
2387	----	Not Detected [<2]	Not Detected [<8]	----	----	Not Detected [<2]
2390	----	----	----	----	----	----
2391	<3	<3	<2	<3	<3	<3
2392	----	----	Not detected	----	----	Not detected
2406	not detected	not detected	----	----	----	not detected
2409	0.839	1.952	not detected	not detected	0.487	0.217
2410	< 5	< 5	< 8	< 5	< 5	< 2
2413	----	----	----	----	----	----
2426	<20	<20	----	<20	<20	<10
2431	----	----	----	----	----	----
2444	----	----	0	----	----	0
2453	----	----	----	----	----	----
2460	----	----	----	----	----	----
2475	not analysed	not analysed	not analysed	not analysed	not analysed	not analysed
2477	----	----	----	----	----	----
2488	----	<10	----	----	----	<10
2492	----	----	----	----	----	----
2500	----	----	----	----	----	----
2504	< 2	< 2	< 5	< 2	< 2	< 2
2511	<10	<10	<10	<10	----	<10
2514	----	----	----	----	----	----
2529	----	----	----	----	----	----
2557	----	< 2	< 8	----	----	< 0.5
2560	<10	<10	Not analysed	<10	<10	<10
2561	0.42	<0.1	----	<0.1	<0.1	<0.1
2563	----	----	----	----	----	----
2564	----	----	----	----	----	----
2568	----	----	<50	----	----	----
2569	ND	ND	ND	ND	ND	ND
2572	<20	<20	----	<20	<20	<20
2573	not detected	not detected	not detected	not detected	not detected	not detected
2590	< L.O.Q.	< L.O.Q.	----	< L.O.Q.	< L.O.Q.	< L.O.Q.
2591	----	----	----	----	----	----
2603	----	0.6054	----	----	1.0740	----
2608	----	----	----	----	----	----
2621	<5	<5	<5	<5	<5	<5
2624	16.0	f+	< 40	not determined	not determined	< 4
2630	----	not detected	not detected	----	----	not detected
2643	----	----	----	----	----	----
2644	----	----	----	----	----	----
2649	<10	<10	<10	<10	<10	<10
2672	<5	<5	<5	<5	<5	<5
2674	----	< 10	< 10	----	----	< 10
2678	----	----	----	----	----	----
2734	not detected	not detected	not detected	not detected	not detected	not detected
2736	----	----	----	----	----	----
2794	not detected	not detected	not analysed	not detected	not detected	not detected
2804	<20	<20	----	<20	<20	<20
2826	----	<20	<20	----	----	<20
2829	----	----	----	----	----	----
2833	Not detected	Not detected	Not applicable	Not detected	Not analysed	Not detected
2835	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2853	----	----	----	----	----	----
2864	not detected	not detected	not detected	not detected	not detected	not detected
2870	----	----	----	----	29.32	f+
2885	----	----	----	----	----	N.D.
2900	not analysed	not analysed	not analysed	not analysed	not analysed	not analysed
3100	<10	<10	<10	<10	<10	<10
3110	<15	<15	----	----	----	<15
3111	----	----	----	----	----	0
3116	not detected	not detected	----	----	----	not detected
3118	ND	ND	ND	ND	ND	ND
3121	----	----	----	----	----	----
3122	< 5	< 5	< 5	< 5	< 5	< 5
3124	<0,5	1.52	----	<0,2	<1	<0,1
3146	not detected	not detected	----	not detected	not detected	not detected
3153	not detected	not detected	not detected	not detected	not detected	not detected
3154	----	3.482	----	----	3.562	----
3160	<16	<16	----	<16	<16	----
3163	not detected	not detected	not analysed	not detected	not detected	not detected
3166	----	----	----	----	----	----

Lab	Sb	Cr	Cr6+	Co	Cu	Hg
3172	< 10	< 10	----	< 10	----	< 10
3182	not analysed	<5	not detected	not analysed	not analysed	not detected
3183	----	----	----	----	----	----
3185	<10	<10	<10	<10	<10	<10
3190	<2	<2	<1	<2	<2	<2
3197	<10	<10	<10	<10	<10	<10
3210	<25	<10	----	<10	<10	<0.02
3214	<5	<5	<8	<5	<5	<2
3216	0.584	0.819	Not analysed	0.017	0.575	Not detected
3225	Not analysed	Not analysed	Not analysed	Not analysed	Not analysed	Not analysed
3228	----	<10	----	----	----	<10
3237	----	----	----	----	----	----
3239	----	not detected	not detected	----	----	not detected
3246	not detected	not detected	not detected	not detected	not detected	not detected
3248	<10	<10	<10	<10	<10	<10
6146	Not detected	Not detected	Not determined	Not detected	Not detected	Not detected
8005	not detected	not detected	----	----	----	not detected
8008	----	----	----	----	----	----

Lab 2131 possibly a false positive test result for Sb

Lab 2316 possibly a false positive test result for Cr

Lab 2624 possibly a false positive test result for Sb

Lab 2870 possibly a false positive test result for Cu

Other reported Metals in sample #20651; results in mg/kg - continued -

Lab	Ni	Mn	Sn
110	----	----	----
210	----	----	----
230	----	----	----
339	<1.5	<1.5	<12.5
348	----	----	----
362	----	----	----
523	<2	<2	<2
551	----	----	----
623	not detected	not detected	not detected
826	N.D.	N.D.	----
840	<2	<2	<2
841	< 10	< 10	< 10
1051	----	----	----
1082	0.313	0.335	0.0405
1126	----	----	----
1527	0.173	----	----
1842	not analysed	not analysed	not analysed
1910	not analysed	not analysed	not analysed
2108	not detected	not detected	not detected
2115	----	0.38	0.26
2118	0.184	0.256	0.272
2120	< 33	----	----
2129	----	----	----
2131	0.8285	0.4295	60.777 f+
2132	----	----	<10.0
2137	----	----	----
2138	Not Detected	Not Detected	Not Detected
2139	< 10	< 10	< 10
2165	Not applicable	Not applicable	Not applicable
2176	not analysed	not analysed	not analysed
2182	----	----	----
2184	Not applicable	Not applicable	Not applicable
2201	<10	<10	<10
2202	N.D.	N.D.	N.D.
2216	----	----	----
2218	----	----	----
2229	----	----	----
2230	<10	<10	<10
2232	Not applicable	Not applicable	Not applicable
2236	----	----	----
2250	not detected	not detected	not detected
2255	n.d	n.d	n.d
2256	ND	ND	ND
2258	not analysed	not analysed	not analysed
2264	5.397	----	----
2265	----	----	----
2266	< detect. limit	< detect. limit	< detect. limit
2271	ND	ND	ND
2284	----	----	----
2290	<20	<20	<20
2293	Not detected	Not Analysed	Not Analysed
2294	Not analysed	Not analysed	Not analysed
2297	<10	<10	<10
2301	<10	<10	<10
2310	Not detected	Not detected	Not detected
2311	Not Detected	Not Detected	Not Detected
2314	----	----	----
2316	57.2	f+ 6.6	----
2320	3.6	0.94	<1.0
2339	----	<5	<5
2347	<5	<5	<5
2350	<5	<5	<10
2352	----	----	----
2353	<10	NA	<10
2355	<10	<10	<10
2357	<10	<10	<10
2358	n.d.	N/A	n.d.
2363	<5	<5	<5
2365	<10	<10	<10
2366	<50	<50	<10
2369	<5	<5	<5
2370	<6	<6	<3
2372	<2	<6	<2
2374	not detected	not analyzed	not analyzed
2375	<10	<10	<10

Lab	Ni	Mn	Sn
2378	----	----	----
2379	----	----	----
2380	ND	ND	ND
2381	Out Cap	Out Cap	Out Cap
2382	<10.0	NA	NA
2384	----	----	----
2385	<5	<5	<5
2387	----	----	----
2390	----	----	----
2391	<3	<3	<3
2392	----	----	----
2406	not detected	----	----
2409	0.105	0.189	not analyzed
2410	< 5	< 5	< 5
2413	----	----	----
2426	<20	----	----
2431	----	----	----
2444	----	----	----
2453	----	----	----
2460	----	----	----
2475	not analysed	not analysed	not analysed
2477	----	----	----
2488	----	----	----
2492	----	----	----
2500	----	----	----
2504	<2	<2	<2
2511	<10	<10	<10
2514	----	----	----
2529	----	----	----
2557	----	----	----
2560	<10	<10	<10
2561	<0.1	1.62	0.99
2563	----	----	----
2564	----	----	----
2568	----	----	----
2569	ND	ND	ND
2572	<20	----	----
2573	not detected	not detected	not detected
2590	< L.O.Q.	< L.O.Q.	< L.O.Q.
2591	----	----	----
2603	0.3492	----	----
2608	----	----	----
2621	<5	<5	<5
2624	< 7	not determined	< 10
2630	----	----	----
2643	----	----	----
2644	----	----	----
2649	<10	<10	<10
2672	<5	<5	<5
2674	----	----	----
2678	----	----	----
2734	not detected	not detected	not detected
2736	----	----	----
2794	not detected	not detected	not detected
2804	<20	----	----
2826	----	----	----
2829	----	----	----
2833	Not analysed	Not detected	Not analysed
2835	Not Detected	Not Detected	15.1
2853	----	----	----
2864	not detected	not detected	not detected
2870	3.42	----	1.16
2885	----	----	----
2900	not analysed	not analysed	not analysed
3100	<10	<10	<10
3110	----	----	----
3111	----	----	----
3116	----	----	----
3118	ND	ND	ND
3121	----	----	----
3122	< 5	< 5	< 5
3124	1.68	<50	----
3146	not detected	not detected	not detected
3153	not detected	not detected	not detected
3154	----	----	0.821
3160	<16	<16	<16
3163	not detected	not detected	not detected

Lab	Ni	Mn	Sn
3166	----	----	----
3172	< 10	----	< 10
3182	not analysed	not analysed	not analysed
3183	----	----	----
3185	<10	<10	<10
3190	<2	<2	<2
3197	<10	<10	<10
3210	<10	<10	<10
3214	<5	<5	<5
3216	0.787	1.292	0.031
3225	Not analysed	Not analysed	Not analysed
3228	----	----	----
3237	----	----	----
3239	----	----	----
3246	not detected	not detected	not detected
3248	<10	<10	<10
6146	Not detected	Not detected	Not detected
8005	----	----	----
8008	----	----	----

Lab 2316 possibly a false positive test result for Ni
 Lab 2131 possibly a false positive test result for Sn

APPENDIX 3 Analytical details

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
110	---	---	---	---	---	---	---
210	No	---	Further cut	Microwave	---	---	ICP-MS
230	---	---	---	---	---	---	---
339	No	0.1	Used as received	Acid digestion + microwave	Adding HNO ₃ + H ₂ O ₂ Step1: 120°C 11min Step2: 240°C 9min Step3: 240°C 8min Adding HF after cooling Step4: 200°C 8min Step5: 200°C 5min	4mL HNO ₃ + 1mL H ₂ O ₂ 1mL HF 10 ml HNO ₃ (67-69%); 2 ml H ₂ O ₂ (30-32%); 3 ml HCl (32-35%)	ICP-AES
348	Yes	0.15	Further cut	Acid Digestion + microwave	850 W max, Temperature max: 210°C, Time: 1h 30 min (including cool down step)	---	ICP-MS
362	Yes	0.5000	Further cut	Acid Digestion	195 °C	H ₂ SO ₄ ; H ₂ O ₂	ICP-OES
523	Yes	0.2	Further cut	Acid Digestion	---	HNO ₃ /HCl	ICP-MS
551	---	---	---	---	---	---	---
623	Yes	0.5	Further cut	Microwave	Time 40 minutes Temperature min 140 C and max 210 C	65% H ₂ O ₂ 30%	ICP-MS
826	---	---	Further grinded	---	5 minutes RAM to 190°C, HOLD 1 minute, 8 minutes RAM to 210°C, HOLD 45 minutes	---	---
840	Yes	1	Further cut	Acid Digestion	---	HNO ₃ +HCl+H ₂ O ₂	UV-VIS
841	Yes	0.1009	Further cut	Microwave	---	---	ICP-OES
1051	Yes	---	Further cut	---	---	---	---
1082	No	0.2	Used as received	Microwave Since XRF was used, the metals were not released from the sample	25min --> 50C 15min --> 100C 15min --> 180C 5min --> 220C	65 w% HNO ₃ 5ml + H ₂ O 3ml	ICP-MS
1126	Yes	2 mm thick	Further cut	Not applicable for our in house method	Not applicable for our in house method	XRF	
1527	Yes	0.25	Further cut	Microwave	step1 5min 70°C step2 5min 95°C	nitric acid+hydrogen peroxide	ICP-MS
1842	Yes	---	Used as received	---	step3 5min 135°C step4 6min 175°C step5 16min 210°C step6 20min 220°C vent 15min	---	XRF
1910	Yes	0.1 (Pb) 0.8 (Cd)	Further cut	microwave (Pb) - dry ashing (Cd)	3h to achieve 500C (Cd) (dry ashing) Ramp to 120°C, hold for 0 min Ramp to 180°C, hold for 5 min Ramp to 200°C, hold for 5 min Ramp to 230°C, hold for 30 min	8 ml HNO ₃ (Pb) (microwave) 8 ml HNO ₃ (concentrated) + 2 ml HCl (concentrated)	AAS
2108	Yes	0.1	Further cut	Acid Digestion	5 min Max. 1000 W 140 °C 5 min	HNO ₃	ICP-MS
2115	Yes	0.15	Further cut	Microwave	Max. 1000 W 190 °C 5 min Max. 1000 W 210 °C 35 min Max. 1000 W 210 °C	---	ICP-MS
2118	Yes	0.1 to standard used: EN 1122: 0.5g CPSC-CH-E1002-08.3: 0.3g EN 16711-1: 0,15g	Further cut	Microwave	Power Ramp Pressure Temperature Hold 800 W 20 800 psi 140°C 5 min 800 W 25 800 psi 200°C 40 min	See remarks HNO ₃ conc.	ICP-MS
2120	Yes	0.05 - 0.200	Further cut	Microwave	in 40 min heated up to 230°C	Nitric Acid & Hydrogen peroxide	ICP-MS
2129	Yes	0.05 - 0.200	Further cut	Microwave	1200 W for 15 min 1200 W for 30 min 450 W for 15 min	HNO ₃ + H ₂ O ₂	ICP-MS
2131	Yes	0.12	Used as received	Microwave	210 °C, 30 min	only HNO ₃ (for microwave)	ICP-MS
2132	Yes	0.15	Further grinded	Microwave	Ramping time 30 min. ramping Temp. 190°C Holding time 15 min.(at 190°C)	HNO ₃ + H ₂ O ₂	ICP-OES
2137	Yes	2g (0.1g / test)	Used as received	Microwave	---	only HNO ₃ (for microwave)	ICP-OES
2138	Yes	0.2	Further cut	Microwave	step1: temp 25 to 110°C 10 min step2: temp 110 to 180°C 10 min step3: temp 180 to 230°C 10 min step4: maintain temp 230°C 7 min	---	ICP-OES
2139	Yes	0.1	Further cut	Microwave	temp to 220°C in 30 minutes, keep 10 minutes	nitric acid 20 % only Nitric acid and hydrochloric acid.	ICP-OES
2165	Yes	0.1 ICP-OES: 0.1	Further cut	Microwave	temp to 220°C in 30 minutes, keep 10 minutes	HNO ₃ and H ₂ O ₂	ICP-OES
2176	Yes	UV-VIS: 0.15	Further cut	Microwave	220 - 235 °C		ICP-OES

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
2182	Yes	Lead: 0.15 Cadmium: 0.50	Further cut	Microwave	For Lead Time ,min Temperature, °C Voltage, W 20 210 1000 10 210 1000 ---	5mL of concentrated nitric acid for lead ---	ICP-OES
2184	---	---	---	---	room temperature to 205°C,25 min(800W),Cool down 30min	7MLHNO3/1H2o2	---
2201	Yes	0.2	Further cut	Microwave	Microwave program / 1hour / 240°C tempo 200C in 20 min then hold for 10 minutes at 200C	Nitric acid + Hydrogen peroxide	ICP-OES
2202	Yes	0.4	Used as received	Microwave	Microwave program / 1hour / 240°C tempo 200C in 20 min then hold for 10 minutes at 200C	ICP-OES	ICP-OES
2216	Yes	0.1	Further grinded	Acid Digestion	5 mL nitric acid	ICP-OES	ICP-OES
2218	Yes	0.15	Further cut	Microwave	---	---	ICP-OES
		#20650: ASTM F963-17: 0.1012/ 0.1013 EN1122: 0.1018/ 0.0999 #20651 ASTM F963-17: 0.1020/ 0.0999 EN1122:			Microwave procedure: step1:100°C, 2min , 10atm step2:140°C, 3min, 20atm step3:170°C, 3min, 25atm step4:190°C, 4min, 35atm step5:210°C, 15min, 40atm 210°C Stage Temperature (°C) Hold Time (min) Ramp Time (min) 1 150 0 10 2 190 0 5 3 210 30 5	All samples are tested by microwave, we used nitric acid and hydrogen peroxide.	ICP-OES
2229	Yes	0.0991/0.1003	Further cut	Microwave	HNO3+HF	ICP-OES	ICP-OES
2230	Yes	0.2	Further cut	Microwave	Nitric acid and hydrogen peroxide	ICP-OES	ICP-OES
2232	Yes	0.2	Further cut	Microwave	4.5mL Conc. nitric acid 1.5mL Conc. hydrochloric acid 1.0mL 50% Hydrogen peroxide	ICP-OES	ICP-OES
2236	Yes	0.1	Further grinded	Microwave	Ramp to 205°C for 25 minutes Hold at 205°C for 25 minutes	3 different: Nitric acid only; Hydrochloric acid/Nitric acid mixture and for Cadmium a mixture from sulfuric acid and nitric acid	ICP-OES
2250	Yes	0.1 0.0947 S #20650 and 0.0916 S	Used as received	Microwave	Step 1: 10 min; 1000 W to reach 110 °C; Step 2: 10 min, 1000 W to reach 180 °C and Step 3: 30 min, 1000 W to hold 180 °C	ICP-MS	ICP-MS
2255	Yes	#20651	Further cut	Microwave	Temp 180°C ; Ramping 15 mins , Hold for 10 mins at same temp	HNO3 10 mL HNO3 + 1 mL H2O2	ICP-MS
2256	Yes	0.2016	Further cut	Microwave	60 min, 210°C 150°C Ramp Time 10 Min. Hold Time 0 Min. -190°C Ramp Time 5 Min. Hold Time 0 Min. -210°C Ramp Time 5 Min. Hold Time 60 Min.	ICP-AES	ICP-AES
2258	Yes	#20650 0.2005 #20651 0.2002	Further cut	Microwave	n/a Nitric acid concentrated	ICP-OES	AAS
2264	No	150	Further grinded	Microwave	210°C / 45 minutes First Step 5 minutes, 160C° Second Step 10 minutes, 200C°	ICP-OES	ICP-OES
2265	Yes	0.1	Further cut	Microwave	---	HNO3	ICP-MS
2266	Yes	0.2	Further cut	Microwave	120°C 5min 150°C 5min 180°C 10min 210°C 30min	---	ICP-MS
2271	Yes	0.1	Further cut	Other	120°C 5min, 150°C 5min, 180°C 10min, 210°C 30min	HNO3+HCL 8mLHNO3+2.5mLH2O2	ICP-OES
2284	Yes	0.15	Further cut	Microwave	10min, 210°C 30min	ICP-OES	ICP-OES
2290	Yes	---	---	---	20°C to 200°C / 10°C per min, total 30 min.	---	---
2293	Yes	0.3 #20650: 0.0952 #20651:	Further cut	Microwave	Temperature ramp 200°C in 25 minutes Hold 10 minutes These conditions for both samples	Nitric acid	ICP-MS
2294	Yes	0.1005	Further cut	Microwave	up to 190°C within 20min, and keep for 15mins	NHO3 concentrated in both samples 8mL HNO3+2mL H2O2	ICP-OES
2297	Yes	0.2	Further cut	Microwave	150,190,210 60 menit	---	ICP-OES
2301	Yes	0.2	Further cut	Microwave	180°C	10ML	ICP-MS
2310	Yes	0.25	Further cut	Microwave			ICP-MS

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
2311	Yes	0.1	Further cut	Microwave	15-100-10hold 10-150-10hold 10-	---	ICP-MS
2314	Yes	0.2	Further cut	Acid Digestion	200-30hold 180	HNO3	AAS
		0.2 heavy metal; 0.15 Hexavalent Chromium			200°C for 40 minutes @ energy 1000W then cooling it for 10 minutes	Microwave digestion with acid combination of HNO3 and HF.	
2316	Yes		Further cut	Microwave			ICP-OES
2320	---	---	---	---	---	---	---
2339	No	0.2	Further cut	Microwave	---	---	ICP-AES
2347	Yes	---	---	---	---	---	---
		Cr6+: #20650: 0.1, #20651: 0.15 Cd: #20650, #20651: 0.5 Pb, Hg: #20650, #20651: 0.2 Sb, Cr, Co, Cu, Ni, Mn, Sn: #20650, #20651: 0.2			Cr6+ : #20650(60°C 1Hr), #20651(Ambient --> 155°C : 10min, 155°C holding : 1.5 Hr, cooling time : 20 min) Cd : #20650, #20651 (210°C) Pb, Hg, Sb, Cr, Co, Cu, Ni, Mn, Sn : #20650, #20651 (Ambient --> 80°C :5min, 80°C --> 40°C : 5min, 40°C --> 210°C : 15min, 210°C holding : 15min, cooling time : 20	Cd : #20650, #20651(H2SO4 + H2O2) Pb, Hg, Sb, Cr, Co, Cu, Ni, Mn, Sn : #20650, #20651 (HNO3 + HF)	
2350	Yes	#20651: 0.2	Further cut	Microwave	The temperature was raised to 130°C for 6 min and keep for 3 min 5 min, 130°C rise to 160°C and keep for 3 min 5 min, 160°C rise to 180 °C and keep for 15 min 210 C , 40 minutes		ICP-OES
2352	Yes	0.25	Other	Microwave		HNO3:HCL=4:1	ICP-OES
2353	Yes	0.25	Further cut	Acid Digestion		HNO3	ICP-OES
2355	Yes	0.25	Further cut	Acid Digestion	200°C for 30mins	7ml HNO3+3ml HCl	ICP-OES
2357	---	---	---	---	---	---	---
2358	Yes	0.25	Further cut	Microwave	210 degree C	HNO3	ICP-OES
2363	Yes	0.25	Further cut	Microwave	max temperature £°210°C Nr t [h£°m£°s] T [jæ] 1 00£°10£°00 165 2 00£°04£°00 165 3 00£°07£°00 205 4 00£°30£°00 205	HNO3+HCl+HF	ICP-OES
2365	Yes	0.2	Further cut	Microwave	rise to 165°C within 10min, then hold 4min; rise to 205°C within 7min, then hold 30min, cooled to r.m. finally	HNO3+H2O2+HBF4	ICP-OES
2366	Yes	0.2	Further cut	Microwave			ICP-OES
2369	---	---	---	---	Step 1: Heat from room temperature to 170¢J within 10 minutes. Step 2: Maintain at 170¢XC for 5 minutes. Step 3: Heat from 170¢XC to 190¢XC in 10 minutes. Step 3: Maintain at 190¢XC for 25 minutes.	---	---
2370	Yes	0.1~0.15 g	Further grinded	Acid Digestion		HNO3 & HF.	ICP-OES
2372	No	0.1	Further cut	Microwave		HNO3	ICP-OES
2374	Yes	0.2g EN 1122; others : 0.1g	Used as received	For EN 1122: hot plate others: microwave	7 min ramp to 150°C and hold for 3 min; 5 min ramp to 180°C and hold for 3 min; 5 min ramp to (200;¡Ä5) ¡æ and hold for 15 min.	6mL HNO3+ 2mL HCL Nitric Acid - Peroxide	ICP-OES
2375	Yes	0.15	Used as received	Microwave	---		ICP-MS
2378	Yes	0.1	Used as received	Microwave	210°C,20min	---	ICP-OES
		0.25g For 2 samples of Chromium as Cr6+ used 0.15g			180 degree Celsius For sample 20650 of Chromium as Cr6+ used 63 degree Celsius/ 6 hr. For sample 20651 of Chromium as Cr6+ used 155 degree Celsius/ 5 hr	HNO3+H2O2	
2379	Yes	0.15g	Further grinded	Microwave	Time Duration (in min) Temperature 3 140 2 140 5 190 5 190 5 210 20 210	None for Chromium as Cr6+	ICP-OES
2380	Yes	0.25	Further cut	Microwave	Step 1: 3 min & 140°C Step 2: 2 min & 140°C Step 3: 5 min & 190°C Step 4: 5 min & 190°C Step 5: 5 min & 210°C	HNO3 & H2O2	ICP-OES
2381	Yes	0.25	Further cut	Microwave	Step 6: 20 min & 210°C		AAS
2382	Yes	0.1	Further cut	Microwave	1、7min ramp to 150°C and hold for 3 min 2、5min ramp to 180°C	1ml HCL+5ml HNO3+1ml HF	ICP-OES

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
					and hold for 3 min 3、5min ramp to 200°C and hold for 15 min		
2384	Yes	0.2	Further grinded	Microwave	---	Nitric Acid	AAS
2385	Yes	0.1	Further cut	Microwave	---	---	ICP-OES
2387	Yes	0.2 #20650: 0.2045 #20651: 0.2016	Further grinded	Microwave	---	---	ICP-OES
2390	Yes	0.10	Further cut	Microwave	20 minutes ramp time to 210 c20 minutes hold time (like stage 6 of the method) Step Target temp. (°C) Pressure max (bar) Ramp time (min) Hold time (min) Power (%) 1 150 60 10 2 90 2 200 60 5 10 90 3 50 60 10 0 0 4 50 60 0 0 5 50 60 0 0 Temperature at 180+-5 degree Celsius, pressure 400 PSI, Ramping 20 min, Holding 20 min	10 ml Conc HNO3 5ml HNO3 + 5ml H2SO4 for sample #20650 10ml HNO3 for sample #20651	ICP-OES
2391	Yes	1.0	Further grinded	Microwave	Temperature is increased to 200°C within 20 minutes, then hold for 30 minutes, and cool down for 60 minutes at last.	HNO3 : H2O2 10 : 1	ICP-OES
2406	Yes	0.1	Further cut	Microwave	step1 : 120°C 10 min step2 : 180°C 15 min 200 °C 25 min ramping, 25 min holding		ICP-OES
2409	No	0.2	Further cut	Microwave	HDXRF Analysis at Room Temperature	Nitric acid	ICP-MS
2410	Yes	0.2	Further cut	Microwave	Other	HNO3, HF	ICP-OES
2413	Yes	Not Applicable	Used as received			Not Applicable	XRF
2426	---	---	---		---	---	---
2431	Yes	0.1 0.2 for Hg, Pb, Cd and 0.1 for	Further cut	Microwave	10 mins to 200C, hold 20mins 1st heatup: 0 to 220°C; 60 min, 2nd holding: 220°C; 60 min,	HNO3, HCl	ICP-MS
2444	Yes	Cr6+	Further cut	Acid Digestion	3rd cooldown: 220 to 40°C;15 min	HNO3 : H2O2	ICP-OES
2453	Yes	0.1	Further cut	Microwave	Microwave - Cd Microwave - Pb	---	ICP-OES
2460	Yes	0.2-Pb 0.5-Cd	Further cut		Rampa a 200°C en 20 min, mantener 10 min – Pb Alta temperatura durante 1 h - Cd	H2SO4 y H2O2 -Cd HNO3- Pb	AAS
2475	Yes	0.1394	Used as received	Microwave	200°C, 20 MIN, 10 MIN	---	ICP-OES
2477	Yes	0.15	Further cut	Microwave		HNO3	ICP-OES
2488	Yes	---	Further cut	Microwave		---	ICP-MS
2492	Yes	0.1	Used as received	Microwave		---	AAS
2500	Yes	0.2003	Further cut	Microwave	Temp.:210°C Hold time: 60 min Heat 10 min to 200°C Stand for 30C at 200C	---	ICP-OES
2504	Yes	0.2	Further cut	Microwave	Cool down to RT 20 min	-	ICP-OES
2511	---	---	---	Microwave	---	---	ICP-MS
2514	Yes	20650=0.3861; 20651=0.3493	Further cut	Microwave	200°C in 20 min and holding at 200°C for 10 min	Nitric acid	AAS
2529	---	---	---		---	---	---
2557	Yes	0.1	Further cut	Other	210°C-30min	HNO3	ICP-MS
2560	Yes	0.1	Further cut	Microwave	65 mins and 190°C	Not applicable 20650 - Nitric Acid only 20651 - 8ml nitric acid, 1 ml Hydrochloric acid.	ICP-MS
2561	Yes	0.05	Further cut	Microwave	Ramp 20 mins Room temperature to 200°C 10 min hold		ICP-MS
2563	---	---	---		---	---	---
2564	Yes	---	Further cut	Acid Digestion	---		AAS
2568	Yes	0.1	Further grinded	Other	---		UV-VIS
2569	Yes	0.25	Further cut	Microwave	---		ICP-OES
2572	---	---	---		microwave method: increasing the temp. to 200°C in approx. 20 min., and holding for 25 mins.	8mL of HNO3 and 1mL of H2O2	---
2573	Yes	2	Further cut	Microwave	Step1: 11 mins to reach 120°C Step2: 2 mins stable at 120°C Step3: 9 mins to reach to 220°C Step4: 10 mins stable at 220°C		ICP-OES
2590	Yes	0.2	Further cut	Microwave			ICP-MS
2591	Yes	0.25	Further cut	Microwave		HNO3 8mL Nitric acid + 2mL	ICP-MS
2603	Yes	0.2	Further cut	Microwave	Microwave digestion temperature:200°C; Keep the time:30 minutes.	Hydrogen peroxide	ICP-MS
2608	---	---	---		---	---	---

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
2621	Yes	0.1-0.2	Further cut	Microwave	---	5N2NO3: 2H2O2	ICP-OES Cd Pb (AAS) Others (ICP-OES or XRF)
2624	No	2	Further cut	Microwave	---	HCl 6 ml + HNO3 2 ml + H2O2 2 ml + H2O 5 ml	
2630	Yes	0.05	Further grinded	Microwave	Step Heating(min) Target temp(°C) Lasting(min) 1. 14 120±5 1 2. 7 150±5 10 3. 10 180±5 10 4. Cool to room temp		ICP-OES
2643	Yes	0.15	Further cut	Acid Digestion	---	Sulfuric acid/Nitric acid	ICP-OES
2644	Yes	0.5	Further cut	Microwave	140°C 5 min 250°C 30 min 50°C 10 min	---	ICP-OES
2649	Yes	1	Further cut	Microwave	210°C / 40 min Step1: Ramp 30min /Hold 20 min / Temp 120°C / Power 700W Step2: Ramp 25 min / Hold 30 min / Temp 230°C / Power 1100W	HNO3 + HCL	ICP-MS
2672	Yes	0.1	Further cut	Microwave	Microwave digestion program: increase temp to 200°C and hold	HNO3 / HCl	ICP-MS
2674	Yes	0.1~0.2	Further cut	Microwave	---	10 mL HNO3	ICP-OES
2678	---	---	---	Microwave	---	---	ICP-OES
2734	Yes	2g each	Further cut	Microwave	---	---	
2736	---	---	---	Microwave	---	---	
2794	Yes	not applicable; XRF analyzed	Used as received	---	---	---	XRF
2804	Yes	0.5	Used as received	Microwave	Increase to at least 200°C within 20 mins, then hold for 10 mins. 30 mins cool down.		ICP-OES
2826	Yes	0.2	Used as received	Microwave	Increase to at least 200°C within 20mins, then hold for 10 mins. 30 mins cool down.		ICP-OES
2829	Yes	150	Further cut	Microwave	---	---	ICP-OES
2833	No	---	Used as received	---	---	---	XRF
2835	Yes	0.2000	Further cut	Microwave	Step 1:145°C Ramp 7 mins hold 3 mins Step 2: 210°C Ramp 10 mins hold 5 mins	HNO3 and H2O2	ICP-OES
2853	Yes	0.1	Further cut	Microwave	Step 3: 50°C hold 10 mins raise to 200°C in 25 minutes, then hold for 20 minutes	3% v/v Nitric acid 5mL HNO3+1.5mL HF+1mL HCl+1mL H2O	ICP-MS
2864	Yes	0.25	Further cut	Microwave	195°C 45min Level max.Power Power Ramp Pressure Temperature Stirrer Hold 1 1600 W 100% 10:00 08.30 bar 120 3 00:00 2 1600 W 100% 15:00 08.30 bar 180 3 05:00 3 1600 W 100% 15:00 08.30 bar 200 3 05:00 4 1600 W 100% 15:00 08.30 bar 230 3 30:00:00	Nitric acid and Hydrogen peroxide	ICP-OES
2870	Yes	0.1	Further cut	Microwave	15:00 08.30 bar 230 3 30:00:00		ICP-MS
2885	No	0.25	Further cut	Microwave	Step 1: Incorporates reaching 210 °C for 20 mins; Step 2: Remaning at 210 degree for 15 minns; Step 3: Cooling for 20 mins,	---	ICP-OES
2900	No	0.2 CPSC-CH- E1002-08.3: 0.1 EN1122:2001: 0.5 IEC 62321-7-	Further cut	Microwave	Sample to 205°C in 20 minutes, and holding at 205°C for 12 minutes.	HNO3 + HF	ICP-OES
3100	Yes	2:2017: 0.1	Further grinded	Microwave		CPSC-CH- E1002-08.3 was used HNO3 EN 1122:2001 was used	Other, please specify in the remarks below
3110	Yes	0.1	Further cut	Microwave	---	H2SO4,H2O2	ICP-OES
3111	No	0.2	Used as received	Microwave	---	HNO3 alone	ICP-OES
3116	--	--	--	Microwave	---	---	--
3118	Yes	0.15 Cr(VI) 2g	Further cut	Microwave	0 – 1300W in 10 min. 1300 W in 30 min. Cool to 70 °C	---	ICP-MS
3121	No	Metals 0,1g	Used as received	Microwave	210°C, 20 min	HNO3 + HCl	ICP-MS
3122	Yes	0.5	Further cut	Microwave	A power controlled program was used with 210-230 degrees Celsius in minimum 10 minutes.	HNO3	ICP-OES
3124	Yes	150-200	Used as received	Acid Digestion		HNO3 + HF + HCL.	ICP-MS

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
3146	Yes	0.25	Further cut	Microwave	Step 1: 02:30 min -> 70 °C Step 2: 12:00 min -> 200 °C Step 3: 18:00 min -> 250 °C Stage 1: Ramp from room temp to 140°C, hold for 5 mins. Stage 2: Ramp temp from 140°C to 200°C in 5 mins, hold for 20 mins 20 min to 200°C, 25 min holding at 200°C, cooling to room temperature	4.5 ml HNO3 (65%) + 1.5 ml HCl (25%)	ICP-MS
3153	Yes	0.1	Further grinded	Microwave	Step 1: 210°C in 20 mins Step 2: Hold at 210°C for 15 mins	3 HNO3 + 1 HCl	ICP-OES
3154	Yes	0.2	Further cut	Microwave	Step 1: 210°C in 20 mins Step 2: Hold at 210°C for 15 mins	HNO3/H2O2	ICP-MS
3160	Yes	0.15	Further cut	Microwave	Step 1: 210°C in 20 mins	---	ICP-MS
3163	No	1	Used as received	---	Step 2: Hold at 210°C for 15 mins	x	XRF
3166	---	---	---	---	---	---	---
3172	Yes	0.1	Used as received	Acid Digestion	210°C Step 1 Room temp to 200°C use time 15 min and hold time 10 min Step 2 200°C to 180°C use time 10 min	HNO3	ICP-OES
3182	Yes	0.2 #20650: 0.5- 2.0 #20651: 0.5-0.8	Further cut	Microwave	Nitric acid and Hydrogen peroxide	ICP-OES	
3183	Yes	0.2 #20650: 0.5- 2.0 #20651: 0.5-0.8	Used as received	Acid Digestion	HNO3 and H2O2	ICP-AES	
3185	Yes	0.1	Further cut	Microwave	Increasing the temp of each sample to 210°C in 20 mins, and hold at 210°C for 10 mins Heat up to 120°C in 6 min, hold 2min; Heat up to 160°C in 5 min, hold 2min; Heat up to 190° in 7 min, hold 30min.	NA	UV-VIS
3190	Yes	0.1	Further cut	Microwave	HNO3 :HCL 3:1	ICP-OES	
3197	Yes	---	Further cut	Microwave	---	ICP-OES	
3210	Yes	0.1	Further cut	Acid Digestion	HNO3+HCl 9 ml HNO3+ 3 ml HCl+ 1 ml HF	AAS	
3214	Yes	0.15	Further grinded	Microwave	210°C, Microwave 1 hr Digestión ácida en microondas analítico con 10 ml de HNO3 concentrado ultrapuro (70%) La muestra obtenida se enrasa a 25 ml y se diluye 1:10 con solución acuosa 2%	ICP-OES	
3216	No	0.2	Used as received	Acid Digestion acid digestion for EN1122 and microwave digestion for CPSC-CH-E1002-08.3	Programa de digestión: rampa de 15 min para alcanzar 190°C y se mantiene a 190°C 15 minutos más.	HNO3 para su análisis por ICPMS (Modelo Agilent 7500).	ICP-MS
3225	Yes	EN1122: 0.5 CPSC-CH- E1002-08.3:	0.1	Further cut	120°C for 5 mins, 160°C for 1 min, 190°C for 4 min Power Time (mins) Temp (°C) Step 1 1000W 3:30 120 Step 2 1000W 15:00 180 Step 3 1000W 15:00 200	HNO3	ICP-OES
3228	Yes	0.1	Further cut	Microwave	Stage 1:temp 160°C ,Ramp time 15 min, hold time 5 min, 1800 W Stage 2:temp 220°C ,Ramp time 10 min, hold time 10 min, 1800 W	ICP-OES	
3237	Yes	0.2 0.2g heavy metals ,Cd,Pb,Hg,Cr using the ICP-OES. 1.25g for Cr6+ using UV-Vis	Further cut	Microwave	Sample was digested for 40 min at a temp from 120-200°C. Alkaline digestion solution (NaOH and Na2CO3) was used for the analysis of Cr6+. Sample extracted at 90°C for 180 min - Increase temperature to 185°C in 25 min. - Keep in 25 min. - Cool down to RT	9 ml HNO3+3 ml HCl	ICP-MS
3239	Yes	UV-Vis	Further grinded	Microwave	- Increase temperature to 185°C in 25 min. - Keep in 25 min. - Cool down to RT 130°C for 1 min, then 170°C for 1 min, then 190°C 15 min	5 ml of nitric acid and 0.1 ul H2O2.	ICP-OES
3246	No	0.1 to 0.2	Used as received	Microwave	Nitric acid 65% H2SO4 and H2O2	ICP-MS	
3248	Yes	0.15 for microwave	Further cut	Microwave		AAS	

lab	ISO 17025 accredited	sample intake (g)	sample preparation	release method	temperature program	acid mixture	analyze technique
		digestion 0.2 for EN1122 Different spots were measured					
6146	Yes	0.2	Used as received	Other	Not performed	Not performed	XRF
8005	Yes	0.1	Further cut	Microwave	--	--	ICP-OES
8008	Yes	0.5	Further cut	Acid Digestion	--	--	AAS

APPENDIX 4**Number of participants per country**

6 labs in BANGLADESH
1 lab in BELGIUM
1 lab in BRAZIL
1 lab in BULGARIA
1 lab in CANADA
3 labs in DENMARK
1 lab in ESTONIA
1 lab in FINLAND
6 labs in FRANCE
10 labs in GERMANY
2 labs in GUATEMALA
21 labs in HONG KONG
6 labs in INDIA
3 labs in INDONESIA
7 labs in ITALY
2 labs in JAPAN
3 labs in MALAYSIA
1 lab in MAURITIUS
4 labs in MEXICO
1 lab in MOROCCO
26 labs in P.R. of CHINA
2 labs in PAKISTAN
1 lab in PERU
1 lab in POLAND
2 labs in PORTUGAL
1 lab in ROMANIA
3 labs in SINGAPORE
9 labs in SOUTH KOREA
5 labs in SPAIN
1 lab in SRI LANKA
1 lab in SWITZERLAND
7 labs in TAIWAN
7 labs in THAILAND
2 labs in THE NETHERLANDS
2 labs in TUNISIA
4 labs in TURKEY
7 labs in U.S.A.
3 labs in UNITED KINGDOM
6 labs in VIETNAM

APPENDIX 5**Abbreviations**

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported

Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 EN1122:01, "Determination of Cadmium in plastics with the method of the wet decomposition".
- 3 ASTM D4004:98, "Determination of Metal Content by Flame Atomic Absorption (AAS) analysis"
- 4 ASTM E178:02
- 5 ASTM E1301:03
- 6 ISO5725:86
- 7 ISO5725, parts 1-6, 1994
- 8 ISO13528:05
- 9 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 10 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 11 Horwitz, W and Albert, R, Journal of AOAC International, 79-3, 589 (1996)
- 12 IP367:96
- 13 DIN38402 T41/42
- 14 W.J. Conover. Practical; Nonparametric Statistics. J. Wiley&Sons. NY., 302. (1971)
- 15 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 16 J.N. Miller, Analyst, 118, 455, (1993)
- 17 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 18 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364 (2002)
- 19 R.G. Visser, Reliability of proficiency test results for metals and phthalates in plastics, Accred Qual Assur, 14, 29-34 (2009)
- 20 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier procedure, Technometrics, 25(2), 165-172, (1983)