

Results of Proficiency Test  
Hydraulic Oil (used)  
November 2019

Organised by: Institute for Interlaboratory Studies  
Spijkenisse, the Netherlands

Author: ing. M. Meijer  
Correctors: ing. A.S. Noordman-de Neef & ing. R.J. Starink  
Report: iis19L10

January 2020

-- Empty page --

**CONTENTS**

1	INTRODUCTION .....	4
2	SET UP .....	4
2.1	ACCREDITATION .....	4
2.2	PROTOCOL.....	4
2.3	CONFIDENTIALITY STATEMENT.....	5
2.4	SAMPLES.....	5
2.5	STABILITY OF THE SAMPLES .....	6
2.6	ANALYZES.....	7
3	RESULTS .....	7
3.1	STATISTICS.....	8
3.2	GRAPHICS.....	8
3.3	Z-SCORES.....	9
4	EVALUATION.....	9
4.1	EVALUATION PER SAMPLE AND PER TEST .....	10
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES .....	14
4.3	COMPARISON OF PROFICIENCY TEST OF NOVEMBER 2019 WITH PREVIOUS PTS .....	16

## Appendices:

1.	Data, statistical and graphic results.....	18
2.	Number of participants per country .....	63
3.	Abbreviations and literature.....	64

## 1 INTRODUCTION

Since 2003 the Institute for Interlaboratory Studies (iis) organizes a proficiency test for the analyzes of used Hydraulic Oil every year. During the annual program 2019/2020 it was decided to continue the round robin for the analyzes of used Hydraulic Oil. This interlaboratory study contains also a PT for the determination of Metals in used Hydraulic Oil.

In the PT for the regular analyzes 62 laboratories in 41 different countries registered for participation. In the PT on Metals 50 laboratories in 35 different countries registered for participation. In total 64 laboratories in 43 different countries registered for at least one of the two PTs. See appendix 2 for the number of participants per country.

In this report, the results of the 2019 used Hydraulic Oil proficiency test are presented and discussed. This report is also available as PDF file from the iis website [www.iisnl.com](http://www.iisnl.com).

## 2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory. In this proficiency test the participants received depending on the registration: one liter glass bottle with used Hydraulic Oil labelled #19236 for the round with regular analyzes and/or a 50 mL PE bottle with used Hydraulic Oil labelled #19237 for the wear Metals round. Participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

### 2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

### 2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website [www.iisnl.com](http://www.iisnl.com), from the FAQ page.

## 2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

## 2.4 SAMPLES

### Preparation of samples for the PT on regular analyzes in used Hydraulic Oil, iis19L10

Approximately 90 liter of used Hydraulic Oil was obtained from a local supplier. After homogenisation 84 amber glass bottles of one liter were filled and labelled #19236. The homogeneity of the subsamples #19236 was checked by determination of Density at 15°C in accordance with ASTM D4052 and Kinematic Viscosity at 40°C according to ASTM D445 on 8 stratified randomly selected samples.

	Density at 15°C in kg/L	Kinematic Viscosity at 40°C in mm <sup>2</sup> /s
Sample #19236-1	0.86921	68.05
Sample #19236-2	0.86921	68.14
Sample #19236-3	0.86922	68.11
Sample #19236-4	0.86920	68.09
Sample #19236-5	0.86921	68.09
Sample #19236-6	0.86920	68.03
Sample #19236-7	0.86921	68.12
Sample #19236-8	0.86920	68.10

Table 1: homogeneity test results of subsamples #19236

From the above test results the repeatabilities were calculated and compared with 0.3 times the reproducibilities of the reference test methods in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15°C in kg/L	Kinematic Viscosity at 40°C in mm <sup>2</sup> /s
r (observed)	0.00002	0.101
reference test method	ISO12185:96	iis memo 1401
0.3 x R (ref. test method)	0.00015	0.368

Table 2: evaluation of the repeatabilities of subsamples #19236

The calculated repeatabilities were lower than 0.3 times the reproducibilities of the reference test methods. Therefore, homogeneity of the subsamples was assumed.

Preparation of samples for PT on Metals in used Hydraulic Oil, iis19L10M

Approximately 10 liter of used Hydraulic Oil was obtained from a local supplier who also did also add different elements. After homogenisation 70 PE bottles of 50 mL were filled and labelled #19237. The homogeneity of the subsamples #19237 was checked by the determination of Copper and Nickel in accordance with ASTM D5185 on 8 stratified randomly selected samples.

	Copper as Cu in mg/kg	Nickel as Ni in mg/kg
Sample #19237-1	15	12
Sample #19237-2	15	12
Sample #19237-3	15	12
Sample #19237-4	15	12
Sample #19237-5	15	12
Sample #19237-6	15	13
Sample #19237-7	15	12
Sample #19237-8	16	13

Table 3: homogeneity test results of subsamples #19237

From the test results of table 3 the repeatabilities were calculated and compared with 0.3 times the reproducibilities of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Copper as Cu in mg/kg	Nickel as Ni in mg/kg
r (observed)	0.99	1.30
reference test method	D5185:18	D5185:18
0.3 x R (ref. test method)	1.09	1.58

Table 4: evaluation of the repeatabilities of subsamples #19237

The calculated repeatabilities were in agreement with 0.3 times the corresponding reproducibilities of the reference test method. Therefore, homogeneity of the subsamples was assumed.

Depending on the registration to each of the participating laboratories one 1 liter bottle labelled #19236 and/or one 50 mL PE bottle labelled #19237 was dispatched on October 16, 2019. An SDS was added to the sample package.

**2.5 STABILITY OF THE SAMPLES**

The stability of used Hydraulic Oil packed in amber glass and PE bottles was checked. The material was found sufficiently stable for the period of the proficiency test.

## 2.6 ANALYZES

The participants were requested to determine on sample #19236: Acid Number (Total), Density at 15°C, Flash Point PMcc, Kinematic Viscosity at 40 and 100°C, Viscosity Stabinger at 40 and 100°C, Sulfur, Water and Level of Contamination (counts/mL and scale number). And on sample #19237 23 elements, wear metals and additives: Al, Ba, B, Cd, Cr, Cu, Fe, Pb, Li, Mg, Mn, Mo, Ni, K, Si, Ag, Na, Sn, Ti, V, Ca, P and Zn. Also some additional questions were asked about Acid Number (Total) and Level of Contamination.

It was explicitly requested to treat the samples as if they were routine samples and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations. Also, some analytical details were asked.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the appropriate reference test methods that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website [www.iisnl.com](http://www.iisnl.com).

## 3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the test result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

### 3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...>' or '>...>' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

According to ISO5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

### 3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also a normal Gauss curve was projected over the Kernel Density Graph for reference.



### 3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation of this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used. In some cases, a reproducibility based on former iis proficiency tests could be used. When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$Z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The  $Z_{(\text{target})}$  scores are listed in the test result tables in appendix 1.

Absolute values for  $z < 2$  are very common and absolute values for  $z > 3$  are very rare. The usual interpretation of z-scores is as follows:

- $|z| < 1$  good
- $1 < |z| < 2$  satisfactory
- $2 < |z| < 3$  questionable
- $3 < |z|$  unsatisfactory

## 4 EVALUATION

In this proficiency test no major problems were encountered with sample dispatch. For the round with the regular analyzes, one participant reported the test results after the final reporting date and one other participant did not report any test results at all.

For the Metals round one participant reported the test results after the final reporting date and five other participants did not report any test results at all. Not all participants were able to report all analyzes requested.

In total, 63 participants reported in total 1402 numerical test results. Observed were 87 outlying test results, which is 6.2%. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as "not OK" or "suspect". The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

#### 4.1 EVALUATION PER SAMPLE AND PER TEST

In this section the results are discussed per sample and per test. The test methods, which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the original data. The abbreviations, used in these tables, are explained in appendix 3.

In the iis PT reports the ASTM test methods are referred to with a number e.g. ASTM D7647 and an added designation for the year that the test method was adopted or revised e.g. ASTM D7647:10. If applicable a designation in parentheses is added to designate the year of reapproval e.g. ASTM D7647:10(2018). In the tables of appendix 1 only the test method number and year of adoption or revision will be used.

Unfortunately, a suitable reference test method providing the precision data is not available for all determinations. For the tests that have no available precision data the calculated reproducibility was compared against the reproducibility estimated from the Horwitz equation.

##### **Sample #19236**

Acid Number (Total): This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D664-A:18e2 using Buffer End Point 60 mL. However, the calculated reproducibility is not in agreement with the precision data of the procedures Buffer End Point 125 mL and Inflection Point 60 and 125 mL.

Density at 15°C: This determination was problematic for some laboratories. Eight statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.

Flash Point PMcc: This determination may not be problematic depending on the test method used. One statistical outlier was observed and one other test result was excluded. The calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ASTM D93:19 procedure A, but not with procedure B. When the test results of procedures A and B are evaluated separately both calculated reproducibilities are in full agreement with the requirements of ASTM D93:19 procedures A or B respectively.

Kin.Viscosity at 40°C: The precision statement given in ASTM D445 for used (in-service) formulated oils appears to be very strict, see previous iis reports iis17L08 and iis18L10. Therefore, the target reproducibility being used is calculated from the reproducibilities found in previous iis PTs on used oils (see lit. 17). This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the average reproducibility found for used oils in previous iis PTs (iis memo 1401), and in this case also in (full) agreement with the requirements of ASTM D445:19.

Kin.Viscosity at 100°C: See the explanation about selection of the target reproducibility at Kinematic Viscosity at 40°C.

This determination was not problematic. Five statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the average reproducibility found for used oils in previous iis PTs (iis memo 1401), but not in agreement with the requirements of ASTM D445:19.

Viscosity Stabinger at 40°C: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D7042:16e3.

Viscosity Stabinger at 100°C: This determination was not problematic. No statistical outliers were observed but three test results were excluded. The calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D7042:16e3.

Sulfur: This determination was problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the requirements of ASTM D4294:16e1.

Water: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D6304:16e1.

Level of Contamination: When the data of the last four round robins are evaluated no significant differences between the test results from ISO11500 and the test results of ASTM D7647 could be found. The reproducibility of ASTM D7647 is used for the calculation of the z-scores because ISO11500 does not mention a reproducibility but only 'Maximum allowable difference' in Annex A. For calibration of equipment all participants that responded used ISO11171.

This determination was problematic. In total nineteen statistical outliers were observed over six parameters (15 for counts/mL and 4 for scale number) and another twenty-six test results were excluded. The calculated reproducibilities after rejection of the suspect data are not in agreement with the requirements of ASTM D7647:10(2018).

The reference method for the analyzes of wear metals is test method ASTM D5185:18. For some metals i.e. Ba and K the consensus value of the group is either above or below the respective application ranges on which the requirements of ASTM D5185:18 are based. However, it was decided to use the reproducibility from ASTM D5185:18 as the calculated reproducibility (after the rejection of the statistical outliers) is in general in good agreement with the requirements of ASTM D5185:18. More discussion per metal is given below.

### **Sample #19237**

- Aluminum: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.
- Barium: This determination was not problematic. Two statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.
- Boron: This determination was not problematic. Three statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.
- Cadmium: This determination may be problematic. No statistical outliers were observed. However, the calculated reproducibility is not in agreement with the estimated reproducibility using the Horwitz equation.
- Chromium: This determination was not problematic. One statistical outlier was observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ASTM D5185:18.
- Copper: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.
- Iron: This determination was not problematic. Two statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.
- Lead: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.

- Lithium: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the estimated reproducibility using the Horwitz equation.
- Magnesium: This determination was not problematic. Two statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.
- Manganese: This determination was not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D5185:18.
- Molybdenum: This determination was not problematic. No statistical outliers were observed but one test result was excluded. The calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.
- Nickel: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.
- Potassium: Twenty-four participants agreed on a level of <1 mg/kg. Therefore, no z-scores were calculated.
- Silicon: This determination was not problematic. Four statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.
- Silver: This determination was not problematic. Three statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.
- Sodium: This determination was not problematic. Three statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ASTM D5185:18.
- Tin: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.

Titanium: This determination was not problematic. Two statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.

Vanadium: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D5185:18.

Calcium: This determination was not problematic. Two statistical outliers were observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the estimated reproducibility using the Horwitz equation, but not with the strict requirements of ASTM D5185:18.

Phosphorus: This determination was not problematic. One statistical outlier was observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in agreement with the requirements of ASTM D5185:18.

Zinc: This determination was not problematic. One statistical outlier was observed and one other test result was excluded. However, the calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ASTM D5185:18.

#### 4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average result, the calculated reproducibility ( $2.8 \cdot$  standard deviation) and the reproducibility derived from literature reference test methods (in casu ASTM and ISO reference test methods) are presented in the next tables.

Parameter	unit	n	average	2.8 * sd	R(lit)
Acid Number (Total)	mg KOH/g	44	0.38	0.20	0.22
Density at 15°C	kg/L	41	0.8692	0.0003	0.0005
Flash Point PMcc	°C	37	213.7	14.5	15.2
Kinematic Viscosity at 40°C	mm <sup>2</sup> /s	44	68.314	0.819	1.230
Kinematic Viscosity at 100°C	mm <sup>2</sup> /s	39	10.478	0.142	0.231
Viscosity Stabinger at 40°C	mm <sup>2</sup> /s	15	68.376	0.480	0.859
Viscosity Stabinger at 100°C	mm <sup>2</sup> /s	16	10.478	0.095	0.116
Sulfur	mg/kg	25	695	179	130
Water	mg/kg	47	64.8	73.6	206.4
L. of Contamination $\geq 4 \mu\text{m}$ (c)	counts/mL	23	221	392	252

Parameter	unit	n	average	2.8 * sd	R(lit)
L. of Contamination $\geq 6 \mu\text{m}$ (c)	counts/mL	23	61.4	150.9	47.1
L. of Contamination $\geq 14 \mu\text{m}$ (c)	counts/mL	23	6.8	17.0	9.3
L. of Contamination $\geq 4 \mu\text{m}$ (c)	scale no.	22	14.6	2.7	1.7
L. of Contamination $\geq 6 \mu\text{m}$ (c)	scale no.	22	12.5	3.0	1.2
L. of Contamination $\geq 14 \mu\text{m}$ (c)	scale no.	22	9.2	3.7	2.0

Table 5: performance evaluation sample #19236

Element	unit	n	average	2.8 * sd	R(lit)
Aluminum as Al	mg/kg	41	12.3	3.8	7.3
Barium as Ba	mg/kg	36	25.2	4.0	11.5
Boron as B	mg/kg	31	11.8	5.4	13.3
Cadmium as Cd	mg/kg	28	11.3	4.1	3.5
Chromium as Cr	mg/kg	42	12.0	3.6	3.7
Copper as Cu	mg/kg	40	14.0	2.6	3.4
Iron as Fe	mg/kg	41	15.0	3.2	4.5
Lead as Pb	mg/kg	37	12.1	3.5	7.4
Lithium as Li	mg/kg	12	54.3	4.5	13.3
Magnesium as Mg	mg/kg	39	23.3	5.6	8.1
Manganese as Mn	mg/kg	39	11.3	2.6	2.4
Molybdenum as Mo	mg/kg	38	11.6	2.7	3.6
Nickel as Ni	mg/kg	43	11.7	2.8	5.1
Potassium as K	mg/kg	21	0.6	2.1	(5.7)
Silicon as Si	mg/kg	38	11.8	2.5	7.6
Silver as Ag	mg/kg	35	11.7	3.2	4.1
Sodium as Na	mg/kg	36	12.1	6.1	6.5
Tin as Sn	mg/kg	37	10.5	4.0	9.0
Titanium as Ti	mg/kg	40	11.8	3.2	8.0
Vanadium as V	mg/kg	42	11.6	2.8	4.2
Calcium as Ca	mg/kg	36	77.5	14.2	18.0
Phosphorus as P	mg/kg	40	266	48	70
Zinc as Zn	mg/kg	39	309	48	45

Table 6: : performance evaluation sample #19237

Results between brackets should be used with due care because of low concentration of element

Without further statistical calculations, it can be concluded that for many tests there is a good compliance of the group of participants with the relevant reference test methods. The problematic tests have been discussed in paragraph 4.1.

### 4.3 COMPARISON OF THE PROFICIENCY TEST OF NOVEMBER 2019 WITH PREVIOUS PTS

	November 2019	November 2018	November 2017	November 2016	November 2015
Number of reporting laboratories	63	52	57	51	55
Number of test results	1402	1053	1202	1168	1126
Number of statistical outliers	87	49	89	29	43
Percentage outliers	6.2%	4.7%	7.4%	2.5%	3.8%

Table 7: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the reference test methods. The conclusions are given in the following table.

Determination	November 2019	November 2018	November 2017	November 2016	November 2015
Acid Number (Total)	+/-	+/-	+	+	+
Density at 15°C	+	+/-	+/-	+/-	+/-
Flash Point PMcc	+/-	+/-	+/-	-	+
Kinematic Viscosity at 40°C	+	+	+	+	++
Kinematic Viscosity at 100°C	+	+	+	+/-	+/-
Viscosity Stabinger at 40°C	+	+/-	+	+/-	++
Viscosity Stabinger at 100°C	+	-	-	+/-	+
Sulfur	-	--	-	-	--
Water	++	++	++	++	++
L. of Contamination – counts/mL	-	-	--	-	--
L. of Contamination – scale no.	-	-	-	-	n.e.
Aluminum as Al	+	+	++	++	++
Barium as Ba	++	++	+	++	+
Boron as B	++	++	++	n.e.	n.e.
Cadmium as Cd	-	n.e.	+	n.e.	n.e.
Chromium as Cr	+/-	n.e.	+/-	+	+
Copper as Cu	+	+	+	+/-	+
Iron as Fe	+	+/-	+	+	++
Lead as Pb	++	++	++	++	++
Lithium as Li	++	n.e.	+	(--)	-
Magnesium as Mg	+	+	+	+	+/-
Manganese as Mn	+/-	+/-	--	+/-	+/-
Molybdenum as Mo	+	++	+/-	+	+/-
Nickel as Ni	+	+/-	++	++	++
Potassium as K	(+)	++	+	n.e.	n.e.
Silicon as Si	++	+	+	++	++



Determination	November 2019	November 2018	November 2017	November 2016	November 2015
Silver as Ag	+	n.e.	+	+/-	+
Sodium as Na	+/-	+/-	-	+/-	+/-
Tin as Sn	++	+	++	++	++
Titanium as Ti	++	n.e.	++	++	++
Vanadium as V	+	n.e.	+	+	++
Calcium as Ca *)	+	--	--	-	-
Phosphorus as P	+	-	-	+	+/-
Zinc as Zn	+/-	-	+/-	-	--

Table 8: comparison determinations against the reference test methods

\*) Based on Horwitz for 2019 and ASTM D5185 for earlier PTs

In the table above the following performance categories were used:



- ++: group performed much better than the reference test method
- + : group performed better than the reference test method
- +/-: group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e.: not evaluated

**APPENDIX 1**

Determination of Acid Number (Total) on sample #19236; results in mg KOH/g



lab	method	value	mark	z(targ)	end point	volume (mL)	remarks
178	D664-A	0.44		0.77	----	60	
179	D664-A	0.42		0.51	IP	125	
225	D974	0.33		-0.66	----	----	
237		----		----	----	----	
255		----		----	----	----	
257		----		----	----	----	
311		----		----	----	----	
317	D974	0.35		-0.40	IP	----	
325	D664-A	0.325		-0.72	BEP (pH 10)	125	
331	D664Mod.	0.47		1.16	IP	60	
335		----		----	----	----	
339		----		----	----	----	
349	D664-A	0.37		-0.14	IP	125	
360	D664-A	0.455		0.96	IP	60	
398		----		----	----	----	
432		----		----	----	----	
442		----		----	----	----	
451	D664-A	0.41		0.38	BEP (pH 10)	60	
496	D664-A	0.32		-0.79	BEP (pH 10)	60	
526	D974	0.35		-0.40	----	----	
541	D974	0.308		-0.94	----	----	
562	D664	0.52		1.81	----	----	
575	D664-A	0.37		-0.14	BEP (pH 10)	60	
603		----		----	----	----	
614		----		----	----	----	
621	D664-A	0.326		-0.71	IP	60	
633	D664-A	0.63	DG(0.05)	3.23	IP	125	
634	D664-A	0.495		1.48	----	----	
780	D664-A	0.63	DG(0.05)	3.23	IP	60	
862	D664-A	0.32		-0.79	----	----	
863	D664-A	0.36		-0.27	----	----	
912	D664-A	0.5		1.55	----	----	
962	D974	0.32		-0.79	----	----	
963	D664-A	0.306		-0.97	BEP (pH 11)	60	
975	D664-A	0.313		-0.88	IP	60	
994	D664-A	0.41	C	0.38	IP	125	fr. 0.61
1059	ISO6619	0.41		0.38	BEP (pH 11)	60	
1146	D664-A	0.517		1.77	BEP (pH 11)	125	
1166	D664-A	0.3358		-0.58	IP	60	
1265	D664-A	0.3247		-0.73	IP	60	
1297	D664-A	0.34		-0.53	IP	125	
1397		----		----	----	----	
1435	D664-A	0.29		-1.18	BEP (pH 10)	100	
1505	D974	0.3895		0.11	----	----	
1531		----		----	----	----	
1660	D664-A	0.393		0.16	BEP (pH 11)	60	
1720		----		----	----	----	
1740	D664-A	0.54		2.07	IP	60	
1743	D664-A	0.34		-0.53	BEP (pH 11)	60	
1788		----		----	----	----	
1857	D664-A	0.473		1.20	IP	125	
1875	ISO6618	0.343		-0.49	----	----	
1888	D664-A	0.528		1.91	BEP (pH 11)	60	
1900	D664-A	0.282		-1.28	----	60	
1920	D664-A	0.308		-0.94	IP	125	
2493	ISO6618	0.347		-0.44	IP	125	
6002	D664-A	0.315		-0.85	BEP (pH 10)	60	
6016	D664-A	0.440		0.77	----	----	
6257		----		----	----	----	
6261	D664-A	0.34		-0.53	----	----	
6284	D974	0.346		-0.45	----	----	
7011	D664-A	0.36		-0.27	----	----	

normality	OK	IP only	BEP only
n	44	suspect	suspect
outliers	2	17	12
mean (n)	0.3807	0	0
st.dev. (n)	0.07239	0.4143	0.3770
R(calc.)	0.2027	0.10512	0.07901
st.dev.(D664-A:18e2, BEP 60mL)	0.07713	0.2943	0.2212
R(D664-A:18e2, BEP 60mL)	0.2160	----	0.07640
Compare		----	0.2139
R(D664-A:18e2, IP 60mL)	0.1822	0.1953	----
R(D664-A:18e2, IP 125mL)	0.0799	0.0873	----
R(D664-A:18e2, BEP 125mL)	0.1145	----	0.1133





Determination of Density at 15°C on sample #19236; results in kg/L

lab	method	value	mark	z(targ)	remarks
178	D4052	0.8836	R(0.01)	80.81	
179	D4052	0.8692		0.17	
225		----		----	
237	D4052	0.8697	R(0.01)	2.97	
255		----		----	
257		----		----	
311	D4052	0.8692		0.17	
317	D4052	0.8692		0.17	
325	D4052	0.8692		0.17	
331	ISO12185	0.8683	R(0.01)	-4.87	
335	D1298	0.8692		0.17	
339		----		----	
349	D4052	0.8691		-0.39	
360	D4052	0.8692		0.17	
398	ISO12185	0.8692	C	0.17	reported 869.2 kg/L
432	D4052	0.86926		0.51	
442		----		----	
451	D4052	0.8688		-2.07	
496	ISO12185	0.86928		0.62	
526	D4052	0.8692		0.17	
541	D4052	0.86920		0.17	
562		----		----	
575		----		----	
603	D4052	0.8692		0.17	
614	D4052	0.8692		0.17	
621	D4052	0.8692		0.17	
633	D4052	0.8686	R(0.01)	-3.19	
634	D4052	0.8692		0.17	
780	ISO12185	0.8691		-0.39	
862	D4052	0.8691		-0.39	
863		----		----	
912	ISO12185	0.8692		0.17	
962	D4052	0.8692		0.17	
963	D4052	0.8692		0.17	
975	D4052	0.8692		0.17	
994	ISO12185	0.8692		0.17	
1059	ISO12185	0.8692		0.17	
1146	D4052	0.8692		0.17	
1166		----		----	
1265	D4052	0.8691		-0.39	
1297	D4052	0.86917		0.01	
1397	ISO12185	0.8692		0.17	
1435	D7042	0.8686	R(0.01)	-3.19	
1505	D7042	0.8667	R(0.01)	-13.83	
1531		----		----	
1660	D7042	0.8679	R(0.01)	-7.11	
1720		----		----	
1740	D7042	0.8690		-0.95	
1743	ISO12185	0.8690		-0.95	
1788	D4052	0.869145	C	-0.13	first reported 0.869145 kg/m <sup>3</sup>
1857	ISO12185	0.8695		1.85	
1875	D7042	0.8692		0.17	
1888		----		----	
1900	D4052	0.8692	C	0.17	first reported 0.8692 kg/m <sup>3</sup>
1920		----		----	
2493	ISO12185	0.86917		0.01	
6002	ISO3675	0.8690		-0.95	
6016	D4052	0.8692		0.17	
6257	ISO12185	0.86899		-1.00	
6261	D4052	0.8699	R(0.01)	4.09	
6284	D4052	0.869312		0.80	
7011	D1298	0.8691		-0.39	
	normality	not OK			
	n	41			
	outliers	8			
	mean (n)	0.86917			
	st.dev. (n)	0.000106			
	R(calc.)	0.00030			
	st.dev.(ISO12185:96)	0.000179			
	R(ISO12185:96)	0.00050			



Determination of Flash Point PMcc on sample #19236; results in °C



lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D93-A	211		-0.49	
225	D93-B	218.0		0.80	
237	D93-B	210		-0.67	
255		----		----	
257		----		----	
311		----		----	
317		----		----	
325	D93-A	217.0		0.62	
331	D93-A	209.4		-0.78	
335		----		----	
339	D93-B	>190		----	
349	D93-A	214		0.06	
360	D93-A	209.0		-0.86	
398		----		----	
432	D93-A	215.5		0.34	
442		----		----	
451	D93-A	214		0.06	
496		----		----	
526	D93-B	212		-0.30	
541	D93-A	210.50		-0.58	
562		----		----	
575		----		----	
603	D93-A	214		0.06	
614		----		----	
621	D93-A	214.0		0.06	
633		----		----	
634	D93-B	214.0		0.06	
780	D93-A	208.0		-1.04	
862	D93-A	200		-2.52	
863		----		----	
912		----		----	
962	D93-A	210		-0.67	
963	D93-A	210.0		-0.67	
975	D93-A	211.0		-0.49	
994	D93-A	209.5		-0.77	
1059	ISO2719-A	214.5		0.16	
1146	D93-A	212.5		-0.21	
1166		----		----	
1265	D93	210		-0.67	
1297	D93-B	218.5		0.90	
1397		----		----	
1435	D93-A	221.0		1.36	
1505	D93-A	227.5		2.56	
1531	D93-B	160	R(0.01)	-9.90	
1660		----		----	
1720		----		----	
1740		----		----	
1743	ISO2719-A	216		0.43	
1788	D93-B	209.0		-0.86	
1857	D93-B	208.5		-0.95	
1875	ISO2719-B	214.5		0.16	
1888	ISO2719-A	222		1.54	
1900	D7094	216.7		0.56	
1920		----		----	
2493	D93-A	217.2		0.66	
6002	ISO2719-A	223.0		1.73	
6016	D93-A	219.2		1.02	
6257	ISO2719-A	208.0		-1.04	
6261	D93-A	216.0		0.43	
6284	D92	211	ex,C	-0.49	first reported 231, D93-A
7011		----		----	
			<u>D93-A &amp; ISO 2719-A only</u>	<u>D93-B &amp; ISO 2719-B only</u>	
	normality	suspect	OK	OK	
	n	37	27	8	
	outliers	1 +1ex	0	1	
	mean (n)	213.65	213.84	213.06	
	st.dev. (n)	5.170	5.654	3.868	
	R(calc.)	14.48	15.83	10.83	
	st.dev.(D93-A:19)	5.418	5.421	----	
	R(D93-A:19)	15.17	15.18	----	
Compare					
	R(D93-B:19)	10	----	10	



Determination of Kinematic Viscosity at 40°C on sample #19236; results in mm<sup>2</sup>/s



lab	method	value	mark	z(targ)	remarks
178	D445	68.3		-0.03	
179	D445	68.31		-0.01	
225	D445	68.87		1.27	
237	D445	68.18		-0.30	
255	D7279 corrected to D445	67.66		-1.49	
257	D7279 corrected to D445	67.62		-1.58	
311	D445	68.09		-0.51	
317	D445	68.74		0.97	
325	D445	68.245		-0.16	
331	D7279Mod.	66.39	R(0.01)	-4.38	
335		----		----	
339		----		----	
349	D445	68.22		-0.21	
360	D445	68.383		0.16	
398		----		----	
432		----		----	
442		----		----	
451	D7279 corrected to D445	68.2	C	-0.26	first reported as Kinematic Viscosity at 100°C
496	D445	68.244		-0.16	
526	D445	68.27		-0.10	
541	D445	68.119		-0.44	
562	D7279	67.9825		-0.75	
575	D7279 corrected to D445	68.1		-0.49	
603		----		----	
614	D445	68.27		-0.10	
621	D445	68.24		-0.17	
633	D7279 corrected to D445	68.158		-0.35	
634	D7279 corrected to D445	68.11	C	-0.46	first reported 66.495
780	D445	68.26		-0.12	
862	D445	68.291		-0.05	
863	D7279	68.34		0.06	
912	D445	68.50		0.42	
962		----		----	
963	D445	68.63		0.72	
975	D445	68.51		0.45	
994	D445	68.24		-0.17	
1059	ISO3104	67.89		-0.96	
1146	D445	68.263		-0.12	
1166	ISO3104	68.8436		1.21	
1265	D7279 corrected to D445	68.11		-0.46	
1297		----		----	
1397		----		----	
1435		----		----	
1505	D445	68.441		0.29	
1531		----		----	
1660		----		----	
1720		----		----	
1740	D445	68.40		0.20	
1743	D7279 corrected to D445	68.36		0.11	
1788	D445	68.4707		0.36	
1857	D445	68.912		1.36	
1875		----		----	
1888	D445	66.40	C,R(0.01)	-4.36	first reported 70.26
1900	D445	68.876		1.28	
1920	D445	68.149		-0.37	
2493	D445	68.92		1.38	
6002	ISO3104	68.37		0.13	
6016		----		----	
6257	ISO3104	68.1711		-0.32	
6261		----	W	----	test result withdrawn, reported 66.44
6284	D445	68.46	C	0.33	first reported 63.859887
7011	D445	68.080		-0.53	
	normality	OK			
	n	44			
	outliers	2			
	mean (n)	68.3136			
	st.dev. (n)	0.29247			
	R(calc.)	0.8189			
	st.dev.(iis memo 1401)	0.43916			
	R(iis memo 1401)	1.2296			see appendix 3, ref. 17
Compare					
	R(D445:19)	0.8567			







Determination of Kinematic Viscosity at 100°C on sample #19236; results in mm<sup>2</sup>/s

lab	method	value	mark	z(targ)	remarks
178	D445	10.50		0.27	
179	D445	10.54		0.75	
225	D445	10.60		1.48	
237	D445	10.49		0.15	
255	D7279 corrected to D445	10.47		-0.10	
257		----		----	
311	D445	10.46		-0.22	
317		----		----	
325	D445	10.485		0.09	
331	D7279Mod.	10.57		1.12	
335		----		----	
339		----		----	
349	D445	10.50		0.27	
360	D445	10.501		0.28	
398		----		----	
432	D445	10.48	C	0.03	first reported as Kinematic Viscosity at 40°C
442		----		----	
451	D7279 corrected to D445	10.45	C	-0.34	first reported as Kinematic Viscosity at 40°C
496	D445	10.472		-0.07	
526	D445	10.43		-0.58	
541	D445	10.447		-0.38	
562	D7279	10.48		0.03	
575	D7279 corrected to D445	10.46		-0.22	
603		----		----	
614	D445	10.53		0.63	
621	D445	10.53		0.63	
633	D7279 corrected to D445	10.282	R(0.05)	-2.38	
634	D7279 corrected to D445	10.404		-0.90	
780	D445	10.43		-0.58	
862	D445	10.496		0.22	
863	D7279	10.49		0.15	
912	D445	10.39		-1.07	
962		----		----	
963	D445	10.48		0.03	
975	D445	10.47		-0.10	
994	D445	10.48		0.03	
1059	ISO3104	10.45		-0.34	
1146	D445	10.486		0.10	
1166		----		----	
1265	D7279 corrected to D445	10.34		-1.67	
1297		----		----	
1397		----		----	
1435		----		----	
1505	D445	10.422		-0.68	
1531		----		----	
1660		----		----	
1720		----		----	
1740	D445	10.48		0.03	
1743	D7279 corrected to D445	10.47		-0.10	
1788	D445	10.4444		-0.41	
1857	D445	10.422		-0.68	
1875		----		----	
1888	D445	10.75	C,R(0.05)	3.31	first reported 11.18
1900	D445	10.29	R(0.05)	-2.28	
1920		----		----	
2493	D445	10.96	C,R(0.01)	5.86	first reported 11.02
6002	ISO3104	10.560		1.00	
6016		----		----	
6257	ISO3104	10.5584		0.98	
6261	D445	10.46	C	-0.22	first reported 10.21
6284	D445	10.24	C,R(0.05)	-2.89	first reported 9.9740256
7011	D445	10.510		0.39	
	normality	OK			
	n	39			
	outliers	5			
	mean (n)	10.4779			
	st.dev. (n)	0.05054			
	R(calc.)	0.1415			
	st.dev.(iis memo 1401)	0.08233			
	R(iis memo 1401)	0.2305			see appendix 3, ref. 17
Compare					
	R(D445:19)	0.1046			





Determination of Viscosity Stabinger at 40°C on sample #19236; results in mm<sup>2</sup>/s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D7042	68.59		0.70	
225		----		----	
237		----		----	
255		----		----	
257		----		----	
311		----		----	
317		----		----	
325		----		----	
331		----		----	
335		----		----	
339		----		----	
349		----		----	
360	D7042	68.387		0.04	
398		----		----	
432		----		----	
442		----		----	
451		----		----	
496	D7042	68.480		0.34	
526		----		----	
541		----		----	
562		----		----	
575		----		----	
603	D7042	67.736	DG(0.05)	-2.09	
614		----		----	
621		----		----	
633		----		----	
634		----		----	
780	D7042	67.76	DG(0.05)	-2.01	
862		----		----	
863		----		----	
912		----		----	
962		----		----	
963	D7042	68.42		0.14	
975	D7042	68.43		0.18	
994	D7042	68.23		-0.48	
1059	D7042	68.42		0.14	
1146		----		----	
1166		----		----	
1265		----		----	
1297	D7042	68.047		-1.07	
1397	D7042	67.23	C,G(0.05)	-3.74	first reported 67.33
1435	D7042	68.601		0.73	
1505	D7042	68.269		-0.35	
1531		----		----	
1660	D7042	68.096		-0.91	
1720		----		----	
1740	D7042	68.45		0.24	
1743		----		----	
1788		----		----	
1857		----		----	
1875	D7042	68.319		-0.19	
1888		----		----	
1900		----		----	
1920		----		----	
2493		----		----	
6002		68.28		-0.31	
6016	D7042	68.620		0.80	
6257		----		----	
6261		----	W	----	test result withdrawn, reported 66.38
6284		----		----	
7011		----		----	
normality		OK			
n		15			
outliers		3			
mean (n)		68.3759			
st.dev. (n)		0.17135			
R(calc.)		0.4798			
st.dev.(D7042:16e3)		0.30660			
R(D7042:16e3)		0.8585			



Determination of Viscosity Stabinger at 100°C on sample #19236; results in mm<sup>2</sup>/s



lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D7042	10.45		-0.68	
225		----		----	
237		----		----	
255		----		----	
257		----		----	
311		----		----	
317		----		----	
325		----		----	
331		----		----	
335		----		----	
339		----		----	
349		----		----	
360	D7042	10.510		0.77	
398		----		----	
432		----		----	
442		----		----	
451		----		----	
496	D7042	10.508		0.73	
526		----		----	
541		----		----	
562		----		----	
575		----		----	
603	D7042	10.385	ex	-2.25	excluded as outlier at Viscosity Stabinger at 40°C
614		----		----	
621		----		----	
633		----		----	
634		----		----	
780	D7042	10.44	ex	-0.92	excluded as outlier at Viscosity Stabinger at 40°C
862		----		----	
863		----		----	
912		----		----	
962		----		----	
963	D7042	10.50		0.53	
975	D7042	10.50		0.53	
994	D7042	10.49		0.29	
1059	D7042	10.46		-0.44	
1146		----		----	
1166		----		----	
1265		----		----	
1297	D7042	10.463		-0.36	
1397	D7042	10.36	ex	-2.85	excluded as outlier at Viscosity Stabinger at 40°C
1435	D7042	10.528		1.21	
1505	D7042	10.476		-0.05	
1531		----		----	
1660	D7042	10.393		-2.06	
1720		----		----	
1740	D7042	10.49		0.29	
1743		----		----	
1788		----		----	
1857		----		----	
1875	D7042	10.502		0.58	
1888		----		----	
1900		----		----	
1920		----		----	
2493		----		----	
6002		10.440		-0.92	
6016	D7042	10.488		0.24	
6257		----		----	
6261	D7042	10.45		-0.68	
6284		----		----	
7011		----		----	
	normality	suspect			
	n	16			
	outliers	0 +3ex			
	mean (n)	10.4780			
	st.dev. (n)	0.03381			
	R(calc.)	0.0947			
	st.dev.(D7042:16e3)	0.04136			
	R(D7042:16e3)	0.1158			



Determination of Sulfur on sample #19236; results in mg/kg



lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D4294	823		2.74	
225		----		----	
237	D4294	755		1.28	
255		----		----	
257		----		----	
311		----		----	
317		----		----	
325	INH-6443	661	C	-0.74	first reported 0.0661mg/kg
331	ISO8754	658.25		-0.80	
335		----		----	
339	INH-050	<1000		----	
349	D2622	611		-1.81	
360	ISO8754	711		0.34	
398	D4294	680		-0.33	
432		----		----	
442		----		----	
451		----		----	
496		----		----	
526		----		----	
541		----		----	
562		----		----	
575		----		----	
603		----		----	
614		----		----	
621	D4294	711		0.34	
633		----		----	
634	D4294	691.5		-0.08	
780	D4294	735		0.85	
862	D2622	684.9		-0.22	
863		----		----	
912	D4294	688		-0.16	
962		----		----	
963		----		----	
975	D4294	680		-0.33	
994		----		----	
1059	ISO14596Mod.	650		-0.97	
1146	D4294	640		-1.19	
1166	In house	617.0		-1.68	
1265		----		----	
1297	D4294	756.15		1.31	
1397		----		----	
1435		----		----	
1505	D5185	559.9		-2.91	
1531		----		----	
1660		----		----	
1720		----		----	
1740	D4294	670		-0.54	
1743		----		----	
1788		----		----	
1857	D4294	739.8		0.96	
1875		783.7		1.90	
1888	D5185	797.7	C	2.20	first reported 1309.7
1900		----		----	
1920		----		----	
2493	ISO10304-1	605.8		-1.92	
6002	D5185	756		1.31	
6016		----		----	
6257	ISO8754	716	C	0.45	first reported 160.55 test result withdrawn, reported 0.27
6261		----	W	----	
6284		----		----	
7011		----		----	
	normality	OK			
	n	25			
	outliers	0			
	mean (n)	695.3			
	st.dev. (n)	64.08			
	R(calc.)	179.4			
	st.dev.(D4294:16e1)	46.54			
	R(D4294:16e1)	130.3			





## Determination of Water on sample #19236; results in mg/kg




lab	method	value	mark	z(targ)	remarks
178	D6304-C	52		-0.17	
179	D6304-C	31		-0.46	
225		----		----	
237	D6304-C	51.2		-0.18	
255		----		----	
257		----		----	
311	D6304-A	70		0.07	
317	D6304-A	54		-0.15	
325	D6304-C	28		-0.50	
331	D6304Mod.	38.33		-0.36	
335	D6304-A	70		0.07	
339		----		----	
349	D6304-A	107		0.57	
360	D6304-A	43.4		-0.29	
398	D6304-C	83.4		0.25	
432		----		----	
442	IP438	62.0		-0.04	
451	D6304-C	204	R(0.01)	1.89	
496	D6304-C	40		-0.34	
526	D6304-A	55.84		-0.12	
541	D6304-A	90.5		0.35	
562		----		----	
575		----		----	
603	D6304-C	93.0		0.38	
614	D6304-C	40		-0.34	
621	D6304-A	58		-0.09	
633	D6304-C	30.7		-0.46	
634	D6304-A	120.5		0.76	
780	D6304-C	259	R(0.01)	2.63	
862	D6304-C	67		0.03	
863	D6304-C	73		0.11	
912	D6304-C	53		-0.16	
962	D6304-A	126		0.83	
963	D6304-C	114		0.67	
975	D6304-A	112		0.64	
994	IP439	205	R(0.01)	1.90	
1059	D6304-C	50		-0.20	
1146	D6304-C	43		-0.30	
1166	D6304-A	53		-0.16	
1265	D6304-C	50.61		-0.19	
1297	D6304-A	49.2		-0.21	
1397	ISO12937	110		0.61	
1435	D6304-A	72.0		0.10	
1505	D6304-A	<10		----	
1531	D6304-A	59.0		-0.08	
1660	EN60814	68		0.04	
1720		----		----	
1740	D6304-C	60		-0.07	
1743	ISO12937	40		-0.34	
1788		----		----	
1857	D6304-A	69.7		0.07	
1875	ISO12937	122		0.78	
1888	EN60814	84.23		0.26	
1900	D6304-C	58		-0.09	
1920	D6304-C	55		-0.13	
2493	ISO12937	40		-0.34	
6002	In house	65.15		0.00	
6016	D6304-A	47		-0.24	
6257		40.2		-0.33	
6261	D6304-A	46.0		-0.26	
6284		----		----	
7011		----		----	
	normality	OK			
	n	47			
	outliers	3			
	mean (n)	64.808			
	st.dev. (n)	26.2735			
	R(calc.)	73.566			
	st.dev.(D6304:16e1)	73.7040			
	R(D6304:16e1)	206.371			



Determination of Level of Contamination on sample #19236; results in counts/mL

lab	method	≥ 4 μm (c)	mark	z(targ)	≥ 6 μm (c)	mark	z(targ)	≥ 14 μm (c)	mark	z(targ)
178	ISO4406	369		1.65	104		2.53	7		0.06
179		----		----	----		----	----		----
225		----		----	----		----	----		----
237		----		----	----		----	----		----
255		----		----	----		----	----		----
257		----		----	----		----	----		----
311		----		----	----		----	----		----
317		----		----	----		----	----		----
325		----		----	----		----	----		----
331	ISO4406Mod.	393.40	ex	1.92	108.8	ex	2.82	8.2	ex	0.42
335		----		----	----		----	----		----
339		----		----	----		----	----		----
349		----		----	----		----	----		----
360	ISO4406	101.8		-1.32	20.6		-2.42	3.7		-0.94
398		----		----	----		----	----		----
432		----		----	----		----	----		----
442		----		----	----		----	----		----
451	D7647	161		-0.66	28		-1.98	2		-1.45
496	D7647	85		-1.51	12		-2.93	1		-1.75
526		----		----	----		----	----		----
541	ISO11500	187.9		-0.36	55.9		-0.32	5.63		-0.35
562		----		----	----		----	----		----
575	D7647	230		0.11	47		-0.85	5		-0.54
603		----		----	----		----	----		----
614	ISO4406	417	C	2.18	115	C	3.19	7	C	0.06
621	ISO4407	1660	R(0.01)	16.00	258	R(0.05)	11.68	2	ex	-1.45
633	D7647	737.4	R(0.05)	5.75	274.0	R(0.05)	12.63	16.0	ex	2.78
634	ISO11500	266		0.51	60		-0.08	6		-0.24
780	ISO11500	142		-0.87	31		-1.80	5		-0.54
862	ISO11500	488		2.97	194		7.88	17		3.08
863	ISO11500	381		1.78	175		6.75	19		3.68
912	D7647	178		-0.47	63.5		0.13	9.9		0.94
962		----		----	----		----	----		----
963		----		----	----		----	----		----
975	D7647	55		-1.84	11		-2.99	1		-1.75
994		----		----	----		----	----		----
1059	D7647	284		0.71	115		3.19	18		3.38
1146	ISO11500	1800	R(0.01)	17.56	370	R(0.05)	18.33	13	ex	1.87
1166	ISO4407	102	C	-1.32	24.40	C	-2.20	2.50	C	-1.30
1265	ISO11500	433.6		2.37	42.5		-1.12	3.2		-1.09
1297		----		----	----		----	----		----
1397	D7647	74.8		-1.62	17.2		-2.62	6.5		-0.09
1435	D7647	183.40		-0.41	39.67		-1.29	3.53		-0.99
1505	D7647	794	R(0.05)	6.38	310	R(0.05)	14.77	55	R(0.01)	14.54
1531		----		----	----		----	----		----
1660		465		2.72	150		5.27	21.4		4.41
1720		----		----	----		----	----		----
1740		----		----	----		----	----		----
1743		----		----	----		----	----		----
1788		----		----	----		----	----		----
1857		----		----	----		----	----		----
1875		----		----	----		----	----		----
1888		3332.17	R(0.01)	34.59	1497.43	R(0.01)	85.30	546.50	R(0.01)	162.85
1900	D7596	109		-1.24	28		-1.98	5		-0.54
1920	D7596	83.740		-1.52	22.806		-2.29	2.045		-1.43
2493		----		----	----		----	----		----
6002	D7596	3088.6	C,R(0.01)	31.88	1627.8	C,R(0.01)	93.05	42.4	C,R(0.01)	10.74
6016	ISO4406	79.5		-1.57	27.3		-2.02	3		-1.15
6257		----		----	----		----	----		----
6261	ISO4407	194.8		-0.29	27.4		-2.02	2.0		-1.45
6284		----		----	----		----	----		----
7011		----		----	----		----	----		----
	normality	OK			suspect			suspect		
	n	23			23			23		
	outliers	6 +1ex			6 +1ex			3 +4ex		
	mean (n)	220.502			61.360			6.800		
	st.dev. (n)	139.8215			53.8951			6.0871		
	R(calc.)	391.500			150.906			17.044		
	st.dev.(D7647:10)	89.9520			16.8352			3.3142		
	R(D7647:10)	251.866			47.139			9.280		




Lab 331 test results excluded as statistical outliers in scale number  
 Lab 614 first reported as ISO 4406 scale number  
 Lab 621 test result excluded as two other related test results are statistical outliers  
 Lab 633 test result excluded as two other related test results are statistical outliers  
 Lab 1146 test result excluded as two other related test results are statistical outliers  
 Lab 1166 first reported 10200 at  $\geq 4\mu\text{m}$ , 2440 at  $\geq 6\mu\text{m}$ , 250 at  $\geq 14\mu\text{m}$   
 Lab 6002 first reported 4088.6 at  $\geq 4\mu\text{m}$ , 1267.8 at  $\geq 6\mu\text{m}$ , 52.4 at  $\geq 14\mu\text{m}$



Determination of Level of Contamination acc. to ISO4406 scale on sample #19236;  
 results in scale number

lab	method	≥ 4 μm (c)	mark	z(targ)	≥ 6 μm (c)	mark	z(targ)	≥ 14 μm (c)	mark	z(targ)
178	ISO4406	16		2.22	14		3.46	10		1.13
179		----		----	----		----	----		----
225		----		----	----		----	----		----
237		----		----	----		----	----		----
255		----		----	----		----	----		----
257		----		----	----		----	----		----
311		----		----	----		----	----		----
317		----		----	----		----	----		----
325		----		----	----		----	----		----
331	ISO4406Mod.	6	E,R(0.01)	-14.07	6	E,R(0.01)	-15.00	4	E,R(0.05)	-7.18
335		----		----	----		----	----		----
339		----		----	----		----	----		----
349		----		----	----		----	----		----
360	ISO4406	14		-1.04	12		-1.15	9		-0.25
398		----		----	----		----	----		----
432		----		----	----		----	----		----
442		----		----	----		----	----		----
451	ISO4406	15		0.59	12		-1.15	8		-1.64
496	ISO4406	14		-1.04	11		-3.46	6	E	-4.41
526	ISO4406	15		0.59	12		-1.15	9		-0.25
541	ISO4406	15		0.59	13		1.15	10		1.13
562		----		----	----		----	----		----
575	ISO4406	15		0.59	13		1.15	9		-0.25
603		----		----	----		----	----		----
614		----		----	----		----	----		----
621	ISO4406	18	ex	5.48	16	E,ex	8.08	8	ex	-1.64
633	ISO4406	17	ex	3.85	15	ex	5.77	11	ex	2.52
634	ISO4406	15		0.59	13		1.15	10		1.13
780	ISO11500	14		-1.04	12		-1.15	9		-0.25
862	ISO4406	16		2.22	15		5.77	11		2.52
863		----		----	----		----	----		----
912	D7647	15		0.59	13		1.15	10		1.13
962		----		----	----		----	----		----
963		----		----	----		----	----		----
975	ISO4406	13		-2.67	11		-3.46	7		-3.02
994		----		----	----		----	----		----
1059	ISO4406	15		0.59	14		3.46	11		2.52
1146	ISO4406	18	ex	5.48	16	ex	8.08	11	ex	2.52
1166	ISO4406	14		-1.04	12		-1.15	8		-1.64
1265	ISO4406	16		2.22	13		1.15	9		-0.25
1297		----		----	----		----	----		----
1397	D7647	13		-2.67	11		-3.46	10		1.13
1435		15		0.59	12		-1.15	9		-0.25
1505	ISO4406	17	ex	3.85	15	ex	5.77	13	ex	5.29
1531		----		----	----		----	----		----
1660	ISO4406	16		2.22	14		3.46	12		3.90
1720		----		----	----		----	----		----
1740	ISO4406	16	ex,C	2.22	16	ex,C	8.08	14	R(0.05)	6.67
1743		----		----	----		----	----		----
1788		----		----	----		----	----		----
1857		----		----	----		----	----		----
1875		----		----	----		----	----		----
1888	ISO4406	19	ex	7.11	18	ex	12.70	16	ex	9.44
1900	D7647	14		-1.04	12		-1.15	9		-0.25
1920	ISO4406	14		-1.04	12		-1.15	8		-1.64
2493		----		----	----		----	----		----
6002	ISO4406	19	ex	7.11	18	ex,C	12.70	13	ex	5.29
6016	ISO4406	13		-2.67	12		-1.15	9		-0.25
6257		----		----	----		----	----		----
6261	ISO4406	15		0.59	12		-1.15	9	E	-0.25
6284		----		----	----		----	----		----
7011		----		----	----		----	----		----
	normality	OK			OK			OK		
	n	22			22			22		
	outliers	1 +7ex			1 +7ex			2 +6ex		
	mean (n)	14.64			12.50			9.18		
	st.dev. (n)	0.953			1.058			1.332		
	R(calc.)	2.67			2.96			3.73		
	st.dev.(D7647:10)	0.614			0.433			0.722		
	R(D7647:10)	1.72			1.21			2.02		

- Lab 331 no correct conversion from counts/mL to ISO4406 scale number respectively 16, 14, 10
- Lab 496 no correct conversion from counts/mL to ISO4406 scale number  $\geq 14 \mu\text{m}$  (c) =7
- Lab 621 test results excluded as statistical outlier in counts/mL
- Lab 621 no correct conversion from counts/mL to ISO4406 scale number  $\geq 6 \mu\text{m}$  (c) =15
- Lab 633 test results excluded as statistical outlier in counts/mL
- Lab 1146 test results excluded as statistical outlier in counts/mL
- Lab 1505 test results excluded as statistical outlier in counts/mL
- Lab 1740 test results excluded as statistical outlier at related scale number  $\geq 14 \mu\text{m}$  (c)
- Lab 1888 test results excluded as statistical outlier in counts/mL
- Lab 6002 test results excluded as statistical outlier in counts/mL
- Lab 6261 no correct conversion from counts/mL to ISO4406 scale number  $\geq 14 \mu\text{m}$  (c) =8




Determination of Aluminum as Al on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		-0.13	
179		----		----	
237	D5185	12.40		0.03	
255	D6595	13.45		0.43	
257	D6595	13.43		0.42	
311		----		----	
325	D5185	12		-0.13	
331	D5185Mod.	12		-0.13	
339		----		----	
349	D5185	13		0.26	
360	D5185	11.6		-0.28	
398		----		----	
432	D5185	12.4		0.03	
451	D5185	12		-0.13	
496	D5185	11.35		-0.38	
526	D5185	2.843	C,R(0.01)	-3.64	first reported 3
541	D5185	11.6		-0.28	
575	D6595	16		1.41	
603		----		----	
614	D5185	12.3		-0.01	
633	D6595	12.0		-0.13	
634	D6595	15.98		1.40	
780	D5185	12		-0.13	
862	D5185	11		-0.51	
863	D5185	11		-0.51	
912	D5185	11.6		-0.28	
962	D5185	12.37		0.02	
963	D5185	11.52		-0.31	
975	D5185	11.64		-0.26	
994	D5185	11.5		-0.32	
1059	In house	14		0.64	
1146	D5185	11.66		-0.26	
1166	In house	11.93		-0.15	
1265	D6595	15.322		1.15	
1297		----		----	
1372	D5185	11.46		-0.33	
1435	D5185	11.25		-0.41	
1505	D5185	11.39		-0.36	
1660	D5185	2.2	R(0.01)	-3.88	
1740	D5185	12		-0.13	
1743	D5185	12		-0.13	
1857	D5185	12.9		0.22	
1875		10		-0.89	
1888	D5185	13.2		0.33	
1900	D5185	15.51		1.22	
2493	In house	11.7		-0.24	
6002	D5185	11.8		-0.20	
6016		----		----	
6080	D5185	13		0.26	
6261	D5185	10.301		-0.78	

normality not OK  
n 41  
outliers 2  
mean (n) 12.331  
st.dev. (n) 1.3741  
R(calc.) 3.848  
st.dev.(D5185:18) 2.6079  
R(D5185:18) 7.302

application range 6 – 40 mg/kg






Determination of Barium as Ba on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	25		-0.05	
179		----		----	
237	D5185	21.63	C	-0.87	first reported 14.59
255		----		----	
257	D6595	27.06		0.45	
311		----		----	
325	D5185	25		-0.05	
331	D5185Mod.	25		-0.05	
339	INH-165	24		-0.29	
349	D5185	26		0.19	
360	D5185	25.5		0.07	
398		----		----	
432	D5185	23.4		-0.44	
451	D5185	24		-0.29	
496	D5185	24.67		-0.13	
526	D5185	21	ex	-1.03	excluded as many statistical outliers in other related metal analyzes
541	D5185	25.07		-0.03	
575	D6595	29		0.92	
603		----		----	
614	D5185	26.33		0.27	
633	D6595	25.6		0.10	
634	D6595	31.57	R(0.01)	1.55	
780	D5185	25		-0.05	
862	D5185	25		-0.05	
863	D5185	26		0.19	
912	D5185	25.1		-0.03	
962	D5185	24.28		-0.23	
963	D5185	24.86		-0.08	
975	D5185	24.84		-0.09	
994	D5185	23.5		-0.42	
1059		----		----	
1146	D5185	25.49		0.07	
1166	In house	24.46		-0.18	
1265	D6595	32.445	R(0.01)	1.76	
1297		----		----	
1372	D5185	25.55		0.08	
1435	D5185	26.68		0.36	
1505	D5185	27.79		0.63	
1660		----		----	
1740	D5185	25.8		0.14	
1743	D5185	26.5		0.31	
1857	D5185	25.52		0.08	
1875		----		----	
1888	D5185	23.9		-0.32	
1900	D5185	26.067		0.21	
2493	In house	26.9		0.41	
6002	D5185	22.0		-0.78	
6016		----		----	
6080	D5185	25		-0.05	
6261		----		----	

normality suspect  
n 36  
outliers 2 +1ex  
mean (n) 25.208  
st.dev. (n) 1.4290  
R(calc.) 4.001  
st.dev.(D5185:18) 4.1031  
R(D5185:18) 11.489

application range 0.5 – 4 mg/kg




Determination of Boron as B on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	13		0.25	
179		----		----	
237	D5185	9.42		-0.51	
255		----		----	
257	D6595	9.85		-0.42	
311		----		----	
325	D5185	14		0.46	
331	D5185Mod.	12		0.04	
339	INH-165	<50		----	
349	D5185	12		0.04	
360	D5185	14.4		0.54	
398		----		----	
432	D5185	12.3		0.10	
451	D5185	12		0.04	
496	D5185	13.99		0.45	
526	D5185	11	ex	-0.17	excluded as many statistical outliers in other related metal analyzes
541	D5185	19.8	R(0.05)	1.67	
575	D6595	14		0.46	
603		----		----	
614	D5185	12.8		0.20	
633	D6595	28.7	R(0.01)	3.55	
634	D6595	9.68		-0.45	
780	D5185	14		0.46	
862	D5185	12		0.04	
863	D5185	12		0.04	
912	D5185	11.3		-0.11	
962		----		----	
963	D5185	11.85		0.00	
975	D5185	11.84		0.00	
994	D5185	9.0		-0.59	
1059		----		----	
1146		----		----	
1166		----		----	
1265	D6595	11.148		-0.14	
1297		----		----	
1372	D5185	11.97		0.03	
1435	D5185	13.30		0.31	
1505		----		----	
1660		----		----	
1740	D5185	11.6		-0.05	
1743	D5185	16		0.88	
1857	D5185	10.1		-0.36	
1875		----		----	
1888	D5185	11.3		-0.11	
1900	D5185	6.437		-1.13	
2493	In house	9.21		-0.55	
6002	D5185	11.2		-0.13	
6016		----		----	
6080	D5185	13		0.25	
6261	D5185	20.413	R(0.05)	1.80	

normality suspect  
n 31  
outliers 3 + 1ex  
mean (n) 11.829  
st.dev. (n) 1.9304  
R(calc.) 5.405  
st.dev.(D5185:18) 4.7590  
R(D5185:18) 13.325


application range 4 – 30 mg/kg



Determination of Cadmium as Cd on sample #19237; results in mg/kg


lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
237	D5185	11.87		0.43	
255		----		----	
257		----		----	
311		----		----	
325		----		----	
331	D5185Mod.	11		-0.26	
339	INH-165	11		-0.26	
349	D5185	14		2.12	
360	D5185	11.6		0.21	
398		----		----	
432	D5185	12.1		0.61	
451	D5185	10		-1.06	
496	D5185	11.11		-0.18	
526		----		----	
541		----		----	
575		----		----	
603		----		----	
614	D5185	11.94		0.48	
633	D6595	12.1		0.61	
634	D6595	8.87		-1.96	
780	D5185	12		0.53	
862	D5185	11		-0.26	
863	D5185	12		0.53	
912	D5185	12.0		0.53	
962		----		----	
963	D5185	11.45		0.09	
975		----		----	
994	D5185	10.0		-1.06	
1059		----		----	
1146		----		----	
1166	In house	10.04		-1.03	
1265	D6595	10.368		-0.77	
1297		----		----	
1372	D5185	12.26		0.74	
1435	D5185	12.31		0.78	
1505	D5185	13.08		1.39	
1660		----		----	
1740		----		----	
1743	D5185	12.4		0.85	
1857	D5185	12.4		0.85	
1875		----		----	
1888	D5185	9.9		-1.14	
1900		----		----	
2493	In house	8.29		-2.42	
6002	D5185	8.21	C	-2.48	first reported 6.21
6016		----		----	
6080	D5185	14		2.12	
6261		----		----	

normality OK  
n 28  
outliers 0  
mean (n) 11.332  
st.dev. (n) 1.4694  
R(calc.) 4.114  
st.dev.(Horwitz) 1.2582  
R(Horwitz) 3.523



Determination of Chromium as Cr on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		-0.03	
179		----		----	
237	D5185	11.91		-0.10	
255	D6595	13.82		1.35	
257	D6595	13.57		1.16	
311		----		----	
325	D5185	12		-0.03	
331	D5185Mod.	12		-0.03	
339	INH-165	11		-0.79	
349	D5185	13		0.73	
360	D5185	12.1		0.04	
398		----		----	
432	D5185	12.7		0.50	
451	D5185	11		-0.79	
496	D5185	11.067		-0.74	
526	D5185	10	ex	-1.55	excluded as many statistical outliers in other related metal analyzes
541	D5185	11.6		-0.34	
575	D6595	15		2.24	
603		----		----	
614	D5185	12.09		0.04	
633	D6595	9.68		-1.79	
634	D6595	13.83		1.35	
780	D5185	12		-0.03	
862	D5185	11		-0.79	
863	D5185	12		-0.03	
912	D5185	12.1		0.04	
962	D5185	11.85		-0.15	
963	D5185	12.18		0.10	
975	D5185	11.8		-0.18	
994	D5185	11.5		-0.41	
1059	In house	12		-0.03	
1146	D5185	11.41		-0.48	
1166	In house	10.69		-1.02	
1265	D6595	15.510		2.63	
1297		----		----	
1372	D5185	11.81		-0.18	
1435	D5185	10.52		-1.15	
1505	D5185	13.26		0.92	
1660	D5185	2.0	R(0.01)	-7.61	
1740	D5185	11.3		-0.56	
1743	D5185	13		0.73	
1857	D5185	12.3		0.19	
1875		10		-1.55	
1888	D5185	14.67	C	1.99	first reported 20.6
1900	D5185	12.27		0.17	
2493	In house	10.4		-1.24	
6002	D5185	10.5		-1.17	
6016		----		----	
6080	D5185	13		0.73	
6261	D5185	10.363		-1.27	
normality		OK			
n		42			
outliers		1 + 1ex			
mean (n)		12.043			
st.dev. (n)		1.2954			
R(calc.)		3.627			
st.dev.(D5185:18)		1.3200			
R(D5185:18)		3.696			
application range 1 – 40 mg/kg					




Determination of Copper as Cu on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	14		0.01	
179		----		----	
237	D5185	14.69		0.59	
255	D6595	14.46		0.40	
257	D6595	13.89		-0.08	
311		----		----	
325	D5185	14		0.01	
331	D5185Mod.	14		0.01	
339	INH-165	12.5		-1.24	
349	D5185	16		1.68	
360	D5185	13.7		-0.24	
398		----		----	
432	D5185	14.3		0.26	
451	D5185	12		-1.66	
496	D5185	12.93		-0.88	
526	D5185	9.660	C,R(0.01)	-3.61	first reported 10
541	D5185	13.6		-0.32	
575	D6595	15		0.85	
603		----		----	
614	D5185	14.65		0.56	
633	D6595	12.8		-0.99	
634	D6595	13.78		-0.17	
780	D5185	14		0.01	
862	D5185	15		0.85	
863	D5185	14		0.01	
912	D5185	14.3		0.26	
962	D5185	14.38		0.33	
963	D5185	13.31		-0.56	
975	D5185	13.48		-0.42	
994	D5185	13.0		-0.82	
1059	In house	15		0.85	
1146	D5185	13.76		-0.19	
1166	In house	12.44		-1.29	
1265	D6595	15.741		1.47	
1297		----		----	
1372	D5185	14.21		0.19	
1435	D5185	14.19		0.17	
1505	D5185	14.46		0.40	
1660	D5185	2.8	R(0.01)	-9.33	
1740	D5185	13.8		-0.15	
1743	D5185	14		0.01	
1857	D5185	14.4		0.35	
1875		10	R(0.01)	-3.32	
1888	D5185	13.2		-0.65	
1900	D5185	15.318		1.11	
2493	In house	11.8		-1.82	
6002	D5185	14.3		0.26	
6016		----		----	
6080	D5185	15		0.85	
6261		----	W	----	test result withdrawn, reported 9.427

normality OK  
n 40  
outliers 3  
mean (n) 13.985  
st.dev. (n) 0.9287  
R(calc.) 2.600  
st.dev.(D5185:18) 1.1987  
R(D5185:18) 3.356

application range 2 – 160 mg/kg




Determination of Iron as Fe on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	15		-0.01	
179		-----		-----	
237	D5185	12.94		-1.28	
255	D6595	16.43		0.87	
257	D6595	15.58		0.35	
311		-----		-----	
325	D5185	15		-0.01	
331	D5185Mod.	15		-0.01	
339	INH-165	17.5		1.53	
349	D5185	17		1.22	
360	D5185	14.8		-0.13	
398		-----		-----	
432	D5185	15.3		0.17	
451	D5185	12		-1.86	
496	D5185	14.05		-0.60	
526	D5185	11.371	ex,C	-2.25	fr. 11, excluded as many statistical outliers in other related metal analyzes
541	D5185	14.9		-0.07	
575	D6595	15		-0.01	
603		-----		-----	
614	D5185	15.14		0.08	
633	D6595	14.3		-0.44	
634	D6595	15.36		0.21	
780	D5185	15		-0.01	
862	D5185	15		-0.01	
863	D5185	15		-0.01	
912	D5185	15.0		-0.01	
962	D5185	15.73		0.44	
963	D5185	15.50		0.30	
975	D5185	15.45		0.27	
994	D5185	12.9		-1.31	
1059	In house	13		-1.24	
1146	D5185	14.76		-0.16	
1166	In house	14.38		-0.39	
1265	D6595	15.643		0.39	
1297	In house	16.189		0.72	
1372	D5185	15.25		0.14	
1435	D5185	14.85		-0.10	
1505	D5185	16.28		0.78	
1660	D5185	2.4	R(0.01)	-7.78	
1740	D5185	15.4		0.24	
1743	D5185	16		0.61	
1857	D5185	15.0		-0.01	
1875		10	R(0.01)	-3.09	
1888	D5185	15.3		0.17	
1900	D5185	15.110		0.06	
2493	In house	14.4		-0.38	
6002	D5185	12.3		-1.68	
6016		-----		-----	
6080	D5185	17		1.22	
6261		-----	W	-----	test result withdrawn, reported 9.938

normality suspect  
n 41  
outliers 2 +1ex  
mean (n) 15.018  
st.dev. (n) 1.1602  
R(calc.) 3.249  
st.dev.(D5185:18) 1.6223  
R(D5185:18) 4.543

application range 2 – 140 mg/kg




Determination of Lead as Pb on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	13		0.33	
179		----		----	
237	D5185	10.73		-0.53	
255	D6595	13.58		0.55	
257	D6595	13.87		0.66	
311		----		----	
325	D5185	12		-0.05	
331	D5185Mod.	12		-0.05	
339		----		----	
349	D5185	15		1.09	
360	D5185	11.9		-0.09	
398		----		----	
432	D5185	12.4		0.11	
451	D5185	11		-0.43	
496	D5185	10.755		-0.52	
526	D5185	3.702	C,R(0.01)	-3.20	first reported 4
541	D5185	11.7		-0.16	
575		----		----	
603		----		----	
614	D5185	12.72		0.23	
633	D6595	12.3		0.07	
634	D6595	14.26		0.81	
780	D5185	12		-0.05	
862	D5185	11		-0.43	
863	D5185	13		0.33	
912		----		----	
962		----		----	
963	D5185	10.76		-0.52	
975	D5185	10.74		-0.53	
994	D5185	12.0		-0.05	
1059	In house	13		0.33	
1146	D5185	11.13		-0.38	
1166	In house	11.68		-0.17	
1265	D6595	14.183		0.78	
1297		----		----	
1372	D5185	12.07		-0.02	
1435	D5185	11.09		-0.39	
1505	D5185	13.96		0.70	
1660	D5185	2.5	R(0.01)	-3.66	
1740	D5185	11.7		-0.16	
1743	D5185	13		0.33	
1857	D5185	13.0		0.33	
1875		----		----	
1888	D5185	12.7		0.22	
1900	D5185	10.220		-0.72	
2493	In house	11.0		-0.43	
6002	D5185	9.97		-0.82	
6016		----		----	
6080	D5185	13		0.33	
6261	D5185	10.159		-0.75	

normality OK  
n 37  
outliers 2  
mean (n) 12.124  
st.dev. (n) 1.2611  
R(calc.) 3.531  
st.dev.(D5185:18) 2.6307  
R(D5185:18) 7.366


application range 10 – 160 mg/kg



Determination of Lithium as Li on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
237	D5185	38.95	C,G(0.01)	-3.23	first reported 63.45
255		----		----	
257		----		----	
311		----		----	
325		----		----	
331	D5185Mod.	53		-0.28	
339		----		----	
349	D5185	54		-0.07	
360		----		----	
398		----		----	
432		----		----	
451	D5185	43	G(0.01)	-2.38	
496	D5185	52.59		-0.37	
526		----		----	
541		----		----	
575		----		----	
603		----		----	
614	D5185	52.7		-0.35	
633	D6595	53.1		-0.26	
634		----		----	
780		----		----	
862	D5185	55		0.14	
863		----		----	
912		----		----	
962		----		----	
963		----		----	
975		----		----	
994		----		----	
1059		----		----	
1146	D5185	56.85		0.52	
1166		----		----	
1265	D6595	55.928		0.33	
1297		----		----	
1372		----		----	
1435	D5185	55.22		0.18	
1505		----		----	
1660		----		----	
1740		----		----	
1743	D5185	57		0.56	
1857		----		----	
1875		----		----	
1888		----		----	
1900		----		----	
2493	In house	53.7		-0.14	
6002	D6595	53.09		-0.26	
6016		----		----	
6080		----		----	
6261		----	W	----	test result withdrawn, reported <1

normality OK  
n 12  
outliers 2  
mean (n) 54.348  
st.dev. (n) 1.6034  
R(calc.) 4.490  
st.dev.(Horwitz) 4.7658  
R(Horwitz) 13.344






Determination of Magnesium as Mg on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	25		0.60	
179		----		----	
237	D5185	20.98		-0.78	
255	D6595	25.37		0.73	
257	D6595	26.57		1.14	
311		----		----	
325	D5185	23		-0.09	
331	D5185Mod.	23		-0.09	
339	INH-165	<50		----	
349	D5185	25		0.60	
360	D5185	22.92		-0.11	
398		----		----	
432	D5185	24.4		0.40	
451	D5185	19		-1.47	
496	D5185	22.28		-0.34	
526	D5185	20	ex	-1.12	excluded as many statistical outliers in other related metal analyzes
541	D5185	23.58		0.11	
575	D6595	25		0.60	
603		----		----	
614	D5185	24.1		0.29	
633	D6595	17.7		-1.91	
634	D6595	27.80		1.57	
780	D5185	23.6		0.12	
862	D5185	24		0.26	
863	D5185	26		0.95	
912	D5185	23.0		-0.09	
962	D5185	23.39		0.05	
963	D5185	23.43		0.06	
975	D5185	23.46		0.07	
994	D5185	21.0		-0.78	
1059		----		----	
1146	D5185	23.21		-0.01	
1166	In house	22.21		-0.36	
1265	D6595	30.507	R(0.05)	2.50	
1297		----		----	
1372	D5185	23.98		0.25	
1435	D5185	23.43		0.06	
1505	D5185	25.75		0.86	
1660	D5185	4.6	R(0.01)	-6.43	
1740	D5185	24		0.26	
1743	D5185	24.8		0.53	
1857	D5185	22.24		-0.35	
1875		20		-1.12	
1888	D5185	20.7		-0.88	
1900	D5185	21.272		-0.68	
2493	In house	23.5		0.09	
6002	D5185	22.6		-0.23	
6016		----		----	
6080	D5185	24		0.26	
6261	D5185	21.583		-0.58	

normality OK  
 n 39  
 outliers 2 +1ex  
 mean (n) 23.253  
 st.dev. (n) 2.0060  
 R(calc.) 5.617  
 st.dev.(D5185:18) 2.8997  
 R(D5185:18) 8.119

application range 5 – 1700 mg/kg




Determination of Manganese as Mn on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	11		-0.33	
179		----		----	
237	D5185	10.06		-1.44	
255	D6595	12.28		1.17	
257	D6595	12.56		1.50	
311		----		----	
325	D5185	11		-0.33	
331	D5185Mod.	11		-0.33	
339	INH-165	11.5		0.26	
349	D5185	12		0.84	
360	D5185	10.9		-0.45	
398		----		----	
432	D5185	12.2		1.08	
451	D5185	10		-1.51	
496	D5185	11.39		0.13	
526		----		----	
541	D5185	11.1		-0.21	
575	D6595	11	C	-0.33	first reported 14
603		----		----	
614	D5185	11.7		0.49	
633	D6595	10.7		-0.68	
634	D6595	12.95		1.96	
780	D5185	11		-0.33	
862	D5185	11		-0.33	
863	D5185	11		-0.33	
912	D5185	11.3		0.02	
962	D5185	10.77		-0.60	
963	D5185	11.59		0.36	
975	D5185	11.60		0.37	
994	D5185	10.0		-1.51	
1059	In house	11		-0.33	
1146	D5185	11.51		0.27	
1166	In house	13.47		2.57	
1265	D6595	16.010	R(0.01)	5.56	
1297		----		----	
1372	D5185	11.62		0.40	
1435	D5185	10.92		-0.43	
1505	D5185	12.19		1.07	
1660		----		----	
1740	D5185	10.8		-0.57	
1743	D5185	12		0.84	
1857	D5185	12.1		0.96	
1875		----		----	
1888	D5185	10		-1.51	
1900		----		----	
2493	In house	11.2		-0.10	
6002	D5185	9.35		-2.27	
6016		----		----	
6080	D5185	13		2.02	
6261	D5185	9.253		-2.39	

normality OK  
n 39  
outliers 1  
mean (n) 11.282  
st.dev. (n) 0.9343  
R(calc.) 2.616  
st.dev.(D5185:18) 0.8505  
R(D5185:18) 2.381

application range 5 – 700 mg/kg




Determination of Molybdenum as Mo on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		0.32	
179		----		----	
237	D5185	10.57		-0.78	
255		----		----	
257	D6595	12.04		0.35	
311		----		----	
325	D5185	11		-0.45	
331	D5185Mod.	12		0.32	
339	INH-165	10.5		-0.83	
349	D5185	13		1.09	
360	D5185	11.4		-0.14	
398		----		----	
432	D5185	11.8		0.17	
451	D5185	14		1.86	
496	D5185	11.24		-0.26	
526	D5185	9	ex	-1.98	excluded as many statistical outliers in other related metal analyzes
541	D5185	11.3		-0.22	
575	D6595	11		-0.45	
603		----		----	
614	D5185	11.8		0.17	
633	D6595	8.77		-2.16	
634	D6595	12.35		0.59	
780	D5185	12		0.32	
862	D5185	11		-0.45	
863	D5185	11		-0.45	
912	D5185	11.3		-0.22	
962	D5185	11.60		0.01	
963	D5185	11.77		0.15	
975	D5185	11.73		0.11	
994	D5185	11.9		0.25	
1059		----		----	
1146		----		----	
1166	In house	11.16		-0.32	
1265	D6595	11.936		0.27	
1297		----		----	
1372	D5185	11.42		-0.12	
1435	D5185	10.35		-0.95	
1505	D5185	12.91		1.02	
1660		----		----	
1740	D5185	11.3		-0.22	
1743	D5185	12		0.32	
1857	D5185	12.2		0.48	
1875		----		----	
1888	D5185	13.1		1.17	
1900	D5185	12.013		0.33	
2493	In house	10.8		-0.60	
6002	D5185	10.5		-0.83	
6016		----		----	
6080	D5185	13		1.09	
6261	D5185	10.305		-0.98	

normality suspect  
n 38  
outliers 0 +1ex  
mean (n) 11.581  
st.dev. (n) 0.9496  
R(calc.) 2.659  
st.dev.(D5185:18) 1.3010  
R(D5185:18) 3.643

application range 5 – 200 mg/kg




Determination of Nickel as Ni on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		0.16	
179		----		----	
237	D5185	9.99		-0.93	
255	D6595	13.66		1.07	
257	D6595	12.56		0.47	
311		----		----	
325	D5185	12		0.16	
331	D5185Mod.	12		0.16	
339	INH-165	11		-0.38	
349	D5185	13		0.71	
360	D5185	11.5		-0.11	
398		----		----	
432	D5185	12.0		0.16	
451	D5185	11		-0.38	
496	D5185	11.175		-0.29	
526	D5185	8	R(0.05)	-2.02	
541	D5185	11.5		-0.11	
575	D6595	13		0.71	
603		----		----	
614	D5185	11.8		0.06	
633	D6595	11.3		-0.22	
634	D6595	13.36		0.91	
780	D5185	11		-0.38	
862	D5185	11		-0.38	
863	D5185	11		-0.38	
912	D5185	11.3		-0.22	
962	D5185	11.96		0.14	
963	D5185	11.59		-0.06	
975	D5185	11.63		-0.04	
994	D5185	10.0		-0.93	
1059	In house	12		0.16	
1146	D5185	11.27		-0.23	
1166	In house	11.23		-0.26	
1265	D6595	13.897		1.20	
1297	In house	12.243		0.30	
1372	D5185	11.83		0.07	
1435	D5185	12.57		0.48	
1505	D5185	12.62		0.50	
1660	D5185	2.3	R(0.01)	-5.13	
1740	D5185	11.2		-0.27	
1743	D5185	13		0.71	
1857	D5185	12.6		0.49	
1875		10		-0.93	
1888	D5185	11.6		-0.05	
1900	D5185	11.310		-0.21	
2493	In house	10.4		-0.71	
6002	D5185	9.73		-1.07	
6016		----		----	
6080	D5185	13		0.71	
6261	D5185	10.182		-0.83	

normality OK  
n 43  
outliers 2  
mean (n) 11.698  
st.dev. (n) 1.0085  
R(calc.) 2.824  
st.dev.(D5185:18) 1.8323  
R(D5185:18) 5.130

application range 5 – 40 mg/kg




Determination of Potassium as K on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	0		----	
179		----		----	
237	D5185	<1		----	
255		----		----	
257	D6595	0.12		----	
311		----		----	
325	D5185	0		----	
331	D5185Mod.	0		----	
339	INH-165	<50		----	
349	D5185	1		----	
360	D5185	0		----	
398		----		----	
432	D5185	1		----	
451	D5185	< 1		----	
496	D5185	0.53		----	
526	D5185	7	R(0.01)	----	
541	D5185	<40		----	
575		----		----	
603		----		----	
614	D5185	2.0		----	
633	D6595	0.01		----	
634	D6595	0.008		----	
780	D5185	<1		----	
862	D5185	<1		----	
863	D5185	<1		----	
912		----		----	
962		----		----	
963		----		----	
975		----		----	
994		----		----	
1059		----		----	
1146		----		----	
1166	In house	1.45		----	
1265	D6595	0.025		----	
1297		----		----	
1372	D5185	0.57		----	
1435	D5185	2.155		----	
1505		----		----	
1660	D5185	2.1		----	
1740	D5185	<0.1		----	
1743	D5185	0.47		----	
1857	D5185	0.75		----	
1875		----		----	
1888		----		----	
1900	D5185	0		----	
2493	In house	0.00		----	
6002	D5185	0.207		----	
6016		----		----	
6080	D5185	<1		----	
6261	D5185	9.880	R(0.01)	----	

normality suspect  
n 21  
outliers 2  
mean (n) 0.590  
st.dev. (n) 0.7513  
R(calc.) 2.104  
st.dev.(D5185:18) (2.0229)  
R(D5185:18) (5.664)

application range 40 – 1200 mg/kg




Determination of Silicon as Si on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		0.07	
179		----		----	
237	D5185	10.53		-0.47	
255	D6595	13.21		0.52	
257	D6595	12.43		0.23	
311		----		----	
325	D5185	12		0.07	
331	D5185Mod.	12		0.07	
339		----		----	
349	D5185	12		0.07	
360	D5185	11.4		-0.15	
398		----		----	
432	D5185	12.5		0.25	
451	D5185	10		-0.67	
496	D5185	11.57		-0.09	
526	D5185	4.163	C,R(0.01)	-2.82	first reported 4
541	D5185	11.4		-0.15	
575	D6595	12		0.07	
603		----		----	
614	D5185	11.8		0.00	
633	D6595	13.2		0.51	
634	D6595	14.98	C,R(0.05)	1.17	first reported 18.14
780	D5185	12		0.07	
862	D5185	11		-0.30	
863	D5185	12		0.07	
912	D5185	11.5		-0.11	
962	D5185	11.36		-0.17	
963	D5185	11.39		-0.15	
975	D5185	11.81		0.00	
994	D5185	11.3		-0.19	
1059	In house	11		-0.30	
1146	D5185	11.61		-0.07	
1166	In house	11.09		-0.27	
1265	D6595	13.269		0.54	
1297		----		----	
1372	D5185	10.93		-0.32	
1435	D5185	10.85		-0.35	
1505	D5185	12.81		0.37	
1660	D5185	2.5	R(0.01)	-3.43	
1740	D5185	11.5		-0.11	
1743	D5185	13		0.44	
1857	D5185	12.8		0.36	
1875		----		----	
1888	D5185	12.4		0.22	
1900	D5185	12.158		0.13	
2493	In house	0.00	R(0.01)	-4.35	
6002	D5185	10.3		-0.56	
6016		----		----	
6080	D5185	14		0.81	
6261	D5185	10.673		-0.42	

normality OK  
n 38  
outliers 4  
mean (n) 11.810  
st.dev. (n) 0.8865  
R(calc.) 2.482  
st.dev.(D5185:18) 2.7128  
R(D5185:18) 7.596

application range 8 – 50 mg/kg




Determination of Silver as Ag on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		0.21	
179		----		----	
237	D5185	12.87		0.80	
255		----		----	
257		----		----	
311		----		----	
325	D5185	12		0.21	
331	D5185Mod.	11		-0.48	
339		----		----	
349	D5185	14		1.58	
360	D5185	11.4		-0.20	
398		----		----	
432	D5185	12.2		0.34	
451	D5185	10		-1.16	
496	D5185	11.11		-0.40	
526	D5185	6.512	C,R(0.01)	-3.55	first reported 7
541	D5185	11.9		0.14	
575	D6595	11		-0.48	
603		----		----	
614	D5185	12.1		0.28	
633	D6595	10.8		-0.61	
634	D6595	9.42		-1.56	
780	D5185	11		-0.48	
862	D5185	12		0.21	
863	D5185	12		0.21	
912	D5185	12.4		0.48	
962	D5185	11.36		-0.23	
963	D5185	11.40		-0.20	
975	D5185	11.39		-0.21	
994	D5185	11.5		-0.13	
1059		----		----	
1146	D5185	11.79		0.06	
1166	In house	16.54	R(0.01)	3.31	
1265	D6595	11.450		-0.17	
1297		----		----	
1372	D5185	11.80		0.07	
1435	D5185	12.22		0.36	
1505	D5185	12.57		0.60	
1660	D5185	2.6	R(0.01)	-6.22	
1740	D5185	11		-0.48	
1743	D5185	12		0.21	
1857	D5185	14.7		2.05	
1875		----		----	
1888		----		----	
1900	D5185	11.950		0.17	
2493	In house	8.90		-1.91	
6002	D5185	10.1		-1.09	
6016		----		----	
6080	D5185	14		1.58	
6261	D5185	12.069		0.25	

normality suspect  
n 35  
outliers 3  
mean (n) 11.697  
st.dev. (n) 1.1591  
R(calc.) 3.246  
st.dev.(D5185:18) 1.4621  
R(D5185:18) 4.094

application range 0.5 – 50 mg/kg




Determination of Sodium as Na on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		-0.04	
179		----		----	
237	D5185	9.76		-1.01	
255	D6595	12.45		0.16	
257		----		----	
311		----		----	
325	D5185	11		-0.47	
331	D5185Mod.	11		-0.47	
339	INH-165	<50		----	
349	D5185	12		-0.04	
360	D5185	12.6		0.23	
398		----		----	
432	D5185	11.8		-0.12	
451	D5185	8		-1.77	
496	D5185	9.64		-1.06	
526	D5185	5.276	ex,C	-2.95	fr. 5, excluded as many statistical outliers in other related metal analyzes
541	D5185	11.3		-0.34	
575	D6595	14	C	0.83	first reported 19
603		----		----	
614	D5185	12.7		0.27	
633	D6595	13.4		0.57	
634	D6595	16.99		2.13	
780	D5185	12		-0.04	
862	D5185	12		-0.04	
863	D5185	12		-0.04	
912	D5185	11.7		-0.17	
962	D5185	11.17		-0.40	
963	D5185	11.31		-0.33	
975	D5185	11.36		-0.31	
994	D5185	9.5		-1.12	
1059		----		----	
1146	D5185	12.19		0.05	
1166	In house	16.60		1.96	
1265	D6595	18.105		2.61	
1297		----		----	
1372	D5185	12.16		0.03	
1435	D5185	12.39		0.13	
1505		----		----	
1660	D5185	1.1	R(0.01)	-4.77	
1740	D5185	12.1		0.01	
1743	D5185	12		-0.04	
1857	D5185	14.1		0.88	
1875		10		-0.90	
1888	D5185	45.56	C,R(0.01)	14.53	first reported 33.2
1900	D5185	20.428	R(0.05)	3.62	
2493	In house	12.1		0.01	
6002	D5185	12.4		0.14	
6016		----		----	
6080	D5185	14		0.83	
6261	D5185	7.088		-2.17	

normality suspect  
n 36  
outliers 3 +1ex  
mean (n) 12.081  
st.dev. (n) 2.1651  
R(calc.) 6.062  
st.dev.(D5185:18) 2.3042  
R(D5185:18) 6.452

application range 7 – 70 mg/kg






Determination of Tin as Sn on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	11		0.16	
179		----		----	
237	D5185	8.27		-0.69	
255		----		----	
257		----		----	
311		----		----	
325	D5185	9		-0.46	
331	D5185Mod.	10		-0.15	
339	INH-165	9		-0.46	
349	D5185	12		0.47	
360	D5185	9.6		-0.27	
398		----		----	
432	D5185	10.3		-0.06	
451	D5185	9		-0.46	
496	D5185	10.40		-0.03	
526	D5185	1.591	C,R(0.01)	-2.76	first reported 1
541	D5185	<10		----	
575	D6595	12		0.47	
603		----		----	
614	D5185	11.16		0.21	
633	D6595	12.4		0.60	
634	D6595	11.13		0.20	
780	D5185	9.6		-0.27	
862	D5185	10		-0.15	
863	D5185	10		-0.15	
912	D5185	10		-0.15	
962	D5185	9.80		-0.21	
963	D5185	10.06		-0.13	
975	D5185	10.11		-0.12	
994	D5185	9.5		-0.31	
1059	In house	10		-0.15	
1146	D5185	10.39		-0.03	
1166	In house	13.86		1.05	
1265	D6595	13.384		0.90	
1297		----		----	
1372	D5185	10.14		-0.11	
1435	D5185	9.976		-0.16	
1505		----		----	
1660	D5185	2.2	R(0.01)	-2.57	
1740	D5185	9.8		-0.21	
1743	D5185	11		0.16	
1857	D5185	10.6		0.04	
1875		----		----	
1888	D5185	12.8		0.72	
1900	D5185	10.552		0.02	
2493	In house	7.92		-0.80	
6002	D5185	10.4		-0.03	
6016		----		----	
6080	D5185	14		1.09	
6261	D5185	8.727		-0.55	

normality OK  
n 37  
outliers 2  
mean (n) 10.483  
st.dev. (n) 1.4317  
R(calc.) 4.009  
st.dev.(D5185:18) 3.2193  
R(D5185:18) 9.014

application range 10 – 40 mg/kg




Determination of Titanium as Ti on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	12		0.08	
179		----		----	
237	D5185	10.11		-0.58	
255	D6595	13.29		0.54	
257	D6595	12.87		0.39	
311		----		----	
325	D5185	12		0.08	
331	D5185Mod.	11		-0.27	
339	INH-165	10.5		-0.44	
349	D5185	13		0.43	
360	D5185	12.0		0.08	
398		----		----	
432	D5185	12.1		0.12	
451	D5185	10		-0.62	
496	D5185	11.04		-0.25	
526	D5185	9	ex	-0.97	excluded as many statistical outliers in other related metal analyzes
541	D5185	11.3		-0.16	
575	D6595	11		-0.27	
603		----		----	
614	D5185	12.1		0.12	
633	D6595	11.7		-0.02	
634	D6595	13.71		0.68	
780	D5185	12		0.08	
862	D5185	12		0.08	
863	D5185	11		-0.27	
912	D5185	11.8		0.01	
962	D5185	11.38		-0.14	
963	D5185	11.95		0.07	
975	D5185	11.74		-0.01	
994	D5185	11.5		-0.09	
1059		----		----	
1146	D5185	11.32		-0.16	
1166	In house	10.30		-0.52	
1265	D6595	13.848		0.73	
1297		----		----	
1372	D5185	12.33		0.20	
1435	D5185	12.11		0.12	
1505	D5185	12.07		0.11	
1660	D5185	2.3	R(0.01)	-3.33	
1740	D5185	11.7		-0.02	
1743	D5185	13		0.43	
1857	D5185	12.3		0.19	
1875		20	R(0.01)	2.90	
1888	D5185	9.1		-0.94	
1900	D5185	13.171		0.49	
2493	In house	10.9		-0.30	
6002	D5185	12.2		0.15	
6016		----		----	
6080	D5185	14		0.79	
6261	D5185	9.152		-0.92	

normality OK  
n 40  
outliers 2 +1ex  
mean (n) 11.765  
st.dev. (n) 1.1372  
R(calc.) 3.184  
st.dev.(D5185:18) 2.8442  
R(D5185:18) 7.964

application range 5 – 40 mg/kg




Determination of Vanadium as V on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	11		-0.43	
179		----		----	
237	D5185	10.42		-0.82	
255	D6595	13.26		1.09	
257	D6595	12.56		0.62	
311		----		----	
325	D5185	13		0.91	
331	D5185Mod.	11		-0.43	
339		----		----	
349	D5185	13		0.91	
360	D5185	11.7		0.04	
398		----		----	
432	D5185	11.8		0.11	
451	D5185	10		-1.10	
496	D5185	11.05		-0.40	
526	D5185	8.045	C,R(0.05)	-2.42	first reported 8
541	D5185	11.4		-0.16	
575	D6595	13		0.91	
603		----		----	
614	D5185	12.0		0.24	
633	D6595	10.0		-1.10	
634	D6595	14.07		1.63	
780	D5185	12		0.24	
862	D5185	12		0.24	
863	D5185	11		-0.43	
912	D5185	11.5		-0.10	
962	D5185	12.23		0.40	
963	D5185	11.87		0.15	
975	D5185	11.87		0.15	
994	D5185	11.2		-0.30	
1059	In house	12		0.24	
1146	D5185	11.36		-0.19	
1166	In house	10.06		-1.06	
1265	D6595	12.779		0.76	
1297	In house	11.270		-0.25	
1372	D5185	11.64		0.00	
1435	D5185	11.55		-0.06	
1505	D5185	11.95		0.21	
1660	D5185	2.4	R(0.01)	-6.21	
1740	D5185	11		-0.43	
1743	D5185	12		0.24	
1857	D5185	12.1		0.31	
1875		10		-1.10	
1888	D5185	11.12	C	-0.35	first reported 16.2
1900	D5185	12.587		0.64	
2493	In house	10.8		-0.57	
6002	D5185	11.7		0.04	
6016		----		----	
6080	D5185	13		0.91	
6261	D5185	9.094		-1.71	

normality OK  
n 42  
outliers 2  
mean (n) 11.641  
st.dev. (n) 1.0170  
R(calc.) 2.848  
st.dev.(D5185:18) 1.4880  
R(D5185:18) 4.166

application range 1 – 50 mg/kg




Determination of Calcium as Ca on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	77		-0.08	
179		----		----	
237	D5185	75.32		-0.34	
255		----		----	
257	D6595	69.97		-1.17	
311		----		----	
325	D5185	72		-0.86	
331	D5185Mod.	74		-0.55	
339		----		----	
349	D5185	79		0.23	
360		----		----	
398		----		----	
432	D5185	80.7		0.49	
451	D5185	90		1.93	
496	D5185	73.45		-0.63	
526	D5185	63	ex	-2.26	excluded as many statistical outliers in other related metal analyzes
541	D5185	74.72		-0.44	
575	D6595	92		2.24	
603		----		----	
614	D5185	78.9		0.21	
633	D6595	71.8		-0.89	
634		----		----	
780	D5185	77.3		-0.04	
862	D5185	75		-0.39	
863	D5185	78		0.07	
912		----		----	
962	D5185	73.12		-0.68	
963	D5185	75.17		-0.37	
975	D5185	75.73		-0.28	
994	D5185	75.4		-0.33	
1059	In house	78		0.07	
1146	D5185	75.78		-0.27	
1166	In house	85.57		1.25	
1265	D6595	82.352		0.75	
1297		----		----	
1372	D5185	77.34		-0.03	
1435	D5185	76.00		-0.24	
1505	D5185	83.9		0.99	
1660	D5185	14	R(0.01)	-9.86	
1740	D5185	80		0.38	
1743	D5185	80.9		0.52	
1857	D5185	71.41		-0.95	
1875		80		0.38	
1888	D5185	72.2		-0.83	
1900	D5185	82.789		0.82	
2493	In house	105	C,R(0.01)	4.26	first reported 138
6002	D6595	77.6		0.01	
6016		----		----	
6080	D5185	79		0.23	
6261	D5185	69.795		-1.20	

normality suspect  
n 36  
outliers 2 + 1ex  
mean (n) 77.534  
st.dev. (n) 5.0551  
R(calc.) 14.154  
st.dev.(Horwitz) 6.4449  
R(Horwitz) 18.046

Compare  
R(D5185:18) 4.290




Determination of Phosphorus as P on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	264		-0.08	
179		----		----	
237	D5185	272.1		0.24	
255	D6595	255		-0.44	
257	D6595	245.68		-0.81	
311		----		----	
325	D5185	273		0.28	
331	D5185Mod.	258		-0.32	
339	INH-165	259		-0.28	
349	D5185	274		0.32	
360	D5185	259.4		-0.27	
398		----		----	
432	D5185	276.1		0.40	
451	D5185	291		0.99	
496	D5185	259.347		-0.27	
526	D5185	252	ex	-0.56	excluded as many statistical outliers in other related metal analyzes
541	D5185	268.8		0.11	
575	D6595	302		1.43	
603		----		----	
614	D5185	277.4		0.45	
633	D6595	331.6	R(0.05)	2.62	
634		----		----	
780	D5185	246		-0.80	
862	D5185	275		0.36	
863	D5185	269		0.12	
912	D5185	255		-0.44	
962	D5185	275.16		0.36	
963	D5185	270.1		0.16	
975	D5185	256.3		-0.39	
994	D5185	250		-0.64	
1059	In house	267		0.04	
1146	D5185	266.8		0.03	
1166	In house	269.01		0.12	
1265	D6595	283.24		0.68	
1297		----		----	
1372	D5185	281.00		0.60	
1435	D5185	273.2		0.28	
1505	D5185	282.6		0.66	
1660		----		----	
1740	D5185	272		0.24	
1743	D5185	282		0.64	
1857	D5185	220.3		-1.83	
1875		280		0.56	
1888	D5185	229		-1.48	
1900	D5185	295.09		1.16	
2493	In house	246	C	-0.80	first reported 178
6002	D5185	235		-1.24	
6016		----		----	
6080	D5185	275		0.36	
6261	D5185	254.711		-0.45	

normality OK  
n 40  
outliers 1 +1ex  
mean (n) 266.083  
st.dev. (n) 17.0465  
R(calc.) 47.730  
st.dev.(D5185:18) 25.0507  
R(D5185:18) 70.142

application range 10 – 1000 mg/kg




Determination of Zinc as Zn on sample #19237; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178	D5185	319		0.63	
179		----		----	
237	D5185	304.3		-0.27	
255	D6595	296		-0.78	
257	D6595	284.88		-1.47	
311		----		----	
325	D5185	319		0.63	
331	D5185Mod.	329		1.25	
339		----		----	
349	D5185	328		1.19	
360	D5185	300.2		-0.53	
398		----		----	
432	D5185	322.0		0.82	
451	D5185	301		-0.48	
496	D5185	302.3		-0.40	
526	D5185	274	ex	-2.14	excluded as many statistical outliers in other related metal analyzes
541	D5185	312.6		0.24	
575	D6595	293		-0.97	
603		----		----	
614	D5185	320.4		0.72	
633	D6595	279.3		-1.81	
634		----		----	
780	D5185	303		-0.35	
862	D5185	309		0.02	
863	D5185	309		0.02	
912	D5185	300		-0.54	
962	D5185	306.71		-0.12	
963	D5185	317.0		0.51	
975	D5185	291.8		-1.04	
994		----		----	
1059	In house	311		0.14	
1146	D5185	311.9		0.20	
1166	In house	319.17		0.64	
1265	D6595	345.92		2.29	
1297		----		----	
1372	D5185	329.60		1.29	
1435	D5185	316.6		0.49	
1505	D5185	327.5		1.16	
1660	D5185	75	R(0.01)	-14.40	
1740	D5185	299		-0.60	
1743	D5185	329		1.25	
1857	D5185	327.3		1.14	
1875		320		0.69	
1888	D5185	283		-1.58	
1900	D5185	328.92		1.24	
2493	In house	272		-2.26	
6002	D5185	274		-2.14	
6016		----		----	
6080	D5185	304		-0.29	
6261	D5185	293.864		-0.92	

normality OK  
n 39  
outliers 1 +1ex  
mean (n) 308.725  
st.dev. (n) 17.0367  
R(calc.) 47.703  
st.dev.(D5185:18) 16.2349  
R(D5185:18) 45.458

application range 60 – 1600 mg/kg



**APPENDIX 2****Number of participants per country**

1 lab in ARGENTINA  
1 lab in AUSTRALIA  
1 lab in AUSTRIA  
1 lab in AZERBAIJAN  
2 labs in BELGIUM  
1 lab in BULGARIA  
1 lab in CHILE  
2 labs in CHINA, People's Republic  
1 lab in COLOMBIA  
1 lab in COTE D'IVOIRE  
1 lab in CROATIA  
1 lab in EGYPT  
4 labs in FRANCE  
2 labs in GERMANY  
4 labs in GREECE  
1 lab in HUNGARY  
1 lab in INDIA  
1 lab in INDONESIA  
1 lab in IRAN, Islamic Republic of  
2 labs in ITALY  
1 lab in KAZAKHSTAN  
2 labs in MALAYSIA  
1 lab in MEXICO  
1 lab in MOROCCO  
3 labs in NETHERLANDS  
1 lab in NIGERIA  
1 lab in NORWAY  
1 lab in OMAN  
2 labs in PHILIPPINES  
2 labs in POLAND  
1 lab in PORTUGAL  
1 lab in QATAR  
2 labs in RUSSIAN FEDERATION  
2 labs in SAUDI ARABIA  
1 lab in SINGAPORE  
1 lab in SLOVENIA  
1 lab in SPAIN  
1 lab in SUDAN  
2 labs in SWEDEN  
2 labs in TANZANIA  
1 lab in UNITED ARAB EMIRATES  
2 labs in UNITED KINGDOM  
2 labs in UNITED STATES OF AMERICA

## APPENDIX 3

### Abbreviations

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= possibly an error in calculations
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
SDS	= Safety Data Sheet

### Literature

- 1 iis Interlaboratory Studies, Protocol for the Organization, Statistics and Evaluation, June 2018
- 2 ASTM E178:02
- 3 ASTM E1301:95(2003)
- 4 ISO 5725:86
- 5 ISO 5725, parts 1-6, 1994
- 6 ISO 13528:05
- 7 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 8 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 9 IP 367:84
- 10 DIN 38402 T41/42
- 11 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 12 J.N. Miller, Analyst, 118, 455, (1993)
- 13 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 14 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analys, 127, 1359-1364 (2002)
- 15 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)
- 16 Horwitz, R. Albert, J. AOAC Int, 79\_3, 589, (1996)
- 17 iis memo 1401, R.G. Visser and C. Nijssen-Wester, Estimation of reproducibility and measurement uncertainty of a viscosity test method from proficiency test data, Accred Qual Assur 20, 125-129, (2015)