

Results of Proficiency Test
Toluene
February 2020

Organized by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: ing. M. Meijer
Correctors: ing. A.S. Noordman-de Neef & ing. R.J. Starink
Report: iis20C04

May 2020

CONTENTS

1	INTRODUCTION	3
2	SET UP	3
2.1	ACCREDITATION	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALITY STATEMENT	3
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES.....	4
2.6	ANALYZES	5
3	RESULTS	5
3.1	STATISTICS	5
3.2	GRAPHICS	6
3.3	Z-SCORES	6
4	EVALUATION	7
4.1	EVALUATION PER TEST	7
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	9
4.3	COMPARISON OF THE PROFICIENCY TEST OF FEBRUARY 2020 WITH PREVIOUS PTS	10

Appendices:

1.	Data, statistical and graphic results	11
2.	Number of participants per country	24
3.	Abbreviations and literature.....	25

1 INTRODUCTION

Since 1999, the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the analysis of Toluene based on the scope of the latest version of ASTM D841 every year. During the annual proficiency testing program 2019/2020, it was decided to continue the round robin for the analysis of Toluene.

In this interlaboratory study 36 laboratories in 18 different countries registered for participation. See appendix 2 for the number of participants per country. In this report, the results of this proficiency test for Toluene are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency tests (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory. It was decided to send one sample of one liter of Toluene labelled #20011. The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

A batch of approximately 75 liters of Toluene was obtained from a local chemical supplier. After homogenization 69 amber glass bottles of 1 liter were filled and labelled #20011. The homogeneity of the subsamples was checked by determination of Density at 20°C in accordance with ASTM D4052 on 8 stratified randomly selected subsamples.

	Density at 20°C in kg/L
sample #20011-1	0.86677
sample #20011-2	0.86678
sample #20011-3	0.86680
sample #20011-4	0.86679
sample #20011-5	0.86680
sample #20011-6	0.86683
sample #20011-7	0.86685
sample #20011-8	0.86677

Table 1: homogeneity test results of subsamples #20011

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 20°C in kg/L
r (observed)	0.00008
reference test method	ISO12185:96
0.3 * R(reference test method)	0.00015

Table 2: evaluation of repeatability of subsamples #20011

The calculated repeatability was in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

To each of the participating laboratories one subsample with Toluene labelled #20011 was sent on January 29, 2020. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of Toluene packed in amber glass bottles was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYZES

The participants were requested to determine on the Toluene sample #20011: Acid Wash Color, Appearance, Color Pt/Co, Copper Corrosion, Density at 20°C, Distillation (IBP, 50% recovered, DP and range), Purity by GC, Benzene, Nonaromatics, Total Impurities, Refractive Index at 25°C and Water.

It was explicitly requested to treat the sample as if it was a routine sample and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results, a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the appropriate reference test methods that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalysis). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the test result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

According to ISO5725 the original test results per determination were submitted to Dixon's, Grubbs' and/or Rosner's outlier tests. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve was projected over the Kernel Density Graph for reference.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ASTM or ISO reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation of this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used. In some cases, a reproducibility based on former iis proficiency tests could be used. When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z(\text{target}) = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The z(target) scores are listed in the test result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. The usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

During the execution of this proficiency test some serious problems occurred. Due to COVID-19 outbreak 10 participants were not able to report because of restricted dispatch to some countries. It was decided that for those 10 participants the deadline was extended and an additional data entry portal round was created. Not all laboratories were able to perform all analyzes requested. Finally, 26 participants reported 253 numerical test results. Observed were 16 statistically outlying test results, which is 6.3%. In proficiency studies, outlier percentages of 3% - 7.5% are quite normal.

Not all original data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

4.1 EVALUATION PER TEST

In this section the test results are discussed per test. The test methods, which were used by the various laboratories, were taken into account for explaining the observed differences when possible and applicable. These methods are also in the tables together with the reported test results. The abbreviations, used in these tables, are explained in appendix 3.

In the iis PT reports, ASTM methods are referred to with a number e.g. D1218 and an added designation for the year that the method was adopted or revised e.g. D1218:12. If applicable, a designation in parentheses is added to designate the year of reapproval e.g. D1218:12(2016). In the results tables of appendix 1 only the method number and year of adoption or revision e.g. D1218:12 will be used.

Unfortunately, a suitable reference test method providing the precision data is not available for all determinations. For these the calculated reproducibility was compared against the reproducibility estimated from the Horwitz equation.

Sample #20011

Acid Wash Color: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D848:18.

Appearance: This determination was not problematic. All participants agreed about the appearance of the sample which was bright, clear and free of suspended matter (Pass).

Color Pt/Co: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D5386:16 or ASTM D1209:05(2019).

Copper Corrosion: This determination was not problematic. All participants agreed on a result of 1a/b (Pass).

Density at 20°C: This determination was not problematic. One statistical outlier was observed. However, the calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ISO12185:96.

Distillation: This determination may be problematic for a number of laboratories. In total seven statistical outliers were observed and two other test results were excluded. However, all calculated reproducibilities after rejection of the suspect data are in agreement with the requirements of automated or manual modes of ASTM D850:18.

Purity: This determination was problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D7504:20.

Benzene: This determination was not problematic. Two statistical outliers were observed. However, the calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D7504:20.

Nonaromatics: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D7504:20.

Total Impurities: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the estimated reproducibility calculated using the Horwitz equation.

Refractive Index: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ASTM D1218:12(2016).

Water: This determination was problematic. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D1064:16.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the relevant reference test method or as declared by the estimated target reproducibility using the Horwitz equation and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility ($2.8 \cdot$ standard deviation) and the target reproducibility derived from literature reference test methods (in casu ASTM test methods) or the estimated target reproducibility are presented in the next table.

Parameter	unit	n	average	2.8 * sd	R(lit.)
Acid Wash Color (acid layer)		21	0.5	0.9	1.9
Appearance		23	Pass (B&C)	n.a.	n.a.
Color Pt/Co		21	3.4	3.0	5.1
Copper Corrosion		18	1 (1a,1b)	n.a.	n.a.
Density at 20°C	kg/L	23	0.86679	0.0002	0.0005
Distillation, IBP	°C	15	110.2	0.3	0.6
Distillation, 50% recovered	°C	16	110.6	0.2	0.2
Distillation, DP	°C	17	110.9	0.3	0.5
Purity	%M/M	23	99.783	0.032	0.013
Benzene	mg/kg	24	430.4	63.0	59.6
Nonaromatics	mg/kg	25	625.8	275.5	565.6
Total Impurities	mg/kg	16	2170	430	433
Refractive Index at 25°C		17	1.4940	0.0008	0.0005
Water	mg/kg	17	80.4	19.3	12.8

Table 3: performance evaluation sample #20011

Without further statistical calculations it can be concluded that for most of the tests there is a good compliance of the group of participating laboratories with the relevant reference test methods. The tests that are problematic have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF FEBRUARY 2020 WITH PREVIOUS PTS

	February 2020	February 2019	March 2018	March 2017	March 2016
Number of reporting laboratories	26	35	36 *)	67	59
Number of test results	253	284	267	743	793
Number of statistical outliers	16	14	10	32	19
Percentage of statistical outliers	6.3%	4.9%	3.8%	4.3%	2.4%

Table 4: comparison with previous proficiency tests

*) from March 2018 the Toluene results are reported separately from Benzene, hence the lower number of reporting laboratories.

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared against the requirements of the respective reference test methods. The conclusions are given the following table.

	February 2020	February 2019	March 2018	March 2017	March 2016
Acid Wash Color (acid layer)	++	++	++	+	++
Color Pt/Co	+	++	++	+	++
Copper Corrosion	n.e.	n.e.	n.e.	n.e.	n.e.
Density at 20°C	++	++	++	++	++
Distillation	+	+	+/-	+/-	+
Purity	--	-	+	n.e.	+
Benzene	+/-	(--)	n.e.	--	+/-
Nonaromatics	++	++	+/-	+	++
Total Impurities	+/-	n.e.	n.e.	n.e.	n.e.
Refractive Index at 25°C	-	+	+	+	-
Water	-	n.e.	n.e.	n.e.	n.e.

Table 5: comparison determinations against the reference test methods

*) Reproducibility between brackets is based on a much higher level than present in sample

In the table above the following performance categories were used:

- ++ : group performed much better than the reference test method
- + : group performed better than the reference test method
- +/- : group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e. : not evaluated



APPENDIX 1

Determination of Acid Wash Color (acid layer) on sample #20011;

lab	method	reported test value	iis conversion*	mark	z(targ)	remarks
52	D848	0+	0.25		-0.30	
150	D848	0	0		-0.66	
171	D848	Pass	----		----	
311	D848	0+	0.25		-0.30	
323	D848	-1	0.75		0.44	
333		----	----		----	
334	D848	1-	0.75		0.44	
343	D848	1-	0.75		0.44	
445	D848	1-	0.75		0.44	
551		----	----		----	
555		----	----		----	
663	D848	No. 0	0		-0.66	
823	D848	1-	0.75		0.44	
840	D848	0+	0.25		-0.30	
855		----	----		----	
862		----	----		----	
864		----	----		----	
865		----	----		----	
866		----	----		----	
870		----	----		----	
912	D848	-1	0.75	C	0.44	first reported I
913	D848	1-	0.75		0.44	
1040	D848	<1	0.75		0.44	
1041	D848	1-	0.75		0.44	
1062	D848	0	0		-0.66	
1065		----	----		----	
1079	D848	0+	0.25		-0.30	
1151		----	----		----	
1530	D848	< 1	0.75		0.44	
1741		----	----		----	
1852	D848	0	0		-0.66	
1982	D848	0	0		-0.66	
6203		----	----		----	
6262	D848	1-	0.75		0.44	
6315		----	----		----	
6321	D848	0+	0.25		-0.30	
normality		OK				
n		21				
outliers		0				
mean (n)		0.45				
st.dev. (n)		0.332				
R(calc.)		0.93				
st.dev.(D848:18)		0.682				
R(D848:18)		1.91				

*) In the calculation of the mean, standard deviation, reproducibility and for the graphs, a reported value of 'y-', '-y' or '<y' is changed into y-0.25 (for example 1- into 0.75) and 'y+' is changed into y+0.25 (for example 0+ into 0.25).

**) The precision data of Benzene is used



Determination of Appearance on sample #20011;

lab	method	value	mark	z(targ)	remarks
52	D4176	Pass		----	
150	E2680	Pass		----	
171	E2680	Pass		----	
311	E2680	pass		----	
323	E2680	clear and bright		----	
333		----		----	
334	EN15769	Clear and colourless		----	
343	E2680	PASS		----	
445	D4176	C&B		----	
551		----		----	
555		----		----	
663	Visual	Bright & Clear		----	
823	E2680	Pass		----	
840	E2680	Pass		----	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	E2680	PASS		----	
913	E2680	Pass		----	
1040	Visual	clear & bright visual free		----	
1041	Visual	CBFSM		----	
1062	Visual	pass		----	
1065		----		----	
1079	Visual	Bright&Clear		----	
1151		----		----	
1530	Visual	clear & bright		----	
1741		----		----	
1852	Visual	clear & colourless		----	
1982	Visual	waterclear		----	
6203		----		----	
6262	Visual	Clear and Bright		----	
6315	Visual	clear,bright		----	
6321	D4176	Pass - Clear and Bright		----	
	n	23			
	mean (n)	Pass (B&C)			



Abbreviations:

C&B = clear and bright

CBFSM = clear and bright and free from suspended matter

Determination of Color Pt/Co on sample #20011;

lab	method	value	mark	z(targ)	remarks
52	D5386	3		-0.20	
150	D5386	4		0.35	
171	D1209	5		0.90	
311	D5386	3		-0.20	
323	D5386	3		-0.20	
333	D5386	4		0.35	
334	D1209	5		0.90	
343	D5386	5		0.90	
445	D6045	<5		----	
551		----		----	
555		----		----	
663	D5386	2		-0.74	
823	D5386	3		-0.20	
840	D1209	3		-0.20	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D5386	2		-0.74	
913	D5386	3		-0.20	
1040	D5386	<5		----	
1041	ISO6271	4.8		0.79	
1062	D5386	2		-0.74	
1065		----		----	
1079	D5386	2.7		-0.36	
1151		----		----	
1530	D1209	< 3		----	
1741		----		----	
1852		----		----	
1982	ISO6271	4.2		0.46	
6203	D1209	4.0		0.35	
6262	D5386	1.8		-0.85	
6315	ISO6271	2		-0.74	
6321	D1209	4		0.35	
normality		OK			
n		21			
outliers		0			
mean (n)		3.36			
st.dev. (n)		1.085			
R(calc.)		3.04			
st.dev.(D5386:16)		1.828			
R(D5386:16)		5.12			
Compare:					
R(D1209:05)		7			





Determination of Copper Corrosion on sample #20011;

lab	method	value	mark	z(targ)	remarks
52	D849	1a		----	
150	D849	1a		----	
171	D849	1a		----	
311	D849	pass		----	
323	D849	1A		----	
333		----		----	
334	D849	Pass		----	
343		----		----	
445	D849	1a		----	
551		----		----	
555		----		----	
663	D849	1a		----	
823	D849	1a		----	
840	D849	1a		----	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D849	1A		----	
913	D849	1a		----	
1040		----		----	
1041		----		----	
1062	D849	1B		----	
1065		----		----	
1079	D849	1A		----	
1151		----		----	
1530	D849	1a		----	
1741		----		----	
1852		----		----	
1982	D849	1A		----	
6203		----		----	
6262	D849	1A		----	
6315	DIN EN ISO 2160	1		----	
6321		----		----	
	n	18			
	mean (n)	1a/b (Pass)			

Determination of Density at 20°C on sample #20011: results in kg/L

lab	method	value	mark	z(targ)	remarks
52	D4052	0.8668		0.07	
150	D4052	0.8667		-0.49	
171	D4052	0.8668		0.07	
311	D4052	0.8668		0.07	
323	D4052	0.8667		-0.49	
333	ISO12185	0.8667		-0.49	
334	ISO12185	0.8669		0.63	
343	ISO12185	0.8668		0.07	
445	D4052	0.8668		0.07	
551		----		----	
555		----		----	
663	D4052	0.86676		-0.16	
823	ISO12185	0.86682		0.18	
840	D4052	0.86680		0.07	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D4052	0.8668		0.07	
913	D4052	0.8668		0.07	
1040	ISO12185	0.86683		0.24	
1041		----		----	
1062	D4052	0.8663	R(0.01)	-2.73	
1065		----		----	
1079	ISO12185	0.86682		0.18	
1151		----		----	
1530	ISO12185	0.86670		-0.49	
1741		----		----	
1852	ISO12185	0.8668		0.07	
1982	D4052	0.8668		0.07	
6203	ISO12185	0.86686		0.40	
6262	ISO12185	0.8669		0.63	
6315	ISO12185	0.86673		-0.32	
6321	ISO12185	0.8667		-0.49	
normality		OK			
n		23			
outliers		1			
mean (n)		0.86679			
st.dev. (n)		0.000060			
R(calc.)		0.00017			
st.dev.(ISO12185:96)		0.000179			
R(ISO12185:96)		0.0005			









Determination of Distillation on sample #20011; results in °C

lab	method	IBP	mark	z(targ)	50% rec	mark	z(targ)	DP	mark	z(targ)	range
52	D850-automated	110.2		0.04	110.6		0.34	111.0		0.78	0.8
150	D850-automated	110.1		-0.44	110.6		0.34	110.9		0.16	0.8
171	D850	109.1	G(0.01)	-5.26	110.6	ex	0.34	111.2	DG(0.01)	2.00	----
311	D850-automated	110.2		0.04	110.6		0.34	110.9		0.16	0.7
323	D850-manual	110.3		0.52	110.6		0.34	110.9		0.16	0.6
333	D850	110.3		0.52	110.6		0.34	110.9		0.16	0.6
334	D850-automated	110.1		-0.44	110.6		0.34	110.9		0.16	0.8
343	D850-automated	110.1		-0.44	110.6		0.34	110.9		0.16	0.8
445	D850-automated	110.2		0.04	110.6		0.34	110.9		0.16	0.7
551		----		----	----		----	----		----	----
555		----		----	----		----	----		----	----
663	D850-automated	110.2		0.04	110.6		0.34	110.9		0.16	0.7
823		----		----	----		----	----		----	----
840	D850-automated	110.16		-0.15	110.60		0.34	111.02		0.90	0.86
855		----		----	----		----	----		----	----
862		----		----	----		----	----		----	----
864		----		----	----		----	----		----	----
865		----		----	----		----	----		----	----
866		----		----	----		----	----		----	----
870		----		----	----		----	----		----	----
912	D1078-manual	110.2		0.04	110.6		0.34	110.8		-0.45	0.6
913		----		----	----		----	----		----	----
1040		----		----	----		----	----		----	----
1041		----		----	----		----	----		----	----
1062	D850-automated	108.7	G(0.05)	-7.19	110.4		-3.25	110.7		-1.07	2.0
1065		----		----	----		----	----		----	----
1079	D850-automated	110.1		-0.44	110.6		0.34	110.9		0.16	0.8
1151		----		----	----		----	----		----	----
1530	D850-automated	110.10		-0.44	110.30	G(0.01)	-5.05	110.70		-1.07	0.6
1741		----		----	----		----	----		----	----
1852		----		----	----		----	----		----	----
1982	D850-automated	110.41		1.05	110.60		0.34	110.93		0.35	0.52
6203	D850-manual	110.2		0.04	110.6		0.34	110.8		-0.45	0.6
6262	D850-automated	109.5	C,G(0.01)	-3.33	110.5	C	-1.46	110.8	C	-0.45	1.3
6315	D850-automated	110.8	ex	2.93	110.9	G(0.05)	5.72	111.3	DG(0.01)	2.62	0.5
6321		----		----	----		----	----		----	----
	normality	not OK			not OK			OK			
	n	15			16			17			
	outliers	3 +1ex			2 +1ex			2			
	mean (n)	110.19			110.58			110.87			
	st.dev. (n)	0.090			0.054			0.088			
	R(calc.)	0.25			0.15			0.25			
	st.dev.(D850-A:18)	0.208			0.056			0.163			
	R(D850-A:18)	0.58			0.16			0.46			
Compare:											
	R(D850-M:18)	0.41			0.65			0.65			

Lab 171 test result excluded for 50% recovered as statistical outliers in other related distillation parameters



Lab 6262 first reported 110.4 for IBP / 111.0 for 50% rec / 111.3 for DP / 0.9 for range

Lab 6315 test result excluded for IBP as statistical outliers in other related distillation parameters





Determination of Purity by GC on sample #20011; results in %M/M

lab	method	value	mark	z(targ)	remarks
52	D7504	99.78		-0.60	
150	D7504	99.77	C	-2.74	first reported 99.75
171	D7504	99.78		-0.60	
311	D7504	99.79		1.54	
323	D7504	99.79		1.54	
333	D7504	99.78		-0.60	
334	D2360	99.795		2.61	
343	D2360	99.77	C	-2.74	first reported 99.81
445	D6526	99.7770		-1.24	
551		----		----	
555		----		----	
663	D7504	99.780		-0.60	
823	D2360	99.7952		2.65	
840	D7504	99.780		-0.60	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D7504	99.80	C	3.68	first reported 99.82
913	D7504	99.80		3.68	
1040	D6526	99.7733		-2.04	
1041		99.818	R(0.05)	7.53	
1062	D2360	99.8030		4.32	
1065		----		----	
1079	D7504	99.8002		3.72	
1151		----		----	
1530	D7504	99.810	R(0.05)	5.82	
1741		----		----	
1852	D7504	99.7812		-0.35	
1982	D2360	99.766		-3.60	
6203	D7504	99.7687		-3.02	
6262	D7504	99.7729		-2.12	
6315	D7504	99.7727		-2.17	
6321	D2360	99.779		-0.82	
normality		OK			
n		23			
outliers		2			
mean (n)		99.7828			
st.dev. (n)		0.01148			
R(calc.)		0.0321			
st.dev.(D7504:20)		0.00467			
R(D7504:20)		0.0131			





Determination of Benzene on sample #20011; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	D7504	444		0.64	
150	D7504	466		1.68	
171	D7504	486		2.62	
311	D7504	440		0.45	
323	D7504	362	DG(0.05)	-3.21	
333	D7504	403		-1.29	
334	D2360	340	DG(0.05)	-4.25	
343	D2360	421	C	-0.44	first reported 410
445	D6526	404		-1.24	
551		----		----	
555		----		----	
663	D7504	431		0.03	
823	D2360	405		-1.19	
840	D7504	436.6		0.29	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D7504	420		-0.49	
913	D7504	439		0.41	
1040	D6526	445		0.69	
1041		408		-1.05	
1062	D2360	448		0.83	
1065		----		----	
1079	D7504	393		-1.76	
1151	In-house	395.38		-1.64	
1530	D7504	421		-0.44	
1741		----		----	
1852	D7504	428	C	-0.11	reported 0.0428 mg/kg
1982	D2360	432	C	0.08	first reported 0.0432 mg/kg
6203	D7504	423		-0.35	
6262	D7504	443		0.59	
6315	D7504	446		0.74	
6321	D2360	450.5		0.95	
	normality	OK			
	n	24			
	outliers	2			
	mean (n)	430.35			
	st.dev. (n)	22.502			
	R(calc.)	63.00			
	st.dev.(D7504:20)	21.266			
	R(D7504:20)	59.55			
	Compare				
	R(Horwitz)	77.39			





Determination of Nonaromatics on sample #20011; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	D7504	675		0.24	
150	D7504	786		0.79	
171	D7504	617		-0.04	
311	D7504	610		-0.08	
323	D7504	537		-0.44	
333	D7504	657		0.15	
334	D2360	580		-0.23	
343	D2360	722	C	0.48	first reported 308
445	D6526	662		0.18	
551		----		----	
555		----		----	
663	D7504	640		0.07	
823	D2360	567		-0.29	
840	D7504	666.9		0.20	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D7504	394		-1.15	
913	D7504	571		-0.27	
1040	D6526	677		0.25	
1041		554		-0.36	
1062	D2360	402		-1.11	
1065		----		----	
1079	D7504	585		-0.20	
1151		----		----	
1530	D7504	512		-0.56	
1741		----		----	
1852	D7504	667	C	0.20	first reported 0.065 mg/kg
1982	D2360	775	C	0.74	first reported 0.0775 mg/kg
6203	D7504	737		0.55	
6262	D7504	705		0.39	
6315	D7504	692		0.33	
6321	D2360	654.4		0.14	
normality		OK			
n		25			
outliers		0			
mean (n)		625.81			
st.dev. (n)		98.381			
R(calc.)		275.47			
st.dev.(D7504:20)		202.013			
R(D7504:20)		565.64			





Determination of Total Impurities on sample #20011; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	D7504	2200		0.19	
150	D7504	2500		2.14	
171		----		----	
311	D7504	2100		-0.45	
323	D7504	1991		-1.16	
333		----		----	
334	D2360	2050		-0.78	
343		----		----	
445	D6526	2230		0.39	
551		----		----	
555		----		----	
663		----		----	
823		----		----	
840	D7504	2201.4		0.20	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912		----		----	
913		----		----	
1040	D6526	2267		0.63	
1041		----		----	
1062	D2360	1971		-1.29	
1065		----		----	
1079	D7504	1996		-1.13	
1151		----		----	
1530	D7504	1933		-1.53	
1741		----		----	
1852	D7504	2188	C	0.12	reported 0.2188 mg/kg
1982	D2360	2340	C	1.10	first reported 0.2340 mg/kg
6203		----		----	
6262	D7504	2271		0.65	
6315	D7504	2275		0.68	
6321	D2360	2207.6		0.24	
normality		OK			
n		16			
outliers		0			
mean (n)		2170.06			
st.dev. (n)		153.416			
R(calc.)		429.56			
st.dev.(Horwitz 2 comp)		154.496			
R(Horwitz 2 comp)		432.59			





Determination of Refractive Index at 25°C on sample #20011;

lab	method	value	mark	z(targ)	remarks
52		----		----	
150	D1218	1.4938		-1.12	
171	D1218	1.4940		0.00	
311	D1218	1.4938		-1.12	
323	D1218	1.4939		-0.56	
333	D1218	1.4940		0.00	
334	D1218	1.4941		0.56	
343	D1218	1.4968	C,G(0.01)	15.68	first reported 1.4969
445	D1218	1.4946		3.36	
551		----		----	
555		----		----	
663	D1218	1.4943		1.68	
823	D1218	1.49380		-1.12	
840	D1218	1.49388		-0.68	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	D1218	1.4940		0.00	
913		----		----	
1040		----		----	
1041		----		----	
1062		----		----	
1065		----		----	
1079	D1218	1.49400		0.00	
1151		----		----	
1530	D1218	1.49430		1.68	
1741		----		----	
1852		1.4940		0.00	
1982	DIN 51423-1	1.49413		0.72	
6203	D1218	1.4933		-3.92	
6262	D1218	1.4941		0.56	
6315		----		----	
6321		----		----	
normality		not OK			
n		17			
outliers		1			
mean (n)		1.49400			
st.dev. (n)		0.000275			
R(calc.)		0.00077			
st.dev.(D1218:12)		0.000179			
R(D1218:12)		0.0005			



Determination of Water on sample #20011; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	D6304	75		-1.18	
150	E1064	89		1.89	
171	E1064	89		1.89	
311	E1064	80		-0.08	
323	E1064	76		-0.96	
333		----		----	
334	E1064	77		-0.74	
343		----		----	
445	E1064	74.3		-1.33	
551		----		----	
555		----		----	
663		----		----	
823	E1064	76		-0.96	
840	E1064	92.2		2.59	
855		----		----	
862		----		----	
864		----		----	
865		----		----	
866		----		----	
870		----		----	
912	E203	110	C,DG(0.05)	6.49	first reported 118
913		----		----	
1040	DIN 51777	88		1.67	
1041		----		----	
1062	D6304	92		2.54	
1065		----		----	
1079	E1064	73.35		-1.54	
1151		----		----	
1530	E1064	76.1		-0.94	
1741		----		----	
1852		102	DG(0.05)	4.74	
1982	E1064	79		-0.30	
6203	D6304	82		0.35	
6262	E1064	73		-1.62	
6315	DIN EN ISO12937	74.6		-1.27	
6321	E1064	130	G(0.01)	10.87	
normality		OK			
n		17			
outliers		3			
mean (n)		80.39			
st.dev. (n)		6.880			
R(calc.)		19.26			
st.dev.(D1064:16)		4.565			
R(D1064:16)		12.78			



APPENDIX 2

Number of participants per country

4 labs in BELGIUM
2 labs in BRAZIL
1 lab in CANADA
6 labs in CHINA, People's Republic
2 labs in FRANCE
6 labs in GERMANY
2 labs in INDIA
1 lab in ISRAEL
1 lab in NETHERLANDS
1 lab in ROMANIA
1 lab in SAUDI ARABIA
1 lab in SERBIA
1 lab in SOUTH KOREA
1 lab in SPAIN
1 lab in THAILAND
2 labs in UNITED KINGDOM
2 labs in UNITED STATES OF AMERICA
1 lab in VIETNAM

APPENDIX 3

Abbreviations

C	= final test result after checking of first reported suspect result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= possibly an error in calculations
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
SDS	= Safety Data Sheet

Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 W. Horwitz and R. Albert, J. AOAC Int., 79, 3, 589, (1996)
- 3 ASTM E178:02
- 4 ASTM E1301:03
- 5 ISO13528:05
- 6 ISO5725:86
- 7 ISO5725, parts 1-6, 1994
- 8 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 9 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 10 IP367:84
- 11 DIN38402 T41/42
- 12 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 13 J.N. Miller, Analyst, 118, 455, (1993)
- 14 Analytical Methods Committee, Technical brief, No 4, January 2001.
- 15 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry 2002, 127, 1359-1364 (2002)
- 16 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)