

**Results of Proficiency Test
Engine Oil (fresh)
June 2021**

Organized by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: ing. M. Meijer
Correctors: ing. A.S. Noordman-de Neef & ing. G.A. Oosterlaken-Buijs
Report: iis21L07

September 2021

CONTENTS

1	INTRODUCTION	3
2	SET UP.....	3
2.1	ACCREDITATION.....	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALITY STATEMENT	3
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES	4
2.6	ANALYZES	5
3	RESULTS.....	5
3.1	STATISTICS	5
3.2	GRAPHICS	6
3.3	Z-SCORES.....	7
4	EVALUATION	7
4.1	EVALUATION PER TEST	8
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	11
4.3	COMPARISON OF THE PROFICIENCY TEST OF JUNE 2021 WITH PREVIOUS PTS	12

Appendices:

1.	Data, statistical and graphic results.....	14
2.	Number of participants per country	68
3.	Abbreviations and literature	69

1 INTRODUCTION

Since 1997 the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the analysis of fresh Engine Oil (Lubricating Oil) in accordance with the latest version of SAE and ASTM D4485 every year. During the annual proficiency testing program 2020/2021 it was decided to continue the round robin for the analysis of Engine Oil (fresh).

In this interlaboratory study 79 laboratories in 43 countries registered for participation. See appendix 2 for the number of participants per country. In this report the results of the Engine Oil (fresh) proficiency test are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory.

It was decided to send one 1L bottle and one 0.5L bottle with the same Engine Oil (fresh) and both bottles labelled #21075.

The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

A batch of approximately 180 liters fresh Diesel Engine Oil (monograde) was obtained from a third supplier. After homogenization 124 amber glass bottles of 1L and 124 amber glass bottles of 0.5L were filled and all bottles were labelled #21075.

The homogeneity of the subsamples was checked by determination of Density at 15°C in accordance with ISO12185 on 8 stratified randomly selected subsamples.

	Density at 15°C in kg/L
Sample #21075-1	0.88413
Sample #21075-2	0.88413
Sample #21075-3	0.88413
Sample #21075-4	0.88413
Sample #21075-5	0.88413
Sample #21075-6	0.88413
Sample #21075-7	0.88413
Sample #21075-8	0.88413

Table 1: homogeneity test results of subsamples #21075

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15°C in kg/L
r (observed)	0.00000
reference test method	ISO12185:96
0.3 x R (reference test method)	0.00015

Table 2: evaluation of the repeatability of subsamples #21075

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

To each of the participating laboratories a set of a 1L and a 0.5L bottle of Engine Oil (fresh) both items labelled #21075 were sent on May 12, 2021. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of the fresh Engine Oil packed in amber glass bottles was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYZES

The participants were requested to determine: Total Acid Number, Base Number (HClO_4 titration), Color ASTM, Conradson Carbon Residue, Ramsbottom Carbon Residue, Carbon Residue (Micro method), Density at 15°C, Evaporation loss by Noack, Flash Point C.O.C., Flash Point PMcc, Foaming Tendency, Foam Stability, Kinematic Viscosity (40°C and 100°C), Viscosity Index, Kinematic Viscosity Stabinger (40°C and 100°C), Viscosity HTHS, Nitrogen, Pour Point (Manual and Automated), Sulfated Ash, Sulfur, Water, Calcium, Phosphorus and Zinc.

Some additional questions were asked about Total Acid Number and Foaming Characteristics.

It was explicitly requested to treat the sample as if it was a routine sample and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as ‘<...’ or ‘>...’ were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either ‘unknown’, ‘OK’, ‘suspect’ or ‘not OK’. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

The assigned value is determined by consensus based on the test results of the group of participants after rejection of the statistical outliers and/or suspect data.

According to ISO13528 all (original received or corrected) results per determination were submitted to outlier tests. In the iis procedure for proficiency tests, outliers are detected prior to calculation of the mean, standard deviation and reproducibility. For small data sets, Dixon (up to 20 test results) or Grubbs (up to 40 test results) outlier tests can be used. For larger data sets (above 20 test results) Rosner’s outlier test can be used. Outliers are marked by D(0.01) for the Dixon’s test, by G(0.01) or DG(0.01) for the Grubbs’ test and by R(0.01) for the Rosner’s test. Stragglers are marked by D(0.05) for the Dixon’s test, by G(0.05) or DG(0.05) for the Grubbs’ test and by R(0.05) for the Rosner’s test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms.

Also, a normal Gauss curve (dotted line) was projected over the Kernel Density Graph (smooth line) for reference. The Gauss curve is calculated from the consensus value and the corresponding standard deviation.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements, e.g. ISO reproducibilities, the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The $z_{(\text{target})}$ scores are listed in the test result tables in appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this proficiency test some problems were encountered with the dispatch of the samples. Seven participants reported test results after the final reporting date and three other participants did not report any test results. Not all participants were able to report all tests requested.

In total 76 participants reported 1156 numerical test results. Observed were 36 outlying test results, which is 3.1%. In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

Not all data sets proved to have a normal Gaussian distribution. These are referred to as "not OK" or "suspect". The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

4.1 EVALUATION PER TEST

In this section the reported test results are discussed per test. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the original data. The abbreviations, used in these tables, are explained in appendix 3.

In the iis PT reports ASTM test methods are referred to with a number (e.g. D2270) and an added designation for the year that the test method was adopted or revised (e.g. D2270:10). If applicable, a designation in parentheses is added to designate the year of reapproval (e.g. D2270:10(2016)). In the test results tables of appendix 1 only the method number and year of adoption or revision (e.g. D2270:10) will be used.

Total Acid Number: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D664-A:18e2 Inflection Point 60mL nor with the other end point modes, except for BEP at 60mL which is in agreement. When evaluated separately for the type of end point the calculated reproducibility of the group using Inflection Point was still not in agreement with the requirements of D664-A:18e2 at both titration volumes, but the group using BEP (pH 10 and 11) is in agreement at both titration volumes.

Base Number (HClO₄ titration): This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ASTM D2896-A:15, forward mode.

Color ASTM: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D1500:12(2017).

Conradson Carbon Residue: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D189:06(2019).

Ramsbottom Carbon Residue: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D524:15(2019).

Carbon Residue (Micro method): This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D4530:15(2020).

Density at 15°C: This determination was problematic for a number of laboratories. Five statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in full agreement with the requirements of ISO12185:96.

Evaporation loss by Noack: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ASTM D5800:20 procedure B, but is in agreement with procedure A.

When evaluated separately for the test results of procedure A and B of ASTM D5800:20 the calculated reproducibility after rejection of the statistical outlier is still not in agreement with the requirements of procedure B.

Flash Point C.O.C.: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D92:18.

Flash Point PMcc: This determination was not problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D93-A:20.

Foaming Tendency: This determination was problematic. Six statistical outliers were observed over the three sequences. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D892:18 for all three sequences.

Foam Stability: This determination was not problematic. All reporting participants agreed on a test result of 0 (Nil). Therefore, no z-scores were calculated.

Kinematic Viscosity at 40°C: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D445:21.

Kinematic Viscosity at 100°C: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D445:21.

Viscosity Index: This determination was not problematic. No statistical outliers were observed but two test results were excluded. The calculated reproducibility after rejection of the suspect data is in full agreement with the requirements of ASTM D2270:10(2016).

Kinematic Viscosity Stabinger at 40°C: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D7042:21.

Kinematic Viscosity Stabinger at 100°C: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D7042:21.

Viscosity HTHS: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D4683:17.

Nitrogen: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D5762:18a.

Pour Point Manual: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D97:17b.

Pour Point Automated: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ASTM D5950:14(2020).

Sulfated Ash: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D874:13a(2018).

Sulfur: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D4294:16e1.

Water: This determination was problematic. Two statistical outliers were observed. A new version of ASTM D6304 was published in 2020 with major changes. In the 2016 version one precision statement was mentioned for test results based on mass with a broad application range and one based on volume. In the 2020 version all precision statements are based on mass with three different procedures (A - direct injection, B - oven accessory and C - evaporation accessory) each with a different application range. In ASTM D6304:20 the reproducibility for all three procedures A, B and C is much stricter compared to ASTM D6304:16e1. Therefore, it was decided not to calculate z-scores due to a large variation in the test results compared with the stricter requirements.

Calcium: This determination was problematic. Six statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D5185:18.

Phosphorus: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D5185:18.

Zinc: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of ASTM D5185:18.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility ($2.8 * \text{standard deviation}$) and the target reproducibility derived from literature reference test methods (in casu ASTM and ISO test methods) are presented in the next table.

Parameter	unit	n	average	$2.8 * \text{sd}$	R(lit)
Total Acid Number	mg KOH/g	45	2.41	1.13	0.83
Base Number (HClO ₄ titration)	mg KOH/g	49	7.78	0.55	0.54
Color ASTM		57	2.9	0.6	1
Conradson Carbon Residue	%M/M	17	0.94	0.16	0.23
Ramsbottom Carbon Residue	%M/M	5	0.88	0.16	0.13
Carbon Residue (Micro method)	%M/M	24	0.95	0.10	0.19
Density at 15°C	kg/L	60	0.8841	0.0005	0.0005
Evaporation loss by Noack	%M/M	18	5.16	0.91	0.82
Flash Point C.O.C.	°C	57	244.6	22.3	18
Flash Point PMcc	°C	47	201.1	5.9	14.3
Foaming Tendency, Sequence I	mL	29	0.5	3.9	0.6
Foaming Tendency, Sequence II	mL	34	15.2	23.0	18.7
Foaming Tendency, Sequence III	mL	29	0.3	2.8	1.1
Foam Stability, Sequence I	mL	34	0	n.e.	n.e.
Foam Stability, Sequence II	mL	34	0	n.e.	n.e.
Foam Stability, Sequence III	mL	34	0	n.e.	n.e.
Kinematic Viscosity at 40°C	mm ² /s	62	131.04	1.81	1.60
Kinematic Viscosity at 100°C	mm ² /s	59	14.27	0.14	0.20
Viscosity Index		62	107.4	2.0	2
Kin. Viscosity Stabinger at 40°C	mm ² /s	23	131.32	1.77	1.54
Kin. Viscosity Stabinger at 100°C	mm ² /s	21	14.31	0.08	0.14
Viscosity HTHS	mPa·s	6	4.25	0.23	0.18
Nitrogen	mg/kg	14	809	254	215
Pour Point Manual	°C	37	-11.0	5.1	9
Pour Point Automated 1°C int.	°C	21	-11.6	5.4	4.5
Sulfated Ash	%M/M	35	0.83	0.22	0.12
Sulfur	mg/kg	36	4812	622	453
Water	mg/kg	38	242	399	(91)
Calcium as Ca	mg/kg	44	501	72	49
Phosphorus as P	mg/kg	52	1000	161	136
Zinc as Zn	mg/kg	53	1100	206	184

Table 3: reproducibilities of tests on sample #21075

Results between brackets should be used with due care

Without further statistical calculations it can be concluded that for many tests there is not a good compliance of the group of participants with the reference test methods. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF JUNE 2021 WITH PREVIOUS PTS

	June 2021	June 2020	June 2019	June 2018	June 2017
Number of reporting laboratories	76	62	75	81	67
Number of test results	1156	961	1157	1337	940
Number of statistical outliers	36	34	49	37	45
Percentage of statistical outliers	3.1%	3.5%	4.2%	2.8%	4.8%

Table 4: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared to the requirements of the reference test methods. The conclusions are given in the following table.

Determination	June 2021	June 2020	June 2019	June 2018	June 2017
Total Acid Number	-	--	--	-	--
Base Number (HClO ₄ titration)	+/-	+	+	-	-
Color ASTM	+	+	+	+	-
Conradson Carbon Residue	+	+	+	+/-	+
Ramsbottom Carbon Residue	-	n.e.	(--)	+	+
Carbon Residue (Micro method)	+	+	-	+	+/-
Density at 15°C	+/-	+	+/-	+/-	+
Evaporation loss by Noack	-	+	--	-	+
Flash Point C.O.C.	-	+/-	+/-	+	-
Flash Point PMcc	++	+	+	+	+/-
Foaming Tendency	-	-	+/-	+/-	+/-
Kinematic Viscosity at 40°C	-	+	+	+	+
Kinematic Viscosity at 100°C	+	+	+/-	+	+
Viscosity Index	+/-	-	-	+/-	+/-
Kin. Viscosity Stabinger at 40°C	-	+/-	+/-	+/-	-
Kin. Viscosity Stabinger at 100°C	+	+	+/-	-	-
Viscosity HTS	-	-	+/-	+	+
Nitrogen	-	-	--	-	-
Pour Point Manual	+	+/-	+	+/-	-
Pour Point Automated 1°C int.	-	--	+	+/-	-
Sulfated Ash	-	+/-	-	-	+
Sulfur	-	-	(--)	-	-
Water	(--)	+	+	+	+/-

Determination	June 2021	June 2020	June 2019	June 2018	June 2017
Calcium as Ca	-	-	+	-	+
Phosphorus as P	-	-	-	-	n.e.
Zinc as Zn	-	+	+/-	+/-	n.e.

Table 5: comparison determinations against the reference test methods

Results between brackets: no z-scores were calculated



The following performance categories were used:

- ++ : group performed much better than the reference test method
- + : group performed better than the reference test method
- +/- : group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e. : not evaluated

APPENDIX 1**Determination of Total Acid Number on sample #21075; results in mg KOH/g**

lab	method	value	mark	z(targ)	end point type	titration volume	remarks
178		----		----	---	---	
179	D664-A	2.38		-0.10	Inflection Point	60 mL	
211		----		----	---	---	
219		----		----	---	---	
237	D664-A	2.3314		-0.26	---	---	
254		----		----	---	---	
257		----		----	---	---	
309	D664-A	2.12		-0.98	---	---	
325	D664-A	2.74		1.12	Buffer End Point (pH 11)	125 mL	
329	D664-A	2.84		1.46	Inflection Point	125 mL	
333	D664-A	2.4		-0.03	Inflection Point	125 mL	
339		----		----	---	---	
349	D664-A	2.77		1.22	Buffer End Point (pH 10)	125 mL	
360	D664-A	2.380		-0.10	Inflection Point	60 mL	
398		----		----	---	---	
421	ISO6619	2.6		0.65	---	---	
432		----		----	---	---	
496	D664-A	2.09		-1.08	Buffer End Point (pH 10)	60 mL	
614		----		----	---	---	
633	D664-A	2.64		0.78	Inflection Point	125 mL	
634	D664-A	1.93		-1.62	---	---	
657	D664-A	2.22	C	-0.64	Inflection Point	125 mL	fr. 0.55
780	D664-A	2.56		0.51	Buffer End Point (pH 10)	60 mL	
823	D664-A	2.35		-0.20	Inflection Point	125 mL	
840	D664-B	2.10		-1.05	Inflection Point	60 mL	
862	D664-A	2.59		0.61	Inflection Point	60 mL	
875	D664-A	2.29		-0.40	---	---	
902	D664-A	2.692		0.96	Inflection Point	60 mL	
912	D664-A	3.5		3.69	---	---	
913		----		----	---	---	
922	D664-A	2.65		0.82	Inflection Point	125 mL	
962		----		----	---	---	
963	D664-A	3.12		2.41	Inflection Point	60 mL	
974		----		----	---	---	
994	D664-A	2.52		0.38	---	125 mL	
1011		----		----	---	---	
1017		----		----	---	---	
1026	D664-A	1.94		-1.59	Buffer End Point (pH 10)	125 mL	
1047	ISO6619	2.6		0.65	Inflection Point	125 mL	
1059	ISO6619	2.61		0.68	Buffer End Point (pH 11)	60 mL	
1146	D664-A	2.741		1.12	Buffer End Point (pH 10)	125 mL	
1150		----		----	---	---	
1173		----		----	---	---	
1213		----		----	---	---	
1235		----		----	---	---	
1243	ISO6618	1.95		-1.55	---	---	
1316	D664-A	2.27		-0.47	Buffer End Point (pH 10)	60 mL	
1320		----		----	---	---	
1326		----		----	---	---	
1328	D664-A	2.695		0.97	Inflection Point	125 mL	
1384	ISO6618	1.645		-2.59	---	---	
1412	D664-A	2.48		0.24	Buffer End Point (pH 10)	125 mL	
1431		----		----	---	---	
1433	D664-A	2.4665		0.20	Inflection Point	125 mL	
1438		2.119		-0.98	---	---	
1460	D664-A	1.5605		-2.87	Inflection Point	60 mL	
1510		----		----	---	---	
1564	D664-A	2.30		-0.37	Inflection Point	60 mL	
1650	D664-A	2.94		1.80	Inflection Point	125 mL	
1740	D664-A	2.59		0.61	---	---	
1748		----		----	---	---	
1799		----		----	---	---	
1850		----		----	---	---	
1864	D664-A	2.43		0.07	Inflection Point	60 mL	
1877		----		----	---	---	
1957	D664-A	2.485		0.26	Buffer End Point (pH 11)	125 mL	
1969		----		----	---	---	
6016	D664-A	1.20		-4.09	Inflection Point	60 mL	
6032		----		----	---	---	
6035		----		----	---	---	
6181		----		----	---	---	
6183		----		----	---	---	
6197	D664-A	2.25		-0.54	Inflection Point	60 mL	
6266		----		----	---	---	



lab	method	value	mark	z(targ)	end point type	titration volume	remarks
6310	D664-B	2.39		-0.06	Buffer End Point (pH 10)	60 mL	
6324		----		----	---		---
6358	D664-A	2.141		-0.91	Buffer End Point (pH 10)	125 mL	
6380	D664-A	2.7833		1.27	Inflection Point	125 mL	
6395		----		----	---		---
normality		suspect			Inflection Point only	BEP only	
n		45			not OK	OK	
outliers		0			22	12	
mean (n)		2.4089			0	0	
st.dev. (n)		0.40271			2.4358	2.4348	
R(calc.)		1.1276			0.42534	0.27459	
st.dev.(D664-A:18e2, IP 60mL)		0.29535			1.1909	0.7689	
R(D664-A:18e2, IP 60mL)		0.8270			0.29805	---	
Compare					0.8345	---	
R(D664-A:18e2, IP 125mL)		0.5474			0.5538	---	
R(D664-A:18e2, BEP 60mL)		1.3069			---	1.3206	
R(D664-A:18e2, BEP 125mL)		0.7963			---	0.8053	



Determination of Base Number (HClO₄ titration) on sample #21075; results in mg KOH/g

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211	D2896-B forward	8.1		1.66	
219		----		----	
237	D2896-A back	7.993		1.11	
254		----		----	
257		----		----	
309	D2896-A forward	7.42		-1.84	
325	D2896-B forward	7.7		-0.40	
329	D2896-A forward	7.76		-0.09	
333		----		----	
339		----		----	
349		----		----	
360	D2896-B forward	7.84		0.32	
398		----		----	
421	ISO3771	7.8		0.12	
432	D2896-B back	7.5		-1.43	
496	D2896-B back	7.75		-0.14	
614	D2896-B forward	7.75		-0.14	
633	D2896-A forward	7.78		0.01	
634	D2896-B forward	7.8		0.12	
657	D2896-B forward	7.5		-1.43	
780	D2896-B forward	7.70		-0.40	
823	D2896-A back	7.8		0.12	
840	D2896-B forward	7.88		0.53	
862	D2896-B forward	7.7		-0.40	
875		----		----	
902	D2896-B forward	8.24		2.38	
912	D2896	8.1		1.66	
913		----		----	
922	D2896-B forward	8.0		1.15	
962		----		----	
963	D2896-A forward	7.42		-1.84	
974	D2896-A back	7.6		-0.91	
994		----		----	
1011	D2896	8.0		1.15	
1017		----		----	
1026	D2896-B back	7.60		-0.91	
1047	PN-76/C04163	7.63		-0.76	
1059	ISO3771	7.9		0.63	
1146	D2896-A forward	7.753		-0.13	
1150	BDS13727	7.6482		-0.66	
1173	In house	8.0		1.15	
1213	D2896-B forward	7.76		-0.09	
1235		----		----	
1243		----		----	
1316	D4739	6.9	R(0.01)	-4.51	
1320	D2896-B forward	7.60		-0.91	
1326	D2896-B forward	7.89		0.58	
1328	D2896	7.63		-0.76	
1384	ISO3771	7.34		-2.25	
1412		----		----	
1431		----		----	
1433	D2896-B forward	7.936		0.82	
1438		8.288		2.63	
1460	D2896	7.560		-1.12	
1510		----		----	
1564	D2896-B back	7.72		-0.29	
1650	D2896-A forward	7.85		0.37	
1740	ISO3771	7.95		0.89	
1748	D2896-A forward	7.65		-0.65	
1799		----		----	
1850		----		----	
1864	D2896-B forward	7.77		-0.04	
1877		----		----	
1957	D2896-A back	7.830		0.27	
1969		----		----	
6016		----		----	
6032	D2896-B forward	7.77		-0.04	
6035		----		----	
6181	D2896-A forward	7.81		0.17	
6183	D2896-A forward	7.78		0.01	
6197	D2896-B forward	7.70		-0.40	
6266	ISO3771	6.918	R(0.01)	-4.42	
6310	D2896-B back	7.86		0.43	

lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	D2896-A forward	7.73		-0.24	
6380		----		----	
6395		----		----	
	normality	OK			
	n	49			
	outliers	2			
	mean (n)	7.777			
	st.dev. (n)	0.1960			
	R(calc.)	0.549			
	st.dev.(D2896-A:15 forward)	0.1944			
	R(D2896-A:15 forward)	0.544			





Determination of Color ASTM on sample #21075

lab	method	value	iis conversion *	mark	z(targ)	remarks
178		----	----		----	
179	D1500	L3.0	2.75		-0.47	
211	D1500	L3.0	2.75		-0.47	
219	D1500	3.0	3.0		0.23	
237	D1500	3.0	3.0		0.23	
254	D1500	2.5	2.5		-1.17	
257		----	----		----	
309	D1500	<3.0	2.75		-0.47	
325	D6045	L3.0	2.75		-0.47	
329	D1500	3.2	3.2		0.79	
333		----	----		----	
339		----	----		----	
349	D6045	L3.0	2.75		-0.47	
360	ISO2049	3.0	3.0		0.23	
398	D1500	L3.0	2.75		-0.47	
421		----	----		----	
432	ISO2049	L3.0	2.75		-0.47	
496	D1500	3.0	3.0		0.23	
614	D1500	<3.0	2.75		-0.47	
633	D1500	L2.5	2.25		-1.87	
634	D1500	3.0	3.0		0.23	
657	D1500	3.0	3.0		0.23	
780	D1500	3.0	3.0		0.23	
823	D1500	3	3		0.23	
840	D1500	3.5	3.5		1.63	
862	D1500	L3.0	2.75		-0.47	
875	D6045	2.9	2.9		-0.05	
902	D1500	2.9	2.9		-0.05	
912		----	----		----	
913		----	----		----	
922	D1500	L3.0	2.75		-0.47	
962	D1500	2.9	2.9		-0.05	
963	D1500	2.9	2.9		-0.05	
974	D1500	L3.0	2.75		-0.47	
994	D1500	3.0	3.0		0.23	
1011	D6045	L3.0	2.75		-0.47	
1017		----	----		----	
1026	D1500	3.0	3.0		0.23	
1047		----	----		----	
1059	D1500	3.0	3.0		0.23	
1146		----	----		----	
1150		----	----		----	
1173		----	----		----	
1213		----	----		----	
1235	ISO2049	L3.0	2.75		-0.47	
1243	ISO2049	3.0	3.0		0.23	
1316	D1500	3.0	3.0		0.23	
1320	D1500	3	3		0.23	
1326	D1500	3.0	3.0		0.23	
1328	D1500	3.0	3.0		0.23	
1384	ISO2049	2.5	2.5		-1.17	
1412	D1500	3	3		0.23	
1431	D1500	3.0	3.0		0.23	
1433	D6045	3.0	3.0		0.23	
1438		----	----		----	
1460	D6045	3.0	3.0		0.23	
1510	D1500	L3.0	2.75		-0.47	
1564	D1500	L3.5	3.25		0.93	
1650		----	----		----	
1740	D1500	3	3		0.23	
1748	D1500	2.5	2.5		-1.17	
1799		----	----		----	
1850	D1500	L3.0	2.75		-0.47	
1864		----	----		----	
1877	D6045	3.0	3.0		0.23	
1957	D1500	3.0	3.0		0.23	
1969		----	----		----	
6016	D1500	3	3		0.23	
6032	D1500	3.0	3.0		0.23	
6035	D6045	3.8	3.8	C,R(0.01)	2.47	first reported 5.4
6181	D1500	3.4	3.4		1.35	
6183	D1500	3.1	3.1		0.51	
6197	D1500	3.0	3.0		0.23	
6266	D1500	4.0	4.0	R(0.01)	3.03	

lab	method	value	iis conversion *	mark	z(targ)	remarks
6310	D1500	L3.5	3.25		0.93	
6324		----	----		----	
6358	D1500	3.0	3.0		0.23	
6380	D1500	3.0	3.0		0.23	
6395		----	----		----	
normality		suspect				
n		57				
outliers		2				
mean (n)		2.92				
st.dev. (n)		0.211				
R(calc.)		0.59				
st.dev.(D1500:12)		0.357				
R(D1500:12)		1				



*) In the calculation of the mean, standard deviation and the reproducibility in this column, a reported value of 'L y' is changed tot y-0.25
(for example L3.5 into 3.25)



Determination of Conradson Carbon Residue on sample #21075; results in %M/M

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D189	1.02		0.98	
211		----		----	
219		----		----	
237	D189	1.024		1.03	
254	D189	0.867		-0.91	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D189	0.97		0.36	
398		----		----	
421	ISO6615	0.88		-0.75	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634	D189	0.925		-0.20	
657	D189	0.96		0.24	
780	D189	0.95		0.11	
823	D189	1.00		0.73	
840	D189	0.864		-0.95	
862		----		----	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922	D189	0.95		0.11	
962		----		----	
963		----		----	
974	D189	0.90		-0.50	
994	D189	0.953		0.15	
1011		----		----	
1017		----		----	
1026		----		----	
1047		----		----	
1059		----		----	
1146		----		----	
1150	ISO6615	0.88		-0.75	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320		----		----	
1326		----		----	
1328		----		----	
1384		----		----	
1412	D189	0.90		-0.50	
1431		----		----	
1433		----		----	
1438		----		----	
1460		----		----	
1510		----		----	
1564		----		----	
1650	D189	0.920		-0.26	
1740		----		----	
1748		----		----	
1799		----		----	
1850	D189	1.03		1.11	
1864		----		----	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197		----		----	
6266		----		----	


lab	method	value	mark	z(targ)	remarks
6310		----		----	
6324		----		----	
6358		----		----	
6380		----		----	
6395		----		----	
	normality	OK			
	n	17			
	outliers	0			
	mean (n)	0.941			
	st.dev. (n)	0.0553			
	R(calc.)	0.155			
	st.dev.(D189:06)	0.0807			
	R(D189:06)	0.226			



Determination of Ramsbottom Carbon Residue on sample #21075; results in %M/M

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D524	0.81		-1.46	
211		----		----	
219		----		----	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360		----		----	
398		----		----	
421		----		----	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634		----		----	
657	D524	0.84		-0.80	
780		----		----	
823		----		----	
840		----		----	
862		----		----	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974	D524	0.86		-0.35	
994		----		----	
1011	D524	0.95		1.64	
1017		----		----	
1026	D524	0.92	C	0.97	first reported 0.03
1047		----		----	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320		----		----	
1326		----		----	
1328		----		----	
1384		----		----	
1412		----		----	
1431		----		----	
1433		----		----	
1438		----		----	
1460		----		----	
1510		----		----	
1564		----		----	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864		----		----	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197		----		----	
6266		----		----	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		-----		-----	
6358		-----		-----	
6380		-----		-----	
6395		-----		-----	
	normality	unknown			
n		5			
outliers		0			
mean (n)		0.876			
st.dev. (n)		0.0577			
R(calc.)		0.162			
st.dev.(D524:15)		0.04516			
R(D524:15)		0.126			



Determination of Carbon Residue (Micro method) on sample #21075; results in %M/M

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237	D4530	1.02		1.01	
254		----		----	
257		----		----	
309		----		----	
325	D4530	0.978		0.38	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D4530	0.95		-0.05	
398		----		----	
421		----		----	
432		----		----	
496	D4530	0.9795		0.40	
614		----		----	
633		----		----	
634		----		----	
657	D4530	0.96		0.11	
780	D4530	0.95		-0.05	
823	ISO10370	1.00		0.71	
840		----		----	
862	D4530	0.94		-0.20	
875	D4530	0.875		-1.18	
902	D4530	0.949		-0.06	
912		----		----	
913		----		----	
922	D4530	0.93		-0.35	
962		----		----	
963		----		----	
974	D4530	0.96		0.11	
994		----		----	
1011		----		----	
1017		----		----	
1026	D4530	0.99		0.56	
1047	ISO10370	0.95		-0.05	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213	D4530	<0.1		<-12.91	possibly a false negative test result?
1235		----		----	
1243		----		----	
1316	D4530	0.98		0.41	
1320	D4530	0.98		0.41	
1326		----		----	
1328		----		----	
1384		----		----	
1412		----		----	
1431	D4530	0.915		-0.57	
1433	D4530	0.96		0.11	
1438		----		----	
1460	D524	1.0		0.71	
1510	D4530	0.290	R(0.01)	-10.03	
1564	D4530	0.93		-0.35	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864	ISO10370	0.88		-1.10	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035	ISO10370	0.89		-0.95	
6181		----		----	
6183		----		----	
6197		----		----	
6266		----		----	

lab	method	value	mark	z(targ)	remarks
6310		----		----	
6324		----		----	
6358	ISO10370	0.93		-0.35	
6380	D4530	0.9754		0.34	
6395		----		----	
	normality	OK			
	n	24			
	outliers	1			
	mean (n)	0.953			
	st.dev. (n)	0.0375			
	R(calc.)	0.105			
	st.dev.(D4530:15)	0.0661			
	R(D4530:15)	0.185			





Determination of Density at 15°C on sample #21075; results in kg/L

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D1298	0.8850	R(0.01)	4.81	
211	D4052	0.8839		-1.35	
219	D1298	0.8838		-1.91	
237	D4052	0.8845		2.01	
254	D4052	0.8844		1.45	
257		----		----	
309	D4052	0.88414		0.00	
325	D4052	0.8842		0.33	
329	D4052	0.8842		0.33	
333	D4052	0.8840		-0.79	
339		----		----	
349	D4052	0.8841		-0.23	
360	ISO12185	0.8841		-0.23	
398	ISO12185	0.8841		-0.23	
421	ISO12185	0.8846		2.57	
432	D4052	0.8844		1.45	
496	ISO12185	0.88409		-0.28	
614	D4052	0.8841		-0.23	
633	D1298	0.8838		-1.91	
634	D4052	0.8838		-1.91	
657	D4052	0.8844		1.45	
780	ISO12185	0.8841		-0.23	
823	D4052	0.88414		0.00	
840	D4052	0.88395		-1.07	
862	D4052	0.8842		0.33	
875	D4052	0.8844		1.45	
902	D4052	0.8841		-0.23	
912	ISO12185	0.8842		0.33	
913		----		----	
922	D4052	0.8841		-0.23	
962		----		----	
963	D4052	0.8841		-0.23	
974	D4052	0.8842		0.33	
994	ISO12185	0.8842		0.33	
1011	D4052	0.8843		0.89	
1017		----		----	
1026	D4052	0.8841		-0.23	
1047	ISO12185	0.88409		-0.28	
1059	ISO12185	0.8841		-0.23	
1146	D4052	0.8841		-0.23	
1150	ISO12185	0.88414		0.00	
1173		----		----	
1213	D4052	0.88415		0.05	
1235	ISO12185	0.884105		-0.20	
1243	ISO12185	0.8842		0.33	
1316	D4052	0.8845		2.01	
1320	ISO12185	0.88477	R(0.05)	3.52	
1326	D4052	0.8841	C	-0.23	reported 884.1 kg/L
1328		----		----	
1384	ISO3675	0.884		-0.79	
1412	D4052	0.8841		-0.23	
1431	D4052	0.88321	R(0.01)	-5.21	
1433	ISO12185	0.8842		0.33	
1438		0.880	R(0.01)	-23.19	
1460	D4052	0.88381		-1.85	
1510	ISO12185	0.8850	R(0.01)	4.81	
1564	D4052	0.8841		-0.23	
1650	D4052	0.8842		0.33	
1740		----		----	
1748	D4052	0.8839		-1.35	
1799		----		----	
1850	D4052	0.8841		-0.23	
1864	ISO12185	0.88417		0.16	
1877	D4052	0.88400		-0.79	
1957		----		----	
1969		----		----	
6016	D4052	0.8841		-0.23	
6032	D4052	0.88433		1.06	
6035	ISO12185	0.8842	C	0.33	first reported 884.7 kg/m³
6181	ISO12185	0.8841		-0.23	
6183	D4052	0.884161		0.11	
6197	D4052	0.8841		-0.23	
6266	D4052	0.88424		0.56	
6310	D4052	0.8842		0.33	

lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO12185	0.884141		0.00	
6380	D4052	0.88407		-0.40	
6395		----		----	



normality suspect
n 60
outliers 5
mean (n) 0.88414
st.dev. (n) 0.000163
R(calc.) 0.00046
st.dev.(ISO12185:96) 0.000179
R(ISO12185:96) 0.0005



Determination of Evaporation loss by Noack on sample #21075; results in %M/M

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237	D5800-B	5.9		2.53	
254		----		----	
257		----		----	
309		----		----	
325	D5800-B	5.0		-0.56	
329		----		----	
333	CEC L-40-93	5.1		-0.21	
339		----		----	
349		----		----	
360		----		----	
398		----		----	
421		----		----	
432		----		----	
496	D5800-B	5.0		-0.56	
614		----		----	
633		----		----	
634		----		----	
657	D5800-B	5.7		1.84	
780		----		----	
823		----		----	
840		----		----	
862	D5800-B	4.9		-0.90	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974	D5800-B	5.2		0.13	
994		----		----	
1011	CEC L-40-93	4.5		-2.27	
1017		----		----	
1026	CEC L-40-93	5.1		-0.21	
1047	DIN51581	5.35		0.64	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	DIN51581	5.00		-0.56	
1316		----		----	
1320		----		----	
1326	D5800-A	5.0		-0.56	
1328		----		----	
1384		----		----	
1412		----		----	
1431		----		----	
1433		----		----	
1438		----		----	
1460		----		----	
1510	D5800-B	5.26		0.33	
1564	DIN51581	5.4		0.81	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850	DIN51581	4.9		-0.90	
1864		----		----	
1877	D5800-B	5.1		-0.21	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197	D5800-B	5.0		-0.56	
6266		----		----	



lab	method	value	mark	z(targ)	remarks
6310		----		----	
6324		----		----	
6358	D5800-A	5.51		1.19	
6380	D5800-B	7.44	G(0.01)	7.80	
6395		----		----	
				<u>D5800-A/DIN51581 only</u>	<u>D5800-B only</u>
normality		suspect		unknown	suspect
n		18		6	9
outliers		1		0	1
mean (n)		5.162		5.193	5.229
st.dev. (n)		0.3233		0.2563	0.3457
R(calc.)		0.905		0.718	0.968
st.dev.(D5800-B:20)		0.2921		---	0.2950
R(D5800-B:20)		0.818		---	0.826
Compare					
R(D5800-A:20)		0.945		0.950	---



Determination of Flash Point C.O.C. on sample #21075; results in °C

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D92	242		-0.41	
211	D92	250		0.83	
219	D92	246.0		0.21	
237	D92	238		-1.03	
254		----		----	
257		----		----	
309		----		----	
325	D92	244		-0.10	
329	D92	226		-2.90	
333	D92	256		1.77	
339		----		----	
349	D92	244		-0.10	
360	D92	244		-0.10	
398	D92	232		-1.97	
421	ISO2592	233		-1.81	
432		----		----	
496	D92	242		-0.41	
614		----		----	
633	D92	253.68		1.41	
634	D92	245		0.06	
657	D92	238		-1.03	
780	D92	232		-1.97	
823	D92	245		0.06	
840	D92	240		-0.72	
862	D92	248		0.52	
875	D92	250		0.83	
902		----		----	
912	D92	242		-0.41	
913		----		----	
922	D92	238		-1.03	
962	D92	246		0.21	
963	D92	244.0		-0.10	
974	D92	238		-1.03	
994	D92	242.0		-0.41	
1011	D92	234		-1.65	
1017	D92	242		-0.41	
1026	D92	254		1.46	
1047		----		----	
1059	ISO2592	242		-0.41	
1146		----		----	
1150	ISO2592	255		1.61	
1173		----		----	
1213	D92	250		0.83	
1235		----		----	
1243	ISO2592	238		-1.03	
1316	D92	256		1.77	
1320	D92	242		-0.41	
1326	D92	243		-0.25	
1328	D92	240		-0.72	
1384	ISO2592	254		1.46	
1412	D92	239.0		-0.88	
1431	D92	246.0		0.21	
1433	D92	248		0.52	
1438		----		----	
1460	D92	230		-2.28	
1510	D92	242.87		-0.27	
1564	D92	242		-0.41	
1650	D92	255	C	1.61	first reported 155
1740		----		----	
1748	D92	254		1.46	
1799		----		----	
1850	ISO2592	252		1.15	
1864	D92	236.0		-1.34	
1877	D92	234		-1.65	
1957	D92	238		-1.03	
1969	ISO2592	256.13		1.79	
6016		----		----	
6032		----		----	
6035	ISO2592	262		2.70	
6181		----		----	
6183	D92	262.0		2.70	
6197	D92	250		0.83	
6266	D92	246.0		0.21	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO2592	251		0.99	
6380	D92	251.5		1.07	
6395		----		----	
	normality	OK			
	n	57			
	outliers	0			
	mean (n)	244.63			
	st.dev. (n)	7.972			
	R(calc.)	22.32			
	st.dev.(D92:18)	6.429			
	R(D92:18)	18			



Determination of Flash Point PMcc on sample #21075; results in °C

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D93-A	204.0		0.57	
211		----		----	
219		----		----	
237	D93-A	202.0		0.18	
254	D93-A	210	C,R(0.01)	1.75	first reported 192
257	D93-A	201		-0.02	
309	D93-A	200.0		-0.21	
325	D93-A	203		0.38	
329	D93-A	204		0.57	
333		----		----	
339		----		----	
349	D93-A	200		-0.21	
360	D93-A	201.0		-0.02	
398		----		----	
421	ISO2719-A	200.5		-0.11	
432		----		----	
496		----		----	
614	D93-A	196		-1.00	
633	D93-A	201.50		0.08	
634	D93-A	202.0		0.18	
657	D93-A	204		0.57	
780	D93-A	191.5	R(0.01)	-1.88	
823	ISO2719-A	198		-0.60	
840	D93-A	200.0		-0.21	
862	D93-A	200.0		-0.21	
875	D93-A	200.5		-0.11	
902	D93-A	202		0.18	
912		----		----	
913		----		----	
922	D93-A	202		0.18	
962	D93-A	198.0		-0.60	
963	D93-A	202.0		0.18	
974	D93-A	204		0.57	
994	D93-A	201.0		-0.02	
1011		----		----	
1017	D93-A	203		0.38	
1026	D93-A	206.0		0.96	
1047		----		----	
1059	ISO2719-A	202.0		0.18	
1146	D93-A	200.0		-0.21	
1150	ISO2719-B	202		0.18	
1173	D93-A	202.0		0.18	
1213	D93-A	198		-0.60	
1235		----		----	
1243	ISO2719-A	202		0.18	
1316		----		----	
1320		----		----	
1326		----		----	
1328		----		----	
1384	ISO2719-A	216	R(0.01)	2.93	
1412	D93-A	200.0		-0.21	
1431	D93-A	202.5		0.28	
1433	D93-A	198		-0.60	
1438		----		----	
1460	D93-B	200		-0.21	
1510	D93-A	199.147		-0.38	
1564	D93-A	200.0		-0.21	
1650	D93-A	197.8		-0.64	
1740		----		----	
1748	D93-A	205		0.77	
1799		----		----	
1850	ISO2719-A	203		0.38	
1864	D93-A	202.9		0.36	
1877	D93-A	201		-0.02	
1957		----		----	
1969	ISO2719-A	213.20	R(0.01)	2.38	
6016	D93-A	201.5		0.08	
6032		----		----	
6035	ISO2719-A	202.0		0.18	
6181		----		----	
6183	D93-A	200.0		-0.21	
6197		----		----	
6266		----		----	
6310		----		----	


lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO2719-A	199.0		-0.41	
6380	D93-A	197.5		-0.70	
6395		----		----	
	normality	OK			
n		47			
outliers		4			
mean (n)		201.08			
st.dev. (n)		2.113			
R(calc.)		5.92			
st.dev.(D93-A:20)		5.099			
R(D93-A:20)		14.28			



Determination of Foaming Tendency, 5 min blowing period on sample #21075; results in mL

lab	method	sample used	diffuser	Seq I	mark	z(targ)	Seq II	mark	z(targ)	Seq III	mark	z(targ)
178		---	---	----		----	----		----	----		----
179		As received	Metal	----		----	20		0.73	----		----
211	D892	As received	Stone	0		-2.12	20		0.73	0		-0.68
219	D892	---	---	0		-2.12	20		0.73	0		-0.68
237	D892	---	---	0		-2.12	0		-2.27	0		-0.68
254		---	---	----		----	----		----	----		----
257		---	---	----		----	----		----	----		----
309	D892	---	---	0		-2.12	30		2.22	0		-0.68
325	D892	As received	Metal	0		-2.12	20		0.73	0		-0.68
329	D892	After agitation	Stone	0		-2.12	20		0.73	0		-0.68
333		---	---	----		----	----		----	----		----
339		---	---	----		----	----		----	----		----
349		---	---	----		----	----		----	----		----
360	D892 (Altern.)	As received	Stone	5		21.04	10		-0.77	0		-0.68
398	D892	After agitation	Stone	0		-2.12	0		-2.27	0		-0.68
421		---	---	----		----	----		----	----		----
432	D892	As received	Stone	0		-2.12	0		-2.27	0		-0.68
496	D892	As received	Metal	0		-2.12	20		0.73	0		-0.68
614		---	---	----		----	----		----	----		----
633		---	---	----		----	----		----	----		----
634		---	---	----		----	----		----	----		----
657	D892	As received	Metal	0		-2.12	30		2.22	0		-0.68
780		---	---	----		----	----		----	----		----
823	D892	As received	Stone	3.3		13.16	15.6		0.07	2.4		5.69
840	D892	After agitation	Metal	10	R(1)	44.21	20		0.73	10	R(1)	25.84
862	D892	As received	Metal	0		-2.12	15		-0.02	0		-0.68
875		---	---	----		----	----		----	----		----
902	D892	After agitation	Metal	0		-2.12	30		2.22	0		-0.68
912		---	---	----		----	----		----	----		----
913		---	---	----		----	----		----	----		----
922	D892	As received	Stone	<10		----	15		-0.02	<10		----
962		---	---	----		----	----		----	----		----
963		---	---	----		----	----		----	----		----
974	D892	As received	Metal	0		-2.12	10		-0.77	0		-0.68
994		---	---	----		----	----		----	----		----
1011		---	---	----		----	----		----	----		----
1017		---	---	----		----	----		----	----		----
1026	D892	After agitation	Metal	0		-2.12	20		0.73	0		-0.68
1047	D892	As received	Stone	0		-2.12	20		0.73	0		-0.68
1059		---	---	----		----	----		----	----		----
1146		---	---	----		----	----		----	----		----
1150		---	---	----		----	----		----	----		----
1173		---	---	----		----	----		----	----		----
1213	D892	After agitation	Metal	5		21.04	5		-1.52	5		12.58
1235		---	---	----		----	----		----	----		----
1243	D892	As received	Stone	0		-2.12	20		0.73	0		-0.68
1316		---	---	----		----	----		----	----		----
1320		---	---	----		----	----		----	----		----
1326		---	---	----		----	----		----	----		----
1328		---	---	----		----	----		----	----		----
1384	ISO6247	As received	Metal	0		-2.12	20		0.73	0		-0.68
1412		---	---	----		----	----		----	----		----
1431	D892	---	0			-2.12	20		0.73	0		-0.68
1433	D892	After agitation	Stone	0		-2.12	10		-0.77	0		-0.68
1438		---	---	----		----	----		----	----		----
1460	D892	As received	Metal	0		-2.12	20		0.73	0		-0.68
1510		---	---	----		----	----		----	----		----
1564	D892	As received	Metal	10	R(1)	44.21	10		-0.77	10	R(1)	25.84
1650		---	---	----		----	----		----	----		----
1740		---	---	----		----	----		----	----		----
1748		---	---	----		----	----		----	----		----
1799		After agitation	Metal	10	R(1)	44.21	10		-0.77	10	R(1)	25.84
1850	ISO6247	As received	Stone	0		-2.12	15		-0.02	0		-0.68
1864		After agitation	Metal	0		-2.12	0		-2.27	0		-0.68
1877		---	---	----		----	----		----	----		----
1957		As received	Metal	---		----	----		----	----		----
1969		---	---	----		----	----		----	----		----
6016	D892	After agitation	Stone	0		-2.12	20		0.73	0		-0.68
6032		---	---	----		----	----		----	----		----
6035		---	---	----		----	----		----	----		----
6181		---	---	----		----	----		----	----		----
6183		---	---	----		----	----		----	----		----
6197	D892	After agitation	Metal	0		-2.12	10		-0.77	0		-0.68
6266		---	---	----		----	----		----	----		----
6310	D892	After agitation	Metal	0		-2.12	10		-0.77	0		-0.68

lab	method	sample used	diffuser	Seq I	mark	z(targ)	Seq II	mark	z(targ)	Seq III	mark	z(targ)
6324	---	---	---	---	---	---	---	---	---	---	---	---
6358	D892 (Altern.)	As received	Stone	0		-2.12	10		-0.77	0		-0.68
6380	---	---	---	---	---	---	---	---	---	---	---	---
6395	---	---	---	---	---	---	---	---	---	---	---	---
normality				not OK			OK			not OK		
n				29			34			29		
outliers				3			0			3		
mean (n)				0.46			15.16			0.26		
st.dev. (n)				1.399			8.210			1.015		
R(calc.)				3.92			22.99			2.84		
st.dev.(D892:18)				0.216			6.668			0.377		
R(D892:18)				0.60			18.67			1.06		



Determination of Foam Stability, 10 min settling point on sample #21075; results in mL



lab	method	Seq I	mark	z(targ)	Seq II	mark	z(targ)	Seq III	mark	z(targ)
178		----		----	----		----	----		----
179	D892	0		----	0		----	0		----
211	D892	0		----	0		----	0		----
219	D892	0		----	0		----	0		----
237	D892	0		----	0		----	0		----
254		----		----	----		----	----		----
257		----		----	----		----	----		----
309	D892	0		----	0		----	0		----
325	D892	0		----	0		----	0		----
329	D892	0		----	0		----	0		----
333		----		----	----		----	----		----
339		----		----	----		----	----		----
349		----		----	----		----	----		----
360	D892 (Altern.)	0		----	0		----	0		----
398	D892	0		----	0		----	0		----
421		----		----	----		----	----		----
432	D892	0		----	0		----	0		----
496	D892	0		----	0		----	0		----
614		----		----	----		----	----		----
633		----		----	----		----	----		----
634		----		----	----		----	----		----
657	D892	0		----	0		----	0		----
780		----		----	----		----	----		----
823	D892	0		----	0		----	0		----
840	D892	0		----	0		----	0		----
862	D892	0		----	0		----	0		----
875		----		----	----		----	----		----
902	D892	0		----	0		----	0		----
912		----		----	----		----	----		----
913		----		----	----		----	----		----
922	D892	0		----	0		----	0		----
962		----		----	----		----	----		----
963		----		----	----		----	----		----
974	D892	0		----	0		----	0		----
994		----		----	----		----	----		----
1011		----		----	----		----	----		----
1017		----		----	----		----	----		----
1026	D892	0		----	0		----	0		----
1047	D892	0		----	0		----	0		----
1059		----		----	----		----	----		----
1146		----		----	----		----	----		----
1150		----		----	----		----	----		----
1173		----		----	----		----	----		----
1213	D892	0		----	0		----	0		----
1235		----		----	----		----	----		----
1243	D892	0		----	0		----	0		----
1316		----		----	----		----	----		----
1320		----		----	----		----	----		----
1326		----		----	----		----	----		----
1328		----		----	----		----	----		----
1384	ISO6247	0		----	0		----	0		----
1412		----		----	----		----	----		----
1431	D892	0		----	0		----	0		----
1433	D892	0		----	0		----	0		----
1438		----		----	----		----	----		----
1460	D892	0		----	0		----	0		----
1510		----		----	----		----	----		----
1564	D892	0		----	0		----	0		----
1650		----		----	----		----	----		----
1740		----		----	----		----	----		----
1748		----		----	----		----	----		----
1799	D892	0		----	0		----	0		----
1850	ISO6247	0		----	0		----	0		----
1864		0		----	0		----	0		----
1877		----		----	----		----	----		----
1957		----		----	----		----	----		----
1969		----		----	----		----	----		----
6016	D892	0		----	0		----	0		----
6032		----		----	----		----	----		----
6035		----		----	----		----	----		----
6181		----		----	----		----	----		----
6183		----		----	----		----	----		----
6197	D892	0		----	0		----	0		----
6266		----		----	----		----	----		----
6310	D892	0		----	0		----	0		----

lab	method	Seq I	mark	z(targ)	Seq II	mark	z(targ)	Seq III	mark	z(targ)
6324		-----		-----	-----		-----	-----		-----
6358	D892 (Altern.)	0		-----	0		-----	0		-----
6380		-----		-----	-----		-----	-----		-----
6395		-----		-----	-----		-----	-----		-----
n		34		34			34			
mean (n)		0		0			0			

Determination of Kinematic Viscosity at 40°C on sample #21075; results in mm²/s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D445	130.80		-0.42	
211	D445	131.15		0.20	
219	D7279 corr. to D445	131.4		0.64	
237	D445	129.7		-2.34	
254	D445	131.7		1.16	
257	D7279 corr. to D445	132.1		1.86	
309	D445	130.9		-0.24	
325	D445	130.8		-0.42	
329	D445	130.69		-0.61	
333	D445	130.9		-0.24	
339		----		----	
349	D445	130.8		-0.42	
360	D445	131.17		0.23	
398		----		----	
421	ISO3104	131.9		1.51	
432	D445	131.1		0.11	
496	D445	130.88		-0.28	
614		----		----	
633	D7279 corr. to D445	131.00		-0.07	
634	D445	130.0		-1.82	
657	D445	130.9		-0.24	
780	D445	131.3		0.46	
823	D445	131.1		0.11	
840	D445	130.74		-0.52	
862	D445	132.4		2.39	
875	D445	132.3		2.21	
902		----		----	
912	D445	131.8		1.34	
913		----		----	
922	D445	131.0		-0.07	
962	D445	131.6	C	0.99	first reported 133.9
963	D445	130.1		-1.64	
974	D445	130.88		-0.28	
994	D445	131.7		1.16	
1011		----		----	
1017	D445	131.7		1.16	
1026	D445	130.7		-0.59	
1047	ISO3104	131.2		0.28	
1059	ISO3104	130.5		-0.94	
1146	D445	131.45		0.72	
1150	ISO3104	130.2382		-1.40	
1173	D445	131.3		0.46	
1213	D445	131.3	C	0.46	first reported 14.24
1235		----		----	
1243	D7279 corr. to D445	131.205		0.29	
1316	ISO3104	130.4		-1.12	
1320		----		----	
1326	D445	129.85		-2.08	
1328		----		----	
1384	ISO3104	130.6		-0.77	
1412	D445	131.22		0.32	
1431		----		----	
1433	D445	131.3		0.46	
1438		131.6		0.99	
1460	D445	130.17		-1.52	
1510	D445	130.73		-0.54	
1564	D445	130.934		-0.18	
1650	D445	130.895		-0.25	
1740		----		----	
1748		----		----	
1799		----		----	
1850	ISO3104	131.1		0.11	
1864	D445	131.1		0.11	
1877	D445	130.9		-0.24	
1957	D445	132.22		2.07	
1969	ISO3104	130.9594		-0.14	
6016		----		----	
6032	D7279 corr. to D445	131.697		1.16	
6035	ISO3104	130.2		-1.47	
6181	ISO3104	129.0		-3.57	
6183	D445	131.0		-0.07	
6197	D445	130.4		-1.12	
6266	D7042	132.0		1.69	



lab	method	value	mark	z(targ)	remarks
6310	D7279 corr. to D445	131.3		0.46	
6324		-----		-----	
6358	D445	130.6		-0.77	
6380	D445	131.738		1.23	
6395		-----		-----	
	normality	OK			
	n	62			
	outliers	0			
	mean (n)	131.037			
	st.dev. (n)	0.6468			
	R(calc.)	1.811			
	st.dev.(D445:21)	0.5709			
	R(D445:21)	1.599			



Determination of Kinematic Viscosity at 100°C on sample #21075; results in mm²/s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D445	16.79	R(0.01)	35.87	
211	D445	14.30		0.46	
219	D7279 corr. to D445	14.26		-0.11	
237	D445	14.28		0.18	
254	D445	14.18		-1.25	
257	D7279 corr. to D445	14.25		-0.25	
309	D445	14.27		0.03	
325	D445	14.27		0.03	
329	D445	14.249		-0.26	
333	D445	14.22		-0.68	
339		----		----	
349	D445	14.27		0.03	
360	D445	14.210		-0.82	
398		----		----	
421	ISO3104	14.29		0.32	
432	D445	14.35		1.17	
496	D445	14.287		0.28	
614		----		----	
633	D7279 corr. to D445	14.177		-1.29	
634	D445	14.08	R(0.05)	-2.67	
657	D445	14.33		0.89	
780	D445	14.23		-0.54	
823	ISO3104	14.18		-1.25	
840	D445	14.285		0.25	
862	D445	14.28		0.18	
875	D445	14.29		0.32	
902		----		----	
912	D445	14.21		-0.82	
913		----		----	
922	D445	14.30		0.46	
962	D445	14.30	C	0.46	first reported 14.76
963	D445	14.26		-0.11	
974	D445	14.23		-0.54	
994	D445	14.35		1.17	
1011		----		----	
1017	D445	14.27		0.03	
1026	D445	14.28		0.18	
1047	ISO3104	14.18		-1.25	
1059	ISO3104	14.22		-0.68	
1146	D445	14.275		0.10	
1150	ISO3104	14.2597		-0.11	
1173		----		----	
1213	D445	14.24	C	-0.39	first reported 131.3
1235		----		----	
1243	D7279 corr. to D445	14.307		0.56	
1316	ISO3104	14.24		-0.39	
1320		----		----	
1326	D445	14.135		-1.89	
1328		----		----	
1384	ISO3104	14.29		0.32	
1412	D445	14.30		0.46	
1431		----		----	
1433	D445	14.28		0.18	
1438		14.30		0.46	
1460	D445	14.265		-0.04	
1510	D445	14.2045		-0.90	
1564	D445	14.262		-0.08	
1650	D445	14.256		-0.17	
1740		----		----	
1748		----		----	
1799		----		----	
1850	ISO3104	14.27		0.03	
1864	D445	14.27		0.03	
1877	D445	14.25		-0.25	
1957	D445	14.425		2.24	
1969	ISO3104	14.2592		-0.12	
6016		----		----	
6032	D7279 corr. to D445	14.335		0.96	
6035	ISO3104	14.26		-0.11	
6181	ISO3104	14.3		0.46	
6183	D445	14.30		0.46	
6197	D445	14.25		-0.25	
6266	D7042	14.37		1.46	
6310	D7279 corr. to D445	14.25		-0.25	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	D445	14.22		-0.68	
6380	D445	14.359		1.30	
6395		----		----	
	normality	suspect			
n	59				
outliers	2				
mean (n)	14.268				
st.dev. (n)	0.0509				
R(calc.)	0.142				
st.dev.(D445:21)	0.0703				
R(D445:21)	0.197				



Determination of Viscosity Index on sample #21075;

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D2270	139	ex	44.19	test result excluded, outlier in KV 100°C
211	D2270	107.6		0.23	
219	D2270	107		-0.61	
237	D2270	109		2.19	
254	D2270	106		-2.01	
257	D2270	106		-2.01	
309	D2270	108		0.79	
325	D2270	108		0.79	
329	D2270	108		0.79	
333	D2270	107		-0.61	
339		----		----	
349	D2270	108		0.79	
360	ISO2909	106.6		-1.17	
398		----		----	
421	ISO2909	107		-0.61	
432	D2270	108		0.79	
496	D2270	107.8		0.51	
614		----		----	
633	D2270	106.83	E	-0.85	iis calculated 106.35
634	D2270	106	ex	-2.01	test result excluded, outlier in KV 100°C
657	D2270	108.4		1.35	
780	D2270	107		-0.61	
823	D2270	106		-2.01	
840	D2270	107.98		0.76	
862	D2270	106		-2.01	
875	D2270	106		-2.01	
902		----		----	
912	D2270	106		-2.01	
913		----		----	
922	D2270	107.9		0.65	
962	D2270	107.3	C	-0.19	first reported 110.919
963	D2270	108.3		1.21	
974	D2270	107		-0.61	
994	D2270	108		0.79	
1011	D2270	108		0.79	
1017	D2270	107		-0.61	
1026	D2270	108		0.79	
1047	ISO2909	106		-2.01	
1059	ISO2909	107		-0.61	
1146	D2270	107		-0.61	
1150	ISO2909	108.42		1.38	
1173		----		----	
1213	D2270	107		-0.61	
1235		----		----	
1243	ISO2909	107.8		0.51	
1316	D2270	108		0.79	
1320		----		----	
1326	D2270	107		-0.61	
1328		----		----	
1384	ISO2909	108		0.79	
1412	D2270	107.7		0.37	
1431	D2270	107.5		0.09	
1433	ISO2909	108	E	0.79	iis calculated 107
1438		107.3		-0.19	
1460	D2270	108.279		1.18	
1510	D2270	106.9		-0.75	
1564	D2270	108		0.79	
1650	D2270	107.5		0.09	
1740		----		----	
1748	D2270	107.33		-0.15	
1799		----		----	
1850	ISO2909	108	E	0.79	iis calculated 107
1864	D2270	107		-0.61	
1877	D2270	107		-0.61	
1957	D2270	108.3		1.21	
1969	ISO2909	107.44		0.01	
6016	D2270	107.32		-0.16	
6032	D2270	108		0.79	
6035	ISO2909	108		0.79	
6181		----		----	
6183	D2270	107.5		0.09	
6197	D2270	108		0.79	
6266	D2270	107.7		0.37	
6310	D2270	107		-0.61	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	D2270	107.3		-0.19	
6380	D2270	107.96		0.74	
6395		----		----	
	normality	OK			
n		62			
outliers		0 +2ex			
mean (n)		107.43			
st.dev. (n)		0.715			
R(calc.)		2.00			
st.dev.(D2270:10)		0.714			
R(D2270:10)		2			



Determination of Kinematic Viscosity Stabinger at 40°C on sample #21075; results in mm²/s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D7042	131.37		0.09	
398		----		----	
421	D7042	132.1		1.41	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634		----		----	
657		----		----	
780		----		----	
823	D7042	130.8		-0.95	
840	D7042	131.34		0.03	
862		----		----	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922	D7042	130.7		-1.13	
962		----		----	
963	D7042	129.9		-2.58	
974		----		----	
994	D7042	131.2		-0.22	
1011	D7042	131.3		-0.04	
1017		----		----	
1026		----		----	
1047		----		----	
1059	D7042	132.1		1.41	
1146		----		----	
1150	D7042	130.98		-0.62	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320	D7042	130.8		-0.95	
1326		----		----	
1328		----		----	
1384	PN-C-04358	130.3		-1.86	
1412	D7042	131.51		0.34	
1431	D7042	131.32		0.00	
1433	D7042	131.6		0.50	
1438		131.6		0.50	
1460	D7042	131.27		-0.09	
1510		----		----	
1564		----		----	
1650		----		----	
1740		----		----	
1748	D7042	131.87		0.99	
1799		----		----	
1850		----		----	
1864		----		----	
1877	D7042	131.4		0.14	
1957		----		----	
1969		----		----	
6016	D7042	131.51		0.34	
6032		----		----	
6035		----		----	
6181	D7042	133.0		3.05	
6183	D7042	131.4		0.14	
6197		----		----	
6266		----		----	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358		----		----	
6380	D7042	131.04		-0.51	
6395		----		----	
	normality	suspect			
n		23			
outliers		0			
mean (n)		131.322			
st.dev. (n)		0.6317			
R(calc.)		1.769			
st.dev.(D7042:21)		0.5510			
R(D7042:21)		1.543			



Determination of Kinematic Viscosity Stabinger at 100°C on sample #21075; results in mm²/s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D7042	14.290		-0.35	
398		----		----	
421	D7042	14.33		0.45	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634		----		----	
657		----		----	
780		----		----	
823	D7042	14.35		0.85	
840	D7042	14.284		-0.47	
862		----		----	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922	D7042	14.29		-0.35	
962		----		----	
963	D7042	14.29	C	-0.35	first reported 14.19
974		----		----	
994	D7042	14.36		1.05	
1011	D7042	14.29		-0.35	
1017		----		----	
1026		----		----	
1047		----		----	
1059	D7042	14.27		-0.75	
1146		----		----	
1150	D7042	14.312		0.09	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320	D7042	14.29		-0.35	
1326		----		----	
1328		----		----	
1384	PN-C-01358	14.26		-0.95	
1412	D7042	14.33		0.45	
1431	D7042	14.289		-0.37	
1433	D7042	14.32		0.25	
1438		14.38		1.45	
1460	D7042	14.611	R(0.01)	6.07	
1510		----		----	
1564		----		----	
1650		----		----	
1740		----		----	
1748	D7042	14.321		0.27	
1799		----		----	
1850		----		----	
1864		----		----	
1877	D7042	14.30		-0.15	
1957		----		----	
1969		----		----	
6016	D7042	14.292		-0.31	
6032		----		----	
6035		----		----	
6181	D7042	14.3		-0.15	
6183	D7042	14.31		0.05	
6197		----		----	
6266		----		----	
6310		----		----	


lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358		----		----	
6380	D7042	14.184	R(0.05)	-2.47	
6395		----		----	
	normality	OK			
	n	21			
	outliers	2			
	mean (n)	14.308			
	st.dev. (n)	0.0297			
	R(calc.)	0.083			
	st.dev.(D7042:21)	0.0500			
	R(D7042:21)	0.140			



Determination of Viscosity HTHS on sample #21075; results in mPa·s

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360		----		----	
398		----		----	
421		----		----	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634		----		----	
657		----		----	
780		----		----	
823		----		----	
840		----		----	
862	D4683	4.20		-0.75	
875		----		----	
902		----		----	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974		----		----	
994		----		----	
1011		----		----	
1017		----		----	
1026	D5481	4.28		0.53	
1047	PN-C-04098	4.17		-1.23	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320		----		----	
1326	D5481	4.16		-1.39	
1328		----		----	
1384		----		----	
1412		----		----	
1431		----		----	
1433		----		----	
1438		----		----	
1460		----		----	
1510		----		----	
1564	D4683	4.343		1.53	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864		----		----	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197	D5481	4.33		1.32	
6266		----		----	
6310		----		----	
6324		----		----	



lab	method	value	mark	z(targ)	remarks
6358		-----		-----	
6380		-----		-----	
6395		-----		-----	
	normality	unknown			
	n	6			
	outliers	0			
	mean (n)	4.247			
	st.dev. (n)	0.0811			
	R(calc.)	0.227			
	st.dev.(D4683:17)	0.0625			
	R(D4683:17)	0.175			



Determination of Nitrogen on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219		----		----	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325	D5762	846		0.48	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360		----		----	
398		----		----	
421		----		----	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634		----		----	
657	D5762	711		-1.27	
780	D3228	840		0.41	
823	D5762	825		0.21	
840	D3228	882.7		0.96	
862	D5762	934		1.63	
875	D5762	819		0.13	
902	D5762	631		-2.31	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974		----		----	
994	D5762	873		0.83	
1011		----		----	
1017		----		----	
1026	D5762	852		0.56	
1047		----		----	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320		----		----	
1326	D5762	779		-0.39	
1328		----		----	
1384	In house	732	C	-1.00	first reported 1131
1412	D5762	920		1.45	
1431		----		----	
1433		----		----	
1438		----		----	
1460		----		----	
1510		----		----	
1564	D5762	679.5		-1.68	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864		----		----	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197		----		----	
6266		----		----	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		-----		-----	
6358		-----		-----	
6380		-----		-----	
6395		-----		-----	
	normality	OK			
	n	14			
	outliers	0			
	mean (n)	808.87			
	st.dev. (n)	90.679			
	R(calc.)	253.90			
	st.dev.(D5762:18a)	76.843			
	R(D5762:18a)	215.16			



Determination of Pour Point Manual on sample #21075; results in °C

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D97	-18	R(0.05)	-2.18	
211	D97	-15		-1.25	
219	D97	-9		0.62	
237	D97	-9		0.62	
254	D97	-12		-0.31	
257		----		----	
309		----		----	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D97	-12		-0.31	
398		----		----	
421	ISO3016	-12		-0.31	
432		----		----	
496		----		----	
614	D97	-9		0.62	
633		----		----	
634		----		----	
657	D97	-12		-0.31	
780	D97	-9		0.62	
823	ISO3016	-12		-0.31	
840	D97	-9		0.62	
862	D97	-12		-0.31	
875	D97	-9		0.62	
902	D97	-9		0.62	
912	D97	-12		-0.31	
913		----		----	
922	D97	-18	R(0.05)	-2.18	
962	D97	-9		0.62	
963	D97	-12		-0.31	
974	D97	-12		-0.31	
994	D97	-9		0.62	
1011	D97	-12		-0.31	
1017		----		----	
1026		----		----	
1047		----		----	
1059	ISO3016	-12		-0.31	
1146	D97	-9		0.62	
1150	ISO3016	-12		-0.31	
1173		----		----	
1213		----		----	
1235		----		----	
1243		----		----	
1316		----		----	
1320	D97	-12		-0.31	
1326	D97	-9		0.62	
1328	D97	-12		-0.31	
1384		----		----	
1412	D97	-9		0.62	
1431	D97	-14.0		-0.94	
1433	ISO3016	-11.67		-0.21	
1438		----		----	
1460		----		----	
1510		----		----	
1564		----		----	
1650		----		----	
1740		----		----	
1748		----		----	
1799	D97	-9		0.62	
1850	ISO3016	-12		-0.31	
1864	D97	-9		0.62	
1877		----		----	
1957		----		----	
1969	ISO3016	-15		-1.25	
6016		----		----	
6032		----		----	
6035	ISO3016	-12		-0.31	
6181		----		----	
6183		----		----	
6197	D97	-9		0.62	
6266		----		----	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO3016	-12		-0.31	
6380	D97	-12		-0.31	
6395		----		----	
	normality	OK			
n		37			
outliers		2			
mean (n)		-10.99			
st.dev. (n)		1.822			
R(calc.)		5.10			
st.dev.(D97:17b)		3.214			
R(D97:17b)		9			



Determination of Pour Point Automated 1°C interval on sample #21075; results in °C

lab	method	value	mark	z(targ)	remarks
178		----		----	
179		----		----	
211		----		----	
219	D5950	-9		1.62	
237		----		----	
254		----		----	
257		----		----	
309		----		----	
325	D5950	-12		-0.24	
329	D5950	-9		1.62	
333		----		----	
339		----		----	
349	D5950	-12		-0.24	
360		----		----	
398		----		----	
421		----		----	
432		----		----	
496	D5950	-9		1.62	
614		----		----	
633		----		----	
634		----		----	
657	D5950	-13		-0.87	
780	D5950	-10		1.00	
823	D5950	-13		-0.87	
840		----		----	
862		----		----	
875		----		----	
902	D5950	-9		1.62	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974		----		----	
994		----		----	
1011	D6892	-12		-0.24	
1017	D5950	-9	C	1.62	first reported -27
1026	D5950	-12		-0.24	
1047		----		----	
1059		----		----	
1146		----		----	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	D7346	-15.5		-2.42	
1316	D5950	-11.0		0.38	
1320		----		----	
1326		----		----	
1328		----		----	
1384		----		----	
1412		----		----	
1431		----		----	
1433	D5950	-13		-0.87	
1438		----		----	
1460	D5950	-19.0	R(0.05)	-4.60	
1510		----		----	
1564	D5950	-12		-0.24	
1650	D5950	-13		-0.87	
1740		----		----	
1748	D7346	-15		-2.11	
1799		----		----	
1850		----		----	
1864	D5950	-11		0.38	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032	D5950	-12		-0.24	
6035		----		----	
6181		----		----	
6183		----		----	
6197		----		----	
6266		----		----	
6310	D5950	-12.3		-0.43	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358		----		----	
6380		----		----	
6395		----		----	
	normality	OK			
	n	21			
	outliers	1			
	mean (n)	-11.61			
	st.dev. (n)	1.912			
	R(calc.)	5.35			
	st.dev.(D5950:14)	1.607			
	R(D5950:14)	4.5			



Determination of Sulfated Ash on sample #21075; results in %M/M

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D874	0.821		-0.14	
211		----		----	
219		----		----	
237	D874	0.802		-0.58	
254	D874	0.9365		2.48	
257		----		----	
309	D874	0.829		0.04	
325		----		----	
329		----		----	
333		----		----	
339		----		----	
349		----		----	
360	D874	0.82		-0.17	
398		----		----	
421	ISO3987	0.77		-1.30	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634	D874	0.884		1.29	
657	D874	0.73		-2.21	
780	D874	0.807		-0.46	
823	D874	0.80		-0.62	
840	D874	0.868		0.92	
862	D874	0.78		-1.08	
875	D874	0.9059		1.79	
902	D874	0.98		3.47	
912		----		----	
913		----		----	
922		----		----	
962		----		----	
963		----		----	
974	D874	0.89		1.43	
994	D874	0.82		-0.17	
1011		----		----	
1017		----		----	
1026	D874	0.91		1.88	
1047	ISO3987	0.79		-0.85	
1059		----		----	
1146	D874	0.9842		3.57	
1150	ISO3987	0.74		-1.98	
1173		----		----	
1213	D874	0.84		0.29	
1235	ISO3987	0.7710		-1.28	
1243		----		----	
1316	D874	0.64		-4.26	
1320		----		----	
1326	D874	0.8105		-0.38	
1328		----		----	
1384		----		----	
1412	D874	0.84		0.29	
1431		----		----	
1433	D874	0.8625		0.80	
1438		----		----	
1460	D874	1.0		3.93	
1510	D874	0.895		1.54	
1564		----		----	
1650	D874	0.7722		-1.25	
1740		----		----	
1748	D874	0.842		0.33	
1799		----		----	
1850	ISO3987	0.77		-1.30	
1864	D874	0.778		-1.12	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181	ISO3987	0.745		-1.87	
6183		----		----	
6197		----		----	
6266		----		----	
6310		----		----	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO3987	0.7470		-1.83	
6380		----		----	
6395	In house	0.775		-1.19	
	normality	OK			
n		35			
outliers		0			
mean (n)		0.827			
st.dev. (n)		0.0780			
R(calc.)		0.219			
st.dev.(D874:13a)		0.0440			
R(D874:13a)		0.123			



Determination of Sulfur on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D4294	4982		1.05	
211		----		----	
219		----		----	
237	D4294	4490		-1.99	
254		----		----	
257		----		----	
309	D4294	4930		0.73	
325	D5185	4589		-1.37	
329	D4294	4620		-1.18	
333	D4294	4630		-1.12	
339	INH-050	4400		-2.54	
349	D2622	4661		-0.93	
360	D5453	4822		0.06	
398	ISO8754	5170		2.21	
421		----		----	
432		----		----	
496		----		----	
614		----		----	
633		----		----	
634	D4294	4610	C	-1.24	first reported 0.461 mg/kg
657	D4294	4770		-0.26	
780	D4294	4887		0.47	
823		----		----	
840	D5453	4929	C	0.73	first reported 0.4929 mg/kg
862	D2622	4870		0.36	
875	D4294	4830	C	0.11	reported 0.483 mg/kg
902	D4294	5300		3.02	
912		----		----	
913		----		----	
922	D4294	4839		0.17	
962		----		----	
963		----		----	
974		----		----	
994	D4294	4974		1.00	
1011		----		----	
1017		----		----	
1026	D2622	4820		0.05	
1047		----		----	
1059	ISO14596	4400		-2.54	
1146	D4294	4400	C	-2.54	first reported 440
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	ISO8754	5080		1.66	
1316	D7751	4710		-0.63	
1320		----		----	
1326	D4292	4718		-0.58	
1328		----		----	
1384	In house	4968		0.97	
1412	D4294	4944		0.82	
1431		----		----	
1433		----		----	
1438		----		----	
1460	D4294	4718		-0.58	
1510	D4294	4710		-0.63	
1564	D4294	5280		2.89	
1650	D4294	4960	C	0.92	first reported 0.496 mg/kg
1740		----		----	
1748		----		----	
1799		----		----	
1850	ISO8754	4800		-0.07	
1864	ISO8754	4975		1.01	
1877		----		----	
1957		----		----	
1969		----		----	
6016		----		----	
6032		----		----	
6035		----		----	
6181		----		----	
6183	D2622	4981.00		1.05	
6197		----		----	
6266		----		----	
6310	D7751	4780		-0.19	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358		----		----	
6380	D5185	4669.54		-0.88	
6395		----		----	
	normality	OK			
	n	36			
	outliers	0			
	mean (n)	4811.6			
	st.dev. (n)	222.12			
	R(calc.)	621.9			
	st.dev.(D4294:16e1)	161.93			
	R(D4294:16e1)	453.4			



Determination of Water on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D6304-C	204		----	
211		----		----	
219		----		----	
237	D6304-C	105		----	
254		----		----	
257	D6304-A	243		----	
309	D6304-C	242		----	
325	D6304-C	271		----	
329		----		----	
333	D6304-A	630		----	
339		----		----	
349	D6304-C	198		----	
360	D6304-B	196		----	
398	ISO12937	420		----	
421	D6304-B	57		----	
432		----		----	
496	D6304-C	168		----	
614	D6304-B	116		----	
633		----		----	
634	D6304-A	139	C	----	first reported 680
657	D6304-A	194	C	----	first reported 0.20 %M/M
780	D6304-B	106		----	
823	D6304-C	201		----	
840	D6304-B	23.6		----	
862	D6304-B	280		----	
875		----		----	
902	D6304-C	419.1		----	
912	D6304	235		----	
913		----		----	
922	D6304-A	450		----	
962	D6304-C	42		----	
963	D6304-C	70		----	
974		----		----	
994	D6304-C	195		----	
1011		----		----	
1017		----		----	
1026	D6304-C	175		----	
1047		----		----	
1059	D6304-B	180		----	
1146	D6304-C	280		----	
1150		----		----	
1173	In house	421.7	C	----	first reported 971.9
1213		----		----	
1235		----		----	
1243	ISO12937	205		----	
1316	D6304-B	240		----	
1320		----		----	
1326		----		----	
1328		----		----	
1384	In house	1182	C,R(0.01)	----	first reported 2622
1412		----		----	
1431		----		----	
1433	ISO12937	455.37		----	
1438		----		----	
1460	D6304-A	200		----	
1510		----		----	
1564	D6304-C	251		----	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864		----		----	
1877	D6304-C	271		----	
1957	D6304-C	1153	C,R(0.01)	----	first reported 1365
1969		----		----	
6016	D6304-A	448		----	
6032		----		----	
6035		----		----	
6181	D6304-C	180		----	
6183		----		----	
6197		----		----	
6266		----		----	
6310	D6304-C	116		----	



lab	method	value	mark	z(targ)	remarks
6324		----		----	
6358	ISO12937	565.7		----	
6380		----		----	
6395		----		----	
normality		OK			
n		38			
outliers		2			
mean (n)		241.933			
st.dev. (n)		142.3934			
R(calc.)		398.701			
st.dev.(D6304-C:20)		(32.4973)			
R(D6304-C:20)		(90.992)			range 20 - 360 mg/kg
Compare					
R(D6304-A:20)		(112.255)			range 20 - 25000 mg/kg
R(D6304-B:20)		(222.307)			range 30 - 2100 mg/kg
R(D6304:16e1)		(454.873)			



Determination of Calcium as Ca on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D4951	500		-0.07	
211		----		----	
219	D5185	520		1.09	
237	D5185	446.1		-3.18	
254		----		----	
257	D6595	352.49	R(0.01)	-8.58	
309	D5185	510.6		0.54	
325	D5185	472		-1.68	
329		496		-0.30	
333	D5185	515		0.80	
339	INH-165	508		0.39	
349	D5185	477		-1.39	
360	D5185	480		-1.22	
398		----		----	
421	D5185	527		1.49	
432		----		----	
496	D5185	610.9	R(0.01)	6.33	
614	D5185	492.5		-0.50	
633	D6595	297.5	R(0.01)	-11.75	
634	D6595	385	C,R(0.01)	-6.70	first reported 256
657	D5185	519		1.03	
780	D5185	497		-0.24	
823	D5185	486		-0.88	
840		----		----	
862	D5185	506		0.28	
875	D5185	454		-2.72	
902		----		----	
912	D5185	517		0.91	
913		----		----	
922	D5185	538		2.12	
962	D5185	492		-0.53	
963	D5185	507.64		0.37	
974		491		-0.59	
994	D5185	551		2.87	
1011	D5185	515		0.80	
1017		----		----	
1026	D5185	495		-0.36	
1047	D5185	504		0.16	
1059	In house	496		-0.30	
1146	D5185	502.72		0.09	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	DIN51399	538		2.12	
1316	D5185	510		0.51	
1320		----		----	
1326	D5185	470		-1.80	
1328	D5185	543		2.41	
1384	D5185	551		2.87	
1412	D5185	510		0.51	
1431		521		1.14	
1433		----		----	
1438		2700	C,R(0.01)	126.85	first reported 270
1460	D5185	457		-2.55	
1510		----		----	
1564	D4951	470		-1.80	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850	In house	450		-2.95	
1864		534		1.89	
1877		----		----	
1957	D5185	499		-0.13	
1969		----		----	
6016	D6595	298.37	R(0.01)	-11.70	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197	D4951	502		0.05	
6266		----		----	



lab	method	value	mark	z(targ)	remarks
6310	D7751	499		-0.13	
6324		----		----	
6358	D5185	477.3		-1.38	
6380		504.540		0.19	
6395		----		----	
	normality	OK			
	n	44			
	outliers	6			
	mean (n)	501.17			
	st.dev. (n)	25.558			
	R(calc.)	71.56			
	st.dev.(D5185:18)	17.335			
	R(D5185:18)	48.54			



Determination of Phosphorus as P on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D4951	950		-1.03	
211		----		----	
219	D5185	972		-0.58	
237	D5185	974	C	-0.53	first reported 1365
254	D5185	953.763		-0.95	
257	D6595	950.73		-1.01	
309	D5185	938.6		-1.26	
325	D5185	986		-0.29	
329		990		-0.20	
333	D5185	1030		0.62	
339	INH-165	980		-0.41	
349	D5185	949		-1.05	
360	D5185	978		-0.45	
398		----		----	
421	D5185	1011		0.23	
432		----		----	
496	D5185	1158.8		3.27	
614	D5185	932.7		-1.38	
633	D6595	1083.5		1.72	
634	D6595	934		-1.36	
657	D5185	998		-0.04	
780	D5185	1055		1.13	
823	D5185	940		-1.23	
840	D5185	976		-0.49	
862	D5185	964		-0.74	
875	D5185	955		-0.93	
902	D5185	1036		0.74	
912	D5185	1095		1.96	
913		----		----	
922	D5185	992		-0.16	
962	D5185	998		-0.04	
963	D5185	1013.0		0.27	
974		953		-0.97	
994	D5185	1102		2.10	
1011	D5185	968		-0.66	
1017		----		----	
1026	D5185	1058		1.20	
1047	D5185	955		-0.93	
1059	In house	957		-0.88	
1146	D5185	965.25		-0.71	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	DIN51399	990		-0.20	
1316	D5185	1080		1.65	
1320		----		----	
1326	D5185	940		-1.23	
1328	D5185	995		-0.10	
1384	D5185	1082		1.69	
1412	D5185	1043		0.89	
1431		1100		2.06	
1433		----		----	
1438		886		-2.35	
1460	D5185	1095		1.96	
1510		----		----	
1564	D4951	1006		0.12	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850		----		----	
1864		1095		1.96	
1877		----		----	
1957	D5185	957		-0.88	
1969		----		----	
6016	D6595	972.44		-0.57	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197	D4951	999		-0.02	
6266		----		----	



lab	method	value	mark	z(targ)	remarks
6310	D7751	982		-0.37	
6324		----		----	
6358	D5185	961.0		-0.80	
6380		1061.613		1.27	
6395		----		----	
	normality	OK			
	n	52			
	outliers	0			
	mean (n)	999.95			
	st.dev. (n)	57.628			
	R(calc.)	161.36			
	st.dev.(D5185:18)	48.562			
	R(D5185:18)	135.97			



Determination of Zinc as Zn on sample #21075; results in mg/kg

lab	method	value	mark	z(targ)	remarks
178		----		----	
179	D4951	1070		-0.45	
211		----		----	
219	D5185	1125		0.38	
237	D5185	1013		-1.32	
254	D5185	973.773		-1.92	
257	D6595	1037		-0.96	
309	D5185	1095.3		-0.07	
325	D5185	1095		-0.07	
329		1126		0.40	
333	D5185	1088		-0.18	
339	INH-165	1071		-0.44	
349	D5185	1033		-1.02	
360	D5185	1008		-1.40	
398		----		----	
421	D5185	1185		1.30	
432		----		----	
496	D5185	1305		3.13	
614	D5185	1107		0.11	
633	D6595	897.4		-3.08	
634	D6595	1192		1.40	
657	D5185	1100		0.00	
780	D5185	1115		0.23	
823	D5185	1083		-0.26	
840	D5185	1110		0.16	
862	D5185	1107		0.11	
875	D5185	960		-2.13	
902	D5185	1119		0.29	
912	D5185	975		-1.90	
913		----		----	
922	D5185	1126		0.40	
962	D5185	1047		-0.80	
963	D5185	1072.69		-0.41	
974		1018		-1.25	
994	D5185	1098		-0.03	
1011	D5185	1109		0.14	
1017		----		----	
1026	D5185	1183		1.27	
1047	D5185	1104		0.06	
1059	In house	1142		0.64	
1146	D5185	1169.5		1.06	
1150		----		----	
1173		----		----	
1213		----		----	
1235		----		----	
1243	DIN51399	1200		1.53	
1316	D5185	1110		0.16	
1320		----		----	
1326	D5185	1120		0.31	
1328	D5185	1150		0.76	
1384	D5185	1204		1.59	
1412	D5185	1184		1.28	
1431		1070		-0.45	
1433		----		----	
1438		955		-2.20	
1460	D5185	1118		0.28	
1510		----		----	
1564	D4951	1140		0.61	
1650		----		----	
1740		----		----	
1748		----		----	
1799		----		----	
1850	In house	1110		0.16	
1864		1185		1.30	
1877		----		----	
1957	D5185	1093		-0.10	
1969		----		----	
6016	D6595	1047.3		-0.80	
6032		----		----	
6035		----		----	
6181		----		----	
6183		----		----	
6197	D4951	1102		0.03	
6266		----		----	

lab	method	value	mark	z(targ)	remarks
6310	D7751	1150		0.76	
6324		----		----	
6358	D5185	1205		1.60	
6380		1084.74		-0.23	
6395		----		----	
	normality	OK			
	n	53			
	outliers	0			
	mean (n)	1099.77			
	st.dev. (n)	73.627			
	R(calc.)	206.15			
	st.dev.(D5185:18)	65.668			
	R(D5185:18)	183.87			



APPENDIX 2**Number of participants per country**

1 lab in AUSTRALIA	2 labs in PAKISTAN
1 lab in AUSTRIA	1 lab in PERU
2 labs in AZERBAIJAN	2 labs in PHILIPPINES
5 labs in BELGIUM	5 labs in POLAND
2 labs in BULGARIA	1 lab in PORTUGAL
3 labs in CHINA, People's Republic	1 lab in ROMANIA
1 lab in CROATIA	2 labs in RUSSIAN FEDERATION
1 lab in CZECH REPUBLIC	3 labs in SAUDI ARABIA
2 labs in FRANCE	2 labs in SINGAPORE
5 labs in GERMANY	1 lab in SLOVAKIA
2 labs in GREECE	1 lab in SLOVENIA
2 labs in INDIA	1 lab in SOUTH KOREA
1 lab in ISRAEL	2 labs in SPAIN
1 lab in ITALY	1 lab in SWEDEN
1 lab in JORDAN	2 labs in TANZANIA
1 lab in KAZAKHSTAN	1 lab in TUNISIA
1 lab in KENYA	2 labs in TURKEY
1 lab in MACEDONIA	1 lab in UNITED ARAB EMIRATES
2 labs in MALAYSIA	3 labs in UNITED KINGDOM
2 labs in MOROCCO	2 labs in UNITED STATES OF AMERICA
3 labs in NETHERLANDS	2 labs in VIETNAM
1 lab in NIGERIA	

APPENDIX 3**Abbreviations**

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)/R(1)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= calculation difference between reported test result and result calculated by iis
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
f+?	= possibly a false positive test result?
f-?	= possibly a false negative test result?
SDS	= Safety Data Sheet

Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 ISO5725:86
- 3 ISO5725 parts 1-6:94
- 4 ISO13528:05
- 5 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 6 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 7 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 8 J.N. Miller, Analyst, 118, 455, (1993)
- 9 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 10 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 11 W. Horwitz and R. Albert, J. AOAC Int, 79.3, 589-621, (1996)
- 12 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)