

Institute for  
Interlaboratory Studies

## Results of Proficiency Test Jet Fuel A1 September 2022

**Organized by:** Institute for Interlaboratory Studies  
Spijkenisse, the Netherlands

**Author:** ing. G.A. Oosterlaken-Buijs  
**Correctors:** ing. M. Meijer & ing. R.J. Starink  
**Approved by:** ing. A.S. Noordman-de Neef

**Report:** iis22J02

November 2022

**CONTENTS**

1	INTRODUCTION .....	3
2	SET UP .....	3
2.1	ACCREDITATION.....	4
2.2	PROTOCOL.....	4
2.3	CONFIDENTIALITY STATEMENT .....	4
2.4	SAMPLES .....	4
2.5	STABILITY OF THE SAMPLES.....	9
2.6	ANALYZES .....	9
3	RESULTS.....	9
3.1	STATISTICS .....	10
3.2	GRAPHICS .....	11
3.3	Z-SCORES.....	11
4	EVALUATION .....	12
4.1	EVALUATION PER SAMPLE AND PER TEST.....	12
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	16
4.3	COMPARISON OF THE PROFICIENCY TEST OF SEPTEMBER 2022 WITH PREVIOUS PTS.....	18
Appendices:		
1.	Data, statistical and graphic results .....	20
2.	z-scores of Distillation .....	86
3.	Equipment used in Particle Size distribution.....	88
4.	z-scores Particle Size Distribution .....	90
5.	Number of participants per country.....	91
6.	Abbreviations and literature .....	92

## 1 INTRODUCTION

Since 1995 the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the analysis of Jet Fuel A1 twice a year based on the latest version of the "Aviation Fuel Quality Requirements for Jointly Operated Systems (AFQRJOS)", sometimes referred to as the "Joint Fuelling System Check List for Jet A-1". The interlaboratory study on Jet Fuel was extended with separate PTs for the determination of Particle Size Distribution, BOCLE, FAME, JFTOT and Particulate Contamination. During the annual proficiency testing program 2022/2023 it was decided to continue the round robin for the analysis of Jet Fuel A1.

In this interlaboratory study registered for participation:

- 151 laboratories in 63 countries for regular analyzes in Jet Fuel A1 iis22J02
- 27 laboratories in 15 countries on the BOCLE determination iis22J02BOCLE
- 60 laboratories in 33 countries on the Particle Size determination iis22J02PS
- 74 laboratories in 36 countries on the FAME determination iis22J02FAME
- 93 laboratories in 48 countries on the JFTOT determination iis22J02JF
- 54 laboratories in 31 countries on the Particulate Contamination determination iis22J02CP

In total 168 laboratories in 66 countries registered for participation in one or more proficiency tests, see appendix 5 for the number of participants per country. In this report the results of the Jet Fuel A1 proficiency tests are presented and discussed. This report is also electronically available through the iis website [www.iisnl.com](http://www.iisnl.com).

## 2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory.

In this proficiency test the participants received, depending on the registration, from one up to seven different samples of Jet Fuel, see table below.

Sample ID	PT ID	Quantity	Purpose
#22150	iis22J02	2x 1 L	Regular analyzes
#22151	iis22J02BOCLE	1x 250 mL	BOCLE
#22152	iis22J02PS	1x 0.5 L	Particle Size Distribution
#22153	iis22J02FAME	1x 100 mL	FAME
#22154	iis22J02FAME	1x 100 mL	FAME
#22155	iis22J02JF	1x 1 L, 70% filled	JFTOT
#22156	iis22J02CP	4x 1 L	Particulate Contamination

Table 1: Jet Fuel samples used in PT iis22J02

Participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

## 2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

## 2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website [www.iisnl.com](http://www.iisnl.com), from the FAQ page.

## 2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

## 2.4 SAMPLES

For the preparation of the sample for the regular analyzes in Jet Fuel A1 a batch of approximately 500 liters of Jet Fuel A1 was obtained from a third party. After homogenization 350 amber glass bottles of 1 L were filled and labelled #22150.

The homogeneity of the subsamples was checked by the determination of Density at 15 °C in accordance with ASTM D4052 on 10 stratified randomly selected subsamples.

	Density at 15 °C in kg/m <sup>3</sup>
sample #22150-1	803.40
sample #22150-2	803.40
sample #22150-3	803.39
sample #22150-4	803.40
sample #22150-5	803.39
sample #22150-6	803.40
sample #22150-7	803.39
sample #22150-8	803.40
sample #22150-9	803.40
sample #22150-10	803.39

Table 2: homogeneity test results of subsamples #22150

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15 °C in kg/m <sup>3</sup>
r (observed)	0.01
reference test method	ASTM D4052:22
0.3 x R (reference test method)	0.15

Table 3: evaluation of the repeatability of subsamples #22150

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the sample for the BOCLE determination in Jet Fuel A1 a batch of approximately 13 liters of Jet Fuel A1 was obtained from a local refinery. After homogenization 45 amber glass bottles of 250 mL were filled and labelled #22151. The homogeneity of the subsamples was checked by the determination of Density at 15 °C in accordance with ASTM D4052 on 8 stratified randomly selected subsamples.

	Density at 15 °C in kg/m <sup>3</sup>
sample #22151-1	814.13
sample #22151-2	814.13
sample #22151-3	814.12
sample #22151-4	814.14
sample #22151-5	814.13
sample #22151-6	814.14
sample #22151-7	814.13
sample #22151-8	814.13

Table 4: homogeneity test results of subsamples #22151

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15 °C in kg/m <sup>3</sup>
r (observed)	0.02
reference test method	ASTM D4052:22
0.3 x R (reference test method)	0.15

Table 5: evaluation of the repeatability of subsamples #22151

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the sample for the Particle Size determination in Jet Fuel A1 a batch of approximately 85 liters of Jet Fuel A1 was obtained from a third party. After homogenization 100 amber glass bottles of 0.5 L were filled and labelled #22152. Each bottle was spiked with 1 mL of Lube oil which contained suspended Arizona Dust before filling with Jet Fuel A1. The homogeneity of the subsamples was checked by the determination of Particle Size Distribution in accordance with IP565 on 8 stratified randomly selected subsamples.

	> 4 $\mu\text{m}$ (c) counts/mL	> 6 $\mu\text{m}$ (c) counts/mL	> 14 $\mu\text{m}$ (c) counts/mL
sample #22152-1	9191	3413	134
sample #22152-2	9436	3589	156
sample #22152-3	9534	3583	147
sample #22152-4	9494	3580	142
sample #22152-5	9726	3672	126
sample #22152-6	9459	3496	124
sample #22152-7	9664	3668	153
sample #22152-8	9677	3655	153

Table 6: homogeneity test results of subsamples #22152

From the above test results the relative standard deviations (RSD) were calculated and compared with 0.3 times the corresponding average relative standard deviation obtained from seventeen iis PTs of IP565 test data from 2014 - 2022 in agreement with the procedure of ISO13528, Annex B2 in the next table.

	> 4 $\mu\text{m}$ (c)	> 6 $\mu\text{m}$ (c)	> 14 $\mu\text{m}$ (c)
RSD% (observed)	2	3	9
reference method	iis PTs	iis PTs	iis PTs
0.3 x RSD% (reference method)	5	7	10

Table 7: evaluation of the relative standard deviations of subsamples #22152

The calculated relative standard deviations are in agreement with 0.3 times the corresponding average relative standard deviation obtained from the previous iis PTs. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the samples for the FAME determination in jet Fuel A1 it was decided to prepare two different PT samples: one with a low(er) and one with a high(er) level of FAME. For the low(er) level of FAME a batch of approximately 20 liters of Jet Fuel A1 was made available from the retain materials from earlier PTs on Jet Fuel A1 FAME. After homogenization 95 amber glass bottles of 100 mL were filled and labelled #22153. For the high(er) level of FAME a batch of approximately 20 liters of Jet Fuel A1 was made available from the retain materials from earlier PTs on Jet Fuel A1 FAME. After homogenization 95 amber glass bottles of 100 mL were filled and labelled #22154. The homogeneity of the subsamples #22153 and #22154 was checked by the determination of FAME in accordance with method IP585 on 8 stratified randomly selected subsamples.

	FAME in mg/kg #22153	FAME in mg/kg #22154
sample 1	20.7	48.1
sample 2	20.3	48.3
sample 3	19.3	46.6
sample 4	20.1	46.9
sample 5	20.6	47.3
sample 6	20.7	46.2
sample 7	20.1	40.9 D(0.01)
sample 8	20.3	47.6

Table 8: homogeneity test results of subsamples #22153 and #22154

From the above test results the repeatabilities were calculated and compared with 0.3 times the corresponding reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	FAME in mg/kg #22153	FAME in mg/kg #22154
r (observed)	1.3	2.2
reference test method	IP585:10	IP585:10
0.3 x R (reference test method)	1.8	3.9

Table 9: evaluation of the repeatabilities of subsamples #22153 and #22154

Subsample #22154-7 is a Dixon outlier and therefore excluded for statistical evaluation of the homogeneity.

The calculated repeatabilities are in agreement with 0.3 times the corresponding reproducibility of the reference test methods. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the sample for the JFTOT determination in Jet Fuel A1 a batch of approximately 140 liters of Jet Fuel was made positive. After homogenization 130 amber glass bottles of 1 L were filled at a level of 70% and labelled #22155.

The homogeneity of the subsamples was checked by the determination of Density at 15 °C in accordance with ASTM D4052 on 8 stratified randomly selected subsamples.

	Density at 15 °C in kg/m <sup>3</sup>
sample #22155-1	803.41
sample #22155-2	803.42
sample #22155-3	803.41
sample #22155-4	803.42
sample #22155-5	803.41
sample #22155-6	803.41
sample #22155-7	803.42
sample #22155-8	803.41

Table 10: homogeneity test results of subsamples #22155

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15 °C in kg/m <sup>3</sup>
r (observed)	0.01
reference test method	ASTM D4052:22
0.3 x R (reference test method)	0.15

Table 11: evaluation of the repeatability of subsamples #22155

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the sample for Particulate Contamination determination in Jet Fuel A1 a batch of approximately 400 liters of Jet Fuel A1 was obtained from a third party. After homogenization 290 amber glass bottles of 1 L were filled and labelled #22156.

The homogeneity of the subsamples was checked by the determination of Density at 15 °C in accordance with ASTM D4052 on 8 stratified randomly selected subsamples.

	Density at 15 °C in kg/m <sup>3</sup>
sample #22156-1	791.44
sample #22156-2	791.44
sample #22156-3	791.44
sample #22156-4	791.44
sample #22156-5	791.45
sample #22156-6	791.44
sample #22156-7	791.44
sample #22156-8	791.44

Table 12: homogeneity test results of subsamples #22156

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15 °C in kg/m <sup>3</sup>
r (observed)	0.01
reference test method	ASTM D4052:22
0.3 x R (reference test method)	0.15

Table 13: evaluation of the repeatability of subsamples #22156

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.



Depending on the registration of the participant the appropriate set of PT samples was sent on August 17, 2022. An SDS was added to the sample package.

## 2.5 STABILITY OF THE SAMPLES

The stability of Jet Fuel A1 packed in the amber glass bottles was checked. The type of bottle was chosen in accordance with ASTM D4306:20. The material has been found sufficiently stable for the period of the proficiency test.

## 2.6 ANALYZES

The participants were requested to determine on sample #22150: Appearance, Total Acidity, Aromatics by FIA, Mono Aromatics (MAH), Di Aromatics (DAH) and Total Aromatics by HPLC, Color Saybolt (automated and manual), Copper Corrosion 2 hrs at 100 °C, Density at 15 °C, Distillation at 760 mmHg (IBP, Temperature at 10%, 50%, 90% recovered and FBP), Existent Gum (unwashed), Flash Point, Freezing Point, Kinematic Viscosity at -20 °C, Mercaptan Sulfur as S, MSEP, Naphthalenes, Smoke Point, Specific Energy (Net) on Sulfur free basis and Total Sulfur.

On sample #22151 it was requested to determine: Wear Scar Diameter.

On sample #22152 it was requested to determine: Particle Size Distribution.

On samples #22153 and #22154 it was requested to determine: FAME content.

On sample #22155 it was requested to determine: Copper and JFTOT.

On sample #22156 it was requested to determine: Particulate Contamination.

It was explicitly requested to treat the samples as if they were routine samples and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website [www.iisnl.com](http://www.iisnl.com).

## 3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination

Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

### 3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...>' or '>...>' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

The assigned value is determined by consensus based on the test results of the group of participants after rejection of the statistical outliers and/or suspect data.

According to ISO13528 all (original received or corrected) results per determination were submitted to outlier tests. In the iis procedure for proficiency tests, outliers are detected prior to calculation of the mean, standard deviation and reproducibility. For small data sets, Dixon (up to 20 test results) or Grubbs (up to 40 test results) outlier tests can be used. For larger data sets (above 20 test results) Rosner's outlier test can be used. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

### 3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve (dotted line) was projected over the Kernel Density Graph (smooth line) for reference. The Gauss curve is calculated from the consensus value and the corresponding standard deviation.

### 3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements (derived from e.g. ISO or ASTM test methods), the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$Z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The  $Z_{(\text{target})}$  scores are listed in the test result tables in appendix 1.

Absolute values for  $z < 2$  are very common and absolute values for  $z > 3$  are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z  < 1$	good
$1 <  z  < 2$	satisfactory
$2 <  z  < 3$	questionable
$3 <  z $	unsatisfactory

## 4 EVALUATION

In this proficiency test some problems were encountered with the dispatch of the samples. For the regular analyzes Jet Fuel A1 thirteen participants reported test results after the final reporting date and seventeen other participants did not report any test results.

For the BOCLE round eight participants did not report any test results.

For the Particle Size Distribution round three participants reported test results after the final reporting date and sixteen other participants did not report any test results.

For the FAME round nine participants reported test results after the final reporting date and nineteen other participants did not report any test results.

For the JFTOT round six participants reported test results after the final reporting date and twentythree other participants did not report any test results.

For the Particulate Contamination round four participants reported test results after the final reporting date and eleven other participants did not report any test results.

Not all participants were able to report all tests requested.

In total 168 participants reported 2754 numerical test results. Observed were 48 outlying test results, which is 1.7%. In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

Not all data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

### 4.1 EVALUATION PER SAMPLE AND PER TEST

In this section the reported test results are discussed per sample and per test. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the original data in appendix 1. The abbreviations, used in these tables, are explained in appendix 6.

In the iis PT reports ASTM test methods are referred to with a number (e.g. D3242) and an added designation for the year that the test method was adopted or revised (e.g. D3242:11). When a method has been reapproved an “R” will be added and the year of approval (e.g. D3242:11R17).

Since the Joint Fuelling System Check List for Jet-A1 is continuously updated, the participants are advised to monitor the updates. The latest version at the time of this Round Robin is DEF STAN 91-091/Issue 14, March 2022 and ASTM D1655:22. One must keep in mind that ISO test methods are not mentioned in the “Checklist”.

#### **sample #22150**

Appearance: This determination was not problematic. All reporting participants agreed about the appearance being Clear and Bright (Pass).

Total Acidity: This determination was problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D3242:11R17.

Aromatics by FIA: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with ASTM D1319:20a.

Mono Aromatics (MAH) by HPLC: This determination was problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D6379:21e1.

Di Aromatics (DAH) by HPLC: This determination was very problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not at all in agreement with the requirements of ASTM D6379:21e1.

Total Aromatics by HPLC in %M/M: This determination was problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D6379:21e1.

Total Aromatics by HPLC in %V/V: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ASTM D6379:21e1.

Color Saybolt (automated): The determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D6045:20.  
When the test results were evaluated separately for a 50 mm and 100 mm cell both calculated reproducibilities are in agreement.

Color Saybolt (manual): The determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of of ASTM D156:15.

Copper Corrosion: This determination was not problematic. All reporting participants agreed on a test result of 1 (1a/1b).

Density at 15 °C: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier in agreement with the requirements of ASTM D4052:22.

Distillation at 760 mmHg: This determination may be problematic dependent on the method used. Five statistical outliers were observed over five parameters. The calculated reproducibilities after rejection of the statistical outliers are all in agreement with the requirements of the automated method of ASTM D86:20b.  
When compared to the manual method requirements of ASTM D86:20b the calculated reproducibility is only for IBP not in agreement.

Existent Gum (unwashed): This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with ASTM D381:22 and IP540:08R19.

Flash Point: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of IP170:21.

Freezing Point: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D2386:19.

Kinematic Viscosity at -20 °C: This determination was very problematic. Seven statistical outliers were observed. It was decided not to calculate z-scores as the calculated reproducibility was too large compared to the requirements ASTM D445:21e1 and IP71-1:20. The requirements of ASTM D445 and IP71 were updated and these requirements became very strict compared to previous versions of the requirements ASTM D445:19 and IP71-1:96 respectively.

Mercaptan Sulfur: This determination was not problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D3227:16.

MSEP: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D3948:20.

Naphthalenes: This determination was not problematic. Six statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D1840:07R17 procedure B and A.

Smoke Point: This determination may be problematic depending on the procedure used. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D1322:22 manual procedure but not with the strict requirements of ASTM D1322:22 automated procedure. When the test results were evaluated separately for the manual or the automated procedure the calculated reproducibility for the automated procedure is still not in agreement with the respective requirements of ASTM D1322:22 automated procedure.

Specific Energy: This determination was not problematic. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D3338:20a.

Total Sulfur: This determination may be problematic depending on the procedure used. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D5453:19a nor with ASTM D2622:21 but is in agreement with ASTM D4294:21.

### **sample #22151**

BOCLE: This determination may be problematic depending on the procedure used. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ASTM D5001:19e1 semi-automated but is not in agreement with ASTM D5001:19e1 full-automated.

When the test results were evaluated separately for semi-automated and full-automated methods, the calculated reproducibility for full-automated is still not in agreement with the requirements of ASTM D5001:19e1 full-automated.

### **sample #22152 Particle Size Distribution Determination**

The Joint Fuelling System Check List for Jet-A1 lists test methods IP565 and IP577 as the reference test methods to determine the Particle Size Distribution in Jet Fuel A1.

Almost all reporting participants mentioned to have used IP565. Four participants used IP577 and one participant used IP564 which is not mentioned in the Checklist as test method since 2020. The test results from IP577 and IP564 were excluded from statistical evaluation as it was observed in previous iis PTs that IP577 as well as IP564 give deviating results compared to IP565.

It is observed that all of the reporting participants used ISO11171 for the calibration and almost all of the reporting participants used ISO4406 for calculating the scale numbers from the counts per mL.

IP counts: This determination may be problematic depending on the particle size channel used. No statistical outliers were observed but thirty-nine test results were excluded over six parameters. The calculated reproducibilities after rejection of the suspect data are in agreement with the requirements of IP565:13 for  $\geq 6 \mu\text{m(c)}$  and  $\geq 14 \mu\text{m(c)}$  but are not in agreement for  $\geq 4 \mu\text{m(c)}$ ,  $\geq 21 \mu\text{m(c)}$ ,  $\geq 25 \mu\text{m(c)}$  and  $\geq 30 \mu\text{m(c)}$ .

ISO class: The determination expressed in ISO scale numbers may be problematic. No statistical outliers were observed but twenty-one test results were excluded over three parameters. The calculated reproducibilities after rejection of the suspect data are in agreement with the indicative requirements of IP565:13 Annex C for particle size channels  $\geq 4 \mu\text{m(c)}$  and  $\geq 6 \mu\text{m(c)}$  but is not in agreement for particle size channel  $\geq 14 \mu\text{m(c)}$ .

**sample #22153 FAME content determination**

GCMS/HPLC: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of IP585:21 or IP590:10. When the test results were evaluated separately for IP585 or IP590 the calculated reproducibilities are in agreement with the respective requirements of IP585:21 or IP590:10.

FTIR: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of IP583:15.

**sample #22154 FAME content determination**

GCMS/HPLC: This determination was problematic. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of IP585:21 or IP590:10. When the test results were evaluated separately for IP585 and IP590 the calculated reproducibilities are still not in agreement with the respective requirements.

FTIR: This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of IP583:15.

**sample #22155 JFTOT determination**

Copper: Only three participants reported a numeric test result. Therefore, no z-scores are calculated.

JFTOT: Using the criteria from AFQRJOS on the test results (including the laboratories that did not report a pass or fail) almost all laboratories would rate the sample as a fail, while 5 laboratories would rate it as a pass.

**sample #22156**

Particulate Contamination: This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in full agreement with the requirements of the ASTM D5452:20.

**4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES**

A comparison has been made between the reproducibility as declared by the reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility ( $2.8 \times$  standard deviation) and the target reproducibility derived from reference methods are presented in the next tables.



Parameter	unit	n	average	2.8 * sd	R(lit)
Appearance		101	Cl&Br (Pass)	n.a.	n.a.
Total Acidity	mg KOH/g	83	0.0015	0.0021	0.0016
Aromatics by FIA	%V/V	58	18.9	2.2	3.1
Mono Aromatics (MAH) by HPLC	%M/M	34	21.4	1.7	1.3
Di Aromatics (DAH) by HPLC	%M/M	36	1.04	0.34	0.16
Total Aromatics by HPLC	%M/M	34	22.4	2.0	1.4
Total Aromatics by HPLC	%V/V	45	20.2	1.7	1.2
Color Saybolt (automated)		41	30.0	0.6	1.2
Color Saybolt (manual)		44	29.9	0.8	2
Copper Corrosion 2 hrs at 100 °C		104	1 (1a/1b)	n.a.	n.a.
Density at 15 °C	kg/m <sup>3</sup>	123	803.4	0.3	0.5
Initial Boiling Point	°C	119	150.5	6.6	8.3
Temp at 10% recovered	°C	119	170.3	2.6	3.7
Temp at 50% recovered	°C	120	195.7	2.2	3.0
Temp at 90% recovered	°C	119	232.9	2.8	3.5
Final Boiling Point	°C	117	252.9	4.1	7.1
Existent Gum (unwashed)	mg/100mL	58	0.6	0.9	3.1
Flash Point	°C	123	43.4	3.1	3.2
Freezing Point	°C	103	-55.5	1.9	2.5
Kinematic Viscosity at -20 °C	mm <sup>2</sup> /s	80	3.738	0.101	(0.018)
Mercaptan Sulfur as S	%M/M	74	0.0004	0.0002	0.0003
MSEP		95	95.9	6.2	7.0
Naphthalenes	%V/V	59	0.72	0.05	0.08
Smoke Point	mm	99	22.7	3.0	3.6
Specific Energy (Net)	MJ/kg	63	43.196	0.040	0.046
Total Sulfur	mg/kg	99	352	67	47
BOCLE	mm	18	0.65	0.05	0.06

Table 14: reproducibilities of tests on samples #22150 and #22151

For results between brackets no z-scores are calculated.

Parameter	unit	n	average	2.8 * sd	R(lit)
Particle Size ≥4 μm (c)	counts/mL	37	14672	3575	1660
Particle Size ≥6 μm (c)	counts/mL	36	4405	1032	978
Particle Size ≥14 μm (c)	counts/mL	36	155	89	89
Particle Size ≥21 μm (c)	counts/mL	36	20	26	18
Particle Size ≥25 μm (c)	counts/mL	36	7	13	8
Particle Size ≥30 μm (c)	counts/mL	35	2	5	3
Particle Size ≥4 μm (c)	ISO scale	31	21.0	0.0	1.0
Particle Size ≥6 μm (c)	ISO scale	31	19.1	0.7	1.0
Particle Size ≥14 μm (c)	ISO scale	31	14.5	1.6	1.4

Table 15: reproducibilities of tests on sample #22152 according to IP565

Parameter	unit	n	average	2.8 * sd	R(lit)
FAME content GCMS/HPLC	mg/kg	35	23.4	8.0	6.8
FAME content FTIR	mg/kg	19	31.2	4.8	6.5

Table 16: reproducibilities of tests on sample #22153

Parameter	unit	n	average	2.8 * sd	R(lit)
FAME content GCMS/HPLC	mg/kg	34	56.2	19.9	15.3
FAME content FTIR	mg/kg	19	63.6	12.4	8.1

Table 17: reproducibilities of tests on sample #22154

Parameter	unit	n	average	2.8 * sd	R(lit)
Copper as Cu	µg/kg	4	>500	n.e.	n.e.
VTR (visual)		44	>2	n.a.	n.a.
ITR (interferometric)	Nm	3	>85	n.a.	n.a.
ETR (elliptometric)	nm	5	>85	n.a.	n.a.
Delta P	mmHg	55	>25	n.a.	n.a.
JFTOT Evaluation (Pass/Fail)		51	Fail	n.a.	n.a.
Particulate Contamination	mg/L	42	0.34	0.32	0.34

Table 18: reproducibilities of tests on sample #22155 and #22156

Without further statistical calculations it can be concluded that for many tests there is a good compliance of the group of participants with the reference test methods. The problematic tests have been discussed in paragraph 4.1.

#### 4.3 COMPARISON OF THE PROFICIENCY TEST OF SEPTEMBER 2022 WITH PREVIOUS PTS

	September 2022	March 2022	September 2021	March 2021	September 2020
Number of reporting laboratories	168	80	160	91	152
Number of test results	2754	1400	3091	1676	2992
Number of statistical outliers	48	28	42	58	94
Percentage of statistical outliers	1.7%	2.0%	1.4%	3.5%	3.1%

Table 19: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared to the requirements of the reference test methods. The conclusions are given in the following table.

Parameter	September 2022	March 2022	September 2021	March 2021	September 2020
Total Acidity	-	+/-	-	-	-
Aromatics by FIA	+	+	+	+	+
Aromatics by HPLC	-	-	-	+	+
Color Saybolt (automated)	++	--	--	--	--

Parameter	September 2022	March 2022	September 2021	March 2021	September 2020
Color Saybolt (manual)	++	-	--	-	-
Density at 15 °C	+	++	++	++	++
Distillation at 760 mmHg	+	+	+	+	+
Existent Gum (unwashed)	++	++	++	++	++
Flash Point	+/-	+/-	+	+/-	+/-
Freezing Point	+	+	+	+	+
Kinematic Viscosity at -20 °C	(--)	-	-	-	-
Mercaptan Sulfur	+	+	+/-	+	+
MSEP	+	+	+	+	+
Naphthalenes	+	+	+/-	+/-	+
Smoke Point	+	++	+	++	++
Specific Energy (Net)	+	+/-	+/-	+/-	-
Total Sulfur	-	-	-	+	+
BOCLE	+	n.a.	-	n.a.	+/-
PC					
IP565 cumulative counts/mL	-	-	-	--	-
IP565 ISO scale numbers	+	+	+	+	+
FAME content	-	n.a.	-	n.a.	-
JFTOT finding correct Pass/Fail	+	n.a.	+	n.a.	+
Particulate Contamination	+/-	n.a.	+	n.a.	+/-

Table 20: comparison of determinations to the reference test methods

The following performance categories were used:

- ++ : group performed much better than the reference test method
- + : group performed better than the reference test method
- +/- : group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e. : not evaluated

**APPENDIX 1**

**Determination of Appearance on sample #22150;**

lab	method	value	z(targ)	lab	method	value	z(targ)
52	Visual	C & B	----	1062		----	----
62	Visual	pass	----	1064	Visual	C&B	----
120	Visual	C&B	----	1065		----	----
140	Visual	C&B	----	1082		----	----
150	Visual	Clear, Bright, free from solid matter	----	1097	Visual	Clair et limpide	----
159	Visual	C&B	----	1105	Visual	C&B	----
169	Visual	Pass	----	1109		----	----
171	Visual	Clear & Bright	----	1121	Visual	Clear & Bright	----
175	D4176	C+B (Pass)	----	1126		----	----
177		----	----	1140	Visual	C&B	----
215	Visual	claire et limpide	----	1182		----	----
221	Visual	Clear & Brighth	----	1191	D4176-1	Pass	----
224		----	----	1205		----	----
225	Visual	Clear & Bright	----	1237	Visual	clear and bright	----
228	Visual	clear and bright	----	1275	Visual	CBFSMW *)	----
235	Visual	C&B	----	1279		----	----
237	Visual	C&B	----	1299	Visual	CL&BR	----
238	Visual	Bright & Clear	----	1316		----	----
253	Visual	Clear & Bright	----	1318	Visual	Clear and bright	----
254	Visual	Clear & bright	----	1320	Visual	Clear&Bright	----
256	Visual	Clear & bright	----	1357	Visual	Clear & Bright	----
258		----	----	1372	Visual	C&B	----
273	Visual	Clear & Bright	----	1397		----	----
311	Visual	clear	----	1399		----	----
317	Visual	Br/Cl	----	1412	Visual	Clear&Bright	----
323	Visual	ACBFFSOUW	----	1417	Visual	Clear & Bright	----
328	Visual	C&B	----	1444	Visual	Clear	----
333		----	----	1455	Visual	Bright and Clear	----
334	Visual	clear & bright	----	1496		----	----
335	Visual	CL&B	----	1538		----	----
365	Visual	C+B	----	1575	Visual	C&B	----
372	Visual	C & B	----	1585	Visual	clear & bright	----
391	Visual	C&B	----	1586	Visual	clear & bright	----
396	Visual	Clear & Bright	----	1587		Br&Cl	----
399	Visual	C&B	----	1610	Visual	Bright & Clear	----
440	Visual	Clear and Bright	----	1613	Visual	B&C	----
445	Visual	C&B	----	1616	Visual	Clear & Bright	----
447	Visual	Clear & Bright	----	1631		----	----
460		----	----	1634	Visual	C&B	----
467	Visual	Clear & Bright	----	1636	Visual	C&B	----
480		----	----	1669	Visual	CyB	----
496	Visual	clear&bright	----	1683	D6986	pasa	----
603		Clear & Bright	----	1688	Visual	clear	----
608	Visual	C&B and visually free from solid matter	----	1715		----	----
631	Visual	clear & bright	----	1720		----	----
657	Visual	Bright & Clear	----	1724	Visual	Clear and bright	----
736	Visual	clear and bright	----	1730		----	----
823	Visual	C&B	----	1757		----	----
824	Visual	Clear & Bright	----	1776		----	----
851	D4176	Bright & Clear	----	1833	In house	C&B	----
854		----	----	1852	Visual	Bright, clear	----
862		----	----	1913		----	----
869		----	----	1944	Visual	clear	----
904	Visual	C&B	----	1961		----	----
914	Visual	Clear and Bright	----	2130	Visual	C&B	----
962	Visual	Clear & Bright	----	2133	Visual	Clear & Bright	----
963	Visual	Bright & Clear	----	6028		----	----
970	Visual	Clear & Bright	----	6041		----	----
974	Visual	Clear & Bright	----	6054		----	----
994	Visual	c & b	----	6075	Visual	Clear & Bright	----
995	Visual	C&B	----	6114	Visual	Clear & Bright	----
996		----	----	6135		----	----
997	Visual	Clear&Bright	----	6142	Visual	C&B	----
1011		----	----	6203	Visual	Clear & Bright	----
1016		----	----	6244		----	----
1019		----	----	6266		----	----
1023	Visual	Clear and bright	----	6274		----	----
1039	Visual	Clear & Bright	----	6312		----	----
1040		----	----	6315	Visual	clear, bright	----
1049	Visual	Br & Cl	----	6321		----	----
1059	Visual	clear & bright	----	6324	Visual	clear and bright	----

lab	method	value	z(targ)	lab	method	value	z(targ)
6332	Visual	Clear and Bright	----	6421		----	----
6344		----	----	6438	Visual	Clear, bright and free from solid matter	----
6346		----	----	6479		----	----
6364	Visual	Clear & Bright	----	6487		----	----
6384		----	----				
n		101					
mean (n)		Clear and Bright (Pass)					



\*) CBFSMW = Clear, bright and visually free from solid matter and undissolved water at ambient fuel temperature

## Determination of Total Acidity on sample #22150; results in mg KOH/g

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D3242	0.001		-0.88	1062	D3242	0		-2.67
62	D3242	0.001		-0.88	1064	D3242	0.0028		2.33
120	D3242	0.00169	C	0.35	1065	D664-A	0.003		2.69
140	D3242	0.0011		-0.70	1082	D3242	0.000814		-1.21
150		----		----	1097	D3242	0.0019		0.73
159	D3242	0.00169	C	0.35	1105	D3242	0.00055		-1.68
169	D3242	0.003		2.69	1109		----		----
171	D3242	0.001		-0.88	1121		----		----
175		----		----	1126		----		----
177	D3242	0.001		-0.88	1140	D3242	0.001		-0.88
215		----		----	1182		----		----
221		----		----	1191	D3242	0.000488		-1.79
224		----		----	1205		----		----
225	D3242	0.0012		-0.52	1237		----		----
228	D3242	0.00132		-0.31	1275	IP354	0.001		-0.88
235	D3242	0.002		0.90	1279		----		----
237	D3242	0.001		-0.88	1299	D3242	0.001		-0.88
238		----		----	1316	D3242	0.001		-0.88
253	D3242	0.0017		0.37	1318	D3242	0.0005		-1.77
254		----		----	1320	D3242	0.0006		-1.59
256		----		----	1357	D3242	0.001		-0.88
258		----		----	1372		0.0011		-0.70
273	D3242	0.0011		-0.70	1397		----		----
311	D3242	0.001		-0.88	1399		----		----
317		----		----	1412	D3242	0.003		2.69
323	D3242	0.001		-0.88	1417		----		----
328		----		----	1444		----		----
333	D3242	0.001		-0.88	1455	D3242	0.0003		-2.13
334	D3242	0.002		0.90	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D3242	0.001		-0.88	1585	D3242	0.0013		-0.35
391	D3242	0.002		0.90	1586	D3242	0.0023	C	1.44
396	D3242	0.0012		-0.52	1587		----		----
399		----		----	1610	IP354	0.001		-0.88
440		----		----	1613	D3242	0.0016		0.19
445		----		----	1616	D3242	0.0012		-0.52
447	D3242	0.033	R(0.01)	56.23	1631		----		----
460		----		----	1634	D3242	0.00125		-0.43
467	D3242	0.00087		-1.11	1636	D3242	0.00198		0.87
480		----		----	1669	D3242	0.032	R(0.01)	54.44
496	D3242	0.0005		-1.77	1683	D3242	0.009	C,R(0.01)	13.40
603		----		----	1688		----		----
608		----		----	1715		----		----
631	D3242	0.00199		0.89	1720	D3242	0.0016		0.19
657	D3242	0.002		0.90	1724	D3242	0.002		0.90
736	D3242	0.0022		1.26	1730		----		----
823	D3242	0.0012		-0.52	1757	D3242	0.001		-0.88
824	D3242	0.00053		-1.72	1776	D3242	0.0026		1.97
851	D3242	0.00213		1.14	1833		0.0015		0.01
854		----		----	1852	D3242	0.0010		-0.88
862		----		----	1913		----		----
869		----		----	1944		----		----
904	D3242	0.0015		0.01	1961		----		----
914	D3242	0.0008		-1.24	2130	IP354	0.00358		3.72
962	D3242	0.0021		1.08	2133	D3242	0.00063		-1.54
963	D3242	0.0018		0.55	6028		----		----
970		----		----	6041		----		----
974	D3242	0.0018		0.55	6054		----		----
994	D3242	0.002		0.90	6075	D3242	0.0037		3.94
995	D3242	0.001		-0.88	6114	D3242	0.0014		-0.17
996		----		----	6135		----		----
997	D3242	0.0014		-0.17	6142		----		----
1011	D3242	0.002		0.90	6203	D3242	0.0020		0.90
1016	D3242	0.0010019		-0.88	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039	D3242	0.002		0.90	6312		----		----
1040	D3242	0.021	R(0.01)	34.81	6315	D3242	0.0027		2.15
1049	D3242	0.00163		0.24	6321	IP354	<0.001		----
1059	D3242	0.003		2.69	6324	D3242	0.0010		-0.88

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D3242	0.002		0.90	6421		----		----
6344	D3242	0.0025		1.80	6438		----		----
6346		----		----	6479		----		----
6364	D3242	0.00111		-0.68	6487		----		----
6384	D3242	0.0005		-1.77					
normality		OK							
n		83							
outliers		4							
mean (n)		0.00149							
st.dev. (n)		0.000761							
R(calc.)		0.00213							
st.dev.(D3242:11R17)		0.000560							
R(D3242:11R17)		0.00157							

Lab 120 first reported 0.008  
 Lab 159 first reported 0.006  
 Lab 1586 first reported 0.0053  
 Lab 1683 first reported 0.007





## Determination of Aromatics by FIA on sample #22150; results in %V/V

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D1319	19.6		0.64	1062		----		----
62	D1319	18.2		-0.61	1064	D1319	20.4		1.35
120		----		----	1065		----		----
140		----		----	1082		----		----
150		----		----	1097	D1319	18.87		-0.01
159		----		----	1105	D1319	19.24		0.32
169	D1319	17.5		-1.23	1109		----		----
171		----		----	1121		----		----
175		----		----	1126		----		----
177	D1319	18.0		-0.79	1140	D1319	18.9655		0.07
215		----		----	1182		----		----
221		----		----	1191		----		----
224		----		----	1205		----		----
225		----		----	1237	EN15553	17.51		-1.22
228		----		----	1275	IP156	18.9		0.01
235	D1319	18.5674		-0.28	1279		----		----
237	D1319	18.0		-0.79	1299	D1319	18.7		-0.17
238		----		----	1316		----		----
253	D1319	18.72		-0.15	1318		----		----
254	D1319	18.4		-0.43	1320	D1319	19.3		0.37
256	D1319	18.1		-0.70	1357	D1319	18.5		-0.34
258		----		----	1372		19.36		0.42
273	D1319	17.92		-0.86	1397		----		----
311		----		----	1399		----		----
317		----		----	1412		----		----
323	D1319	19.3		0.37	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D1319	18.5		-0.34
334		----		----	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D1319	18.9		0.01	1585		----		----
391	D1319	19.1		0.19	1586	D1319	19.0		0.10
396		----		----	1587		----		----
399		----		----	1610	IP156	19.6		0.64
440		----		----	1613	D1319	19.07		0.16
445		----		----	1616	D1319	18.25		-0.57
447		----		----	1631	D1319	19.11		0.20
460		----		----	1634		----		----
467	D1319	18.3		-0.52	1636	D1319	18.05		-0.74
480		----		----	1669		----		----
496		----		----	1683	D1319	17.30		-1.41
603		----		----	1688		----		----
608		----		----	1715		----		----
631	D1319	17.78		-0.98	1720		----		----
657	D1319	19.7		0.72	1724	D1319	19.3		0.37
736	D1319	19.21		0.29	1730		----		----
823	D1319	19.9		0.90	1757		----		----
824	D1319	19.577		0.62	1776		----		----
851		----		----	1833		19.46		0.51
854		----		----	1852	D1319	19.008		0.11
862		----		----	1913		----		----
869		----		----	1944	D1319	18.26		-0.56
904		----		----	1961		----		----
914	D1319	18.6		-0.25	2130	IP156	17.7		-1.05
962	D1319	17.9		-0.88	2133	D1319	20.96		1.85
963	D1319	18.3		-0.52	6028		----		----
970	D1319	19.5		0.55	6041		----		----
974		----		----	6054		----		----
994		----		----	6075	D1319	19.43		0.48
995	D1319	20.45		1.39	6114	D1319	18.75		-0.12
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D1319	18.8		-0.08	6203	D1319	19.68		0.71
1016		----		----	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039		----		----	6312		----		----
1040	D1319	19.45		0.50	6315		----		----
1049		----		----	6321	IP156	21.8	R(0.05)	2.59
1059	D1319	19.3		0.37	6324		----		----



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D1319	20.3		1.26	6421		----		----
6344		----		----	6438	D1319	17.9		-0.88
6346		----		----	6479		----		----
6364	D1319	19.612		0.65	6487		----		----
6384	D1319	19.3		0.37					
normality		OK							
n		58							
outliers		1							
mean (n)		18.886							
st.dev. (n)		0.7910							
R(calc.)		2.215							
st.dev.(D1319:20a)		1.1241							
R(D1319:20a)		3.148							





## Determination of Mono Aromatics (MAH) by HPLC on sample #22150; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52		----		----	1062		----		----
62		----		----	1064	D6379	21.42		0.08
120		----		----	1065	IP436	21.6		0.48
140		----		----	1082		----		----
150		----		----	1097		----		----
159		----		----	1105	D6379	20.92		-1.04
169		----		----	1109		----		----
171	D6379	18.0	R(0.01)	-7.54	1121		----		----
175		----		----	1126	EN12916	21.3		-0.19
177		----		----	1140		----		----
215		----		----	1182		----		----
221		----		----	1191	D6379	23.592	R(0.05)	4.92
224		----		----	1205	D8267	21.349		-0.08
225		----		----	1237		----		----
228		----		----	1275	IP436	20.0931		-2.88
235		----		----	1279		----		----
237		----		----	1299		----		----
238		----		----	1316		----		----
253		----		----	1318	D6379	20.44		-2.10
254		----		----	1320	D6379	21.64		0.57
256		----		----	1357	D6379	----		----
258		----		----	1372		----		----
273		----		----	1397		----		----
311		----		----	1399		----		----
317		----		----	1412		----		----
323		----		----	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D6379	23.2	C	4.04
334	D6379	21.9		1.15	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D6379	21.07		-0.70	1585	D6379	21.4		0.03
391		----		----	1586	D6379	22.71		2.95
396		----		----	1587		----		----
399		----		----	1610		----		----
440		----		----	1613		----		----
445		----		----	1616	D6379	21.873		1.09
447		----		----	1631		----		----
460		----		----	1634	D6379	20.5		-1.97
467	D6379	20.86		-1.17	1636		----		----
480		----		----	1669		----		----
496	D6379	21.52		0.30	1683		----		----
603		----		----	1688		----		----
608		----		----	1715		----		----
631		----		----	1720		----		----
657	IP436	21.7		0.70	1724		----		----
736	D6379	20.69121		-1.54	1730		----		----
823	D6379	21.6		0.48	1757		----		----
824	D6379	21.3		-0.19	1776		----		----
851	D6379	20.59		-1.77	1833		----		----
854		----		----	1852	D6379	21.248		-0.30
862		----		----	1913		----		----
869		----		----	1944	D6379	21.04		-0.77
904		----		----	1961		----		----
914		----		----	2130		----		----
962	D6379	22.045		1.47	2133		----		----
963	D6379	21.10		-0.63	6028		----		----
970		----		----	6041		----		----
974	D6379	22.01		1.39	6054		----		----
994	D6379	21.06		-0.72	6075	D6379	21.51	C	0.28
995	D6379	21.993		1.36	6114		----		----
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D6379	21.16		-0.50	6203	D6379	20.97		-0.92
1016		----		----	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039		----		----	6312		----		----
1040		----		----	6315	IP436	21.58		0.44
1049	D6379	21.682		0.66	6321		----		----
1059		----		----	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		----		----	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384		----		----					

normality suspect  
 n 34  
 outliers 2  
 mean (n) 21.385  
 st.dev. (n) 0.6223  
 R(calc.) 1.742  
 st.dev.(D6379:21e1) 0.4488  
 R(D6379:21e1) 1.257

Lab 1455 first reported 18.6 %V/V  
 Lab 6075 first reported 19.07





## Determination of Di Aromatics (DAH) by HPLC on sample #22150; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52		----		----	1062		----		----
62		----		----	1064	D6379	0.96		-1.42
120		----		----	1065	IP436	1.04		-0.03
140		----		----	1082		----		----
150		----		----	1097		----		----
159		----		----	1105	D6379	1.13		1.53
169		----		----	1109		----		----
171	D6379	0.9		-2.46	1121		----		----
175		----		----	1126	EN12916	1.1		1.01
177		----		----	1140		----		----
215		----		----	1182		----		----
221		----		----	1191	D6379	1.080		0.66
224		----		----	1205	D8267	0.914		-2.22
225		----		----	1237		----		----
228		----		----	1275	IP436	0.8760		-2.87
235		----		----	1279		----		----
237		----		----	1299		----		----
238		----		----	1316		----		----
253		----		----	1318	D6379	1.19		2.57
254		----		----	1320	D6379	1.14		1.70
256		----		----	1357	D1319	----		----
258		----		----	1372		----		----
273		----		----	1397		----		----
311		----		----	1399		----		----
317		----		----	1412	D6379	0.7		-5.93
323		----		----	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D6379	1.0	C	-0.73
334	D6379	1.6	R(0.01)	9.68	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D6379	1.05		0.14	1585	D6379	1.01		-0.55
391		----		----	1586	D6379	0.91		-2.29
396		----		----	1587		----		----
399		----		----	1610		----		----
440		----		----	1613		----		----
445		----		----	1616	D6379	0.985		-0.99
447		----		----	1631		----		----
460		----		----	1634	D6379	1.0		-0.73
467	D6379	1.05		0.14	1636		----		----
480		----		----	1669		----		----
496	D6379	1.14		1.70	1683		----		----
603		----		----	1688		----		----
608		----		----	1715		----		----
631		----		----	1720		----		----
657	IP436	1.3		4.48	1724		----		----
736	D6379	0.98060		-1.06	1730		----		----
823	D6379	1.2		2.74	1757		----		----
824	D6379	1.1		1.01	1776		----		----
851	D6379	0.83		-3.67	1833		----		----
854		----		----	1852	D6379	1.142		1.74
862		----		----	1913		----		----
869		----		----	1944	D6379	1.08		0.66
904		----		----	1961		----		----
914		----		----	2130		----		----
962	D6379	1.144		1.77	2133		----		----
963	D6379	1.17		2.22	6028		----		----
970		----		----	6041		----		----
974	D6379	1.12		1.36	6054		----		----
994	D6379	1.095		0.92	6075	D6591	0.88	C	-2.81
995	D6379	1.139		1.68	6114		----		----
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D6379	0.98		-1.07	6203	D6379	0.97		-1.25
1016		----		----	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039		----		----	6312		----		----
1040		----		----	6315	IP436	1.17		2.22
1049	D6379	1.030		-0.20	6321		----		----
1059		----		----	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		----		----	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384		----		----					

normality OK  
 n 36  
 outliers 1  
 mean (n) 1.042  
 st.dev. (n) 0.1217  
 R(calc.) 0.341  
 st.dev.(D6379:21e1) 0.0577  
 R(D6379:21e1) 0.162

Lab 1455 first reported 0.8 %V/V  
 Lab 6075 first reported 0.63





Determination of Total Aromatics by HPLC on sample #22150; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52		----		----	1062		----		----
62		----		----	1064	D6379	22.37		-0.01
120		----		----	1065		----		----
140		----		----	1082		----		----
150		----		----	1097		----		----
159		----		----	1105	D6379	22.10		-0.57
169		----		----	1109		----		----
171	D6379	18.9	R(0.01)	-7.17	1121		----		----
175		----		----	1126	EN12916	22.6		0.46
177		----		----	1140		----		----
215		----		----	1182		----		----
221		----		----	1191		----		----
224		----		----	1205	D8267	22.263		-0.23
225		----		----	1237		----		----
228		----		----	1275	IP436	20.9691		-2.90
235		----		----	1279		----		----
237		----		----	1299		----		----
238		----		----	1316		----		----
253		----		----	1318	D6379	21.64		-1.52
254		----		----	1320	D6379	22.59		0.44
256		----		----	1357		----		----
258		----		----	1372		----		----
273		----		----	1397		----		----
311		----		----	1399		----		----
317		----		----	1412		----		----
323		----		----	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D6379	24.3	C	3.97
334	D6379	23.5		2.32	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D6379	22.12		-0.53	1585	D6379	22.4		0.05
391		----		----	1586	D6379	23.62		2.56
396		----		----	1587		----		----
399		----		----	1610		----		----
440		----		----	1613		----		----
445		----		----	1616	D6379	22.858		0.99
447		----		----	1631		----		----
460		----		----	1634	D6379	21.5		-1.81
467	D6379	21.92		-0.94	1636		----		----
480		----		----	1669		----		----
496	D6379	22.66		0.59	1683		----		----
603		----		----	1688		----		----
608		----		----	1715		----		----
631		----		----	1720		----		----
657	IP436	23.0		1.29	1724		----		----
736	D6379	21.67181		-1.45	1730		----		----
823	D6379	22.8		0.87	1757		----		----
824	D6379	22.4		0.05	1776		----		----
851	D6379	21.42		-1.97	1833		----		----
854		----		----	1852	D6379	22.390		0.03
862		----		----	1913		----		----
869		----		----	1944	D6379	22.12		-0.53
904		----		----	1961		----		----
914		----		----	2130		----		----
962	D6379	23.19		1.68	2133		----		----
963	D6379	22.26		-0.24	6028		----		----
970		----		----	6041		----		----
974	D6379	23.13		1.55	6054		----		----
994	D6379	22.15		-0.47	6075	D6379	22.39	C	0.03
995	D6379	22.21		-0.34	6114		----		----
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D6379	22.15		-0.47	6203	D6379	21.94		-0.90
1016		----		----	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039		----		----	6312		----		----
1040		----		----	6315	IP436	22.74		0.75
1049	D6379	22.712		0.69	6321		----		----
1059		----		----	6324	D6379	20.70	C	-3.46

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		----		----	6421	D6379	13.0	R(0.01)	-19.33
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384		----		----					

normality suspect  
 n 34  
 outliers 2  
 mean (n) 22.376  
 st.dev. (n) 0.7227  
 R(calc.) 2.024  
 st.dev.(D6379:21e1) 0.4850  
 R(D6379:21e1) 1.358

Lab 1455 first reported 19.5 %V/V  
 Lab 6075 first reported 19.70  
 Lab 6324 first reported 20.07





## Determination of Total Aromatics by HPLC on sample #22150; results in %V/V

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52		----		----	1062	D6379	20.82		1.55
62		----		----	1064	D6379	19.36		-1.89
120		----		----	1065		----		----
140		----		----	1082	D6379	19.607		-1.31
150		----		----	1097		----		----
159		----		----	1105	D6379	20.0		-0.38
169		----		----	1109		----		----
171		----		----	1121		----		----
175		----		----	1126		----		----
177		----		----	1140		----		----
215		----		----	1182		----		----
221		----		----	1191	D6379	19.82		-0.81
224		----		----	1205	D8267	20.011		-0.36
225		----		----	1237		----		----
228		----		----	1275	IP436	19.0		-2.74
235		----		----	1279		----		----
237		----		----	1299	IP436	20.3		0.33
238		----		----	1316	D6379	19.5		-1.56
253		----		----	1318	D6379	19.64		-1.23
254		----		----	1320	D6379	20.47		0.73
256		----		----	1357		----		----
258		----		----	1372		----		----
273		----		----	1397	D6379	21.08		2.17
311		----		----	1399		----		----
317		----		----	1412		----		----
323		----		----	1417		----		----
328		----		----	1444		----		----
333	D6379	20.6		1.03	1455		----		----
334	D6379	21.1		2.22	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D6379	19.91		-0.59	1585	D6379	20.3		0.33
391	D6379	19.9		-0.62	1586	D6379	20.04		-0.29
396		----		----	1587		----		----
399		----		----	1610		----		----
440		----		----	1613		----		----
445		----		----	1616	D6379	20.830		1.58
447	IP436	20.7		1.27	1631		----		----
460		----		----	1634	D6379	19.5		-1.56
467	D6379	19.80		-0.85	1636		----		----
480		----		----	1669	D6379	20.9		1.74
496	D6379	20.62		1.08	1683	D6379	19.6		-1.33
603		----		----	1688	D6379	20.1		-0.15
608		----		----	1715		----		----
631		----		----	1720		18.62		-3.64
657	IP436	20.8		1.51	1724		----	W	----
736		----		----	1730		----		----
823		----		----	1757		----		----
824	D6379	20.1		-0.15	1776	D6379	19.51		-1.54
851	D6379	19.36		-1.89	1833		20.46		0.70
854		----		----	1852	D6379	19.291		-2.06
862		----		----	1913		----		----
869		----		----	1944	D6379	19.75		-0.97
904		----		----	1961		----		----
914		----		----	2130		----		----
962	D6379	21.03		2.05	2133		----		----
963	D6379	20.18		0.04	6028		----		----
970		----		----	6041		----		----
974	D6379	20.90		1.74	6054		----		----
994		----		----	6075	D6379	20.65	C	1.15
995	D6379	20.99		1.96	6114		----		----
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D6379	20.1		-0.15	6203	D6379	19.90		-0.62
1016	IP436	19.652		-1.20	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039	D6379	20.62		1.08	6312		----		----
1040		----		----	6315		----		----
1049	D6379	20.756		1.40	6321		----		----
1059		----		----	6324	D6379	17.90	R(0.05)	-5.34



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		----		----	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384	D6379	21.1		2.22					
	normality	OK							
	n	45							
	outliers	1							
	mean (n)	20.162							
	st.dev. (n)	0.6226							
	R(calc.)	1.743							
	st.dev.(D6379:21e1)	0.4235							
	R(D6379:21e1)	1.186							

Lab 1724 test results withdrawn, reported 21.4  
 Lab 6075 first reported 17.16





Determination of Color Saybolt (automated) on sample #22150;

lab	method	cell size	value	mark	z(targ)	lab	method	cell size	value	mark	z(targ)
52	D6045	100 mm	>30		----	1062	D6045	50 mm	30		0.11
62			----		----	1064	D6045	50 mm	30		0.11
120	D6045	50 mm	30		0.11	1065			----		----
140			----		----	1082	D6045		>30		----
150	D6045		>30		----	1097			----		----
159			----		----	1105	D6045	50 mm	30		0.11
169			----		----	1109			----		----
171	D6045		30		0.11	1121	D6045	100 mm	> 30		----
175			----		----	1126			----		----
177			----		----	1140	D6045	50 mm	30		0.11
215			----		----	1182			----		----
221			----		----	1191	D6045	100 mm	>30		----
224			----		----	1205			----		----
225			----		----	1237	D6045	50 mm	>30		----
228			----		----	1275			----		----
235			----		----	1279			----		----
237			----		----	1299	D6045	100 mm	30		0.11
238			----		----	1316			----		----
253			----		----	1318	D6045	100 mm	30		0.11
254			----		----	1320	D6045	50 mm	>30		----
256			----		----	1357	D6045	50 mm	30		0.11
258			----		----	1372		100 mm	30		0.11
273			----		----	1397			----		----
311			----		----	1399			----		----
317			----		----	1412			----		----
323	D6045	50 mm	30		0.11	1417			----		----
328			----		----	1444			----		----
333	D6045	100 mm	30		0.11	1455			----		----
334	D6045		30		0.11	1496			----		----
335			----		----	1538			----		----
365	D6045	50 mm	30		0.11	1575			----		----
372	D6045	50 mm	30		0.11	1585		100 mm	30		0.11
391			----		----	1586	D6045	50 mm	30		0.11
396	D6045	50 mm	30		0.11	1587	D6045	50 mm	30		0.11
399			----		----	1610	D6045	50 mm	30		0.11
440			----		----	1613	D6045		>30		----
445	D6045	50 mm	30		0.11	1616	D6045	100 mm	30		0.11
447	D6045	100 mm	30		0.11	1631			----		----
460			----		----	1634			----		----
467			----		----	1636			----		----
480			----		----	1669	D6045		>30		----
496	D6045		>30		----	1683	D6045	50 mm	>30		----
603			----		----	1688	D6045		30		0.11
608			----		----	1715			----		----
631	D6045	50 mm	>30		----	1720		50 mm	30		0.11
657	D6045	100 mm	30		0.11	1724	D6045		30	C	0.11
736			----		----	1730			----		----
823	D6045	50 mm	30		0.11	1757			----		----
824	D6045	50 mm	30		0.11	1776	D6045		30		0.11
851			----		----	1833			30		0.11
854			----		----	1852	D6045	10 mm	30		0.11
862			----		----	1913			----		----
869			----		----	1944	D6045	50 mm	30		0.11
904	D6045		29		-2.15	1961			----		----
914	D6045	10 mm	30		0.11	2130			----		----
962	D6045	50 mm	>30		----	2133			----		----
963	D6045		30		0.11	6028			----		----
970			----		----	6041			----		----
974	D6045	100 mm	30		0.11	6054			----		----
994			----		----	6075			----		----
995			----		----	6114			----		----
996			----		----	6135			----		----
997			----		----	6142			----		----
1011	D6054	100 mm	>30		----	6203	D6045	50 mm	>30		----
1016	D6045	100 mm	>30		----	6244			----		----
1019			----		----	6266	D6045		>30	C	----
1023	D6045	50 mm	>30		----	6274			----		----
1039			----		----	6312			----		----
1040			----		----	6315	D6045	50 mm	>30		----
1049	D6045	50 mm	30		0.11	6321	D6045	50 mm	30		0.11
1059	D6045	50 mm	30		0.11	6324			----		----

lab	method	cell size	value	mark	z(targ)	lab	method	cell size	value	mark	z(targ)
6332			----		----	6421			----		----
6344			----		----	6438	D6045	10 mm	29		-2.15
6346			----		----	6479			----		----
6364			----		----	6487			----		----
6384	D6045	33 mm	>30		----						
	normality		not OK				<u>only 50 mm cell</u>		<u>only 100 mm cell</u>		
	n		41				unknown		unknown		
	outliers		0				21		9		
	mean (n)		29.95				0		0		
	st.dev. (n)		0.218				30.00		30.00		
	R(calc.)		0.61				0.000		0.000		
	st.dev.(D6045:20)		0.443				0.00		0.00		
	R(D6045:20)		1.24				0.443		0.443		
							1.24		1.24		

Lab 1724 first reported 25+  
 Lab 6266 first reported <0.5





Determination of Color Saybolt (manual) on sample #22150;

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52		----		----	1062		----		----
62	D156	29		-1.27	1064		----		----
120		----		----	1065		----		----
140	D156	30		0.13	1082		----		----
150	D156	30		0.13	1097	NF M 07-003	30		0.13
159		----		----	1105		----		----
169		----		----	1109		----		----
171	D156	30		0.13	1121	D156	> 30		----
175		----		----	1126		----		----
177	D156	30	C	0.13	1140		----		----
215	D156	30		0.13	1182		----		----
221		----		----	1191		----		----
224		----		----	1205		----		----
225		----		----	1237		----		----
228	D156	30		0.13	1275	D156	30		0.13
235	D156	<30		----	1279		----		----
237	D156	29		-1.27	1299		----		----
238	D156	29		-1.27	1316	D156	>30		----
253	D156	30		0.13	1318		----		----
254	D156	30		0.13	1320	D156	>30		----
256	D156	<30		----	1357		----		----
258	D156	30		0.13	1372		----		----
273	D156	30		0.13	1397		----		----
311		----		----	1399		----		----
317	D156	30		0.13	1412		----		----
323		----		----	1417		----		----
328	D156	30		0.13	1444		----		----
333		----		----	1455	D156	30		0.13
334	D156	30		0.13	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D156	30		0.13	1585		30		0.13
391	D156	30		0.13	1586		----		----
396		----		----	1587		----		----
399	D156	28	G(0.01)	-2.67	1610		----		----
440	D156	>30		----	1613		----		----
445		----		----	1616		----		----
447		----		----	1631		----		----
460		----		----	1634	D156	> 30		----
467	D156	30		0.13	1636	D156	>30		----
480		----		----	1669		----		----
496		----		----	1683		----		----
603	D156	30		0.13	1688		----		----
608	D156	30		0.13	1715		----		----
631		----		----	1720		----		----
657	D156	30		0.13	1724	D156	30	C	0.13
736	D156	>30		----	1730		----		----
823	D156	30		0.13	1757		----		----
824	D156	30		0.13	1776		----		----
851	D156	30		0.13	1833		----		----
854		----		----	1852		----		----
862		----		----	1913		----		----
869		----		----	1944	D156	30		0.13
904	D156	29		-1.27	1961		----		----
914		----		----	2130		----		----
962	D156	>30		----	2133		----		----
963		----		----	6028		----		----
970	D156	30		0.13	6041		----		----
974	D156	30		0.13	6054		----		----
994	D156	30		0.13	6075	D156	>30		----
995	D156	30		0.13	6114	D156	30		0.13
996		----		----	6135		----		----
997	D156	30		0.13	6142		----		----
1011		----		----	6203		----		----
1016		----		----	6244		----		----
1019		----		----	6266	D156	30.0		0.13
1023		----		----	6274		----		----
1039	D156	30		0.13	6312		----		----
1040		----		----	6315		----		----
1049		----		----	6321		----		----
1059	D156	30		0.13	6324	D156	30		0.13

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D156	30		0.13	6421		----		----
6344	D156	>30		----	6438		----		----
6346		----		----	6479		----		----
6364	D156	30		0.13	6487		----		----
	normality	not OK							
	n	44							
	outliers	1							
	mean (n)	29.91							
	st.dev. (n)	0.291							
	R(calc.)	0.81							
	st.dev.(D156:15)	0.714							
	R(D156:15)	2							

Lab 177 first reported 26  
 Lab 1724 first reported 25+



Determination of Copper Corrosion 2 hrs at 100 °C on sample #22150;

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D130	1a		----	1062	D130	1A		----
62	D130	1a		----	1064	D130	1a		----
120		----		----	1065		----		----
140		----		----	1082		----		----
150		----		----	1097	ISO2160	1a		----
159	D130	1a		----	1105	D130	1a		----
169	D130	1A		----	1109		----		----
171	D130	1a		----	1121	D130	1a		----
175		----		----	1126		----		----
177	D130	1a		----	1140	IP154	1A		----
215	D130	1A		----	1182		----		----
221	D130	1		----	1191		----		----
224		----		----	1205		----		----
225	D130	1a		----	1237		----		----
228	D130	1A		----	1275	IP154	1A		----
235	D130	1a		----	1279		----		----
237	D130	1A		----	1299	D130	1A		----
238	D130	1a		----	1316	D130	1a		----
253	D130	1A		----	1318	D130	1a		----
254	D130	1a		----	1320	D130	1a		----
256	D130	1A		----	1357	D130	1a		----
258		1a		----	1372	D130	1b		----
273	D130	1a		----	1397		----		----
311	D130	1A		----	1399		----		----
317	D130	1a		----	1412	D130	1a		----
323	D130	1A		----	1417	D130	1B		----
328	D130	1		----	1444		----		----
333	D130	1a		----	1455	D130	1A		----
334	D130	1b		----	1496		----		----
335	D130	1		----	1538		----		----
365	IP154	1a		----	1575	D130	1a		----
372	D130	1a		----	1585	D130	1a		----
391	D130	1A		----	1586	D130	1A		----
396	D130	1A		----	1587	D130	1a		----
399	D130	1A		----	1610	D130	1a		----
440		----		----	1613	D130	1a		----
445	IP154	1a		----	1616	D130	1a		----
447	IP154	1a		----	1631		----		----
460		----		----	1634	D130	1a		----
467	D130	1a		----	1636	D130	1a		----
480		----		----	1669	D130	1a		----
496	ISO2160	1a		----	1683	D130	1a		----
603	D130	1A		----	1688	D130	1A		----
608	D130	1a		----	1715		----		----
631	D130	1a		----	1720	D130	1a		----
657	D130	1a		----	1724	D130	1a		----
736	D130	1a		----	1730		----		----
823	D130	1b		----	1757	D130	1a		----
824	D130	1a		----	1776		----		----
851	D130	1b		----	1833		1a		----
854		----		----	1852	D130	1A		----
862		----		----	1913		----		----
869		----		----	1944	D130	1a		----
904	ISO2160	1a		----	1961	D130	1a		----
914	D130	1a		----	2130	IP154	1a		----
962	D130	1a		----	2133	D130	1a		----
963	D130	1a		----	6028		----		----
970	D130	1a		----	6041		----		----
974	D130	1a		----	6054		----		----
994	D130	1a		----	6075	D130	1a		----
995	D130	1a		----	6114	D130	1a		----
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D130	1a		----	6203	D130	1A		----
1016		----		----	6244		----		----
1019		----		----	6266		----		----
1023	IP154	1a		----	6274		----		----
1039	ISO2160	1A		----	6312		----		----
1040		----		----	6315	D130	1		----
1049	D130	1A		----	6321	IP154	1A		----
1059	D130	1a		----	6324	D130	1a		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D130	1a		----	6421	D130	1		----
6344	ISO2160	1A		----	6438	D130	1a		----
6346		----		----	6479		----		----
6364	D130	1A		----	6487		----		----
6384	D130	1a		----					
	n	104							
	mean (n)	1 (1a/1b)							

Determination of Density at 15 °C on sample #22150; results in kg/m<sup>3</sup>



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D4052	803.4		0.01	1062	D4052	803.4		0.01
62	D4052	803.4		0.01	1064	D4052	803.3		-0.55
120	D4052	803.4		0.01	1065	D4052	803.6		1.13
140	D4052	803.39		-0.05	1082	ISO12185	803.7		1.69
150	D4052	803.4		0.01	1097	ISO12185	803.38		-0.10
159	D4052	803.37		-0.16	1105	D4052	803.40		0.01
169	D4052	803.2	C	-1.11	1109		----		----
171	D4052	803.5		0.57	1121	D4052	803.4		0.01
175	D4052	803.4		0.01	1126	D4052	803.40		0.01
177	D4052	803.7	C	1.69	1140	IP365	803.3		-0.55
215	D1298	803.3		-0.55	1182	ISO12185	803.625		1.27
221	D4052	803.5		0.57	1191	ISO12185	803.44		0.23
224		----		----	1205	ISO12185	803.41		0.07
225	D4052	803.4		0.01	1237	ISO12185	803.2		-1.11
228	D4052	803.4		0.01	1275	IP365	803.4		0.01
235	D4052	803.4		0.01	1279		----		----
237	D4052	803.4		0.01	1299	D4052	803.3		-0.55
238	D4052	803.4		0.01	1316	D4052	803.3		-0.55
253	D4052	803.3		-0.55	1318	D4052	803.38		-0.10
254	D4052	803.3		-0.55	1320	D4052	803.4		0.01
256	D4052	803.4		0.01	1357	D4052	803.5		0.57
258		803.4		0.01	1372		803.44		0.23
273	D4052	803.2		-1.11	1397		----		----
311	D4052	803.4		0.01	1399		----		----
317	D4052	803.4		0.01	1412	D4052	803.4		0.01
323	D4052	803.4		0.01	1417	IP365	803.4		0.01
328	D4052	803.6		1.13	1444		----		----
333	D4052	803.5		0.57	1455	D4052	803.4		0.01
334	D4052	803.4		0.01	1496		----		----
335	D4052	803.4		0.01	1538		----		----
365	IP365	803.3		-0.55	1575		----		----
372	D4052	803.4		0.01	1585	ISO12185	803.41		0.07
391	D4052	803.3		-0.55	1586	D4052	803.5		0.57
396	D4052	803.5		0.57	1587	D4052	803.42		0.12
399	D4052	803.4		0.01	1610	IP365	803.4		0.01
440		803.4		0.01	1613	D4052	803.41		0.07
445	D4052	803.5		0.57	1616	D4052	803.3		-0.55
447	D4052	803.4		0.01	1631	D4052	803.4		0.01
460		----		----	1634	D4052	803.4		0.01
467	D4052	803.36		-0.21	1636	D4052	803.4		0.01
480	D4052	803.4		0.01	1669	D4052	803.3		-0.55
496	D4052	803.39		-0.05	1683	D4052	803.4		0.01
603	D4052	803.4		0.01	1688	D4052	804.0	R(0.01)	3.37
608	D4052	803.4		0.01	1715	ISO12185	803.6		1.13
631	D4052	803.53		0.74	1720	D4052	803.7		1.69
657	D4052	803.3		-0.55	1724	D1298	803.2		-1.11
736	D4052	803.4		0.01	1730	D4052	803.33		-0.38
823	D4052	803.4		0.01	1757	D7777	803.3		-0.55
824	D4052	803.4		0.01	1776	ISO12185	803.5		0.57
851	D4052	803.3		-0.55	1833		803.4		0.01
854		----		----	1852	D4052	803.41		0.07
862		----		----	1913		----		----
869		----		----	1944	D4052	803.40		0.01
904	D4052	803.5		0.57	1961		----		----
914	D4052	803.3		-0.55	2130	D4052	803.3		-0.55
962	D4052	803.4		0.01	2133	D4052	803.43		0.18
963	D4052	803.4		0.01	6028		----		----
970	D4052	803.3		-0.55	6041		----		----
974	D4052	803.3		-0.55	6054		----		----
994	D4052	803.4		0.01	6075	D4052	803.2		-1.11
995	D4052	803.5		0.57	6114	D4052	803.3		-0.55
996		----		----	6135		----		----
997	D4052	803.4		0.01	6142	ISO12185	803.4		0.01
1011	D4052	803.4		0.01	6203	D4052	803.4		0.01
1016		----		----	6244		----		----
1019	D4052	803.35		-0.27	6266	D4052	803.528		0.73
1023	D4052	803.5	C	0.57	6274		----		----
1039	ISO12185	803.4		0.01	6312		----		----
1040	D4052	803.44		0.23	6315	D4052	803.4		0.01
1049	D4052	803.34		-0.33	6321	IP365	803.4		0.01
1059	D4052	803.4		0.01	6324	D4052	803.4		0.01



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D4052	803.3		-0.55	6421	D4052	803.4		0.01
6344	ISO12185	803.4		0.01	6438	D4052	803.4		0.01
6346		-----		-----	6479		-----		-----
6364	D4052	803.3		-0.55	6487		-----		-----
6384	D4052	803.4		0.01					

normality not OK  
 n 123  
 outliers 1  
 mean (n) 803.398  
 st.dev. (n) 0.0912  
 R(calc.) 0.255  
 st.dev.(D4052:22) 0.1786  
 R(D4052:22) 0.50

Lab 169 first reported 804.0  
 Lab 177 first reported 804.1  
 Lab 1023 first reported 0.8035 kg/m<sup>3</sup>













Determination of Distillation at 760 mmHg on sample #22150; results in °C

lab	method	IBP	mark	10% rec	mark	50% rec	mark	90% rec	mark	FBP	mark	Res.	Loss
52	D86-automated	145.5		170.1		195.3		232.4		253.1		1.2	0.3
62	D86-automated	151.7		171.3		196.0		233.3		253.8	C	1.0	0.3
120	D86-automated	151.8		169.9		194.9		233.6		251.3		1.2	2.8
140	D86-automated	151.3		170.5		195.3		231.9		253.5		1.2	0.1
150	D86-automated	144.7		169.0		195.9		235.4		251.0		1.2	2.2
159	D86-automated	151.2		171.3		196.7		233.6		254.9		1.2	0.8
169	D86-automated	149.5		171.0		195.8		232.7		253.2		1.2	0.4
171	D86-automated	150.0		170.5		195.8		232.4		253.8		0.8	0.6
175	D86-automated	153.4		170.6		198.0		236.7	R5	257.2		1.0	1.2
177	D86-automated	145.5		169.8		195.3		231.9		251.5	C	1.3	0.3
215	D86-manual	149.0		170.0		197.0		233.0		251.0		1.2	0.8
221	D86-automated	155.5		172.3		196.7		233.6		254.4		1.2	0.5
224		----		----		----		----		----		----	----
225	D86-manual	151.0		170.0		196.0		234.0		254.0		1.4	0.6
228	D86-manual	152.0		172.0		196.0		232.0		255.0		0.7	0.3
235	D86-automated	147.3		168.6		194.1		231.7		249.5		1.2	0.8
237	D86-manual	154.0		173.0		196.0		233.0		252.0		1.0	0.5
238		----		----		----		----		----		----	----
253	D86-manual	148.0		168.0		196.0		232.0		251.0		1.0	0.7
254	D86-manual	151.0		170.0		195.0		234.0		254.0		----	----
256	D86-manual	151.0		170.0		196.0		233.0		253.0		----	----
258	D86-automated	155.4		171.1		195.4		230.7		252.9		1.2	0.6
273	D86-automated	151.2		171.0		196.9		233.5		256.4		1.0	0.6
311	D86-automated	152.1		170.4		196.1		233.8		253.4		1.2	0.6
317	D86-automated	153.2		170.2		196.0		233.7		255.4		0.3	1.3
323	D86-automated	150.5		171.2		195.7		233.2		252.2		1.1	0.4
328	D86-automated	150.5		170.0		195.4		232.1		252.1		1.3	0.1
333	D86-automated	146.6		169.3		194.4		231.2		251.0		1.0	0.4
334	D86-automated	149.0		170.4		195.9		233.3		252.2		1.1	1.0
335	D86-automated	151.5		170.3		196.5		234.3		251.9		1.1	1.2
365	IP123-automated	148.8		169.3		195.3		233.6		252.8		1.4	0.9
372	D86	150.9		170.4		195.7		232.8		252.9		1.3	0.3
391		----		----		----		----		----		----	----
396		----		----		----		----		----		----	----
399		----		----		----		----		----		----	----
440	D86-automated	150.8		170.7		196.1		234.5		254.2		1.3	0.5
445	D86-automated	148.1	C	168.6		194.6		232.7		250.4		1.2	0.9
447		151.6		170.2		195.4		233.8		252.7		1.2	0.9
460		----		----		----		----		----		----	----
467	D86-automated	153.1		170.6		195.9		232.8		252.4		1.3	0.5
480	D86-automated	150.8		171.45		196.0		232.6		251.85		1.15	0.45
496	D86-automated	153.7		171.8		195.4		233.4		253.6		1.2	1.0
603	D86-automated	152.4		169.2		193.9		232.0		250.0		1.2	1.3
608	D86-automated	145.9		170.3		196.5		233.3		252.3		1.2	0.6
631	D86-manual	151.0		168.0		194.0	C	231.0		249.5		1.0	1.2
657	D86-automated	154.4		170.8		196.6		233.2		254.0		1.0	0.6
736	D86-manual	150.0		169.5		195.0		233.0		250.0		1.2	0.8
823	D86-automated	151.5		171.2		196.7		233.8		255.1		0.8	0.7
824	D86-automated	150.0		171.8		196.1		233.7		253.8		1.2	0.5
851	D86-automated	148.6		170.3		195.2		232.4		251.8		1.2	0.6
854		----		----		----		----		----		----	----
862		----		----		----		----		----		----	----
869		----		----		----		----		----		----	----
904	D86-automated	150.1		169.4		196.4	C	233.4		253.5		----	----
914	D86-automated	150.3		169.9		195.5		233.3		257.2		1.2	0.8
962	D86-automated	150.8		169.4		194.3		231.0		251.0		1.2	0.7
963	D86-automated	147.7		170.3		195.3		231.5		251.7		1.0	1.0
970	D86-manual	150.0		169.0		195.0		231.0		252.0		1.0	1.0
974	D86-automated	149.0		169.1		195.0		232.1		252.0		1.0	1.0
994	D86-manual	150.0		170.0		195.0		232.0		255.0		1.0	0.2
995	D86-manual	149.5		170.5		195.5		235.0		255.5		1.3	0.3
996		----		----		----		----		----		----	----
997	D86-manual	149.0		171.0		196.0		235.0		255.5		1.3	0.2
1011	D86-automated	149.9		170.6		195.5		234.2		253.4		----	----
1016		----		----		----		----		----		----	----
1019		----		----		----		----		----		----	----
1023		151.7		170.5		194.7		232.6		252.7		1.1	0.8
1039	D86-automated	152.2		171.2		196.3		233.1		251.8		1.2	0.7
1040	D86-automated	148.9		169.0		195.3		233.3		253.2		1.2	0.7
1049	D86-automated	152.6		171.3		196.4		234.3		253.5		1.2	0.8
1059	D86-automated	150.9		170.7		195.8		232.6		253.6		1.2	0.3
1062	D86-automated	151.5		170.1		196.1		233.1		253.7		1.2	0.5

lab	method	IBP	mark	10% rec	mark	50% rec	mark	90% rec	mark	FBP	mark	Res.	Loss
1064	D86-automated	152.8		170.8		195.7		233.4		254.8		1.3	0.4
1065		146.6		166.6	R5	195.7		233.8		275.4	R1	1.4	0.6
1082	ISO3405-automated	150.4		171.4		196.4		233.8		254.2		1.3	C 0.3 C
1097	ISO3405-automated	152.6		169.6		195.2		233.4		252.6		1.2	0.5
1105	D86-automated	148.6		171.7		196.2		232.9		252.3		1.1	0.1
1109		----		----		----		----		----		----	----
1121	D86-automated	145.7		170.4		195.1		232.1		252.0		1.2	0.5
1126	ISO3405-automated	152.4		170.9		195.5		231.4		251.7		1.1	0.5
1140	D86-automated	146.3		169.6		195.2		232.4		253.5		0.4	0.5
1182		155.4	C	172.1		197.2		234.9		254.3	C	0.8	1.1
1191	ISO3405-automated	146.4		171.5		195.8		232.9		252.7		1.3	0
1205	D86-automated	152.0		170.8		196.3		232.5		252.4		1.2	0.2
1237	ISO3405-automated	149.8		170.4		195.4		232.0		252.3		1.2	0.5
1275	IP123-automated	149.4		169.8		195.0		232.4		251.8		1.2	0.5
1279		----		----		----		----		----		----	----
1299	D86-automated	151.2		170.9		196.4		234.3		254.5		1.2	0.7
1316	D86-automated	150.4		170.8		196.2		233.1		253.6		1.2	0.2
1318	D86-automated	154.9		171.1		195.8		233.4		253.8		1.2	0.5
1320	D86-automated	147.2	C	169.5		195.2		231.5		252.0		1.2	0.1
1357	D86-automated	151.3		169.8		194.1		232.9		253.8		1.2	1.5
1372	D86-automated	151.9		168.8		194.1		231.2		253.9		1.1	0.9
1397		152.6		171.3		195.7		231.9		251.4		1.2	0.4
1399		----		----		----		----		----		----	----
1412	D86-manual	150.0		170.0		195.5		233.0		253.0		----	----
1417	IP123-automated	152.3		170.8		197.3		235.7		253.0		1.5	1.5
1444		----		----		----		----		----		----	----
1455	D86-automated	151.0		171.2		196.1		233.0		253.2		1.2	0.4
1496		----		----		----		----		----		----	----
1538		----		----		----		----		----		----	----
1575		----		----		----		----		----		----	----
1585	D86-automated	148.1		170.0		195.1		231.8		251.5		1.2	0.5
1586	D86-automated	151.3		170.4		195.9		233.2		252.9		1.3	0.6
1587	D86-automated	150.4		170.4		195.6		232.3		252.3		1.2	0.2
1610	IP123-automated	152.4		171.5		196.3		233.6		254.0		1.2	0.1
1613	D86-automated	152.8		171.2		196.2		232.7		253.9		1.2	0.3
1616	D86-automated	148.8		169.4		194.2		231.3		252.1		1.2	0.4
1631	D86-automated	147.1		169.4		195.2		232.7		253.4		1.2	0.4
1634		150.9		170.5		195.6		232.8		253.3		1.2	0.1
1636	D86-automated	149.7		171.1		196.5		234.2		251.8		1.2	0.7
1669	D86-automated	146.2		168.7		194.9		232.1		253.1		1.2	0.2
1683	D86-automated	151.2		169.2		194.4		232.0		250.1		1.2	1.1
1688		----		169.0		196.0		231.0		252.0		----	----
1715	ISO3405-automated	154.3		169.8		195		234		253.7		1	0.6
1720	D86-automated	149.1		170.6		195.9		232.9		258.9	C,R1	0.9	0.2
1724	D86-automated	148.9		169.4		194.5		231		252.5		1.2	0.6
1730		----		----		----		----		----		----	----
1757	D86-automated	151.3		170.4		194.8		231.9		250.4		1.2	1
1776	ISO3405-automated	148.9		169.0		195.1		232.8		251.8		1.1	1.2
1833		146.9		169.2		194.6		232		252.2		1.2	0.4
1852	D86-automated	150.9		170.5		196.1		233.0		253.7		1.2	0.7
1913		----		----		----		----		----		----	----
1944	D86-automated	149.6		170.5		195.9		233.0		253.5		1.2	0.5
1961		----		----		----		----		----		----	----
2130		151.5		170.6		196.3		234.0		253.5		1.2	0.9
2133	D86-automated	151.3		170.9		196.4		233.6		254.1		1.2	0.9
6028		----		----		----		----		----		----	----
6041		----		----		----		----		----		----	----
6054		----		----		----		----		----		----	----
6075	D86-automated	147.6		170.9		196.7		234.0		252.9		1.2	0.8
6114	D86-automated	153.4		170.8		195.6		233.2		252.9		1.3	0.6
6135		----		----		----		----		----		----	----
6142	ISO3405-automated	148.7		169.3		194.6		231.6		250.6		1.2	0.8
6203	D86-automated	148.6		170.0		195.8		233.2		252.6		1.2	0.9
6244	D86-manual	154.2		171.4		196.2		232.4	C	262.2	C,R1	1.4	0.6
6266	D86-automated	151.98		171.46		195.6		232.2		252.35	C	1.4	0.199
6274		----		----		----		----		----		----	----
6312		----		----		----		----		----		----	----
6315	D86-automated	149.4		170.4		195.7		233.1		253.4		1.0	0.8
6321	IP123-automated	151.6		170.6		195.8		233.0		252.8		1.4	0.2
6324	D86-manual	152.5		169.5		196.0	C	233.0		254.1		1.0	0.5
6332	D86-manual	154.5		171.5		196.5		234		253.5		1.2	0.8
6344	D86-automated	151.4		170.6		196.0		232.8		253.3		1.2	0.6
6346		----		----		----		----		----		----	----
6364	D86-automated	152.4		169.2		194.2		231.6		252.3		1.4	1.2

lab	method	IBP	mark	10% rec	mark	50% rec	mark	90% rec	mark	FBP	mark	Res.	Loss
6384	D86-automated	149.2		170.8		196.1		234.6		254.4		1.2	0.9
6421	D86	154		172		197		233		255		----	----
6438	D86-automated	146.7		168.7		194.3		232.5		250.9		1.2	0.9
6479		----		----		----		----		----		----	----
6487		----		----		----		----		----		----	----
	normality	OK		OK		OK		OK		OK			
	n	119		119		120		119		117			
	outliers	0		1		0		1		3			
	mean (n)	150.49		170.34		195.65		232.90		252.91			
	st.dev. (n)	2.347		0.933		0.770		1.005		1.458			
	R(calc.)	6.57		2.61		2.16		2.81		4.08			
	st.dev.(D86-A:20b)	2.956		1.338		1.071		1.248		2.536			
	R(D86-A:20b)	8.28		3.75		3.0		3.49		7.1			
	compare												
	R(D86-M:20b)	4.65		3.00		2.84		3.32		3.94			



Lab 62 first reported 257.8  
 Lab 177 first reported 351.5  
 Lab 445 first reported 141.8  
 Lab 631 first reported 192.5  
 Lab 904 first reported 198.4  
 Lab 1082 first reported 98.4 and 1.3 respectively  
 Lab 1182 first reported 157.1 and 257.6 respectively  
 Lab 1320 first reported 142.7  
 Lab 1720 first reported 259.8  
 Lab 6244 first reported 236.5 and 265.4 respectively  
 Lab 6266 first reported 182.566  
 Lab 6324 first reported 193.0



## Determination of Existent Gum (unwashed) on sample #22150; results in mg/100 mL

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D381	<1		----	1062	D381	0		-0.53
62	D381	<1		----	1064	D381	0		-0.53
120		----		----	1065		----		----
140		----		----	1082	IP540	0.18		-0.37
150	D381	<1		----	1097	IP540	0.3		-0.26
159	D381	1.0		0.37	1105	D381	0.40		-0.17
169	D381	<1		----	1109		----		----
171		----		----	1121		----		----
175		----		----	1126		----		----
177	D381	1		0.37	1140	IP540	<1		----
215	D381	0.6		0.01	1182		----		----
221	D381	<1		----	1191	IP541	0.22		-0.33
224		----		----	1205		----		----
225	D381	0.8		0.19	1237		----		----
228	D381	1.0		0.37	1275	IP540	0.1		-0.44
235	D381	0.4		-0.17	1279		----		----
237	D381	1.0		0.37	1299	IP540	0.2		-0.35
238		----		----	1316	D381	<1		----
253	IP540	<1.0		----	1318	IP540	0.2		-0.35
254	D381	0.6		0.01	1320	D381	1.19		0.54
256	IP540	0.6		0.01	1357	D381	<0.5		----
258	D381	0.2		-0.35	1372	IP540	0.6		0.01
273	D381	0.6		0.01	1397		----		----
311	IP540	<1		----	1399		----		----
317	D381	<1		----	1412	D381	<1		----
323	D381	1.0		0.37	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D381	< 1		----
334	D381	<0.5		----	1496		----		----
335		----		----	1538	IP540	<1		----
365	IP540	1.0		0.37	1575	D381	0.6		0.01
372	IP540	0.6		0.01	1585	D381	0.45		-0.12
391		----		----	1586	IP540	0.5		-0.08
396	IP540	<1.0		----	1587	IP540	0.2		-0.35
399	D381	0.6		0.01	1610	IP540	<1		----
440		----		----	1613	D381	0.6		0.01
445		----		----	1616	IP540	0.8		0.19
447	D381	1		0.37	1631		----		----
460		----		----	1634	D381	< 1		----
467	IP540	<1		----	1636	IP540	0.8		0.19
480		----		----	1669		----		----
496		----		----	1683	D381	<1		----
603	D381	0.4		-0.17	1688		----		----
608	D381	<1		----	1715		----		----
631	D381	0.7		0.10	1720	D381	1.0		0.37
657	D381	<1		----	1724	IP540	<1		----
736	D381	0.5		-0.08	1730		----		----
823	D381	1.0		0.37	1757	D381	0.8		0.19
824	D381	0.8		0.19	1776	IP540	1		0.37
851	IP540	<1		----	1833	IP540	<1		----
854		----		----	1852	D381	0.0001		-0.53
862		----		----	1913		----		----
869		----		----	1944		----		----
904		----		----	1961		----		----
914	D381	0.6		0.01	2130	IP540	0.4		-0.17
962	IP540	0.7		0.10	2133	D381	0.0		-0.53
963	D381	0.6		0.01	6028		----		----
970	D381	0.5		-0.08	6041		----		----
974	IP540	0.7		0.10	6054		----		----
994	D381	0.8		0.19	6075	IP540	0.0		-0.53
995		----		----	6114	IP540	1		0.37
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D381	1		0.37	6203	D381	0.8		0.19
1016	D381	<1		----	6244		----		----
1019		----		----	6266		----		----
1023	IP540	0.8		0.19	6274		----		----
1039	ISO6246	<1		----	6312		----		----
1040		----		----	6315	IP540	<1		----
1049	D381	0		-0.53	6321	IP540	<1		----
1059	D381	<1		----	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	IP540	0.7		0.10	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384	D381	<1		----					
	normality	OK							
	n	58							
	outliers	0							
	mean (n)	0.59							
	st.dev. (n)	0.332							
	R(calc.)	0.93							
	st.dev.(D381:22)	1.109							
	R(D381:22)	3.11							
	compare								
	R(IP540:08R19)	1.82							



## Determination of Flash Point on sample #22150; results in °C



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D56	44.5		0.96	1062	IP170	43.0		-0.35
62	D56	42.5	C	-0.79	1064	IP170	44.0		0.52
120	D56	44.4		0.87	1065		----		----
140		----		----	1082	ISO13736	43.5		0.09
150	D56	44.0		0.52	1097	ISO13736	42.1		-1.14
159	D56	45.0		1.40	1105	IP170	44.0		0.52
169	D56	43.0		-0.35	1109		----		----
171	D56	46.5		2.71	1121	IP170	41.9		-1.31
175	D93	46		2.27	1126	D93	47.0		3.15
177	D56	42.0		-1.23	1140	IP170	43.5		0.09
215	IP170	43.0		-0.35	1182	D93	43.1	C	-0.26
221		----		----	1191	ISO13736	44.75		1.18
224		----		----	1205	D56	43.2		-0.18
225	IP170	42.0		-1.23	1237	D56	43.0		-0.35
228	IP170	43.0		-0.35	1275	IP170	42.0		-1.23
235	IP170	43.0		-0.35	1279		----		----
237	IP170	43.0		-0.35	1299	IP170	43.5		0.09
238	IP170	44.0		0.52	1316	IP170	43.5		0.09
253	IP170	42.5		-0.79	1318	IP170	43.0		-0.35
254	IP170	44.0		0.52	1320	D56	43.0		-0.35
256	IP170	43.0		-0.35	1357	IP170	43.5		0.09
258	IP170	42.5		-0.79	1372		44		0.52
273	IP170	43.5		0.09	1397		----		----
311	IP170	43.5		0.09	1399		----		----
317	IP170	43.5		0.09	1412	D93	43.0		-0.35
323	IP170	44.5		0.96	1417	IP170	43.5		0.09
328	IP170	42.0		-1.23	1444	D93	43.0		-0.35
333	IP170	43.5		0.09	1455	IP170	43.0		-0.35
334	IP170	42.5		-0.79	1496		----		----
335	IP170	42.5		-0.79	1538		----		----
365	IP170	42.425		-0.85	1575	D56	42.3		-0.96
372	IP170	43.0		-0.35	1585	ISO13736	45.0		1.40
391	IP170	42		-1.23	1586	IP170	43.0		-0.35
396	IP170	44.5		0.96	1587	IP170	43.0		-0.35
399	IP170	43		-0.35	1610	IP170	44.0		0.52
440	D93	44.0		0.52	1613	D56	43.5		0.09
445	IP170	42.0		-1.23	1616	IP170	43.5		0.09
447	IP170	43.0		-0.35	1631		----		----
460		----		----	1634	IP170	43.2		-0.18
467	IP170	43.0		-0.35	1636	IP170	43.5		0.09
480	D56	44.0		0.52	1669	IP170	44.5		0.96
496	D3828	43.2		-0.18	1683	IP170	43.0		-0.35
603		----		----	1688	D56	44.0		0.52
608	IP170	45.0		1.40	1715	D56	43		-0.35
631	D56	41.0		-2.10	1720	D3828	42.1		-1.14
657	IP170	42.0		-1.23	1724	IP170	44		0.52
736	D56	43.0		-0.35	1730	D56	42.5		-0.79
823	IP170	43.0		-0.35	1757	D56	42.0		-1.23
824	IP170	43.0		-0.35	1776	IP170	43.0		-0.35
851	IP170	42.5		-0.79	1833		44.5		0.96
854		----		----	1852	IP170	44.5		0.96
862		----		----	1913		----		----
869		----		----	1944	ISO13736	43.0		-0.35
904	D56	43.0		-0.35	1961		----		----
914	IP170	42.0		-1.23	2130	IP170	44.0		0.52
962	IP170	42.0		-1.23	2133	D93	45.0		1.40
963	IP170	42.5		-0.79	6028		----		----
970	IP170	42.5		-0.79	6041		----		----
974	IP170	42.5		-0.79	6054		----		----
994	D56	43.5		0.09	6075	IP170	41.2		-1.93
995	IP170	43.5		0.09	6114	IP170	43.5		0.09
996		----		----	6135		----		----
997	IP170	44.0		0.52	6142	ISO13736	44		0.52
1011	IP170	42.5		-0.79	6203	D56	45.5		1.84
1016	IP170	42.5		-0.79	6244	D56	44.5		0.96
1019	ISO3679	41.3		-1.84	6266	IP170	45.2		1.57
1023	IP170	45.0		1.40	6274		----		----
1039	IP170	44.5		0.96	6312	IP170	41.9		-1.31
1040		----		----	6315	IP170	43.0		-0.35
1049	ISO13736	43.0		-0.35	6321	IP170	43.0		-0.35
1059	IP170	43.0		-0.35	6324	D93	44.5		0.96



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	IP170	43.5		0.09	6421	IP170	46		2.27
6344	ISO2719	45		1.40	6438		----		----
6346		----		----	6479		----		----
6364	IP170	43.5	C	0.09	6487	ISO13736	44.6		1.05
6384	IP170	47.0		3.15					

normality suspect  
 n 123  
 outliers 0  
 mean (n) 43.40  
 st.dev. (n) 1.116  
 R(calc.) 3.13  
 st.dev.(IP170:21) 1.143  
 R(IP170:21) 3.2

Lab 62 first reported 48.0  
 Lab 1182 first reported 46  
 Lab 6364 first reported 39.5





Determination of Freezing Point on sample #22150; results in °C

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D5972	-56.6		-1.27	1062	D7153	-55.4		0.07
62	D5972	-56.3		-0.93	1064	D7153	-55.1		0.41
120	D5972	-58.0	R(0.05)	-2.84	1065		----		----
140		----		----	1082	IP529	-55.1		0.41
150		----		----	1097	IP529	-55.1		0.41
159	D2386	-54.8		0.75	1105	D7153	-55.5		-0.04
169	D2386	-57.0	C	-1.72	1109		----		----
171	D2386	-55.0		0.52	1121	D2386	-54.99		0.53
175		----		----	1126		----		----
177	D2386	-57		-1.72	1140	D7153	-54.9		0.63
215	D2386	-55.0		0.52	1182	D5972	-55.1		0.41
221	D2386	-56		-0.60	1191	IP529	-55.3		0.19
224		----		----	1205		----		----
225		----		----	1237	D2386	-55.5		-0.04
228	D2386	-55.5		-0.04	1275	IP529	-55.1		0.41
235	D2386	-56		-0.60	1279		----		----
237	D2386	-55.0		0.52	1299	D2386	-54.0		1.64
238		----		----	1316	D7153	-55.5		-0.04
253	D7153	-55.1		0.41	1318	D7153	-55.1		0.41
254	D2386	-56.0		-0.60	1320	D5972	-56.2		-0.82
256	D2386	-56.5		-1.16	1357	D5972	-56.5		-1.16
258	D2386	-54.2		1.42	1372	D7153	-55.7		-0.26
273	D2386	-57		-1.72	1397	D5972	-55.0		0.52
311	D2386	-55.5		-0.04	1399		----		----
317	D2386	-55.0		0.52	1412	D2386	-54		1.64
323	D2386	-55.0		0.52	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D7153	-56.0		-0.60
334	D7153	-55.1		0.41	1496		----		----
335	IP529	-55.3		0.19	1538	D5972	-55.0		0.52
365	IP16	-56.6		-1.27	1575		----		----
372	D2386	-55.5		-0.04	1585	D2386	-55.0		0.52
391	D2386	-56.5		-1.16	1586	D7153	-54.5		1.08
396	D2386	-55.0		0.52	1587	IP529	-55.3		0.19
399	D7153	-55.7		-0.26	1610		----		----
440		----		----	1613	D7153	-55.6		-0.15
445	D2386	-55.5		-0.04	1616	D7153	-55.6		-0.15
447	D2386	-55.0		0.52	1631	D7153	-55.6		-0.15
460		----		----	1634	D2386	-55.5		-0.04
467	D2386	-55.0		0.52	1636	D2386	-55.5		-0.04
480		----		----	1669	D2386	-56.6		-1.27
496		----		----	1683	D2386	-56.5		-1.16
603	D2386	-55.5		-0.04	1688	D2386	-55.0		0.52
608		----		----	1715	D5972	-54.8		0.75
631	D5972	-56.0		-0.60	1720	D7153	-55.5		-0.04
657	D7153	-55.8		-0.37	1724	D2386	-55.6		-0.15
736	D2386	-55.0		0.52	1730	D2386	-55.5		-0.04
823	D2386	-55.5		-0.04	1757		----		----
824	D2386	-56.0		-0.60	1776	IP529	-54.8		0.75
851	D2386	-54.5		1.08	1833	D7153	-55.3		0.19
854		----		----	1852	D7153	-54.9		0.63
862		----		----	1913		----		----
869		----		----	1944		----		----
904		----		----	1961		----		----
914	D7153	-55.4		0.07	2130	IP529	-54.9		0.63
962	D2386	-56		-0.60	2133	D7153	-55.9		-0.49
963	D2386	-55.5		-0.04	6028		----		----
970	D2386	-55.0		0.52	6041		----		----
974	D2386	-55.0		0.52	6054		----		----
994	D2386	-55		0.52	6075	IP529	-54.6		0.97
995	D2386	-55.0		0.52	6114	D2386	-55.5		-0.04
996		----		----	6135		----		----
997	D2386	-55.0		0.52	6142		----		----
1011	D2386	-58.0	R(0.05)	-2.84	6203	D2386	-55.0		0.52
1016	D5972	-56.1		-0.71	6244	D2386	-56.4		-1.05
1019		----		----	6266		----		----
1023	D7153	-55.1		0.41	6274		----		----
1039	IP529	-55.1		0.41	6312		----		----
1040	D7153	-54.9		0.63	6315	D7153	-54.9		0.63
1049	D7153	-54.8		0.75	6321	D5972	-56.2		-0.82
1059	D2386	-55.0		0.52	6324	D2386	-56.0		-0.60

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D2386	-57		-1.72	6421	D2386	-57	C	-1.72
6344	D2386	-55		0.52	6438		----		----
6346		----		----	6479		----		----
6364	D2386	-55.5		-0.04	6487		----		----
6384	D2386	-57.0		-1.72					
	normality	OK							
	n	103							
	outliers	2							
	mean (n)	-55.47							
	st.dev. (n)	0.673							
	R(calc.)	1.88							
	st.dev.(D2386:19)	0.893							
	R(D2386:19)	2.5							

Lab 169 first reported -46.0  
 Lab 6421 reported 57





Determination of Kinematic Viscosity at -20 °C on sample #22150; results in mm<sup>2</sup>/s

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D445	3.723		----	1062	D445	3.788		----
62	D445	3.714		----	1064	D445	3.741		----
120	D445	3.725		----	1065				----
140				----	1082	ISO3104	3.73311631		----
150	D445	3.803		----	1097	ISO3104	3.7416		----
159	D445	3.681		----	1105	D445	3.7214		----
169	D445	3.878	C,R(0.05)	----	1109				----
171				----	1121				----
175	D445	3.733		----	1126				----
177	D445	3.753		----	1140	D445	4.116	R(0.01)	----
215				----	1182	D7042	4.0437	C,R(0.01)	----
221				----	1191	ISO3104	3.732102656		----
224				----	1205				----
225	D445	3.713		----	1237	ISO3104	3.7162		----
228				----	1275	IP71	3.743		----
235	D445	3.766		----	1279				----
237	D445	3.775		----	1299	D445	3.759		----
238				----	1316	ISO3104	3.7185		----
253	D445	3.7669		----	1318	D7042	3.715		----
254	D445	3.748	C	----	1320	D445	3.690		----
256				----	1357				----
258				----	1372	D7042	3.7129		----
273				----	1397	D7042	3.751		----
311	D445	3.737		----	1399				----
317				----	1412				----
323	D445	3.747		----	1417	D445	3.776		----
328				----	1444	D445	3.761		----
333	D445	3.732		----	1455	D7042	3.719		----
334	D445	3.734		----	1496				----
335				----	1538				----
365				----	1575				----
372	D445	3.733		----	1585	D445	3.715		----
391				----	1586	D445	3.663		----
396	D445	3.736		----	1587	D445	3.9000	R(0.05)	----
399				----	1610	D7042	3.6631		----
440				----	1613	D445	3.7353		----
445	D445	3.736		----	1616	D445	3.755		----
447	D445	3.798		----	1631	D7042	3.7737		----
460				----	1634	D445	3.724		----
467	D7042	3.7462		----	1636	D445	3.736		----
480				----	1669	D445	3.840		----
496				----	1683	D445	3.722		----
603				----	1688	D445	4.1	R(0.01)	----
608				----	1715				----
631				----	1720	D445	3.747		----
657	D445	3.712		----	1724	D445	3.844		----
736	D445	3.725		----	1730				----
823	D445	3.742		----	1757	D7042	3.7566		----
824	D445	3.726		----	1776	D445	3.6528		----
851	D445	3.745		----	1833		3.753		----
854				----	1852	D7945	3.731		----
862				----	1913				----
869				----	1944				----
904				----	1961				----
914	D445	3.726		----	2130	IP71	3.726		----
962				----	2133	D445	3.7323		----
963	D445	3.7932	C	----	6028				----
970				----	6041				----
974	D445	3.720		----	6054				----
994	D445	3.798	C	----	6075	D445	3.7444		----
995	D445	3.801	C	----	6114	D445	3.739		----
996				----	6135				----
997				----	6142				----
1011	D445	3.759		----	6203	D445	3.6952		----
1016	D445	3.7155		----	6244				----
1019	EN16896	3.629	R(0.05)	----	6266	D7042	3.644		----
1023				----	6274				----
1039	D7945	3.734		----	6312				----
1040	D7042	3.734		----	6315	D7042	3.726		----
1049	D445	3.729		----	6321	IP71	3.734		----
1059				----	6324				----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		-----		-----	6421	D445	3.810		-----
6344	ISO3104	3.671		-----	6438		-----		-----
6346		-----		-----	6479		-----		-----
6364	D445	3.8742	R(0.05)	-----	6487		-----		-----
6384	D445	3.730		-----					

normality suspect  
 n 80  
 outliers 7  
 mean (n) 3.73803  
 st.dev. (n) 0.036241  
 R(calc.) 0.10147  
 st.dev. (D445:21e1) (0.006558)  
 R(D445:21e1) (0.01836)  
 compare  
 R (IP71-1:20) (0.01836) procedure A  
 R(D445:19) (0.07102)  
 R (IP71-1:96) (0.07102)

Lab 169 first reported 3.619  
 Lab 254 first reported 3.981  
 Lab 963 first reported 3.627  
 Lab 994 first reported 4.057  
 Lab 995 first reported 3.998  
 Lab 1182 first reported 3.5339





## Determination of Mercaptan Sulfur as S on sample #22150; results in %M/M

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D3227	0.0004		-0.17	1062	D3227	0.0004		-0.17
62	D3227	0.0004		-0.17	1064	D3227	0.00035		-0.59
120	D3227	0.00037		-0.42	1065	D3227	0.00034		-0.68
140	D3227	0.00032	C	-0.85	1082		----		----
150		----		----	1097	ISO3012	0.00039		-0.25
159	D3227	0.0004		-0.17	1105	D3227	0.000364		-0.47
169	D3227	0.0006		1.54	1109		----		----
171	D3227	0.0003		-1.02	1121		----		----
175	D3227	0.0004		-0.17	1126		----		----
177	D3227	0.0004		-0.17	1140	IP342	0.00196	R(0.01)	13.17
215		----		----	1182		----		----
221		----		----	1191	ISO3012	0.000421	C	0.01
224		----		----	1205		----		----
225		----		----	1237		----		----
228		----		----	1275	IP342	0.0003		-1.02
235	D3227	0.00042		0.00	1279		----		----
237	D3227	0.00055		1.12	1299	D3227	0.0003		-1.02
238		----		----	1316		----		----
253	D3227	0.0005		0.69	1318	D3227	0.0004		-0.17
254		----		----	1320	D3227	0.00043		0.09
256		----		----	1357	D3227	0.0003		-1.02
258		----		----	1372		0.00122	R(0.01)	6.84
273	D3227	0.0004		-0.17	1397		----		----
311	D3227	0.00036		-0.51	1399		----		----
317		----		----	1412	UOP163	0.0004		-0.17
323	D3227	0.00035		-0.59	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D3227	0.0005		0.69
334	D3227	0.0004		-0.17	1496		----		----
335	D3227	0.00043		0.09	1538		----		----
365		----		----	1575		----		----
372	D3227	0.00052		0.86	1585	UOP163	0.00035		-0.59
391	D3227	0.00047		0.43	1586	D3227	0.0005		0.69
396		----		----	1587		----		----
399		----		----	1610	IP342	0.0004		-0.17
440		----		----	1613	D3227	0.00059		1.46
445	D3227	0.0004		-0.17	1616	D3227	0.00044		0.18
447	D3227	0.0006		1.54	1631		----		----
460		----		----	1634	D3227	0.0004605		0.35
467	D3227	0.00050		0.69	1636	D3227	0.00040		-0.17
480		----		----	1669	D3227	0.000329		-0.77
496	D3227	0.00033		-0.76	1683	D3227	0.0006		1.54
603		----		----	1688	D3227	0.0003		-1.02
608		----		----	1715		----		----
631		----		----	1720		----		----
657	D3227	<0.0003		----	1724	D3227	0.0004		-0.17
736	D3227	0.0003		-1.02	1730		----		----
823	D3227	0.0004		-0.17	1757		----		----
824	D3227	0.00042		0.00	1776		----		----
851	D3227	0.000455		0.30	1833		0.0004		-0.17
854		----		----	1852	D3227	0.0004		-0.17
862		----		----	1913		----		----
869		----		----	1944		----		----
904		----		----	1961		----		----
914	D3227	0.00051		0.77	2130		----		----
962	D3227	0.0005		0.69	2133	D3227	0.00012	R(0.05)	-2.56
963	D3227	0.0005		0.69	6028		----		----
970		----		----	6041		----		----
974	D3227	0.0005		0.69	6054		----		----
994	D3227	0.00044		0.18	6075	D3227	0.0005		0.69
995	D3227	0.00048		0.52	6114	D3227	0.00043		0.09
996		----		----	6135		----		----
997		----		----	6142	IP342	0.0004		-0.17
1011	D3227	0.0004		-0.17	6203	D3227	0.000436		0.14
1016	D3227	0.000378		-0.35	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039	IP342	0.000461		0.36	6312		----		----
1040	D3227	0.000396		-0.20	6315	D3227	0.00046		0.35
1049	D3227	0.00040742		-0.10	6321	IP342	0.0005		0.69
1059	D3227	0.0005		0.69	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D3227	0.0003		-1.02	6421	D3227	0.001	R(0.01)	4.96
6344		----		----	6438	UOP163	0.00030		-1.02
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384	D3227	0.00038		-0.34					
	normality	OK							
	n	74							
	outliers	4							
	mean (n)	0.000419							
	st.dev. (n)	0.0000761							
	R(calc.)	0.000213							
	st.dev.(D3227:16)	0.0001170							
	R(D3227:16)	0.000328							

Lab 140 first reported 0.0011  
 Lab 1191 reported 4.21 %M/M





Determination of MSEP on sample #22150;

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D3948	94		-0.76	1062	D3948	85	R(0.01)	-4.35
62	D3948	96		0.04	1064	D7224	96		0.04
120	D3948	98		0.84	1065		----		----
140	D3948	92		-1.56	1082	D3948	97		0.44
150	D3948	98		0.84	1097	D3948	96		0.04
159	D3948	97		0.44	1105	D3948	93		-1.16
169	D3948	91		-1.96	1109		----		----
171	D3948	90		-2.35	1121	D3948	99		1.23
175		----		----	1126		----		----
177	D7224	98		0.84	1140	D3948	95		-0.36
215	D3948	97		0.44	1182		----		----
221		----		----	1191	D3948	92		-1.56
224		----		----	1205		----		----
225	D3948	95		-0.36	1237		----		----
228	D3948	99.0		1.23	1275	D3948	98		0.84
235	D7224	97		0.44	1279		----		----
237	D3948	96		0.04	1299	D3948	98		0.84
238		----		----	1316	D3948	92		-1.56
253	D3948	96		0.04	1318	D3948	95		-0.36
254		----		----	1320	D3948	96		0.04
256	D7224	99		1.23	1357	D3948	93		-1.16
258	D3948	97		0.44	1372		98		0.84
273		----		----	1397	D7224	96		0.04
311		----		----	1399		----		----
317	D7224	93		-1.16	1412		----		----
323	D3948	95		-0.36	1417		----		----
328	D7224	97		0.44	1444		----		----
333	D3948	95		-0.36	1455	D7224	95		-0.36
334	D3948	93		-1.16	1496		----		----
335	D7224	97		0.44	1538	D7224	98		0.84
365	D7224	95		-0.36	1575		----		----
372		----		----	1585	D3948	98		0.84
391		----		----	1586	D3948	95		-0.36
396	D3948	99		1.23	1587	D7224	93		-1.16
399	D3948	97		0.44	1610	D3948	94		-0.76
440		----		----	1613	D3948	95		-0.36
445	D7224	96		0.04	1616	D3948	93		-1.16
447	D3948	98		0.84	1631	D7224	97		0.44
460		----		----	1634	D7224	95		-0.36
467	D3948	90	C	-2.35	1636		----		----
480		----		----	1669	D3948	98		0.84
496	D3948	93.5		-0.96	1683	D3948	99		1.23
603		----		----	1688	D3948	99		1.23
608	D3948	99		1.23	1715		----		----
631	D7224	94		-0.76	1720	D3948	100		1.63
657	D3948	99		1.23	1724	D3948	94		-0.76
736	D3948	95		-0.36	1730		----		----
823	D3948	98		0.84	1757	D3948	96		0.04
824	D3948	97		0.44	1776	D3948	97		0.44
851	D3948	93		-1.16	1833		95		-0.36
854		----		----	1852	D3948	97		0.44
862		----		----	1913		----		----
869		----		----	1944		----		----
904	D3948	96		0.04	1961	D3948	97		0.44
914	D3948	99		1.23	2130	D7224	91		-1.96
962	D3948	94		-0.76	2133	D3948	94		-0.76
963	D3948	97		0.44	6028		----		----
970	D3948	95		-0.36	6041		----		----
974	D7224	95		-0.36	6054		----		----
994		----		----	6075	D7224	97.5		0.64
995		----		----	6114	D7224	98		0.84
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D3948	95		-0.36	6203	D3948	95		-0.36
1016	D3948	95		-0.36	6244		----		----
1019		----		----	6266		----		----
1023	D3948	97		0.44	6274		----		----
1039	D3948	96		0.04	6312		----		----
1040	D3948	98		0.84	6315	D3948	98		0.84
1049	D7224	97		0.44	6321	D3948	96		0.04
1059	D3948	95		-0.36	6324	D3948	95		-0.36



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D3948	98		0.84	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384	D3948	97		0.44					
	normality	OK							
	n	95							
	outliers	1							
	mean (n)	95.91							
	st.dev. (n)	2.206							
	R(calc.)	6.18							
	st.dev.(D3948:20)	2.508							
	R(D3948:20)	7.02							

Lab 467 first reported 87





## Determination of Naphthalenes on sample #22150; results in %V/V

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D1840-A	0.71		-0.44	1062	D1840-A	0.74		0.65
62	D1840-A	0.74	C	0.65	1064	D1840-A	0.735		0.47
120		----		----	1065	D1840-A	0.76		1.37
140		----		----	1082		----		----
150		----		----	1097	D1840-A	0.729		0.25
159	D1840-B	0.9038	C,R(0.01)	6.58	1105	D1840-A	0.704		-0.65
169	D1840-B	0.73		0.29	1109		----		----
171	D1840-B	0.79	R(0.01)	2.46	1121	D1840-B	0.721		-0.04
175	D1840-B	0.74		0.65	1126		----		----
177	D1840-B	0.73	C	0.29	1140		----		----
215		----		----	1182		----		----
221		----		----	1191	D1840	0.7120737		-0.36
224		----		----	1205		----		----
225		----		----	1237	D1840-B	0.722		0.00
228		----		----	1275	D1840-A	0.7117		-0.37
235		----		----	1279		----		----
237	D1840-B	0.72		-0.07	1299	D1840-B	0.70		-0.80
238		----		----	1316	D1840-B	0.72		-0.07
253	D1840-B	0.70		-0.80	1318	D1840-A	0.70		-0.80
254		----		----	1320		----		----
256		----		----	1357		----		----
258		----		----	1372		----		----
273		----		----	1397		----		----
311		----		----	1399		----		----
317		----		----	1412		----		----
323	D1840-A	0.72		-0.07	1417		----		----
328		----		----	1444		----		----
333	D1840-B	0.69		-1.16	1455	D1840-B	0.74		0.65
334	D1840-B	0.70	C	-0.80	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D1840-B	0.739		0.61	1585	D1840-B	0.794	R(0.01)	2.61
391	D1840-A	0.71		-0.44	1586	D1840-A	0.76		1.37
396		----		----	1587		----		----
399		----		----	1610	D1840-B	0.71		-0.44
440		----		----	1613	D1840-A	0.74		0.65
445	D1840-A	0.72		-0.07	1616	D1840-B	0.716		-0.22
447		----		----	1631		----		----
460		----		----	1634	D1840-A	0.71		-0.44
467	D1840-B	0.748		0.94	1636	D1840-B	0.717		-0.18
480		----		----	1669	D1840-B	0.55	R(0.01)	-6.23
496		----		----	1683	D1840-B	0.76	C	1.37
603		----		----	1688		----		----
608		----		----	1715		----		----
631		----		----	1720	D1840-B	0.72		-0.07
657	D1840-A	0.72		-0.07	1724		----		----
736	D1840-B	0.732		0.36	1730		----		----
823	D1840-A	0.74		0.65	1757		----		----
824	D1840-B	0.72		-0.07	1776	D1840-A	0.74		0.65
851	D1840	0.73		0.29	1833		----		----
854		----		----	1852	D1840-A	0.7105		-0.42
862		----		----	1913		----		----
869		----		----	1944		----		----
904		----		----	1961		----		----
914	D1840-A	0.73		0.29	2130	D1840-A	0.72		-0.07
962	D1840-A	0.70		-0.80	2133	D1840-A	0.691		-1.12
963	D1840-A	0.74		0.65	6028		----		----
970		----		----	6041		----		----
974	D1840-A	0.71		-0.44	6054		----		----
994	D1840-A	0.714		-0.29	6075	D1840-A	0.72		-0.07
995	D1840-B	0.70		-0.80	6114	D1840-A	0.73		0.29
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D1840-B	0.72		-0.07	6203		----		----
1016	D1840-A	0.72		-0.07	6244		----		----
1019		----		----	6266		----		----
1023		----		----	6274		----		----
1039	D1840-B	0.72		-0.07	6312		----		----
1040	D1840-A	0.994	R(0.01)	9.85	6315	D1840-A	0.717		-0.18
1049	D1840-A	0.711		-0.40	6321	D1840-B	0.70		-0.80
1059	D1840-B	0.74		0.65	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332		----		----	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364		----		----	6487		----		----
6384	D1840-B	0.86	R(0.01)	5.00					
	normality	OK							
	n	59							
	outliers	6							
	mean (n)	0.7220							
	st.dev. (n)	0.01657							
	R(calc.)	0.0464							
	st.dev.(D1840-B:07R17)	0.02761							
	R(D1840-B:07R17)	0.0773							
	compare								
	R(D1840-A:07R17)	0.0515							

Lab 62 first reported 0.94  
 Lab 159 first reported 0.658  
 Lab 177 first reported 0.89  
 Lab 334 first reported 1.13  
 Lab 1683 first reported 0.8013





## Determination of Smoke Point on sample #22150; results in mm

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D1322-manual	21.5		-0.96	1062	D1322-manual	23.5		0.59
62	D1322-manual	21.0		-1.34	1064	D1322-automated	22.6		-0.11
120	D1322-automated	23.1		0.28	1065	D1322-automated	22.2		-0.41
140	D1322-automated	22.2		-0.41	1082	D1322-automated	22.6		-0.11
150	D1322-automated	21.6		-0.88	1097	D1322-automated	22.0		-0.57
159	D1322-automated	24.4		1.28	1105	D1322-automated	22.80		0.05
169	D1322-automated	21.9		-0.65	1109		----		----
171	D1322-automated	21.9		-0.65	1121	D1322-manual	23.0		0.20
175		----		----	1126		----		----
177	D1322-automated	22.4		-0.26	1140	D1322-manual	21.0		-1.34
215		----		----	1182		----		----
221		----		----	1191	D1322	22.55		-0.14
224		----		----	1205		----		----
225		----		----	1237		----		----
228	D1322-automated	21.6		-0.88	1275	IP598-automated	22.2		-0.41
235	D1322-manual	24.8		1.59	1279		----		----
237	D1322-automated	21.3		-1.11	1299	D1322-automated	22.3		-0.34
238		----		----	1316	D1322-automated	22.5		-0.18
253	D1322-manual	23		0.20	1318	D1322-automated	22.8		0.05
254	D1322-manual	23		0.20	1320	D1322-automated	21.9		-0.65
256	D1322-manual	23		0.20	1357	D1322-manual	24		0.97
258		24		0.97	1372	D1322-automated	22.1		-0.49
273	D1322-manual	23.2		0.36	1397		----		----
311	D1322-manual	21.8		-0.72	1399		----		----
317		----		----	1412	D1322-manual	25.0		1.75
323	D1322-automated	22.0		-0.57	1417		----		----
328		----		----	1444		----		----
333	D1322-automated	21.8		-0.72	1455	D1322-automated	23.0		0.20
334	D1322-automated	22.4		-0.26	1496		----		----
335	D1322-automated	21.9		-0.65	1538	D1322-automated	21.9		-0.65
365	IP57-manual	22.96		0.17	1575	D1322-manual	24.2		1.13
372	D1322-automated	22.5		-0.18	1585	D1322-automated	21.95		-0.61
391	D1322-manual	22.5		-0.18	1586	D1322-manual	24		0.97
396		----		----	1587	D1322-automated	22.6		-0.11
399		----		----	1610	IP598-manual	24		0.97
440	D1322-manual	25.0		1.75	1613	D1322-automated	22.3		-0.34
445		----		----	1616	D1322-automated	22.5		-0.18
447	D1322-manual	25		1.75	1631		----		----
460		----		----	1634	D1322-automated	21.9		-0.65
467	D1322-manual	23.98		0.96	1636	D1322-automated	22.6		-0.11
480		----		----	1669	D1322-automated	22.6		-0.11
496		----		----	1683	D1322-automated	22.4		-0.26
603		----		----	1688	D1322-automated	22		-0.57
608		----		----	1715	D1322-manual	23.2		0.36
631	D1322-automated	21.6		-0.88	1720	D1322-automated	21.3		-1.11
657	D1322-automated	21.2		-1.19	1724	D1322-manual	21		-1.34
736	D1322-manual	22.9		0.13	1730		----		----
823	D1322-automated	22.3		-0.34	1757	D1322-manual	25.0		1.75
824	D1322-automated	21.9		-0.65	1776	D1322-automated	22.1		-0.49
851	D1322-manual	24.5		1.36	1833		22.2		-0.41
854		----		----	1852	D1322-automated	21.931		-0.62
862		----		----	1913		----		----
869		----		----	1944	D1322-manual	24.7		1.52
904	D1322-manual	24		0.97	1961		----		----
914	D1322-automated	21.3		-1.11	2130		----		----
962	D1322-manual	22		-0.57	2133	D1322-manual	25.33		2.00
963	D1322-manual	22.0		-0.57	6028		----		----
970	D1322-manual	23		0.20	6041		----		----
974	D1322-automated	23.0		0.20	6054		----		----
994	D1322-manual	24.5		1.36	6075	D1322-automated	22.5		-0.18
995	D1322-manual	23.9		0.90	6114	D1322-manual	22.5		-0.18
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D1322-automated	21.7		-0.80	6203	D1322-manual	24.1		1.05
1016	IP598-automated	22.1		-0.49	6244		----		----
1019		----		----	6266	D1322-automated	22.1		-0.49
1023		----		----	6274		----		----
1039	D1322-automated	22.2		-0.41	6312		----		----
1040		----		----	6315	D1322-automated	22.6		-0.11
1049	D1322-automated	22.2		-0.41	6321	D1322-automated	22.5		-0.18
1059	D1322-automated	22.6		-0.11	6324	D1322-manual	24.0		0.97

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D1322-manual	24.5	C	1.36	6421	D1322	24		0.97
6344	D1322-manual	21.5	C	-0.96	6438		----		----
6346		----		----	6479		----		----
6364	D1322-manual	24.0		0.97	6487		----		----
6384	D1322-automated	22.3		-0.34					

		<u>only D1322-manual</u>	<u>only D1322-auto</u>
normality	OK	OK	not OK
n	99	38	53
outliers	0	0	0
mean (n)	22.74	23.40	22.22
st.dev. (n)	1.060	1.257	0.540
R(calc.)	2.97	3.52	1.51
st.dev.(D1322-M:22)	1.295	1.317	---
R(D1322-M:22)	3.63	3.69	---
compare			
R(D1322-A:22)	0.87	---	0.86

Lab 6332 first reported 25.5  
 Lab 6344 first reported 25.4





Determination of Specific Energy (Net) on Sulfur free basis on sample #22150; results in MJ/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D3338	43.179		-1.06	1062	D3338	43.159		-2.28
62	D3338	43.206		0.58	1064	D3338	43.216		1.19
120		----		----	1065		----		----
140		----		----	1082	D3338	43.18004		-1.00
150		----		----	1097	D3338	43.200		0.22
159		----		----	1105	D3338	43.20		0.22
169	D3338	43.220		1.43	1109		----		----
171	D3338	43.202		0.34	1121		----		----
175		----		----	1126		----		----
177	D3338	43.207		0.64	1140		----		----
215		----		----	1182		----		----
221		----		----	1191		----		----
224		----		----	1205		----		----
225		----		----	1237	D3338	43.227	C	1.86
228		----		----	1275	D3338	43.189		-0.45
235	D3338	43.203		0.40	1279		----		----
237	D3338	43.223		1.62	1299	D3338	43.2	C	0.22
238		----		----	1316	D3338	43.208		0.70
253	D3338	43.2025		0.37	1318	D3338	43.200		0.22
254	D3338	43.20		0.22	1320	D3338	43.19		-0.39
256		----		----	1357	D3338	43.217		1.25
258		----		----	1372		43.1869		-0.58
273		----		----	1397		----		----
311		----		----	1399		----		----
317		----		----	1412		----		----
323	D3338	43.186		-0.63	1417		----		----
328		----		----	1444		----		----
333		----		----	1455	D3338	43.200	C	0.22
334	D3338	43.182		-0.88	1496		----		----
335		----		----	1538		----		----
365		----		----	1575		----		----
372	D3338	43.205		0.52	1585	D3338	43.199		0.16
391		----		----	1586	D3338	43.187	C	-0.57
396		----		----	1587		----		----
399		----		----	1610	D3338	43.196		-0.03
440		----		----	1613	D3338	43.203		0.40
445		----		----	1616	D3338	43.207		0.64
447		----		----	1631	D3338	43.185		-0.70
460		----		----	1634	D3338	43.213		1.01
467	D3338	43.2139		1.06	1636	D3338	43.210		0.83
480		----		----	1669		----		----
496		----		----	1683		----		----
603		----		----	1688		----		----
608		----		----	1715		----		----
631		----		----	1720	D3338	43.21		0.83
657	D3338	43.195		-0.09	1724	D3338	43.19	C	-0.39
736	D3338	43.196		-0.03	1730		----		----
823	D3338	43.192		-0.27	1757	D4529	43.209		0.77
824	D3338	43.196		-0.03	1776	D3338	43.18		-1.00
851	D3338	43.199		0.16	1833		43.175	C	-1.30
854		----		----	1852	D3338	43.1980		0.10
862		----		----	1913		----		----
869		----		----	1944		----		----
904		----		----	1961		----		----
914	D3338	43.209		0.77	2130	D3338	43.218		1.31
962		----		----	2133	D3338	43.1728		-1.44
963	D3338	43.21		0.83	6028		----		----
970		----		----	6041		----		----
974	D3338	43.175		-1.30	6054		----		----
994	D3338	43.18		-1.00	6075	D3338	43.2014		0.30
995	D3338	43.169		-1.67	6114	D3338	43.208		0.70
996		----		----	6135		----		----
997		----		----	6142		----		----
1011	D3338	43.193		-0.21	6203	D3338	43.1904		-0.37
1016		----		----	6244		----		----
1019	D4809	43.297	R(0.01)	6.12	6266		----		----
1023		----		----	6274		----		----
1039	D3338	43.186		-0.63	6312		----		----
1040		----		----	6315	D3338	43.14	R(0.05)	-3.43
1049	D3338	43.17897		-1.06	6321	D3338	43.145	R(0.05)	-3.13
1059	D3338	43.185		-0.70	6324		----		----

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D3338	43.2		0.22	6421		----		----
6344		----		----	6438		----		----
6346		----		----	6479		----		----
6364	D3338	43.1757		-1.26	6487		----		----
6384	D3338	43.181		-0.94					
normality		OK							
n		63							
outliers		3							
mean (n)		43.1964							
st.dev. (n)		0.01428							
R(calc.)		0.0400							
st.dev.(D3338:20a)		0.01643							
R(D3338:20a)		0.046							

Lab 1237 first reported 43.282  
 Lab 1299 first reported 46.76  
 Lab 1455 first reported 43200 MJ/kg  
 Lab 1586 first reported 43187 MJ/kg  
 Lab 1724 first reported 42.9  
 Lab 1833 reported 43175 MJ/kg





## Determination of Total Sulfur on sample #22150; results in mg/kg

lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
52	D5453	326		-1.54	1062	D4294	399		2.80
62	D5453	304		-2.85	1064	D5453	363.1		0.67
120		----		----	1065	D5453	340		-0.71
140	D2622	357.49		0.33	1082	ISO8754	355.7		0.23
150	D5453	358		0.37	1097	D5453	357.03		0.31
159		----		----	1105	D4294	341.70		-0.60
169	D4294	372	C	1.20	1109		----		----
171	D2622	364		0.72	1121	IP336	367		0.90
175		----		----	1126	ISO20846	347.0		-0.29
177	D4294	322		-1.78	1140	IP336	375		1.38
215		----		----	1182		----		----
221		----		----	1191	ISO8754	353.9		0.12
224		----		----	1205	ISO14596	352.9		0.06
225		----		----	1237	ISO8754	330		-1.30
228		----		----	1275	D4294	340.58		-0.67
235	D5453	320.2		-1.88	1279		----		----
237	D4294	338		-0.82	1299	ISO8754	360	C	0.48
238		----		----	1316	D4294	141.6	C,R(0.01)	-12.50
253		----		----	1318	D5453	349.8		-0.12
254	D4294	353.8		0.12	1320		----		----
256		----		----	1357	D5453	370		1.08
258		172	R(0.01)	-10.69	1372	D4294	357.6		0.34
273	D5453	355		0.19	1397	D2622	350		-0.11
311	D2622	348		-0.23	1399		----		----
317		----		----	1412	D4294	391		2.33
323	D5453	314		-2.25	1417	In house	390		2.27
328	ISO20847	354		0.13	1444		----		----
333	D4294	357		0.31	1455	D2622	357		0.31
334	D5453	368		0.96	1496		----		----
335	D4294	339	C	-0.76	1538	D5453	365.45		0.81
365		----		----	1575	D4294	380		1.67
372	D5453	358		0.37	1585	D4294	361.5		0.57
391	D4294	361		0.54	1586	ISO20846	370		1.08
396		----		----	1587		----		----
399	D4294	359.8		0.47	1610	IP336	353		0.07
440	D5453	280		-4.27	1613	D4294	350.0		-0.11
445	D5453	340.49		-0.68	1616	D4294	360		0.48
447	IP336	350	C	-0.11	1631		----		----
460		----		----	1634	D5453	307		-2.67
467	D4294	370		1.08	1636	D4294	355		0.19
480	D4294	342.07		-0.58	1669	D4294	343		-0.53
496	D2622	349.9		-0.12	1683	D4294	350		-0.11
603		----		----	1688	D4294	370		1.08
608	D4294	374.50		1.35	1715		----		----
631	D4294	357.5		0.34	1720	D5453	322.96		-1.72
657	D5453	355		0.19	1724	IP336	357	C	0.31
736	ISO20884	355		0.19	1730		----		----
823	D5453	338		-0.82	1757		----		----
824	D4294	326		-1.54	1776	ISO20846	310		-2.49
851	D4294	370		1.08	1833		352		0.01
854		----		----	1852	D5453	336		-0.94
862		----		----	1913		----		----
869		----		----	1944	D5453	374.21		1.33
904	D4294	316		-2.13	1961		----		----
914	D4294	374		1.32	2130	IP336	300		-3.08
962		----		----	2133	D4294	313.1		-2.30
963	D4294	367		0.90	6028		----		----
970	D4294	350		-0.11	6041		----		----
974	D4294	350		-0.11	6054		----		----
994	D5453	381		1.73	6075	D5453	383		1.85
995	D5453	353		0.07	6114	D4294	347		-0.29
996		----		----	6135		----		----
997		----		----	6142	ISO20846	415.08		3.76
1011	D4294	370	C	1.08	6203	D2622	340		-0.71
1016		----		----	6244		----		----
1019	D1552	497	R(0.01)	8.63	6266		----		----
1023	D4294	400	C	2.86	6274		----		----
1039	D2622	351		-0.05	6312		----		----
1040		318.5		-1.98	6315	D2622	359		0.42
1049	D5453	356.3		0.26	6321	IP336	404		3.10
1059	ISO14596	320		-1.89	6324	D4294	398.0		2.74



lab	method	value	mark	z(targ)	lab	method	value	mark	z(targ)
6332	D4294	320	C	-1.89	6421	D5453	315		-2.19
6344		----		----	6438		----		----
6346		----		----	6479	IP600	360.8		0.53
6364	D4294	320		-1.89	6487		----		----
6384	D4294	349		-0.17					
	normality	OK							
	n	99							
	outliers	3							
	mean (n)	351.86							
	st.dev. (n)	23.767							
	R(calc.)	66.55							
	st.dev.(D5453:19a)	16.820							
	R(D5453:19a)	47.10							
	compare								
	R(D4294:21)	84.00							
	R(D2622:21)	46.95							



Lab 169 first reported 275  
 Lab 335 first reported 0.0339 mg/kg  
 Lab 447 reported 0.035 mg/kg  
 Lab 1011 reported 0.0370 mg/kg  
 Lab 1023 first reported 0.04 mg/kg  
 Lab 1299 first reported 0.036 mg/kg  
 Lab 1316 reported 0.01416 mg/kg  
 Lab 1724 first reported 0.042 mg/kg  
 Lab 6332 first reported 0.032 mg/kg



Determination of BOCLE on sample #22151; results in mm

lab	method	value	mark	z(targ)	remarks
150	D5001 full-automated	0.65		-0.03	
171	D5001 semi-automated	0.58	G(0.05)	-3.32	
237	D5001 semi-automated	0.67		0.91	
323		----		----	
334	D5001 semi-automated	0.66		0.44	
496	D5001 full-automated	0.668		0.81	
862		----		----	
963	D5001 full-automated	0.693		1.99	
1016	D5001 semi-automated	0.6355		-0.71	
1040	D5001 semi-automated	0.627		-1.11	
1064	D5001 full-automated	0.635		-0.74	
1082	D5001 semi-automated	0.66		0.44	
1135	D5001 full-automated	0.62		-1.44	
1279		----		----	
1316	D5001 semi-automated	0.652		0.06	
1320	D5001 semi-automated	0.65		-0.03	
1357	D5001 full-automated	0.65		-0.03	
1399		----		----	
1496		----		----	
1538	D5001 full-automated	0.65		-0.03	
1631	D5001 semi-automated	0.6465		-0.20	
1724	D5001 full-automated	0.67		0.91	
1833	D5001 semi-automated	0.64		-0.50	
1913		----		----	
6041		----		----	
6274		----		----	
6344	D5001 semi-automated	0.635		-0.74	









		<u>only semi-automated</u>	<u>only full-automated</u>
normality	OK	OK	OK
n	18	10	8
outliers	1	1	0
mean (n)	0.6507	0.6476	0.6545
st.dev. (n)	0.0178	0.01343	0.02248
R(calc.)	0.0498	0.0376	0.0629
st.dev.(D5001:19e1 (semi-automated))	0.0213	0.02115	---
R(D5001:19e1 (semi-automated))	0.0597	0.0592	---
compare			
R(D5001:19e1 (full-automated))	0.0335	---	0.0340







Determination of Particle Size Distribution on sample #22152 acc. to IP565, results in counts/mL

lab	method	≥4 μm (c)	m	≥6 μm(c)	m	≥14 μm (c)	m	≥21 μm (c)	m	≥25 μm (c)	m	≥30 μm (c)	m
140		----		----		----		----		----		----	
150		----		----		----		----		----		----	
171	IP565	13626		3873		155		23		8		4	
225		----		----		----		----		----		----	
237		----		----		----		----		----		----	
311	IP565	13800		4121		115		12		2		<1	
323		----		----		----		----		----		----	
333	IP565	15571		4691		176		21		5		2	
334	IP565	14541		4296		177		26		9		3	
335		----		----		----		----		----		----	
372	IP565	16462		5144	ex	218		41.6		17.2		5.8	
447		----		----		----		----		----		----	
657	IP565	12568.4		4464.3		186.6		37.0		15.1		3.7	
823	IP565	17091.4		5336.5		226.9		41.3	C	16.6	C	7.1	
824	IP565	17542.7		5018.4		214.4		39.5		15.6		5.5	
862		----		----		----		----		----		----	
963	IP565	14533	ex,C	4275		134		17.3		5.6		1.0	
974	IP565	14250		4140		120		15		5		2	
1011	IP565	13629		4462		148		24		8		3	
1016	IP565	14684.0		4215.8		124.7		10.6		2.3		0.9	
1049	IP577	146016	ex	53725.7	ex	2436	ex	461.7	ex	166.0	ex	45.3	ex
1062	IP565	14008.5		4037.2	C	118.7		7.7		1.2		0.2	
1064	IP565	14545.9		4544.7		109.1		3.4		0.5		0.2	
1065	IP565	12733		4034.8		132.7		19.5		6.2		1.3	
1097	IP564	9279.9	ex	2929.7	ex	57.1	ex	22.5	ex	19.3	ex	18.0	ex
1109		----		----		----		----		----		----	
1135	IP565	12831		4390		79		10		2		0	
1140	IP565	15874.5		4740.1		158.0		15.9		3.4		0.9	
1191		15278.2		4257.8		170.4		16.5		4.9		0.9	
1279		----		----		----		----		----		----	
1299	IP577	9514.7	ex	2523.2	ex	40.9	ex	6.8	ex	2.7	ex	0.9	ex
1316	IP577	11123.6	ex	3562.4	ex	135.4	ex	25.1	ex	13.8	ex	5.2	ex
1318	IP565	14789.5		4273.4		188.8		28.1		11.2		4.0	
1320		----		----		----		----		----		----	
1357	IP565	14576		4014		161	ex	11.2		3.5		1.3	
1397		----		----		----		----		----		----	
1399		----		----		----		----		----		----	
1402	IP565	13664.7		4106.4		133.1		24.2		11.3		5.5	
1455	IP565	13577		4594		145		15		3		0	
1496		----		----		----		----		----		----	
1538	IP565	13514.6		----		----		----		----		----	
1585	IP565	15507.8		4336.5		138.8		14.2		5.0		1.5	
1587	IP565	13344.5		3691.8		183.5		25.2		9.1		3.2	
1610	IP565	17080.4		4843.3		149.3		16.4		3.9		0.7	
1613	IP565	15371.4		4958.7		164.2		28.9		11.1		3.4	
1634	IP565	14598		4754		143		23		10		3	
1636	IP565	12655.4		4227.6		148.7		15.7		4.0		0.4	
1724	IP565	14880.5		4302		193.8		24.7		11.3		4.2	
1833	IP565	15585	ex	4406	ex	193.5	ex	39	ex	15	ex	5.4	ex
1852	IP565	16104.5		4804.0		154.1		21.7		6.9		2.1	
1857	IP565	16036		4825		168		----		----		----	
1913		----		----		----		----		----		----	
1961	IP577	10765.2	ex	3568.0	ex	113.8	ex	14.3	ex	3.6	ex	0.7	ex
2130	IP565	14489.1		3675.6		165.6		21.0		7.3		1.9	
6075	IP565	14597.7		4208.3		127.3		15.5		3.8		1.2	
6112	IP565	15359.7		4661.9		131.6		15.3		2.9		0.2	
6203	IP565	15670.4		4539.8		146.9		15.5		4.4		1.1	
6274		----		----		----		----		----		----	
6315	IP565	13354		4569		169		23		7		1	
6321	IP565	14653.6		4293.2		174.6		12.1		3.6		0.9	
normality		OK		OK		OK		OK		OK		OK	
n		37		36		36		36		36		35	
outliers		0 + 7ex		0 + 7ex		0 + 7ex		0 + 6ex		0 + 6ex		0 + 6ex	
mean (n)		14671.93		4404.89		155.24		20.33		6.86		2.20	
st.dev. (n)		1276.815		368.700		31.898		9.125		4.466		1.856	
R(calc.)		3575.08		1032.36		89.31		25.55		12.51		5.20	
st.dev.(IP565:13)		592.985		349.123		31.846		6.579		2.878		1.234	
R(IP565:13)		1660.36		977.54		89.17		18.42		8.06		3.45	

- Lab 372 test result excluded as test result in counts/mL didn't match test result in ISO scale number
- Lab 823 first reported 49.2 and 20.2 respectively
- Lab 963 test result excluded as test result in counts/mL didn't match test result in ISO scale number, first reported 8433
- Lab 1049 test results excluded as test method IP577 was used and test results in counts/mL didn't match test results in ISO scale number, see also §4.1
- Lab 1062 first reported 40037.2
- Lab 1097 test results excluded as test method IP564 was used, see also §4.1
- Lab 1299 test results excluded as test method IP577 was used, see also §4.1
- Lab 1316 test results excluded as test method IP577 was used, see also §4.1
- Lab 1357 test result excluded as test result in counts/mL didn't match test result in ISO scale number
- Lab 1833 test results excluded as test results in counts/mL didn't match test results in ISO scale number
- Lab 1961 test results excluded as test method IP577 was used, see also §4.1










Determination of Particle Size Distribution on sample #22152 acc. to IP565, in ISO scale numbers

lab	method	≥4µm (c)	mark	z(targ)	≥6µm (c)	mark	z(targ)	≥14 µm (c)	mark	z(targ)
140		----		----	----		----	----		----
150		----		----	----		----	----		----
171		----		----	----		----	----		----
225		----		----	----		----	----		----
237		----		----	----		----	----		----
311	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
323		----		----	----		----	----		----
333	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
334	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
335		----		----	----		----	----		----
372	ISO4406 acc. to IP565	21		0.00	19	ex	-0.18	15		1.12
447		----		----	----		----	----		----
657	ISO4406	21		0.00	19		-0.18	15		1.12
823	ISO4406 acc. to IP565	21		0.00	20		2.56	15		1.12
824	ISO4406	21		0.00	20		2.56	15		1.12
862		----		----	----		----	----		----
963	ISO4406	20	ex	-2.73	19		-0.18	14		-0.92
974	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
1011		----		----	----		----	----		----
1016		----		----	----		----	----		----
1049	ISO4406 acc. to IP577	21	ex	0.00	20	ex	2.56	15	ex	1.12
1062		21		0.00	19		-0.18	14		-0.92
1064	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
1065		----		----	----		----	----		----
1097	ISO4406 acc. to IP564	20	ex	-2.73	19	ex	-0.18	13	ex	-2.96
1109		----		----	----		----	----		----
1135	ISO4406 acc. to IP565	21		0.00	19		-0.18	13		-2.96
1140		----		----	----		----	----		----
1191	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
1279		----		----	----		----	----		----
1299	ISO4406 acc. to IP577	20	ex	-2.73	19	ex	-0.18	13	ex	-2.96
1316	ISO4406 acc. to IP577	21	ex	0.00	19	ex	-0.18	14	ex	-0.92
1318	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
1320		----		----	----		----	----		----
1357	ISO4406	21		0.00	19		-0.18	14	ex	-0.92
1397		----		----	----		----	----		----
1399		----		----	----		----	----		----
1402	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
1455		21		0.00	19		-0.18	14		-0.92
1496		----		----	----		----	----		----
1538		----		----	----		----	----		----
1585	IP565	21		0.00	19		-0.18	14		-0.92
1587	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
1610		21		0.00	19		-0.18	14		-0.92
1613	IP565	21		0.00	19		-0.18	15		1.12
1634	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
1636	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
1724	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
1833	ISO4406 acc. to IP565	12	ex	-24.59	11	ex	-22.03	10	ex	-9.08
1852		21		0.00	19		-0.18	14		-0.92
1857	IP565	21		0.00	19		-0.18	15		1.12
1913		----		----	----		----	----		----
1961	ISO4406 acc. to IP577	21	ex	0.00	19	ex	-0.18	14	ex	-0.92
2130	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
6075	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
6112	ISO4406	21	C	0.00	19	C	-0.18	14	C	-0.92
6203	ISO4406 acc. to IP565	21		0.00	19		-0.18	14		-0.92
6274		----		----	----		----	----		----
6315	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
6321	ISO4406 acc. to IP565	21		0.00	19		-0.18	15		1.12
	normality	unknown			not OK			OK		
	n	31			31			31		
	outliers	0 + 7ex			0 + 7ex			0 + 7ex		
	mean (n)	21.00			19.06			14.45		
	st.dev. (n)	0.000			0.250			0.568		
	R(calc.)	0.00			0.70			1.59		
	st.dev.(IP565:13)	0.366			0.366			0.490		
	R(IP565:13)	1.03			1.03			1.37		

Lab 372 test result excluded as test result in counts/mL didn't match test result in ISO scale number  
 Lab 963 test result excluded as test result in counts/mL didn't match test result in ISO scale number  
 Lab 1049 test results excluded as test method IP577 was used and test results in counts/mL didn't match test results in ISO scale number, see also §4.1  
 Lab 1097 test results excluded as test method IP564 was used, see also §4.1  
 Lab 1299 test results excluded as test method IP577 was used, see also §4.1  
 Lab 1316 test results excluded as test method IP577 was used, see also §4.1  
 Lab 1833 test results excluded as test results in counts/mL didn't match test results in ISO scale number  
 Lab 1961 test results excluded as test method IP577 was used, see also §4.1  
 Lab 6112 first reported 15359.7, 4661.9 and 131.6 respectively





## Determination of FAME content – GCMS/HPLC on sample #22153; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
62		----		----	
140		----		----	
171	IP585	24.2		0.33	
228		----		----	
235	IP585	22.69		-0.29	
237		----		----	
253		----		----	
254		----		----	
311	IP585	23.6		0.08	
317		----		----	
323	IP585	24.7		0.54	
328	IP585	29.0		2.30	
333	IP585	24		0.25	
334	IP585	22.3		-0.45	
335	IP585	26.6		1.32	
372	IP590	23.4		0.00	
399	IP585	22.22		-0.48	
445	IP585	22.25		-0.47	
447		----		----	
460		----		----	
467		----		----	
496	IP585	18.38		-2.06	
631		----		----	
657		----		----	
823	IP585	22.8		-0.25	
824	IP585	28	C	1.89	first reported 35.78
851		----		----	
862		----		----	
974		----		----	
1011	IP585	26.3		1.19	
1016	IP585	23.89		0.20	
1023		----		----	
1040	IP599	30.8		3.04	
1049		----		----	
1062		----		----	
1064	IP585	20.67		-1.12	
1097		----		----	
1135	IP585	25.704		0.95	
1140	IP585	23.63		0.10	
1275	IP585	26.67		1.35	
1279		----		----	
1299	IP585	22.2		-0.49	
1316	IP585	21.71		-0.69	
1318	IP585	22.54		-0.35	
1399		----		----	
1402	IP585	19.3		-1.69	
1496		----		----	
1538		----		----	
1585	IP599	19.05		-1.79	
1587		----		----	
1610		----		----	
1613	IP599	24.255		0.35	
1631	IP590	20.28		-1.28	
1634	IP585	21.34		-0.85	
1724		----		----	
1833	IP590	23.80		0.17	
2130		----	W	----	test result withdrawn, reported 11.23
2133	IP585	19.21		-1.72	
6041		----		----	
6064		----		----	
6065		----		----	
6066		----		----	
6075		----		----	
6103		----		----	
6112		----		----	
6114	IP590	23.0		-0.16	
6203		----		----	
6274		----		----	
6299	IP585	26.2		1.15	
6315	IP585	20.216		-1.31	
6321		----		----	
6379	IP585	24.0		0.25	
6384	IP585	54.8	R(0.01)	12.92	





	OK	<u>only IP585</u> OK	<u>only IP590</u> not OK
normality	OK	OK	not OK
n	35	28	4
outliers	1	1	0
mean (n)	23.397	23.369	22.620
st.dev. (n)	2.8393	2.6499	1.5938
R(calc.)	7.950	7.420	4.463
st.dev.(IP585:21)	2.4314	2.4287	---
R(IP585:21)	6.808	6.800	---
compare			
R(IP590:10)	5.540	---	5.383



Determination of FAME content – FTIR on sample #22153; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
62	D7797	33.5		0.97	
140		----		----	
171		----		----	
228	D7797	32.23		0.42	
235		----		----	
237		----		----	
253	IP583	27.53		-1.60	
254	D7797	33.73		1.07	
311		----		----	
317		----		----	
323		----		----	
328		----		----	
333		----		----	
334		----		----	
335		----		----	
372		----		----	
399		----		----	
445		----		----	
447		----		----	
460		----		----	
467	IP583	33.40		0.93	
496		----		----	
631	D7797	18.61	R(0.01)	-5.43	
657		----		----	
823		----		----	
824		----		----	
851	D7797	32.39		0.49	
862		----		----	
974	IP583	30.1		-0.49	
1011		----		----	
1016		----		----	
1023	D7797	28.98		-0.97	
1040		----		----	
1049	IP583	32.11		0.37	
1062	IP583	31		-0.11	
1064		----		----	
1097	IP583	31.25		0.00	
1135		----		----	
1140		----		----	
1275		----		----	
1279		----		----	
1299	IP583	33.2		0.84	
1316		----		----	
1318		----		----	
1399		----		----	
1402	IP583	31.1		-0.06	
1496		----		----	
1538	D7797	29.98		-0.54	
1585		----		----	
1587	IP583	31.17		-0.03	
1610		----		----	
1613		----		----	
1631		----		----	
1634		----		----	
1724		----		----	
1833		----		----	
2130		----		----	
2133		----		----	
6041		----		----	
6064		----		----	
6065		----		----	
6066		----		----	
6075		----		----	
6103	D7797	30.145		-0.47	
6112	IP583	29.15		-0.90	
6114		----		----	
6203	D7797	32.54		0.56	
6274		----		----	
6299		----		----	
6315		----		----	
6321	IP583	30.15		-0.47	
6379		----		----	
6384		----		----	



normality	OK
n	19
outliers	1
mean (n)	31.245
st.dev. (n)	1.7185
R(calc.)	4.812
st.dev.(IP583:15)	2.3282
R(IP583:15)	6.519



Determination of FAME content – GCMS/HPLC on sample #22154; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
62		----		----	
140		----		----	
171	IP585	59.0		0.51	
228		----		----	
235	IP585	53.23		-0.55	
237		----		----	
253		----		----	
254		----		----	
311	IP585	56.4		0.03	
317		----		----	
323	IP585	60.2		0.73	
328	IP585	62.5		1.15	
333	IP585	57		0.14	
334	IP585	55.3		-0.17	
335	IP585	65.5		1.70	
372	IP590	56.0		-0.04	
399	IP585	40.01		-2.97	
445		----		----	
447		----		----	
460		----		----	
467		----		----	
496	IP585	46.45		-1.79	
631		----		----	
657		----		----	
823	IP585	56.0		-0.04	
824	IP585	66	C	1.79	first reported 64.74
851		----		----	
862		----		----	
974		----		----	
1011	IP585	56.7		0.09	
1016	IP585	62.30		1.11	
1023		----		----	
1040	IP599	75.7		3.57	
1049		----		----	
1062		----		----	
1064	IP585	50.06		-1.13	
1097		----		----	
1135	IP585	56.92		0.13	
1140	IP585	62.26		1.10	
1275	IP585	66.12		1.81	
1279		----		----	
1299	IP585	51.2		-0.92	
1316		----		----	
1318	IP585	52.43		-0.70	
1399		----		----	
1402	IP585	45.1		-2.04	
1496		----		----	
1538		----		----	
1585	IP599	50.1		-1.12	
1587		----		----	
1610		----		----	
1613	IP599	61.576		0.98	
1631	IP590	46.43		-1.80	
1634	IP585	53.3		-0.54	
1724	IP590	51.69		-0.83	
1833	IP590	49.57		-1.22	
2130		----	W	----	test result withdrawn, reported 22.72
2133	IP585	50.98		-0.96	
6041		----		----	
6064		----		----	
6065		----		----	
6066		----		----	
6075		----		----	
6103		----		----	
6112		----		----	
6114	IP590	58.5		0.42	
6203		----		----	
6274		----		----	
6299	IP585	61.4		0.95	
6315	IP585	59.113		0.53	
6321		----		----	
6379	IP585	56.9		0.12	
6384		----		----	



		<u>only IP585</u>	<u>only IP590</u>
normality	OK	OK	OK
n	34	25	5
outliers	0	0	29
mean (n)	56.234	56.219	52.438
st.dev. (n)	7.1136	6.6395	4.8558
R(calc.)	19.918	18.591	13.596
st.dev.(IP585:21)	5.4558	5.4545	---
R(IP585:21)	15.276	15.273	---
compare			
R(IP590:10)	12.179	---	11.412



Determination of FAME content – FTIR on sample #22154; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
62	D7797	65.5		0.66	
140		----		----	
171		----		----	
228	D7797	63.22		-0.12	
235		----		----	
237		----		----	
253	IP583	56.82		-2.33	
254	D7797	70.84		2.50	
311		----		----	
317		----		----	
323		----		----	
328		----		----	
333		----		----	
334		----		----	
335		----		----	
372		----		----	
399		----		----	
445		----		----	
447		----		----	
460		----		----	
467	IP583	65.09		0.52	
496		----		----	
631	D7797	35.20	R(0.01)	-9.78	
657		----		----	
823		----		----	
824		----		----	
851	D7797	62.01		-0.54	
862		----		----	
974	IP583	61.8		-0.61	
1011		----		----	
1016		----		----	
1023	D7797	58.24		-1.84	
1040		----		----	
1049	IP583	68.0		1.52	
1062	IP583	62.7		-0.30	
1064		----		----	
1097	IP583	62.71		-0.30	
1135		----		----	
1140		----		----	
1275		----		----	
1279		----		----	
1299	IP583	67.8		1.46	
1316		----		----	
1318		----		----	
1399		----		----	
1402	IP583	53.4		-3.51	
1496		----		----	
1538	D7797	61.21		-0.82	
1585		----		----	
1587	IP583	63.76		0.06	
1610		----		----	
1613		----		----	
1631		----		----	
1634		----		----	
1724		----		----	
1833		----		----	
2130		----		----	
2133		----		----	
6041		----		----	
6064		----		----	
6065		----		----	
6066		----		----	
6075		----		----	
6103	D7797	67.38		1.31	
6112	IP583	63.46		-0.04	
6114		----		----	
6203	D7797	70.58		2.41	
6274		----		----	
6299		----		----	
6315		----		----	
6321	IP583	63.44		-0.05	
6379		----		----	
6384		----		----	

normality	OK
n	19
outliers	1
mean (n)	63.577
st.dev. (n)	4.4114
R(calc.)	12.352
st.dev.(IP583:15)	2.9017
R(IP583:15)	8.125



Determination of Copper as Cu on sample #22155; results in µg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
120		----		----	
140		----		----	
150		----		----	
159		----		----	
171		----		----	
175		----		----	
177		----		----	
225		----		----	
228		----		----	
235		----		----	
237		----		----	
253		----		----	
254		----		----	
256		----		----	
311		----		----	
323		----		----	
334		----		----	
335		----		----	
372		----		----	
391		----		----	
398		----		----	
399		----		----	
440		----		----	
445		----		----	
447		----		----	
467	INH-7111	616		----	
496		----		----	
631		----		----	
657		----		----	
671		----		----	
823		----		----	
824		----		----	
851		----		----	
862		----		----	
869		----		----	
922		----		----	
962		----		----	
963		----		----	
974		----		----	
1011		----		----	
1016		----		----	
1039		----		----	
1040		----		----	
1049		----		----	
1064		----		----	
1097		----		----	
1105		----		----	
1109		----		----	
1146	D5185	713.5		----	
1191		----		----	
1237		----		----	
1279		----		----	
1299		----		----	
1316	D7111	>500		----	
1320		----		----	
1357		----		----	
1372		----		----	
1399		----		----	
1412		----		----	
1417		----		----	
1455		----		----	
1496		----		----	
1538		----		----	
1585		----		----	
1586		----		----	
1587		----		----	
1610		----		----	
1613		----		----	
1616		----		----	
1631		----		----	
1634		----		----	
1720		----		----	
1724		----		----	
1730		----		----	



lab	method	value	mark	z(targ)	remarks
1833		----		----	
1852		----		----	
1854		----		----	
1913		----		----	
1961		----		----	
2130		----		----	
6028		----		----	
6041		----		----	
6054		----		----	
6075		----		----	
6079	D7111	519		----	
6114		----		----	
6203		----		----	
6274		----		----	
6315		----		----	
6321		----		----	
6324		----		----	
6344		----		----	
	n	4			
	mean (n)	>500			

Determination of JFTOT at 260 °C on sample #22155; Visual tube rating (VTR), Interferometric tube rating (ITR) in nm and Ellipsometric tube rating (ETR) in nm, Delta P in mmHg, Evaluation Pass/Fail





lab	method	VTR	ITR	ETR	Delta P	Time to 25 mmHg (min)	Pumped Vol. (mL)	Heater Temp. (°C)	Pass/Fail
52		----	----	----	----	----	----	----	----
120	D3241	3	39.5	----	4.1 C	16	500	----	----
140		----	----	----	----	----	----	----	----
150	D3241	4	----	----	251	29 C	500	260	----
159	D3241	C&B	----	----	2.0 C	----	550	260	Pass
171	D3241	>4	----	----	251	18 C	450	260	Fail
175	D3241	3	----	----	250	12	440	260	Fail
177		----	----	----	----	----	----	----	----
225		----	----	----	----	----	----	----	----
228	D3241	2	----	----	2.8	16	450	260	Pass
235	D3241	4	----	----	----	30 C	500	260	Fail
237	D3241	<4	----	----	>25 C	----	460	260	FAIL
253	D3241	4	----	----	250.1	11.87	450	260	Fail
254	D3241	<4	----	----	248	18	450	260	Fail
256	D3241	>3	----	----	251.3	15.15	440	260	Fail
311	D3241	4	----	----	280	----	460	260	FAIL
323	D3241	4	----	----	4.0	150	494	260	fail
334	D3241	4	67	----	250	15	485	260	----
335	D3241	4	52.1	----	280	14	510	260	----
372	D3241	4A	----	45	250	16	449	260	Fail
391		----	----	----	----	----	----	----	----
398	D3241	4	----	----	260	15	----	260	----
399	D3241	<4	----	61.73	280.1	26 C	450	260	Fail
440		----	----	----	----	----	----	----	----
445	IP323/D3241	3A	----	----	280	11.58	400	260	Fail
447		----	----	----	----	----	----	----	----
467	D3241	----	----	93.56	252.7	19.69	453	260.0	fail
496		----	----	----	----	----	----	----	----
631	D3241	<4	----	----	<1.0	----	500	260	Fail
657	D3241	4	----	195.84	250.4 C	13.52	440	260	Fail
671	D3241	2A	----	55.52	7.0 C	32	450	260	----
823	D3241	3	----	----	6 C	30	450	260	pass
824	D3241	<4	----	----	280.1	23	510	260	fail
851	D3241	<4	72.3	----	280.0	7.59	450	260	Fail
862		----	----	----	----	----	----	----	----
869		----	----	----	----	----	----	----	----
922		----	----	----	>25	19	----	260	Fail
962		----	----	----	----	----	----	----	----
963	D3241	>4	----	----	----	12.56 C	450	260	Fail
974	D3241	4	----	----	>250	17	450	260	Fail
1011	D3241	4	----	----	250	18	450	260	Fail
1016	D3241	4	----	----	251.5	----	433	260	----
1039		----	76.1	----	253.5	16.63	459	260	fail
1040		----	42.1	----	252.7	----	----	----	Fail
1049	D3241	3A	----	48.32	252	17	438	260	Fail
1064	D3241	A [white deposit]	----	200.82	250.3	19.5	443	260	Fail
1097	D3241	<4A	----	----	251.2	13 min 17 s	457	260	Fail
1105	D3241	> 4	85.6	----	251.1	26.13	434.18	260	FAIL
1109		----	----	----	----	----	----	----	----
1146	D3241	3	77.2	----	250	13.15	456	260	fail
1191	D3241	4	80.8	----	252.2	14.6	470	260	pass
1237	D3241	3	----	----	280.1	19	450	260	Fail
1279		----	----	----	----	----	----	----	----
1299	D3241	4	105	----	280	19	----	260	FAIL
1316		----	----	----	----	----	----	----	----
1320	D3241	4	----	----	250.4	13.95	423	260	Fail
1357	D3241	----	----	----	<25 C	----	460	260	Pass
1372		<1	----	----	280	10.53	460	260	Fail
1399		----	----	----	----	----	----	----	----
1412	D3241	4	----	----	0	----	450	260	Fail
1417	IP323	>3	----	----	280.1	----	450	260	Fail
1455	D3241	Abnormal	----	----	252.0	12:26	417.4	260	Fail
1496		----	----	----	----	----	----	----	----
1538	D3241	<4	----	----	280.1	29	450	260	----
1585	D3241/IP323	4	----	----	253.1	12.97	455	260	Fail
1586	IP323	1	----	----	----	W	510	260	Pass
1587	D3241	4	----	----	>25	13min 34sec C	446	260	fail
1610	IP323	<2	----	----	187.5	121.1	443	260	Fail
1613	D3241	4	----	----	280.2	----	450	260	Fail
1616	D3241	3A	----	----	280	15.9	450	260	Fail
1631		----	----	----	----	----	----	----	----
1634	D3241	4	----	----	251	----	500	260	Fail

lab	method	VTR	ITR	ETR	Delta P	Time to 25 mmHg (min)	Pumped Vol. (mL)	Heater Temp. (°C)	Pass/Fail
1720		4	----	----	280	14:54	----	260	----
1724	D3241	1	----	----	280.1 C	15.8 C	510	260	pass
1730	D3241	4	130	----	>25	----	450	260	FAIL
1833	D3241	4	----	----	250.5	----	----	260	----
1852	D3241/IP323	4	----	194.16	251.1	12:16	463.9	260	Fail
1854	D3241	<4	----	----	251	18	450	260	fail
1913		----	----	----	----	----	----	----	----
1961	D3241	3A	----	----	280.1	27	450	260	Fail
2130		----	----	----	----	----	----	----	----
6028		----	----	----	----	----	----	----	----
6041		----	----	----	----	----	----	----	----
6054		----	----	----	----	----	----	----	----
6075		----	----	----	----	----	----	----	----
6079	D3241	----	45.24	304.11	253.7	15.2	451	260	Fail
6114	D3241	4A	----	----	0.8	----	450	260	Fail
6203	D3241	>4	----	----	100.2	10	500	260.0	fail
6274		----	----	----	----	----	----	----	----
6315		----	----	50.5	251.8	17	700	260	Fail
6321	IP323	4	72.4	----	280.1	10.50	440	260	Fail
6324	D3241	<2.0	----	----	>25 C	<20	455	260	Fail C
6344	D3241	3	----	----	-1	15	450	260	----
n		44	3	5	55				
Mean (n)		>2	>85	>85	>25				
Pass									7
Fail									51

Lab 120 first reported 0  
 Lab150 first reported 150  
 Lab 159 first reported 0.1  
 Lab 171 first reported 150  
 Lab 235 first reported 40  
 Lab 237 first reported 0.2  
 Lab 399 first reported 150  
 Lab 657 first rported 5.5

Lab 671 first reported 100.0  
 Lab 823 first reported 4  
 Lab 963 first reported 0  
 Lab 1357 first reported <2 and 1 respectively  
 Lab 1586 test result withdrawn, reported 0 and 0 respectively  
 Lab 1587 first reported 1334  
 Lab 1724 first reported 0 and 0 respectively  
 Lab 6324 first reported 1 and Pass respectively



Pass according to specification AFQRJOS is when VTR is less than 3 (no peacock or abnormal color), ITR/ETR is maximum 85 and Delta P is maximum 25.



Determination of Particulate Contamination on sample #22156; Particulate matter in mg/L, Filtration time in minutes and Volume of fuel filtered in mL

lab	method	part. matter	mark	z(targ)	filtr. time	vol. filtered	remarks
140		----		----	----	----	
150		----		----	----	----	
169		----		----	----	----	
171	D5452	0.30		-0.34	----	1000	
334	D5452	0.21		-1.08	----	3.9	
335	D5452	0.25		-0.75	----	4000	
365	IP423	0.15		-1.57	8	4000	
372	D5452	0.308		-0.28	14	3900	
398	D5452	0.342		0.00	----	3800	
399	D5452	0.39		0.40	15	3785	
440	D5452	0.45		0.89	----	4.0	
496	D5452	0.399		0.47	----	----	
631	D5452	0.25		-0.75	7	4000	
657	D5452	0.3709		0.24	35	3828	
851	D5452	0.30		-0.34	5	4000	
962	D5452	0.42		0.64	25	4000	
963	D5452	0.32		-0.18	15	4000	
974	D5452	0.47		1.05	5	3800	
1011	D5452	0.28		-0.50	----	4000	
1039	D5452	0.35		0.07	4	4000	
1049	D5452	0.475		1.09	----	----	
1059	D5452	0.25		-0.75	79	3990	
1064	D5452	0.15		-1.57	3.6	3970	
1109		----		----	----	----	
1135	D5452	0.26		-0.67	2.87	1959	
1279		----		----	----	----	
1299	D5452	0.05		-2.39	----	3790	
1316	D5452	0.35		0.07	----	900	
1320	D5452	0.28		-0.50	----	3945	
1357	D5452	0.50	C	1.30	----	4000	first reported 0.9
1399		----		----	----	----	
1414		----		----	----	----	
1455	D5452	0.33		-0.09	4000	40	
1538	D5452	<1.0	C	----	----	4000	first reported 0.7
1586	D5452	0.5		1.30	----	4000	
1587	D5452	0.62		2.28	----	3700	
1610		----		----	----	----	
1616	D5452	0.3725		0.25	2.50	4000	
1631	D5452	0.238		-0.85	----	----	
1634	D5452	0.30		-0.34	13.5	3900	
1724	IP423	0.6		2.12	----	----	
1833	IP423	0.3		-0.34	4	3800	
1852	D5452	0.30		-0.34	3.44	4000	
1913		----		----	----	----	
1961	D5452	0.31		-0.26	----	4000	
2130	D5452	0.25		-0.75	12.40.25	4000	
6041		----		----	----	----	
6075	D5452	0.39		0.40	12	3906.6	
6103	D5452	0.475		1.09	21	4000	
6203	D5452	0.375		0.27	8.4	4000	
6274		----		----	----	----	
6315	D5452	0.38		0.32	12	3980	
6321	IP423	0.31		-0.26	----	3900	
6421	D5452	0.42		0.64	----	----	

normality OK  
n 42  
outliers 0  
mean (n) 0.342  
st.dev. (n) 0.1137  
R(calc.) 0.318  
st.dev.(D5452:20) 0.1219  
R(D5452:20) 0.341



**APPENDIX 2** z-scores of Distillation at 760 mmHg

lab	IBP	10% rec	50% rec	90% rec	FBP
52	-1.69	-0.18	-0.33	-0.40	0.07
62	0.41	0.72	0.32	0.32	0.35
120	0.44	-0.33	-0.70	0.56	-0.64
140	0.27	0.12	-0.33	-0.80	0.23
150	-1.96	-1.00	0.23	2.00	-0.76
159	0.24	0.72	0.98	0.56	0.78
169	-0.33	0.50	0.14	-0.16	0.11
171	-0.17	0.12	0.14	-0.40	0.35
175	0.99	0.20	2.19	3.04	1.69
177	-1.69	-0.40	-0.33	-0.80	-0.56
215	-0.50	-0.25	1.26	0.08	-0.76
221	1.70	1.47	0.98	0.56	0.59
224	----	----	----	----	----
225	0.17	-0.25	0.32	0.88	0.43
228	0.51	1.24	0.32	-0.72	0.82
235	-1.08	-1.30	-1.45	-0.96	-1.35
237	1.19	1.99	0.32	0.08	-0.36
238	----	----	----	----	----
253	-0.84	-1.75	0.32	-0.72	-0.76
254	0.17	-0.25	-0.61	0.88	0.43
256	0.17	-0.25	0.32	0.08	0.03
258	1.66	0.57	-0.24	-1.77	-0.01
273	0.24	0.50	1.16	0.48	1.37
311	0.55	0.05	0.42	0.72	0.19
317	0.92	-0.10	0.32	0.64	0.98
323	0.00	0.65	0.04	0.24	-0.28
328	0.00	-0.25	-0.24	-0.64	-0.32
333	-1.32	-0.77	-1.17	-1.36	-0.76
334	-0.50	0.05	0.23	0.32	-0.28
335	0.34	-0.03	0.79	1.12	-0.40
365	-0.57	-0.77	-0.33	0.56	-0.05
372	0.14	0.05	0.04	-0.08	-0.01
391	----	----	----	----	----
396	----	----	----	----	----
399	----	----	----	----	----
440	0.11	0.27	0.42	1.28	0.51
445	-0.81	-1.30	-0.98	-0.16	-0.99
447	0.38	-0.10	-0.24	0.72	-0.08
460	----	----	----	----	----
467	0.88	0.20	0.23	-0.08	-0.20
480	0.11	0.83	0.32	-0.24	-0.42
496	1.09	1.09	-0.24	0.40	0.27
603	0.65	-0.85	-1.64	-0.72	-1.15
608	-1.55	-0.03	0.79	0.32	-0.24
631	0.17	-1.75	-1.54	-1.52	-1.35
657	1.32	0.35	0.88	0.24	0.43
736	-0.17	-0.62	-0.61	0.08	-1.15
823	0.34	0.65	0.98	0.72	0.86
824	-0.17	1.09	0.42	0.64	0.35
851	-0.64	-0.03	-0.42	-0.40	-0.44
854	----	----	----	----	----
862	----	----	----	----	----
869	----	----	----	----	----
904	-0.13	-0.70	0.70	0.40	0.23
914	-0.06	-0.33	-0.14	0.32	1.69
962	0.11	-0.70	-1.26	-1.52	-0.76
963	-0.94	-0.03	-0.33	-1.12	-0.48
970	-0.17	-1.00	-0.61	-1.52	-0.36
974	-0.50	-0.92	-0.61	-0.64	-0.36
994	-0.17	-0.25	-0.61	-0.72	0.82
995	-0.33	0.12	-0.14	1.68	1.02
996	----	----	----	----	----
997	-0.50	0.50	0.32	1.68	1.02
1011	-0.20	0.20	-0.14	1.04	0.19
1016	----	----	----	----	----
1019	----	----	----	----	----
1023	0.41	0.12	-0.89	-0.24	-0.08
1039	0.58	0.65	0.60	0.16	-0.44
1040	-0.54	-1.00	-0.33	0.32	0.11
1049	0.71	0.72	0.70	1.12	0.23
1059	0.14	0.27	0.14	-0.24	0.27
1062	0.34	-0.18	0.42	0.16	0.31

lab	IBP	10% rec	50% rec	90% rec	FBP
1064	0.78	0.35	0.04	0.40	0.74
1065	-1.32	-2.79	0.04	0.72	8.87
1082	-0.03	0.79	0.70	0.72	0.51
1097	0.71	-0.55	-0.42	0.40	-0.12
1105	-0.64	1.02	0.51	0.00	-0.24
1109	----	----	----	----	----
1121	-1.62	0.05	-0.52	-0.64	-0.36
1126	0.65	0.42	-0.14	-1.20	-0.48
1140	-1.42	-0.55	-0.42	-0.40	0.23
1182	1.66	1.32	1.44	1.60	0.55
1191	-1.38	0.87	0.14	0.00	-0.08
1205	0.51	0.35	0.60	-0.32	-0.20
1237	-0.23	0.05	-0.24	-0.72	-0.24
1275	-0.37	-0.40	-0.61	-0.40	-0.44
1279	----	----	----	----	----
1299	0.24	0.42	0.70	1.12	0.63
1316	-0.03	0.35	0.51	0.16	0.27
1318	1.49	0.57	0.14	0.40	0.35
1320	-1.11	-0.62	-0.42	-1.12	-0.36
1357	0.27	-0.40	-1.45	0.00	0.35
1372	0.48	-1.15	-1.45	-1.36	0.39
1397	0.71	0.72	0.04	-0.80	-0.60
1399	----	----	----	----	----
1412	-0.17	-0.25	-0.14	0.08	0.03
1417	0.61	0.35	1.54	2.24	0.03
1444	----	----	----	----	----
1455	0.17	0.65	0.42	0.08	0.11
1496	----	----	----	----	----
1538	----	----	----	----	----
1575	----	----	----	----	----
1585	-0.81	-0.25	-0.52	-0.88	-0.56
1586	0.27	0.05	0.23	0.24	-0.01
1587	-0.03	0.05	-0.05	-0.48	-0.24
1610	0.65	0.87	0.60	0.56	0.43
1613	0.78	0.65	0.51	-0.16	0.39
1616	-0.57	-0.70	-1.36	-1.28	-0.32
1631	-1.15	-0.70	-0.42	-0.16	0.19
1634	0.14	0.12	-0.05	-0.08	0.15
1636	-0.27	0.57	0.79	1.04	-0.44
1669	-1.45	-1.22	-0.70	-0.64	0.07
1683	0.24	-0.85	-1.17	-0.72	-1.11
1688	----	-1.00	0.32	-1.52	-0.36
1715	1.29	-0.40	-0.61	0.88	0.31
1720	-0.47	0.20	0.23	0.00	2.36
1724	-0.54	-0.70	-1.08	-1.52	-0.16
1730	----	----	----	----	----
1757	0.27	0.05	-0.80	-0.80	-0.99
1776	-0.54	-1.00	-0.52	-0.08	-0.44
1833	-1.21	-0.85	-0.98	-0.72	-0.28
1852	0.14	0.12	0.42	0.08	0.31
1913	----	----	----	----	----
1944	-0.30	0.12	0.23	0.08	0.23
1961	----	----	----	----	----
2130	0.34	0.20	0.60	0.88	0.23
2133	0.27	0.42	0.70	0.56	0.47
6028	----	----	----	----	----
6041	----	----	----	----	----
6054	----	----	----	----	----
6075	-0.98	0.42	0.98	0.88	-0.01
6114	0.99	0.35	-0.05	0.24	-0.01
6135	----	----	----	----	----
6142	-0.60	-0.77	-0.98	-1.04	-0.91
6203	-0.64	-0.25	0.14	0.24	-0.12
6244	1.26	0.79	0.51	-0.40	3.66
6266	0.50	0.84	-0.05	-0.56	-0.22
6274	----	----	----	----	----
6312	----	----	----	----	----
6315	-0.37	0.05	0.04	0.16	0.19
6321	0.38	0.20	0.14	0.08	-0.05
6324	0.68	-0.62	0.32	0.08	0.47
6332	1.36	0.87	0.79	0.88	0.23
6344	0.31	0.20	0.32	-0.08	0.15
6346	----	----	----	----	----
6364	0.65	-0.85	-1.36	-1.04	-0.24

lab	IBP	10% rec	50% rec	90% rec	FBP
6384	-0.44	0.35	0.42	1.36	0.59
6421	1.19	1.24	1.26	0.08	0.82
6438	-1.28	-1.22	-1.26	-0.32	-0.79
6479	----	----	----	----	----
6487	----	----	----	----	----



**APPENDIX 3 Equipment used in Particle Size distribution**

lab	Equipment	Test method based on equipment	Test method reported	Calibration method reported
140				
150				
171	Stanhope-Seta	IP565	IP565	
225				
237				
311	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
323				
333	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
334	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
335				
372	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
447				
657	Stanhope-Seta	IP565	IP565	ISO4406
823	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
824	Stanhope-Seta	IP565	IP565	ISO4406
862				
963	Stanhope-Seta	IP565	IP565	ISO4406
974	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1011	Stanhope-Seta	IP565	IP565	
1016	Stanhope-Seta	IP565	IP565	
1049	Pamas	IP577	IP577	ISO4406 acc. to IP577
1062	Stanhope-Seta	IP565	IP565	
1064	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1065	Stanhope-Seta	IP565	IP565	
1097	Parker Hannifin	IP564	IP564	ISO4406 acc. to IP564
1109				
1135	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1140	Stanhope-Seta	IP565	IP565	
1191				ISO4406 acc. to IP565
1279				
1299	Pamas	IP577	IP577	ISO4406 acc. to IP577
1316	Pamas	IP577	IP577	ISO4406 acc. to IP577
1318	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1320				
1357	Parker Hannifin	IP564	IP565	ISO4406
1397				
1399				
1402	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1455	Stanhope-Seta	IP565	IP565	
1496				
1538	Stanhope-Seta	IP565	IP565	
1585	Stanhope-Seta	IP565	IP565	IP565
1587	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1610	Stanhope-Seta	IP565	IP565	
1613	Stanhope-Seta	IP565	IP565	IP565
1634	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1636	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1724		IP565	IP565	ISO4406 acc. to IP565
1833	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
1852	Stanhope-Seta	IP565	IP565	
1857	Stanhope-Seta	IP565	IP565	IP565
1913				
1961	Pamas	IP577	IP577	ISO4406 acc. to IP577
2130	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
6075	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
6112	Stanhope-Seta	IP565	IP565	ISO4406
6203	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
6274				
6315	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565
6321	Stanhope-Seta	IP565	IP565	ISO4406 acc. to IP565

**APPENDIX 4 z-scores of Particle Size Distribution, counts/mL**

lab	$\geq 4 \mu\text{m}$	$\geq 6 \mu\text{m}$	$\geq 14 \mu\text{m}$	$\geq 21 \mu\text{m}$	$\geq 25 \mu\text{m}$	$\geq 30 \mu\text{m}$
140	----	----	----	----	----	----
150	----	----	----	----	----	----
171	-1.76	-1.52	-0.01	0.41	0.40	1.46
225	----	----	----	----	----	----
237	----	----	----	----	----	----
311	-1.47	-0.81	-1.26	-1.27	-1.69	----
323	----	----	----	----	----	----
333	1.52	0.82	0.65	0.10	-0.65	-0.16
334	-0.22	-0.31	0.68	0.86	0.74	0.65
335	----	----	----	----	----	----
372	3.02	2.12	1.97	3.23	3.59	2.92
447	----	----	----	----	----	----
657	-3.55	0.17	0.98	2.53	2.86	1.21
823	4.08	2.67	2.25	3.19	3.39	3.97
824	4.84	1.76	1.86	2.91	3.04	2.67
862	----	----	----	----	----	----
963	-0.23	-0.37	-0.67	-0.46	-0.44	-0.97
974	-0.71	-0.76	-1.11	-0.81	-0.65	-0.16
1011	-1.76	0.16	-0.23	0.56	0.40	0.65
1016	0.02	-0.54	-0.96	-1.48	-1.58	-1.06
1049	221.50	141.27	71.62	67.08	55.30	34.93
1062	-1.12	-1.05	-1.15	-1.92	-1.97	-1.62
1064	-0.21	0.40	-1.45	-2.57	-2.21	-1.62
1065	-3.27	-1.06	-0.71	-0.13	-0.23	-0.73
1097	-9.09	-4.23	-3.08	0.33	4.32	12.80
1109	----	----	----	----	----	----
1135	-3.10	-0.04	-2.39	-1.57	-1.69	-1.79
1140	2.03	0.96	0.09	-0.67	-1.20	-1.06
1191	1.02	-0.42	0.48	-0.58	-0.68	-1.06
1279	----	----	----	----	----	----
1299	-8.70	-5.39	-3.59	-2.06	-1.45	-1.06
1316	-5.98	-2.41	-0.62	0.72	2.41	2.43
1318	0.20	-0.38	1.05	1.18	1.51	1.46
1320	----	----	----	----	----	----
1357	-0.16	-1.12	0.18	-1.39	-1.17	-0.73
1397	----	----	----	----	----	----
1399	----	----	----	----	----	----
1402	-1.70	-0.85	-0.70	0.59	1.54	2.67
1455	-1.85	0.54	-0.32	-0.81	-1.34	-1.79
1496	----	----	----	----	----	----
1538	-1.95	----	----	----	----	----
1585	1.41	-0.20	-0.52	-0.93	-0.65	-0.57
1587	-2.24	-2.04	0.89	0.74	0.78	0.81
1610	4.06	1.26	-0.19	-0.60	-1.03	-1.22
1613	1.18	1.59	0.28	1.30	1.47	0.97
1634	-0.12	1.00	-0.38	0.41	1.09	0.65
1636	-3.40	-0.51	-0.21	-0.70	-0.99	-1.46
1724	0.35	-0.29	1.21	0.66	1.54	1.62
1833	1.54	0.00	1.20	2.84	2.83	2.59
1852	2.42	1.14	-0.04	0.21	0.01	-0.08
1857	2.30	1.20	0.40	----	----	----
1913	----	----	----	----	----	----
1961	-6.59	-2.40	-1.30	-0.92	-1.13	-1.22
2130	-0.31	-2.09	0.33	0.10	0.15	-0.25
6075	-0.13	-0.56	-0.88	-0.73	-1.06	-0.81
6112	1.16	0.74	-0.74	-0.77	-1.38	-1.62
6203	1.68	0.39	-0.26	-0.73	-0.85	-0.89
6274	----	----	----	----	----	----
6315	-2.22	0.47	0.43	0.41	0.05	-0.97
6321	-0.03	-0.32	0.61	-1.25	-1.13	-1.06

**APPENDIX 5****Number of participants per country**

1 lab in AUSTRALIA	1 lab in MARTINIQUE
3 labs in AZERBAIJAN	1 lab in MAURITIUS
5 labs in BELGIUM	1 lab in MOROCCO
1 lab in BULGARIA	1 lab in MOZAMBIQUE
2 labs in CANADA	8 labs in NETHERLANDS
8 labs in CHINA, People's Republic	2 labs in NIGERIA
3 labs in COLOMBIA	1 lab in NORTH MACEDONIA, Republic of
1 lab in COSTA RICA	2 labs in NORWAY
2 labs in COTE D'IVOIRE	2 labs in OMAN
1 lab in CROATIA	2 labs in PAKISTAN
2 labs in DENMARK	1 lab in PHILIPPINES
1 lab in DJIBOUTI	3 labs in POLAND
1 lab in EGYPT	2 labs in PORTUGAL
1 lab in ESTONIA	1 lab in QATAR
3 labs in FINLAND	2 labs in ROMANIA
6 labs in FRANCE	2 labs in RUSSIAN FEDERATION
3 labs in GEORGIA	5 labs in SAUDI ARABIA
6 labs in GERMANY	1 lab in SENEGAL
4 labs in GREECE	1 lab in SINGAPORE
1 lab in GUINEA REPUBLIC	2 labs in SLOVAKIA
1 lab in GUAM	2 labs in SLOVENIA
1 lab in HONG KONG	2 labs in SOUTH AFRICA
1 lab in HUNGARY	4 labs in SPAIN
1 lab in INDIA	1 lab in SUDAN
2 labs in IRELAND	5 labs in SWEDEN
1 lab in ISRAEL	3 labs in TANZANIA
4 labs in ITALY	1 lab in TOGO
1 lab in JORDAN	2 labs in TUNISIA
1 lab in KAZAKHSTAN	5 labs in TURKEY
3 labs in KENYA	1 lab in TURKMENISTAN
2 labs in KOREA, Republic of	3 labs in UNITED ARAB EMIRATES
2 labs in MALAYSIA	14 labs in UNITED KINGDOM
1 lab in MALTA	8 labs in UNITED STATES OF AMERICA

## APPENDIX 6

### Abbreviations

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01) or R(1)	= outlier in Rosner's outlier test
R(0.05) or R(5)	= straggler in Rosner's outlier test
E	= calculation difference between reported test result and result calculated by iis
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
f+?	= possibly a false positive test result?
f-?	= possibly a false negative test result?
SDS	= Safety Data Sheet

### Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 ISO5725:86
- 3 ISO5725 parts 1-6:94
- 4 ISO13528:05
- 5 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 6 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 7 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 8 J.N. Miller, Analyst, 118, 455, (1993)
- 9 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 10 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 11 W. Horwitz and R. Albert, J. AOAC Int, 79.3, 589-621, (1996)
- 12 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)