



Institute for
Interlaboratory Studies

Results of Proficiency Test AdBlue (32% Urea solution) May 2023

Organized by: Institute for Interlaboratory Studies
Spijkenisse, the Netherlands

Author: Mrs. E.R. Montenij-Bos
Correctors: ing. R.J. Starink & ing. G.A. Oosterlaken-Buijs
Approved by: ing. A.S. Noordman-de Neef

Report: iis23G05

August 2023

CONTENTS

1	INTRODUCTION	3
2	SET UP	3
2.1	QUALITY SYSTEM.....	3
2.2	PROTOCOL.....	3
2.3	CONFIDENTIALITY STATEMENT	3
2.4	SAMPLES	4
2.5	STABILITY OF THE SAMPLES.....	4
2.6	ANALYZES	4
3	RESULTS	5
3.1	STATISTICS	5
3.2	GRAPHICS	6
3.3	Z-SCORES	6
4	EVALUATION	7
4.1	EVALUATION PER TEST	7
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	8
4.3	COMPARISON OF THE PROFICIENCY TEST OF MAY 2023 WITH PREVIOUS PTS	9

Appendices:

1.	Data, statistical and graphic results	11
2.	Summary of reported elements	20
3.	Number of participants per country.....	21
4.	Abbreviations and literature	22

1 INTRODUCTION

Since 2017 the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the analysis of AdBlue 32% Urea solution based on the latest version of ISO22241 part 1 every year. During the annual proficiency testing program 2022/2023 it was decided to continue the round robin for the analysis of AdBlue 32% Urea solution.

In this interlaboratory study 35 laboratories in 23 countries registered for participation, see appendix 3 for the number of participants per country. In this report the results of the AdBlue 32% Urea solution proficiency test are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory.

It was decided to send one sample AdBlue 32% Urea solution in a 1-liter HDPE wide-neck bottle labelled #23077.

The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 QUALITY SYSTEM

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, has implemented a quality system based on ISO/IEC17043:2010. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

A batch of approximately 50 liters of AdBlue 32% Urea solution was obtained from a local supplier. After homogenization 45 wide-neck HDPE bottles of 1 L were filled and labelled #23077.

The homogeneity of the subsamples was checked by determination of Density at 20 °C in accordance with ASTM D4052 on 4 stratified randomly selected subsamples.

	Density at 20 °C in kg/m ³
sample #23077-1	1089.00
sample #23077-2	1088.99
sample #23077-3	1088.98
sample #23077-4	1089.01

Table 1: homogeneity test results of subsamples #23077

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 20 °C in kg/m ³
r (observed)	0.04
reference test method	ISO12185:96
0.3 x R (reference test method)	0.15

Table 2: evaluation of the repeatability of subsamples #23077

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

To each of the participating laboratories one sample AdBlue 32% Urea solution labelled #23077 was sent on May 3, 2023. An SDS was added to the sample package.

2.5 STABILITY OF THE SAMPLES

The stability of AdBlue 32% Urea solution packed in the HDPE bottles was checked. The material was found sufficiently stable for the period of the proficiency test.

2.6 ANALYZES

The participants were requested to determine: Aldehyde as Formaldehyde, Alkalinity as NH₃, Biuret, Density at 20 °C, Insoluble matter, Phosphate as PO₄, Refractive index at 20 °C, Urea content (by total Nitrogen and by refractive index) and Trace elements (Al, Ca, Cr, Cu, Fe, Mg, Ni, K, Na and Zn).

It was explicitly requested to treat the sample as if it was a routine sample and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis/. The reported test results are tabulated per determination in appendices 1 and 2 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the result tables in appendices 1 and 2. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

The assigned value is determined by consensus based on the test results of the group of participants after rejection of the statistical outliers and/or suspect data.

According to ISO13528 all (original received or corrected) results per determination were submitted to outlier tests. In the iis procedure for proficiency tests, outliers are detected prior to calculation of the mean, standard deviation and reproducibility. For small data sets, Dixon (up to 20 test results) or Grubbs (up to 40 test results) outlier tests can be used. For larger data sets (above 20 test results) Rosner's outlier test can be used. Outliers are marked by $D(0.01)$ for the Dixon's test, by $G(0.01)$ or $DG(0.01)$ for the Grubbs' test and by $R(0.01)$ for the Rosner's test. Stragglers are marked by $D(0.05)$ for the Dixon's test, by $G(0.05)$ or $DG(0.05)$ for the Grubbs' test and by $R(0.05)$ for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve (dotted line) was projected over the Kernel Density Graph (smooth line) for reference. The Gauss curve is calculated from the consensus value and the corresponding standard deviation.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements (derived from e.g. ISO or ASTM test methods), the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The $z_{(\text{target})}$ scores are listed in the result tables of appendix 1.

Absolute values for $z < 2$ are very common and absolute values for $z > 3$ are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z < 1$	good
$1 < z < 2$	satisfactory
$2 < z < 3$	questionable
$3 < z $	unsatisfactory

4 EVALUATION

In this proficiency test no problems were encountered with the dispatch of the samples. One participant reported test results after the final reporting date and two other participants did not report any test results. Not all participants were able to report all tests requested. In total 33 participants reported 204 numerical test results. Observed were 10 outlying test results, which is 4.9%. In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

All data sets proved to have a normal Gaussian distribution.

4.1 EVALUATION PER TEST

In this section the reported test results are discussed per test. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the original data in appendices 1 and 2. The abbreviations, used in these tables, are explained in appendix 4.

Aldehyde as Formaldehyde: This determination was not problematic. All reporting participants agreed on a test result of less than 0.5 mg/kg. Therefore, no z-scores are calculated.

Alkalinity as NH₃: This determination was not problematic. All reporting participants agreed on a test result of less than 0.1 %M/M. Therefore, no z-scores are calculated.

Biuret: This determination was not problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with the requirements of ISO22241-2:19.

Density at 20 °C: This determination was problematic for a number of laboratories. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.

Insoluble matter: This determination was not problematic. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO22241-2:19.

Phosphate as PO₄: This determination was not problematic. Almost all reporting participants agreed on a test result of less than 0.25 mg/kg. Therefore, no z-scores are calculated.

Refractive index at 20 °C: This determination was not problematic. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO22241-2:19.

Urea content (by total Nitrogen): This determination was not problematic. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ISO22241-2:19.

Urea content (by refractive index): This determination was problematic. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ISO22241-2:19.

The majority of the participants agreed on a concentration near or below the limit of detection for the trace elements mentioned in paragraph 2.6. Therefore, no z-scores are calculated. The reported test results of these elements are given in appendix 2.

4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility (2.8 * standard deviation) and the target reproducibility derived from reference methods are presented in the next table.

Parameter	unit	n	average	2.8 * sd	R(lit)
Aldehyde as Formaldehyde	mg/kg	31	<0.5	n.e.	n.e.
Alkalinity as NH ₃	%M/M	31	<0.1	n.e.	n.e.
Biuret	%M/M	31	0.150	0.028	0.044
Density at 20 °C	kg/m ³	27	1089.0	0.3	0.5
Insoluble matter	mg/kg	22	2.2	4.3	8.2
Phosphate as PO ₄	mg/kg	27	<0.25	n.e.	n.e.
Refractive index at 20 °C		29	1.3824	0.0002	0.0003
Urea content (by total Nitrogen)	%M/M	7	32.1	0.3	1.1
Urea content (by refractive index)	%M/M	30	32.2	0.3	0.2

Table 3: reproducibilities of tests on sample #23077

Without further statistical calculations, it can be concluded that for many tests there is a good compliance of the group of participants with the reference test methods. The problematic tests have been discussed in paragraph 4.1.

4.3 COMPARISON OF THE PROFICIENCY TEST OF MAY 2023 WITH PREVIOUS PTS

	May 2023	May 2022	May 2021	May 2020	June 2019
Number of reporting laboratories	33	24	16	18	18
Number of test results	204	174	121	161	192
Number of statistical outliers	10	9	7	8	7
Percentage of statistical outliers	4.9%	5.2%	5.8%	5.0%	3.6%

Table 4: comparison with previous proficiency tests

In proficiency tests, outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared to the requirements of the reference test methods. The conclusions are given in the following table.

	May 2023	May 2022	May 2021	May 2020	June 2019
Aldehyde as Formaldehyde	n.e.	(++)	+/-	+/-	+/-
Alkalinity as NH ₃	n.e.	n.e.	n.e.	++	++
Biuret	+	+	+	++	+/-
Density at 20 °C	+	-	+	+	+/-
Insoluble matter	+	+	+	+/-	+/-
Phosphate as PO ₄	n.e.	(--)	-	n.e.	n.e.
Refractive index at 20 °C	+	+/-	+	+	++
Urea content (by total Nitrogen)	++	+	n.e.	+	n.e.
Urea content (by refractive index)	-	-	+/-	+	-
Elements	n.e.	n.e.	--	n.e.	n.e.

Table 5: comparison determinations to the reference test methods

For results between brackets no z-scores are calculated.

The following performance categories were used:

- ++ : group performed much better than the reference test method
- + : group performed better than the reference test method
- +/- : group performance equals the reference test method
- : group performed worse than the reference test method
- : group performed much worse than the reference test method
- n.e. : not evaluated

APPENDIX 1**Determination of Aldehyde as Formaldehyde on sample #23077; result in mg/kg**



lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex F	0.02		----	
309	ISO22241-2 Annex F	<0.5		----	
331	ISO22241-2 Annex F	0.1		----	
334	ISO22241-2 Annex F	0.04		----	
381	ISO22241-2 Annex F	<0,5		----	
398	ISO22241-2 Annex F	0.07		----	
405	ISO22241-2 Annex F	<0.004		----	
420	ISO22241-2 Annex F	<0,5		----	
455		----		----	
496	ISO22241-2 Annex F	0.1		----	
541	ISO22241-2 Annex F	<0.5		----	
551	ISO22241-2 Annex F	0.01		----	
862	ISO22241-2 Annex F	<0.5		----	
863	ISO22241-2 Annex F	0.05		----	
864	ISO22241-2 Annex F	<0.5		----	
912	ISO22241-2 Annex F	<0.50		----	
994		----		----	
1271		----		----	
1299		----		----	
1397	ISO22241-2 Annex F	<0,5		----	
1557	ISO22241-2 Annex F	0.1		----	
1615	ISO22241-2 Annex F	<0.50		----	
1656	ISO22241-2 Annex F	<0.5		----	
1807	ISO22241-2 Annex F	<0,5		----	
2568	ISO22241-2 Annex F	<0.5		----	
6015	ISO22241-2 Annex F	0.05		----	
6039	ISO22241-2 Annex F	0.050		----	
6256	ISO22241-2 Annex F	0.120		----	
6330	ISO22241-2 Annex F	0.0097		----	
6439	ISO22241-2 Annex F	0.06		----	
6441	ISO22241-2 Annex F	<0.1		----	
6504	ISO22241-2 Annex F	<0,1		----	
6508	ISO22241-2 Annex F	0.200	C	----	first reported 0.340
6512	ISO22241-2 Annex F	<0,5		----	
6517	ISO22241-2 Annex F	< 0,479		----	
	n	31			
	mean (n)	<0.5			

Determination of Alkalinity as NH₃ on sample #23077; result in %M/M

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex D	<0.01	C	----	first reported 0.004
309	ISO22241-2 Annex D	<0.10		----	
331	ISO22241-2 Annex D	0.0		----	
334	ISO22241-2 Annex D	<0.10		----	
381	ISO22241-2 Annex D	<0,1		----	
398	ISO22241-2 Annex D	0.04		----	
405	ISO22241-2 Annex D	0.046		----	
420	ISO22241-2 Annex D	0.04		----	
455	ISO22241-2 Annex D	0.04890		----	
496	ISO22241-2 Annex D	0.04		----	
541	ISO22241-2 Annex D	<0.1		----	
551	ISO22241-2 Annex D	0.05		----	
862	ISO22241-2 Annex D	<0.1		----	
863	ISO22241-2 Annex D	0.033		----	
864	ISO22241-2 Annex D	<0.10		----	
912	ISO22241-2 Annex D	<0.10		----	
994		----		----	
1271		----		----	
1299		----		----	
1397	ISO22241-2 Annex D	<0,1		----	
1557	ISO22241-2 Annex D	0.046	C	----	first reported 0.00476
1615	ISO22241-2 Annex D	<0.10		----	
1656	ISO22241-2 Annex D	<0.1		----	
1807	ISO22241-2 Annex D	0.025		----	
2568	ISO22241-2 Annex D	<0.1		----	
6015	ISO22241-2 Annex D	0.040		----	
6039	ISO22241-2 Annex D	0.0438		----	
6256	ISO22241-2 Annex D	0.0419		----	
6330	ISO22241-2 Annex D	0.0367		----	
6439	ISO22241-2 Annex D	0.04		----	
6441	ISO22241-2 Annex D	0.0		----	
6504	ISO22241-2 Annex D	0.047		----	
6508		----		----	
6512	ISO22241-2 Annex D	0.04		----	
6517	ISO22241-2 Annex D	0.039		----	
	n	31			
	mean (n)	<0.1			



Determination of Biuret on sample #23077; result in %M/M

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex E	0.15		-0.03	
309	ISO22241-2 Annex E	0.13		-1.30	
331	ISO22241-2 Annex E	0.1	R(0.01)	-3.21	
334	ISO22241-2 Annex E	0.15		-0.03	
381	ISO22241-2 Annex E	0.157		0.42	
398	ISO22241-2 Annex E	0.150		-0.03	
405	ISO22241-2 Annex E	0.14		-0.67	
420	ISO22241-2 Annex E	0.137		-0.86	
455	ISO22241-2 Annex E	0.14000		-0.67	
496	ISO22241-2 Annex E	0.15		-0.03	
541	ISO22241-2 Annex E	0.14		-0.67	
551	ISO22241-2 Annex E	0.1557		0.33	
862	ISO22241-2 Annex E	0.16		0.61	
863	ISO22241-2 Annex E	0.158		0.48	
864	ISO22241-2 Annex E	0.16		0.61	
912	ISO22241-2 Annex E	0.17		1.24	
994		----		----	
1271		----		----	
1299		----		----	
1397	ISO22241-2 Annex E	0.15		-0.03	
1557	ISO22241-2 Annex E	0.1562		0.37	
1615	ISO22241-2 Annex E	0.15		-0.03	
1656	ISO22241-2 Annex E	0.14		-0.67	
1807	ISO22241-2 Annex E	0.16		0.61	
2568	ISO22241-2 Annex E	0.1469		-0.23	
6015	ISO22241-2 Annex E	0.165		0.93	
6039	ISO22241-2 Annex E	0.14		-0.67	
6256	ISO22241-2 Annex E	0.170		1.24	
6330	ISO22241-2 Annex E	0.1483		-0.14	
6439	ISO22241-2 Annex E	0.14		-0.67	
6441	ISO22241-2 Annex E	0.1		-0.67	
6504	ISO22241-2 Annex E	0.166		0.99	
6508	ISO22241-2 Annex E	0.15		-0.03	
6512	ISO22241-2 Annex E	0.15		-0.03	
6517	ISO22241-2 Annex E	0.144		-0.41	
	normality	OK			
	n	31			
	outliers	1			
	mean (n)	0.1505			
	st.dev. (n)	0.01013			
	R(calc.)	0.0284			
	st.dev.(ISO22241-2:19)	0.01571			
	R(ISO22241-2:19)	0.044			





Determination of Density at 20 °C on sample #23077; result in kg/m³

lab	method	value	mark	z(targ)	remarks
273	D4052	1088.9		-0.45	
309	D4052	1089.12		0.79	
331	ISO12185	1088.8		-1.01	
334	ISO12185	1089.0		0.11	
381	ISO12185	1089.1		0.67	
398	ISO12185	1088.97		-0.05	
405	ISO12185	1089.1		0.67	
420	ISO12185	1088.9		-0.45	
455	ISO12185	1088.9		-0.45	
496	ISO12185	1088.89		-0.50	
541	ISO12185	1088.9		-0.45	
551	D4052	1089.0		0.11	
862	ISO12185	1088.9		-0.45	
863	ISO12185	1088.91		-0.39	
864	ISO12185	1088.9		-0.45	
912		----		----	
994		----		----	
1271	ISO12185	1088.95		-0.17	
1299		----		----	
1397	ISO12185	1088.9		-0.45	
1557	ISO3675	1089.7	C,R(0.01)	4.03	first reported 1090.0
1615	D4052	1088.96		-0.11	
1656	D4052	1089.0		0.11	
1807	ISO12185	1089.0		0.11	
2568	ISO12185	1089.1		0.67	
6015	ISO12185	1089.10		0.67	
6039	ISO12185	1088.95		-0.17	
6256		----		----	
6330	ISO3675	1089.10		0.67	
6439	ISO3675	1089.0	C	0.11	first reported 1088.5
6441	ISO12185	1089.2		1.23	
6504	ISO3675	1090.0	R(0.01)	5.71	
6508	In house	1089.70	R(0.01)	4.03	
6512	ISO12185	1088.9		-0.45	
6517		----		----	
	normality	OK			
	n	27			
	outliers	3			
	mean (n)	1088.979			
	st.dev. (n)	0.0962			
	R(calc.)	0.269			
	st.dev.(ISO12185:96)	0.1786			
	R(ISO12185:96)	0.5			



Determination of Insoluble matter on sample #23077; result in mg/kg

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex G	<1		----	
309	ISO22241-2 Annex G	0.9		-0.45	
331	ISO22241-2 Annex G	2.5		0.09	
334	ISO22241-2 Annex G	<1		----	
381	ISO22241-2 Annex G	6.2		1.35	
398	ISO22241-2 Annex G	1.50		-0.25	
405	ISO22241-2 Annex G	0.96		-0.43	
420	ISO22241-2 Annex G	0.476		-0.60	
455	ISO22241-2 Annex G	50.50969	R(0.01)	16.44	
496	ISO22241-2 Annex G	0.00		-0.76	
541	ISO22241-2 Annex G	<1.0		----	
551		----		----	
862	ISO22241-2 Annex G	4.1		0.64	
863	ISO22241-2 Annex G	3.00		0.26	
864	ISO22241-2 Annex G	3.0		0.26	
912	ISO22241-2 Annex G	4.7		0.84	
994		----		----	
1271	ISO22241-2 Annex G	2.5		0.09	
1299		----		----	
1397	ISO22241-2 Annex G	1.69		-0.19	
1557	ISO22241-2 Annex G	3.11		0.30	
1615	ISO22241-2 Annex G	0.86		-0.47	
1656		----		----	
1807	ISO22241-2 Annex G	2.2		-0.01	
2568	ISO22241-2 Annex G	<0.1		----	
6015	ISO22241-2 Annex G	2.595		0.12	
6039	ISO22241-2 Annex G	8.7586	R(0.05)	2.22	
6256	ISO22241-2 Annex G	0.1090		-0.72	
6330	ISO22241-2 Annex G	3.5633		0.45	
6439	ISO22241-2 Annex G	1.35		-0.30	
6441	ISO22241-2 Annex G	1.23	C	-0.34	first reported 12.3
6504	ISO22241-2 Annex G	2.64		0.14	
6508		----		----	
6512	ISO22241-2 Annex G	<1		----	
6517		----		----	
normality		OK			
n		22			
outliers		2			
mean (n)		2.2356			
st.dev. (n)		1.54226			
R(calc.)		4.3183			
st.dev.(ISO22241-2:19)		2.93571			
R(ISO22241-2:19)		8.220			





Determination of Phosphate as PO₄ on sample #23077; result in mg/kg

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex H	0.40	C	----	First reported 1.11; Possibly a false positive test result?
309	ISO22241-2 Annex H	<0.25		----	
331	ISO22241-2 Annex H	0.05		----	
334	ISO22241-2 Annex H	0.13		----	
381	ISO22241-2 Annex H	<0,05		----	
398	ISO22241-2 Annex H	0.03		----	
405	ISO22241-2 Annex I	<0.003		----	
420	ISO22241-2 Annex H	<0,05		----	
455		----		----	
496	ISO22241-2 Annex H	0.01		----	
541	ISO22241-2 Annex H	0.07		----	
551		----		----	
862	ISO22241-2 Annex H	<0.05		----	
863	ISO22241-2 Annex H	0.02		----	
864	ISO22241-2 Annex H	<0.05		----	
912	ISO22241-2 Annex H	<0.05		----	
994		----		----	
1271		----		----	
1299		----		----	
1397	ISO22241-2 Annex H	<0,05		----	
1557	ISO22241-2 Annex H	0.05		----	
1615	ISO22241-2 Annex H	<0.05		----	
1656		----		----	
1807		----		----	
2568	ISO22241-2 Annex H	<0.05		----	
6015	ISO22241-2 Annex H	0.000		----	
6039	ISO22241-2 Annex H	0.014		----	
6256	ISO22241-2 Annex H	0.11		----	
6330	ISO22241-2 Annex H	0.1067		----	
6439	ISO22241-2 Annex H	0.068		----	
6441	ISO22241-2 Annex H	0.1		----	
6504	ISO22241-2 Annex I	<0,01		----	
6508		----		----	
6512	ISO22241-2 Annex I	<0,1		----	
6517	ISO22241-2 Annex H	0.173		----	
	n	27			
	mean (n)	<0.25			



Determination of Refractive index at 20 °C on sample #23077;

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex C	1.3824	C	0.34	first reported 1.3820
309	ISO22241-2 Annex C	1.38233		-0.25	
331	ISO22241-2 Annex C	1.3824		0.34	
334	ISO22241-2 Annex C	1.3823		-0.50	
381	ISO22241-2 Annex C	1.38247		0.94	
398	ISO22241-2 Annex C	1.38232		-0.34	
405		----		----	
420	ISO22241-2 Annex C	1.38233		-0.25	
455	ISO22241-2 Annex C	1.38231		-0.42	
496	ISO22241-2 Annex C	1.38186	R(0.01)	-4.24	
541	ISO22241-2 Annex C	1.38233		-0.25	
551	ISO22241-2 Annex C	1.38239		0.26	
862	ISO22241-2 Annex C	1.3823		-0.50	
863	ISO22241-2 Annex C	1.38234		-0.17	
864	ISO22241-2 Annex C	1.3824		0.34	
912	ISO22241-2 Annex C	1.3825		1.19	
994		----		----	
1271	ISO22241-2 Annex C	1.38237		0.09	
1299		----		----	
1397	ISO22241-2 Annex C	1.3824		0.34	
1557	ISO22241-2 Annex C	1.3822		-1.35	
1615	ISO22241-2 Annex C	1.38234		-0.17	
1656	ISO22241-2 Annex C	1.3823		-0.50	
1807	ISO22241-2 Annex C	1.3823		-0.50	
2568	ISO22241-2 Annex C	1.3824		0.34	
6015	ISO22241-2 Annex C	1.38235		-0.08	
6039	ISO22241-2 Annex C	1.38235		-0.08	
6256	ISO22241-2 Annex C	1.38231		-0.42	
6330	ISO22241-2 Annex C	1.3820	R(0.01)	-3.05	
6439	ISO22241-2 Annex C	1.38244		0.68	
6441	ISO22241-2 Annex C	1.4		0.34	
6504	ISO22241-2 Annex C	1.38243		0.60	
6508	ISO22241-2 Annex C	1.3827	R(0.01)	2.89	
6512	ISO22241-2 Annex C	1.3823		-0.50	
6517	ISO22241-2 Annex C	1.38241		0.43	
normality		OK			
n		29			
outliers		3			
mean (n)		1.382360			
st.dev. (n)		0.0000627			
R(calc.)		0.000176			
st.dev.(ISO22241-2:19)		0.0001179			
R(ISO22241-2:19)		0.00033			





Determination of Urea content (by total Nitrogen) on sample #23077; result in %M/M

lab	method	value	mark	z(target)	remarks
273	ISO22241-2 Annex B	32.0		-0.36	
309		----		----	
331		----		----	
334	ISO22241-2 Annex B	32.1		-0.09	
381	ISO22241-2 Annex B	32.0		-0.36	
398		----		----	
405	ISO22241-2 Annex B	32.22		0.23	
420		----		----	
455		----		----	
496		----		----	
541		----		----	
551		----		----	
862	ISO22241-2 Annex B	32.1		-0.09	
863		----		----	
864		----		----	
912		----		----	
994		----		----	
1271		----		----	
1299		----		----	
1397	ISO22241-2 Annex B	32.2		0.17	
1557		----		----	
1615		----		----	
1656		----		----	
1807		----		----	
2568		----		----	
6015		----		----	
6039		----		----	
6256		----		----	
6330	ISO22241-2 Annex B	32.321		0.50	
6439		----		----	
6441		----		----	
6504		----		----	
6508		----		----	
6512		----		----	
6517		----		----	
normality		unknown			
n		7			
outliers		0			
mean (n)		32.134			
st.dev. (n)		0.1190			
R(calc.)		0.333			
st.dev.(ISO22241-2:19)		0.3761			
R(ISO22241-2:19)		1.053			



Determination of Urea content (by refractive index) on sample #23077; result in %M/M

lab	method	value	mark	z(targ)	remarks
273	ISO22241-2 Annex C	32.2		0.28	
309	ISO22241-2 Annex C	32.2		0.28	
331	ISO22241-2 Annex C	32.3		1.61	
334	ISO22241-2 Annex C	32.2		0.28	
381	ISO22241-2 Annex C	32.2		0.28	
398	ISO22241-2 Annex C	32.1		-1.04	
405		----		----	
420	ISO22241-2 Annex C	32.1		-1.04	
455	ISO22241-2 Annex C	32.37		2.54	
496	ISO22241-2 Annex C	31.8	R(0.05)	-5.02	
541	ISO22241-2 Annex C	32.1		-1.04	
551	ISO22241-2 Annex C	32.15		-0.38	
862	ISO22241-2 Annex C	32.1		-1.04	
863	ISO22241-2 Annex C	32.12		-0.78	
864	ISO22241-2 Annex C	32.1		-1.04	
912	ISO22241-2 Annex C	32		-2.37	
994		----		----	
1271	ISO22241-2 Annex C	32.37		2.54	
1299		----		----	
1397		----		----	
1557	ISO22241-2 Annex C	32.25		0.95	
1615	ISO22241-2 Annex C	32.13		-0.64	
1656	ISO22241-2 Annex C	32.1		-1.04	
1807	ISO22241-2 Annex C	32.1		-1.04	
2568	ISO22241-2 Annex C	32.1		-1.04	
6015	ISO22241-2 Annex C	32.13		-0.64	
6039	ISO22241-2 Annex C	32.241		0.83	
6256	ISO22241-2 Annex C	32.0913		-1.16	
6330	ISO22241-2 Annex C	32.3467		2.23	
6439	ISO22241-2 Annex C	32.2		0.28	
6441	ISO22241-2 Annex C	32.2		0.28	
6504	ISO22241-2 Annex C	32.17		-0.11	
6508	ISO22241-2 Annex C	32.42	C	3.20	first reported 31.84
6512	ISO22241-2 Annex C	32.1		-1.04	
6517	ISO22241-2 Annex C	32.168		-0.14	
normality		OK			
n		30			
outliers		1			
mean (n)		32.179			
st.dev. (n)		0.1001			
R(calc.)		0.280			
st.dev.(ISO22241-2:19)		0.0754			
R(ISO22241-2:19)		0.211			



APPENDIX 2

Summary of reported elements on sample #23077; result in mg/kg

lab	method	Al	Ca	Cr	Cu	Fe	Mg	Ni	K	Na	Zn
273	*)	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
309	*)	<0.2	<0.5	<0.2	<0.2	<0.5	<0.5	<0.2	<0.2	<0.2	<0.2
331		----	----	----	----	----	----	----	----	----	----
334	*)	0.03	<0.01	0.01	<0.01	<0.01	0.01	<0.01	0.04	<0.01	0.05
381	*)	<0,1	<0,2	<0,2	<0,2	<0,1	<0,4	<0,2	<0,2	<0,2	<0,2
398		----	----	----	----	----	----	----	----	----	----
405	*)	0.018	0.038	0.023	0.012	<0.001	<0.001	<0.001	0.083	0.15	0.043
420	*)	0.10	0.13	<0.1 C	0.03	0.03	<0.1 C	<0,03	<0,03	0.17	0.03
455	*)	0	0.6545	0	0	0	0	0	0.24492	0.22883	0
496		----	----	----	----	----	----	----	----	----	----
541		----	----	----	----	----	----	----	----	----	----
551	*)	<0.021	0.2563	0.0296	<0.023	0.0249	0.0082	<0.029	0.039	0.0714	0.0654
862	*)	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	0.2	<0.1
863	*)	0.01	0.00	0.02	0.01	0.00	0.00	0.00	0.00	0.17	0.04
864	*)	<0.1	<0.2	<0.2	<0.2	<0.1	<0.4	<0.2	<0.2	<0.2	<0.2
912	*)	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
994		----	----	----	----	----	----	----	----	----	----
1271		----	----	----	----	----	----	----	----	----	----
1299		----	----	----	----	----	----	----	----	----	----
1397	*)	0.090	0.54 C	0.045	0.060	0.050	0.050	0.060	0.145	0.74	0.17 C
1557	*)	<0,5	0.47 C	0.05	0.18 C	0.33 C	0.48 C	0.03	<0,5	<0,5	0.05
1615	*)	<0.10	<0.10	<0.10	<0.10	<0.10	<0.10	<0.10	<0.10	0.15	<0.10
1656		<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
1807		----	----	----	----	----	----	----	----	----	----
2568	*)	<0.5	<0.5	<0.2	<0.2	<0.5	<0.5	<0.2	<0.5	<0.5	<0.2
6015	*)	0.000	0.005	0.020	0.000	0.000	0.000	0.010	0.000	0.000	0.070
6039		0.023 C	0.112	0.028	0.016	0.005	0.011	0.013	0.1563	0.220	0.07
6256	*)	0.100	0.113	0.058	0.024	0.018	0.022	0.035	0.030	0.209	0.116
6330	*)	0.00	0.00	0.0233	0.0247	0.00	0.00	0.028	0.1362	0.1580	0.0210
6439	*)	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	0.045	0.188	<0.05
6441	*)	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1
6504	*)	<0,01	<0,01	<0,01	<0,01	<0,01	<0,01	<0,01	<0,01	<0,01	0.04
6508		----	----	----	----	----	----	----	----	----	----
6512	*)	<0,05	0.08	<0,05	<0,05	<0,05	<0,05	<0,05	<0,2	<0,1	<0,05
6517	*)	0.023	0.015	0.011	0.009	0.010	0.014	0.020	0.044	0.196	0.079

*) ISO22241-2 Annex I

Lab 420 first reported Cr 0.11; Mg 0.07

Lab 1397 first reported Ca 0.675; Zn 0.215

Lab 1557 first reported Ca 0.47; Cu 0.18; Fe 0.33; Mg 0.48

Lab 6039 first reported Al 0.023

APPENDIX 3

Number of participants per country

1 lab in ARGENTINA
1 lab in AUSTRIA
1 lab in AZERBAIJAN
1 lab in BOSNIA and HERZEGOVINA
1 lab in BRAZIL
1 lab in BULGARIA
3 labs in CHINA, People's Republic
1 lab in CROATIA
1 lab in CZECH REPUBLIC
2 labs in FRANCE
3 labs in GERMANY
2 labs in INDIA
2 labs in ITALY
1 lab in JAPAN
1 lab in MALAYSIA
1 lab in NETHERLANDS
1 lab in POLAND
1 lab in ROMANIA
4 labs in SERBIA
1 lab in SOUTH AFRICA
2 labs in SPAIN
1 lab in TURKEY
2 labs in UNITED KINGDOM

APPENDIX 4

Abbreviations

C	= final test result after checking of first reported suspect test result
D(0.01)	= outlier in Dixon's outlier test
D(0.05)	= straggler in Dixon's outlier test
G(0.01)	= outlier in Grubbs' outlier test
G(0.05)	= straggler in Grubbs' outlier test
DG(0.01)	= outlier in Double Grubbs' outlier test
DG(0.05)	= straggler in Double Grubbs' outlier test
R(0.01)	= outlier in Rosner's outlier test
R(0.05)	= straggler in Rosner's outlier test
E	= calculation difference between reported test result and result calculated by iis
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
f+?	= possibly a false positive test result?
f-?	= possibly a false negative test result?
SDS	= Safety Data Sheet

Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 ISO5725:86
- 3 ISO5725 parts 1-6:94
- 4 ISO13528:05
- 5 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 6 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 7 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 8 J.N. Miller, Analyst, 118, 455, (1993)
- 9 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 10 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 11 W. Horwitz and R. Albert, J. AOAC Int, 79.3, 589-621, (1996)
- 12 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)