



Institute for  
Interlaboratory Studies

## Results of Proficiency Test Vacuum Gasoil (VGO) November 2023

Organized by: Institute for Interlaboratory Studies  
Spijkenisse, the Netherlands

Author: Mr. R.J. Starink, BSc  
Correctors: Mr. M. Meijer, BSc & Mrs. E.R. Montenij-Bos  
Approved by: Mr. R.J. Starink, BSc

Report: iis23G08

February 2024

**CONTENTS**

1	INTRODUCTION .....	3
2	SET UP .....	3
2.1	ACCREDITATION .....	3
2.2	PROTOCOL.....	4
2.3	CONFIDENTIALITY STATEMENT .....	4
2.4	SAMPLES .....	4
2.5	STABILITY OF THE SAMPLES.....	5
2.6	ANALYZES .....	6
3	RESULTS .....	6
3.1	STATISTICS .....	6
3.2	GRAPHICS .....	7
3.3	Z-SCORES .....	8
4	EVALUATION .....	8
4.1	EVALUATION PER SAMPLE AND PER TEST .....	9
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES.....	12
4.3	COMPARISON OF THE PROFICIENCY TEST OF NOVEMBER 2023 WITH PREVIOUS PTS .....	13

## Appendices:

1.	Data, statistical and graphic results .....	15
2.	Number of participants per country.....	53
3.	Abbreviations and literature.....	54

## 1 INTRODUCTION

Since 2013 the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the analysis of Vacuum Gasoil (VGO) based on the latest version of ISO8217 every year. During the annual proficiency testing program of 2023 it was decided to continue the round robin for the analysis of Vacuum Gasoil (VGO).

In this interlaboratory study registered for participation:

- 56 laboratories in 26 countries for regular analyzes in VGO iis23G08
- 40 laboratories in 23 countries on metal analyzes in VGO iis23G08M

In total 56 laboratories in 26 countries registered for participation in one or two proficiency tests, see appendix 2 for the number of participants per country. In this report the results of the Vacuum Gasoil (VGO) proficiency tests are presented and discussed. This report is also electronically available through the iis website [www.iisnl.com](http://www.iisnl.com).

## 2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to a laboratory that has performed the tests in accordance with for ISO/IEC17043 relevant requirements of ISO/IEC17025.

In this proficiency test the participants received, depending on the registration, from one up to two different samples of Vacuum Gasoil, see table below.

Sample ID	PT ID	Quantity	Purpose
#23250	iis23G08	1x 1 L	Regular analyzes
#23251	iis23G08M	1x 250 mL	Metal analyzes

Table 1: Vacuum Gasoil samples used in PT iis23G08 and iis23G08M.

The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

### 2.1 ACCREDITATION

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, is accredited in agreement with ISO/IEC17043:2010 (R007), since January 2000, by the Dutch Accreditation Council (Raad voor Accreditatie). This PT falls under the accredited scope. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

## 2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website [www.iisnl.com](http://www.iisnl.com), from the FAQ page.

## 2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

## 2.4 SAMPLES

For the preparation of the sample for the regular analyzes in Vacuum Gasoil a batch of approximately 100 liters of Vacuum Gasoil was obtained from a third party. After homogenization 80 amber glass bottles of 1 L were filled and labelled #23250. The homogeneity of the subsamples was checked by determination of Density at 15 °C in accordance with ISO12185 on 8 stratified randomly selected subsamples.

	Density at 15 °C in kg/m <sup>3</sup>
sample #23250-1	911.6
sample #23250-2	911.6
sample #23250-3	911.7
sample #23250-4	911.7
sample #23250-5	911.7
sample #23250-6	911.7
sample #23250-7	911.7
sample #23250-8	911.7

Table 2: homogeneity test results of subsamples #23250

From the above test results the repeatability was calculated and compared with 0.3 times the reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Density at 15 °C in kg/m <sup>3</sup>
r (observed)	0.13
reference test method	ISO12185:96
0.3 x R (reference test method)	0.45

Table 3: evaluation of the repeatability of subsamples #23250

The calculated repeatability is in agreement with 0.3 times the reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

For the preparation of the sample for the metal analyzes in Vacuum Gasoil a batch of approximately 20 liters of Vacuum Gasoil was obtained from a third party. To this batch detectable levels of Calcium, Vanadium and Nickel were added. After homogenization 75 PE bottles of 250 mL were filled and labelled #23251.

The homogeneity of the subsamples was checked by determination of Nickel and Vanadium in accordance with IP501 on 8 stratified randomly selected subsamples.

	Nickel in mg/kg	Vanadium in mg/kg
sample #23251-1	12.6	10.5
sample #23251-2	12.6	10.9
sample #23251-3	12.1	10.7
sample #23251-4	13.5	11.3
sample #23251-5	12.6	10.8
sample #23251-6	13.8	10.8
sample #23251-7	12.6	10.6
sample #23251-8	13.0	11.3

Table 4: homogeneity test results of subsamples #23251

From the above test results the repeatabilities were calculated and compared with 0.3 times the corresponding reproducibility of the reference test method in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Nickel in mg/kg	Vanadium in mg/kg
r (observed)	1.6	0.8
reference test method	IP621:16	IP621:16
0.3 x R (reference test method)	2.5	1.3

Table 5: evaluation of the repeatabilities of subsamples #23251

The calculated repeatabilities are in agreement with 0.3 times the corresponding reproducibility of the reference test method. Therefore, homogeneity of the subsamples was assumed.

Depending on the registration of the participant the appropriate set of PT samples was sent on November 8, 2023. An SDS was added to the sample package.

## 2.5 STABILITY OF THE SAMPLES

The stability of Vacuum Gasoil packed in amber glass bottles and PE bottles was checked. The material was found sufficiently stable for the period of the proficiency test.

## 2.6 ANALYZES

The participants were requested to determine on sample #23250: Total Acid Number, Aniline Point, Asphaltenes, Carbon Residue (micro method), Density at 15 °C, Flash Point PMcc, Kinematic Viscosity at 50 °C and 100 °C, Nitrogen, Pour Point (Manual and Automated), Total Sulfur, Simulated Distillation and Distillation at 10 mmHg (IBP, 10% rec, 30% rec, 50% rec, 70% rec, 90% rec and FBP).

On sample #23251 it was requested to determine: Aluminum, Silicon, Sum Aluminum and Silicon, Arsenic, Copper, Iron, Nickel, Sodium, Vanadium and Calcium.

It was also requested to report some analytical details on the determination of Total Acid Number.

It was explicitly requested to treat the samples as if they were routine samples and to report the test results using the indicated units on the report form and not to round the test results, but report as much significant figures as possible. It was also requested not to report 'less than' test results, which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations.

To get comparable test results a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website [www.iisnl.com](http://www.iisnl.com).

## 3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal [www.kpmd.co.uk/sgs-iis/](http://www.kpmd.co.uk/sgs-iis/). The reported test results are tabulated per determination in appendix 1 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the result tables in appendix 1. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

### 3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For the statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a data set does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

The assigned value is determined by consensus based on the test results of the group of participants after rejection of the statistical outliers and/or suspect data.

According to ISO13528 all (original received or corrected) results per determination were submitted to outlier tests. In the iis procedure for proficiency tests, outliers are detected prior to calculation of the mean, standard deviation and reproducibility. For small data sets, Dixon (up to 20 test results) or Grubbs (up to 40 test results) outlier tests can be used. For larger data sets (above 20 test results) Rosner's outlier test can be used. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by R(0.01) for the Rosner's test. Stragglers are marked by D(0.05) for the Dixon's test, by G(0.05) or DG(0.05) for the Grubbs' test and by R(0.05) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirement based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1. was met for all evaluated tests. Therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

### 3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve (dotted line) was projected over the Kernel Density

Graph (smooth line) for reference. The Gauss curve is calculated from the consensus value and the corresponding standard deviation.

### 3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements (derived from e.g. ISO or ASTM test methods), the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

The target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

$$Z_{(\text{target})} = (\text{test result} - \text{average of PT}) / \text{target standard deviation}$$

The  $Z_{(\text{target})}$  scores are listed in the test result tables in appendix 1.

Absolute values for  $z < 2$  are very common and absolute values for  $z > 3$  are very rare. Therefore, the usual interpretation of z-scores is as follows:

$ z  < 1$	good
$1 <  z  < 2$	satisfactory
$2 <  z  < 3$	questionable
$3 <  z $	unsatisfactory

## 4 EVALUATION

In this proficiency test some problems were encountered with the dispatch of the samples. The reporting time on the data entry portal was extended with another week.

In the Vacuum Gasoil regular round robin ten participants reported test results after the extended reporting date and six other participants did not report any test results.

In the Vacuum Gasoil Metals round robin seven participants reported test results after the extended reporting date and five other participants did not report any test results.

Not all participants were able to report all tests requested.

In total 51 participants reported 898 numerical test results. Observed were 46 outlying test results, which is 5.1%. In proficiency tests outlier percentages of 3% - 7.5% are quite normal.



Not all data sets proved to have a normal Gaussian distribution. These are referred to as “not OK” or “suspect”. The statistical evaluation of these data sets should be used with due care, see also paragraph 3.1.

#### 4.1 EVALUATION PER SAMPLE AND PER TEST

In this section the reported test results are discussed per sample and per test. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the original data in appendix 1. The abbreviations, used in these tables, are explained in appendix 3.

In the iis PT reports ASTM test methods are referred to with a number (e.g. D4530) and an added designation for the year that the test method was adopted or revised (e.g. D4530:15). When a method has been reapproved an “R” will be added and the year of approval (e.g. D4530:15R20).

Although VGO is an important feedstock for cracking installations there are only a few analytical test methods specifically designed for the analysis of VGO. Most parameters are to be determined by using methods that are intended for residual fuel oil and blending components. Where applicable precision data for Fuel Oil is used.

##### **sample #23250**

Total Acid Number: The majority of the laboratories reported to use Inflection Point with titration volume of 125 mL. Therefore, the z-scores are calculated with the reproducibility of ASTM D664-A:18e2 for Inflection Point at titration volume 125 mL.

The group of participants had difficulty to meet the target requirements. Six statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D664-A:18e2 for IP at 125 mL and BEP at 60 mL and 125 mL. However, calculated reproducibility is in agreement with the requirements of ASDTM D664-A:18e2 for IP at 60 mL. The mean is below the application range of ASTM D664:18e2, therefore no z-scores are calculated.

Aniline Point: The group of participants had difficulty to meet the target requirements. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of ASTM D611:23.

Asphaltenes: The group of participants met the target requirements. All reporting participants agreed on a level <0.5 %M/M which is below the application range of test method IP143:04R21. Therefore, no z-scores are calculated.

Carbon Residue (micro method): The group of participants met the target requirements. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ASTM D4530:15R20.

Density at 15 °C: The group of participants met the target requirements. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of ISO12185:96.

Flash Point PMcc: The group of participants met the target requirements. Two statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with requirements of ASTM D93-B:20.

Kinematic Viscosity at 50 °C: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D445:23.

Kinematic Viscosity at 100 °C: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of ASTM D445:23.

Nitrogen: The group of participants met the target requirements. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with ASTM D5762:18a.

Pour Point Manual: The group of participants met the target requirements. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is in agreement with ASTM D97:17bR22.

Pour Point Automated: The group of participants had difficulty to meet the target requirements. No statistical outliers were observed. The calculated reproducibility is not in agreement with ASTM D5950:14R20.

Total Sulfur: The group of participants met the target requirements. Five statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of ASTM D4294:21.

Simulated Distillation: The group of participants met the target requirements. Over seven parameters nine statistical outliers were observed and one other test result was excluded. All the calculated reproducibilities after rejection of the suspect data are in agreement with the requirements of ASTM D6352:19e1.

Distillation at 10 mmHg as AET: The group of participants met the target requirements. Over seven parameters four statistical outliers were observed. Allmost all calculated reproducibilities after rejection of the statistical outliers are in agreement with the requirements of ASTM D1160:18, except for 90% recovered.

**sample #23251**

- Aluminum as Al: The group of participants may have had difficulty to meet the target requirements. No statistical outliers were observed. The calculated reproducibility is not in agreement with the requirements of IP501:05R19. The mean is below the application range of IP501:05R19, therefore no z-scores are calculated.
- Silicon as Si: The group of participants may have had difficulty to meet the target requirements. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is not in agreement with the requirements of IP501:05R19. The mean is below the application range of IP501:05R19, therefore no z-scores are calculated.
- Sum Aluminum and Silicon: The group of participants may have had difficulty to meet the target requirements. One statistical outlier was observed. The calculated reproducibility after rejection of the statistical outlier is not in agreement with the requirements of IP501:05R19. The mean is below the application range of IP501:05R19, therefore no z-scores are calculated.
- Arsenic as As: All reporting participants agreed on a level <10 mg/kg. Therefore, no z-scores are calculated.
- Copper as Cu: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with IP621:16.
- Iron as Fe: The group of participants met the target requirements. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of IP621:16.
- Nickel as Ni: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of IP621:16.
- Sodium as Na: The group of participants met the target requirements. Three statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of IP621:16.
- Vanadium as V: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of IP621:16.
- Calcium as Ca: The group of participants met the target requirements. No statistical outliers were observed. The calculated reproducibility is in agreement with the requirements of IP621:16.

## 4.2 PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES

A comparison has been made between the reproducibility as declared by the reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility ( $2.8 \cdot$  standard deviation) and the target reproducibility derived from reference methods are presented in the next table.

Parameter	unit	n	average	2.8 * sd	R(lit)
Total Acid Number	mg KOH/g	21	0.042	0.027	(0.008)
Aniline Point	°C	19	89.4	2.2	1
Asphaltenes	%M/M	33	<0.5	n.e.	n.e.
Carbon Residue (micro method)	%M/M	29	0.15	0.07	0.15
Density at 15 °C	kg/m <sup>3</sup>	45	912.2	1.2	1.5
Flash Point PMcc	°C	36	192.1	6.7	10
Kinematic Viscosity at 50 °C	mm <sup>2</sup> /s	45	37.92	1.65	3.21
Kinematic Viscosity at 100 °C	mm <sup>2</sup> /s	39	7.442	0.189	0.898
Nitrogen	mg/kg	27	1071	251	285
Pour Point Manual	°C	38	38.1	5.1	9
Pour Point Automated 3 °C int.	°C	10	36.5	8.8	6.1
Total Sulfur	%M/M	38	0.041	0.004	0.009
Simulated Distillation					
Initial Boiling Point	°C	11	263.9	17.8	49.1
Temp 10% recovered	°C	10	361.0	3.3	7.1
Temp 30% recovered	°C	10	407.5	3.5	5.9
Temp 50% recovered	°C	10	438.3	4.6	6.4
Temp 70% recovered	°C	11	474.1	6.4	7.2
Temp 90% recovered	°C	11	525.5	6.2	10.5
Final Boiling Point	°C	11	604.4	21.3	38.1
Distillation at 10 mmHg as AET					
Initial Boiling Point	°C	21	282.7	41.3	49.4
Temp 10% recovered	°C	24	381.1	12.5	17.3
Temp 30% recovered	°C	24	419.3	7.4	11.3
Temp 50% recovered	°C	24	445.3	8.0	10.8
Temp 70% recovered	°C	24	476.8	9.8	9.6
Temp 90% recovered	°C	24	526.7	17.8	10.3
Final Boiling Point	°C	21	548.5	16.9	26.9

Table 6: reproducibilities of tests on sample #23250

For results between brackets no z-scores are calculated.

Parameter	unit	n	average	2.8 * sd	R(lit)
Aluminum as Al	mg/kg	24	4.2	1.9	(1.4)
Silicon as Si	mg/kg	17	2.1	1.2	(0.7)
Sum Aluminum and Silicon	mg/kg	20	6.4	3.1	(1.6)
Arsenic as As	mg/kg	3	<10	n.e.	n.e.
Copper as Cu	mg/kg	23	1.5	0.7	2.2
Iron as Fe	mg/kg	31	3.4	1.2	3.9
Nickel as Ni	mg/kg	35	12.4	3.8	8.1
Sodium as Na	mg/kg	32	2.9	1.3	4.0
Vanadium as V	mg/kg	35	11.9	1.9	4.7
Calcium as Ca	mg/kg	30	10.1	2.8	2.9

Table 7: reproducibilities of tests on sample #23251

For results between brackets no z-scores are calculated.

Without further statistical calculations it can be concluded that for many tests there is a good compliance of the group of participants with the reference test methods. The problematic tests have been discussed in paragraph 4.1.

#### 4.3 COMPARISON OF THE PROFICIENCY TEST OF NOVEMBER 2023 WITH PREVIOUS PTS

	November 2023	November 2022	December 2021	December 2020	December 2019
Number of reporting laboratories	51	57	63	64	63
Number of test results	898	1048	1276	1172	1211
Number of statistical outliers	46	47	35	37	69
Percentage of statistical outliers	5.1%	4.5%	2.7%	3.2%	5.7%

Table 8: comparison with previous proficiency tests

In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency tests was compared to the requirements of the reference test methods. The conclusions are given the following table.

Parameter	November 2023	November 2022	December 2021	December 2020	December 2019
Total Acid Number	(--)	+/-	-	-	-
Aniline Point	--	-	--	-	--
Asphaltenes	n.e.	n.e.	n.e.	n.e.	n.e.
Carbon Residue (micro method)	++	-	-	-	+
Density at 15 °C	+	+	+/-	+	+/-
Flash Point PMcc	+	+	+	+	+
Kinematic Viscosity at 50 °C	+	++	+	++	++
Kinematic Viscosity at 100 °C	++	++	++	++	++
Nitrogen	++	+	+	-	+

Parameter	November 2023	November 2022	December 2021	December 2020	December 2019
Pour Point Manual	+	-	+	++	+
Pour Point Automated 3 °C int.	-	-	+/-	+/-	-
Total Sulfur	++	-	+	+/-	+/-
Simulated Distillation	+	+	+	-	-
Distillation at 10 mmHg as AET	+	+	+	+/-	+/-
Aluminum as Al	(-)	+	+	+/-	+/-
Silicon as Si	(--)	--	--	--	-
Sum Aluminum and Silicon	+/-	+/-	--	-	+/-
Arsenic as As	n.e.	n.e.	n.e.	n.e.	--
Copper as Cu	++	n.e.	n.e.	+/-	++
Iron as Fe	++	++	+/-	+/-	+/-
Nickel as Ni	++	++	+	+	+
Sodium as Na	++	+	+	+/-	+/-
Vanadium as V	++	+	++	++	+/-
Calcium as Ca	+/-	n.e.	-	n.e.	+/-

Table 9: comparison of determinations to the reference test methods

For results between brackets no z-scores are calculated.



The following performance categories were used:

- ++ : group performed much better than the reference test method.
- + : group performed better than the reference test method.
- +/- : group performance equals the reference test method.
- : group performed worse than the reference test method.
- : group performed much worse than the reference test method.
- n.e. : not evaluated.

**APPENDIX 1**

**Determination of Total Acid Number on sample #23250; results in mg KOH/g**



lab	method	value	mark	z(targ)	End Point	volume	remarks
52	D664-A	<0.10		----	---	---	
62	D664-A	<0.10		----	Inflection Point	125 mL	
120		----		----	---	---	
140	D664-A	<0.10		----	Inflection Point	60 mL	
150	D664-A	<0.1	C	----	Inflection Point	125 mL	First reported 1.5
154		----		----	---	---	
158	D664-A	<0.10		----	Inflection Point	125 mL	
159		----		----	---	---	
171	D664-A	<0.10		----	---	---	
225		----		----	---	---	
311	D664-A	<0.10		----	Inflection Point	125 mL	
313		----		----	---	---	
323	D664-A	< 0.10		----	---	---	
333		----		----	---	---	
356		----		----	---	---	
381	D664-A	0.03		----	---	---	
445	D664-A	0.040		----	Inflection Point	60 mL	
494	D664-A	0.133	R(0.01)	----	Buffer End Point pH 10	60 mL	
608	D664-A	0.35	R(0.01)	----	Inflection Point	125 mL	
657	D664-A	0.03		----	Inflection Point	125 mL	
710		----		----	---	---	
736	D664-A	0.041		----	Inflection Point	125 mL	
750		----		----	---	---	
752	D664-A	0.0307		----	Inflection Point	60 mL	
753		----		----	---	---	
778		----		----	---	---	
781	D664-A	0.05		----	Inflection Point	125 mL	
785		----		----	---	---	
798		----		----	---	---	
823		----		----	---	---	
872		----		----	---	---	
873	D664-A	0.044		----	Buffer End Point pH 10	125 mL	
874	D664-A	0.050		----	Buffer End Point pH 10	125 mL	
875	D664-A	0.03		----	---	---	
994	D664-A	0.045		----	Inflection Point	60 mL	
995	D664-A	0.05		----	Inflection Point	125 mL	
1026	D664-A	0.015	R(0.05)	----	---	---	
1065	D664-A	0.018	R(0.05)	----	---	---	
1081		----		----	---	---	
1140	IP177	0.06		----	Inflection Point	60 mL	
1191	ISO6618	0.03		----	---	---	
1205		----		----	---	---	
1357		----		----	---	---	
1372		----		----	---	---	
1556	D664-A	0.042		----	---	---	
1677	D664-A	0.03		----	Inflection Point	125 mL	
1862	D664-A	0.0491	C	----	Inflection Point	125 mL	First reported 0.491
1950	D664-A	0.05		----	Inflection Point	125 mL	
1967	D664-A	0.012	R(0.05)	----	Buffer End Point pH 10	125 mL	
6024		----		----	---	---	
6025	D664-A	0.036		----	Inflection Point	125 mL	
6026	D664-A	0.0337		----	Inflection Point	125 mL	
6112	D664-A	0.04		----	---	---	
6195	D664-A	0.08	R(0.05)	----	Inflection Point	125 mL	
6203	D664-A	0.06		----	Inflection Point	125 mL	
6447		----		----	Inflection Point	125 mL	
					---	---	
	normality	OK					
	n	21					
	outliers	6					
	mean (n)	0.0415					
	st.dev. (n)	0.00981					
	R(calc.)	0.0275					
	st.dev.(D664-A:18e2 IP 125 mL)	(0.00283)					
	R(D664-A:18e2 IP 125 mL)	(0.0079)			Application range ASTM D664: 0.1 – 150 mg KOH/g		
Compare							
	R(D664-A:18e2 IP 60 mL)	(0.0296)					
	R(D664-A:18e2 BEP 125 mL)	(0.0111)					
	R(D664-A:18e2 BEP 60 mL)	(0.0248)					





## Determination of Aniline Point on sample #23250; results in °C

lab	method	value	mark	z(targ)	remarks
52	D611-E	88.55		-2.33	
62		----		----	
120		----		----	
140	D611-E	90.45		2.99	
150	D611-E	88.80		-1.63	
154		----		----	
158	D611-E	90.30	C	2.57	First reported 91.45
159		----		----	
171	D611-E	91.35		5.51	
225		----		----	
311	D611-E	88.45		-2.61	
313		----		----	
323	D611-A	89.80		1.17	
333		----		----	
356		----		----	
381		----		----	
445	D611-A	88.40		-2.75	
494		----		----	
608	D611-A	92.4	R(0.05)	8.45	
657	D611-B	88.65		-2.05	
710		----		----	
736	D611-E	88.75		-1.77	
750		----		----	
752		----		----	
753		----		----	
778		----		----	
781	D611-E	89.45		0.19	
785	D611-E	89.3		-0.23	
798		----		----	
823		----		----	
872		----		----	
873	D611-E	89.40		0.05	
874	D611-E	88.95		-1.21	
875	D611-E	89.65		0.75	
994		----		----	
995		----		----	
1026		----		----	
1065		----		----	
1081		----		----	
1140	D611-E	89.0		-1.07	
1191		----		----	
1205		----		----	
1357		----		----	
1372		----		----	
1556		----		----	
1677	D611-A	90.00		1.73	
1862	D611-B	90.10		2.01	
1950		----		----	
1967		----		----	
6024		----		----	
6025		----		----	
6026		----		----	
6112	D611-E	88.9		-1.35	
6195	D611-A	N/A		----	
6203		----		----	
6447		----		----	
	normality	OK			
	n	19			
	outliers	1			
	mean (n)	89.382			
	st.dev. (n)	0.7950			
	R(calc.)	2.226			
	st.dev.(D611:23)	0.3571			
	R(D611:23)	1			





Determination of Asphaltenes on sample #23250; results in %M/M

lab	method	value	mark	z(targ)	remarks
52	IP143	<0.05		----	
62		----		----	
120	IP143	0.02	C	----	First reported 0.25
140	IP143	<0.05		----	
150	D6560	<0.05		----	
154	IP143	<0.05		----	
158	D6560	<0.05		----	
159		----		----	
171	IP143	<0.05		----	
225		----		----	
311		----		----	
313		----		----	
323	IP143	< 0.50		----	
333		----		----	
356	IP143	Less 0.50		----	
381	INH-642	0.017		----	
445	IP143	0.06		----	
494	IP143	0.022		----	
608	D6560	0.04		----	
657	IP143	0.02		----	
710		----		----	
736	IP143	<0.5		----	
750		----		----	
752	INH-642	0.02036		----	
753		----		----	
778		----		----	
781	IP143	0.02		----	
785	IP143	0.05		----	
798		----		----	
823		----		----	
872		----		----	
873	IP143	Less 0.5		----	
874	IP143	<0.50		----	
875	IP143	< 0.50		----	
994	IP143	<0.5		----	
995	IP143	<0.5		----	
1026		----		----	
1065	D6560	0.02		----	
1081		----		----	
1140	IP143	0.37		----	
1191		----		----	
1205		----		----	
1357		----		----	
1372		----		----	
1556	IP143	0.03		----	
1677	IP143	0.04		----	
1862	IP143	0.04		----	
1950	IP143	<0.50		----	
1967	In house	0.0265	C	----	Reported 265 %M/M (probably unit error)
6024		----		----	
6025		----		----	
6026	In house	0.0294		----	
6112		----		----	
6195	IP143	0.32		----	
6203	IP143	0.01228		----	
6447		----		----	
				----	
n		33			
mean (n)		<0.5			Application range: IP143 : 0.50 – 30 %M/M



## Determination of Carbon Residue (micro method) on sample #23250; results in %M/M

lab	method	value	mark	z(targ)	remarks
52	D4530	0.13		-0.36	
62	D4530	0.19	C	0.78	First reported 0.43
120		----		----	
140	D4530	<0.10		----	
150	D4530	0.11		-0.74	
154	D4530	0.17		0.40	
158	D4530	0.11		-0.74	
159		----		----	
171	D4530	0.22	DG(0.05)	1.35	
225	D4530	0.13		-0.36	
311	D4530	0.13		-0.36	
313		----		----	
323	D4530	0.13		-0.36	
333		----		----	
356		----		----	
381		----		----	
445	D4530	0.17		0.40	
494	D4530	0.25	DG(0.05)	1.92	
608	D4530	0.15		0.02	
657	D4530	0.19		0.78	
710	ISO10370	0.17		0.40	
736	D4530	0.15		0.02	
750		----		----	
752		----		----	
753		----		----	
778		----		----	
781	D4530	0.122		-0.51	
785	D4530	0.12		-0.55	
798		----		----	
823		----		----	
872		----		----	
873	D4530	0.15		0.02	
874	D4530	0.15		0.02	
875	D4530	0.15		0.02	
994		----		----	
995		----		----	
1026		----		----	
1065		----		----	
1081	ISO10370	0.34	G(0.01)	3.63	
1140	D4530	0.17		0.40	
1191	ISO10370	0.15103		0.04	
1205		----		----	
1357		----		----	
1372		----		----	
1556	ISO10370	0.144		-0.09	
1677	D4530	0.19		0.78	
1862	D4530	0.139		-0.19	
1950	D4530	0.124		-0.47	
1967	D4530	0.1215		-0.52	
6024		----		----	
6025	D4530	0.136		-0.25	
6026	ISO10370	0.1456		-0.06	
6112		----		----	
6195	D4530	0.2		0.97	
6203	D4530	0.1752		0.50	
6447		----		----	
	normality	OK			
	n	29			
	outliers	3			
	mean (n)	0.1489			
	st.dev. (n)	0.02525			
	R(calc.)	0.0707			
	st.dev.(D4530:15R20)	0.05264			
	R(D4530:15R20)	0.1474			



Determination of Density at 15 °C on sample #23250; results in kg/m<sup>3</sup>



lab	method	value	mark	z(targ)	remarks
52	D4052	911.7		-0.84	
62	D4052	911.8		-0.66	
120	D4052	912.6		0.84	
140	D4052	912.0		-0.28	
150	D4052	912.4		0.46	
154	D4052	912.4		0.46	
158	D4052	912.4		0.46	
159		----		----	
171	ISO12185	912.4		0.46	
225	D4052	911.8		-0.66	
311	ISO12185	911.7		-0.84	
313	ISO12185	912.0		-0.28	
323	ISO12185	911.6		-1.03	
333	ISO12185	912.4		0.46	
356	ISO12185	911.7		-0.84	
381	D4052	909.3	R(0.01)	-5.32	
445	ISO12185	912.5		0.65	
494	ISO12185	913.1		1.77	
608	D4052	911.8		-0.66	
657	D4052	911.9		-0.47	
710	ISO12185	912.3		0.28	
736	D1298	911.7		-0.84	
750		----		----	
752	ISO3675	912.4		0.46	
753	ISO12185	912.4		0.46	
778	ISO12185	912.5		0.65	
781	ISO12185	912.5		0.65	
785	ISO12185	911.9		-0.47	
798		----		----	
823		----		----	
872	ISO12185	912.2		0.09	
873	ISO12185	912.2		0.09	
874	ISO12185	912.2		0.09	
875	D1298	912.0		-0.28	
994	ISO12185	911.7		-0.84	
995	ISO12185	912.2		0.09	
1026	D4052	880.2	R(0.01)	-59.64	
1065	D4052	913.2		1.96	
1081	D4052	911.6		-1.03	
1140	IP365	912.6		0.84	
1191	ISO12185	912.5		0.65	
1205		----		----	
1357		----		----	
1372		----		----	
1556	ISO12185	911.7		-0.84	
1677	D4052	911.6		-1.03	
1862	ISO12185	912.4		0.46	
1950	ISO12185	912.2		0.09	
1967	D1298	912.1		-0.10	
6024	D1298	912.5		0.65	
6025	ISO12185	912.3	C	0.28	Reported 0.9123 kg/m3
6026	ISO12185	912.6		0.84	
6112	ISO12185	912.3		0.28	
6195	D4052	910.8		-2.52	
6203	D4052	914.6	R(0.01)	4.57	
6447		----		----	
	normality	suspect			
	n	45			
	outliers	3			
	mean (n)	912.15			
	st.dev. (n)	0.435			
	R(calc.)	1.22			
	st.dev.(ISO12185:96)	0.536			
	R(ISO12185:96)	1.5			



## Determination of Flash Point PMcc on sample #23250; results in °C



lab	method	value	mark	z(targ)	remarks
52	D93-A	188.0		-1.14	
62	D93-B	>150		----	
120		----		----	
140	D93-B	>110.0		----	
150		----		----	
154	D93-B	>110		----	
158	D93-B	>110		----	
159		----		----	
171	D93-B	>110		----	
225		----		----	
311	D93-B	191.0		-0.30	
313	D93-B	194.0		0.54	
323	D93-B	189.5		-0.72	
333	D93-B	189.0		-0.86	
356	D93-B	194.0		0.54	
381		----		----	
445	D93-B	188.0		-1.14	
494	D93-B	201.0	R(0.05)	2.50	
608	D93-B	189		-0.86	
657	D93-B	192		-0.02	
710	D93-B	192.0		-0.02	
736	D93-B	191.0		-0.30	
750		----		----	
752	D93-B	190.13		-0.55	
753	D93-B	194.0		0.54	
778	D93-B	195.5		0.96	
781	D93-A	193.0		0.26	
785	D93-B	196.5		1.24	
798		----		----	
823		----		----	
872	D93-B	192.0		-0.02	
873	D93-B	193.0		0.26	
874	D93-B	194.0		0.54	
875	D93-B	191.0		-0.30	
994	D93-B	194.0		0.54	
995	D93-B	194.0		0.54	
1026	D93-A	196.0		1.10	
1065	D93-A	190		-0.58	
1081		189.0		-0.86	
1140	D93-B	>100.0		----	
1191	ISO2719	192.0		-0.02	
1205	D93-B	194.8		0.76	
1357		----		----	
1372		----		----	
1556	ISO2719	190.0		-0.58	
1677	D93-B	193.0		0.26	
1862	D93-B	192.0		-0.02	
1950	D93-B	188.0		-1.14	
1967	D93-B	191.5		-0.16	
6024	D93-B	195.5		0.96	
6025	D93-B	191.0		-0.30	
6026	D93-B	193.0		0.26	
6112		----		----	
6195	D93-B	130.0	R(0.01)	-17.38	
6203	D93-B	194.5		0.68	
6447		----		----	
	normality	OK			
	n	36			
	outliers	2			
	mean (n)	192.081			
	st.dev. (n)	2.3881			
	R(calc.)	6.687			
	st.dev.(D93-B:20)	3.5714			
	R(D93-B:20)	10			







Determination of Kinematic Viscosity at 50 °C on sample #23250; results in mm<sup>2</sup>/s

lab	method	value	mark	z(targ)	remarks
52	D445	38.01		0.08	
62	D445	36.71		-1.06	
120	D445	38.85		0.81	
140		----		----	
150		----		----	
154	D445	37.60		-0.28	
158	D445	37.79		-0.11	
159		----		----	
171	D445	39.12		1.05	
225	D445	37.75		-0.15	
311	D445	37.84		-0.07	
313	D445	37.85		-0.06	
323	D445	37.79		-0.11	
333	D445	37.81		-0.10	
356	D445	37.73		-0.17	
381	D445	37.77		-0.13	
445	D445	37.62		-0.26	
494	D445	37.726		-0.17	
608	D445	38.38		0.40	
657	D445	37.94		0.02	
710	D445	38.223		0.26	
736	D445	37.73		-0.17	
750		----		----	
752	D445	37.84		-0.07	
753	D445	37.75		-0.15	
778	D445	39.65		1.51	
781	D445	37.39		-0.46	
785	D445	38.94		0.89	
798		----		----	
823		----		----	
872	D445	37.84		-0.07	
873	D445	37.52		-0.35	
874	D445	37.91		-0.01	
875	D445	37.84		-0.07	
994	D445	38.08		0.14	
995		----		----	
1026	ISO3104	38.01		0.08	
1065	D445	37.59		-0.29	
1081		----		----	
1140	D445	36.62		-1.14	
1191	ISO3104	38.32		0.35	
1205	ISO3104	37.42575		-0.43	
1357		----		----	
1372		----		----	
1556	ISO3104	37.352		-0.50	
1677	D445	39.62		1.48	
1862	D445	38.02		0.09	
1950	D445	37.81		-0.10	
1967	D445	37.986		0.06	
6024	D445	37.59	C	-0.29	Test results Kinematic Viscosity 50°C and 100°C were mixed-up
6025	D445	38.05		0.11	
6026	ISO3104	37.59		-0.29	
6112	D445	37.96		0.03	
6195	D445	37.22		-0.61	
6203	D7042	38.294		0.33	
6447		----		----	
	normality	not OK			
	n	45			
	outliers	0			
	mean (n)	37.9212			
	st.dev. (n)	0.59077			
	R(calc.)	1.6541			
	st.dev.(D445:23)	1.14590			
	R(D445:23)	3.2085			





Determination of Kinematic Viscosity at 100 °C on sample #23250; results in mm<sup>2</sup>/s

lab	method	value	mark	z(targ)	remarks
52	D445	7.425		-0.05	
62	D445	7.508		0.21	
120	D7042	7.378		-0.20	
140		----		----	
150		----		----	
154		----		----	
158		----		----	
159		----		----	
171	D445	7.515		0.23	
225	D445	7.339		-0.32	
311	D445	7.485		0.13	
313	D445	7.342		-0.31	
323	D445	7.449		0.02	
333	D445	7.440		-0.01	
356		----		----	
381	D445	7.417		-0.08	
445	D445	7.406		-0.11	
494	D445	7.387		-0.17	
608	D445	7.51		0.21	
657	D445	7.461		0.06	
710	D445	7.4903		0.15	
736	D445	7.493		0.16	
750		----		----	
752	D445	7.489		0.15	
753	D445	7.525		0.26	
778		----		----	
781	D445	7.482		0.12	
785	D445	7.395		-0.15	
798		----		----	
823		----		----	
872	D445	7.489		0.15	
873	D445	7.491		0.15	
874	D445	7.528		0.27	
875	D445	7.509		0.21	
994	D445	7.542		0.31	
995		----		----	
1026	D445	7.475		0.10	
1065	D445	7.3826		-0.19	
1081		----		----	
1140	D445	7.253		-0.59	
1191	ISO3104	7.4412		0.00	
1205	ISO3104	7.469745		0.09	
1357		----		----	
1372		----		----	
1556		----		----	
1677	D445	7.475		0.10	
1862	D445	7.376		-0.21	
1950	D445	7.290		-0.48	
1967	D445	7.4839		0.13	
6024	D445	7.450	C	0.02	Test results Kinematic Viscosity 50°C and 100°C were mixed-up
6025	D445	7.398		-0.14	
6026	ISO3104	7.490		0.15	
6112		----		----	
6195	D445	7.380		-0.19	
6203	D7042	7.389		-0.17	
6447		----		----	
	normality	OK			
	n	39			
	outliers	0			
	mean (n)	7.4423			
	st.dev. (n)	0.06752			
	R(calc.)	0.1890			
	st.dev.(D445:23)	0.32055			
	R(D445:23)	0.8975			



## Determination of Nitrogen on sample #23250; results in mg/kg



lab	method	value	mark	z(targ)	remarks
52	D4629	1100		0.29	
62		----		----	
120		----		----	
140		----		----	
150	D5762 Gravimetric	1200		1.27	
154		----		----	
158		----		----	
159		----		----	
171	D5762 Gravimetric	1100		0.29	
225		----		----	
311	D5762 Volumetric	1140		0.68	
313		----		----	
323		----		----	
333	D5762 Gravimetric	1100		0.29	
356		----		----	
381		----		----	
445	D5762 Gravimetric	989.3		-0.80	
494	D5762 Gravimetric	928		-1.40	
608	D5762 Gravimetric	880.54		-1.87	
657	D5762 Gravimetric	1110.6		0.39	
710		----		----	
736	D5762 Volumetric	1100		0.29	
750		----		----	
752		----		----	
753		----		----	
778		----		----	
781	D5762 Gravimetric	1108		0.36	
785		----		----	
798		----		----	
823		----		----	
872		----		----	
873	D4629	990		-0.80	
874	D5762	1000		-0.70	
875	D5762 Volumetric	1020		-0.50	
994	D5762 Volumetric	1103		0.32	
995		----		----	
1026	D4629	1135		0.63	
1065	D4629	1100		0.29	
1081	D4629	870		-1.97	
1140	D4629	1252.2		1.78	
1191	D5762 Gravimetric	1091.1		0.20	
1205		----		----	
1357		----		----	
1372		----		----	
1556	D5762 Gravimetric	1087		0.16	
1677	D5762 Gravimetric	1120		0.48	
1862	D5762 Gravimetric	1109		0.37	
1950		----		----	
1967		----		----	
6024		----		----	
6025	D5762 Volumetric	1010		-0.60	
6026	D5762 Gravimetric	1100		0.29	
6112		----		----	
6195	D4629	1000		-0.70	
6203	D5762 Gravimetric	0.25	R(0.01)	-10.52	
6447	D5762 Volumetric	1170		0.97	
	normality	OK			
	n	27			
	outliers	1			
	mean (n)	1070.88			
	st.dev. (n)	89.531			
	R(calc.)	250.69			
	st.dev.(D5762:18a)	101.734			
	R(D5762:18a)	284.85			



Determination of Pour Point Manual on sample #23250; results in °C



lab	method	value	mark	z(targ)	remarks
52	D97	39		0.27	
62	D97	36		-0.66	
120	D97	30	R(0.01)	-2.53	
140	D97	39		0.27	
150	D97	39		0.27	
154	D97	39		0.27	
158	D97	39		0.27	
159		----		----	
171	D97	33		-1.60	
225	D97	36		-0.66	
311	D97	39		0.27	
313		----		----	
323	D97	36		-0.66	
333		----		----	
356		----		----	
381	ISO3016	42		1.20	
445	D97	39		0.27	
494		----		----	
608	D97	39		0.27	
657	D97	42		1.20	
710	D97	39		0.27	
736	D97	39		0.27	
750		----		----	
752	D97	39		0.27	
753	D97	39		0.27	
778	D97	39		0.27	
781	D97	39		0.27	
785	D97	39.0		0.27	
798		----		----	
823		----		----	
872	D97	36		-0.66	
873	D97	39		0.27	
874	D97	39		0.27	
875	D97	39		0.27	
994	D97	39		0.27	
995		----		----	
1026		----		----	
1065		----		----	
1081		----		----	
1140		----		----	
1191	ISO3016	36		-0.66	
1205		----		----	
1357		----		----	
1372		----		----	
1556	ISO3016	39		0.27	
1677	D97	39		0.27	
1862	D97	36		-0.66	
1950	D97	39		0.27	
1967	D97	36		-0.66	
6024	D97	36		-0.66	
6025	D97	36		-0.66	
6026	D97	36		-0.66	
6112	D97	39		0.27	
6195	D97	36		-0.66	
6203	D97	39		0.27	
6447		----		----	
	normality	OK			
	n	38			
	outliers	1			
	mean (n)	38.13			
	st.dev. (n)	1.833			
	R(calc.)	5.13			
	st.dev.(D97:17bR22)	3.214			
	R(D97:17bR22)	9			







Determination of Pour Point Automated 3 °C interval on sample #23250; results in °C

lab	method	value	mark	z(targ)	remarks
52		----		----	
62		----		----	
120		----		----	
140		----		----	
150		----		----	
154		----		----	
158		----		----	
159		----		----	
171	D5950	39		1.15	
225		----		----	
311		----		----	
313		----		----	
323		----		----	
333	D5950	36		-0.23	
356		----		----	
381		----		----	
445		----		----	
494	D5950	33		-1.61	
608		----		----	
657	D5950	39		1.15	
710		----		----	
736		----		----	
750		----		----	
752		----		----	
753		----		----	
778		----		----	
781	D5950	39		1.15	
785		----		----	
798		----		----	
823		----		----	
872		----		----	
873		----		----	
874		----		----	
875		----		----	
994		----		----	
995		----		----	
1026	D5950	39		1.15	
1065	D5950	30		-2.98	
1081	D5950	35		-0.69	
1140	D5950	36		-0.23	
1191	D5950	39		1.15	
1205		----		----	
1357		----		----	
1372		----		----	
1556		----		----	
1677		----		----	
1862		----		----	
1950		----		----	
1967		----		----	
6024		----		----	
6025		----		----	
6026		----		----	
6112		----		----	
6195		----		----	
6203		----		----	
6447		----		----	
	normality	OK			
	n	10			
	outliers	0			
	mean (n)	36.50			
	st.dev. (n)	3.136			
	R(calc.)	8.78			
	st.dev.(D5950:14R20)	2.179			
	R(D5950:14R20)	6.1			



## Determination of Total Sulfur on sample #23250; results in %M/M

lab	method	value	mark	z(targ)	remarks
52	D4294	0.0434		0.77	
62	D4294	0.0465	R(0.05)	1.71	
120		----		----	
140	D4294	0.0442		1.01	
150	D4294	0.0397		-0.35	
154	D4294	0.0420		0.34	
158		----		----	
159		----		----	
171	D4294	0.0445		1.10	
225		----		----	
311		----		----	
313	D4294	0.054	R(0.01)	3.98	
323	D4294	0.039		-0.56	
333	D4294	0.0404		-0.14	
356	D4294	0.040		-0.26	
381	ISO8754	0.04		-0.26	
445	D4294	0.0430		0.65	
494	ISO8754	0.05	R(0.01)	2.76	
608	D4294	0.038		-0.87	
657	D4294	0.0403		-0.17	
710	D4294	0.040		-0.26	
736	D4294	0.0402		-0.20	
750		----		----	
752	D4294	0.0411		0.07	
753	D4294	0.0386		-0.69	
778	D4294	0.0421		0.37	
781	D4294	0.0416		0.22	
785	D4294	0.0406		-0.08	
798		----		----	
823		----		----	
872	D4294	0.0411		0.07	
873	D4294	0.0401		-0.23	
874	D4294	0.0420		0.34	
875	D4294	0.0393		-0.47	
994	D4294	0.0403		-0.17	
995	D4294	0.039		-0.56	
1026	D2622	0.042		0.34	
1065	D4294	0.0400		-0.26	
1081	D4294	0.0494	C,R(0.01)	2.58	First reported 0.494
1140	IP336	0.05	R(0.01)	2.76	
1191	ISO8754	0.04		-0.26	
1205		----		----	
1357		----		----	
1372		----		----	
1556	ISO8754	0.0398		-0.32	
1677	D4294	0.0407	C	-0.05	First reported 0.407
1862	D4294	0.0414		0.16	
1950	D4294	0.0408		-0.02	
1967	D4294	0.0422		0.40	
6024		----		----	
6025	D4294	0.041		0.04	
6026	D4294	0.0408		-0.02	
6112		----		----	
6195	D4294	0.042		0.34	
6203	D4294	0.0404		-0.14	
6447	D2622	0.0413		0.13	
	normality	OK			
	n	38			
	outliers	5			
	mean (n)	0.04087			
	st.dev. (n)	0.001434			
	R(calc.)	0.00402			
	st.dev.(D4294:21)	0.003304			
	R(D4294:21)	0.00925			



Determination of Simulated Distillation on sample #23250; result in °C


lab	method	IBP	10%	30%	50%	70%	90%	FBP
52	D6352	282.0	362.5	407.0	436.5	470.5	522.5	603.5
62		----	----	----	----	----	----	----
120		----	----	----	----	----	----	----
140		----	----	----	----	----	----	----
150	D6352	261.5	360.0	407.0	438.0	473.5	525.5	607.5
154		----	----	----	----	----	----	----
158		----	----	----	----	----	----	----
159		----	----	----	----	----	----	----
171	D6352	261.5	362.0	408.5	439.5	474.5	527.5	619.5
225		----	----	----	----	----	----	----
311	D6352	262.5	358.5	405.0	435.5	471.0	522.0	596.5
313		----	----	----	----	----	----	----
323	D6352	202.8 G(1)	383.9 G(1)	433.3 G(1)	463.9 G(1)	494.2 G(1)	538.7 G(1)	603.8 ex
333		----	----	----	----	----	----	----
356		----	----	----	----	----	----	----
381		----	----	----	----	----	----	----
445	D7169	266.0	361.5	407.5	438.5	474.5	525.5	603.0
494		----	----	----	----	----	----	----
608		----	----	----	----	----	----	----
657	D6352	262.2	360.6	407.2	438.2	473.5	523.9	604.2
710		----	----	----	----	----	----	----
736		----	----	----	----	----	----	----
750		----	----	----	----	----	----	----
752		----	----	----	----	----	----	----
753		----	----	----	----	----	----	----
778		----	----	----	----	----	----	----
781	D6352	259.7	360.4	407.2	437.4	473.7	524.2	597.8
785		----	----	----	----	----	----	----
798		----	----	----	----	----	----	----
823		----	----	----	----	----	----	----
872		----	----	----	----	----	----	----
873		----	----	----	----	----	----	----
874		----	----	----	----	----	----	----
875		----	----	----	----	----	----	----
994		----	----	----	----	----	----	----
995		----	----	----	----	----	----	----
1026		----	----	----	----	----	----	----
1065		----	----	----	----	----	----	----
1081	IP480	259	362	409	440	476	529 C	609
1140	D7169	260.8	361.5	409.4	441.1	476.0	526.3	612.3
1191		----	----	----	----	----	----	----
1205		----	----	----	----	----	----	----
1357		----	----	----	----	----	----	----
1372		----	----	----	----	----	----	----
1556	ISO3924	265.7	370.9 G(1)	416.0 G(1)	445.9 G(5)	478.7	527.9	592.3
1677		----	----	----	----	----	----	----
1862		----	----	----	----	----	----	----
1950		----	----	----	----	----	----	----
1967		----	----	----	----	----	----	----
6024		----	----	----	----	----	----	----
6025		----	----	----	----	----	----	----
6026		----	----	----	----	----	----	----
6112		----	----	----	----	----	----	----
6195		----	----	----	----	----	----	----
6203	D6352	262.5	361.5	407.0	438.0	473.5	526.5	602.5
6447		----	----	----	----	----	----	----
	normality	not OK	suspect	OK	OK	OK	OK	OK
	n	11	10	10	10	11	11	11
	outliers	1	2	2	2	1	1	0 + 1ex
	mean (n)	263.95	361.05	407.48	438.27	474.13	525.53	604.37
	st.dev. (n)	6.363	1.190	1.249	1.643	2.291	2.217	7.616
	R(calc.)	17.82	3.33	3.50	4.60	6.42	6.21	21.32
	st.dev.(D6352:19e1)	17.536	2.536	2.107	2.286	2.571	3.75	13.607
	R(D6352:19e1)	49.1	7.1	5.9	6.4	7.2	10.5	38.1


Lab 1081 first reported 553

Lab 323: ex = test result excluded as 4 or more test results are statistical outliers.

z-scores for Simulated Distillation

lab	IBP	10%	30%	50%	70%	90%	FBP
52	1.03	0.57	-0.23	-0.77	-1.41	-0.81	-0.06
62	----	----	----	----	----	----	----
120	----	----	----	----	----	----	----
140	----	----	----	----	----	----	----
150	-0.14	-0.41	-0.23	-0.12	-0.24	-0.01	0.23
154	----	----	----	----	----	----	----
158	----	----	----	----	----	----	----
159	----	----	----	----	----	----	----
171	-0.14	0.37	0.48	0.54	0.14	0.53	1.11
225	----	----	----	----	----	----	----
311	-0.08	-1.01	-1.18	-1.21	-1.22	-0.94	-0.58
313	----	----	----	----	----	----	----
323	-3.49	9.01	12.25	11.21	7.81	3.51	-0.04
333	----	----	----	----	----	----	----
356	----	----	----	----	----	----	----
381	----	----	----	----	----	----	----
445	0.12	0.18	0.01	0.10	0.14	-0.01	-0.10
494	----	----	----	----	----	----	----
608	----	----	----	----	----	----	----
657	-0.10	-0.18	-0.13	-0.03	-0.24	-0.43	-0.01
710	----	----	----	----	----	----	----
736	----	----	----	----	----	----	----
750	----	----	----	----	----	----	----
752	----	----	----	----	----	----	----
753	----	----	----	----	----	----	----
778	----	----	----	----	----	----	----
781	-0.24	-0.26	-0.13	-0.38	-0.17	-0.35	-0.48
785	----	----	----	----	----	----	----
798	----	----	----	----	----	----	----
823	----	----	----	----	----	----	----
872	----	----	----	----	----	----	----
873	----	----	----	----	----	----	----
874	----	----	----	----	----	----	----
875	----	----	----	----	----	----	----
994	----	----	----	----	----	----	----
995	----	----	----	----	----	----	----
1026	----	----	----	----	----	----	----
1065	----	----	----	----	----	----	----
1081	-0.28	0.37	0.72	0.76	0.73	0.93	0.34
1140	-0.18	0.18	0.91	1.24	0.73	0.21	0.58
1191	----	----	----	----	----	----	----
1205	----	----	----	----	----	----	----
1357	----	----	----	----	----	----	----
1372	----	----	----	----	----	----	----
1556	0.10	3.88	4.04	3.34	1.78	0.63	-0.89
1677	----	----	----	----	----	----	----
1862	----	----	----	----	----	----	----
1950	----	----	----	----	----	----	----
1967	----	----	----	----	----	----	----
6024	----	----	----	----	----	----	----
6025	----	----	----	----	----	----	----
6026	----	----	----	----	----	----	----
6112	----	----	----	----	----	----	----
6195	----	----	----	----	----	----	----
6203	-0.08	0.18	-0.23	-0.12	-0.24	0.26	-0.14
6447	----	----	----	----	----	----	----









Determination of Distillation at 10 mmHg as AET on sample #23250; result in °C

lab	method	IBP	10%	30%	50%	70%	90%	FBP
52		----	----	----	----	----	----	----
62		----	----	----	----	----	----	----
120		----	----	----	----	----	----	----
140		----	----	----	----	----	----	----
150	D1160	333 DG(5)	392	425	452	483	536	596 G(1)
154		----	----	----	----	----	----	----
158		----	----	----	----	----	----	----
159		----	----	----	----	----	----	----
171	D1160	298	385	421	443	473	517	540
225		----	----	----	----	----	----	----
311	D1160	304	380	418	445	478	527	555
313		----	----	----	----	----	----	----
323	D1160	309	379	417	447	480	534	558
333		----	----	----	----	----	----	----
356	D1160	276	385	414	448	476	524	553
381		----	----	----	----	----	----	----
445		----	----	----	----	----	----	----
494	D1160	95.6 G(1)	376.5	416.4	444.4	476.9	531.8	554.8
608		----	----	----	----	----	----	----
657		----	----	----	----	----	----	----
710	D1160	291.0	372.5	418.0	442.5	474.0	522.0	555.0
736	D1160	297.8	384.3	418.5	445.5	474.5	524.4	550.4
750		----	----	----	----	----	----	----
752		----	----	----	----	----	----	----
753		----	----	----	----	----	----	----
778		----	----	----	----	----	----	----
781	D1160	283	382	420	443	478	524	539
785	D1160	271.1	379.7	419.1	447.0	481.1	532.4	----
798		----	----	----	----	----	----	----
823		----	----	----	----	----	----	----
872		----	----	----	----	----	----	----
873	D1160	283	380	419	445	475	526	543
874	D1160	279.1	374.2	418.7	442.0	475.0	526.0	543.0
875	D1160	260.9	372.1	415.1	445.2	477.0	528.8	539.0
994	D1160	284.0	383.0	420.0	445.0	472.0	515.0	546.0
995	D1160	277.0	380.5	418.0	442.0	468.0	510.5	539.0
1026		----	----	----	----	----	----	----
1065		----	----	----	----	----	----	----
1081		----	----	----	----	----	----	----
1140		----	----	----	----	----	----	----
1191		----	----	----	----	----	----	----
1205		----	----	----	----	----	----	----
1357		----	----	----	----	----	----	----
1372		----	----	----	----	----	----	----
1556		----	----	----	----	----	----	----
1677	D1160	295	386	420	446	478	527	555
1862	D1160	288	384	423	449	481	533	550
1950	D1160	272	378	417	442	472	520	548
1967	D1160	270	380	419	443	477	530	549
6024	D1160	258	381	424	450	480	532	550
6025	D1160	268.0	384.0	423.0	443.0	478.0	530.0	549.0
6026	D1160	267	382	419	442	477	530	549
6112	D1160	338.0 DG(5)	383.8	418.3	445.4	476.3	525.5	554.2
6195	D1160	303.8	382.2	421.0	449.7	481.8	533.5	----
6203		----	----	----	----	----	----	----
6447		----	----	----	----	----	----	----
	normality	OK	OK	OK	OK	OK	OK	OK
	n	21	24	24	24	24	24	21
	outliers	3	0	0	0	0	0	1
	mean (n)	282.65	381.12	419.25	445.28	476.77	526.66	548.54
	st.dev. (n)	14.750	4.461	2.648	2.854	3.516	6.366	6.034
	R(calc.)	41.30	12.49	7.41	7.99	9.85	17.83	16.90
	st.dev.(D1160:18)	17.660	6.179	4.033	3.856	3.438	3.674	9.605
	R(D1160:18)	49.45	17.30	11.29	10.80	9.63	10.29	26.89

z-scores Distillation at 10 mmHg as AET

lab	IBP	10%	30%	50%	70%	90%	FBP
52	----	----	----	----	----	----	----
62	----	----	----	----	----	----	----
120	----	----	----	----	----	----	----
140	----	----	----	----	----	----	----
150	2.85	1.76	1.42	1.74	1.81	2.54	4.94
154	----	----	----	----	----	----	----
158	----	----	----	----	----	----	----
159	----	----	----	----	----	----	----
171	0.87	0.63	0.43	-0.59	-1.10	-2.63	-0.89
225	----	----	----	----	----	----	----
311	1.21	-0.18	-0.31	-0.07	0.36	0.09	0.67
313	----	----	----	----	----	----	----
323	1.49	-0.34	-0.56	0.45	0.94	2.00	0.98
333	----	----	----	----	----	----	----
356	-0.38	0.63	-1.30	0.71	-0.23	-0.72	0.46
381	----	----	----	----	----	----	----
445	----	----	----	----	----	----	----
494	-10.59	-0.75	-0.71	-0.23	0.04	1.40	0.65
608	----	----	----	----	----	----	----
657	----	----	----	----	----	----	----
710	0.47	-1.39	-0.31	-0.72	-0.81	-1.27	0.67
736	0.86	0.52	-0.19	0.06	-0.66	-0.62	0.19
750	----	----	----	----	----	----	----
752	----	----	----	----	----	----	----
753	----	----	----	----	----	----	----
778	----	----	----	----	----	----	----
781	0.02	0.14	0.18	-0.59	0.36	-0.72	-0.99
785	-0.65	-0.23	-0.04	0.45	1.26	1.56	----
798	----	----	----	----	----	----	----
823	----	----	----	----	----	----	----
872	----	----	----	----	----	----	----
873	0.02	-0.18	-0.06	-0.07	-0.52	-0.18	-0.58
874	-0.20	-1.12	-0.14	-0.85	-0.52	-0.18	-0.58
875	-1.23	-1.46	-1.03	-0.02	0.07	0.58	-0.99
994	0.08	0.30	0.18	-0.07	-1.39	-3.17	-0.26
995	-0.32	-0.10	-0.31	-0.85	-2.55	-4.40	-0.99
1026	----	----	----	----	----	----	----
1065	----	----	----	----	----	----	----
1081	----	----	----	----	----	----	----
1140	----	----	----	----	----	----	----
1191	----	----	----	----	----	----	----
1205	----	----	----	----	----	----	----
1357	----	----	----	----	----	----	----
1372	----	----	----	----	----	----	----
1556	----	----	----	----	----	----	----
1677	0.70	0.79	0.18	0.19	0.36	0.09	0.67
1862	0.30	0.47	0.93	0.96	1.23	1.73	0.15
1950	-0.60	-0.50	-0.56	-0.85	-1.39	-1.81	-0.06
1967	-0.72	-0.18	-0.06	-0.59	0.07	0.91	0.05
6024	-1.40	-0.02	1.18	1.22	0.94	1.45	0.15
6025	-0.83	0.47	0.93	-0.59	0.36	0.91	0.05
6026	-0.89	0.14	-0.06	-0.85	0.07	0.91	0.05
6112	3.13	0.43	-0.24	0.03	-0.14	-0.32	0.59
6195	1.20	0.18	0.43	1.15	1.46	1.86	----
6203	----	----	----	----	----	----	----
6447	----	----	----	----	----	----	----








## Determination of Aluminum as Al, Silicon as Si and sum Al and Si on sample #23251; results in mg/kg

lab	method	Al	mark	z(targ)	Si	mark	z(targ)	Sum Al+Si	mark	z(targ)
52	IP501	<5		----	<10		----	<15		----
62	IP501	<5		----	<10		----	5		----
120		----		----	----		----	----		----
150	IP501	<5		----	<10		----	<15		----
154	IP501	<5		----	<10	C	----	<15	C	----
158	IP501	<5		----	<10		----	----		----
171		----		----	----		----	----		----
225	IP501	4		----	2		----	6		----
311		----		----	----		----	----		----
323	IP501	< 5		----	< 10		----	< 15		----
333	IP501	5.6		----	1.6		----	7.2		----
381	D5184	2.9		----	----		----	----		----
445	IP501	5.0		----	5.0	DG(5)	----	10.0	G(5)	----
494	IP501	4		----	2		----	6		----
608		----		----	----		----	----		----
657	IP501	5.2		----	<10		----	5.2		----
736	IP501	3.3		----	1.6		----	4.9		----
752		----		----	----		----	----		----
781	IP501	<5		----	<10		----	<15		----
785	IP470	3.5		----	2.1		----	5.6		----
823		----		----	----		----	----		----
873		----		----	----		----	----		----
874	IP501	4		----	2		----	6		----
875	IP501	4		----	2		----	6		----
994	IP501	3.9		----	1.4		----	5.3		----
995	IP470	3.6		----	1.4		----	----		----
1026	IP501	4		----	2		----	6		----
1065		----		----	----		----	----		----
1140	IP501	5.457		----	2.898		----	8.355		----
1191	IP501	3.43		----	2.06		----	5.49		----
1357	IP501	4.3		----	2.4		----	6.7		----
1556		----		----	----		----	----		----
1677	IP501	4		----	4	DG(5)	----	8		----
1862	IP501	3.66		----	<10		----	<15		----
1950	IP470	4.3		----	2.1		----	6.4		----
1967	IP470	3.99		----	2.67		----	6.66		----
6024	IP470	4.2		----	3.8	DG(5)	----	8.0		----
6026	IP470	4.6		----	3.6	DG(5)	----	8.2		----
6195	IP501	4.4		----	2.8		----	7.2		----
6203	IP501	5.085		----	1.8		----	----		----
	normality	OK			OK			OK		
	n	24			17			20		
	outliers	0			4			1		
	mean (n)	4.18			2.05			6.41		
	st.dev. (n)	0.685			0.441			1.094		
	R(calc.)	1.92			1.24			3.06		
	st.dev.(IP501:05R19)	(0.504)			(0.243)			(0.559)		
	R(IP501:05R19)	(1.41)			(0.68)			(1.57)		
	Application range	5 – 150 mg/kg			10 – 250 mg/kg					

Lab 154: First reported 12 and 15





## Determination of Arsenic as As on sample #23251; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52		----		----	
62		----		----	
120		----		----	
150		----		----	
154		----		----	
158		----		----	
171		----		----	
225		----		----	
311		----		----	
323		----		----	
333		----		----	
381	INH-118	5.1		----	
445		----		----	
494		----		----	
608		----		----	
657		----		----	
736		----		----	
752		----		----	
781		----		----	
785		----		----	
823		----		----	
873		----		----	
874		----		----	
875		----		----	
994		----		----	
995		----		----	
1026		----		----	
1065		----		----	
1140		----		----	
1191		----		----	
1357		----		----	
1556		----		----	
1677	IP621	3.8		----	
1862		----		----	
1950		----		----	
1967		----		----	
6024		----		----	
6026		----		----	
6195		----		----	
6203	In house	4.52		----	
	n	3			
	mean (n)	<10			

Determination of Copper as Cu on sample #23251; results in mg/kg



lab	method	value	mark	z(targ)	remarks
52		----		----	
62		----		----	
120		----		----	
150		----		----	
154		----		----	
158	IP501	1		-0.58	
171		----		----	
225		----		----	
311	IP621	1.6		0.17	
323	IP501	1		-0.58	
333		----		----	
381	INH-118	1.6		0.17	
445	IP621	1.6		0.17	
494	IP501	1		-0.58	
608		----		----	
657	IP501	<1		----	
736	IP501	1.4		-0.08	
752		----		----	
781	IP621	1.5		0.04	
785	IP470	1.5		0.04	
823		----		----	
873	IP621	1.6		0.17	
874	IP621	2.0		0.66	
875		----		----	
994	IP501	1.5		0.04	
995	IP621	1.44		-0.03	
1026		----		----	
1065		----		----	
1140		----		----	
1191	In house	1.55		0.10	
1357	IP501	1.2		-0.33	
1556	IP621	1.57		0.13	
1677	IP621	1.5		0.04	
1862	IP621	1.45		-0.02	
1950	IP621	1.5		0.04	
1967	IP470	1.43		-0.05	
6024		----		----	
6026	IP470	1.5		0.04	
6195	IP501	1.6		0.17	
6203	IP501	1.71		0.30	

normality suspect  
n 23  
outliers 0  
mean (n) 1.467  
st.dev. (n) 0.2324  
R(calc.) 0.651  
st.dev.(IP621:16) 0.8033  
R(IP621:16) 2.249



Determination of Iron as Fe on sample #23251; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	IP501	3		-0.32	
62	IP501	3		-0.32	
120		----		----	
150	IP501	4		0.40	
154	IP501	6	G(0.05)	1.84	
158	IP501	3		-0.32	
171		----		----	
225	IP501	4		0.40	
311	IP621	3.1		-0.24	
323	IP501	4		0.40	
333	IP501	3.6		0.12	
381		----		----	
445	IP621	3.7		0.19	
494	IP501	4		0.40	
608		----		----	
657	IP501	3.0		-0.32	
736	IP501	3.1		-0.24	
752		----		----	
781	IP621	3.1		-0.24	
785	IP470	3.6		0.12	
823		----		----	
873	IP621	3.8		0.26	
874	IP621	3.9		0.33	
875	IP501	3		-0.32	
994	IP501	3.3		-0.10	
995	IP621	3.9		0.33	
1026	IP501	4		0.40	
1065	IP470	3.5		0.04	
1140	IP501	5.554	G(0.05)	1.52	
1191	ISO10478	2.55		-0.64	
1357	IP501	3.7		0.19	
1556	IP621	3.01		-0.31	
1677	IP621	3.3		-0.10	
1862	IP501	3.19		-0.18	
1950	IP621	3.2		-0.17	
1967	IP470	3.53		0.07	
6024	IP470	3		-0.32	
6026	IP470	3.4		-0.03	
6195	IP501	4.1		0.48	
6203	IP501	5.205	G(0.01)	1.27	
normality		OK			
n		31			
outliers		3			
mean (n)		3.438			
st.dev. (n)		0.4243			
R(calc.)		1.188			
st.dev.(IP621:16)		1.3904			
R(IP621:16)		3.893			







Determination of Nickel as Ni on sample #23251; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	IP501	13		0.22	
62	IP501	11		-0.47	
120		----		----	
150	IP501	11		-0.47	
154	IP501	11		-0.47	
158	IP501	10		-0.82	
171		----		----	
225	IP501	13		0.22	
311	IP621	12.2		-0.06	
323	IP501	12		-0.13	
333	IP501	12.9		0.18	
381	D7111	12.1		-0.09	
445	IP621	12.6		0.08	
494	IP501	13		0.22	
608		----		----	
657	IP501	13.2		0.29	
736	IP501	12.0		-0.13	
752		----		----	
781	IP621	12.0		-0.13	
785	IP470	12.4		0.01	
823		----		----	
873	IP621	12.8		0.15	
874	IP621	12.9		0.18	
875	IP501	13		0.22	
994	IP501	11.5		-0.30	
995	IP621	12.7		0.11	
1026	IP501	15		0.91	
1065	IP470	14.6		0.77	
1140	IP501	13.74		0.47	
1191	ISO10478	9.22		-1.09	
1357	IP501	10.4		-0.68	
1556	IP621	12.81		0.15	
1677	IP621	12.1		-0.09	
1862	IP501	12.08		-0.10	
1950	IP621	12.6		0.08	
1967	IP470	12.74		0.13	
6024	IP470	10		-0.82	
6026	IP470	13		0.22	
6195	IP501	15.8		1.18	
6203	IP501	12.705		0.11	



normality suspect  
n 35  
outliers 0  
mean (n) 12.374  
st.dev. (n) 1.3472  
R(calc.) 3.772  
st.dev.(IP621:16) 2.8973  
R(IP621:16) 8.112



Determination of Sodium as Na on sample #23251; results in mg/kg



lab	method	value	mark	z(targ)	remarks
52	IP501	3		0.06	
62	IP501	2		-0.63	
120		----		----	
150	IP501	5	R(0.01)	1.45	
154	IP501	4		0.76	
158	IP501	3		0.06	
171		----		----	
225	IP501	2		-0.63	
311	IP621	2.6		-0.21	
323	IP501	5	R(0.01)	1.45	
333	IP501	3.6		0.48	
381	D7111	3.0		0.06	
445	IP621	6.6	R(0.01)	2.56	
494	IP501	3		0.06	
608		----		----	
657	IP501	3.1		0.13	
736	IP501	2.5		-0.28	
752		----		----	
781	IP621	2.5		-0.28	
785	IP470	2.5		-0.28	
823		----		----	
873	IP621	2.7		-0.15	
874	IP621	2.8		-0.08	
875	IP501	3		0.06	
994	IP501	2.5		-0.28	
995	IP621	2.9		-0.01	
1026	IP501	3		0.06	
1065	IP470	4.0		0.76	
1140	IP501	3.191		0.20	
1191	ISO10478	2.72		-0.13	
1357	IP501	3.7		0.55	
1556	IP621	3.11		0.14	
1677	IP621	3.0		0.06	
1862	IP501	2.82		-0.06	
1950	IP621	2.7		-0.15	
1967	IP470	2.75		-0.11	
6024	IP470	3		0.06	
6026	IP470	2.8		-0.08	
6195	IP501	2.9		-0.01	
6203	IP501	2.7		-0.15	

normality suspect  
n 32  
outliers 3  
mean (n) 2.909  
st.dev. (n) 0.4525  
R(calc.) 1.267  
st.dev.(IP621:16) 1.4417  
R(IP621:16) 4.037





Determination of Vanadium as V on sample #23251; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	IP501	12		0.06	
62	IP501	11		-0.53	
120		----		----	
150	IP501	11		-0.53	
154	IP501	12		0.06	
158	IP501	11		-0.53	
171		----		----	
225	IP501	12		0.06	
311	IP621	11.5		-0.23	
323	IP501	12		0.06	
333	IP501	11.4		-0.29	
381	D7111	12.6		0.42	
445	IP621	11.5		-0.23	
494	IP501	12		0.06	
608		----		----	
657	IP501	13.1		0.71	
736	IP501	10.5		-0.82	
752		----		----	
781	IP621	11.6		-0.17	
785	IP470	12.5		0.36	
823		----		----	
873	IP621	12.0		0.06	
874	IP621	11.8		-0.06	
875	IP501	12		0.06	
994	IP501	11.7		-0.11	
995	IP621	12.2		0.18	
1026	IP501	12		0.06	
1065	IP470	12.1		0.12	
1140	IP501	13.13		0.73	
1191	ISO10478	10.48		-0.83	
1357	IP501	11.6		-0.17	
1556	IP621	12.11		0.13	
1677	IP621	12.0		0.06	
1862	IP501	12.22		0.19	
1950	IP621	12.9		0.59	
1967	IP470	12.10		0.12	
6024	IP470	11		-0.53	
6026	IP470	12.0		0.06	
6195	IP501	13.2		0.77	
6203	IP501	12.085		0.11	
normality		OK			
n		35			
outliers		0			
mean (n)		11.895			
st.dev. (n)		0.6633			
R(calc.)		1.857			
st.dev.(IP621:16)		1.6961			
R(IP621:16)		4.749			



Determination of Calcium as Ca on sample #23251; results in mg/kg

lab	method	value	mark	z(targ)	remarks
52	IP501	9		-1.02	
62	IP501	9		-1.02	
120		----		----	
150	IP501	9		-1.02	
154	IP501	10		-0.05	
158	IP501	9		-1.02	
171		----		----	
225	IP501	9		-1.02	
311		----		----	
323	IP501	10		-0.05	
333	IP501	9.7		-0.34	
381	D7111	9.0		-1.02	
445	IP621	11.0		0.92	
494	IP501	12		1.89	
608		----		----	
657	IP501	9.1		-0.92	
736	IP501	9.4		-0.63	
752		----		----	
781	IP501	11.3		1.21	
785	IP470	11.2		1.11	
823		----		----	
873		----		----	
874	IP621	11.8		1.69	
875		----		----	
994	IP501	9.6		-0.44	
995	IP621	9.2		-0.83	
1026	IP501	10		-0.05	
1065		----		----	
1140	IP501	10.11		0.05	
1191	In house	10.49		0.42	
1357	IP5	9.8		-0.25	
1556		----		----	
1677	IP621	10.5		0.43	
1862	IP501	9.07		-0.95	
1950	IP470	9.8		-0.25	
1967	IP470	9.05		-0.97	
6024	IP470	12		1.89	
6026	IP470	11.0		0.92	
6195	IP501	10.2		0.14	
6203	IP501	11.285		1.19	
normality		OK			
n		30			
outliers		0			
mean (n)		10.053			
st.dev. (n)		0.9893			
R(calc.)		2.770			
st.dev.(IP501:05R19)		1.0309			
R(IP501:05R19)		2.887			



## APPENDIX 2

### Number of participants per country

1 lab in AZERBAIJAN  
2 labs in BELGIUM  
2 labs in CANADA  
1 lab in COTE D'IVOIRE  
1 lab in EGYPT  
1 lab in ESTONIA  
1 lab in FINLAND  
1 lab in FRANCE  
1 lab in GEORGIA  
1 lab in GERMANY  
1 lab in GREECE  
1 lab in ISRAEL  
2 labs in KAZAKHSTAN  
1 lab in KOREA, Republic of  
1 lab in MALAYSIA  
1 lab in MALTA  
5 labs in NETHERLANDS  
1 lab in OMAN  
1 lab in ROMANIA  
17 labs in RUSSIAN FEDERATION  
1 lab in SERBIA  
1 lab in SINGAPORE  
1 lab in SWEDEN  
1 lab in UNITED ARAB EMIRATES  
2 labs in UNITED KINGDOM  
7 labs in UNITED STATES OF AMERICA

## APPENDIX 3

### Abbreviations

C	= final test result after checking of first reported suspect test result
D(0.01) / D(1)	= outlier in Dixon's outlier test
D(0.05) / D(5)	= straggler in Dixon's outlier test
G(0.01) / G(1)	= outlier in Grubbs' outlier test
G(0.05) / G(5)	= straggler in Grubbs' outlier test
DG(0.01) / DG(1)	= outlier in Double Grubbs' outlier test
DG(0.05) / DG(5)	= straggler in Double Grubbs' outlier test
R(0.01) / R(1)	= outlier in Rosner's outlier test
R(0.05) / R(5)	= straggler in Rosner's outlier test
E	= calculation difference between reported test result and result calculated by iis
W	= test result withdrawn on request of participant
ex	= test result excluded from statistical evaluation
n.a.	= not applicable
n.e.	= not evaluated
n.d.	= not detected
fr.	= first reported
f+?	= possibly a false positive test result?
f-?	= possibly a false negative test result?
SDS	= Safety Data Sheet

### Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 ISO5725:86
- 3 ISO5725 parts 1-6:94
- 4 ISO13528:05
- 5 M. Thompson and R. Wood, J. AOAC Int, 76, 926, (1993)
- 6 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 7 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 8 J.N. Miller, Analyst, 118, 455, (1993)
- 9 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 10 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, 127, 1359-1364, (2002)
- 11 W. Horwitz and R. Albert, J. AOAC Int, 79.3, 589-621, (1996)
- 12 Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, 25(2), 165-172, (1983)

Address: Malledijk 18, P.O. Box 200, 3200 AE Spijkenisse, The Netherlands  
Telephone number: +31 (0)88 214 45 41  
Email address: [nl.iis@sgs.com](mailto:nl.iis@sgs.com)  
Website: [www.iisnl.com](http://www.iisnl.com)

Institute for Interlaboratory Studies is a full member of SGS Nederland B.V. and registered at the Chamber of Commerce under number: 24226722. Unless otherwise agreed, all orders are executed in accordance with the SGS general conditions.