

Results of Proficiency Test Total Metals in Metal/Metal Alloy June 2023

Institute for Interlaboratory Studies Spijkenisse, the Netherlands Organized by:

Author: ing. M. Meijer

ing. A. Ouwerkerk & ing. R.J. Starink ing. A.S. Noordman-de Neef Correctors:

Approved by:

Report: iis23V24

September 2023

CONTENTS

1	INTRODUCTION	3
2	SET UP	3
2.1	QUALITY SYSTEM	3
2.2	PROTOCOL	4
2.3	CONFIDENTIALITY STATEMENT	4
2.4	SAMPLES	4
2.5	ANALYZES	6
3	RESULTS	6
3.1	STATISTICS	6
3.2	GRAPHICS	7
3.3	Z-SCORES	8
4	EVALUATION	8
4.1	EVALUATION PER PART OF THE SAMPLE AND PER ELEMENT	9
4.2	PERFORMANCE EVALUATION FOR THE GROUP OF LABORATORIES	11
4.3	COMPARISON OF THE PROFICIENCY TEST OF JUNE 2023 WITH PREVIOUS PTS	11
4.4	EVALUATION OF THE ANALYTICAL DETAILS	12
5	DISCUSSION	13
6	CONCLUSION	13
Appe	ndices:	
1.	Data, statistical and graphic results of #23630-1	14
2.	Data, statistical and graphic results of #23630-2	
3.	Other reported Elements	
4. -	Analytical details	
5. 2	Number of participants per country	
3.	Abbreviations and literature	44

1 Introduction

Heavy metals can be found in everything from jewelry and watch cases to electronic components. While Lead, Cadmium, Mercury and other heavy metals serve a purpose they are highly toxic to humans. Nickel is one of the most abundant metallic elements, likely to be found in most metals and metal alloys in trace quantities, including coins. Lead is a heavy metal that has often been used in jewelry, to make the article heavier, brighten colors and to stabilize or soften plastic. But Lead does not break down in the environment and accumulates in the human body.

Cadmium is also a heavy metal that has been used for over a century in both fashion and fine jewelry products. Small amounts of Cadmium may be added to metal alloys to impart specific technical and functional attributes to the metal alloys. It may be present in jewelry as part of the metal alloy, solder or gold coating for electroforming / electroplating, or as a pigment or stabilizer in non-metal components.

The legislation covering the restrictions on metals is found in Regulation (EC) No 1907/2006 concerning the Registration, Evaluation, Authorization and Restriction of Chemicals (REACH). Presence of Lead has been limited in the Consumer Product Safety Improvement Act (CPSIA) of 2008. The limit of Lead is 100 mg/kg. In REACH there are limits mentioned for Cadmium of 100 mg/kg and for Lead 500 mg/kg.

Since 2021 the Institute for Interlaboratory Studies (iis) organizes a proficiency scheme for the determination of Metals in Metal/Metal Alloy every year. During the annual proficiency testing program 2022/2023 it was decided to continue the proficiency test for the determination of Metals in Metal/Metal Alloy.

In this interlaboratory study 53 laboratories in 19 countries registered for participation, see appendix 5 for the number of participants per country. In this report the results of the Metals in Metal/Metal Alloy proficiency test are presented and discussed. This report is also electronically available through the iis website www.iisnl.com.

2 SET UP

The Institute for Interlaboratory Studies (iis) in Spijkenisse, the Netherlands, was the organizer of this proficiency test (PT). Sample analyzes for fit-for-use and homogeneity testing were subcontracted to an ISO/IEC17025 accredited laboratory.

It was decided to send one metal bracelet labelled #23630.

The participants were requested to report rounded and unrounded test results. The unrounded test results were preferably used for statistical evaluation.

2.1 QUALITY SYSTEM

The Institute for Interlaboratory Studies in Spijkenisse, the Netherlands, has implemented a quality system based on ISO/IEC17043:2010. This ensures strict adherence to protocols for sample preparation and statistical evaluation and 100% confidentiality of participant's data. Feedback from the participants on the reported data is encouraged and customer's satisfaction is measured on regular basis by sending out questionnaires.

2.2 PROTOCOL

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5). This protocol is electronically available through the iis website www.iisnl.com, from the FAQ page.

2.3 CONFIDENTIALITY STATEMENT

All data presented in this report must be regarded as confidential and for use by the participating companies only. Disclosure of the information in this report is only allowed by means of the entire report. Use of the contents of this report for third parties is only allowed by written permission of the Institute for Interlaboratory Studies. Disclosure of the identity of one or more of the participating companies will be done only after receipt of a written agreement of the companies involved.

2.4 SAMPLES

A batch of bracelets was purchased from the local market. Each bracelet was packed separately and labelled #23630. The bracelet could be divided into two parts (see picture):



Part 1: decorated leaves at ends of the bracelet, further referred to as #23630-1.

Part 2: main part of the bracelet, further referred to as #23630-2. The green line indicates the cutting line where the parts can be separated from each other.

The homogeneity of #23630-1 was checked by determination of Cadmium as Cd according to an in house method on 8 stratified randomly selected subsamples.

	Cadmium as Cd in mg/kg
sample #23630-1 sub-1	24620
sample #23630-1 sub-2	22595
sample #23630-1 sub-3	23934
sample #23630-1 sub-4	23161
sample #23630-1 sub-5	24401
sample #23630-1 sub-6	22581
sample #23630-1 sub-7	22153
sample #23630-1 sub-8	26355

Table 1: homogeneity test results of subsamples #23630-1

From the above test results the relative standard deviation (RSD) was calculated and compared with 0.3 times the average relative standard deviation obtained from twenty-four elements which were evaluated in the previous two iis PTs Total Metals in Metal/Metal Alloy in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Cadmium as Cd
RSD% observed	5.9
reference method	iis PTs
0.3 x RSD% (reference method)	6.6

Table 2: evaluation of the relative standard deviation of subsamples #23630-1

The calculated relative standard deviation is in agreement with 0.3 times the average relative standard deviation obtained from the evaluation of the elements from the previous two PTs. Therefore, homogeneity #23630-1 was assumed.

The homogeneity of #23630-2 was checked by determination of Cobalt according to an in house method on 7 stratified randomly selected subsamples.

	Cobalt as Co in mg/kg
sample #23630-2 sub-1	29.59
sample #23630-2 sub-2	25.74
sample #23630-2 sub-3	25.99
sample #23630-2 sub-4	25.93
sample #23630-2 sub-5	25.95
sample #23630-2 sub-6	26.28
sample #23630-2 sub-7	26.75

Table 3: homogeneity test results of subsamples #23630-2

From the above test results the relative standard deviation (RSD) was calculated and compared with 0.3 times the average relative standard deviation obtained from twenty-four elements which were evaluated in the previous two PTs Total Metals in Metal/Metal Alloy in agreement with the procedure of ISO13528, Annex B2 in the next table.

	Cobalt as Co
RSD% observed	5.1
reference method	iis PTs
0.3 x RSD% (reference method)	6.6

Table 4: evaluation of the relative standard deviation of subsamples #23630-2

The calculated relative standard deviation is in agreement with 0.3 times the average relative standard deviation obtained from the evaluation of the elements from the previous two PTs. Therefore, homogeneity of #23630-2 was assumed.

To each of the participating laboratories one bracelet sample #23630 was sent on June 7, 2023.

2.5 ANALYZES

The participants were requested to determine on the parts of sample #23630: total levels of Antimony as Sb, Arsenic as As, Cadmium as Cd, Chromium as Cr, Cobalt as Co, Copper as Cu, Lead as Pb, Manganese as Mn, Mercury as Hg, Nickel as Ni, Selenium as Se, Strontium as Sr, Tin as Sn, Zinc as Zn and Zirconium as Zr.

It was requested to report if the laboratory was accredited for the determined elements and to report some analytical details.

It was explicitly requested to treat the sample as if it was a routine sample and to report the test results using the indicated units on the report form and not to round the test results but report as much significant figures as possible. It was also requested not to report 'less than' test results which are above the detection limit, because such test results cannot be used for meaningful statistical evaluations. Furthermore, some additional instructions were sent on how to measure the different parts of the bracelet sample and it was asked to report the test results for the different parts separately.

To get comparable test results a detailed report form and a letter of instructions are prepared. On the report form the reporting units are given as well as the reference test methods (when applicable) that will be used during the evaluation. The detailed report form and the letter of instructions are both made available on the data entry portal www.kpmd.co.uk/sgs-iis-cts/. The participating laboratories are also requested to confirm the sample receipt on this data entry portal. The letter of instructions can also be downloaded from the iis website www.iisnl.com.

3 RESULTS

During five weeks after sample dispatch, the test results of the individual laboratories were gathered via the data entry portal www.kpmd.co.uk/sgs-iis-cts/. The reported test results are tabulated per determination in appendices 1 to 3 of this report. The laboratories are presented by their code numbers.

Directly after the deadline, a reminder was sent to those laboratories that had not reported test results at that moment. Shortly after the deadline, the available test results were screened for suspect data. A test result was called suspect in case the Huber Elimination Rule (a robust outlier test) found it to be an outlier. The laboratories that produced these suspect data were asked to check the reported test results (no reanalyzes). Additional or corrected test results are used for data analysis and the original test results are placed under 'Remarks' in the test result tables in appendices 1 to 3. Test results that came in after the deadline were not taken into account in this screening for suspect data and thus these participants were not requested for checks.

3.1 STATISTICS

The protocol followed in the organization of this proficiency test was the one as described for proficiency testing in the report 'iis Interlaboratory Studies: Protocol for the Organisation, Statistics and Evaluation' of June 2018 (iis-protocol, version 3.5).

For statistical evaluation the *unrounded* (when available) figures were used instead of the rounded test results. Test results reported as '<...' or '>...' were not used in the statistical evaluation.

First, the normality of the distribution of the various data sets per determination was checked by means of the Lilliefors-test, a variant of the Kolmogorov-Smirnov test and by the calculation of skewness and kurtosis. Evaluation of the three normality indicators in combination with the visual evaluation of the graphic Kernel density plot, lead to judgement of the normality being either 'unknown', 'OK', 'suspect' or 'not OK'. After removal of outliers, this check was repeated. If a dataset does not have a normal distribution, the (results of the) statistical evaluation should be used with due care.

The assigned value is determined by consensus based on the test results of the group of participants after rejection of the statistical outliers and/or suspect data.

According to ISO13528 all (original received or corrected) results per determination were submitted to outlier tests. In the iis procedure for proficiency tests, outliers are detected prior to calculation of the mean, standard deviation and reproducibility. For small data sets, Dixon (up to 20 test results) or Grubbs (up to 40 test results) outlier tests can be used. For larger data sets (above 20 test results) Rosner's outlier test can be used. Outliers are marked by D(0.01) for the Dixon's test, by G(0.01) or DG(0.01) for the Grubbs' test and by F(0.01) for the Rosner's test. Stragglers are marked by F(0.01) for the Dixon's test, by F(0.01) for the Rosner's test. Both outliers and stragglers were not included in the calculations of averages and standard deviations.

For each assigned value, the uncertainty was determined in accordance with ISO13528. Subsequently the calculated uncertainty was evaluated against the respective requirements based on the target reproducibility in accordance with ISO13528. In this PT, the criterion of ISO13528, paragraph 9.2.1 was met for all evaluated tests, therefore, the uncertainty of all assigned values may be negligible and need not be included in the PT report.

Finally, the reproducibilities were calculated from the standard deviations by multiplying them with a factor of 2.8.

3.2 GRAPHICS

In order to visualize the data against the reproducibilities from literature, Gauss plots were made, using the sorted data for one determination (see appendix 1 and 2). On the Y-axis the reported test results are plotted. The corresponding laboratory numbers are on the X-axis. The straight horizontal line presents the consensus value (a trimmed mean). The four striped lines, parallel to the consensus value line, are the +3s, +2s, -2s and -3s target reproducibility limits of the selected reference test method. Outliers and other data, which were excluded from the calculations, are represented as a cross. Accepted data are represented as a triangle.

Furthermore, Kernel Density Graphs were made. This is a method for producing a smooth density approximation to a set of data that avoids some problems associated with histograms. Also, a normal Gauss curve (dotted line) was projected over the Kernel Density Graph (smooth line) for reference. The Gauss curve is calculated from the consensus value and the corresponding standard deviation.

3.3 Z-SCORES

To evaluate the performance of the participating laboratories the z-scores were calculated. As it was decided to evaluate the performance of the participants in this proficiency test (PT) against the literature requirements (derived from e.g. ISO or ASTM test methods), the z-scores were calculated using a target standard deviation. This results in an evaluation independent of the variation in this interlaboratory study.

This target standard deviation was calculated from the literature reproducibility by division with 2.8. In case no literature reproducibility was available, other target values were used, like Horwitz or an estimated reproducibility based on former iis proficiency tests.

When a laboratory did use a test method with a reproducibility that is significantly different from the reproducibility of the reference test method used in this report, it is strongly advised to recalculate the z-score, while using the reproducibility of the actual test method used, this in order to evaluate whether the reported test result is fit-for-use.

The z-scores were calculated according to:

```
z_{\text{(target)}} = \text{(test result - average of PT)} / \text{target standard deviation}
```

The $z_{(target)}$ scores are listed in the test result tables in appendices 1 to 2.

Absolute values for z<2 are very common and absolute values for z>3 are very rare. Therefore, the usual interpretation of z-scores is as follows:

```
|z| < 1 good
1 < |z| < 2 satisfactory
2 < |z| < 3 questionable
3 < |z| unsatisfactory
```

4 **EVALUATION**

In this proficiency test no problems were encountered with the dispatch of the samples. Six participants reported test results after the final reporting date and five other participants did not report any test results. Not all participants did not report all tests requested. In total 48 participants reported 252 numerical test results. Observed were 14 outlying test results, which is 5.6%. In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

All data sets proved to have a normal Gaussian distribution.

EVALUATION PER PART OF THE SAMPLE AND PER ELEMENT

In this section the results are discussed per part of the sample and per element. The test methods which were used by the various laboratories were taken into account for explaining the observed differences when possible and applicable. These test methods are also in the tables together with the reported test results in appendices 1 and 2. The abbreviations, used in these tables, are explained in appendix 6.

Participants used different test methods to determine Metals in Metal/Metal Alloy. Some participants used test method CPSC-CH-E1001.08 (Determination of Total Lead in Metal Products), some used EN16711-1 (Determination of Metal content in Textile) or IEC62321-5 (Determination of Cadmium and Lead in Metals) and others used an in house method. Regretfully, no precision data is mentioned in CPSC-CH-E1001.08. The precision data mentioned in EN16711-1 involve a low concentration of metals tested in textile which may be not applicable here. The Horwitz equation can be used but it is a known fact that when Horwitz is used at higher concentrations, the equation returns a stricter reproducibility than at lower concentrations. Therefore, a target reproducibility based on the Horwitz equation may not be applicable to the higher concentrations of some metals in the bracelet. However, method IEC62321-5 (published in 2014) does contain precision data based on a metal sample (metal parts found in e.g. buttons). The precision was determined for concentrations up to 1000 mg/kg and the RSD found for this range is 10%. It was decided to use this as the target reproducibility for all metals, calculated according to this formula: 0.1 * mean * 2.8.

#23630-1: The decorated leaves at the ends of the bracelet

Cadmium as Cd: This determination was not problematic. Four statistical outliers were observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of IEC62321-5:14.

Copper as Cu: This determination was very problematic. No statistical outliers were observed. A large variation in the reported test results is observed,

therefore no z-scores are calculated. Lead as Pb: This determination was very problematic. In the leaves of the decorated ends of the bracelet solder is present. Solder usually consists of an alloy of Tin and Lead, although nowadays there are also lead-free variants. From

> the test results it seems that some participants included solder in the analysis while other participants did not. This resulted unfortunately in a large variation in the reported test results and therefore no z-scores are

calculated, see also paragraph 5 for more discussion.

This determination was very problematic. Two statistical outliers were Tin as Sn: observed. In the leaves of the decorated ends of the bracelet solder is present. Solder usually consists of an alloy of Tin and Lead, although nowadays there are also lead-free variants. From the test results it seems that some participants included solder in the analysis while other

participants did not. This resulted unfortunately in a large variation in the

reported test results and therefore no z-scores are calculated.

Zinc as Zn: This determination was not problematic. Two statistical outliers were

observed. The calculated reproducibility after rejection of the statistical outliers is in agreement with the requirements of IEC62321-5:14.

The participants agreed on a concentration near or below the limit of detection for all other elements mentioned in paragraph 2.5. Therefore, no z-scores are calculated for these elements. The reported test results are given in appendix 3.

#23630-2: Main part of the bracelet (wrist band)

Arsenic as As: This determination was problematic. No statistical outliers were observed.

The calculated reproducibility is not in agreement with the requirements of

IEC62321-5:14.

Chromium as Cr: This determination was problematic. No statistical outliers were observed.

The calculated reproducibility is not in agreement with the requirements of

IEC62321-5:14.

<u>Cobalt as Co</u>: This determination was problematic. No statistical outliers were observed.

The calculated reproducibility is not in agreement with the requirements of

IEC62321-5:14.

Copper as Cu: This determination was very problematic. One statistical outlier was

observed. A large variation in the reported test results is observed,

therefore no z-scores are calculated.

Manganese as Mn: This determination was not problematic. One statistical outlier was

observed. The calculated reproducibility after rejection of the statistical

outlier is in agreement with the requirements of IEC62321-5:14.

Nickel as Ni: This determination was problematic. One statistical outlier was observed.

The calculated reproducibility after rejection of the statistical outlier is not in

agreement with the requirements of IEC62321-5:14.

<u>Tin as Sn</u>: This determination was problematic. Two statistical outliers were observed.

The calculated reproducibility after rejection of the statistical outliers is not

in agreement with the requirements of IEC62321-5:14.

<u>Zinc as Zn</u>: This determination was very problematic. One statistical outlier was

observed. A large variation in the reported test results is observed,

therefore no z-scores are calculated.

The participants agreed on a concentration near or below the limit of detection for all other elements mentioned in paragraph 2.5. Therefore, no z-scores are calculated for these elements. The reported test results are given in appendix 3.

4.2 Performance evaluation for the group of Laboratories

A comparison has been made between the reproducibility as declared by the reference test method and the reproducibility as found for the group of participating laboratories. The number of significant test results, the average, the calculated reproducibility (2.8 * standard deviation) and the target reproducibility derived from reference methods are presented in the next tables.

Element	unit	n	average	2.8 * sd	R(lit)
Cadmium as Cd	mg/kg	31	866561	127159	242637
Copper as Cu	mg/kg	15	53983	61391	(15115)
Lead as Pb	mg/kg	35	n.e.	n.e.	n.e.
Tin as Sn	mg/kg	11	500	436	(140)
Zinc as Zn	mg/kg	12	53775	12277	(15057)

Table 5: reproducibilities of tests on sample #23630-1 (decorated leaves, ends of the bracelet)

For results between brackets no z-scores are calculated.

Element	unit	n	average	2.8 * sd	R(lit)
Arsenic as As	mg/kg	21	36.2	12.7	10.1
Chromium as Cr	mg/kg	30	216	89	60
Cobalt as Co	mg/kg	18	34.6	13.8	9.7
Copper as Cu	mg/kg	15	90327	109593	(25292)
Manganese as Mn	mg/kg	18	2782	732	779
Nickel as Ni	mg/kg	22	73.8	24.6	20.7
Tin as Sn	mg/kg	13	600	350	168
Zinc as Zn	mg/kg	12	167	338	(47)

Table 6: reproducibilities of tests on sample #23630-2 (main part of the bracelet (wrist band))

For results between brackets no z-scores are calculated.

Without further statistical calculations it can be concluded that for many tests there is not a good compliance of the group of participants with the reference test method. See also the discussion in paragraph 4.1 and 5.

4.3 COMPARISON OF THE PROFICIENCY TEST OF JUNE 2023 WITH PREVIOUS PTS

	June 2023	April 2022	June 2021
Number of reporting laboratories	48	66	56
Number of test results	252	627	350
Number of statistical outliers	14	59	26
Percentage of statistical outliers	5.6%	9.4%	7.4%

Table 7: comparison with previous proficiency tests

In proficiency tests outlier percentages of 3% - 7.5% are quite normal.

The performance of the determinations of the proficiency test was compared to uncertainties observed in PTs over the years, expressed as relative standard deviation (RSD) of the PTs, see next table.

Element	June 2023	April 2022	June 2021
Arsenic as As	12%	20-42%	20%
Cadmium as Cd	5%	17%	15-17%
Chromium as Cr	15%	13-15%	15-26%
Cobalt as Co	14%	12-17%	20%
Copper as Cu	41-43%	4-25%	10-42%
Lead as Pb		29%	24-56%
Manganese as Mn	9%	6-8%	5-9%
Nickel as Ni	12%	13-18%	7-12%
Tin as Sn	21-31%	7-31%	
Zinc as Zn	8-72%	9-80%	7-84%

Table 8: development of the uncertainties over the years

The RSDs observed in this PT are in line with RSDs observed in previous iis PTs.

4.4 EVALUATION OF THE ANALYTICAL DETAILS

For this PT some analytical details were requested which are listed in appendix 4. Based on the answers given by the participants the following can be summarized:

- 82% of the participants mentioned to be ISO/IEC17025 accredited to determine the reported elements for Metals in Metal/Metal Alloys.
- 88% of the participants used 0.5 grams or less as sample intake, most times 0.1 to 0.2 grams.
- 67% of the participants used Aqua Regia (a mixture of concentrated Nitric acid and concentrated Hydrochloric acid) as digestion acid, while 33% used Nitric acid only (in different concentrations).
- 60% of the participants used ICP-OES as technique to quantify the metals, 26% ICP-MS and 7% used AAS. A few used combinations of techniques with ICP-OES, ICP-MS and XRF.

As the majority of the group follow the same analytical procedures no further statistical analysis has been performed.

5 DISCUSSION

In this PT the average of the homogeneity test results for sample #23630-1 is not in line with the average (consensus value) from the PT results. There are several reasons for this. First, the goal of the homogeneity testing is very different from the goal of the evaluation of the reported PT results. To prove the homogeneity of the PT samples, a test method is selected with a high precision (smallest variation). The accuracy (trueness) of the test method is less relevant.

Secondly, the homogeneity testing is done by one laboratory only. The test results of this (ISO/IEC 17025 accredited) laboratory will have a bias (systematic deviation) depending on the test method used. The desire to detect small variations between the PT samples leads to the use of a sensitive test method with high precision, which may be a test method with significant bias. Also each test result reported by the laboratories that participate in the PT will have a bias. However, some will have a positive bias and others a negative bias. These different biases compensate each other in the PT average (consensus value). Therefore, the PT consensus value may deviate from the average of the homogeneity test. At the same time the accuracy of the PT consensus value is more reliable than the accuracy of the average of the results of the homogeneity test.

In this proficiency test for the determination of Metals in Metal/Metal Alloys a bracelet was purchased from a retailer and is therefore a real-life sample. Several participants contacted iis to ask how the leaves of the decoration of sample #23630-1 should be tested, specific whether the solder should be tested or not. The leaves were connected to the rest of the bracelet with apparently a Lead/Tin containing solder. In the letter of instructions iis requests participants to analyze the PT item in the way the laboratory is used to do. It appears that laboratories have different ways of working and some include (partly) the solder into the analysis while other laboratories skip the solder. As a result, the Lead and Tin determinations could not be evaluated because of the large deviation in reported test results. The large variation in these two elements may partly be explained by the lack of clear testing procedures on how to analyze this type of samples. It is recommended that members of technical committees bring this further into discussion to improve the analytical procedure in the test methods to avoid this confusion.


6 CONCLUSION


Each participating laboratory will have to evaluate its performance in this study and decide about any corrective actions if necessary. Therefore, participation on a regular basis in this scheme could be helpful to improve the performance and thus increase of the quality of the analytical results.

APPENDIX 1 – Decoration leaves at the end of the bracelet

Determination of Cadmium as Cd on sample #23630-1; results in mg/kg


Detern	nination of Cadmium	as Cd on san	nple #236		
lab	method	value	mark	z(targ)	remarks
210					
623	In house	856126.27		-0.12	
840					
1126					
2115	EN16711-1	904146.2		0.43	
2121					
	In house	831790		-0.40	
2159					
	EN16711-1	867777		0.01	
	CPSC-CH-E1001-08.3	853061	С	-0.16	first reported 712078.3
	CPSC-CH-E1001-08.3	867126	Ü	0.01	1100100011001112010.0
	CPSC-CH-E1001-08.3	802135		-0.74	
2230	01 00-011- <u>L</u> 1001-00.9			-0.74	
	CPSC-CH-E1001-08.3	860000		-0.08	
	EN16711-1	863000		-0.04	
	EN16711-1	796786		-0.81	
2311	EN16711-1	792778.65		-0.85	
	CPSC-CH-E1001-08.3	829521.18		-0.43	
2326					
2347	0000 011 51				
	CPSC-CH-E1001-08.3	874475		0.09	
2358	EN16711-1	864704		-0.02	
2366					
2373					
2375	EN16711-1	805527		-0.70	
2380	EN16711-1	912786.697		0.53	
2381	EN16711-1	979106.70		1.30	
2382					
2385	In house	798753		-0.78	
2459	EN16711-1	887800		0.25	
2475	In house	854298.53		-0.14	
	CPSC-CH-E1001-08.3	869799.5		0.04	
2503		848000	С	-0.21	first reported 848.000
2511			•		
	CPSC-CH-E1001-08.3	459324.400	C,R(0.01)	-4.70	first reported 675442.28
	CPSC-CH-E1001-08.3	831978.339	0,11(0.01)	-0.40	1110t 10portou 070 112.20
	CPSC-CH-E1001-08.3	929220		0.72	
	In house	932039		0.72	
2637		372000	R(0.01)	-5.71	
2678	III llouse		13(0.01)		
2734	CDSC CH E4004 00 0	 000000 F		0.10	
	CPSC-CH-E1001-08.3	882339.5		0.18	
2741	CDCC CIL E4004 00 0	 502040 20	D(0.04)	4.40	
	CPSC-CH-E1001-08.3	503849.39	R(0.01)	-4.19	
	CPSC-CH-E1001-08.3	957345.2		1.05	
	In house	896380.85		0.34	
2977					
	EN62321-5	876342.95		0.11	
3116					
	IEC62321-3	336600	R(0.01)	-6.12	
3209	In house	865965.51		-0.01	
3214					
3228	CPSC-CH-E1001-08.3	872300		0.07	
	normality	OK			
	n	31			
	outliers	4			
	mean (n)	866561.44			
	st.dev. (n)	45413.930	RSD=5%		
	R(calc.)	127159.00			
	st.dev.(IEC62321-5:14)	86656.144			
	R(IEC62321-5:14)	242637.20			
		= .20020			






Determination of Copper as Cu on sample #23630-1; results in mg/kg

lab	method	value	mark	z(targ)	remarks
210					
623	In house	24707.60			
840					
1126	EN10744 4				5
2115	EN16711-1	81083	С		first reported 181553.6
2121 2146					
2159					
2165					
2182					
2184					
2216					
2230					
2256					
2265	EN10744 4	 5400 7			
	EN16711-1	51367			
2311 2314	EN16711-1 CPSC-CH-E1001-08.3	46989.74 54094.40			
2326	CF3C-CH-E1001-06.3	54094.40			
2347					
2350					
2358	EN16711-1	39560			
2366					
2373					
2375	EN16711-1	55210			
2380	=11.0=11.1				5
2381	EN16711-1	53085.1	С		first reported 126109.60
2382	In house	66398			
2459	EN16711-1	39625			
2475	214107111				
2500					
2503		94700	С		first reported 94.700
2511					
2590	CPSC-CH-E1001-08.3	49251.83			
2591	0000 011 54004 00 0	40000			
	CPSC-CH-E1001-08.3	10029			
2624 2637	In house	83300			
2678	III House				
2734					
2737					
2741					
2810					
2829					
2864	CDCC CH E4004 00 0	60240			
	CPSC-CH-E1001-08.3	60349			
2995 3116					
3172					
3209					
3214					
3228					
	normality	OK			
	n outliere	15			
	outliers mean (n)	0 53983.31			
	st.dev. (n)	21925.303	RSD=41%		
	R(calc.)	61390.85	1.05 4170		
	st.dev.(IEC62321-5:14)	(5398.331)			
	R(IEC62321-5:14)	(15115.33)			







Determination of Lead as Pb on sample #23630-1; results in mg/kg

lab	method	value	mark	z(targ)	remarks
210					
	In house	172.70			
840					
1126					
	EN16711-1	6.18			
2121	CPSC-CH-E1001-08.3	6.554			
2146	In house	22537			
2159					
	EN16711-1	Not detected			
	CPSC-CH-E1001-08.3	not detected			
	CPSC-CH-E1001-08.3	<10			
	CPSC-CH-E1001-08.3	26018			
	CPSC-CH-E1001-08.3	<10			
	CPSC-CH-E1001-08.3	Not detected			
	EN16711-1	3940			
	EN16711-1	<10			
	EN16711-1	11.97			
2314					
2326					
2347					
2350	CPSC-CH-E1001-08.3	< 50			
2358	EN16711-1	not detected			
	CPSC-CH-E1001-08.3	15309			
2373					
	EN16711-1	6400			
	EN16711-1	13.196			
	EN16711-1	14.90			
	CPSC-CH-E1001-08.3	6600			
	In house	11187			
2459		346.53			
2475					
	CPSC-CH-E1001-08.3	ND			
2503					
2511	CPSC-CH-E1001-08.3	<10			
2590	CPSC-CH-E1001-08.3	3.48			
2591	CPSC-CH-E1001-08.3	56984.162			
2622					
2624					
2637	In house	6320			
2678					
2734					
2737					
2741					
2810	CPSC-CH-E1001-08.3	14374.76			
2829	CPSC-CH-E1001-08.3	not detected			
2864	In house	8.35			
2977	CPSC-CH-E1001-08.3	not detected			
2995	EN62321-5	8731.83			
	CPSC-CH-E1001-08.3	<10			
3172					
3209	In house	<10.0			
	EN16711-1	<10			
3228	CPSC-CH-E1001-08.3	<10			
	n	24	11		
	mean (n)	<350	>3500		

Determination of Tin as Sn on sample #23630-1; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210					
623	In house	525.29			
840					
1126	EN40744 4	740.0			
2115	EN16711-1	749.3			
2121 2146					
2159					
2165					
2182					
2184					
2216					
2230					
2256					
2265					
2310	EN16711-1	478			
2311	EN16711-1	498.43			
2314	CPSC-CH-E1001-08.3	574.21			
2326					
2347					
2350	=11.0=11.1				5
2358	EN16711-1	not detected	С		first reported 167 / possibly a false negative test result?
2366					
2373 2375	EN16711-1	125			
2380	EN 107 I I-1	435			
2381					
2382					
2385	In house	2634	G(0.01)		
2459	EN16711-1	196.36	, ,		
2475					
2500					
2503					
2511					
2590 2591					
2622	CPSC-CH-E1001-08.3	382			
2624	O1 00 011 E1001 00.0				
2637	In house	1500	G(0.01)		
2678			, ,		
2734					
2737					
2741					
2810					
2829 2864					
	CPSC-CH-E1001-08.3	422.9			
2995	O1 00-011-E1001-00.5				
3116					
3172	EN16711-1	499			
3209					
3214	EN16711-1	740.34			
3228					
	normality	OK			
	normality n	OK 11			
	outliers	2			
	mean (n)	500.075			
	st.dev. (n)	155.8328	RSD=31%		
	R(calc.)	436.332	0.70		
	st.dev.(IEC62321-5:14)	(50.0075)			
	R(IEC62321-5:14)	(140.021)			






Determination of Zinc as Zn on sample #23630-1; results in mg/kg


	method	value	mark	z(targ)	remarks
210	In harra			0.40	
	In house	51490.35		-0.42	
840 1126					
	EN16711-1	48431.6		-0.99	
2121	LIVIO711-1			-0.55	
2146					
2159					
2165					
2182					
2184					
2216					
2230					
2256					
2265	EN40744 4	 FC004		0.50	
	EN16711-1 EN16711-1	56921		0.59 -0.34	
	CPSC-CH-E1001-08.3	51959.61 52302.14		-0.34	
2326	CF3C-CH-E1001-00.3	J2JU2.14 		-0.27	
2347					
2350					
	EN16711-1	57044		0.61	
2366					
2373					
	EN16711-1	54166		0.07	
2380					
	EN16711-1	59275.50		1.02	
2382	In Italian	44004		4.00	
	In house EN16711-1	44091	C(0.05)	-1.80	
2475	EN 107 11-1	34502.22	G(0.05)	-3.58	
2500					
2503		56300	С		first reported 56.300
2511			· ·		mot reported co.coc
2590					
2591					
	CPSC-CH-E1001-08.3	58151		0.81	
2624					
	In house	23700	G(0.05)	- 5.59	
2678					
2734					
2737 2741					
2810					
2829					
2864					
	CPSC-CH-E1001-08.3	55167		0.26	
2995					
3116					
3172					
3209					
3214					
3228					
	normality	OK			
	n	12			
	outliers	2			
	mean (n)	53774.93			
	st.dev. (n)	4384.591	RSD=8%		
	R(calc.)	12276.86			
	st.dev.(IEC62321-5:14)	5377.493			
	R(IEC62321-5:14)	15056.98			






APPENDIX 2 – Main part of the bracelet Determination of Arsenic as As on sample #23630-2; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210					
623	In house	37.10		0.24	
840	In house	29.66		-1.81	
1126	=>				
2115	EN16711-1	36.36		0.04	
2121					
2146 2159	In house	 45.44		2.55	
2165	III House				
2182					
2184					
2216	CPSC-CH-E1001-08.3	38		0.49	
2230					
2256	=>				
2265		not detected		0.24	possibly a false negative test result?
	EN16711-1 EN16711-1	35 34.81		-0.34 -0.39	
	EN16711-1 EN16711-1	37.51		0.36	
2326	ENTO/TI-T				
2347					
2350					
2358	EN16711-1	35.1		-0.31	
	CPSC-CH-E1001-08.3	38		0.49	
	EN16711-1	36.1		-0.03	
	EN16711-1	35		-0.34	
2380	EN16711-1	32.669 34.10		-0.98 -0.58	
2382	EN16711-1	34.10		-0.56	
2385	In house	42.2		1.65	
2459					
2475					
2500					
	CPSC-CH-E1001-08.3	43.3		1.96	
	EN16711-1	38.681		0.68	
2590 2591	CPSC-CH-E1001-08.3	< L.O.Q.			
2622					
2624					
2637					
2678					
2734					
2737					
2741	CPSC-CH-E1001-08.3	39.733		0.97	
2810 2829					
2864					
	CPSC-CH-E1001-08.3	36.32		0.03	
2995					
3116					
	EN16711-1	26.8		-2.60	
3209					
3214	EN16711-1	28.70		-2.08	
3228					
	normality	ОК			
	n	21			
	outliers	0			
	mean (n)	36.2182			
	st.dev. (n)	4.52634	RSD=12%		
	R(calc.)	12.6738			
	st.dev.(IEC62321-5:14)	3.62182 10.1411			
	R(IEC62321-5:14)	10.1411			






Determination of Chromium as Cr on sample #23630-2; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210					
	In house	201.85		-0.65	
840	In house	218.51		0.12	
1126	III II Guee				
2115	EN16711-1	162.63		-2.47	
2121	LINIO711-1			-2.41	
2146					
2159	In house	194.9		-0.98	
2165		235.7		0.91	
2182	EN16711-1	233.1		0.91	
	CPSC-CH-E1001-08.3	225.2		0.43	
				0.43	
2230	CPSC-CH-E1001-08.3	236		0.93	
	CPSC-CH-E1001-08.3			-2.59	
	EN16711-1	160.025 205.0			
				-0.51	
	EN16711-1	245		1.34	
	EN16711-1	251.78		1.66	
	EN16711-1	258.14		1.95	
2326					
2347					
2350	EN16711 1	 177		1.00	
	EN16711-1	177		-1.80	
	CPSC-CH-E1001-08.3	232		0.74	
	EN16711-1	235.3		0.90	
	EN16711-1	225		0.42	
	EN16711-1	206.274		-0.45	
	EN16711-1	217.90		0.09	
2382		450			
	In house	158		-2.68	
2459	EN16711-1	173.6		-1.96	
2475	CDCC CU F4004 00 2			0.40	
2500	CPSC-CH-E1001-08.3	225.1		0.42	
2503	CPSC-CH-E1001-08.3	180		-1.67	
2511	CDCC CIL E4004 00 2	057.074	0	4.00	first non-order d 40-50
2590	CPSC-CH-E1001-08.3	257.071	С	1.90	first reported 49.58
2591					
2622					
2624					
2637					
2678					
2734	CDSC CH E1001 00 2	253 10		1 70	
2737 2741	CPSC-CH-E1001-08.3	253.19		1.72	
2810					
2829					
2864					
	CPSC-CH-E1001-08.3	242.0		1.21	
	EN62321-5	162		-2.50	
3116	LINUZUZ I-U	102		-2.50	
	EN16711-1	250.7		1.61	
3209		240.36		1.13	
	EN16711-1	240.36		0.13	
	CPSC-CH-E1001-08.3	230		0.13	
3220	31 30-311-E 100 1-00.3	200		0.00	
	normality	OK			
	n	30			
	outliers	0			
	mean (n)	215.9660			
	st.dev. (n)	31.65342	RSD=15%		
	R(calc.)	88.6296	100-1070		
	st.dev.(IEC62321-5:14)	21.59660			
	R(IEC62321-5:14)	60.4705			
		30.1700			






Determination of Cobalt as Co on sample #23630-2; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210	In house	22.40		0.40	
	In house In house	33.19 30.46		-0.40 -1.19	
1126	III IIUuse	30.46		-1.19	
	EN16711-1	26.92		-2.21	
2121				-2.21	
2146					
2159	In house	38.068		1.01	
2165					
2182					
2184					
2216 2230					
2256					
2265					
	EN16711-1	39		1.28	
	EN16711-1	Not Detected			possibly a false negative test result?
	EN16711-1	43.43		2.56	
2326					
2347					
2350	EN16711 1	22.6		2 17	
2358	EN16711-1	23.6		-3.17 	
	EN16711-1	35.1		0.15	
	EN16711-1	38		0.13	
2380					
2381					
2382					
	In house	29.7		-1.41	
	EN16711-1	38.14		1.03	
2475	CPSC-CH-E1001-08.3	37.6		0.87	
	CPSC-CH-E1001-08.3	31.6		-0.86	
	EN16711-1	31.076		-1.01	
	CPSC-CH-E1001-08.3	35.42		0.24	
2591					
2622					
2624					
2637					
2678 2734					
2734 2737					
2741					
2810					
2829					
2864					
	CPSC-CH-E1001-08.3	not detected			possibly a false negative test result?
2995					
3116 3172	EN16711-1	35.4		0.24	
	In house	38.68		1.19	
	EN16711-1	36.99		0.70	
3228					
	normality	OK			
	n cuttions	18			
	outliers mean (n)	0 34.5763			
	st.dev. (n)	4.91634	RSD=14%		
	R(calc.)	13.7658	14/0		
	st.dev.(IEC62321-5:14)	3.45763			
	R(IEC62321-5:14)	9.6814			






Determination of Copper as Cu on sample #23630-2; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210	In house	40000 00			
623	In house	43233.86			
840 1126	In house	103506.6			
2115	EN16711-1	91669	С		first reported 208622.2
2121	211107111		Ü		
2146					
2159					
2165					
2182					
2184 2216					
2230					
2256					
2265					
2310	EN16711-1	89825			
2311	EN16711-1	66229.66			
2314	EN16711-1	87532.43			
2326 2347					
2350					
2358	EN16711-1	38916			
2366					
2373	EN16711-1	78346.3			
2375	EN16711-1	96609			
2380	EN140744 4	77500 50			
2381 2382	EN16711-1	77582.50 			
2385	In house	83810			
2459	EN16711-1	137253.32			
2475					
2500					
2503	CPSC-CH-E1001-08.3	167000	С		first reported 167.000
2511	CDSC CU E1001 09 3	26206.44			
2590 2591	CPSC-CH-E1001-08.3	36396.41			
2622		157000			
2624	In house	981930	G(0.01)		
2637			,		
2678					
2734					
2737					
2741 2810					
2829					
2864					
	CPSC-CH-E1001-08.3	not detected			possibly a false negative test result?
2995					
3116					
3172					
3209 3214					
3228					
3223					
	normality	OK			
	n	15			
	outliers	1			
	mean (n)	90327.34	DSD-420/		
	st.dev. (n) R(calc.)	39140.426 109593.19	RSD=43%		
	st.dev.(IEC62321-5:14)	(9032.734)			
	R(IEC62321-5:14)	(25291.66)			
	,	/			






Determination of Manganese as Mn on sample #23630-2; results in mg/kg


lab	method	value	mark	z(targ)	remarks
210					
	In house	2724.64		-0.21	
840	In house	2685.61		-0.35	
1126					
	EN16711-1	2474.4		-1.10	
2121					
2146					
	In house	2903		0.44	
2165					
2182					
2184					
2216					
2230					
2256					
2265					
	EN16711-1	2938		0.56	
	EN16711-1	3062.84		1.01	
	EN16711-1	3002.25		0.79	
2326					
2347					
2350					
2358					
2366					
2373					
	EN16711-1	2847		0.23	
2380					
	EN16711-1	3235.60		1.63	
2382					
	In house	2254		-1.90	
	EN16711-1	2346.25		-1.57	
2475					
2500	CPSC-CH-E1001-08.3	2919.1		0.49	
2503	CPSC-CH-E1001-08.3	2690	С	-0.33	first reported 2.690
2511					
2590	CPSC-CH-E1001-08.3	2556.55		-0.81	
2591					
2622		1289	G(0.01)	-5.37	
2624					
2637					
2678					
2734					
	CPSC-CH-E1001-08.3	2958.47		0.64	
2741					
2810					
2829					
2864	0000 011 54004 00 0				
	CPSC-CH-E1001-08.3	2622		-0.57	
2995					
3116					
3172	In the same	0707.04			
	In house	2797.01		0.05	
3214	EN16711-1	3054.85		0.98	
3228					
	normality:	OK			
	normality	OK			
	n	18			
	outliers	1			
	mean (n)	2781.754	DCD=00/		
	st.dev. (n)	261.4100	RSD=9%		
	R(calc.)	731.948			
	st.dev.(IEC62321-5:14)	278.1754			
	R(IEC62321-5:14)	778.891			






Determination of Nickel as Ni on sample #23630-2; results in mg/kg


2210	210 623 In house 63.03 .1.46 840 In house 70.11 .0.50 1126	lab me	ethod	value	mark	z(targ)	remarks
623 In house 63.03	623 In house 63.03 -1.46 840 In house 70.11 -0.50 1126					-(
In house	840 In house 70.11 -0.50 -1126		house	63.03		-1.46	
11126	1126						
2115 EN16711-1 70.6 -0.43 2121	2115 EN16711-1						
21121 2146 2159 In house 92.02 2.47 2182	2121 2146 2159 In house 92.02 2.47 2165 2182		N16711-1	70.6		-0.43	
2159 In house 92.02 2.47 2162	2159 In house 92.02 2.47 2165						
2165 2184 2216 2216 2230 2256 2255 2251 2310 2111 2117 2211 2211 2211 2211 2211 22	2184						
2182	2182 2184 2216 2230 2256 2265 2310 EN16711-1 74 0.03 2311 EN16711-1 72.11 -0.22 2326 2347 2350 2358 EN16711-1 53.6 2373 EN16711-1 53.6 2373 EN16711-1 67.2 -0.89 2375 EN16711-1 67.2 -0.89 2375 EN16711-1 64.329 -1.28 2381 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 88.19 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2777 CPSC-CH-E1001-08.3 72.98 1.17 2774 2810 CPSC-CH-E1001-08.3 72.98 1.17 2795 1116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12%	2159 In I	house	92.02		2.47	
2216	2184 2216 2230 2256 2265 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326 2347 2350 2358 EN16711-1 53.6 2373 EN16711-1 74 0.03 2388 EN16711-1 74 0.03 2389 EN16711-1 74 0.03 2380 EN16711-1 74 0.03 2380 EN16711-1 74 0.03 2380 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.000 2459 EN16711-1 229.76 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 82.39 0.01 1.17 0.03 2829 2864 0.00 2995 29864 0.00 2995 29864 0.00 2995 29864 0.00 2995 2995 3116 0.00 20 00 00 00 00 00 00 00 00 00 00 00 0	2165					
2216	2216 2230 2256 2265 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326 2347 2350 2358 EN16711-1 53.6 -2.73 2366 2373 EN16711-1 67.2 -0.89 2373 EN16711-1 74 0.03 2380 EN16711-1 74 0.03 2381 EN16711-1 64.329 -1.28 2381 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2500 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2501 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 82.39 1.17 2742						
2230	2230 2256						
2256 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326	2256 2265 2261 2261 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326 2347 2350 2348 2356 2373 EN16711-1 53.6 -2.73 2366 2373 EN16711-1 67.2						
2265 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326 2327	2265 2310 EN16711-1 74 0.03 2311 EN16711-1 75.10 0.18 2326						
2311 EN16711-1 74 0.03 2314 EN16711-1 72.11 -0.22 2314 EN16711-1 75.10 0.18 2326	2310 EN16711-1 74 0.03 2311 EN16711-1 72.11 -0.22 2314 EN16711-1 75.10 0.18 2326						
2311 EN16711-1 72.11 -0.22 2314 EN16711-1 75.10 0.18 2326	2311 EN16711-1 72.11 -0.22 2314 EN16711-1 75.10 0.18 2326 2347 2350 2358 EN16711-1 53.6 -2.73 2366 2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.000 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 CPSC-CH-E1001-08.3 74.67 0.12 2621 74 0.03 2624 2637 2637 2678 2734 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829 2734 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829 2734 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829						
2314 EN16711-1 75.10 0.18 2326 2327 2337 2358 EN16711-1 53.6 -2.73 2368 2373 EN16711-1 67.2 -0.89 2373 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624	2314 EN16711-1 75.10 0.18 2326						
2326 2347 2350 2358 EN16711-1 53.6 2-73 2368 2373 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.6 EN16711-1 229.6 EN16711-1 229.6 EN16711-1 220.6 EN16711-1 EN167	2326 2347 2350 2358 EN16711-1 53.6 2373 EN16711-1 57.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98						
2347	2347 2350 2358 EN16711-1 53.6 -2.73 2366 2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 88.19 2734 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 RSD=12% RS						
2350	2350 2358 EN16711-1 53.6 2366 2373 EN16711-1 67.2 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 2816 297 2081 297 298 2081 297 298 2081 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 1.02		•				
2385 EN16711-1 53.6 -2.73 2366 EN16711-1 67.2 -0.89 2373 EN16711-1 74 0.03 2380 EN16711-1 66.70 -0.96 2381 EN16711-1 66.70 -0.96 2382	2358 EN16711-1 53.6 -2.73 2366						
2366 2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2383 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624	2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864						
2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382	2373 EN16711-1 67.2 -0.89 2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2737 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2829 2864						
2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2774 2775 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2774 2775 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2866 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2867	2375 EN16711-1 74 0.03 2380 EN16711-1 64.329 -1.28 2381 EN16711-1 66.70 -0.96 2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2628 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2777 CPSC-CH-E1001-08.3 82.39 2786 2797 CPSC-CH-E1001-08.3 81.9 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 81.9 1.96 2829 2864 2878 2879 2804 2816 EN16711-1 81.3 1.02 2829 3214 EN16711-1 81.3 1.02 295 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3214 EN16711-1 81.3 RSD=12% 3228 RSD=12% 3246 RSD=12% 3260 RSD=12% 3276 RSD=12% 3277 RSD=12% 3278 RSD=12% 3278 RSD=12%						
2381 EN16711-1 66.70 -0.96 2382 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624	2381 EN16711-1 66.70 -0.96 2382	2375 EN				0.03	
2382	2382 2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475	2380 EN	N16711-1	64.329		-1.28	
2385 In house 73.8 0.00 2459 EN16711-1 229.76 R(0.01) 21.15 2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591	2385 In house 73.8 0.00	2381 EN	N16711-1	66.70		-0.96	
2459 EN16711-1	2459 EN16711-1						
2475 2500 2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2734 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2877 CPSC-CH-E1001-08.3 72.98 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2475 2500 2503 CPSC-CH-E1001-08.3 85 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
2500	2500		N16711-1	229.76	R(0.01)		
2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637 2678 2734 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2877 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 1.02 3228	2503 CPSC-CH-E1001-08.3 85 1.52 2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591						
2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591	2511 EN16711-1 80.759 0.95 2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624 2637						
2590 CPSC-CH-E1001-08.3 74.67 0.12 2591 2622 74 0.03 2624	2590 CPSC-CH-E1001-08.3 74.67 0.12 2591						
2591	2591 2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 3228 normality NK n 22 outliers mean (n) st.dev. (n) R(calc.) RSD=12% RSD=12% RSD=12%						
2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2622 74 0.03 2624 2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829						
2624	2624 2637 2678 2734 2737 2737 2PSC-CH-E1001-08.3 28.39 2810 2PSC-CH-E1001-08.3 28.19 2829 2864 2977 2PSC-CH-E1001-08.3 2977 2PSC-CH-E1001-08.3 372.98 2977 3116 2995 3116 3172 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 31228 normality N N N N 22 Outliers mean (n) 73.766 st.dev (n) R(calc.) RSD=12% RSD=12% RSD=12%						
2637	2637 2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 3214 EN16711-1 66.96 3228 normality N N N N N 22 Outliers mean (n) st.dev (n) R(calc.) RSD=12% RSD=12% RSD=12%						
2678	2678 2734 2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 3228 normality N N N N N N N OK N 22 Outliers mean (n) st.dev (n) R(calc.) RSD=12% RSD=12% RSD=12%						
2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3218 normality OK n 22 outliers 1	2737 CPSC-CH-E1001-08.3 82.39 1.17 2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3218 normality OK n 22 outliers 1	2741 2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648	2734					
2810 CPSC-CH-E1001-08.3 88.19 1.96 2829 2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2810 CPSC-CH-E1001-08.3 88.19 1.96 2829	2737 CP	PSC-CH-E1001-08.3	82.39		1.17	
2829	2829						
2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2864 2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2977 CPSC-CH-E1001-08.3 72.98 -0.11 2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
2995 3116 3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	2995 3116 3172 EN16711-1 81.3 3209 3214 EN16711-1 66.96 -0.92 3228 normality N N N N N N N N N N N N N N N N N N N						
3116 3172 EN16711-1 81.3 3209 3214 EN16711-1 66.96 -0.92 3228 normality n 22 outliers 1	3116 3172 EN16711-1 3172 EN16711-1 3181.3 3102 3209 3214 EN16711-1 66.96 -0.92 3228 normality Nok Not						
3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 OK n	3172 EN16711-1 81.3 1.02 3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
3209 3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	3209 3214 EN16711-1 66.96 -0.92 3228						
3214 EN16711-1 66.96 -0.92 3228 normality OK n 22 outliers 1	3214 EN16711-1 66.96 -0.92 normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
3228 normality OK n 22 outliers 1	normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
normality OK n 22 outliers 1	normality OK n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
n 22 outliers 1	n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648	-					
n 22 outliers 1	n 22 outliers 1 mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648	nor					
	mean (n) 73.766 st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648	n	:	22			
mean (n) 73.766	st.dev. (n) 8.8030 RSD=12% R(calc.) 24.648						
	R(calc.) 24.648						
					RSD=12%		
	st.dev.(IEC62321-5:14) /.3766						
NIEU02321-3.141 20.034	R(IEC62321-5:14) 20.654	K(I	(IEU02321-5:14)	20.034			






Determination of Tin as Sn on sample #23630-2; results in mg/kg


10	lab	method	value	mark	z(targ)	remarks
623 In house 465.98						
840 In house 688.99 1.16 1126		In house	465.98		-2.23	
2115 EN16711-1 703.9 1.74 2121	840					
2121 2126 2159 In house 408.5 2162 2182 2184 2216 2218 2216 2217 2230 2256 2255 2255 227 2310 2311 2311 2311 2311 2311 2311 2311						
2146		EN16711-1	703.9		1.74	
2159 In house 408.5						
2165 2184 2184 2184 2230 2230 2256 2265 2265 2265 2265 2265 227 2310 2311 2311 2314 2317 2317 2311 2317 2317 2318 2318 2318 2318 2318 2319 2319 2319 2319 2319 2319 2319 2319		In house				
2182 2184 2216 2230		in nouse				
2184						
2216						
2230						
2256 2310 EN16711-1 702 1.71 2311 EN16711-1 579.11 -0.34 2314 EN16711-1 711.54 1.87 2326 2347 2328 2349 2350 2358 EN16711-1 271 C.DG(0.05) -5.48 first reported 150 2360 2373 2375 EN16711-1 670 1.17 2380 2381 2382 2385 In house 560 -0.66 2385 In house 560 -0.66 2459 2475 2500 2503 CPSC-CH-E1001-08.3 548 -0.86 2511 2520 2531 2521 2522 2532 2533 2534 2535 2536 2537 2537 2537 2537 2538 2538 2538 2538 2538 2538 2538 2538	2230					
2265 2310 EN16711-1 702 1.71 2311 EN16711-1 579.11 -0.34 2314 EN16711-1 711.54 1.87 2326 2347 2350 2368 2373 2388 EN16711-1 670						
2311 EN16711-1 579.11 -0.34 2314 EN16711-1 711.54 1.87 2326						
2314 EN16711-1 711.54 1.87 2326 2347 2350 2351 2352 EN16711-1 271 C.DG(0.05) -5.48 first reported 150 2360 2373 2375 EN16711-1 670 1.1.7 2380 2381 2382 2383 In house 560 -0.66 2459 2475 2475 2475 2500 2500 CPSC-CH-E1001-08.3 548 -0.86 2511 2590 2591 2622 39 DG(0.05) -9.35 2624 2637 2638 2734 2737 2678 2734 2737 2741 2737 2741 2741 2737 2741 2741 2737 2741 2741 2737 2741 2737 2741 2738 2739 2739 2734 2737 2741 2741 275 2864 2977 CPSC-CH-E1001-08.3 375.0 -3.75 2879 3172 EN16711-1 617 0.29 2899 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 RSD=21% 8t.dev. (lfc D62321-5:14) 59.9574	2310		702			
2326						
2347		EN16711-1				
2350						
2386 EN16711-1 271 C,DG(0.05) -5.48 first reported 150 2366						
2366 2375 EN16711-1 670 1.17 2380		FN16711-1		C DG(0.05)	-5 12	first reported 150
2373 EN16711-1 670 1.17 2380		LINIO/ I I-I		0,00(0.00)		ilist reported 130
2375 EN16711-1 670 1.17 2380						
2380		EN16711-1				
2382	2380					
2385 In house 560 -0.66 2459	2381					
2459 2475 2500 2503 CPSC-CH-E1001-08.3 548 -0.86 2511						
2475 2503 CPSC-CH-E1001-08.3 548 -0.86 2511		In house				
2500						
2503 CPSC-CH-E1001-08.3 548 -0.86 2511						
2511		CPSC-CH-F1001-08 3				
2590		3. 33 311 21001-00.0				
2591						
2622 39 DG(0.05) -9.35 2624						
2637 2678 2734 2737 2741 2810 2829 2977 CPSC-CH-E1001-08.3 375.0 3116 3172 EN16711-1 617 3209 3214 EN16711-1 784.44 3.08 3228 normality n	2622		39	DG(0.05)	-9.35	
2678 2734 2737 2741 2810 2829 2977 CPSC-CH-E1001-08.3 375.0 -3.75 2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality n 13 outliers p 2 mean (n) St.dev. (n) R(calc.) St.dev. (IEC62321-5:14) 59.9574						
2734						
2737 2741 2810 2829 2864 2977 CPSC-CH-E1001-08.3 375.0 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality n 13 outliers 2 mean (n) 599.574 st.dev. (n) R(calc.) St.dev. (lEC62321-5:14) 59.9574						
2741 2810 2829 2864 2977 CPSC-CH-E1001-08.3 375.0 -3.75 2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
2810 2829 2864 2977 CPSC-CH-E1001-08.3 375.0 -3.75 2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality n 13 outliers mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
2829						
2864 2977 CPSC-CH-E1001-08.3 375.0 -3.75 2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574	2829					
2977 CPSC-CH-E1001-08.3 375.0 -3.75 2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
2995 3116 3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality N N N N N N N N N N N N N N N N N N N	2977	CPSC-CH-E1001-08.3	375.0		-3.75	
3172 EN16711-1 617 0.29 3209 3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
3209 3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574		=1110=111				
3214 EN16711-1 784.44 3.08 3228 normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574		EN16711-1				
normality OK n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574		EN16711 1				
normality		□N 10 / 1 1 - 1				
n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574	3220					
n 13 outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574		normality	OK			
outliers 2 mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
mean (n) 599.574 st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574						
st.dev. (n) 125.0346 RSD=21% R(calc.) 350.097 st.dev.(IEC62321-5:14) 59.9574		mean (n)				
st.dev.(IEC62321-5:14) 59.9574		st.dev. (n)	125.0346	RSD=21%		
st.dev.(IEC62321-5:14) 59.9574						
		st.dev.(IEC62321-5:14)				
N(IEC02321-3.14) 107.001		R(IEC62321-5:14)	167.881			






Determination of Zinc as Zn on sample #23630-2; results in mg/kg

lab	method	value	mark	z(targ)	remarks
210					-
623		377.89			
840	In house	146.70			
1126					
2115	EN16711-1	94.07			
2121					
2146					
2159 2165					
2182					
2184					
2216					
2230					
2256					
2265					
2310		323			
2311		251.36			
2314	EN16711-1	303.47			
2326 2347					
2350					
2358	EN16711-1	30.3			
2366					
2373					
2375	EN16711-1	58			
2380					
2381					
2382	In Lance	40.0			
2385 2459	In house	48.0 826.52	G(0.01)		
2475	EN16711-1		G(0.01)		
2500					
2503	CPSC-CH-E1001-08.3	43.7			
2511					
2590					
2591					
2622		155			
2624					
2637 2678					
2734					
2737					
2741					
2810					
2829					
2864					
2977					
2995					
3116 3172					
3209					
3214	EN16711-1	169.01			
3228					
	normality	OK			
	n	12			
	outliers	1			
	mean (n)	166.7083	DCD-700/		
	st.dev. (n) R(calc.)	120.74050 338.0734	RSD=72%		
	st.dev.(IEC62321-5:14)	(16.67083)			
	R(IEC62321-5:14)	(46.6783)			
		(.0.07 00)			





APPENDIX 3

Other reported Metals in sample #23630-1; results in mg/kg

		sample #23630-1; r			
lab	Sb	As	Cr	Со	Mn
210					
	75.35	Not Detected	Not Detected	Not Detected	Not Detected
840					
1126					
2115					1.83
2121					
2146					
2159					
2165			Not detected		
2182					
2184			<10		
2216	52	None Detected	3	Not Determined	Not Determined
2230					
2256			Not detected		
2265	not analysed	not detected	not detected	not analysed	not analysed
2310	<10	not detected	not detected	not detected	not detected
2311	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2314	9.11				
2326					
2347					
2350	N.A.	N.A.	N.A.	N.A.	N.A.
2358	not detected	not detected	not detected	not detected	not determined
2366					
2373	not analyzed	not analyzed	not analyzed	not analyzed	not analyzed
2375	<10	<10	<10	<10	<10
2380					
2381					
2382	no capability	no capability	no capability	no capability	no capability
2385	<10	<10	<10	<10	128
2459	84.62		ND	ND	ND
2475					
2500			ND	ND	ND
2503	379	336			
2511					
2590	< L.O.Q.		< L.O.Q.	< L.O.Q.	
2591					
2622					
2624					
2637	21	23	150	22	1790
2678					
2734					
2737					
2741					
2810					
2829	not analyzed	not analyzed	not analyzed	not analyzed	not analyzed
2864					
2977	not detected	not detected	not detected	not detected	not detected
2995			not detected		
3116					
3172	< 10	< 10	< 10	< 10	
3209			<10.0	<10.0	<10.0
3214	<10	<10	<10	<10	<10
3228			<10		

Other reported Metals in sample #23630-1; results in mg/kg -continued

lab	Hg	Ni	Se	Sr	Zr
210					
	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
1126					
2115		1.85	3.24		
2121					
2146					
2159					
2165	Not detected				
2182					
2184	<10				
2216	None Detected	Not Determined	None Detected	Not Determined	Not Determined
2230					
2256	Not detected				
2265	not detected	not analysed	not analysed	not analysed	not analysed
2310	not detected	not detected	not detected	not detected	not detected
2311	Not Detected	Not Detected	Not Detected	Not Detected	Not Detected
2314					
2326					
2347					
2350	< 2	N.A.	N.A.	N.A.	N.A.
2358	not detected	not detected	not detected	not determined	not determined
2366					
	not analyzed	not analyzed	not analyzed	not analyzed	not analyzed
2375		<10	<10	<10	<10
2380					
2381					
	no capability	no capability	no capability	no capability	no capability
2385		<10	<10	<10	<10
2459		ND	ND	ND	
2475					
2500					
2503		41.7	508		
2511					
	< L.O.Q.	< L.O.Q.		< L.O.Q.	
2591		242			
2622 2624		212			
	not detected	45	not detected	not detected	not detected
2678		45 	not detected	not detected	not detected
2734					
2737					-
2741					
2810					
	not analyzed	not analyzed	not analyzed	not analyzed	not analyzed
2864					
2977		not detected	not detected	not detected	not detected
	not detected				
3116					
3172		< 10			
3209					
3214		<10	<10	<10	<10
3228					

Other reported Metals in sample #23630-2; results in mg/kg

lab	Sb	Cd	Pb	Hg	Se	Sr	Zr
210							
623		15.69	30.04	Not Detected	Not Detected	Not Detected	Not Detected
840	<10	<10	<10	<10	<10	<10	not analyzed
1126							
2115		1.98	1.59				
2121			0.8825				
2146		33.5	< 100 mg/kg				
2159		Not applicable					
2165		Not detected	Not detected	Not detected			
2182		not detected	not detected				
2184		<10	<10	<10			
2216	None Detected	20	2	None Detected	None Detected	Not Determined	Not Determined
2230			<10				
2256		Not detected	ND	Not detected			
2265	not analysed	not detected	not detected	not detected	not analysed	not analysed	not analysed
2310	not detected	not detected	38.2	not detected	not detected	not detected	not detected
2311	Not Detected	Not Detected	28.38	Not Detected	Not Detected	Not Detected	Not Detected
2314			48.73				
2326							
2347		<5	<20				
2350	N.A.	< 5	< 50	< 2	N.A.	N.A.	N.A.
2358	not detected	not detected	not detected	not detected	not detected	not determined	not determined
2366	<20	<5	<10	<5	<20		
2373	<20	<5	<20	<5	not applicable	not applicable	not applicable
2375	<10	<10	<10	<10	<10	<10	<10
2380							
2381							
2382	no capability	no capability	not detected	no capability	no capability	no capability	no capability
2385		<10	<20	<1	<10	<10	<10
2459	310.01	28.21	98.42			198.00	
2475							
2500		ND	ND	ND			
2503							
2511			<10				
	< L.O.Q.	2.10	< L.O.Q.	< L.O.Q.		< L.O.Q.	
2591		not detected	not detected				
2622		<5	<25				
2624							
2637							
2678 2734							
2734							
2737			<20				
2810		not detected	not detected				
2829	not analyzed	not detected	not detected	not analyzed	not analyzed	not analyzed	not analyzed
2864		41.26	not detected	not detected	not analyzed	not analyzed	
2977		not detected					
2995		not detected	not detected	not detected			
3116			<50				
	< 10	10.6	< 5	< 10			
3209		<5.0	<10.0	<5.0			
3214		<10	<10.0	<10	<10	<10	<10
3228		<10	<10	<10			
0220		. 3		. •			

APPENDIX 4 Analytical details

	Accredited ISO17025	Intake in grams	Digestion acid used	Concentration Acid	Quantify technique used
<u>210</u>	Yes	0.2	HNO3 Aquaregia for Sn	16	ICP-OES
	Yes	0.3	7.5mLHNO3:2.5mLHCl	14%	ICP-OES and ICP-MS
1126					
2115		0.1 g	HNO3		ICP-MS
	Yes	80mg	4.5mL HNO3 + 1.5mL HCl	HNO3 69.5% HCI 37%	ICP-MS
2146	No	#23630-1: whole sample (1,9 g) was digested #23630-2: 0,2 g	HNO3	67 % (wt/wt)	ICP-OES
2159		0.2 grams	Nitric acid	% 65 nitric acid	ICP-OES
2165	Yes	0.1g, nearest to 0.1mg	Aqua regia	8%(V/V) Aqua regia	ICP-OES
2182		0.1g	Nitric acid	67-69%	AAS
2184		0.1g	HNO3 and HCI	24%	ICP-MS
2216		#23630-1 = 0.9758 g #23630-2 = 0.98 g	Nitric and hydrochloric acids		ICP-MS
2230		0.2388 g	aqua regia	40%	ICP-OES
2256		23630-1:0.0726g 23630-2:0.0656g	conc. HNO3 & conc. HCl	Acid mixture (10% nitric acid + 4% hydrochloric acid)	ICP-OES
2265	Yes	0,1	HNO3 / HCI	3/1	ICP-OES
2310		0.25	Conc.Nitric acid	69%	ICP-MS
2311		0.1	Nitric Acid	69	ICP-OES
2314		0.1g	Nitric acid	69%	ICP-OES
2326		0.05	11000/110:	000/	
2347 2350		0.25 approximately 0.2g	Nitric acid & hydrochloric	20% Nitric acid 70% , hydrochloric acid 36%	ICP-OES
2358		0.25	acid HNO3	65	ICP-MS
2366					
2373		0.1g	HCL,HNO3	HCL:36%~38% HNO3:65%~68%	ICP-OES
2375		0.15 gram	HNO3 + HCI	9 mL + 3 mL	ICP-MS
2380	Yes	0.20 g	HNO3 & HCI	HNO3-65% & HCI-37%	ICP-OES
2381	Yes	0.25	Nitric acid	65%	ICP-OES
2382		1.5g	HNO3&HCL	HNO3: 69.0~72.0% HCL: 36.0~38.0%	AAS
2459		0.2	Aqua Regia	Conc. Acids (Nitric and Hydro Chloric Acid is used for digestion)	ICP-OES / AAS
2475		0.1	2.5 ml HNO3 and 7.5 ml HCl		ICP-MS
2500		0.1 gram	Nitric acid / Hydrochloric acid	Nitric acid 65-70% Hydrochloric acid 37%	ICP-OES
2503					
2511					
2590		0.2 gr	HNO3	4%	ICP-MS
2591 2622	Yes Yes	0,1 grams 0.5117 gr	Nitric Acid Nitric acid and hydrochloric acid		ICP-MS AAS
2624	Yes	0.1506	HNO3	~ 20%	ICP-OES
2637		0.1000	111400	2070	ICP-MS
2678					
2734					
2737	Yes	0.2g	HNO3:HCI=1:3	HNO3:65% HCI : 36%	ICP-OES
2741	Yes	0.1g	HNO3 65%.	3%	ICP-MS
2810		0,8577g sample 1 1,4266g sample 2	Nitric acid / Hydrochloric acid	37% Hydrochloric acid / 65% Nitric acid	ICP-OES
2829	No	0.09	Nitric and hydrochloric acid	HCl 37% and HNO3 65%	ICP-OES
2864	No	0.2g	Nitric acid	65%	ICP-OES
2977	Yes	0,1	HNO3 HCI	7	ICP-OES
2995	No	#22630-1: 0.9494 g #22630-2: 1.0259 g	HNO3 + HCI	HNO3 65% / HCI 37%	ICP-OES
3116		0.1 (for both CPSC and Canada methods)	Conc. nitric acid and conc. hydrochloric acid (for both CPSC & Canada methods)	8 mL HNO3 / 3 mL HCl (for CPSC method) 8 mL HNO3 / 2 mL HCl (for Canada method)	ICP-OES
3172	Yes				
3209			aqua regia (hydrochloric acid ; Nitric acid = 3:1 V/V)	20%	ICP-OES
3214	No	0.15 grams	9ml Hcl, 3ml HNO3 1 ml HF	26%	ICP-OES
3228	Yes	0.1-0.2g	concentrated HNO3 and HCI	10%HNO3+10%HCl in sample solution	ICP-OES

APPENDIX 5

Number of participants per country

- 2 labs in BANGLADESH
- 1 lab in FINLAND
- 2 labs in FRANCE
- 3 labs in GERMANY
- 4 labs in HONG KONG
- 3 labs in INDIA
- 1 lab in INDONESIA
- 9 labs in ITALY
- 1 lab in KOREA, Republic of
- 1 lab in MOROCCO
- 10 labs in P.R. of CHINA
- 2 labs in PAKISTAN
- 1 lab in SPAIN
- 3 labs in TAIWAN
- 1 lab in THE NETHERLANDS
- 3 labs in TUNISIA
- 2 labs in TURKEY
- 2 labs in U.S.A.
- 2 labs in VIETNAM

APPENDIX 6

Abbreviations

C = final test result after checking of first reported suspect test result

 $\begin{array}{ll} D(0.01) &= \text{outlier in Dixon's outlier test} \\ D(0.05) &= \text{straggler in Dixon's outlier test} \\ G(0.01) &= \text{outlier in Grubbs' outlier test} \\ G(0.05) &= \text{straggler in Grubbs' outlier test} \\ DG(0.01) &= \text{outlier in Double Grubbs' outlier test} \\ DG(0.05) &= \text{straggler in Double Grubbs' outlier test} \\ \end{array}$

R(0.01) = outlier in Rosner's outlier test R(0.05) = straggler in Rosner's outlier test

W = test result withdrawn on request of participant ex = test result excluded from statistical evaluation

n.a. = not applicable
n.e. = not evaluated
n.d. = not detected
fr. = first reported

Literature

- 1 iis Interlaboratory Studies, Protocol for the Organisation, Statistics & Evaluation, June 2018
- 2 ISO5725:86
- 3 ISO5725 parts 1-6:94
- 4 ISO13528:05
- 5 M. Thompson and R. Wood, J. AOAC Int, <u>76</u>, 926, (1993)
- 6 W.J. Youden and E.H. Steiner, Statistical Manual of the AOAC, (1975)
- 7 P.L. Davies, Fr. Z. Anal. Chem, 331, 513, (1988)
- 8 J.N. Miller, Analyst, <u>118</u>, 455, (1993)
- 9 Analytical Methods Committee, Technical Brief, No 4, January 2001
- 10 P.J. Lowthian and M. Thompson, The Royal Society of Chemistry, Analyst, <u>127</u>, 1359-1364, (2002)
- 11 W. Horwitz and R. Albert, J. AOAC Int, <u>79, 3,</u> 589-621, (1996)
- Bernard Rosner, Percentage Points for a Generalized ESD Many-Outlier Procedure, Technometrics, <u>25(2)</u>, 165-172, (1983)